



(11)

EP 2 328 480 B1

(12)

EUROPEAN PATENT SPECIFICATION

(45) Date of publication and mention
of the grant of the patent:
06.01.2016 Bulletin 2016/01

(51) Int Cl.:
A61B 8/08 (2006.01) **A61B 5/00** (2006.01)
A61B 8/12 (2006.01)

(21) Application number: **09798815.8**

(86) International application number:
PCT/US2009/051017

(22) Date of filing: **17.07.2009**

(87) International publication number:
WO 2010/009412 (21.01.2010 Gazette 2010/03)

(54) LOW-COST DEVICE FOR C-SCAN PHOTOACOUSTIC IMAGING

KOSTENGÜNSTIGES GERÄT FÜR FOTOAKUSTISCHE C-SCAN-BILDGEBUNG

DISPOSITIF À BAS COÛT POUR IMAGERIE PHOTO-ACOUSTIQUE À BALAYAGE C

(84) Designated Contracting States:
**AT BE BG CH CY CZ DE DK EE ES FI FR GB GR
HR HU IE IS IT LI LT LU LV MC MK MT NL NO PL
PT RO SE SI SK SM TR**

(30) Priority: **18.07.2008 US 81946 P**

(43) Date of publication of application:
08.06.2011 Bulletin 2011/23

(73) Proprietor: **University of Rochester
Rochester, NY 14642 (US)**

(72) Inventors:
• **DOGRA, Vikram, S.
Pittsford Drive, NY 14534 (US)**
• **RAO, Navalgund, A., H., K.
Rochester, NY 14618 (US)**

(74) Representative: **Gillard, Matthew Paul
Withers & Rogers LLP
4 More London Riverside
London
SE1 2AU (GB)**

(56) References cited:
**US-A- 6 102 857 US-A1- 2005 187 471
US-A1- 2007 299 341 US-A1- 2008 123 083**

- **Yadong Wei ET AL: "Photoacoustic tomography imaging using a 4f acoustic lens and peak-hold technology", Opt. Express, 2 April 2008 (2008-04-02), pages 89-95, XP055090658, Retrieved from the Internet: URL: <http://www.opticsinfobase.org/oe/abstr.act.cfm?URI=oe-16-8-5314> [retrieved on 2013-11-28]**
- **YONGHENG HE ET AL: "A novel photoacoustic tomography based on a time-resolved technique and an acoustic lens imaging system", PHYSICS IN MEDICINE AND BIOLOGY, INSTITUTE OF PHYSICS PUBLISHING, BRISTOL GB, vol. 51, no. 10, 21 May 2006 (2006-05-21), pages 2671-2680, XP020095765, ISSN: 0031-9155, DOI: 10.1088/0031-9155/51/10/019**

Note: Within nine months of the publication of the mention of the grant of the European patent in the European Patent Bulletin, any person may give notice to the European Patent Office of opposition to that patent, in accordance with the Implementing Regulations. Notice of opposition shall not be deemed to have been filed until the opposition fee has been paid. (Art. 99(1) European Patent Convention).

Description

Field of the Invention

[0001] The present invention is directed to imaging, e.g., of prostate tissue for detection of cancer and more particularly to photoacoustic imaging and ultrasound imaging to produce C-scans.

Description of Related Art

[0002] Prostate cancer is the most prevalent newly diagnosed malignancy in men, second only to lung cancer in causing cancer-related deaths. Adenocarcinoma of the prostate is the most common malignancy in the Western world. There were a projected 218,890 new cases of prostate cancer diagnosed in the United States in 2007, with an estimated 27,050 deaths. As men age, the risk of developing prostate cancer increases. Prostate cancer has been found incidentally in approximately 30% of autopsy specimens of men in their sixth decade. Seventy to 80% of patients who have prostate cancer are older than 65 years. Clinically localized disease is usually suspected based on an elevated prostate specific antigen (PSA) test or abnormal digital rectal exam (DRE), prompting transrectal ultrasound (TRUS) guided biopsy of the prostate for definitive diagnosis. TRUS however, is not reliable enough to be used solely as a template for biopsy. There are cancers that are not visible (isoechoic) on TRUS. Furthermore, in PSA screened populations, the accuracy of TRUS was only 52% due to false-positive findings encountered. Increased tumor vessels (angiogenesis) have been shown microscopically in prostate cancer compared with benign prostate tissue. Efficacy of color and power Doppler ultrasound has not been demonstrated, probably due to limited resolution and small flow velocities. Elasticity imaging, with its many variants, is a new modality that is currently under extensive investigation. It is evident that given the limitations of the present diagnostic protocols, development of a new imaging modality that improves visualization and biopsy yield of prostate cancer would be beneficial. Furthermore, by making it cost effective, we can place it in the hands of primary care physicians, where it will serve its primary purpose as an adjunct to PSA, DRE, and TRUS.

[0003] The need for tumor visualization is equally critical in the treatment of localized prostate cancer disease. Existing therapeutic strategies, namely external beam radiation, prostate brachytherapy, cryosurgery, and watchful waiting, all will benefit significantly from the development of a new modality that promises better tumor contrast. Thus, prostate cancer continues to be an area in which progress is needed despite recent advancements.

[0004] Appropriate imaging of prostate cancer is a crucial component for diagnosing prostate cancer and its staging, in addition to PSA levels and DRE. The current state of prostate imaging for diagnosis of prostate cancer includes ultrasound, ultrasound-guided prostate biop-

sies, magnetic resonance imaging (MRI), and nuclear scintigraphy. These modalities are helpful, but have drawbacks and limitations. MRI is expensive and not mobile. Nuclear scintillation is expensive, provides low resolution planar images, and there are problems with radiotracer excretion through the kidneys. Both these modalities are not available for general use.

[0005] Ultrasound is not reliable enough to use solely as a template for diagnosing prostate cancer. It has two problems. First, in many cases prostate cancer appears as an isoechoic lesion (similar gray scale value as surrounding tissue) causing high miss rate (Figure 10). Secondly, when it is visible (hyper or hypoechoic), it is not possible to say with certainty if it is cancer or benign because many other noncancer conditions such as prostate atrophy, inflammation of the prostate gland, and benign tumors may also look similar in appearance on ultrasound examination. A biopsy has to be performed on the suspect lesion for definitive diagnosis. Biopsies are uncomfortable and bleeding may result as a complication. Because of poor lesion detection, even the current prostate biopsy techniques miss approximately 30% of prostate cancer. Utility of color flow and power Doppler in conjunction with gray scale ultrasound has been explored, but not successfully. Therefore, there is an urgent need for a new imaging methodology that will be portable, economical to build, and will have widespread utility as a tool for primary screening and diagnosis of prostate cancer.

[0006] Figure 10 shows transrectal sonography of the prostate gland with confirmed prostate cancer. Figure 10 demonstrates the grayscale appearance of the prostate gland (within arrows) in a patient with confirmed multiple zones of prostate cancer. As shown, the whole prostate gland appears heterogeneous and the prostate cancer zone is not identifiable because prostate cancer is similar in appearance to the surrounding prostate gland tissue. The arrowheads correspond to calcifications that can be seen in the prostate gland.

[0007] U.S. Patent Application Publication No. US 2009/0024038 A1 to Arnold teaches an acoustic imaging probe incorporating photoacoustic excitation, but does not overcome all of the above-noted deficiencies of the prior art.

[0008] US2005/018747 A1 discloses a non-invasive imaging system that applies ultrasound and stimulation light to a subject and uses a two dimensional array of transducer elements to pick up photo acoustic signals and reflected ultrasound that arise in response.

[0009] "Photoacoustic tomography imaging using a 4f acoustic lens and peak-hold technology by Yadong Wei et al is an article in Optics Express (14 April 2008, Vol. 16, No. 8, pages 5314-19) that describes a system in which a photoacoustic response is imaged using an acoustic lens and a one dimensional set (i.e. a line) of transducer elements.

Summary of the Invention

[0010] The invention is defined by the appended claims, to which reference should now be made.

[0011] It is therefore an aim for certain embodiments of the invention to provide photoacoustic (PA) imaging technology for the betterment of prostate cancer diagnosis and disease management.

[0012] It is a further aim for certain embodiments the invention to provide a low-cost imaging technology that can take C-scan images of the prostate gland in real time based on the photoacoustic phenomenon.

[0013] It is a yet further aim for certain embodiments of the invention to provide a low-cost imaging technology that can take C-scan images based on both the photoacoustic phenomenon and ultrasound so that the resulting images are automatically coregistered.

[0014] It is a still further aim for embodiments of the invention to provide such a low-cost imaging technology whose applicability is not limited to prostate cancer, but instead can be extended to other organs for cancer detection such as breast cancer, uterine cancer, ovarian cancer, thyroid cancer, etc.

[0015] To enable embodiments of the invention to achieve the above and other aims, the present invention is directed to an imaging technology combining C-scans with the photoacoustic effect. C-scans are two-dimensional (2D) images produced by digitizing the point-by-point signal variation of an interrogating sensor while it is scanned over the surface. The x-y position of the sensor is recorded simultaneously with signal variations. A computer converts point-by-point data into an appropriate image.

[0016] The photoacoustic effect is a phenomenon where acoustic waves are produced by absorbing points or objects of a medium that is irradiated with low-energy nanosecond (ns) pulses of laser light usually in near infrared region (NIR). The absorption of a short optical pulse causes localized heating and rapid thermal expansion, which generates thermoelastic stress waves (ultrasound waves). The ultrasound waves are generated instantaneously and simultaneously everywhere in a three-dimensional (3D) tissue volume irradiated by the laser pulse. The frequency content of the ultrasound waves generated is usually 1 to 20 MHz. It propagates in tissue as spherical waves in all directions. These waves will be referred to as the "PA signal" herein.

[0017] In certain embodiments of the invention, the prostate gland will be stimulated with laser light, resulting in ultrasound waves (photoacoustic effect) that will be focused by an acoustic lens and captured by a specific 2D sensor array and subsequently displayed as a C-scan on a computer screen. The amplitude of the ultrasound waves generated by laser stimulation is proportional to the optical absorption of the tissue element at that spatial location. Variability in tissue absorption results in C-scan image contrast. For example, optical absorption in blood is 10-100 times higher than prostate tissue in the near

infrared region. On this basis it is believed that the ultrasound signature in the C-scan image from prostate cancer will be different than the rest of the prostate gland. It is non echo-based imaging because there is no transmitted ultrasound pulse and contrast in the image does not depend on the echogenicity. Therefore, it is more advantageous than conventional ultrasound because it is based on a different contrast mechanism. It may be considered a hybrid imaging technique, combining ultrasound for imaging and photoacoustic effect for contrast.

[0018] The key task in PA imaging is to determine the laser absorption distribution from the measured PA signal data. Most of the existing methodologies are slow (because the sensor acquires data point-by-point), computationally intensive (because computer based algorithms have to be used to reconstruct the image), suffer from poor SNR, and are expensive to develop and market. Lens focusing of sound energy generated with photoacoustic stimulation will improve the SNR and will project reconstructed coronal plane images (C-scan) of the prostate gland simultaneously.

[0019] The imaging system can be lightweight, small in size, and therefore portable. The entire lens and sensor assembly can fit in a capsule no larger than 1.5x1x1 inches that can be used transrectally. A low power laser source and a computer monitor is all that will be required externally.

[0020] The electrical power consumption of the invention, at least in an embodiment called "Low Cost Device for C-Scan Photoacoustic Imaging of Prostate Gland Cancer" (CSPIP), will be very low because, unlike other existing methodologies under development, lens focusing does not require any electrical power. CSPIP has the potential for near real-time imaging (30 frames per second) unlike other methodologies, because they require some computational time for image reconstruction before it can be displayed. In CSPIP the lens forms a focused image that will be read out with the specially designed sensor at video rates (20-30 frames per second). CSPIP also offers the potential for full 3 dimensional imaging of the prostate through development of zoom acoustical lenses that will enable full resolution z-axis image cataloging and reconstruction as a solid. This will enhance the probability of early detection by enabling an arbitrary number of viewing angles of the prostate structure on the computer screen. Additionally this will also help in prostate cancer therapy planning and management (Radiotherapy and Brachytherapy-radioactive seed implantation).

[0021] The invention has great potential to exploit high contrast in laser NIR absorption between blood and prostate tissue. Angiogenesis (increased tumor blood vessels) of prostate cancer is well established. Therefore, it is believed that we will be able to visualize the cancer and its size better with CSPIP than with transrectal ultrasound (TRUS), because the latter is based only on the echogenicity of cancer tissue.

[0022] CSPIP may also prove to be unique and better

than Doppler ultrasound in detecting angiogenesis, because Doppler flow has poor detection sensitivity for slow flow velocities. In CSPIP, contrast in the image depends on the blood concentration in cancer tissue and not on the blood flow velocity; therefore, it is expected to have better contrast and improved delineation of prostate cancer.

[0023] In summary, CSPIP for prostate cancer detection we have proposed overcomes many shortcomings of the existing methodologies for imaging prostate cancer. It is a low cost, portable, non-echo based device that works on laser NIR stimulation followed by lens-based focusing of the PA signal. In CSPIP images are generated secondary to photoacoustic stimulation of the prostate gland. This system will have better contrast resolution than conventional ultrasound for detection of prostate cancer.

[0024] The present invention differs from that of the *Arnold* reference at least in terms of time gating. The present invention uses time gating to select acoustic signals coming from a particular depth and thereby to provide a C-scan. Since *Arnold* does not use time gating, it cannot produce a C-scan.

Brief Description of the Drawing

[0025] Preferred embodiments of the present invention will be set forth in detail with respect to the drawings, in which:

Fig. 1 is a concept diagram to illustrate the concept of C-scan;

Fig. 2 is a diagram of a stage-1 experiment;

Fig. 3 is a diagram of a stage-2 experiment;

Fig. 4 shows a set of C-scan photoacoustic images of a pencil exposed by a 2 mm laser beam;

Fig. 5A shows a lens-based C-scan photoacoustic imaging breadboard proof of concept;

Fig. 5B shows C-scan planar slices of a photoacoustic signal at different time gate centers;

Fig. 5C shows a C-scan image of the lead pencil embedded in chicken breast tissue;

Fig. 6 is a diagram of a probe according to the first preferred embodiment;

Figs 6A and 6B are diagrams of a variation of the probe of Fig. 6

Fig. 7 is a diagram of a system according to the second preferred embodiment;

Fig. 8 is a diagram of a system according to the third preferred embodiment;

Figs. 8A-8D are diagrams of a system according to a first variation of the third preferred embodiment;

Fig. 8E is a diagram of a system according to a second variation of the third preferred embodiment;

Fig. 9 is a diagram of a system according to the fourth preferred embodiment; and

Fig. 10 is a traditional transrectal sonogram of a prostate with confirmed prostate cancer.

Detailed Description of the Preferred Embodiment

[0026] Preferred embodiments of the present invention will be set forth in detail with reference to the drawings, in which like reference numerals refer to like elements or steps throughout.

[0027] Figure 1 shows C-scan with an ultrasound sensor 102, which is scanned in a scan plane P_S to perform imaging in an imaged plane P_I to produce a signal S from which an image I is reconstructed. From the travel time of the sound waves one can determine the depth z of the reflector or scatterer R , because the speed of sound c is known. Therefore, if we record the sensor signal amplitude at a fixed time $t_0 = 2z_0/c$ and scan it point-by-point in the x-y plane as shown, we essentially image the imaged or object plane P_I at depth z_0 . This is referred to as the C-scan planar image at depth z_0 .

[0028] The PA signal at each pixel is proportional to the incident laser flux and the absorption coefficient of the tissue. Because the absorption coefficient of tissue depends on the laser wavelength, experiments will be performed to determine the best wavelength for optimal contrast in the PA image. Laser wavelengths to be used will be in the near infrared region.

[0029] Figure 2 shows the schematics for the experimental setup. Freshly excised prostate tissue slices of 2-4 mm will be immersed in a saline solution bath. A 10 ns laser pulse at a fixed wavelength λ will be delivered to a small spot (approximate diameter 100 microns) via a fiber optic cable 204.

[0030] According to the American National Standards Institute (ANSI), maximum permissible exposure for human skin is 100 mJ/cm². Laser flux will be kept below this limit at all times. A 20 MHz, 6 mm diameter circular disk transducer with a focal length of 12 mm will be used as a sensor for the PA signal. The transducer focal spot is expected to be 100 microns and will overlap with the laser spot. This overlap region is expected to set the spatial resolution for these experiments. The laser-transducer rigid assembly will be scanned in x-y and z directions with stepper motors under computer control. The step size may vary from 10-50 microns. At each pixel location, the time-gated high frequency PA ultrasound signal will be digitized at 100 MHz and stored for further processing. The time gating can be performed in, or under the control of, a processor by any suitable technique. PA images will be created using several different data extraction schemes applied to the signal captured at each pixel location; they include (i) peak signal value in the time gate (ii) total energy in the time gate, and (iii) Amplitudes at several different temporal frequencies in the PA signal derived via its Fourier transform. These stage 1 experiments will enable determination of the optimal laser wavelength to be used for PA imaging.

[0031] The objective of stage 2 is to finalize the preliminary acoustic lens and hand-held device design generated by and to fabricate and test its ability to produce C-scan planar PA images of the prostate gland in-vitro.

[0032] Figure 3 is a concept diagram that explains the innovative approach to PA imaging and the setup for the stage 2 experiments. Consider a nanosecond laser beam that irradiates the object at time $t=0$. Let there be two point absorbers in plane A and B, respectively. The PA transient wave fronts are shown diverging from the points. The idea is to focus these PA transient wavefronts with an acoustic lens system 306, much like optical focusing of light waves for optical image formation. An appropriately designed lens will converge the wavefronts from the point in plane A to the point in plane A* and the point in plane B to point in plane B* respectively.

[0033] A sensor located in plane A* can capture this energy as a focused image. It takes approximately 6.5 microseconds for ultrasound to travel 1 cm in tissue. It has been shown that the PA transient time signal is typically 1 microsecond or less in duration. Therefore, by time gating the total signal we can discard or minimize the signal coming from planes that are separated more than 1.5 mm from the chosen plane of focus. The assumption here is that the signal from all points in plane A comes to a focus in plane A* and its arrival time will be within the selected time gate width. The signal from all other planes will be discriminated primarily due to the time gate and secondarily due to the spatial spreading of its energy on this sensor plane that does not happen to be their corresponding conjugate image plane.

[0034] Tests of the concept will be described. Fig. 5A shows the setup. The image plane was scanned with a 1.5 mm diameter hydrophone. C-scan image plane was 10x25 pixels. The object was 0.2 mm lead pencil illuminated by a laser beam of 2 mm diameter. The laser generated pulses of wavelength 1064 nm with 10 ns pulse width. The point object was located at an object distance (OD) from the lens. The lens had a focal length of 27.5 mm. Hydrophone was located in the image plane at an image distance (ID) from the lens. An acoustic reflector was used to simulate first embodiment of prostate probe.

[0035] Fig. 5B shows a single PA signal and several time gates chosen for C-Scan. C-Scan planar slices at different time gate centers are also depicted in the figure. Fig. 4 shows the C-Scan image of the point source with fixed OD=55 mm and different ID. This demonstrates that the sharp image is located at an ID of 55 mm corresponding to an OD of 55mm for a lens of focal length 27.5 mm. Fig. 5C depicts a C-Scan image of the lead pencil embedded in a chicken breast tissue.

[0036] Further lens design and fabrication will be performed. *In vitro* testing of the fabricated lens will be performed using phantoms and prostate gland tissue. Sensing of the PA signal will be done using a 0.1 mm diameter hydrophone. The tip of the hydrophone will be scanned in the image plane point-by-point with a pixel resolution of 100 microns. Lens aberrations, depth of field, in plane spatial resolution and resolution between C-scan planes, including signal to noise ratio are the quality metrics that will be tested. Based on preliminary study it is expected that the in plane and between the C-scan planes resolu-

tion of the proposed imaging system will be 250 microns or better.

[0037] An important subcomponent of the CSPIP imaging system is the 2D ultrasound sensor array with a 100 x 100 micron pixel size and a pitch of 100-200 microns. For near real-time processing, there is a need for on-pixel integrated circuit design. No off-the-shelf technology exists at the present time. Currently, complementary metal oxide semiconductor (CMOS) technology appears to be the best platform to work with because it can perform on-pixel processing with action specific integrated circuits (ASIC), is inexpensive to mass produce, consumes less power, and is undergoing rapid developments. This project will look at innovative approaches to chip design where the signal sensing is incorporated into the CMOS process in a way that meets the demands set by the previous stages. Piezoelectric materials such as PZT or PVDF will be laid on top of the CMOS as sensing layer. Sensing of the PA signal at each pixel will take place via the piezoelectric effect that generates charge proportional to the PA signal. CMOS technology with active pixel processing can integrate the charge within a selectable time gate, convert into voltage, and amplify it. The voltage can then be read out pixel by pixel at video rates with existing camera technology. The development work is expected to be centered on the piezoelectric sensor design and its integration into the CMOS technology.

[0038] A system according to the invention can integrate a laser source, acoustic lens, and sensor components into a housing assembly that can be used for *in vitro* and *in vivo* testing. The laser stimulation module, lens system, and the ultrasound sensor will be housed in a capsule suitable for transrectal insertion to image prostate gland cancer. The capsule will be approximately 1.5 x 1 x 1 inches in size. Approximate weight of the capsule when fully assembled will be 12-14 ounces.

[0039] Phantoms will be constructed with point and line targets made from high absorption coefficient material such as India ink or anticoagulated animal blood embedded in a low absorption background material. The objective is to map out the point-spread function of the CSPIP system at various points in the 3D space in front of the lens. A secondary objective is to determine the system signal sensitivity and signal-to-noise ratio (SNR). Linearity of the sensor and image contrast will be evaluated. Finally, 3D imaging of the excised prostate gland will be performed before the pathology study to determine the efficacy of PA imaging system to diagnose prostate cancer. An animal model for prostate cancer will be used for testing. Dogs will be injected with canine transmissible venereal sarcoma cell line that produces neoplasm in 15-40 days. TRUS and our prototype PA imaging device CSPIP will be used transrectally for imaging, comparison, and evaluation.

[0040] A first preferred embodiment of a probe will now be disclosed. The first preferred embodiment, shown in Fig. 6 as 600, is a transrectal prostate probe. Components used for probe construction include: acoustic lens

306, acoustic reflector 608, acoustic window 610, acoustic propagation medium 612, fiber optic cable 614, collimator 616, beam expander 618, optical mirror 620, optical window 622, 2d sensor 102, air 626, probe body 628, and plane wave generator 630. The fiber optic cable is in communication with a laser source (not shown), which can be a photodiode or any other suitable source.

[0041] A simple double-concave lens was designed and fabricated to be used as an acoustic lens. The material was chosen as Acrylic whose density is 1.19 g/cm^3 . The velocity of sound in acrylic medium is $2.75 \text{ mm}/\mu\text{s}$ and the acoustic impedance is about 3.26 MRayl . Other excellent candidate materials for acoustic lenses include RTV Silicone rubber (density: 1.49 g/cm^3 , velocity: $0.92 \text{ mm}/\mu\text{s}$ and acoustic impedance: 1.37 MRayl) and agarose hydrogels (velocity of sound: $1.32 \text{ mm}/\mu\text{s}$, acoustic impedance: 1.57 MRayl). The biconcave lens will converge the wavefront from the object plane (tissue) onto the image plane (sensor) with a magnification dictated by the object plane-lens and lens-image plane distances. The lens has a focal length of 27.5 mm and a diameter of 25 mm with a corresponding f number of 1.1 . It was designed to produce best images when the object plane-lens distance and the corresponding lens-image plane distance were 55 mm with a resulting magnification of 1 .

[0042] Stainless steel was used as the acoustic reflector due to its excellent acoustic properties. It has an acoustic impedance of 46 MRayl which is very high compared to that of water. The velocity of sound through stainless steel medium is $5.79 \text{ mm}/\mu\text{s}$. Stainless steel has a density of 7.89 g/cm^3 . The thickness of the acoustic reflector is 0.254 mm .

[0043] Low density polyethylene material was tested and used as acoustic window material. It has an acoustic impedance of 2.12 MRayl and a density of 0.92 g/cm^3 . Sound travels with a velocity of $1.85 \text{ mm}/\mu\text{s}$ through this material. The window has a thickness of 0.793 mm and accounts for transmission of about 82.8% of the acoustic signals. Other materials that can be used as acoustic window are Hydrogels and TPX (Polymethylpentene). Hydrogels - Agarose gel samples have an acoustic impedance of 1.57 MRayl , very close to that of water. The longitudinal velocity of sound in these materials is about $1.32 \text{ mm}/\mu\text{s}$. Polymethylpentene, commonly called TPX is another excellent material with a transparency of 91% in the visible spectrum and a density of 0.83 g/cm^3 . The speed of sound in TPX is $2.22 \text{ mm}/\mu\text{s}$.

[0044] Water was used as the ultrasonic wave propagation medium. Attenuation of ultrasound in water is 1.65 Np/m and the velocity of sound is $1.497 \text{ mm}/\mu\text{s}$ which is close to that of a biological tissue. Water has an acoustic impedance of 1.494 MRayl and a density of 0.998 gm/cm^3 . Mineral oil can also be used as an alternative for water for this purpose. Water was chosen over mineral oil due to its low attenuation coefficient compared to that of mineral oil which is 2.15 Np/m . The velocity of sound in mineral oil is $1.44 \text{ mm}/\mu\text{s}$. Mineral oil has an acoustic

impedance of 1.19 M Rayl and a density of 0.825 g/cm^3 . Valves 632 for water inlet/outlet were provided in the design that allows for regular maintenance. Any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl can be used as appropriate.

[0045] This lens technology can also be utilized for focusing reflected ultrasound signal similar to conventional ultrasound imaging. Hence, an input ultrasound source (plane wave generator) was added for ultrasound pulse echo C-scan in addition to PA imaging thus making the device, a dual imaging modality.

[0046] The laser source will be a tunable broadband laser with frequencies ranging from $500\text{-}1100 \text{ nm}$ with a typical pulse width of 10 ns and $1\text{-}10 \text{ Hz}$ pulse repetition frequency. The laser pulses will be delivered via a high quality fiber optic cable. The laser beam waist at the output of the fiber-optic cable-collimator assembly is typically 2 mm with intensity below $100 \text{ millijoules/cm}^2$. Beam divergence will be controlled by an optical lens. This beam will be angled by a laser mirror such that the laser beam comes out from an optical window at grazing angle to the outer surface of the probe.

[0047] The probe will have a diameter between 0.75 and 1.25 inches, and a length approximately 5 inches. The acoustic window will be aligned in front of the prostate when inserted into the rectum. With the help of acoustic coupling gel, the PA ultrasound waves generated in the prostate will enter through the acoustic window and guided towards the acoustic reflector. This reflector plate is made from stainless steel which has very high acoustic impedance material compared to water. The acoustic and optical chambers are separate and sealed. The acoustic chamber will contain any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl , for example, water, as an acoustic propagation medium. This device will provide the C-scan images corresponding to a tissue thickness less than 1 mm , usually in the range of $700\text{-}900$ microns over the entire length and width of the prostate by means of the 2d sensor.

[0048] A FC-connectorized high power delivery fiber optic cable is used to deliver the pulsed high power laser beam from an external laser source. It features a pure silica core and a bonded hard polymer cladding. The fiber has a numerical aperture of 0.37 which allows for greater light coupling efficiencies. The fiber's core diameter is $600 \mu\text{m}$ and cladding diameter is $630 \mu\text{m}$. The maximum power density that the fiber can withstand is 100 KW/cm^2 . This fiber is suitable for use in the 500 nm to 1100 nm wavelength range.

[0049] An FC-connectorized fiber optic collimator is used to collimate the diverging output laser beam from the fiber. The numerical aperture of the collimator is 0.25 , with a full beam divergence of 0.048° and $1/e^2$ output beam diameter of 2 mm . It has a focal length of 11 mm .

[0050] Cylindrical and spherical plano-concave lenses are used as beam expanders depending on the application to diverge the collimated beam. Cylindrical lenses

provide four times higher laser exposure on the tissue compared to spherical lenses. The lenses have a focal length of -15 mm which facilitates in illuminating the tissue volume over a width of about 1 cm. The size of the cylindrical lens was 10 mm x 12 mm with a clear aperture greater than 90% of surface dimensions. The spherical lens has a diameter of 12.7 mm with a clear aperture greater than 90% of its diameter.

[0051] A laser mirror of diameter 12.7 mm and thickness 6 mm is used to reflect the diverging beam resulting from the optical lens onto the tissue. It has a high damage threshold of 5 J/cm² for a 10 ns pulse. Reflectivity of the mirror is greater than 99.7% with a clear aperture greater than 80% of diameter.

[0052] A BK7 1/4 wave laser line window of dimensions 35 mm x 15 mm and thickness 3 mm was used as an optical window to allow light propagation onto the tissue.

[0053] A two-dimensional PZT sensor array will be designed and fabricated with on-pixel processing capabilities that can provide C-scan images in real time. The frequency bandwidth of the sensor will be 2-15 MHz. The sensor will provide C-scan images with slice thickness in the range of 700 to 900 microns. Behind the sensor, action specific integrated circuits will be employed to read the voltage at each pixel, amplify, convert it from analog to digital format and read out on the display.

[0054] The laser light beam is allowed to propagate through air before passing out of the optical window. It also ensures that the sensor's backend electronics are insulated from water (or whatever acoustic propagation medium is used).

[0055] The probe body is made of a material called watershed XC11122. Its properties mimic traditional engineering plastics including ABS and PBT. The size of the probe depends mainly on the application. In case of the prostate, it is designed as side-fire probe, and in case of applications like breast, skin, thyroid or transvaginal imaging, it will be designed as end-fire probe or its variables.

[0056] A high bandwidth ultrasonic transducer of 5 MHz nominal center frequency is used as a plane wave generator whose electrical impedance will be well matched to conventional ultrasound pulsar/receivers. The transducer will have a low Q-factor (air-backed into PMMA) and conforms perfectly to the probe's cylindrical surface. Typical thickness of this plane wave generator would be 110 microns (dual layer) with a low electrical impedance of 30 to 100 ohms. It will be bonded with epoxy resin or a transfer adhesive to acoustic window material and then be used.

[0057] A variation of the first preferred embodiment will be described with reference to Figs. 6A and 6B. The variation, shown in Figs. 6A and 6B as 600', uses ultrasound only rather than both ultrasound and the photoacoustic effect and is constructed like the previously described probe 600, except for the omission of elements relating solely to the photoacoustic effect and for any differences described herein. The device 600' will facilitate the de-

finite detection of prostate cancer and generates C-scan multiplanar images. An input ultrasound source (plane wave generator) was added for ultrasound pulse echo C-scan imaging while focusing of each plane happens in real time with the help of acoustic lens.

[0058] The probe will have a diameter between 0.75 and 1 inches, and a length approximately 5 inches. The acoustic window will be aligned in front of the prostate when inserted into the rectum. With the help of acoustic coupling gel, the ultrasound waves reflected from the prostate will enter through the acoustic window and guided towards the acoustic reflector. This reflector plate is made from stainless steel which has very high acoustic impedance material compared to water. Water is used as acoustic propagation medium because of its low ultrasound attenuation coefficient compared to mineral oil. Any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl can be used. This device will provide the C-scan multiplanar images of the prostate gland.

[0059] Another preferred embodiment, shown in Fig. 7 as 700, can be used to evaluate prostate cancer (or cancer of any other organ) *in vitro*. It is constructed like the first preferred embodiment, except as described below.

[0060] When pathologists receive a freshly excised sample with cancer, they are uncertain of the exact location of cancer. In order to determine its location, the specimen will be sliced and visually inspected by naked eye. If any abnormalities are present, the specimens will be examined through histopathology procedure. This device will facilitate the detection of cancer in the tissue for e.g., prostate gland which has been removed by surgery thus aiding the pathologist in locating the cancer. It provides the depth and location of the cancer by calculating all the foci (since the prostate is multifocal) within the prostate gland. This procedure saves a lot of time and will help in appropriate staging of prostate cancer. This device can be used not only for prostate but also for detecting cancer in any other organ of the body such as ovary, testis, uterus, kidney etc. that has been submitted to the pathologist for evaluation. This device will provide the C-scan images corresponding to a tissue thickness less than 1 mm, usually in the range of 700-900 microns over the entire length and width of the prostate.

[0061] The specimen Sp is usually setup in a water bath 734 to ensure maximum acoustic matching between the tissue and the medium. Any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl can be used instead. The acoustic window will be aligned in front of the prostate during the examination. With the help of acoustic coupling gel 733 and acoustic windows 735, the PA ultrasound waves generated in the prostate will enter through the acoustic window and focused onto the sensor by means of the acoustic lens. Water (or other acoustic propagation medium, as noted) is filled inside the probe which acts as the acoustic propagation medium. Scanning optics 736 allow scanning of

the light over the sample.

[0062] A third preferred embodiment, shown in Fig. 8 as 800, can be used for breast imaging. It is constructed like the first and second preferred embodiments, except as described below.

[0063] Breast cancer is the highest incident cancer among the women worldwide. Although the techniques such as x-ray mammography, ultrasound imaging and biopsy are currently used to detect and diagnose breast cancers, they all suffer from drawbacks including high rate of missed positives, use of ionizing radiation and patient discomfort.

[0064] The third preferred embodiment generates C-scan planar images, depicting the entire breast in sequential transverse planes. Focusing of each plane happens in real time with the help of the acoustic lens. The lens will be large enough for the photoacoustic signals generated from the entire breast volume to propagate towards the receiver. A breast holder 837 holds the breast Br.

[0065] A first variation of the third preferred embodiment is shown in Figs. 8A-8D as 800'. The first variation uses the photoacoustic effect only and can be constructed like the system 800 according to the third preferred embodiment, except for the omission of the plane wave generator and any changes set forth herein.

[0066] The device will facilitate the definitive detection of breast cancer and generates C-scan multiplanar images of the breast. The device includes a table that can be built into a hospital bed. It has a vent (sliding door 846) on the top to allow free hanging of the breast that needs to be imaged. During the examination, only one of the breasts (either left or right) will be compressed with a compression plate and the other one that needs to be imaged will be allowed to protrude through the vent provided. The probe will have a diameter between 3 and 5 inches, and a length approximately 3 inches, and it includes an acoustic window that is curved inwards to ensure matching between the breast surface and the acoustic propagation medium (any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl, for example, water) inside the probe. This probe which consists of acoustic lens, water and the sensor will be mounted onto a translation stage that can move horizontally from left to right depending on the side being imaged. Also, this entire stage is provided a vertical translation movement to make sure that the probe is in contact with the breast before the exam is started.

[0067] Focusing of each plane happens in real time with the help of the acoustic lens. The lens will be large enough for the photoacoustic signals generated from the entire breast volume to propagate towards the sensor. The laser source will be a tunable broadband laser with frequencies ranging from 500-1100 nm with a typical pulse width of 10 ns and 1-10 Hz pulse repetition frequency. The laser pulses will be delivered via a bundle of high quality fiber optic cables. The laser beam waist at the output of the fiber-optic cable is typically 2 mm with

intensity below 100 millijoules/cm². The fiber-optic cables are arranged in a circular fashion and are equipped into a ring around the breast. The fiber-optic cables are terminated onto compound lens-beam expanders which can move back and forth for controlled illumination of the breast.

[0068] The laser beams irradiate the breast from all the directions so that the entire breast will become the source of photoacoustic signals. With the help of acoustic coupling gel, the photoacoustic waves generated in the breast enter through the acoustic window and are focused by the acoustic lens onto the sensor and the C-scan images are generated. Water is used as the acoustic propagation medium. Any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl could be used instead. This device will provide the C-scan images corresponding to a tissue thickness less than 1 mm, usually in the range of 700-900 microns over the entire length and width of breast by means of the 2d sensor.

[0069] The acoustic lens will converge the wavefront from the object plane (tissue) onto the image plane (sensor) with a magnification dictated by the object plane-lens and lens-image plane distances. It produces best images when the object plane-lens distance and the corresponding lens-image plane distance are at a distance of twice the lens-focal length with a resulting magnification of 1. The lens will be large enough to ensure that the acoustic signals from the entire breast are received by the sensor.

[0070] The shape of the acoustic window 610' in this case will follow a concave pattern to ensure maximum acoustic match between the breast and the acoustic propagation medium (any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl, for example, water) inside the probe so that the signal-to-noise ratio is maximized.

[0071] A group of high power delivery fiber optic cables will be used to deliver the pulsed high power laser beam from an external laser source. They feature a pure silica core and a bonded hard polymer cladding. The fibers have a larger numerical aperture that allows for greater light coupling efficiencies. The fiber's core diameter is typically 600 μ m and cladding diameter is 630 μ m. The maximum power density that the fiber can withstand is 100 KW/cm². These fibers will be suitable for use in 500 nm to 1100 nm wavelength range.

[0072] Cylindrical and spherical plano-concave lenses will be used as beam expanders 618' to diverge the light beam coming out of the fibers. These lenses will also allow focusing like camera-zoom lenses. Depending on the desired area to be imaged, the area over which the light is irradiated will be controlled with the aid of them. Cylindrical lenses provide four times higher laser exposure on the tissue compared to spherical lenses. The lenses have a focal length of 15 mm which facilitates in illuminating the tissue volume over a width of about 1 cm. The lenses have their clear apertures greater than 90%

of surface dimensions.

[0073] A high energy broadband tunable pulsed laser source 840 with a typical pulse repetition rate of 10 Hz and pulse duration of 10 ns will be employed as the external laser source to generate laser beam of desired size. The exposure energy will be lower than the maximum permissible exposure limits set by ANSI 136.1 standards for the human tissue. A Ti:Sapphire laser coupled to Nd:YAG laser through a frequency doubler will provide a wide range of wavelengths that are suitable for photoacoustic imaging.

[0074] There are two translation stages incorporated in the design. The first one, 842, is for vertical movement of the acoustic lens-sensor configuration along with the circular ring with fiber optic cables to adjust according to the location of the breast. The second one, 844, is for horizontal movement of the detector configuration (along with the lens and the ring with fiber optic cables) to accommodate the left or right breast, which may be accessed through a sliding door 846.

[0075] A second variation of the third preferred embodiment, shown in Fig. 8E as 800", is used for breast ultrasound. It can be constructed like the third preferred embodiment 800, except for the omission of the elements concerned only with the photoacoustic effect and for any differences described herein.

[0076] The device 800" will facilitate the definitive detection of breast cancer and generates C-scan multiplanar images of the breast. An input ultrasound source (plane wave generator) is used for ultrasound pulse echo C-scan imaging. Focusing of each plane happens in real time with the help of the acoustic lens. The lens will be large enough for focusing the ultrasound signals reflected from the entire breast volume to propagate towards the receiver. The breast is compresses during the examination and the C-scan images over the entire volume of the breast are produced by the device. The probe including the acoustic window, acoustic lens and the 2d sensor is brought in contact with the acoustically transparent compression plate from the bottom.

[0077] The probe will have a diameter between 3 and 5 inches, and a length approximately 3 inches. With the help of acoustic coupling gel, when the plane wave generator produces the ultrasound waves, they get reflected from the breast and propagate back through the acoustic window, and will be focused onto the image plane (2d sensor) with the help of an acoustic lens. Water is used as acoustic propagation medium. Any appropriate acoustic propagation medium with acoustic impedance close to 1 M Rayl could be used instead. This device will provide the C-scan images corresponding to a tissue thickness less than 1 mm, usually in the range of 700-900 microns over the entire length and width of breast by means of the 2d sensor.

[0078] The fourth preferred embodiment, shown in Fig. 9 as 900, is a transvaginal photoacoustic and ultrasound C-scan probe. It is constructed like the first through third preferred embodiments, except as described below.

[0079] In the opto-acoustic window 938, an appropriate material will be used that allows both the optical beam and ultrasound waves with minimal transmission losses.

[0080] To diagnose the cause of certain types of infertility, pelvic pain, abnormal bleeding, and menstrual problems, transvaginal ultrasound is used. It is often used to show the lining of the uterus. Though the test may reveal certain conditions such as pelvic infection or monitor the growth of the fetus during the pregnancy, it suffers from false positive results in detecting ovarian cysts, ovarian tumors and the cancers of the uterus, vagina and other pelvic structures.

[0081] The fourth preferred embodiment generates C-scan planar images, depicting the entire uterus or ovaries in sequential planes. Focusing of each plane happens in real time with the help of the acoustic lens.

[0082] The tip of the probe is made optically as well as acoustically transparent with the help of an opto-acoustic window. The delivery of light through two fiber-optic cables ensures that more area is exposed to light thus resulting in the generation of photoacoustic signals from a larger area.

[0083] The probe will have a diameter between 0.75 and 1 inches, and a length approximately 5 inches. The opto-acoustic window will be aligned in front of the uterus when inserted into the vagina. With the help of acoustic coupling gel, the PA ultrasound waves generated in the tissue will enter through the acoustic window and focused onto the sensor by means of acoustic lens.

[0084] While preferred embodiments have been set forth in detail above, those skilled in the art who have reviewed the present disclosure will readily appreciate that other embodiments can be realized within the scope of the invention. For example, numerical examples are illustrative rather than limiting, as are recitations of specific materials. Also, other embodiments can facilitate in detecting and diagnosing the lesions and tumors associated with lymph nodes, the liver, the gall bladder, the pancreas, the thyroid gland, and the skin. Indeed, the present invention could be adapted for any medical or veterinary imaging or even for non-biological imaging. Moreover, while water has been disclosed as an acoustic propagation medium, any other acoustic propagation medium having appropriate acoustic impedance and other qualities could be used instead. Therefore, the present invention should be construed as limited only by the appended claims.

Claims

1. A method for imaging a region of interest, the method comprising:
 - (a) stimulating the region of interest with light from a light source (614) to produce a first set of ultrasound waves through a photo acoustic effect;

- (b) stimulating the region of interest with ultrasound radiation from an ultrasound source (630) to produce a second set of ultrasound waves reflected from the region of interest; the method being characterised that is also comprises:
- (c) focusing the first and second sets of waves through an acoustic lens (306) to produce focused waves;
- (d) detecting the focused waves in a two-dimensional transducer array (102) while applying a time gate to produce time-gated photo acoustic signals, the time-gated photo acoustic signals providing at least one photo acoustic C-scan at any given period of time, step (d) being repeated multiple times to provide real-time imaging;
- (e) detecting the focused waves in the two-dimensional transducer array while applying a time gate to produce time-gated ultrasound signals, the time-gated ultrasound signals providing at least one ultrasound C-scan at any given period of time, step (e) being repeated multiple times to provide said real-time imaging; and
- (f) imaging the region of interest from the at least one photo acoustic C-scan and the at least one ultrasound C-scan provided in steps (d) and (e).
2. The method of claim 1, wherein, during steps (a)-(c), the light source and the transducer array are scanned relative to the region of interest.
3. The method of claim 2, wherein the light source and the transducer array are provided in a rigid assembly, and wherein the rigid assembly is scanned during steps (a)-(c).
4. The method of claim 2, wherein the light source and the transducer array are scanned in two dimensions.
5. The method of claim 2, wherein the light source and the transducer array are scanned in three dimensions.
6. A probe (600) for imaging a region of interest, the probe comprising:
- a light source (614) for stimulating the region of interest with light to produce a first set of ultrasound waves through a photo acoustic effect;
- an ultrasound source (630) for stimulating the region of interest with ultrasound radiation to produce a second set of ultrasound waves reflected from the region of interest;
- the probe being **characterised in that** it also comprises:
- an acoustic lens (306) for focusing the first and second sets of waves through an

acoustic lens to produce focused waves;

a two-dimensional transducer array (102) for detecting the focused first set of waves while applying a time gate to produce time-gated photo acoustic signals, the time-gated photo acoustic signals providing at least one photo acoustic C-scan at any given period of time, said at least one photo acoustic C-scan being repeated multiple times to provide real-time imaging, and for detecting the focused second set of waves while applying a time gate to produce time-gated ultrasound signals, the time-gated ultrasound signals providing at least one ultrasound C-scan at any given period of time, said at least one ultrasound C-scan being repeated multiple times to provide said real-time imaging.

20 Patentansprüche

1. Verfahren zur Abbildung eines Zielbereichs, umfassend

- (a) die Anregung des Zielbereichs mit Licht einer Lichtquelle (614) zur Erzeugung einer ersten Gruppe von Ultraschallwellen durch einen fotoakustischen Effekt;
- (b) die Anregung des Zielbereichs mit Ultraschallstrahlung einer Ultraschallquelle (630) um eine zweite Gruppe von Ultraschallwellen zu erzeugen, die vom Zielbereich reflektiert werden; wobei das Verfahren **dadurch gekennzeichnet ist, dass** es auch
- (c) die Fokussierung der ersten und der zweiten Gruppe von Wellen mittels einer akustischen Linse (306), um fokussierte Wellen zu erzeugen;
- (d) die Erfassung der fokussierten Wellen in einer zweidimensionalen Transduceranordnung (102) während ein Zeittor zur Erzeugung zeitgetakteter fotoakustischer Signale angelegt wird, wobei die zeitgetakteten fotoakustischen Signale in einem beliebigen vorgegebenen Zeitraum mindestens einen fotoakustischen C-Scan liefern, und wobei Schritt (d) zur Bereitstellung einer Echtzeitabbildung mehrmals wiederholt wird;
- (e) die Erfassung der fokussierten Wellen in der zweidimensionalen Transduceranordnung während ein Zeittor zur Erzeugung zeitgetakteter Ultraschallsignale angelegt wird, wobei die zeitgetakteten Ultraschallsignale in einem beliebigen vorgegebenen Zeitraum mindestens einen Ultraschall-C-Scan liefern, wobei Schritt (e) zur Bereitstellung dieser Echtzeitabbildung mehrmals wiederholt wird; und
- (f) Abbildung des Zielbereichs aus dem mindestens einen fotoakustischen C-Scan und dem

mindestens einen Ultraschall-C-Scan, bereitgestellt in den Schritten (d) und (e), umfasst.

2. Verfahren nach Anspruch 1, wobei während der Schritte (a) bis (c) die Lichtquelle und die Transduceranordnung bezüglich des Zielbereichs abtastend bewegt werden. 5
3. Verfahren nach Anspruch 2, wobei die Lichtquelle und die Transduceranordnung in einer starren Baugruppe vorliegen und wobei die starre Baugruppe während der Schritte (a) bis (c) abtastend bewegt wird. 10
4. Verfahren nach Anspruch 2, wobei die Lichtquelle und die Transduceranordnung in zwei Dimensionen abtastend bewegt werden. 15
5. Verfahren nach Anspruch 2, wobei die Lichtquelle und die Transduceranordnung in drei Dimensionen abtastend bewegt werden. 20
6. Eine Sonde (600) zur Abbildung eines Zielbereichs, umfassend: 25
 - eine Lichtquelle (614) zur Anregung des Zielbereichs mit Licht um eine erste Gruppe von Ultraschallwellen durch einen fotoakustischen Effekt zu erzeugen;
 - eine Ultraschallquelle (630) zur Anregung des Zielbereichs mit Ultraschallstrahlung um eine zweite Gruppe von Ultraschallwellen zu erzeugen, die vom Zielbereich reflektiert werden;
 - wobei die Sonde **dadurch gekennzeichnet ist, dass** sie auch umfasst: 30
 - eine akustische Linse (306) zur Fokussierung der ersten und der zweiten Gruppe von Wellen durch eine akustische Linse zur Erzeugung von fokussierten Wellen; 40
 - eine zweidimensionale Transduceranordnung (102) zur Erfassung der fokussierten ersten Wellengruppe, während ein Zeittor zur Erzeugung zeitgetakteter fotoakustischer Signale angelegt wird, wobei die zeitgetakteten fotoakustischen Signale in einem beliebigen vorgegebenen Zeitraum mindestens einen fotoakustischen C-Scan erzeugen, wobei der mindestens eine fotoakustische C-Scan zur Bereitstellung einer Echtzeitabbildung mehrmals wiederholt wird und zur Erfassung der fokussierten zweiten Wellengruppe, während ein Zeittor zur Erzeugung zeitgetakteter Ultraschallsignale angelegt wird, wobei die zeitgetakteten Ultraschallsignale in einem beliebigen vorgegebenen Zeitraum mindestens einen Ultraschall-C-Scan erzeugen, wobei der 50

mindestens eine Ultraschall-C-Scan zur Bereitstellung einer Echtzeitabbildung mehrmals wiederholt wird.

Revendications

1. Dispositif d'imagerie d'une région intéressante, ce procédé comprenant les étapes consistant à :
 - a) stimuler la région intéressante avec de la lumière provenant d'une source de lumière (614) pour produire un premier jeu d'ondes ultrasonores par un effet photo-acoustique,
 - b) stimuler la région intéressante avec une radiation ultrasonore provenant d'une source d'ultrasons (630) pour produire un second jeu d'ondes ultrasonores réfléchies par la région intéressante,
 - ce procédé étant **caractérisé en ce qu'il** comprend également des étapes consistant à :
 - c) focaliser le premier et le second jeux d'ondes au travers d'une lentille acoustique (306) pour produire des ondes focalisées,
 - d) détecter les ondes focalisées dans un réseau de transducteurs bidirectionnel (102) en appliquant un déclenchement chronologique pour produire des signaux photo-acoustiques à déclenchement périodique, les signaux photo-acoustiques à déclenchement périodique permettant d'obtenir au moins d'un balayage photo-acoustique C pendant toute période de temps donnée, l'étape (d) étant répétée de multiples fois pour permettre d'obtenir une imagerie en temps réel,
 - e) détecter les ondes focalisées dans le réseau de transducteurs bidimensionnel en appliquant un déclenchement chronologique pour produire des signaux ultrasonores à déclenchement périodique, les signaux ultrasonores à déclenchement périodique permettant d'obtenir au moins un balayage ultrasonore C pendant toute période de temps donnée, l'étape (e) étant répétée de multiples fois pour permettre d'obtenir l'imagerie en temps réel, et
 - f) réaliser l'imagerie de la région intéressante à partir du balayage photo-acoustique C et du balayage ultrasonore C obtenu dans les étapes (d) et (e).
2. Procédé conforme à la revendication 1, selon lequel au cours des étapes (a)-(c), la source de lumière et le réseau de transducteurs sont balayés par rapport à la région intéressante.
3. Procédé conforme à la revendication 2, selon lequel la source de lumière et le réseau de transducteurs sont réalisés selon un ensemble rigide, et l'ensemble 55

rigide est balayé au cours des étapes (a)-(c).

4. Procédé conforme à la revendication 2, selon lequel la source de lumière et le réseau de transducteurs sont balayés selon deux dimensions. 5

5. Procédé conforme à la revendication 2, selon lequel la source de lumière et le réseau de transducteurs sont balayés selon trois dimensions. 10

6. Sonde (600) pour effectuer l'imagerie d'une région intéressante, cette sonde comprenant :

une source de lumière (614) pour stimuler la région intéressante avec de la lumière pour produire un premier jeu d'ondes ultrasonores par un effet photo-acoustique, 15
une source d'ultrasons (630) pour stimuler la région intéressante avec une radiation ultrasonore pour produire un second jeu d'ondes ultrasonores réfléchies par la région intéressante, 20
cette sonde étant **caractérisée en ce qu'elle** comprend également :

une lentille acoustique (306) pour focaliser 25
le premier et le second jeux d'ondes au travers de cette lentille acoustique pour produire des ondes focalisées,
un réseau de transducteurs bidimensionnel (102) pour détecter le premier jeu d'ondes 30
focalisées en appliquant un déclenchement chronologique pour produire des signaux photo-acoustiques à déclenchement périodique, les signaux photo-acoustiques à déclenchement périodique permettant d'obtenir au moins un balayage photo-acoustique C pendant toute période de temps donnée, ce balayage photo-acoustique C étant répété de multiples fois pour permettre d'obtenir une imagerie en temps réel et pour détecter le second jeu d'ondes focalisées en appliquant un déclenchement chronologique pour produire des signaux ultrasonores à déclenchement périodique, les signaux ultrasonores à déclenchement périodique permettant d'obtenir au moins un balayage ultrasonore C pendant toute période de temps donnée, ce balayage ultrasonore C étant répété de multiples fois pour permettre d'obtenir l'imagerie en temps réel. 50

55

Figure 1




Figure 2





Figure 3






Figure 4: C-scan images of lead pencil exposed by 2mm laser beam




Figure 5 A : Lens based C-scan PA imaging breadboard proof of concept setup




Figure 5β C-Scan planar slices of a PA signal at different time gate centers




Figure 5 C C-Scan image of the lead pencil embedded in chicken breast tissue

Fig. 6
Prostate Photoacoustic and Ultrasound Probe




Fig. 6A
Prostate Ultrasound C-Scan Probe





Fig. 6B






Fig. 7





Fig. 8

Fig. 81+ Breast Photoacoustic C-scan Probe design # 2






Fig. 8C






Fig. 8D


846



800'


Fig. 8 E Breast Ultrasound C-scan Imaging Probe





9511

Figure 1^a Prior Art



REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- US 20090024038 A1, Arnold [0007]
- US 2005018747 A1 [0008]

Non-patent literature cited in the description

- **YADONG WEI et al.** Photoacoustic tomography imaging using a 4f acoustic lens and peak-hold technology. Optics Express, 14 April 2008, vol. 16, 5314-19 [0009]

专利名称(译)	用于c扫描光声成像的低成本设备		
公开(公告)号	EP2328480B1	公开(公告)日	2016-01-06
申请号	EP2009798815	申请日	2009-07-17
[标]申请(专利权)人(译)	罗彻斯特大学		
申请(专利权)人(译)	罗切斯特大学		
当前申请(专利权)人(译)	罗切斯特大学		
[标]发明人	DOGRA VIKRAM S RAO NAVALGUND A H K		
发明人	DOGRA, VIKRAM, S. RAO, NAVALGUND, A., H., K.		
IPC分类号	A61B8/08 A61B5/00 A61B8/12		
CPC分类号	A61B5/0095 A61B5/0059 A61B5/0091 A61B5/415 A61B5/418 A61B5/4312 A61B5/4337 A61B5/4381 A61B5/4571 A61B8/4416 G01S7/52061		
优先权	61/081946 2008-07-18 US		
其他公开文献	EP2328480A4 EP2328480A2		
外部链接	Espacenet		

摘要(译)

用激光刺激前列腺或其他感兴趣区域，产生超声波（光声效应），其由声透镜聚焦并由特定的1或2D传感器阵列捕获并随后在C上扫描显示。显示屏。由激光刺激产生的超声波的幅度与该空间位置处的组织元件的光学吸收成比例。组织吸收的可变性导致C扫描图像对比度。将光声成像与利用施加到感兴趣区域的平面超声波产生的超声C扫描图像组合。

Figure 1

