



EUROPEAN PATENT SPECIFICATION

(45) Date of publication and mention
of the grant of the patent:

12.09.2018 Bulletin 2018/37

(21) Application number: **12839562.1**

(22) Date of filing: **10.10.2012**

(51) Int Cl.:

A61B 1/005 (2006.01)	A61B 1/00 (2006.01)
A61B 1/05 (2006.01)	A61B 1/06 (2006.01)
A61B 5/00 (2006.01)	G02B 23/24 (2006.01)
H04N 5/232 (2006.01)	H04N 5/235 (2006.01)
H04N 9/04 (2006.01)	H04N 5/225 (2006.01)
G06T 7/11 (2017.01)	G06T 7/136 (2017.01)
A61B 1/04 (2006.01)	G06T 7/00 (2017.01)

(86) International application number:

PCT/JP2012/076235

(87) International publication number:

WO 2013/054817 (18.04.2013 Gazette 2013/16)

(54) **ENDOSCOPE SYSTEM AND IMAGE GENERATION METHOD**

ENDOSKOPSYSTEM UND BILDERZEUGUNGSVERFAHREN

SYSTÈME D'ENDOSCOPE ET PROCÉDÉ DE GÉNÉRATION D'IMAGE

(84) Designated Contracting States:

**AL AT BE BG CH CY CZ DE DK EE ES FI FR GB
GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO
PL PT RO RS SE SI SK SM TR**

(30) Priority: **12.10.2011 JP 2011225142**

04.09.2012 JP 2012193909

03.10.2012 JP 2012220933

(43) Date of publication of application:

20.08.2014 Bulletin 2014/34

(73) Proprietor: **Fujifilm Corporation**

Minato-ku

Tokyo 106-8620 (JP)

(72) Inventors:

- **YAMAGUCHI, Hiroshi**
Kanagawa 258-8538 (JP)
- **OZAWA, Satoshi**
Kanagawa 258-8538 (JP)
- **IIDA, Takayuki**
Kanagawa 258-8538 (JP)

(74) Representative: **Klunker IP**

Patentanwälte PartG mbB

Destouchesstraße 68

80796 München (DE)

(56) References cited:

EP-A1- 1 568 307	WO-A2-2007/031946
JP-A- S62 266 028	JP-A- 2010 082 271
JP-B2- 4 762 369	US-A1- 2010 208 047
US-A1- 2011 157 340	

- **TIAN H ET AL: "AUTOMATIC SEGMENTATION ALGORITHM FOR THE EXTRACTION OF LUMEN REGION AND BOUNDARY FROM ENDOSCOPIC IMAGES", MEDICAL AND BIOLOGICAL ENGINEERING AND COMPUTING, SPRINGER, HEILDELBERG, DE, vol. 39, no. 1, 1 January 2001 (2001-01-01), pages 8-14, XP001077972, ISSN: 0140-0118, DOI: 10.1007/BF02345260**
- **Mirko Arnold ET AL: "Automatic Segmentation and Inpainting of Specular Highlights for Endoscopic Imaging", EURASIP Journal on Image and Video Processing, vol. 58, no. 6, 1 January 2010 (2010-01-01), pages 1-12, XP055289896, DE ISSN: 1687-5176, DOI: 10.1109/TITB.2003.813794**
- **SETH A. GROSS ET AL: "Hold On Picasso, Narrow Band Imaging Is Here", AMERICAN JOURNAL OF GASTROENTEROLOGY, vol. 101, no. 12, 1 December 2006 (2006-12-01), pages 2717-2718, XP055289951, US ISSN: 0002-9270, DOI: 10.1111/j.1572-0241.2006.00923.x**

Note: Within nine months of the publication of the mention of the grant of the European patent in the European Patent Bulletin, any person may give notice to the European Patent Office of opposition to that patent, in accordance with the Implementing Regulations. Notice of opposition shall not be deemed to have been filed until the opposition fee has been paid. (Art. 99(1) European Patent Convention).

- NORIYA UEDO ET AL: "ROLE OF NARROWBAND IMAGING FOR DIAGNOSIS OF EARLY-STAGE ESOPHAGOGASTRIC CANCER: CURRENT CONSENSUS OF EXPERIENCED ENDOSCOPISTS IN ASIA-PACIFIC REGION", DIGESTIVE ENDOSCOPY, vol. 23, 29 May 2011 (2011-05-29), pages 58-71, XP055203156, ISSN: 0915-5635, DOI: 10.1111/j.1443-1661.2011.01119.x

Description

BACKGROUND OF THE INVENTION

1. Field of the Invention

[0001] The present invention relates to an endoscope system capable of clearly observing a microstructure such as a pit pattern or an irregular pattern such as hypertrophy, which is formed on body tissue, and an image generation method.

2. Description of the Related Art

[0002] In recent medical treatment, diagnosis or the like using an endoscope apparatus has been widely performed. In this endoscopic diagnosis, not only normal light observation, in which white light of broadband light is used as illumination light within the subject, but also special light observation, in which a lesion, such as cancer, is made clearer than other parts or the position or the size of the lesion is easily intuitively grasped by using the special light having a specific wavelength as illumination light, is performed.

[0003] For example, in JP2001-170009A, using the fact that the degree of penetration in the depth direction of the body tissue and the absorption characteristics of hemoglobin in the blood have a wavelength dependency, a microstructure such as a pit pattern or a microvessel formed in a body tissue surface layer is made clear with blue narrow-band light having a short wavelength, and a thick blood vessel located in a medium-deep layer of the body tissue is made clear with green narrow-band light having a longer wavelength than that of the blue narrow-band light. Blood vessels or superficial microstructures of the surface to medium-deep layers are important clues at the time of differential diagnosis of cancer or degree-of-penetration diagnosis. Therefore, it is possible to greatly improve the accuracy of differentiation and the like by making the blood vessels or the superficial microstructures of the surface to medium-deep layers clear using blue narrow-band light or green narrow-band light.

[0004] In addition, in JP1996-252218A (JP-H08-252218A), a boundary between a lesion part and a normal part is made clear by using the characteristic that the amount of auto-fluorescence emitted from the lesion part, which is thickened due to the lesion such as cancer, is less than the amount of auto-fluorescence from the normal part, which is not thickened, when irradiating the body tissue with excitation light for exciting the auto-fluorescence. By making the boundary between the lesion part and the normal part clear as described above, it becomes easy to grasp the position or the size of the lesion part when performing observation from a distant-view state as at the time of screening.

SUMMARY OF THE INVENTION

[0005] In recent years, there are various kinds of cancer differentiation methods or methods for degree-of-penetration diagnosis. Accordingly, there is not only a case where cancer diagnosis is performed from both a blood vessel pattern, such as a superficial microvessel or a medium-deep layer blood vessel, and an irregular pattern, such as a superficial microstructure or hypertrophy, but also a case where diagnosis is performed by focusing only on the irregular pattern. When performing diagnosis by focusing only on the irregular pattern as described above, it is necessary to reduce the visibility of the blood vessel pattern while improving the visibility of the irregular pattern.

[0006] For making only the irregular pattern clear, there is no description or suggestion in JP2001-170009A. In addition, according to JP1996-252218A (JP-H08-252218A), it is possible to make the hypertrophy of the irregular pattern clear. However, auto-fluorescence used to detect the hypertrophy is weak. Therefore, in order to capture the auto-fluorescence with good sensitivity, a high-sensitivity imaging device such as an EM-CCD is separately required.

[0007] It is an object of the present invention to provide an endoscope system and an image generation method capable of making only irregularities on body tissue, such as a superficial microstructure or a hypertrophy, clear.

[0008] An endoscope system of the present invention includes: the features of claim 1.

[0009] US 2010/208047 A1 discloses an image processing apparatus for processing in vivo images from inside a body cavity. Featured data corresponding to an area of the in-vivo image are calculated and an area whose featured data correspond to a predetermined threshold are extracted as body tissue.

EP 1 568 307 A1 discloses an image processing method an input image is processed by matching it with a template so as to extract a structural component serving as an extraction target in the image.

[0010] Preferably, imaging of the subject is performed by a color imaging device having pixels of a plurality of colors in which respective color separation filters are provided. Preferably, illumination means for sequentially irradiating the subject with light beams of a plurality of colors is further included, and imaging of the subject is performed by a monochrome imaging device whenever sequential irradiation is performed by the illumination means. It is preferable to further include display means for displaying the irregularity image.

[0011] An image generation method of the present invention includes the features of claim 4.

[0012] According to the present invention, the irregularity image obtained by the irregularity image generation means is an image generated by extracting only the information of the irregularities on the subject from the image signal. Therefore, in this irregularity image, not the blood vessel pattern, such as a microvessel, but only the

irregular pattern on the body tissue, such as a superficial microstructure or a hypertrophy, is made clear.

BRIEF DESCRIPTION OF THE DRAWINGS

[0013]

Fig. 1 is a diagram showing an endoscope system of a first embodiment.

Fig. 2 is a diagram showing the internal configuration of the endoscope system of the first embodiment.

Fig. 3 is a graph showing the emission spectrum of white light W that is excited and emitted by excitation light.

Fig. 4A is a diagram showing a non-enlarged image captured when the zoom lens is located at a wide-angle position.

Fig. 4B is a diagram showing an enlarged image captured when the zoom lens is located at a telephoto position.

Fig. 5 is a graph showing the spectral transmittances of color filters of R, G, and B colors.

Fig. 6A is a diagram for explaining the imaging control of a CCD in a normal observation mode, a hypertrophy observation mode, and a hypertrophy and blood vessel observation mode in the first embodiment.

Fig. 6B is a diagram for explaining the imaging control of the CCD in a superficial microstructure observation mode and a microstructure and blood vessel observation mode in the first embodiment.

Fig. 7 is a diagram showing the internal configuration of a special light image generation unit.

Fig. 8 is a diagram for explaining a method of generating a superficial microstructure image.

Fig. 9 is a diagram showing a superficial microstructure enhancement image.

Fig. 10 is a diagram for explaining a method of generating a hypertrophy image.

Fig. 11 is a diagram showing hypertrophy, a part that is not the hypertrophy, and reflected light R1, R2, and R3 from the medium-deep layer blood vessel.

Fig. 12 is a diagram showing a hypertrophy enhancement image.

Fig. 13 is a graph showing the relationship between the brightness ratio B/G and the blood vessel depth.

Fig. 14 is a graph showing the brightness distribution of a part of a B/G image.

Fig. 15 is a graph showing the brightness distribution of a part of a superficial blood vessel extraction image.

Fig. 16 is a graph showing the brightness distribution of a part of a medium-deep layer blood vessel image.

Fig. 17 is a diagram showing a microstructure and blood vessel enhancement image.

Fig. 18 is a diagram showing a hypertrophy and blood vessel enhancement image.

Fig. 19 is a flowchart showing a sequential flow in

the superficial microstructure observation mode or the microstructure and blood vessel observation mode.

Fig. 20 is a flowchart showing a sequential flow in the hypertrophy observation mode or the hypertrophy and blood vessel observation mode.

Fig. 21 is a diagram showing the internal configuration of an endoscope system of a second embodiment.

Fig. 22 is a diagram showing a rotary filter.

Fig. 23 is a graph showing the spectral transmittances of a B filter, a G filter, and an R filter and the emission intensities of broadband light BB and blue narrow-band light BN.

Fig. 24A is a diagram for explaining the imaging control of a CCD in a normal observation mode, a hypertrophy observation mode, and a hypertrophy and blood vessel observation mode in the second embodiment.

Fig. 24B is a diagram for explaining the imaging control of the CCD in a superficial microstructure observation mode and a microstructure and blood vessel observation mode in the second embodiment.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

[0014] As shown in Figs. 1 and 2, an endoscope system 10 of a first embodiment includes: an electronic endoscope 11 (a form of image signal acquisition means) that images the inside of a subject; a processor device 12 that performs various kinds of image processing on the image captured by the electronic endoscope 11; a light source device 13 that supplies light for illuminating the subject to the electronic endoscope 11; and a monitor 14 that displays an image after various kinds of image processing are performed by the processor device 12.

[0015] The electronic endoscope 11 includes a flexible insertion unit 16 that is inserted into the subject, an operating unit 17 provided at the proximal end of the insertion unit 16, and a universal code 18 that makes a connection between the operating unit 17 and the processor device 12 and the light source device 13. A curved portion 19 obtained by connecting a plurality of curved pieces is formed at the distal end of the insertion unit 16. The curved portion 19 is curved in the horizontal and vertical directions by operating an angle knob 21 of the operating unit 17. A distal portion 16a including an optical system for imaging the body cavity and the like is provided at the distal end of the curved portion 19. The distal portion 16a is directed in a desired direction within the subject by the bending operation of the curved portion 19.

[0016] In addition, a mode switch SW 15 for switching to various modes is provided in the operating unit 17. The various modes include a total of five modes of a normal observation mode in which a normal light image obtained by imaging a subject illuminated with white light is displayed on the monitor 14, a superficial microstruc-

ture observation mode in which a superficial microstructure enhancement image emphasizing the microstructure formed on the surface layer of body tissue is displayed on the monitor 14, a hypertrophy observation mode in which a hypertrophy enhancement image that emphasizes a hypertrophy having a thickness from the surface layer to the medium-deep layer in body tissue is displayed on the monitor 14, a microstructure and blood vessel observation mode in which a microstructure and blood vessel enhancement image emphasizing the superficial microstructure and the blood vessel located at a specific depth is displayed on the monitor 14, and a hypertrophy and blood vessel observation mode in which a hypertrophy and blood vessel enhancement image emphasizing hypertrophy and the blood vessel located at a specific depth is displayed on the monitor 14.

[0017] A connector 24 is attached to the universal code 18 on the side of the processor device 12 and the light source device 13. The connector 24 is a composite connector including a communication connector and a light source connector, and the electronic endoscope 11 is detachably connected to the processor device 12 and the light source device 13 through the connector 24.

[0018] As shown in Fig. 2, the light source device 13 includes: an excitation light source 30 that emits excitation light EL having a specific wavelength; a blue narrow-band light source 31 that emits blue narrow-band light BN that is narrowed to a specific wavelength in the blue band; an optical fiber for excitation light 35 on which the excitation light EL from the excitation light source 30 is incident; an optical fiber for blue narrow-band light 36 on which the blue narrow-band light BN from the blue narrow-band light source 31 is incident; a coupler 37 that optically couples the optical fibers 35 and 36 with a light guide 43 in the electronic endoscope; and a light source switching unit 39 for ON/OFF switching of the excitation light source 30 and the blue narrow-band light source 31.

[0019] The excitation light source 30 is a semiconductor light source, such as a laser diode, and emits the excitation light EL having a center wavelength of 445 nm as shown in Fig. 3. The excitation light EL is irradiated to a phosphor 40, which is provided in the distal portion 16a of the electronic endoscope 11, through the optical fiber for excitation light 35, the coupler 37, and the light guide 43. The phosphor 40 is configured to include a plurality of kinds of fluorescent materials (for example, a YAG-based fluorescent material or a fluorescent material, such as BAM ($\text{BaMgAl}_{10}\text{O}_{17}$)) that absorb a part of the excitation light EL and excite and emit fluorescence FL of green to red. The fluorescence FL excited and emitted by the phosphor 40 is combined with the excitation light EL that is transmitted through the phosphor 40 without being absorbed by the phosphor 40, thereby generating white light W.

[0020] The blue narrow-band light source 31 is a semiconductor light source, such as a laser diode, and emits the blue narrow-band light BN having a center wavelength of 405 nm. Since the blue narrow-band light BN

has a degree of penetration to the surface layer of body tissue, the blue narrow-band light BN is used in order to brightly illuminate the superficial microstructure of body tissue, such as a pit pattern.

[0021] The light source switching unit 39 is connected to a controller 59 in the processor device, and switches the excitation light source 30 and the blue narrow-band light source 31 ON (lighting) and OFF (lighting off) according to the mode that is set. When the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode are set, the excitation light source 30 is always ON while the blue narrow-band light source 31 is always OFF. Accordingly, only the white light W is always irradiated to the subject. In contrast, when the superficial microstructure observation mode or the microstructure and blood vessel observation mode is set, the excitation light source 30 is always ON while the blue narrow-band light source 31 is switched ON and OFF in every other frame. Accordingly, irradiation of only the white light W and simultaneous irradiation of the white light W and the blue narrow-band light BN are alternately repeated every other frame.

[0022] The electronic endoscope 11 includes the light guide 43, a CCD 44, an analog processing circuit 45 (AFE: Analog Front End), an imaging control unit 46, and a magnification control unit 47. The light guide 43 is a large-diameter optical fiber, a bundle fiber, or the like, and the incidence end is inserted in the coupler 37 of the light source device and the exit end is directed to the phosphor 40. Light guided by the light guide 43 is irradiated to the subject through the phosphor 40, an irradiation lens 48b, and an illumination window 49. An observation window 50 receives returned light from the subject. The received light is incident on the CCD 44 through a condensing lens 51 and a zoom lens 48a.

[0023] An actuator 48c to move the zoom lens 48a in the optical axis direction is attached to the zoom lens 48a. The driving of the actuator 48c is controlled by the magnification control unit 47 connected to the controller 59. The magnification control unit 47 controls the actuator 48c so that the zoom lens 48a moves to a position corresponding to the magnification set by a zoom operation unit 20. When it is necessary to observe the overall condition in the subject, for example, at the time of screening, the zoom lens 48a is set to a wide-angle position so that a non-enlarged image shown in Fig. 4A is displayed on the monitor 14. On the other hand, when it is necessary to observe the detailed structure of a part to be observed, for example, at the time of differential diagnosis of cancer, the zoom lens 48a is set to a telephoto position so that an enlarged image shown in Fig. 4B is displayed on the monitor 14.

[0024] In the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode, the overall condition in the subject is observed in many cases. Therefore, the zoom lens 48a is set to the wide-angle position in many cases. On the other hand, in the superficial microstructure obser-

vation mode and the microstructure and blood vessel observation mode, an object to be observed is enlarged and observed in many cases. Therefore, the zoom lens 48a is set to the telephoto position in many cases.

[0025] The CCD 44 has an imaging surface 44a on which light from the condensing lens 51 is incident, and performs photoelectric conversion of the light received by the imaging surface 44a and accumulates the signal charges. The accumulated signal charges are read as an imaging signal, and the imaging signal is transmitted to the AFE 45. The CCD 44 is a color CCD, and pixels of three colors of a B pixel in which a color filter of B color is provided, a G pixel in which a color filter of G color is provided, and an R pixel in which a color filter of R color is provided are arrayed on the imaging surface 44a. These color filters of B, G, and R colors have spectral transmittances indicated by curves 52, 53, and 54 shown in Fig. 5.

[0026] The AFE 45 is configured to include a correlated double sampling circuit (CDS), an automatic gain control circuit (AGC), and an analog/digital converter (A/D) (all not shown). The CDS performs correlated double sampling processing on an imaging signal from the CCD 44 to remove noise caused by the driving of the CCD 44. The AGC amplifies an imaging signal from which noise has been removed by the CDS. The A/D converts an imaging signal amplified by the AGC into a digital imaging signal of a predetermined number of bits, and inputs the digital imaging signal to the processor device 12.

[0027] The imaging control unit 46 is connected to the controller 59 in the processor device 12, and transmits a driving signal to the CCD 44 when there is an instruction from the controller 59. The CCD 44 outputs an imaging signal to the AFE 45 at a predetermined frame rate based on the driving signal from the imaging control unit 46.

[0028] In the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode, as shown in Fig. 6A, a step of performing photoelectric conversion of image light of the white light W and accumulating the signal charges and a step of reading the accumulated signal charges are performed within one frame period. This imaging control is repeatedly performed while the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode are set. In each read step, a blue signal B, a green signal G, and a red signal R are respectively output from the B, G, and R pixels of the CCD 44.

[0029] When the superficial microstructure observation mode or the microstructure and blood vessel observation mode is set, as shown in Fig. 6B, a step of performing photoelectric conversion of image light of the white light W and accumulating the signal charges and a step of reading the accumulated signal charges are performed within one frame period. Then, a step of performing photoelectric conversion of image light of the white light W and the blue narrow-band light BN and accumulating the signal charges and a step of reading the

accumulated signal charges are performed within one frame period. This imaging control of two frames is repeatedly performed while the surface observation mode is set. In each read step, a blue signal B1, a green signal G1, and a red signal R1 are respectively output from the B, G, and R pixels of the CCD 44 in the first frame, and a blue signal B2, a green signal G2, and a red signal R2 are output from the B, G, and R pixels, respectively, in the next second frame.

[0030] As shown in Fig. 2, the processor device 12 includes a normal light image generation unit 55, a frame memory 56, a special light image generation unit 57 (a form of irregularity image generation means), and a display control circuit 58. The controller 59 controls each of the units. The normal light image generation unit 55 generates a normal light image from a signal obtained by imaging the image light of the white light W with the electronic endoscope 11. That is, in the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode, a normal light image is generated from the blue signal B, the green signal G, and the red signal R. In the superficial microstructure observation mode or the microstructure and blood vessel observation mode, a normal light image is generated from the blue signal B1, the green signal G1, and the red signal R1. The generated normal light image is temporarily stored in the frame memory 56.

[0031] As shown in Fig. 7, the special light image generation unit 57 includes a superficial microstructure enhancement image generation section 61, a hypertrophy enhancement image generation section 62, and an irregularities and blood vessel enhancement image generation section 63. The superficial microstructure enhancement image generation section 61 includes a brightness signal generation section 61a (first brightness signal generation section) that generates a brightness signal I indicating the average brightness of the subject, a microstructure extraction section 61b (microstructure image generation section) that extracts a superficial microstructure image from the brightness signal I, and an image generation section 61c that generates a superficial microstructure enhancement image by combining the normal light image with the superficial microstructure image extracted by the microstructure extraction section 61b.

[0032] The brightness signal generation section 61a generates the brightness signal I ($= (B2 + G2 + R2)/3$) by dividing the total value ($B2 + G2 + R2$) of the signal values of the signals B2, G2, and R2 of the second frame, which are obtained in the superficial microstructure observation mode or the microstructure and blood vessel observation mode, by 3. Since the brightness signal I is obtained by averaging the signal values of the respective colors, the brightness signal I indicates the average brightness of the subject. As shown in Fig. 8(A), the microstructure extraction section 61b extracts a pixel region, which exceeds a fixed high-brightness threshold value Th1 of the brightness signal I, as a superficial microstructure P. In addition, the microstructure extraction sec-

tion 61b includes a first image generation section that generates a microstructure image by extracting a pixel region exceeding the fixed high-brightness threshold value Th1 of the brightness signal I as the superficial microstructure P. As shown in Fig. 8(B), in a superficial microstructure image in which the superficial microstructure P is extracted, the pixel value of the brightness signal I is used as it is for a pixel region exceeding the high-brightness threshold value Th1, while the pixel value is set to "0" for a pixel region equal to or less than the high-brightness threshold value Th1.

[0033] The reason why the pixel region exceeding the high-brightness threshold value Th1 is set as the superficial microstructure P as described above is as follows. The superficial microstructure is formed of a pit pattern in which micropores formed on the body tissue surface layer gather, for example. Accordingly, when white light or the blue narrow-band light BN is incident on the micropores of the pit pattern, a multiple scattering phenomenon occurs so that the micropores shine brightly. Therefore, a bright region exceeding the high-brightness threshold value Th1 is set as a superficial microstructure. In addition, the superficial microvessel V1 is present between gland ducts in many cases. Since the superficial microvessel shows the strong absorption characteristics for the blue component or the blue narrow-band light BN of white light, the superficial microvessel is displayed darkly on the image. Accordingly, since the brightness of the superficial microvessel V1 is almost less than the high-brightness threshold value Th1, superficial microvessels are hardly present (hardly included) in the superficial microstructure image.

[0034] Then, high-frequency filtering processing is performed on the superficial microstructure image. The brightness value of a part corresponding to the superficial microstructure P changes greatly compared with other parts. Therefore, as shown in Fig. 8(C), a superficial microstructure image 70 in which only the superficial microstructure P is sharply extracted is obtained by performing high-frequency filtering.

[0035] The image generation section 61c generates a superficial microstructure enhancement image 72 shown in Fig. 9 by combining the normal light image with the superficial microstructure image 70 obtained by the microstructure extraction section 61b. The generated superficial microstructure enhancement image 72 is displayed on the monitor 14 by the display control circuit 58. Since the superficial microstructure enhancement image 72 is based on the bright normal light image, it is possible to perform diagnosis while securing the visibility. In the superficial microstructure enhancement image 72, only the superficial microstructure P of the body tissue of the surface layer other than a superficial microvessel V1 is clear. Accordingly, the superficial microstructure enhancement image 72 is an image that is suitable when performing cancer diagnosis by focusing on only the superficial microstructure P.

[0036] As shown in Fig. 7, the hypertrophy enhance-

ment image generation section 62 includes: a gain adjustment section 62a that adjusts the gain of the signals B, G, and R in order to increase a signal including the information of a hypertrophy portion; a brightness signal generation section 62b (second brightness signal generation section) that generates the brightness signal I indicating the average brightness of the subject from the signals B, G, and R after gain adjustment; a hypertrophy extraction section 62c (hypertrophy image generation section) that extracts a hypertrophy image from the brightness signal I; and an image generation section 62d that generates a hypertrophy enhancement image by combining the normal light image with the hypertrophy image extracted by the hypertrophy extraction section.

[0037] The gain adjustment section 62a multiplies a pixel value by the gain of 1 or less for the signal B of the signals B, G, and R obtained in the hypertrophy observation mode or the hypertrophy and blood vessel observation mode, and multiplies a pixel value by the gain of 1 or more for the signals G and R. Since the hypertrophy has a thickness from the surface layer to the medium-deep layer, a lot of information regarding the hypertrophy is included not in the reflected light of light having a short wavelength less than 500 nm but in the reflected light of light having a long wavelength exceeding 500 nm. Therefore, the information regarding the hypertrophy can be made to be noticeable by reducing the pixel value of the signal B including a lot of information of the reflected light of light having a wavelength less than 500 nm and increasing the pixel values of the signals G and R including a lot of information of the reflected light of light having a wavelength exceeding 500 nm.

[0038] The brightness signal generation section 61b generates the brightness signal I ($= (B + G + R)/3$) by dividing the total value ($B + G + R$) of the signal values of the signals B, G, and R after gain adjustment by 3. Since the brightness signal I is obtained by averaging the signal values of the respective colors, the brightness signal I indicates the average brightness of the subject. As shown in Fig. 10(A), the hypertrophy extraction section 62c extracts a pixel region, in which the pixel value is within a medium brightness range M of threshold values Th2 to Th3, of the brightness signal I as hypertrophy Q. In addition, the hypertrophy extraction section 62c includes a second image generation section that extracts a pixel region, in which the pixel value is within a medium brightness range M of threshold values Th2 to Th3, of the brightness signal I as hypertrophy Q. As shown in Fig. 10(B), in a hypertrophy image 80 in which the hypertrophy Q is extracted, the pixel value is used as it is for a pixel region where the pixel value is within the medium brightness range M, while the pixel value is set to "0" for a pixel region where the pixel value is not within the medium brightness range M.

[0039] The reason why the pixel region where the pixel value is within the medium brightness range M is set as the hypertrophy Q is as follows. As shown in Fig. 11, in the hypertrophy Q, the thickness from the surface layer

to the medium-deep layer is increased due to a ridge from the body tissue surface layer, compared with other parts that are not thickened. Accordingly, when emitting light having a wavelength exceeding 500 nm that has a degree of penetration to the medium-deep layer, the amount of reflected light R1 from the hypertrophy Q is reduced compared with the amount of reflected light R2 of other parts that are not thickened. On the other hand, the amount of reflected light R1 from the hypertrophy Q is not reduced compared with the amount of reflected light R3 when emitting light having a wavelength exceeding 500 nm to a mid-depth blood vessel V2. This is because the absorbance of the mucous membrane that forms the hypertrophy Q is lower than the absorbance of hemoglobin in the blood vessel. Therefore, a pixel region where the pixel value is within the medium brightness range M is set as the hypertrophy Q.

[0040] Then, low-frequency to medium-frequency filtering processing is performed on the hypertrophy image 80. A change in the brightness value of the hypertrophy portion is small compared with other parts. Therefore, as shown in Fig. 10(C), the hypertrophy image 80 in which only the hypertrophy is sharply extracted is obtained by performing low-frequency to medium-frequency filtering.

[0041] The image generation section 61d generates a hypertrophy enhancement image 82 shown in Fig. 12 by combining the normal light image with the hypertrophy image 80 obtained by the hypertrophy extraction section 61c. The generated hypertrophy enhancement image 82 is displayed on the monitor 14 by the display control circuit 58. Since the hypertrophy enhancement image 82 is based on the bright normal light image, it is possible to perform diagnosis while securing the visibility. In the hypertrophy enhancement image 82, only the hypertrophy Q of the biological information of the body tissue surface layer is clear. Accordingly, the hypertrophy enhancement image 82 is an image that is suitable when performing cancer diagnosis by focusing on only the hypertrophy.

[0042] As shown in Fig. 7, the irregularities and blood vessel enhancement image generation section 63 includes a B/G image generation section 63a, a blood vessel extraction section 63b, and an irregularities and blood vessel enhancement image generation section 63c. The B/G image generation section 63a generates a B/G image from the signal of the blue component and the signal of the green component among the signals acquired by the imaging of the image light of the white light. In each pixel of the B/G image, information of the brightness ratio B/G obtained by dividing the signal value of the blue signal B by the signal value of the green signal G is included. In the hypertrophy observation mode or the hypertrophy and blood vessel observation mode, the B/G image is generated from B and G. On the other hand, in the superficial microstructure observation mode or the microstructure and blood vessel observation mode, the B/G image is generated from B1 and G1.

[0043] The brightness ratio B/G of each pixel in the

B/G image is relevant to the blood vessel depth. As shown in Fig. 13, there is a proportional relationship in which the brightness ratio B/G increases as the blood vessel depth increases. Accordingly, the magnitude relationship of "brightness of superficial blood vessel < brightness of mucous membrane < brightness of medium-deep layer blood vessel" is satisfied. In addition, as shown in the B/G image of Fig. 14, due to factors such as uneven brightness, there is a case where the brightness of the B/G image is the highest in the middle and decreases toward the periphery from the center. Therefore, the above-described magnitude relationship (brightness of superficial blood vessel < brightness of mucous membrane < brightness of medium-deep layer blood vessel) is locally satisfied, but may not be satisfied on the whole.

[0044] The blood vessel extraction section 63b generates a blood vessel extraction image by extracting a blood vessel located in a specific layer from the B/G image.

The blood vessel extraction is performed by frequency filtering processing. When extracting a superficial blood vessel, high-frequency components, which are frequency band components many of which are present in the superficial blood vessel, are extracted from the B/G image. As a result, as shown in Fig. 15, it is possible to obtain a superficial blood vessel extraction image in which the brightness of the superficial blood vessel is negative and the brightness of a mucous membrane portion is approximately "0". In the superficial blood vessel extraction image, only the superficial blood vessel is extracted sharply. On the other hand, when extracting a medium-deep layer blood vessel, medium-frequency components, which are frequency band components many of which are present in the medium-deep layer blood vessel, are extracted from the B/G image. As a result, as shown in Fig. 16, it is possible to obtain a medium-deep layer blood vessel extraction image in which the brightness of the medium-deep layer blood vessel is positive and the brightness of a mucous membrane portion is approximately "0". In the medium-deep layer blood vessel extraction image, only the medium-deep layer blood vessel is extracted sharply.

[0045] Since the component of the mucous membrane becomes a value close to approximately "0" by performing the frequency filtering processing described above, it is possible to extract only a blood vessel portion. In addition, the above-described magnitude relationship (brightness of superficial blood vessel < brightness of mucous membrane < brightness of medium-deep layer blood vessel) is also satisfied on the whole.

[0046] When the microstructure and blood vessel observation mode is set, the irregularities and blood vessel enhancement image generation section 65 generates a microstructure and blood vessel enhancement image, in which the superficial microstructure and the blood vessel of a specific layer are clear, by combining the superficial microstructure image, the blood vessel extraction image, and the normal light image. The generated microstruc-

ture and blood vessel enhancement image is displayed on the monitor 14 by the display control circuit 58. For example, when a superficial blood vessel extraction image is used as a blood vessel extraction image, a microstructure and blood vessel enhancement image 90 in which both a pit pattern P, which is configured to include a number of gland ducts, and a microvessel V1 between the gland ducts are clear is displayed, as shown in Fig. 17. Thus, by making clear both the pit pattern P and the microvessel V1, it is possible to improve the diagnostic performance in differentiating cancer.

[0047] On the other hand, when the hypertrophy and blood vessel observation mode is set, the irregularities and blood vessel enhancement image generation section 65 generates a hypertrophy and blood vessel enhancement image, in which hypertrophy and the blood vessel of a specific layer are clear, by combining the hypertrophy image, the blood vessel extraction image, and the normal light image. The generated hypertrophy and blood vessel enhancement image is displayed on the monitor 14 by the display control circuit 58. For example, when a medium--deep layer blood vessel extraction image is used as a blood vessel extraction image, a hypertrophy and blood vessel enhancement image 92 in which both the hypertrophy Q and the medium-deep layer blood vessel V2 are clear is displayed, as shown in Fig. 18. Thus, by making clear both the medium-deep layer blood vessel V2 and the hypertrophy Q with a possibility of cancer, it becomes easy to detect a lesion at the time of screening.

[0048] Next, a sequential flow in the superficial microstructure observation mode and the microstructure and blood vessel observation mode will be described with reference to the flowchart shown in Fig. 19.

[0049] When switching to the superficial microstructure observation mode or the microstructure and blood vessel observation mode is performed by the mode switch SW 15, the white light W is irradiated toward the subject. By imaging the subject illuminated with the white light W using the color CCD 44, the blue signal B1, the green signal G1, and the red signal R1 are output from the CCD 44. Then, the white light W and the blue narrow-band light BN are simultaneously irradiated toward the subject. By imaging the subject illuminated with the white light W and the blue narrow-band light BN using the CCD 44, the blue signal B2, the green signal G2, and the red signal R2 are output from the CCD 44.

[0050] Then, a normal light image is generated based on the blue signal B1, the green signal G1, and the red signal R1. Then, a brightness signal I is generated by dividing the total value ($B2 + G2 + R2$) of the signal values of the blue signal B2, G2, and R2 by 3. A superficial microstructure image is generated by extracting a pixel region exceeding the high-brightness threshold value Th1 of the brightness signal I. By performing high-frequency filtering on the generated superficial microstructure image 70, the superficial microstructure in the image is sharpened.

[0051] When the superficial microstructure observa-

tion mode is set, a superficial microstructure enhancement image is generated by combining the normal light image with the superficial microstructure image after frequency filtering. The generated superficial microstructure enhancement image is displayed on the monitor 14. The series of operations described above are repeatedly performed while the superficial microstructure observation mode is set.

[0052] On the other hand, when the microstructure and blood vessel observation mode is set, a B/G image having a brightness ratio B/G between the blue signal B1 and the green signal G1 is generated. After the B/G image is generated, a blood vessel of a specific layer is extracted from the B/G image. As a result, a blood vessel extraction image is obtained. A microstructure and blood vessel enhancement image is generated by combining the blood vessel extraction image, the superficial microstructure enhancement image 70, and the normal light image. The generated microstructure and blood vessel enhancement image is displayed on the monitor 14. The series of operations described above are repeatedly performed while the microstructure and blood vessel observation mode is set.

[0053] Next, a sequential flow in the hypertrophy observation mode and the hypertrophy and blood vessel observation mode will be described with reference to the flowchart shown in Fig. 19.

[0054] When switching to the hypertrophy observation mode or the hypertrophy and blood vessel observation mode is performed by the mode switch SW 15, the white light W is irradiated toward the subject. By imaging the subject illuminated with the white light W using the color CCD 44, the blue signal B, the green signal G, and the red signal R are output from the CCD 44.

[0055] Then, a normal light image is generated based on the blue signal B, the green signal G, and the red signal R. In addition, the gain of the signals B, G, and R is adjusted to reduce the pixel value of the signal B and increase the pixel values of the signals G and R. Then, the brightness signal I is generated by dividing the total value ($B + G + R$) of the signal values of the signals B, G, and R after gain adjustment by 3. A hypertrophy image is generated by extracting a pixel region less than the low-brightness threshold value Th2 of the brightness signal I. By performing low-frequency to medium-frequency filtering on the generated hypertrophy image, the hypertrophy in the image is sharpened.

[0056] When the hypertrophy observation mode is set, a hypertrophy enhancement image is generated by combining the normal light image with the hypertrophy image after frequency filtering. The generated hypertrophy enhancement image is displayed on the monitor 14. The series of operations described above are repeatedly performed while the hypertrophy observation mode is set.

[0057] On the other hand, when the hypertrophy and blood vessel observation mode is set, a B/G image having a brightness ratio B/G between the blue signal B and the green signal G is generated. After the B/G image is

generated, a blood vessel of a specific layer is extracted from the B/G image. As a result, a blood vessel extraction image is obtained. A hypertrophy and blood vessel enhancement image is generated by combining the blood vessel extraction image, the hypertrophy enhancement image, and the normal light image. The generated hypertrophy and blood vessel enhancement image is displayed on the monitor 14. The series of operations described above are repeatedly performed while the hypertrophy and blood vessel observation mode is set.

[0058] In a second embodiment of the present invention, unlike the first embodiment of the simultaneous system using the color CCD, subject imaging is performed in a frame sequential method using the rotary filter of RGB. As shown in Fig. 21, in an endoscope system 100 of the second embodiment, the configuration of a light source device 113 is different from that of the light source device 13 of the first embodiment. In addition, the phosphor 40 is not provided in the distal portion 16a of the electronic endoscope 11. In addition, the configuration of the CCD and the operation of the imaging control unit 46 in the electronic endoscope 11 are different from those in the first embodiment. In addition, the method of generating a normal light image in the microstructure observation mode and the microstructure and blood vessel observation mode is different. Since others are the same as those described in the first embodiment, only the differences from the first embodiment will be described below.

[0059] The light source device 113 includes: a white light source 130 that emits broadband light BB (400 nm to 700 nm); a rotary filter 134 that separates the broadband light BB from the white light source 130 into light beams of three colors of B, G, and R and sequentially supplies these light beams of respective colors to the light guide 43; a semiconductor light source unit 136 that emits the blue narrow-band light BN; and a light joining unit 138 that makes the blue narrow-band light BN join the optical path L1 of the broadband light BB between the rotary filter 134 and the light guide 43.

[0060] The white light source 130 is formed of a xenon lamp, a halogen lamp, a metal halide, or the like. As shown in Fig. 22, the rotary filter 134 is rotatably provided so that a B filter portion 134a, a G filter portion 134b, and an R filter portion 134c are selectively inserted in the optical path L1 of the broadband light BB. The rotary filter 134 has a disc shape, and is divided into three regions, each of which is a fan-shaped region having a central angle of 120°, in the circumferential direction, and the B filter portion 134a, the G filter portion 134b, and the R filter portion 134c are respectively provided in the three regions.

[0061] As shown in Fig. 23, the B filter portion 134a allows B light of the blue band from the broadband light BB to be transmitted therethrough, the G filter portion 134b allows G light of the green band from the broadband light BB to be transmitted therethrough, and the R filter portion 134c allows R light of the red band from the broad-

band light BB to be transmitted therethrough. Therefore, B light, G light, and R light are sequentially emitted from the rotary filter 134 by the rotation of the rotary filter 134.

[0062] The semiconductor light source unit 136 includes a blue narrow-band light source 136a, an optical fiber 136b, and a light source control section 136c. The blue narrow-band light source 136a is a semiconductor light source, such as a laser diode, and emits the blue narrow band BN having a center wavelength of 405 nm as shown in Fig. 23. The blue narrow-band light source 136a is turned on and off according to the control of the light source control section 136c. In the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode, the blue narrow-band light source 236a is always in the OFF state. In the microstructure observation mode and the microstructure and blood vessel observation mode, only the blue narrow-band light source 136a is turned on only when the B filter 134a of the rotary filter 134 is inserted in the optical path L1 of the broadband light BB, and the blue narrow-band light source 136a is turned off when the other G and R filters 134b and 134c are inserted in the optical path L1.

[0063] The blue narrow-band light BN from the blue narrow-band light source 136a is incident on the optical fiber 136b. The blue narrow-band light BN is emitted toward the light joining unit 138 through a collimator lens (not shown). The light joining unit 138 is a dichroic mirror. Light from the rotary filter 134 is transmitted through the light joining unit 138 as it is, while the blue narrow-band light BN from the semiconductor light source unit 136 is reflected from the light joining unit 138. Accordingly, the optical path L2 of the blue narrow-band light BN is made to match the optical path L1 of the broadband light BB. Light emitted from the light joining unit 138 is incident on the light guide 43 through a condensing lens 142.

[0064] A CCD 144 in the electronic endoscope is a monochrome imaging device in which a micro color filter is not provided on the imaging surface, unlike in the first embodiment described above. The imaging control unit 46 that controls the imaging of the CCD 144 also performs a different operation from that in the first embodiment described above.

[0065] In the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode, as shown in Fig. 24A, image light beams of three colors of B, G, and R are sequentially captured and the electric charges are accumulated, and frame sequential imaging signals B, G, and R are sequentially output based on the accumulated electric charges. The series of operations are repeated while the normal observation mode, the hypertrophy observation mode, and the hypertrophy and blood vessel observation mode are set. The frame sequential imaging signals B, G, and R correspond to B, G, and R of the first embodiment, respectively.

[0066] In the microstructure observation mode and the microstructure and blood vessel observation mode, as

shown in Fig. 24B, image light of mixed color light of the light of B color and the blue narrow-band light BN that are simultaneously irradiated, image light of the light of G color, and image light of the light of R color are sequentially captured and the electric charges are accumulated, and frame sequential imaging signals B+BN, G, and R are sequentially output based on the accumulated electric charges. The series of operations are repeated while the surface layer observation mode is set. The frame sequential imaging signals B+BN, G, and R correspond to B2, G2, and R2 of the first embodiment, respectively.

[0067] In the second embodiment, when generating a normal light image in the microstructure observation mode and the microstructure and blood vessel observation mode, a correction signal generation section 55a in the normal light image generation unit 55 generates a correction signal B' by removing the component of the blue narrow-band light BN from the frame sequential imaging signal B+BN, and generates a normal light image using the correction signal B' and the frame sequential imaging signals G and R. The correction signal generation section 55a generates the correction signal B' by multiplying the frame sequential imaging signal B+BN by a correction coefficient k that is determined in advance from the relationship between the amount of light of B color and the amount of blue narrow-band light BN or the like.

[0068] In the first and second embodiments described above, a superficial microstructure enhancement image is generated by combining the normal light image with the superficial microstructure image, and a hypertrophy enhancement image is generated by combining the normal light image with the hypertrophy image. However, the superficial microstructure image or the hypertrophy image may be displayed on the monitor as it is without combination with the normal light image.

[0069] In the first and second embodiments described above, a pixel region where the pixel value exceeds the threshold value Th1 of the brightness signal I is set as the superficial microstructure. Instead of this, however, a pixel region where the pixel value is within the high brightness range N of the threshold values Th4 to Th5 of the brightness signal 1 may be set as the superficial microstructure. The upper limit Th5 of the high brightness range N is set to be lower than the brightness value of halation caused by specular reflection of illumination light. Therefore, it is possible to extract only the superficial microstructure instead of a halation region by performing extraction processing based on the high brightness range N. In addition, the threshold value Th1 is included in the high brightness range N.

Claims

1. An endoscope system, comprising:

an image signal acquisition means for acquiring an image signal by imaging a subject; and
an irregularity image generation means for generating an irregularity image by extracting information of irregularities on the subject from the image signal;

wherein the irregularity image generation means includes a microstructure image generation section (61) that generates a microstructure image as the irregularity image by extracting a microstructure of a body tissue surface layer from the image signals of a plurality of colors obtained by imaging the subject irradiated by white light and blue narrow-band light; the microstructure image generation section includes:

a first brightness signal generation section (61a) that generates a first brightness signal showing an average brightness of the subject based on the image signals of the plurality of colors; and

a microstructure extraction section (61b) that extracts the microstructure image by extracting a pixel region which exceeds a high brightness threshold value (Th1) of the first brightness signal as a microstructure, and

wherein the irregularity image generation means includes a hypertrophy enhancement image generation section (62) that generates a hypertrophy image as the irregularity image by extracting only a hypertrophy, which has a thickness from a body tissue surface layer to a medium-deep layer, including:

a gain adjustment section (62a) adapted to adjust the gain of the image signals of the plurality of colours, obtained by imaging the subject irradiated by white light only,

a second brightness signal generation section (62b) that generates a second brightness signal showing an average brightness of the subject after gain adjustment; and

a hypertrophy extraction section that extracts a hypertrophy image by extracting a pixel region, in which the pixel values are within a medium brightness range (M) of the second brightness signal as the hypertrophy.

2. The endoscope system according to claim 1, wherein high-frequency filtering is performed on the microstructure image.

3. The endoscope system according to claims 1 or 2, further comprising:
a display means for displaying the irregularity image.
4. An image generation method, comprising: 5
- acquiring an image signal by imaging a subject with an imaging device; and
generating an irregularity image by extracting information of irregularities on the subject from the image signal using irregularity image generation means, 10
- wherein the step of generating an irregularity image includes generating a microstructure image as the irregularity image by extracting a microstructure of a body tissue surface layer from the image signals of a plurality of colors obtained by imaging the subject irradiated by white light and blue narrow-band light; 15
- the microstructure image generating step includes: 20
- generating a first brightness signal showing an average brightness of the subject based on the image signals of the plurality of colors; and 25
- extracting the microstructure image by extracting a pixel region which exceeds a high brightness threshold value (Th1) of the first brightness signal as a microstructure, and 30
- wherein the step of generating an irregularity image includes generating a hypertrophy image as the irregularity image by extracting only a hypertrophy, which has a thickness from a body tissue surface layer to a medium-deep layer, further including 35
- a gain adjustment step adjusting the gain of the image signals of the plurality of colours, obtained by imaging the subject irradiated by white light only, 40
- generating a second brightness signal showing an average brightness of the subject after gain adjustment; and
- a hypertrophy extraction step for extracting a hypertrophy image by extracting a pixel region, in which the pixel values are within a medium brightness range (M) of the second brightness signal as the hypertrophy. 45
- 50

Patentansprüche

1. Endoskopsystem, umfassend:
- eine Bildsignal-Erfassungseinrichtung zum Erfassen eines Bildsignals durch Abbilden eines Subjekts; und 55
- eine Unregelmäßigkeitsbild-Erzeugungsein-

richtung zum Erzeugen eines Unregelmäßigkeitsbilds durch Extrahieren von Information über Unregelmäßigkeiten des Subjekts aus dem Bildsignal;

wobei die Unregelmäßigkeitsbild-Erzeugungseinrichtung einen Mikrostrukturbild-Erzeugungsabschnitt (61) aufweist, der ein Mikrostrukturbild als Unregelmäßigkeitsbild dadurch erzeugt, dass er eine Mikrostruktur einer Körpergewebe-Oberflächenschicht aus den Bildsignalen mehrerer Farben extrahiert, die erhalten werden durch Abbilden des mit weißem Licht und blauem, schmalbandigen Licht bestrahlten Objekts;

der Mikrostrukturbild-Erzeugungsabschnitt enthält:

einen ersten Helligkeitssignal-Erzeugungsabschnitt (61a), der ein erstes Helligkeitssignal erzeugt, welches eine durchschnittliche Helligkeit des Objekts darstellt, basierend auf den Bildsignalen der mehreren Farben; und

einen Mikrostruktur-Extrahierabschnitt (61b), der das Mikrostrukturbild dadurch extrahiert, dass er eine Pixelzone, die einen Schwellenwert (Th1) für hohe Helligkeit des ersten Helligkeitssignals überschreitet, als eine Mikrostruktur extrahiert, und wobei die Unregelmäßigkeitsbild-Erzeugungseinrichtung einen Hypertrophieverstärkungsbild-Erzeugungsabschnitt (62) enthält, der ein Hypertrophiebild als Unregelmäßigkeitsbild dadurch erzeugt, dass er nur eine Hypertrophie extrahiert, die eine Dicke von einer Körpergewebe-Oberflächenschicht bis zu einer mitteltiefen Schicht aufweist, enthaltend:

einen Verstärkungseinstellabschnitt (62a), ausgebildet zum Einstellen der Verstärkung der Bildsignale der mehreren Farben, erhalten durch Abbilden des mit nur weißem Licht bestrahlten Subjekts,

einen zweiten Helligkeitssignal-Erzeugungsabschnitt (62b), der ein zweites Helligkeitssignal erzeugt, das eine durchschnittliche Helligkeit des Subjekts nach Verstärkungseinstellung darstellt; und

einen Hypertrophieextrahierabschnitt, der ein Hypertrophiebild dadurch extrahiert, dass er eine Pixelzone, in der die Pixelwerte innerhalb eines mittleren Helligkeitsbereichs (M) des zweiten Helligkeitsbereichs liegen, als die Hypertrophie extrahiert.

2. Endoskopsystem nach Anspruch 1, bei dem eine Hochfrequenzfilterung an dem Mikrostrukturbild ausgeführt wird.
3. Endoskopsystem nach Anspruch 1 oder 2, weiterhin umfassend:
eine Anzeigeeinrichtung zum Anzeigen des Unregelmäßigkeitsbilds.

4. Bilderzeugungsverfahren, umfassend:

Erfassen eines Bildsignals durch Abbilden eines Subjekts mit einer Bildaufnahmeverrichtung; und

Erzeugen eines Unregelmäßigkeitsbilds durch Extrahieren von Information über Unregelmäßigkeiten an dem Subjekt aus dem Bildsignal unter Verwendung einer Unregelmäßigkeitsbild-Erzeugungseinrichtung,

wobei der Schritt des Erzeugens eines Unregelmäßigkeitsbilds das Erzeugen eines Mikrostrukturbilds als das Unregelmäßigkeitsbild durch Extrahieren einer Mikrostruktur einer Körpergewebe-Oberflächenschicht aus den Bildsignalen mehrerer Farben enthält, welche erhalten werden durch Abbilden des Subjekts, das mit weißem Licht und schmalbandigem, blauen Licht bestrahlt wird;

wobei der Mikrostrukturbild-Erzeugungsschritt enthält:

Erzeugen eines ersten Helligkeitssignals, das eine durchschnittliche Helligkeit des Subjekts zeigt, basierend auf den Bildsignalen der mehreren Farben; und

Extrahieren des Mikrostrukturbilds durch Extrahieren einer Pixelzone, die einen Schwellenwert (Th1) hoher Helligkeit des ersten Helligkeitssignals überschreitet, als eine Mikrostruktur, und

wobei der Schritt des Erzeugens eines Unregelmäßigkeitsbilds das Erzeugen eines Hypertrophiebilds als das Unregelmäßigkeitsbild beinhaltet, indem nur eine Hypertrophie extrahiert wird, die eine Dicke von der Körpergewebe-Oberflächenschicht zu einer mitteltiefen Schicht aufweist, weiterhin enthaltend:

einen Verstärkungseinstellschritt des Einstellens der Verstärkung der Bildsignale der mehreren Farben, die durch Abbilden des Subjekts mit lediglich weißem Licht erhalten werden,

Erzeugen eines zweiten Helligkeitssignals, das eine durchschnittliche Helligkeit des Subjekts nach der Verstärkungseinstellung zeigt; und

einen Hypertrophieextrahierschritt des Extrahierens einer Hypertrophiebilds durch Extrahieren einer Pixelzone, in welcher die Pixelwerte innerhalb eines mittleren Helligkeitsbereichs (M) des zweiten Helligkeitssignals liegen, als die Hypertrophie.

10 **Revendications**

1. Système d'endoscope, comprenant :

des moyens d'acquisition de signal d'image destinés à acquérir un signal d'image en imageant un sujet, et

des moyens de génération d'image d'irrégularité, destinés à générer une image d'irrégularité en extrayant des informations d'irrégularités sur le sujet à partir du signal d'image ;

dans lequel les moyens de génération d'image d'irrégularité incluent une section de génération d'image de microstructure (61), laquelle génère une image de microstructure comme image d'irrégularité en extrayant une microstructure d'une couche superficielle de tissu corporel à partir des signaux d'image d'une pluralité de couleurs, obtenus en imageant le sujet irradié par une lumière blanche et une lumière à bande étroite bleue ;

la section de génération d'image de microstructure inclut :

une première section de génération de signal de luminosité (61a), laquelle génère un premier signal de luminosité montrant une luminosité moyenne du sujet sur la base des signaux d'image de la pluralité de couleurs, et

une section d'extraction de microstructure (61b), laquelle extrait l'image de microstructure en extrayant une région de pixels, laquelle dépasse une valeur de seuil de luminosité élevée (Th1) du premier signal de luminosité comme microstructure, et

dans lequel les moyens de génération d'image d'irrégularité incluent une section de génération d'image de mise en valeur d'hypertrophie (62), laquelle génère une image d'hypertrophie comme image d'irrégularité en extrayant uniquement une hypertrophie, laquelle présente une épaisseur allant d'une couche superficielle de tissu corporel à une couche moyennement profonde, incluant :

une section de réglage de gain (62a) apte à régler le gain des signaux d'ima-

ge de la pluralité de couleurs, obtenus en imageant le sujet uniquement par une lumière blanche ;
 une seconde section de génération de signal luminosité (62b), laquelle génère un second signal de luminosité montrant une luminosité moyenne du sujet après réglage de gain, et
 une section d'extraction d'hypertrophie, laquelle extrait une image d'hypertrophie en extrayant une région de pixels, où les valeurs de pixels se trouvent dans les limites d'une plage de luminosité intermédiaire (M) du second signal de luminosité, comme l'hypertrophie.

2. Système d'endoscope selon la revendication 1, dans lequel un filtrage haute fréquence est réalisé sur l'image de microstructure.
3. Système d'endoscope selon la revendication 1 ou 2, comprenant en outre :
des moyens d'affichage destinés à afficher l'image d'irrégularité.
4. Procédé de génération d'image, comprenant les étapes consistant à :

acquérir un signal d'image en imageant un sujet avec un dispositif d'imagerie, et
 générer une image d'irrégularité en extrayant des informations d'irrégularités sur le sujet à partir du signal d'image à l'aide des moyens de génération d'image d'irrégularité,
 dans lequel l'étape pour générer une image d'irrégularité inclut la génération d'une image de microstructure comme image d'irrégularité en extrayant une microstructure d'une couche superficielle de tissu corporel à partir des signaux d'image d'une pluralité de couleurs, obtenus en imageant le sujet irradié par une lumière blanche et une lumière à bande étroite bleue ;
 l'étape de génération d'image de microstructure inclut les étapes consistant à :

générer un premier signal de luminosité montrant une luminosité moyenne du sujet sur la base des signaux d'image de la pluralité de couleurs, et
 extraire l'image de microstructure en extrayant une région de pixels, laquelle dépasse une valeur de seuil de luminosité élevée (Th1) du premier signal de luminosité comme microstructure, et
 dans lequel l'étape de génération d'une image d'irrégularité inclut la génération d'une image d'hypertrophie comme image

d'irrégularité en extrayant uniquement une hypertrophie, laquelle présente une épaisseur allant d'une couche superficielle de tissu corporel à une couche moyennement profonde, incluant en outre :

une étape de réglage de gain réglant le gain des signaux d'image de la pluralité de couleurs, obtenus en imageant le sujet uniquement par une lumière blanche ;
 la génération d'un second signal de luminosité montrant une luminosité moyenne du sujet après réglage de gain, et
 une étape d'extraction d'hypertrophie destinée à extraire une image d'hypertrophie en extrayant une région de pixels, où les valeurs de pixels se trouvent dans les limites d'une plage de luminosité intermédiaire (M) du second signal de luminosité, comme l'hypertrophie.

FIG. 1




FIG. 2




FIG. 3




FIG. 4A




FIG. 4B




FIG. 5




FIG. 6A




FIG. 6B




FIG. 7




FIG. 8




FIG. 9






FIG. 10

(A)



(B)



(C)




FIG. 11




FIG. 12




FIG. 13




FIG. 14




FIG. 15




FIG. 16




FIG. 17




FIG. 18




FIG. 19




FIG. 20




FIG. 21




FIG. 22




FIG. 23




FIG. 24A





FIG. 24B



REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- JP 2001170009 A [0003] [0006]
- JP 8252218 A [0004] [0006]
- JP H08252218 A [0004] [0006]
- US 2010208047 A1 [0009]
- EP 1568307 A1 [0009]

专利名称(译)	内窥镜系统和图像生成方法		
公开(公告)号	EP2767212B1	公开(公告)日	2018-09-12
申请号	EP2012839562	申请日	2012-10-10
[标]申请(专利权)人(译)	富士胶片株式会社		
申请(专利权)人(译)	富士胶片株式会社		
当前申请(专利权)人(译)	富士胶片株式会社		
[标]发明人	YAMAGUCHI HIROSHI OZAWA SATOSHI IIDA TAKAYUKI		
发明人	YAMAGUCHI, HIROSHI OZAWA, SATOSHI IIDA, TAKAYUKI		
IPC分类号	A61B1/005 A61B1/00 A61B1/05 A61B1/06 A61B5/00 G02B23/24 H04N5/232 H04N5/235 H04N9/04 H04N5/225 G06T7/11 G06T7/136 A61B1/04 G06T7/00		
CPC分类号	A61B1/0638 A61B1/00009 A61B1/0051 A61B1/05 A61B1/0669 A61B1/0684 A61B5/0084 G02B23/2461 G06T7/11 G06T7/136 G06T2207/10068 G06T2207/10152 H04N5/23245 H04N5/23296 H04N5/2351 H04N9/045 H04N2005/2255		
优先权	2012193909 2012-09-04 JP 2011225142 2011-10-12 JP 2012220933 2012-10-03 JP		
其他公开文献	EP2767212A1 EP2767212A4		
外部链接	Espacenet		

摘要(译)

仅清楚身体组织上的不规则性，例如表面微结构或肥大。通过使用彩色 CCD 44对用白光W照射的对象进行成像来获得蓝色信号B，绿色信号G和红色信号R。基于这些信号B，G和R，亮度信号I ($(B + G + R) / 3$) 显示对象的平均亮度。将超过亮度信号I的固定阈值Th1的像素区域提取为表面微结构P，例如凹坑图案。通过提取表面微结构P获得的表面微结构图像70显示在监视器14上。

FIG. 1

