



US008727977B2

(12) **United States Patent**
Banet et al.

(10) **Patent No.:** **US 8,727,977 B2**
(45) **Date of Patent:** **May 20, 2014**

(54) **BODY-WORN VITAL SIGN MONITOR**

(75) Inventors: **Matt Banet**, Kihei, HI (US); **Jim Moon**,
Portland, OR (US); **Gunner Trommer**,
Encinitas, CA (US); **Caroline**
DiBattista, San Diego, CA (US)

(73) Assignee: **Sotera Wireless, Inc.**, San Diego, CA
(US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 568 days.

(21) Appl. No.: **12/762,810**

(22) Filed: **Apr. 19, 2010**

(65) **Prior Publication Data**

US 2011/0224557 A1 Sep. 15, 2011

Related U.S. Application Data

(60) Provisional application No. 61/312,624, filed on Mar.
10, 2010.

(51) **Int. Cl.**
A61B 5/00 (2006.01)

(52) **U.S. Cl.**
USPC **600/300; 600/301; 600/545; 600/485**

(58) **Field of Classification Search**
USPC **600/485**
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,086,916 A 5/1978 Freeman et al.
4,263,918 A 4/1981 Swearingen et al.
4,270,547 A 6/1981 Steffen et al.
4,305,400 A 12/1981 Logan
4,577,639 A 3/1986 Simon et al.

4,582,068 A 4/1986 Philipps et al.
4,653,498 A 3/1987 New, Jr. et al.
4,710,164 A 12/1987 Levin et al.
4,722,351 A 2/1988 Philipps et al.
4,802,486 A 2/1989 Goodman et al.
4,807,638 A 2/1989 Sramek
4,905,697 A 3/1990 Heggs et al.
5,025,791 A 6/1991 Niwa
5,140,990 A 8/1992 Jones et al.

(Continued)

FOREIGN PATENT DOCUMENTS

EP 0443267 A1 8/1991
GB 2329250 A 3/1999

(Continued)

OTHER PUBLICATIONS

International Search Report and Written Opinion issued Jul. 22, 2011
in PCT/US2011/027843.

(Continued)

Primary Examiner — Michael Kahelin


Assistant Examiner — Mitchell Alter

(74) *Attorney, Agent, or Firm* — Michael A. Whittaker;
Acuity Law Group, P.C.

(57) **ABSTRACT**

The invention provides a body-worn vital sign monitor that measures a patient's vital signs (e.g. blood pressure, SpO₂, heart rate, respiratory rate, and temperature) while simultaneously characterizing their activity state (e.g. resting, walking, convulsing, falling) and posture (upright, supine). The monitor processes this information to minimize corruption of the vital signs and associated alarms/alerts by motion-related artifacts. It also features a graphical user interface (GUI) rendered on a touchpanel display that facilitates a number of features to simplify and improve patient monitoring and safety in both the hospital and home.

17 Claims, 28 Drawing Sheets



(56)

References Cited

U.S. PATENT DOCUMENTS

5,190,038 A	3/1993	Polson et al.	6,480,729 B2	11/2002	Stone
5,197,489 A	3/1993	Conlan	6,491,647 B1	12/2002	Bridger et al.
5,224,928 A	7/1993	Sibalis et al.	6,503,206 B1	1/2003	Li et al.
5,247,931 A	9/1993	Norwood	6,516,289 B2	2/2003	David
5,289,824 A	3/1994	Mills et al.	6,526,310 B1	2/2003	Carter et al.
5,316,008 A	5/1994	Suga et al.	6,527,729 B1	3/2003	Turcott
5,339,818 A	8/1994	Baker et al.	6,533,729 B1	3/2003	Khair et al.
5,448,991 A	9/1995	Polson et al.	6,541,756 B2	4/2003	Schulz et al.
5,465,082 A	11/1995	Chaco	6,544,173 B2	4/2003	West et al.
5,482,036 A	1/1996	Diab et al.	6,544,174 B2	4/2003	West et al.
5,485,838 A	1/1996	Ukawa et al.	6,546,267 B1	4/2003	Sugiura et al.
5,490,505 A	2/1996	Diab et al.	6,551,252 B2	4/2003	Sackner et al.
5,515,858 A	5/1996	Myllymaki	6,584,336 B1	6/2003	Ali et al.
5,517,988 A	5/1996	Gerhard	6,589,170 B1	7/2003	Flach et al.
5,524,637 A	6/1996	Erickson	6,595,929 B2	7/2003	Stivoric et al.
5,575,284 A	11/1996	Athan et al.	6,605,038 B1	8/2003	Teller et al.
5,588,427 A	12/1996	Tien	6,606,993 B1	8/2003	Wiesmann et al.
5,593,431 A	1/1997	Sheldon	6,616,606 B1	9/2003	Petersen et al.
5,632,272 A	5/1997	Diab et al.	6,650,917 B2	11/2003	Diab et al.
5,645,060 A	7/1997	Yorkey	6,684,090 B2	1/2004	Ali et al.
5,649,543 A	7/1997	Hosaka et al.	6,694,177 B2	2/2004	Eggers et al.
5,680,870 A	10/1997	Hood et al.	6,699,194 B1	3/2004	Diab et al.
5,685,299 A	11/1997	Diab et al.	6,732,064 B1	5/2004	Kadtke et al.
5,709,205 A	1/1998	Bukta	6,745,060 B2	6/2004	Diab et al.
5,743,856 A	4/1998	Oka et al.	6,770,028 B1	8/2004	Ali et al.
5,766,131 A	6/1998	Kondo et al.	6,790,178 B1	9/2004	Mault et al.
5,769,785 A	6/1998	Diab et al.	6,811,538 B2	11/2004	Westbrook et al.
5,800,349 A	9/1998	Isaacson et al.	6,845,256 B2	1/2005	Chin et al.
5,820,550 A	10/1998	Polson et al.	6,850,787 B2	2/2005	Weber et al.
5,848,373 A	12/1998	DeLorme et al.	6,879,850 B2	4/2005	Kimball
5,853,370 A	12/1998	Chance et al.	6,893,396 B2	5/2005	Schulze et al.
5,857,975 A	1/1999	Golub	6,934,571 B2	8/2005	Wiesmann et al.
5,865,755 A	2/1999	Golub	6,947,781 B2	9/2005	Asada et al.
5,865,756 A	2/1999	Peel, III	6,976,958 B2	12/2005	Quy
5,873,834 A	2/1999	Yanagi et al.	6,985,078 B2	1/2006	Suzuki et al.
5,876,353 A	3/1999	Riff	6,997,882 B1	2/2006	Parker et al.
5,895,359 A	4/1999	Peel, III	7,020,508 B2	3/2006	Stivoric et al.
5,899,855 A	5/1999	Brown	7,020,578 B2	3/2006	Sorensen et al.
5,913,827 A	6/1999	Gorman	7,029,447 B2	4/2006	Rantala
5,919,141 A	7/1999	Money et al.	7,041,060 B2	5/2006	Flaherty et al.
5,941,836 A	8/1999	Friedman	7,048,687 B1	5/2006	Reuss et al.
5,964,701 A	10/1999	Asada et al.	7,079,888 B2	7/2006	Oung et al.
5,964,720 A	10/1999	Pelz	7,115,824 B2	10/2006	Lo
5,971,930 A	10/1999	Elghazzawi	7,156,809 B2	1/2007	Quy
6,002,952 A	12/1999	Diab et al.	7,184,809 B1	2/2007	Sterling et al.
6,011,985 A	1/2000	Athan et al.	7,186,966 B2	3/2007	Al-Ali
6,018,673 A	1/2000	Chin et al.	7,194,293 B2	3/2007	Baker, Jr.
6,036,642 A	3/2000	Diab et al.	7,215,984 B2	5/2007	Diab et al.
6,041,783 A	3/2000	Gruenke	7,215,987 B1	5/2007	Sterling et al.
6,057,758 A	5/2000	Dempsey et al.	7,225,007 B2	5/2007	Al-Ali et al.
6,067,462 A	5/2000	Diab et al.	7,241,265 B2	7/2007	Cummings et al.
6,081,735 A	6/2000	Diab et al.	7,257,438 B2	8/2007	Kinast
6,081,742 A	6/2000	Amano et al.	7,296,312 B2	11/2007	Menkedick et al.
6,094,592 A	7/2000	Yorkey et al.	7,299,159 B2	11/2007	Nanikashvili
6,117,077 A	9/2000	Del Mar et al.	7,301,451 B2	11/2007	Hastings
6,129,686 A	10/2000	Friedman	7,314,451 B2	1/2008	Halperin et al.
6,157,850 A	12/2000	Diab et al.	7,351,206 B2	4/2008	Suzuki et al.
6,159,147 A	12/2000	Lichter et al.	7,355,512 B1	4/2008	Al-Ali
6,160,478 A	12/2000	Jacobsen et al.	7,373,191 B2	5/2008	DeLonzor et al.
6,168,569 B1	1/2001	McEwen et al.	7,373,912 B2	5/2008	Self et al.
6,198,394 B1	3/2001	Jacobsen et al.	7,377,794 B2	5/2008	Al-Ali et al.
6,198,951 B1	3/2001	Kosuda et al.	7,382,247 B2	6/2008	Welch et al.
6,199,550 B1	3/2001	Wiesmann et al.	7,383,069 B2	6/2008	Ruchti et al.
6,206,830 B1	3/2001	Diab et al.	7,383,070 B2	6/2008	Diab et al.
6,236,872 B1	5/2001	Diab et al.	7,384,398 B2	6/2008	Gagnadre et al.
6,251,080 B1	6/2001	Henkin et al.	7,400,919 B2	7/2008	Petersen et al.
6,261,247 B1	7/2001	Ishikawa et al.	7,420,472 B2	9/2008	Tran
6,262,769 B1	7/2001	Anderson et al.	7,427,926 B2	9/2008	Sinclair et al.
6,263,222 B1	7/2001	Diab et al.	7,455,643 B1	11/2008	Li et al.
6,287,262 B1	9/2001	Amano et al.	7,468,036 B1	12/2008	Rulkov et al.
6,334,065 B1	12/2001	Al-Ali et al.	7,477,143 B2	1/2009	Albert
6,371,921 B1	4/2002	Caro et al.	7,479,890 B2	1/2009	Lehrman et al.
6,388,240 B2	5/2002	Schulz et al.	7,485,095 B2	2/2009	Shusterman
RE37,852 E	9/2002	Aso et al.	7,502,643 B2	3/2009	Farrington et al.
6,443,890 B1	9/2002	Schulze et al.	7,508,307 B2	3/2009	Albert
			7,509,131 B2	3/2009	Krumm et al.
			7,509,154 B2	3/2009	Diab et al.
			7,522,035 B2	4/2009	Albert
			7,530,949 B2	5/2009	Al Ali et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

7,539,532 B2	5/2009	Tran	2005/0113703 A1	5/2005	Farrington et al.
7,541,939 B2	6/2009	Zadesky et al.	2005/0119586 A1	6/2005	Coyle et al.
7,542,878 B2	6/2009	Nanikashvili	2005/0124866 A1	6/2005	Elaz et al.
7,586,418 B2	9/2009	Cuddihy et al.	2005/0124903 A1	6/2005	Roteliuk et al.
7,598,878 B2	10/2009	Goldreich	2005/0149350 A1	7/2005	Kerr et al.
7,602,301 B1	10/2009	Stirling et al.	2005/0171444 A1	8/2005	Ono et al.
7,616,110 B2	11/2009	Crump et al.	2005/0187796 A1	8/2005	Rosenfeld et al.
7,625,344 B1	12/2009	Brady et al.	2005/0206518 A1	9/2005	Welch et al.
7,628,071 B2	12/2009	Sasaki et al.	2005/0209511 A1	9/2005	Heruth et al.
7,628,730 B1	12/2009	Watterson et al.	2005/0228296 A1	10/2005	Banet
7,641,614 B2	1/2010	Asada et al.	2005/0228298 A1	10/2005	Banet et al.
7,648,463 B1	1/2010	Elhag et al.	2005/0228301 A1	10/2005	Banet et al.
7,656,287 B2	2/2010	Albert et al.	2005/0234317 A1	10/2005	Kiani
7,668,588 B2	2/2010	Kovacs	2005/0240087 A1	10/2005	Keenan et al.
7,670,295 B2	3/2010	Sackner et al.	2005/0261565 A1	11/2005	Lane et al.
7,674,230 B2	3/2010	Reisfeld	2005/0261593 A1	11/2005	Zhang et al.
7,674,231 B2	3/2010	McCombie et al.	2005/0265267 A1	12/2005	Hwang
7,678,061 B2	3/2010	Lee et al.	2005/0283088 A1	12/2005	Bernstein
7,684,954 B2	3/2010	Shahabdeen et al.	2006/0036141 A1	2/2006	Kamath et al.
7,689,437 B1	3/2010	Teller et al.	2006/0047215 A1	3/2006	Newman et al.
7,698,101 B2	4/2010	Alten et al.	2006/0074321 A1	4/2006	Kouchi et al.
7,698,830 B2	4/2010	Townsend et al.	2006/0074322 A1	4/2006	Nitzan
7,698,941 B2	4/2010	Sasaki et al.	2006/0122469 A1	6/2006	Martel
7,715,984 B2	5/2010	Ramakrishnan et al.	2006/0128263 A1	6/2006	Baird
7,725,147 B2	5/2010	Li et al.	2006/0142648 A1	6/2006	Banet et al.
7,782,189 B2	8/2010	Spoonhower et al.	2006/0155589 A1	7/2006	Lane et al.
7,827,011 B2	11/2010	DeVaul et al.	2006/0178591 A1	8/2006	Hempfling
7,976,480 B2	7/2011	Grajales et al.	2006/0200029 A1	9/2006	Evans et al.
7,983,933 B2	7/2011	Karkanas et al.	2006/0252999 A1	11/2006	Devaul et al.
8,047,998 B2	11/2011	Kolluri et al.	2006/0265246 A1	11/2006	Hoag
8,082,160 B2	12/2011	Collins, Jr. et al.	2006/0270949 A1	11/2006	Mathie et al.
8,137,270 B2	3/2012	Keenan et al.	2006/0271404 A1	11/2006	Brown
8,167,800 B2	5/2012	Ouchi et al.	2006/0281979 A1	12/2006	Kim et al.
2001/0004234 A1	6/2001	Petelenz et al.	2007/0010719 A1	1/2007	Huster et al.
2001/0013826 A1	8/2001	Ahmed et al.	2007/0055163 A1	3/2007	Asada et al.
2001/0045395 A1	11/2001	Kitaevich et al.	2007/0066910 A1	3/2007	Inukai et al.
2002/0013517 A1	1/2002	West et al.	2007/0071643 A1	3/2007	Hall et al.
2002/0032386 A1	3/2002	Sackner et al.	2007/0094045 A1	4/2007	Cobbs et al.
2002/0072859 A1 *	6/2002	Kajimoto et al. 702/19	2007/0118056 A1	5/2007	Wang et al.
2002/0151805 A1	10/2002	Sugo et al.	2007/0129769 A1	6/2007	Bourget et al.
2002/0156354 A1	10/2002	Larson	2007/0142715 A1	6/2007	Banet et al.
2002/0170193 A1	11/2002	Townsend et al.	2007/0142730 A1	6/2007	Laermer et al.
2002/0193671 A1	12/2002	Ciurczak et al.	2007/0156456 A1	7/2007	McGillin et al.
2002/0193692 A1	12/2002	Inukai et al.	2007/0161912 A1	7/2007	Zhang et al.
2002/0198679 A1	12/2002	Victor et al.	2007/0185393 A1	8/2007	Zhou et al.
2003/0004420 A1	1/2003	Narimatsu	2007/0188323 A1	8/2007	Sinclair et al.
2003/0097046 A1	5/2003	Sakamaki et al.	2007/0193834 A1	8/2007	Pai et al.
2003/0130590 A1	7/2003	Bui et al.	2007/0208233 A1	9/2007	Kovacs
2003/0135099 A1	7/2003	Al-Ali	2007/0232867 A1	10/2007	Hansmann
2003/0153836 A1	8/2003	Gagnadre et al.	2007/0237719 A1	10/2007	Jones et al.
2003/0158699 A1	8/2003	Townsend et al.	2007/0244376 A1	10/2007	Wang
2003/0167012 A1	9/2003	Friedman et al.	2007/0250261 A1	10/2007	Soehren
2003/0171662 A1	9/2003	O'Connor et al.	2007/0252853 A1	11/2007	Park et al.
2003/0181815 A1	9/2003	Ebner et al.	2007/0255116 A1	11/2007	Mehta et al.
2003/0208335 A1	11/2003	Unuma et al.	2007/0260487 A1	11/2007	Bartfeld et al.
2004/0019288 A1	1/2004	Kinast	2007/0265533 A1	11/2007	Tran
2004/0030261 A1	2/2004	Rantala	2007/0265880 A1	11/2007	Bartfeld et al.
2004/0034293 A1	2/2004	Kimball	2007/0270671 A1	11/2007	Gal
2004/0034294 A1	2/2004	Kimball et al.	2007/0276261 A1	11/2007	Banet et al.
2004/0054821 A1	3/2004	Warren et al.	2007/0282208 A1	12/2007	Jacobs et al.
2004/0073128 A1	4/2004	Hatlestad et al.	2007/0287386 A1	12/2007	Agrawal et al.
2004/0077934 A1	4/2004	Massad	2007/0293731 A1 *	12/2007	Downs et al. 600/300
2004/0111033 A1	6/2004	Oung et al.	2007/0293770 A1	12/2007	Bour et al.
2004/0122315 A1	6/2004	Krill	2007/0293781 A1	12/2007	Sims et al.
2004/0133079 A1	7/2004	Mazar et al.	2008/0004500 A1	1/2008	Cazares et al.
2004/0162466 A1	8/2004	Quy	2008/0004507 A1	1/2008	Williams, Jr. et al.
2004/0162493 A1	8/2004	Mills	2008/0004904 A1	1/2008	Tran
2004/0193063 A1	9/2004	Kimura et al.	2008/0027341 A1	1/2008	Sackner et al.
2004/0225207 A1	11/2004	Bae et al.	2008/0039731 A1	2/2008	McCombie et al.
2004/0267099 A1 *	12/2004	McMahon et al. 600/300	2008/0077026 A1	3/2008	Banet et al.
2005/0027205 A1	2/2005	Tarassenko et al.	2008/0077027 A1	3/2008	Allgeyer
2005/0043598 A1	2/2005	Goode, Jr. et al.	2008/0082001 A1	4/2008	Hatlestad et al.
2005/0059870 A1	3/2005	Aceti	2008/0101160 A1	5/2008	Besson
2005/0070773 A1	3/2005	Chin et al.	2008/0103405 A1	5/2008	Banet et al.
2005/0113107 A1	5/2005	Meunier	2008/0114220 A1	5/2008	Banet et al.
			2008/0132106 A1	6/2008	Burnes et al.
			2008/0139955 A1	6/2008	Hansmann et al.
			2008/0146887 A1	6/2008	Rao et al.
			2008/0146892 A1	6/2008	LeBoeuf et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

2008/0161707 A1 7/2008 Farrington et al.
 2008/0162496 A1 7/2008 Postrel
 2008/0167535 A1 7/2008 Stivoric et al.
 2008/0171927 A1 7/2008 Yang et al.
 2008/0194918 A1 8/2008 Kulik et al.
 2008/0195735 A1 8/2008 Hodges et al.
 2008/0204254 A1 8/2008 Kazuno
 2008/0208013 A1 8/2008 Zhang et al.
 2008/0208273 A1 8/2008 Owen et al.
 2008/0214963 A1 9/2008 Guillemaud et al.
 2008/0221399 A1 9/2008 Zhou et al.
 2008/0221404 A1 9/2008 Tso
 2008/0262362 A1 10/2008 Kolluri et al.
 2008/0275349 A1 11/2008 Halperin et al.
 2008/0281168 A1 11/2008 Gibson et al.
 2008/0281310 A1 11/2008 Dunning et al.
 2008/0287751 A1 11/2008 Stivoric et al.
 2008/0294019 A1 11/2008 Tran
 2008/0319282 A1 12/2008 Tran
 2008/0319327 A1 12/2008 Banet et al.
 2009/0018408 A1 1/2009 Ouchi et al.
 2009/0018409 A1 1/2009 Banet et al.
 2009/0018453 A1 1/2009 Banet et al.
 2009/0040041 A1 2/2009 Janetis et al.
 2009/0054752 A1 2/2009 Jonnalagadda et al.
 2009/0069642 A1 3/2009 Gao et al.
 2009/0076363 A1 3/2009 Bly et al.
 2009/0076397 A1 3/2009 Libbus et al.
 2009/0076398 A1 3/2009 Li et al.
 2009/0076405 A1 3/2009 Amurthur et al.
 2009/0082681 A1 3/2009 Yokoyama et al.
 2009/0112281 A1* 4/2009 Miyazawa et al. 607/46
 2009/0112630 A1 4/2009 Collins, Jr. et al.
 2009/0118590 A1 5/2009 Teller et al.
 2009/0118626 A1 5/2009 Moon et al.
 2009/0131759 A1 5/2009 Sims et al.
 2009/0187085 A1 7/2009 Pav
 2009/0192366 A1 7/2009 Mensinger et al.
 2009/0198139 A1 8/2009 Lewicke et al.
 2009/0221937 A1 9/2009 Smith et al.
 2009/0222119 A1 9/2009 Plahey et al.
 2009/0227877 A1 9/2009 Tran
 2009/0233770 A1 9/2009 Vincent et al.
 2009/0256113 A1 10/2009 Borchers et al.
 2009/0259113 A1* 10/2009 Liu et al. 600/300
 2009/0262074 A1 10/2009 Nasiri et al.
 2009/0264712 A1 10/2009 Baldus et al.
 2009/0287067 A1 11/2009 Dorogusker et al.
 2009/0295541 A1 12/2009 Roof
 2009/0306485 A1 12/2009 Bell
 2009/0306487 A1 12/2009 Crowe et al.
 2009/0306524 A1 12/2009 Muhlsteff et al.
 2009/0312973 A1 12/2009 Hatlestad et al.
 2009/0318779 A1 12/2009 Tran
 2009/0322513 A1 12/2009 Hwang et al.
 2010/0010380 A1 1/2010 Panken et al.
 2010/0030034 A1 2/2010 Schulhauser et al.
 2010/0030085 A1 2/2010 Rojas Ojeda et al.
 2010/0056881 A1 3/2010 Libbus et al.
 2010/0056886 A1 3/2010 Hurtubise et al.
 2010/0113948 A1 5/2010 Yang et al.
 2010/0125188 A1 5/2010 Schilling et al.
 2010/0130811 A1 5/2010 Leuthardt et al.
 2010/0160793 A1 6/2010 Lee et al.
 2010/0160794 A1 6/2010 Banet et al.
 2010/0160795 A1 6/2010 Banet et al.
 2010/0160796 A1 6/2010 Banet et al.
 2010/0160797 A1 6/2010 Banet et al.
 2010/0160798 A1 6/2010 Banet et al.
 2010/0168589 A1 7/2010 Banet et al.
 2010/0210930 A1 8/2010 Saylor
 2010/0217099 A1 8/2010 Leboeuf et al.
 2010/0222649 A1 9/2010 Schoenberg
 2010/0234693 A1 9/2010 Srinivasan et al.
 2010/0234695 A1 9/2010 Morris

2010/0234786 A1 9/2010 Fulkerson et al.
 2010/0241011 A1 9/2010 McCombie et al.
 2010/0280440 A1 11/2010 Skelton et al.
 2010/0298650 A1 11/2010 Moon et al.
 2010/0298651 A1 11/2010 Moon et al.
 2010/0298652 A1 11/2010 McCombie et al.
 2010/0298653 A1 11/2010 McCombie et al.
 2010/0298654 A1 11/2010 McCombie et al.
 2010/0298655 A1 11/2010 McCombie et al.
 2010/0298656 A1 11/2010 McCombie et al.
 2010/0298657 A1 11/2010 McCombie et al.
 2010/0298658 A1 11/2010 McCombie et al.
 2010/0298659 A1 11/2010 McCombie et al.
 2010/0298660 A1 11/2010 McCombie et al.
 2010/0298661 A1 11/2010 McCombie et al.
 2010/0312115 A1 12/2010 Dentinger
 2010/0324384 A1 12/2010 Moon et al.
 2010/0324385 A1 12/2010 Moon et al.
 2010/0324386 A1 12/2010 Moon et al.
 2010/0324387 A1 12/2010 Moon et al.
 2010/0324388 A1 12/2010 Moon et al.
 2010/0324389 A1 12/2010 Moon et al.
 2010/0331640 A1 12/2010 Medina
 2011/0066006 A1 3/2011 Banet et al.
 2011/0066007 A1 3/2011 Banet et al.
 2011/0066008 A1 3/2011 Banet et al.
 2011/0066009 A1 3/2011 Moon et al.
 2011/0066010 A1 3/2011 Moon et al.
 2011/0066037 A1 3/2011 Banet et al.
 2011/0066038 A1 3/2011 Banet et al.
 2011/0066039 A1 3/2011 Banet et al.
 2011/0066043 A1 3/2011 Banet et al.
 2011/0066044 A1 3/2011 Moon et al.
 2011/0066045 A1 3/2011 Moon et al.
 2011/0066050 A1 3/2011 Moon et al.
 2011/0066051 A1 3/2011 Moon et al.
 2011/0066062 A1 3/2011 Banet et al.
 2011/0070829 A1 3/2011 Griffin et al.
 2011/0076942 A1 3/2011 Taveau et al.
 2011/0093281 A1 4/2011 Plummer et al.
 2011/0105862 A1 5/2011 Gies et al.
 2011/0144456 A1 6/2011 Muhlsteff et al.
 2011/0152632 A1 6/2011 Le Neel et al.
 2011/0178375 A1 7/2011 Forster
 2011/0224498 A1 9/2011 Banet et al.
 2011/0224499 A1 9/2011 Banet et al.
 2011/0224500 A1 9/2011 Banet et al.
 2011/0224506 A1 9/2011 Moon et al.
 2011/0224507 A1 9/2011 Banet et al.
 2011/0224508 A1 9/2011 Moon
 2011/0224556 A1 9/2011 Moon et al.
 2011/0224557 A1 9/2011 Banet et al.
 2011/0224564 A1 9/2011 Moon et al.
 2011/0257489 A1 10/2011 Banet et al.
 2011/0257551 A1 10/2011 Banet et al.
 2011/0257552 A1 10/2011 Banet et al.
 2011/0257554 A1 10/2011 Banet et al.
 2011/0257555 A1 10/2011 Banet et al.
 2011/0275907 A1 11/2011 Inciardi et al.
 2012/0065525 A1 3/2012 Douniama et al.
 2012/0123232 A1 5/2012 Najarian et al.

FOREIGN PATENT DOCUMENTS

WO 9932030 A1 7/1999
 WO 2006005169 A1 1/2006
 WO 2007024777 A2 3/2007
 WO 2007143535 A2 12/2007
 WO 2008037820 A1 4/2008
 WO 2008110788 A1 9/2008
 WO 2010135516 A2 11/2010
 WO 2010135518 A1 11/2010
 WO 2010148205 A1 12/2010
 WO 2011032132 A2 3/2011
 WO 2011034881 A1 3/2011
 WO 2011082341 A1 7/2011

(56)

References Cited

FOREIGN PATENT DOCUMENTS

WO 2011112782 A1 9/2011
 WO 2011133582 A1 10/2011

OTHER PUBLICATIONS

Allen et al., Classification of a known sequence of motions and postures from accelerometry data using adapted Gaussian mixture models. *Physiol. Meas.* 2006;27:935-951.

Asada et al., Active Noise Cancellation Using MEMS Accelerometers for Motion-Tolerant Wearable Bio-Sensors. Proceedings of the 26th Annual International Conference of the IEEE EMBS. San Francisco, CA, USA. Sep. 1-5, 2005:2157-2160.

Bowers et al., Respiratory Rate Derived from Principal Component Analysis of Single Lead Electrocardiogram. Computers in Cardiology Conference Proceedings Sep. 2008;35:437-440.

Bussmann et al., Measuring daily behavior using ambulatory accelerometry: The Activity Monitor. *Behav Res Methods Instrum Comput.* Aug. 2001;33(3):349-356.

Cretikos et al., The Objective Medical Emergency Team Activation Criteria: a case-control study. *Resuscitation* Apr. 2007;73(1):62-72.

Espina et al., Wireless Body Sensor Network for Continuous Cuffless Blood Pressure Monitoring. Proceedings of the 3rd IEEE-EMBS. International Summer School and Symposium on Medical Devices and Biosensors. MIT, Boston, USA, Sep. 4-6, 2006:11-15.

Fieselmann et al., Respiratory rate predicts cardiopulmonary arrest for internal medicine patients. *J Gen Intern Med* Jul. 1993;8(7):354-360.

Goldhill et al., A physiologically-based early warning score for ward patients: the association between score and outcome. *Anaesthesia* Jun. 2005;60(6):547-553.

Hung et al., Estimation of Respiratory Waveform Using an Accelerometer. 5th IEEE International Symposium on Biomedical Imaging: From Nano to Macro, May 14-17, 2008:1493-1496.

Jin, A Respiration Monitoring System Based on a Tri-Axial Accelerometer and an Air-Coupled Microphone. Technische Universiteit Eindhoven, University of Technology. Master's Graduation Paper, Electrical Engineering Aug. 25, 2009.

Karantonis et al., Implementation of a Real-Time Human Movement Classifier Using a Triaxial Accelerometer for Ambulatory Monitoring. *IEEE Transactions on Information Technology in Biomedicine*. Jan. 2006;10(1):156-167.

Khambete et al., Movement artefact rejection in impedance pneumography using six strategically placed electrodes. *Physiol. Meas.* 2000;21:79-88.

Khan et al., Accelerometer Signal-based Human Activity Recognition Using Augmented Autoregressive Model Coefficients and Artificial w Neural Nets. 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society. Aug. 20-24, 2008:5172-5175.

Mason, Signal Processing Methods for Non-Invasive Respiration Monitoring. Department of Engineering Science, University of Oxford 2002.

Mathie et al., Classification of basic daily movements using a triaxial accelerometer. *Med Biol Eng Comput.* Sep. 2004;42(5):679-687.

Otto et al., System Architecture of a Wireless Body Area Sensor Network for Ubiquitous Health Monitoring. *Journal of Mobile Multimedia* Jan. 10, 2006;1(4):307-326.

Park et al., An improved algorithm for respiration signal extraction from electrocardiogram measured by conductive textile electrodes using instantaneous frequency estimation. *Med Bio Eng Comput* 2008;46:147-158.

PDF-Pro for iPhone & iPod touch User Manual. ePapyrus Jul. 2009;1:1-25 <http://epapyrus.com/en/files/PDFPro%>.

Seo et al., Performance Improvement of Pulse Oximetry-Based Respiration Detection by Selective Mode Bandpass Filtering. *Ergonomics and Health Aspects of Work with Computers Lecture Notes in Computer Science*, 2007;4566:300-308.

Soh et al., An investigation of respiration while wearing back belts. *Applied Ergonomics* 1997; 28(3):189-192.

Subbe et al., Effect of introducing the Modified Early Warning score on clinical outcomes, cardiopulmonary arrests and intensive care utilization in acute medical admissions. *Anaesthesia* Aug. 2003;58(8):797-802.

Vuorela et al., Two portable long-term measurement devices for ECG and bioimpedance. Second International Conference on Pervasive Computing Technologies for Healthcare.. Jan. 30-Feb. 1, 2008: 169-172.

Wolf et al., Development of a Fall Detector and Classifier based on a Triaxial Accelerometer Demo Board. 2007:210-213.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 30, 2012 in U.S. Appl. No. 12/762,790.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 30, 2012 in U.S. Appl. No. 12/469,236.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 3, 2012 in U.S. Appl. No. 12/469,094.

Restriction Requirement issued by the US Patent and Trademark Office on Feb. 2, 2012 in U.S. Appl. No. 12/469,222.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 27, 2012 in U.S. Appl. No. 12/559,426.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 3, 2012 in U.S. Appl. No. 12/559,039.

Non-Final Office Action issued by the US Patent and Trademark Office on Dec. 29, 2011 in U.S. Appl. No. 12/559,080.

Response to Non-Final Office Action submitted Mar. 19, 2012 in U.S. Appl. No. 12/559,080.

Notice of Allowance issued by the US Patent and Trademark Office on Apr. 2, 2012 in U.S. Appl. No. 12/559,080.

Non-Final Office Action issued by the US Patent and Trademark Office on Dec. 15, 2011 in U.S. Appl. No. 12/560,077.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 8, 2012 in U.S. Appl. No. 12/560,093.

Restriction Requirement issued by the US Patent and Trademark Office on Dec. 14, 2012 in U.S. Appl. No. 12/560,093.

Response to Restriction Requirement submitted Feb. 15, 2012 in U.S. Appl. No. 12/560,093.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 1, 2012 in U.S. Appl. No. 12/560,104.

Restriction Requirement issued by the US Patent and Trademark Office on Jan. 19, 2012 in U.S. Appl. No. 12/469,115.

Response to Restriction Requirement submitted Feb. 15, 2012 in U.S. Appl. No. 12/469,115.

Restriction Requirement issued by the US Patent and Trademark Office on Nov. 14, 2011 in U.S. Appl. No. 12/469,127.

Response to Restriction Requirement submitted Feb. 15, 2012 in U.S. Appl. No. 12/469,127.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 9, 2012 in U.S. Appl. No. 12/469,127.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 3, 2012 in U.S. Appl. No. 12/469,137.

International Preliminary Report on Patentability dated Dec. 1, 2011 issued in PCT/US2010/035554.

International Search Report and Written Opinion dated Sep. 23, 2010 issued in PCT/US2010/035554.

International Preliminary Report on Patentability dated Jan. 5, 2012 issued in PCT/US2010/039000.

International Search Report and Written Opinion dated Sep. 7, 2010 issued in PCT/US2010/039000.

International Search Report and Written Opinion dated Nov. 3, 2010 issued in PCT/US2010/048729.

International Search Report and Written Opinion dated Nov. 5, 2010 issued in PCT/US2010/048866.

International Search Report and Written Opinion dated Mar. 3, 2011 issued in PCT/US2010/062564.

International Search Report and Written Opinion dated Jul. 22, 2011 issued in PCT/US2011/027843.

International Search Report and Written Opinion dated Jul. 20, 2011 issued in PCT/US2011/033100.

Non-Final Office Action issued by the US Patent and Trademark Office on May 26, 2011 in U.S. Appl. No. 12/469,151.

Response to Non-Final Office Action submitted Nov. 25, 2011 in U.S. Appl. No. 12/469,151.

(56)

References Cited**OTHER PUBLICATIONS**

Notice of Allowance issued by the US Patent and Trademark Office on Feb. 1, 2012 in U.S. Appl. No. 12/469,151.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 4, 2011 in U.S. Appl. No. 12/469,182.

Response to Non-Final Office Action submitted Nov. 25, 2011 in U.S. Appl. No. 12/469,182.

Notice of Allowance issued by the US Patent and Trademark Office on Dec. 28, 2011 in U.S. Appl. No. 12/469,182.

International Search Report and Written Opinion dated Oct. 15, 2010 issued in PCT/US2010/035550.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 12, 2012 in U.S. Appl. No. 12/559,429.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 12, 2012 in U.S. Appl. No. 12/559,430.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 24, 2012 in U.S. Appl. No. 12/559,435.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 25, 2012 in U.S. Appl. No. 12/762,733.

Non-Final Office Action issued by the US Patent and Trademark Office on Apr. 27, 2012 in U.S. Appl. No. 12/762,822.

Non-Final Office Action issued by the US Patent and Trademark Office on Mar. 27, 2012 in U.S. Appl. No. 12/559,422.

Mathie, Monitoring and Interpreting Human Movement Patterns using a Triaxial Accelerometer. Faculty of Engineering. The University of New South Wales. PhD Dissertation. Aug. 2003: part1 pp. 1-256.

Mathie, Monitoring and Interpreting Human Movement Patterns using a Triaxial Accelerometer. Faculty of Engineering. The University of New South Wales. PhD Dissertation. Aug. 2003: part2 pp. 256-512.

International Search Report and Written Opinion dated Apr. 27, 2012 as reported in PCT/US2011/067441.

Non-Final Office Action issued by the US Patent and Trademark Office on May 7, 2012 in U.S. Appl. No. 12/469,115.

Non-Final Office Action issued by the US Patent and Trademark Office on May 9, 2012 in U.S. Appl. No. 12/762,836.

Non-Final Office Action issued by the US Patent and Trademark Office on May 10, 2012 in U.S. Appl. No. 12/559,419.

Jackson, Digital Filter Design and Synthesis Using High-Level Modeling Tools. Virginia Polytechnic Institute and State University Thesis. Dec. 1999.

Kim et al., Two Algorithms for Detecting Respiratory Rate from ECG Signal. IFMBE Proceedings 2007;14(6) JC27:4069-4071.

O'Haver, Peak Finding and Measurement, Version 1.6 Oct. 26, 2006. <http://web.archive.org/web/20090205162604/http://terpconnect.umd.edu/~toh/spectrum/PeakFindingandMeasurement.htm>.

Reinvuo et al., Measurement of Respiratory Rate with High-Resolution Accelerometer and EMFit Pressure Sensor. Proceedings of the 2006 IEEE Sensors Applications Symposium Feb 7-9, 2006:192-195.

Non-Final Office Action issued by the US Patent and Trademark Office on May 11, 2012 in U.S. Appl. No. 12/762,846.

Non-Final Office Action issued by the US Patent and Trademark Office on May 11, 2012 in U.S. Appl. No. 12/762,874.

Non-Final Office Action issued by the US Patent and Trademark Office on Jun. 11, 2012 in U.S. Appl. No. 12/469,222.

Non-Final Office Action issued by the US Patent and Trademark Office on Jun. 8, 2012 in U.S. Appl. No. 12/650,383.

Non-Final Office Action issued by the US Patent and Trademark Office on Jun. 8, 2012 in U.S. Appl. No. 12/650,392.

Non-Final Office Action issued by the US Patent and Trademark Office on Jun. 20, 2012 in U.S. Appl. No. 12/762,751.

International Search Report and Written Opinion dated Jun. 29, 2012 issued in PCT/US2012/025640.

Non-Final Office Action issued by the US Patent and Trademark Office on Jul. 5, 2012 in U.S. Appl. No. 12/560,138.

"Signal Strength." Oct. 6, 2008. http://web.archive.org/web/20081006200523/http://en.wikipedia.org/wiki/Signal_strength.

Non-Final Office Action issued by the US Patent and Trademark Office on May 24, 2012 in U.S. Appl. No. 12/560,111.

Restriction Requirement issued by the US Patent and Trademark Office on Apr. 24, 2012 in U.S. Appl. No. 12/469,107.

Response to Restriction Requirement submitted Jun. 14, 2012 in U.S. Appl. No. 12/469,107.

Non-Final Office Action issued by the US Patent and Trademark Office on Jul. 18, 2012 in U.S. Appl. No. 12/650,389.

Chan et al., Noninvasive and Cuffless Measurements of Blood Pressure for Telemedicine. Proceedings of the 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society 2001:3 pages.

Fung, Advisory System for Administration of Phenylephrine Following Spinal Anesthesia for Cesarean Section. Master's Thesis. University of British Columbia 2002: 119 pages.

Liu et al., The Changes in Pulse Transit Time at Specific Cuff Pressures during Inflation and Deflation. Proceedings of the 28th IEEE EMBS Annual International Conference New York City, USA, Aug. 30-Sep. 3, 2006:6404-6405.

Nitzan et al., Effects of External Pressure on Arteries Distal to the Cuff During Sphygmomanometry. IEEE Transactions on Biomedical Engineering, Jun. 2005;52(6):1120-1127.

USB 2.0 Specification Engineering Change Notice. Oct. 20, 2000.

Yan and Zhang, A Novel Calibration Method for Noninvasive Blood Pressure Measurement Using Pulse Transit Time. Proceedings of the 4th IEEE-EMBS International Summer School and Symposium on Medical Devices and Biosensors St Catharine's College, Cambridge, UK, Aug. 19-22, 2007.

Zislin et al., Ways of Improving the Accuracy of Arterial Pressure Oscillometry. Biomedical Engineering 2005;39 (4):174-178.

International Search Report and Written Opinion dated May 29, 2012 issued in PCT/US2012/025648.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 3, 2012 in U.S. Appl. No. 12/762,925.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 3, 2012 in U.S. Appl. No. 12/762,963.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 20, 2012 in U.S. Appl. No. 12/762,777.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 21, 2012 in U.S. Appl. No. 12/469,107.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 24, 2012 in U.S. Appl. No. 12/762,936.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 30, 2012 in U.S. Appl. No. 12/469,202.

Non-Final Office Action issued by the US Patent and Trademark Office on Aug. 31, 2012 in U.S. Appl. No. 12/469,213.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 14, 2012 in U.S. Appl. No. 12/650,374.

Drinnan et al., Relation between heart rate and pulse transit time during paced respiration. Physiol. Meas. Aug. 2001;22(3):425-432.

Flash et al., The Coordination of Arm Movements: An Experimentally Confirmed Mathematical Model. J Neurosci. Jul. 1985;5(7):1688-1703.

Ma and Zhang, A Correlation Study on the Variabilities in Pulse Transit Time, Blood Pressure, and Heart Rate Recorded Simultaneously from Healthy Subjects. Conf Proc IEEE Eng Med Biol Soc. 2005;1:996-999.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 17, 2012 in U.S. Appl. No. 12/469,192.

Gallagher, Comparison of Radial and Femoral Arterial Blood Pressure in Children after Cardiopulmonary Bypass. J Clin Monit. Jul. 1985;1(3):168-171.

Park et al., Direct Blood Pressure Measurements in Brachial and Femoral Arteries in Children. Circulation Feb. 1970; 41(2):231-237.

Talkowski, Quantifying Physical Activity in Community Dwelling Older Adults Using Accelerometry. University of Pittsburgh (Dissertation) 2008:1-91.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 17, 2012 in U.S. Appl. No. 12/650,354.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 21, 2012 in U.S. Appl. No. 12/469,115.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 26, 2012 in U.S. Appl. No. 12/560,104.

(56)

References Cited

OTHER PUBLICATIONS

Packet Definition. The Linux Information Project Jan. 8, 2006 <http://www.linfo.org/packet.html>.

RS-232. Wikipedia Dec. 5, 2008 <http://web.archive.org/web/20081205160754/http://en.wikipedia.org/wiki/RS-232>.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/469,236 dated Sep. 27, 2012.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/487,283 dated Sep. 27, 2012.

Non-Final Office Action issued by the US Patent and Trademark Office on Sep. 28, 2012 in U.S. Appl. No. 12/560,087.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/762,836 dated Oct. 9, 2012.

Non-Final Office Action issued by the US Patent and Trademark Office on Oct. 9, 2012 in U.S. Appl. No. 12/762,726.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/559,429 dated Oct. 12, 2012.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/559,430 dated Oct. 12, 2012.

Final Office Action issued by the US Patent and Trademark Office on Oct. 22, 2012 in U.S. Appl. No. 12/762,822.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/559,435 dated Oct. 23, 2012.

Final Office Action issued by the US Patent and Trademark Office on Oct. 24, 2012 in U.S. Appl. No. 12/599,429.

Final Office Action issued by the US Patent and Trademark Office on Oct. 24, 2012 in U.S. Appl. No. 12/599,430.

Non-Final Office Action issued by the US Patent and Trademark Office on Oct. 23, 2012 in U.S. Appl. No. 12/762,944.

Response to Non-Final Office Action issued in U.S. Appl. No. 12/762,733 dated Oct. 25, 2012.

Final Office Action issued by the US Patent and Trademark Office on Oct. 25, 2012 in U.S. Appl. No. 12/599,426.

Alves et al., CAN Protocol: A Laboratory Prototype for Fieldbus Applications. XIX IMEKO World Congress Fundamental and Applied Metrology Sep. 6-11, 2009, Lisbon, Portugal. 4 pages :454-457 ISBN 978-963-88410-0-1.

Benefits of Digital Sensors. Gems Sensors. Feb. 14, 2008. <http://web.archive.org/web/20080214122230/http://www.sensorland.com/HowPage054.html>.

Final Office Action issued by the US Patent and Trademark Office on Oct. 25, 2012 in U.S. Appl. No. 12/762,790.

Final Office Action issued by the US Patent and Trademark Office on Oct. 26, 2012 in U.S. Appl. No. 12/762,836.

Non-Final Office Action issued by the US Patent and Trademark Office on Oct. 30, 2012 in U.S. Appl. No. 12/559,386.

Non-Final Office Action issued by the US Patent and Trademark Office on Nov. 6, 2012 in U.S. Appl. No. 12/559,379.

Non-Final Office Action issued by the US Patent and Trademark Office on Nov. 6, 2012 in U.S. Appl. No. 12/650,370.

Poon and Zhang, Cuff-Less and Noninvasive Measurements of Arterial Blood Pressure by Pulse Transit Time. Conf Proc IEEE Eng Med Biol Soc. 2005;6:5877-5880.

Non-Final Office Action issued by the US Patent and Trademark Office on Nov. 7, 2012 in U.S. Appl. No. 12/559,392.

Non-Final Office Action issued by the US Patent and Trademark Office on Oct. 24, 2012 in U.S. Appl. No. 12/559,403.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/559,413 on Nov. 9, 2012.

Response to Office Action issued in U.S. Appl. No. 12/762,846 dated Nov. 13, 2012.

Response to Office Action issued in U.S. Appl. No. 12/762,874 dated Nov. 13, 2012.

Response to Office Action issued in U.S. Appl. No. 12/560,111 dated Nov. 26, 2012.

Response to Office Action issued in U.S. Appl. No. 11/930,881 dated Nov. 26, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/559,419 on Nov. 16, 2012.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/559,408 on Nov. 23, 2012.

Response to Office Action issued in U.S. Appl. No. 12/138,199 dated Nov. 29, 2012.

Response to Office Action issued in U.S. Appl. No. 12/650,383 dated Dec. 7, 2012.

Response to Office Action issued in U.S. Appl. No. 12/650,392 dated Dec. 7, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/559,435 on Dec. 12, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/560,111 on Dec. 12, 2012.

Clifford et al., Measuring Tilt with Low-g Accelerometers. Freescale Semiconductor, Inc., 2005:8 pages.

McKneely et al., Plug-and-Play and Network-Capable Medical Instrumentation and Database with a Complete Healthcare Technology Suite: MediCAN. Joint Workshop on High Confidence Medical Devices, Software, and Systems and Medical Device Plug-and-Play Interoperability. 2007:122-129.

Montgomery et al., Lifeguard—A Personal Physiological Monitor for Extreme Environments. Conf Proc IEEE Eng Med Biol Soc. 2004;3:2192-2195.

Thongpithoonrat et al., Networking and Plug-and-Play of Bedside Medical Instruments. Conf Proc IEEE Eng Med Biol Soc. 2008;2008:1514-1517.

Yang et al., Research on Multi-Parameter Physiological Monitor Based on CAN Bus. IFMBE Proceed. 2008;19:417-419.

Zeltwanger, Controller Area Network and CANopen in Medical Equipment. Bus Briefing: Med Dev Manuf Technol. 2002:34-37.

Zitzmann and Schumann, Interoperable Medical Devices Due to Standardized CANopen Interfaces. Joint Workshop on High Confidence Medical Devices, Software, and Systems and Medical Device Plug-and-Play Interoperability. 2007:97-103.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 13/432,976 on Dec. 14, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/762,733 on Dec. 20, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/762,846 on Dec. 20, 2012.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/650,392 on Jan. 3, 2013.

Final Rejection issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/487,283 on Jan. 3, 2013.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 13/292,923 on Jan. 14, 2013.

Notice of Allowance issued by the United States Patent and Trademark Office in U.S. Appl. No. 11/470,708 on Jan. 18, 2013.

International Search Report and Written Opinion issued in PCT/US2012/064302 on Jan. 15, 2013.

Supplemental European Search Report issued in EP 10778376 dated Jan. 31, 2013.

Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/559,039 dated Feb. 11, 2013.

Reddan et al., Intradialytic Blood vol. Monitoring in Ambulatory Hemodialysis Patients: A Randomized Trial. J Am Soc Nephrol. Jul. 2005;16(7):2162-2169.

Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/469,222 dated Feb. 13, 2013.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/650,383 dated Feb. 15, 2013.

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 13/346,408 dated Feb. 25, 2013.

Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/650,389 dated Mar. 14, 2013.

Klabunde, Mean Arterial Pressure. Cardiovascular Physiology Concepts. Mar. 8, 2007. <http://web.archive.org/web/20070308182914/http://www.cvphysiology.com/Blood%20Pressure/BP006.htm>.

Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/762,874 dated Mar. 14, 2013.

(56)

References Cited

OTHER PUBLICATIONS

Non Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 13/196,326 dated Mar. 22, 2013.
De Scalzi et al., Relationship Between Systolic Time Intervals and Arterial Blood Pressure. Clin Cardiol. 1986;9:545-549.

Ahlstrom et al., Noninvasive investigation of blood pressure changes using the pulse wave transit time: a novel approach in the monitoring of hemodialysis patients. J Artif Organs. 2005;8(3):192-197.

Final Office Action issued by the United States Patent and Trademark Office in U.S. Appl. No. 12/762,751 dated Mar. 29, 2013.

* cited by examiner




Fig. 1




Fig. 2B




Fig. 2A




Fig. 3




Fig. 4





Fig. 5






Fig. 7




Fig. 8




Fig. 9




Fig. 10




Fig. 11




Fig. 12




Fig. 13




Fig. 14




Fig. 15




Fig. 16




Fig. 17




Fig. 18



Fig. 19




Fig. 20

Configuration for Indexing cNIBP Measurement (first ~60 seconds) Configuration for cNIBP/RR Measurements (remaining 3 hours, 59 minutes)




Fig. 21A




Fig. 21B




Fig. 22A




Fig. 22B




Fig. 23B




Fig. 23A




Fig. 24




Fig. 25




Fig. 26




Fig. 27




Fig. 28

BODY-WORN VITAL SIGN MONITOR**CROSS REFERENCES TO RELATED APPLICATIONS**

This application claims the benefit of U.S. Provisional Application No. 61/312,624, filed Mar. 10, 2010, entitled "BODY-WORN VITAL SIGN MONITOR", which is incorporated herein in its entirety.

STATEMENT REGARDING FEDERALLY SPONSORED RESEARCH OR DEVELOPMENT

Not Applicable

BACKGROUND OF THE INVENTION**1. Field of the Invention**

The present invention relates to medical devices for monitoring vital signs, e.g., arterial blood pressure.

2. Description of the Related Art

Conventional vital sign monitors are used throughout the hospital, and are particularly commonplace in high-acuity areas such as the intensive care unit (ICU), emergency department (ED), or operating room (OR). Patients in these areas are generally sick and require a high degree of medical attention, typically provided by a relatively high ratio of clinicians compared to lower-acuity areas of the hospital. Outside the ICU and OR, clinicians typically measure vital signs such as systolic, diastolic, and mean arterial blood pressures (SYS, DIA, MAP), respiratory rate (RR), oxygen saturation (SpO₂), heart rate (HR), and temperature (TEMP) with portable or wall-mounted vital sign monitors. It can be difficult to effectively monitor patients in this way, however, because measurements are typically made every few hours, and the patients are often ambulatory and not constrained to a single hospital room. This poses a problem for conventional vital sign monitors, which are typically heavy and unwieldy, as they are not intended for the ambulatory population. To make a measurement, a patient is typically tethered to the monitor with a series of tubes and wires. Some companies have developed ambulatory vital sign monitors with limited capabilities (e.g. cuff-based blood pressure using oscillometry and SpO₂ monitoring), but typically these devices only make intermittent, rather than continuous, measurements. And even these measurements tend to work best on stationary patients, as they are easily corrupted by motion-related artifacts.

Most vital signs monitors feature a user interface that shows numerical values and waveforms associated with the vital signs, alarm parameters, and a 'service menu' that can be used to calibrate and maintain the monitor. Some monitors have internal wireless cards that communicate with a hospital network, typically using protocols such as 802.11b/g.

One of the most important parameters measured with vital signs monitors is blood pressure. In critical care environments like the ICU and OR, blood pressure can be continuously monitored with an arterial catheter inserted in the patient's radial or femoral artery. Alternatively, blood pressure can be measured intermittently with a cuff using oscillometry, or manually by a clinician using auscultation. Most vital sign monitors perform both catheter and cuff-based measurements of blood pressure. Blood pressure can also be monitored continuously with a technique called pulse transit time (PTT), defined as the transit time for a pressure pulse launched by a heartbeat in a patient's arterial system. PTT has been shown in a number of studies to correlate to SYS, DIA, and MAP. In these studies, PTT is typically measured with a

conventional vital signs monitor that includes separate modules to determine both an electrocardiogram (ECG) and SpO₂. During a PTT measurement, multiple electrodes typically attach to a patient's chest to determine a time-dependent ECG component characterized by a sharp spike called the 'QRS complex'. The QRS complex indicates an initial depolarization of ventricles within the heart and, informally, marks the beginning of the heartbeat and a pressure pulse that follows.

SpO₂ is typically measured with a bandage or clothespin-shaped sensor that clips to a patient's finger and includes optical systems operating in both the red and infrared spectral regions. A photodetector measures radiation emitted from the optical systems that transmits through the patient's finger. Other body sites, e.g., the ear, forehead, and nose, can also be used in place of the finger. During a measurement, a microprocessor analyses both red and infrared radiation detected by the photodetector to determine the patient's blood oxygen saturation level and a time-dependent waveform called a photoplethysmograph (PPG). Time-dependent features of the PPG indicate both pulse rate and a volumetric absorbance change in an underlying artery caused by the propagating pressure pulse.

Typical PTT measurements determine the time separating a maximum point on the QRS complex (indicating the peak of ventricular depolarization) and a foot of the PPG waveform (indicating the beginning the pressure pulse). PTT depends primarily on arterial compliance, the propagation distance of the pressure pulse (which is closely approximated by the patient's arm length), and blood pressure. To account for patient-dependent properties, such as arterial compliance, PTT-based measurements of blood pressure are typically 'calibrated' using a conventional blood pressure cuff and oscillometry. Typically during the calibration process the blood pressure cuff is applied to the patient, used to make one or more blood pressure measurements, and then left for future measurements. Going forward, the calibration measurements are used, along with a change in PTT, to measure the patient's continuous blood pressure (cNIBP). PTT typically relates inversely to blood pressure, i.e., a decrease in PTT indicates an increase in blood pressure.

A number of issued U.S. Patents describe the relationship between PTT and blood pressure. For example, U.S. Pat. Nos. 5,316,008; 5,857,975; 5,865,755; and 5,649,543 each describe an apparatus that includes conventional sensors that measure both ECG and PPG waveforms which are then processed to determine PTT.

SUMMARY OF THE INVENTION

To improve the safety of hospitalized patients, particularly those in lower-acuity areas, it is desirable to have a body-worn monitor that continuously measures all vital signs from a patient, provides tools for effectively monitoring the patient, and wirelessly communicates with a hospital's information technology (IT) network. Preferably the monitor operates algorithms featuring: 1) a low percentage of false positive alarms/alerts; and 2) a high percentage of true positive alarms/alerts. The term 'alarm/alert', as used herein, refers to an audio and/or visual alarm generated directly by a monitor worn on the patient's body, or alternatively a remote monitor (e.g., a central nursing station). To accomplish this, the invention provides a body-worn monitor that measures a patient's vital signs (e.g. cNIBP, SpO₂, HR, RR, and TEMP) while simultaneously characterizing their activity state (e.g. resting, walking, convulsing, falling) and posture (upright, supine). The body-worn monitor processes this information

to minimize corruption of the vital signs and associated alarms/alerts by motion-related artifacts.

The body-worn monitor features a graphical user interface (GUI) rendered on a touchpanel display that facilitates a number of features to simplify and improve patient monitoring and safety in both the hospital and home. For example, the monitor features a battery-powered, wrist-worn transceiver that processes motion-related signals generated with an internal motion sensor (e.g. an accelerometer). When the transceiver's battery runs low, the entire unit can be swapped out by simply 'bumping' the original transceiver with a new one having a fully charged battery. Accelerometers within the transceivers detect the 'bump', digitize the corresponding signals, and wirelessly transmit them to a patient data server (PDS) within the hospital's network. There, the signals are analyzed and patient information (e.g. demographic and vital sign data) formerly associated with the original transceiver is re-associated with the new transceiver. A clinician can view the data using a computer functioning as a remote viewing device (RVD), such as a conventional computer on wheels (COW).

The body-worn monitor additionally includes a speaker, microphone, and software that collectively facilitate voice over IP (VOIP) communication. With these features, the wrist-worn transceiver can be used as a two-way communicator allowing, e.g., the patient to alert a clinician during a time of need. Additionally, during medical procedures or diagnoses, the clinician can enunciate annotations directly into the transceiver. These annotations along with vital sign information are wirelessly transmitted to the PDS and ultimately a hospital's electronic medical records (EMR) system, where they are stored and used for post-hoc analysis of the patient. In a related application, the transceiver includes a barcode scanner that, prior to administering medications, scans barcodes associated with the patient, clinician, and medications. The transceiver sends the decoded barcode information back to the PDS, where a software program analyzes it to determine that there are no errors in the medication or the rate at which it is delivered. A signal is then sent from the PDS to the GUI, clearing the clinician to administer the medications.

The body-worn monitor can determine a patient's location in addition to their vital signs and motion-related properties. Typically, the location-determining sensor and the wireless transceiver operate on a common wireless system, e.g. a wireless system based on 802.11a/b/g/n, 802.15.4, or cellular protocols. In this case a location is determined by processing the wireless signal with one or more algorithms known in the art. These include, for example, triangulating signals received from at least three different wireless base stations, or simply estimating a location based on signal strength and proximity to a particular base station. In still other embodiments the location sensor includes a conventional global positioning system (GPS).

VOIP-based communications typically take place between the body-worn monitor and a remote computer or telephone interfaced to the PDS. The location sensor, wireless transceiver, and first and second voice interfaces can all operate on a common wireless system, such as one of the above-described systems based on 802.11 or cellular protocols. In embodiments, the remote computer, for example, can be a monitor that is essentially identical to the transceiver worn by the patient, and can be carried or worn by a clinician. In this case the monitor associated with the clinician features a display wherein the user can select to display information (e.g. vital signs, location, and alarms) corresponding to a particular

patient. This monitor can also include a voice interface so the clinician can communicate with the patient.

The wrist-worn transceiver's touchpanel display can render a variety of different GUIs that query the patient for their pain level, test their degree of 'mentation', i.e. mental activity, and perform other functions to assist and improve diagnosis. Additionally, the transceiver supports other GUIs that allow the patient to order food within the hospital, change the channel on their television, select entertainment content, play games, etc. To help promote safety in the hospital, the GUI can also render a photograph or video of the patient or, in the case of neo-natal patients, their family members.

The body-worn monitor can include a software framework that generates alarms/alerts based on threshold values that are either preset or determined in real time. The framework additionally includes a series of 'heuristic' rules that take the patient's activity state and motion into account, and process the vital signs accordingly. These rules, for example, indicate that a walking patient is likely breathing and has a regular heart rate, even if their motion-corrupted vital signs suggest otherwise.

The body-worn monitor features a series of sensors that attach to the patient to measure time-dependent PPG, ECG, ACC, oscillometric (OSC), and impedance pneumography (IP) waveforms. A microprocessor (CPU) within the monitor continuously processes these waveforms to determine the patient's vital signs, degree of motion, posture and activity level. Sensors that measure these signals typically send digitized information to the wrist-worn transceiver through a serial interface, or bus, operating on a controlled area network (CAN) protocol. The CAN bus is typically used in the automotive industry, and allows different electronic systems to effectively and robustly communicate with each other with a small number of dropped packets, even in the presence of electrically noisy environments. This is particularly advantageous for ambulatory patients that may generate signals with large amounts of motion-induced noise.

Blood pressure is determined continuously and non-invasively using a technique, based on PTT, which does not require any source for external calibration. This technique, referred to herein as the 'Composite Technique', determines blood pressure using PPG, ECG, and OSC waveforms. The Composite Technique is described in detail in the co-pending patent application, the contents of which are fully incorporated herein by reference: BODY-WORN SYSTEM FOR MEASURING CONTINUOUS NON-INVASIVE BLOOD PRESSURE (CNIBP) (U.S. Ser. No. 12/650,354; filed Nov. 15, 2009). In other embodiments, PTT can be calculated from time-dependent waveforms other than the ECG and PPG, and then processed to determine blood pressure. In general, PTT can be calculated by measuring a temporal separation between features in two or more time-dependent waveforms measured from the human body. For example, PTT can be calculated from two separate PPGs measured by different optical sensors disposed on the patient's fingers, wrist, arm, chest, ear, or virtually any other location where an optical signal can be measured using a transmission or reflection-mode optical configuration. In other embodiments, PTT can be calculated using at least one time-dependent waveform measured with an acoustic sensor, typically disposed on the patient's chest. Or it can be calculated using at least one time-dependent waveform measured using a pressure sensor, typically disposed on the patient's bicep, wrist, or finger. The pressure sensor can include, for example, a pressure transducer, piezoelectric sensor, actuator, polymer material, or inflatable cuff.

Specifically, in one aspect, the invention provides a method for monitoring a patient featuring the following steps: (a) associating a first set of vital sign information measured from the patient with a first transceiver that includes a first motion sensor; (b) storing the first set of vital sign information in a computer memory; (c) contacting the first transceiver with a second transceiver that includes a second motion sensor, the contacting causing the first motion sensor to generate a first motion signal and the second motion sensor to generate a second motion signal; (d) processing the first and second motion signals to determine that the first transceiver is to be replaced by the second transceiver; and (e) associating a second set of vital sign information with the patient, the second set of vital sign information measured with the second transceiver.

In embodiments, both the first and second motion sensors are accelerometers that generate time-dependent waveforms (e.g. ACC waveforms). Contacting the two transceivers typically generates waveforms that include individual 'pulses' (e.g. a sharp spike) caused by rapid acceleration and deceleration detected by the respective accelerometers. Typically the pulses are within waveforms generated along the same axes in both transceivers. The pulses can be collectively processed (using, e.g., an autocorrelation algorithm) to determine that they are generated during a common period of time. In embodiments, amplitudes of the first and second pulses are required to exceed a pre-determined threshold value in order for the second transceiver to replace the first transceiver. Pulses that meet this criterion are wirelessly transmitted to a remote server, where they are processed as described above. If the server determines that the second transceiver is ready to replace the first transceiver, it transmits instruction information to the transceivers to guide the replacement process. This instruction information, for example, is displayed by the GUIs of both transceivers. Once the replacement process is complete, vital sign information measured by the second transceiver is stored along with that measured by the first transceiver in a computer memory (e.g. a database) on the remote computer. The vital sign information can include conventional vital signs (e.g. HR, SYS, DIA, RR, and TEMP), along with the time-dependent waveforms used to calculate the vital signs (e.g. PPG, ECG, OSC, IP) and motion-related properties (ACC). Patient demographic information (e.g. name, gender, weight, height, date of birth) can also be associated with both the first and second sets of vital sign information.

In another aspect, the invention provides a method for pairing a patient monitor with a remote display device (e.g. an RVD) using a methodology similar to that described above. The display device is typically a portable display device (e.g. a personal digital assistant, or PDA), or a remote computer, such as a COW or central nursing station. The method includes the following steps: (a) contacting either a display device or an area proximal to the display device with the transceiver to generate a motion signal with its internal accelerometer; (b) transmitting the motion signal to a computer; (c) processing the motion signal with the computer to associate the transceiver with the display device; (d) measuring a set of vital sign information from the patient with the transceiver; and (e) displaying the set of vital sign information on the display device. Here, the act of contacting the display device with the transceiver generates a pulse in the ACC waveform, as described above. Processing done by the computer analyzes both the pulse and a location of the display device to associate it with the transceiver.

Several methods can be used to determine the location of the display device. For example, the wireless transmitter

within the transceiver is configured to operate on a wireless network, and algorithms operating on the remote computer and can analyze signals between the transceiver and wireless access points within the network (e.g. RSSI signals indicating signal strength) to determine an approximate location of the transceiver and thus the display device which it contacts. In embodiments the algorithms can involve, e.g., triangulating at least three RSSI values, or simply estimating location by determining the nearest access point from a single RSSI value. Triangulation typically involves using a map grid that includes known locations of multiple wireless access points and display devices within a region of the hospital; the map grid is determined beforehand and typically stored, e.g., in a database. For example, the approximate location of the transceiver can be determined using triangulation. Then the nearest display device, lying within a known location within a pre-determined radius, is paired with the transceiver. Typically the pre-determined radius is between 1-5 m.

In another aspect, the invention provides a body-worn monitor including first and second sensors attached to the patient, and a processing component that interfaces to both sensors and processes signals from them to calculate at least one vital sign value. A wireless transmitter receives the vital sign value and transmits it over a wireless interface, and additionally provides a two-way communications system configured to transmit and receive audio signals over the same wireless interface. In embodiments, the two-way communications system includes a speaker and a microphone, both of which are integrated into the transceiver. Typically the wireless interface is a hospital-based wireless network using an 802.11 protocol (e.g. 802.11a/b/g/n). A VOIP system typically runs on the wireless network to supply two-way voice communications. Alternatively the wireless network is based on a cellular protocol, such as a GSM or CDMA protocol.

Typically the body-worn monitor features a wrist-worn transceiver that functions as a processing component, and includes a touchpanel display configured to render both patient and clinician interfaces. The touchpanel display is typically a liquid crystal display (LCD) or organic light-emitting diode display (OLED) display with a clear touchpanel utilizing established resistive or capacitive technologies adhered to its front surface. The patient interface is typically rendered by default, and includes a graphical icon that, when initiated, activates the two-way communications system. The clinician interface typically requires a security code (entered using either a 'soft' numerical keypad or through a barcode scanner) to be activated. The transceiver typically includes a strap configured to wrap around the patient's arm, and most typically the wrist; this allows it to be worn like a conventional wristwatch, which is ideal for two-way communications between the patient and a clinician.

In a related aspect, the invention provides a wrist-worn transceiver wherein the two-way communications system described above, or a version thereof, is used as a voice annotation system. Such a system receives audio signals (typically from a clinician), digitizes them, and transmits the resulting digital audio signals, or a set of parameters determined from these signals, over the wireless interface to a computer memory. The audio signals are typically used to annotate vital sign information. They can be used, for example, to indicate when a pharmaceutical compound is administered to the patient, or when the patient undergoes a specific therapy. Typically the voice annotation uses the same speaker used for the two-way communication system. It also may include a speech-to-text converter that converts audio annotations from the clinician into text fields that can be easily stored alongside the vital sign information. In embodi-

ments, both a text field and the original audio annotation are stored in a computer memory (e.g. database), and can be edited once stored. In other embodiments, a pre-determined text field (indicating, e.g., that a specific medication is delivered at a time/date automatically determined by the transceiver) is used to annotate the vital sign information. In still other embodiments, a set of parameters determined from the digital audio signals can include an icon or a numerical value. Annotations in the database can be viewed afterwards using a GUI that renders both the vital sign information (shown, e.g., in a graphical form) and one or more of the annotations (e.g. icon, text field, numerical value, or voice annotation).

In another aspect, the invention provides a wrist-worn transceiver featuring a GUI that the patient can use to indicate their level of pain. Here, the GUI typically includes a touch-panel display configured to render a set of input fields, with each input field in the set indicating a different level of pain. Once contacted, the input fields generate a signal that is processed to determine the patient's level of pain. This signal can be further processed and then wirelessly transmitted to a remote computer for follow-on analysis.

In embodiments, the touchpanel display features a touch-sensitive area associated with each input field that generates a digital signal (e.g. a number) after being contacted. Each input field is typically a unique graphical icon such as a cartoon or numerical value indicating an escalating level of pain. The transceiver can also include a voice annotation system similar to that described above so the patient can specifically describe their pain (e.g. its location) using their own voice. This information can be wirelessly transmitted to a remote computer (e.g. a PDS) featuring a display device (e.g. an RVD). This system can render both vital sign information and a parameter determined from the pain signal, and can additionally include an alarming system that activates an alarm if the pain signal or a parameter calculated therefrom exceeds a pre-determined threshold.

In a related aspect, the invention provides a wrist-worn transceiver that includes a mentation sensor configured to collect data input characterizing the patient's level of mentation (e.g. mental acuity). This information, along with traditional vital signs and the waveforms they are calculated from, is wirelessly transmitted to a remote computer for analysis. In embodiments, the mentation sensor is a touchpanel display that renders a GUI to collect information characterizing the patient's level of mentation. For example, the GUI can render a series of icons, a game, test, or any other graphical or numerical construct that can be used to evaluate mentation. In a specific embodiment, for example, the GUI includes a set of input fields associated with a numerical value. Here, the mentation 'test' features an algorithm to determine if the input fields are contacted by the patient in a pre-determined numerical order. Upon completion, the test results can be evaluated to generate a mentation 'score'. In this aspect, the wrist-worn transceiver also includes a two-way communication system that receives audio information from the patient. This audio information can be used for conventional communication purposes, and can additionally be analyzed to further gauge mentation. As in previous embodiments, the mentation score can be sent with vital sign information to a PDS/RVD for follow-on analysis. These systems may include an alarming system that generates an alarm if the mentation parameter or a parameter calculated therefrom exceeds a pre-determined threshold.

In another aspect, the invention provides a wrist-worn transceiver featuring a motion sensor (e.g. an accelerometer, mercury switch, or tilt switch) that generates a motion signal indicating the transceiver's orientation. The processing com-

ponent within the transceiver processes the motion signal and, in response, orients the GUI so that it can be easily viewed in 'rightside up' configuration, i.e. with text rendered in a conventional manner from left to right. If the transceiver is moved (e.g., so that it is viewed by a clinician instead of a patient), the accelerometers generate new motion signals, and the GUI is 'flipped' accordingly. Typically, for example, the GUI is rendered in either a first orientation or a second orientation, with the two orientations separated by 180 degs., and in some cases by 90 degs. In embodiments, the first orientation corresponds to a 'patient GUI', and the second orientation corresponds to a 'clinician GUI'. This allows, for example, the appropriate GUI to be automatically rendered depending on the transceiver's orientation. The clinician GUI typically includes medical parameters, such as vital signs and waveforms, whereas the patient GUI typically includes non-medical features, such as a 'nurse call button', time/date, and other components described in more detail below.

In preferred embodiments, the motion sensor is a 3-axis accelerometer configured to generate a time-domain ACC waveform. During a measurement, the processing component additionally analyzes the waveform to determine parameters such as the patient's motion, posture, arm height, and degree of motion.

In another aspect of the invention, the wrist-worn transceiver features a display device configured to render at least two GUIs, with the first GUI featuring medical content, and the second GUI featuring non-medical content relating to entertainment, food service, games, and photographs. The photograph, for example, can include an image of the patient or a relative of the patient; this latter case may be particularly useful in neo-natal hospital wards. To capture the photograph, the body-worn monitor may include a digital camera, or a wireless interface to a remote digital camera, such as that included in a portable computer or cellular telephone.

In other embodiments, the second GUI is configured to render menus describing entertainment content, such as television (e.g. different channels or pre-recorded content), movies, music, books, and video games. In this case, the touch-panel display can be used to select the content or, in embodiments, play a specific game. The wireless transmitter within the transceiver is further configured to transmit and receive information from a remote server configured to store digital representations of these media sources. In still other embodiments, the second GUI is configured to display content relating to a food-service menu. Here, the wireless transmitter is further configured to transmit and receive information from a remote server configured to interface with a food-service system.

In another aspect, the invention provides a system for monitoring a patient that includes a vital sign monitor configured to be worn on the patient's body, and a remote computer. The vital sign monitor features connection means (e.g. a flexible strap or belt) configured to attach a transceiver to the patient's body, and sensor with a sensing portion (e.g. electrodes and an optical sensor) that attaches to the patient to measure vital sign information. A mechanical housing included in the transceiver covers a wireless decoder, processing component, and wireless transmitter, and supports a display component. The wireless decoder (e.g. a barcode scanner or radio frequency identification (RFID) sensor) is configured to detect information describing a medication, a medication-delivery rate, a clinician, and the patient. For example, this information may be encoded in a barcode or RFID tag located on the patient, clinician, medication, or associated with an infusion pump. The processing component is configured to process: 1) the vital sign information to

generate a vital sign and a time-dependent waveform; and 2) information received by the wireless decoder to generate decoded information. The wireless transmitter within the mechanical housing receives information from the processing component, and transmits it to a remote computer. In response the remote computer processes the information and transmits an information-containing packet back to the vital sign monitor.

In embodiments, the remote computer performs an analyzing step that compares information describing both the medication and the patient to database information within a database. The database may include, for example, a list of acceptable medications and acceptable medication-delivery rates corresponding to the patient. In some cases both the vital sign information and the decoded information are collectively analyzed and compared to values in the database to affect treatment of the patient. For example, this analysis may determine that a patient with a low blood pressure should not receive medications that further lower their blood pressure. Or it may suggest changing a dosage level of the medication in order to compensate for a high heart rate value. In general, the remote computer can analyze one or more vital sign values corresponding to a patient, along with the patient's demographic information, medical history, and medications, and determine acceptable medications and medication-delivery rates based on this analysis. In response, the computer can transmit a packet back to the vital sign monitor, which renders its contents on the display. The packet can include a message confirming that a particular medication and medication-delivery rate are acceptable for the patient, and may also include a set of instructions for delivering the medication and performing other therapies.

Still other embodiments are found in the following detailed description of the invention, and in the claims.

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 is a schematic drawing showing the wrist-worn transceiver of the invention attached to a patient's wrist;

FIGS. 2A and 2B show, respectively, schematic drawings of the wrist-worn transceiver of FIG. 1 oriented 'rightside up' so that a patient can view the GUI, and 'upside down' so that a clinician can view the GUI;

FIG. 3 shows a schematic drawing of the wrist-worn transceiver of FIG. 1 and a list of features available in both a patient GUI and a clinician GUI;

FIG. 4 shows a schematic drawing of the body-worn monitor featuring sensors for measuring ECG, PPG, ACC, OSC, and IP waveforms, and systems for processing these to determine a patient's vital signs;

FIG. 5 shows a schematic drawing of an IT configuration of the invention where the body-worn monitor of FIG. 4 is connected through a wireless network to a PDS and hospital EMR;

FIGS. 6A and 6B show, respectively, schematic drawings of a new transceiver having a fully charged battery being swapped with an original transceiver having a depleted battery before and after deploying the 'bump' methodology;

FIG. 7 shows a schematic drawing of transceivers undergoing the 'bump' methodology of FIGS. 6A and 6B and wirelessly transmitting their ACC waveforms to the PDS for analysis;

FIG. 8 shows screen captures from a GUI used to guide a clinician through the 'bump' methodology of FIGS. 6A, 6B, and 7;

FIG. 9 shows a schematic drawing of a transceiver being 'bumped' against a RVD in order to pair the two devices;

FIG. 10 shows a map indicating how the transceiver and RVD of FIG. 9 are paired to each other;

FIG. 11 shows a schematic drawing of the wrist-worn transceiver of FIG. 1 being used for voice annotation of a patient's vital sign data;

FIG. 12 shows a schematic drawing of the wrist-worn transceiver of FIG. 11 wirelessly transmitting voice annotations to the PDS for analysis;

FIG. 13 shows screen captures from a GUI used to guide a clinician through the voice annotation methodology of FIGS. 11 and 12;

FIG. 14 shows screen captures from a GUI used when the wrist-worn transceiver functions as a two-way communicator between the patient and a clinician;

FIG. 15 shows a screen capture from a GUI used to render a 'pain index' on the wrist-worn transceiver;

FIG. 16 shows a screen capture from a GUI used to render a mentation test on the wrist-worn transceiver;

FIG. 17 shows a screen capture from a GUI used to render a photograph of the patient on the wrist-worn transceiver;

FIG. 18 shows a screen capture from a GUI used to render a food menu on the wrist-worn transceiver;

FIG. 19 shows a screen capture from a GUI used to render a menu of television channels on the wrist-worn transceiver;

FIG. 20 shows a schematic drawing of the barcode scanner in the wrist-worn transceiver scanning barcodes associated with a patient, clinician, and medication, and sending the decoded barcode information to the PDS;

FIGS. 21A and 21B show three-dimensional images of the body-worn monitor of FIG. 4 attached to a patient with and without, respectively, a cuff-based pneumatic system used for a calibrating indexing measurement;

FIGS. 22A and 22B show, respectively, three-dimensional images of the wrist-worn transceiver before and after receiving cables from other sensors within the body-worn monitor;

FIG. 23A shows a schematic drawing of a patient wearing the body-worn monitor of FIG. 21B and its associated sensors;

FIG. 23B shows graphs of time-dependent ECG, PPG, OSC, ACC, and IP waveforms generated with the body-worn monitor and sensors of FIG. 23A;

FIG. 24 shows screen captures from a GUI used to render vital signs and ECG, PPG, and IP waveforms on the wrist-worn transceiver;

FIG. 25 shows a schematic drawing of the ACC, ECG, pneumatic, and auxiliary systems of the body-worn monitor communicating over the CAN protocol with the wrist-worn transceiver;

FIG. 26 shows an alternate IT configuration of the invention where the wrist-worn transceiver of FIG. 1 communicates with the PDS through a wireless access point connected to the Internet;

FIG. 27 shows an alternate IT configuration of the invention where the wrist-worn transceiver of FIG. 1 communicates with the PDS through a wireless device connected to the Internet; and

FIG. 28 shows an alternate IT configuration of the invention where the wrist-worn transceiver of FIG. 1 communicates with the PDS through an internal cellular modem connected to the Internet.

DETAILED DESCRIPTION OF THE INVENTION

System Overview

FIG. 1 shows a transceiver 72 according to the invention that attaches to a patient's wrist 66 using a flexible strap 90. The transceiver 72 connects through a first flexible cable 92 to

a thumb-worn optical sensor **94**, and through a second flexible cable **82** to an ECG circuit and a series of chest-worn electrodes (not shown in the figure). During a measurement, the optical sensor **94** and chest-worn electrodes measure, respectively, time-dependent optical waveforms (e.g. PPG) and electrical waveforms (e.g. ECG and IP), which are processed as described below to determine vital signs and other physiological parameters such as cNIBP, SpO₂, HR, RR, TEMP, pulse rate (PR), and cardiac output (CO). Once measured, the transceiver **72** wirelessly transmits these and other information to a remote PDS and RVD. The transceiver **72** includes a touchpanel display that renders a GUI **50** which, in turn, displays the vital signs, physiological parameters, and a variety of other features described in detail below. Collectively, the transceiver **72** and GUI **50** incorporate many features that are normally reserved for non-medical applications into a body-worn vital sign monitor that continuously monitors ambulatory patients as they move throughout the hospital.

The transceiver **72** includes an embedded accelerometer that senses its motion and position, and in response can affect properties of the GUI. Referring to FIGS. 2A and 2B, for example, time-resolved ACC waveforms from the accelerometer can be processed with a microprocessor within the transceiver to detect orientation of the touchpanel display. This information can then be analyzed to determine if it is the clinician or patient who is viewing the display. In response, the GUI can 'flip' so that it is properly oriented (i.e. 'rightside up', as opposed to being upside down) for the viewer. For example, as shown in FIG. 2A, when the transceiver **72** is worn on the patient's right wrist **66** the internal accelerometer generates ACC waveforms that are processed by the microprocessor to determine this orientation. The GUI **50A** is adjusted according so that it is always oriented with numbers and text arranged rightside up and read from left to right. When the patient's arm is rotated, as shown in FIG. 2B, the ACC waveforms change accordingly because the accelerometer's axes are swapped with respect to gravity. Such a situation would occur, for example, if a clinician were to orient the patient's arm in order read the transceiver's display. In this case, the ACC waveforms are processed to determine the new orientation, and the GUI **50B** is flipped so it is again rightside up, and can be easily read by the clinician.

The internal accelerometer can also detect if the transceiver is 'bumped' by an external object. In this case, the ACC waveform will feature a sharp 'spike' generated by rapid acceleration and deceleration caused by the bumping process. As described in detail below, such a bumping process can serve as a fiducial marker that initiates a specific event related to the transceiver, such as a battery swap or process that involves pairing the transceiver to an external wireless system or display.

The accelerometer within the transceiver, when combined with other accelerometers within the body-worn monitor, can also be used to determine the patient's posture, activity level, arm height and degree of motion, as described in detail below. Use of one or more accelerometers to detect such motion-related activities is described, for example, in the following patent applications, the contents of which are incorporated herein by reference: BODY-WORN MONITOR FEATURING ALARM SYSTEM THAT PROCESSES A PATIENT'S MOTION AND VITAL SIGNS (U.S. Ser. No. 12/469,182; filed May 20, 2009) and BODY-WORN VITAL SIGN MONITOR WITH SYSTEM FOR DETECTING AND ANALYZING MOTION (U.S. Ser. No. 12/469,094; filed May 20, 2009).

Referring to FIG. 3, in addition to the GUI **50**, the transceiver **72** includes a high-fidelity speaker **120**, a microphone **101**, and a barcode scanner **102** which, respectively, enunciates audible information, measures voice signals from both the patient and a clinician, and scans graphical barcodes to decode numerical information describing the patient and their medication. Signals from these and other components are processed to supply information to either a 'patient GUI' **52** or a 'clinician GUI' **54**. The patient GUI **52**, for example, typically includes features that are decoupled from a standard clinical diagnosis; these include a nurse call button, voice communications, a 'pain' index, a mentation test to estimate the patient's cognitive abilities, meal ordering within the hospital, games, and a controller for entertainment content, e.g. to adjust parameters (e.g. channels, volume) for a standard television set. The clinical GUI **54**, in comparison, includes features that are used for clinical diagnoses and for operating the transceiver in a hospital environment. The primary features of this GUI **54** include displaying vital signs (e.g. cNIBP, SpO₂, HR, RR, TEMP), other medical parameters (e.g. PR, CO), and waveforms (PPG, ECG, IP). Secondary features of the clinical GUI include voice communications, battery-change and pairing operations using the above-described 'bump' methodology, voice annotation of medical records and diagnoses, a method for checking medications using the barcode scanner **102**, and display of a photograph or video describing the patient.

During normal operation, the GUI renders **50** simple icons indicating that the transceiver is powered on and operational (e.g., a 'beating heart'), the strength of the wireless signal (e.g. a series of bars with escalating height), and the battery level (e.g. a cartoon of a battery with a charge-dependent gauge). The transceiver **72** displays these icons until the touchpanel display is contacted by either the patient or a clinician. This process yields the patient GUI **52**, which features a large icon **57** showing a telephone (which is used for nurse call applications, as described below), and a smaller icon **53** showing a lock which, when tapped, enables the clinician to 'unlock' the transceiver and utilize the clinician interface **54**. The transceiver **72** immediately renders a GUI that shows vital signs and waveform information if the patient's physiological condition requires immediate medical attention, e.g. in the case of cardiac arrest.

The clinician interface **54** is password-protected to prevent the patient or any other non-clinician from viewing important and potentially confusing medical information. A password can either be entered as a standard personal identification number (PIN) by tapping keys on a numerical keypad (as shown in FIG. 3), or by simply swiping a barcode printed on the clinician's hospital badge across the barcode scanner **102**. The microprocessor within the transceiver unlocks the clinician interface following either of these events, and enables all the features associated with the interface, which are described in detail below. For example, with this interface the clinician can view vital signs and waveforms to make a medical diagnosis, as described with reference to FIG. 24. If the transceiver's battery charge is running low, the clinician can swap in a new transceiver and transfer data from the original transceiver simply by 'bumping' the two transceivers together, as described with reference to FIGS. 6-8. Medical records can be voice-annotated and stored on the PDS or a hospital's EMR using the process shown in FIGS. 11-13. The patient's medication can be checked by scanning and processing information encoded in barcodes associated with the patient, clinician, and medication, as shown in FIG. 20. All of this functionality is programmed within the transceiver and the body-worn monitor, and can be accomplished without teth-

ering the patient to a conventional vital sign monitor typically mounted on a wall in the hospital or a rolling stand. Ultimately this allows the patient to wear a single body-worn monitor as they transition throughout the various facilities within the hospital, e.g. the ED, ICU, x-ray facility, and operating room.

Hardware in Body-Worn Monitor

FIGS. 4 and 5 show schematic drawings of a body-worn monitor **100** used to measure vital signs from a patient and render the different GUIs described above (FIG. 4), along with a wireless system over which the transceiver **72** sends information through a hospital network **60** to either a remote RVD, e.g. a computer **62** or hand-held device **64** (FIG. 5). Referring to FIG. 4, the body-worn monitor **100** features a wrist-worn transceiver **72** that continuously determines vital signs and motion-related properties from an ambulatory patient in a hospital. The monitor **100** is small, lightweight, and comfortably worn on the patient's body during their stay in the hospital; its specific form factor is described in detail below with reference to FIGS. 21 and 22. It provides continuous monitoring, and features a software framework that determines alarms/alerts if the patient begins to decompensate. Such systems are described in the following co-pending patent applications, the contents of which have been previously incorporated herein by reference: BODY-WORN MONITOR FEATURING ALARM SYSTEM THAT PROCESSES A PATIENT'S MOTION AND VITAL SIGNS (U.S. Ser. No. 12/469,182; filed May 20, 2009) and BODY-WORN VITAL SIGN MONITOR WITH SYSTEM FOR DETECTING AND ANALYZING MOTION (U.S. Ser. No. 12/469,094; filed May 20, 2009). The framework processes both the patient's motion and their vital sign information with algorithms that reduce the occurrence of false alarms.

A combination of features makes the body-worn monitor **100** ideal for ambulatory patients within the hospital. For example, as shown in FIG. 5, the transceiver **72** features a wireless transmitter **224** that communicates through a collection of wireless access points **56** (e.g. routers based on 802.11 protocols) within a hospital network **60**, which includes a PDS. From the PDS **60** data are sent to an RVD (e.g. a portable tablet computer **62**) located at a central nursing station, or to a local computer (e.g. a hand-held PDA **64**) carried by the clinician. In embodiments, data can be sent to the PDA **64** through a peer-to-peer wireless connection. The specific mode of communication can be determined automatically (using, e.g., a signal strength associated with the wireless connection), or manually through an icon on the GUI.

The transceiver **72** features a CPU **222** that communicates through a digital CAN interface, or bus, to external systems featuring ECG **216**, external accelerometers **215b-c**, pneumatic **220**, and auxiliary **245** sensors. Each sensor **215b-c**, **216**, **220**, **245** is 'distributed' on the patient to minimize the bulk and weight normally associated with conventional vital sign monitors, which typically incorporate all electronics associated with measuring vital signs in a single plastic box. Moreover, each of these sensors **215b-c**, **216**, **220**, **245** generate digital signals close to where they actually attach to the patient, as opposed to generating an analog signal and sending it through a relatively long cable to a central unit for processing. This can reduce noise due to cable motion which is often mapped onto analog signals. Cables **240**, **238**, **246** used in the body-worn monitor **210** to transmit packets over the CAN bus typically include five separate wires bundled together with a single protective cladding: the wires supply power and ground to the remote ECG system **216**, accelerometers **215b-c**, pneumatic **220**, and auxiliary systems **245**; provide high/low signal transmission lines for data transmit-

ted over the CAN protocol; and provide a grounded electrical shield for each of these four wires. There are several advantages to this approach. First, a single pair of transmission lines in the cable (i.e. the high/low signal transmission lines) can transmit multiple digital waveforms generated by completely different sensors. This includes multiple ECG waveforms (corresponding, e.g., to vectors associated with three, five, and twelve-lead ECG systems) from the ECG circuit, along with ACC waveforms associated with the x, y, and z axes of accelerometers within the body-worn monitor **100**. The same two wires, for example, can transmit up to twelve ECG waveforms (measured by a twelve-lead ECG system), and six ACC waveforms (measured by the accelerometers **215b-c**). Limiting the transmission line to a pair of conductors reduces the number of wires attached to the patient, thereby decreasing the weight and any cable-related clutter. Second, cable motion induced by an ambulatory patient can change the electrical properties (e.g. electrical impedance) of its internal wires. This, in turn, can add noise to an analog signal and ultimately the vital sign calculated from it. A digital signal, in contrast, is relatively immune to such motion-induced artifacts.

The ECG **216**, pneumatic **220**, and auxiliary **245** systems are stand-alone systems that each includes a separate CPU, analog-to-digital converter, and CAN transceiver. During a measurement, they connect to the transceiver **72** through cables **240**, **238**, **246** and connectors **230**, **228**, **232** to supply digital inputs over the CAN bus. The ECG system **216**, for example, is completely embedded in a terminal portion of its associated cable. Systems for three, five, and twelve-lead ECG monitoring can be swapped in an out simply by plugging the appropriate cable (which includes the ECG system **216**) into a CAN connector **230** on the wrist-worn transceiver **72**, and the attaching associated electrodes to the patient's body.

As described above, the transceiver **72** renders separate GUIs that can be selected for either the patient or a clinician. To do this, it includes a barcode scanner **242** that can scan a barcode printed, e.g., on the clinician's badge. In response it renders a GUI featuring information (e.g. vital signs, waveforms) tailored for a clinician that may not be suitable to the patient. So that the patient can communicate with the clinician, the transceiver **72** includes a speaker **241** and microphone **237** interfaced to the CPU **222** and wireless system **224**. These components allow the patient to communicate with a remote clinician using a standard VOIP protocol. A rechargeable Li-ion battery **239** powers the transceiver **72** for about four days on a single charge. When the battery charge runs low, the entire transceiver **72** is replaced using the 'bump' technique described in detail below.

Three separate digital accelerometers **215a-c** are non-obtrusively integrated into the monitor's form factor; two of them **215b-c** are located on the patient's body, separate from the wrist-worn transceiver **72**, and send digitized, motion-related information through the CAN bus to the CPU **222**. The first accelerometer **215a** is mounted on a circuit board within the transceiver **72**, and monitors motion of the patient's wrist. The second accelerometer **215b** is incorporated directly into the cable **240** connecting the ECG system **216** to the transceiver **72** so that it can easily attach to the patient's bicep and measure motion and position of the patient's upper arm. As described below, this can be used to orient the screen for viewing by either the patient or clinician. Additionally, signals from the accelerometers can be processed to compensate for hydrostatic forces associated with changes in the patient's arm height that affect the monitor's cNIBP measurement, and can be additionally used to calibrate the monitor's blood

pressure measurement through the patient's 'natural' motion. The third accelerometer **215c** is typically mounted to a circuit board that supports the ECG system **216** on the terminal end of the cable, and typically attaches to the patient's chest. Motion and position of the patient's chest can be used to determine their posture and activity states, which as described below can be used with vital signs for generating alarm/alerts. Each accelerometer **215a-c** measures three unique ACC waveforms, each corresponding to a separate axis (x, y, or z) representing a different component of the patient's motion. To determine posture, arm height, activity level, and degree of motion, the transceiver's CPU **222** processes signals from each accelerometer **215a-c** with a series of algorithms, described in the following pending patent applications, the contents of which have been previously incorporated herein by reference: BODY-WORN MONITOR FEATURING ALARM SYSTEM THAT PROCESSES A PATIENT'S MOTION AND VITAL SIGNS (U.S. Ser. No. 12/469,182; filed May 20, 2009) and BODY-WORN VITAL SIGN MONITOR WITH SYSTEM FOR DETECTING AND ANALYZING MOTION (U.S. Ser. No. 12/469,094; filed May 20, 2009). In total, the CPU **222** can process nine unique, time-dependent signals corresponding to the three axes measured by the three separate accelerometers. Algorithms determine parameters such as the patient's posture (e.g., sitting, standing, walking, resting, convulsing, falling), the degree of motion, the specific orientation of the patient's arm and how this affects vital signs (particularly cNIBP), and whether or not time-dependent signals measured by the ECG **216**, optical **218**, or pneumatic **220** systems are corrupted by motion.

To determine blood pressure, the transceiver **72** processes ECG and PPG waveforms using a measurement called with Composite Technique, which is described in the following patent application, the contents of which have been previously incorporated herein by reference: BODY-WORN SYSTEM FOR MEASURING CONTINUOUS NON-INVASIVE BLOOD PRESSURE (cNIBP) (U.S. Ser. No. 12/650,354; filed Nov. 15, 2009). The Composite Technique measures ECG and PPG waveforms with, respectively, the ECG **216** and optical **218** systems. The optical system **218** features a thumb-worn sensor that includes LEDs operating in the red (λ ~660 nm) and infrared (λ ~900 nm) spectral regions, and a photodetector that detects their radiation after it passes through arteries within the patient's thumb. The ECG waveform, as described above, is digitized and sent over the CAN interface to the wrist-worn transceiver **72**, while the PPG waveform is transmitted in an analog form and digitized by an analog-to-digital converter within the transceiver's circuit board. The pneumatic system **220** provides a digitized pressure waveform and oscillometric blood pressure measurements through the CAN interface; these are processed by the CPU **222** to make cuff-based 'indexing' blood pressure measurements according to the Composite Technique. The indexing measurement typically only takes about 40-60 seconds, after which the pneumatic system **220** is unplugged from its connector **228** so that the patient can move within the hospital without wearing an uncomfortable cuff-based system. The optical waveforms measured with the red and infrared wavelengths can additionally be processed to determine SpO₂ values, as described in detail in the following patent application, the contents of which is incorporated herein by reference: BODY-WORN PULSE OXIMETER (U.S. Ser. No. 12/559,379; filed Sep. 14, 2009).

Collectively, these systems **215a-c**, **216**, **218**, and **220** continuously measure the patient's vital signs and motion, and supply information to the software framework that calculates alarms/alerts. A third connector **232** also supports the CAN

bus and is used for auxiliary medical devices **245** (e.g. a glucometer, infusion pump, system for measuring end-tidal CO₂) that is either worn by the patient or present in their hospital room.

Once a measurement is complete, the transceiver **72** uses the internal wireless transmitter **224** to send information in a series of packets to a PDS **60** within the hospital. The wireless transmitter **224** typically operates on a protocol based on 802.11, and can communicate with the PDS **60** through an existing network within the hospital as described above with reference to FIG. 5. Information transmitted by the transceiver alerts the clinician if the patient begins to decompensate. The PDS **60** typically generates this alarm/alert once it receives the patient's vital signs, motion parameters, ECG, PPG, and ACC waveforms, and information describing their posture, and compares these parameters to preprogrammed threshold values. As described in detail below, this information, particularly vital signs and motion parameters, is closely coupled together. Alarm conditions corresponding to mobile and stationary patients are typically different, as motion can corrupt the accuracy of vital signs (e.g., by adding noise), and induce artificial changes in them (e.g., through acceleration of the patient's heart and respiratory rates) that may not be representative of the patient's actual physiology.

Swapping and Pairing Transceivers Using 'Bump' Methodology

FIGS. 6A, 6B, 7, and 8 show how a wrist-worn transceiver **72A** with a depleted battery can be swapped with a similar transceiver **72B** having a fully charged battery using the 'bump' methodology described above. Prior to the swap, as shown in FIG. 6A, both transceivers are readied by activating the appropriate GUI **50C**, **50D** following the screens shown in FIG. 8. This process activates firmware on each transceiver **72A**, **72B** indicating that the swap is about to occur. In response, each transceiver sends a packet through the wireless access point **56** and to the hospital network and PDS **60**. The packet describes a transceiver-specific address, e.g. a MAC address associated with its wireless transmitter. Once this is done, the GUIs **50C**, **50D** on both transceivers **72A**, **72B** indicate to a clinician that they can be 'bumped' together, and that the swap can proceed.

At this point, as shown in FIG. 6A, the new transceiver **72B** (with the fully charged battery) is then bumped against the old transceiver **72A** (with the depleted battery). Internal accelerometers within both transceivers **72A**, **B** detect the bumping process and, in response, independently generate ACC waveforms **130**, **132**, both featuring a sharp spike indicating the rapid acceleration and deceleration due to the bumping process. Typically the ACC waveforms **130**, **132** correspond to the same axes in both transceivers. The ACC waveforms are digitized within each transceiver and then transmitted through the wireless access point **56** to the PDS **60**, where they are stored in a computer memory and analyzed with a software program that is activated when both devices are 'readied', as described above. The software program compares formatted versions of the ACC waveforms **130'**, **132'** to detect the rapid spikes, as shown by the graph **140** in FIG. 7. The rapid spikes in the waveforms **130'**, **132'** should occur within a few microseconds of each other, as indicated by the shaded window **142** in the graph **140**. Other transceivers operating on the network may generate similar motion-related spikes due to movements of the patient wearing them, but the probability that such spikes occur at the exact same time as the transceivers being swapped is extremely low. The software program interprets the concurrence of the spikes as indicating that data stored on the old transceiver **72A** is to be transferred to the new transceiver **72B**. The data, for example,

includes demographic information describing the patient (e.g. their name, age, height, weight, photograph), the medications they are taking, and all the vital sign and waveform information stored in memory in the old transceiver **72A**. Following the bump, this information is associated with the address corresponding to the new transceiver **72B**. At this time information may also be sent from the PDS so it can be stored locally on the new transceiver. When all the relevant information is transferred over, the GUIs **50C**, **50D** on both transceivers **72A**, **72B** indicate that they can be swapped. At this point, cables connected to the optical sensor and ECG electrodes are unplugged from the old transceiver **72A**, and plugged into the new transceiver **72B**. The clinician then attaches the new transceiver to the patient's wrist, and commences measuring vital signs from the patient as described above.

In other embodiments, a time period corresponding to a portion (e.g. a peak value) of the motion-generated spike is determined on each of the wrist-worn transceivers that are bumped together. Each transceiver then sends its time period to the PDS, where they are collectively analyzed to determine if they are sufficiently close in value (e.g. within a few hundred milliseconds). If this criterion is met, software on the PDS assumes that the transceivers are ready to be swapped, and performs the above-described steps to complete this process.

FIG. 8 shows a sequence of screens within the GUI that describe the process for swapping transceivers to the clinician. The process begins when a screen **158** rendered by Device A indicates that its battery is running low of charge. This is indicated by a standard 'low battery' icon located in the upper right-hand corner of the screen **158**, as well as a larger icon located near the bottom of the screen. A time describing the remaining life of the battery appears near this icon when this time is 5 minutes or less. Each transceiver includes a sealed internal Li-ion battery that cannot be easily replaced in the hospital. Instead, the transceiver is inserted in a battery charger that typically includes eight or sixteen ports, each of which charges a separate transceiver. To swap Device A with Device B, the clinician taps the screen **158** to yield a new screen **160** which includes a series of six icons, each related to a unique feature. The icon in the lower left-hand corner shows two interchanging batteries. When tapped, this icon yields a new screen **160** indicating that Device A is ready to be swapped. Device B is then removed from a port in the battery charger, and a sequence of screens **150**, **152**, **154** are initiated as described with reference to Device A.

When Devices A and B both show, respectively, screens **162**, **154**, they are ready to be swapped using the 'bumping' process. At this point, as described above, a clinician 'bumps' Device B into Device A, which in turn generates two ACC waveforms **130**, **132** featuring sharp, time-dependent spikes indicating the bump. The waveforms **130**, **132** include spikes, as shown by the shaded box **142**, which are concurrent in time, and are wirelessly transmitted in a packet that indicates their origin through the pathway shown in FIG. 7 to the PDS. There, they are analyzed by the software program described above to determine that data associated with Device A (e.g. patient information, vital signs) is now associated with Device B. When this association is complete, the PDS transmits a packet back through the pathway shown in FIG. 7 to both Device A and B, indicating that the PDS is ready to transfer the data. Device B then renders a screen **166** asking the clinician to confirm the process. Data is transferred if the clinician taps the 'check' box in the lower right-hand corner of the screen; during this process Device B renders a screen **168** that shows the patient's name to further confirm with the

clinician that the transfer process is valid. When it is complete, Device A is no longer active, meaning it cannot collect data or generate alarms. Device B renders a screen **170** that instructs the clinician to disconnect the optical and electrical sensors from Device A, and to clean this device and insert it into the battery charger. During this process all alarms are paused for Device B. A screen **172** on Device B then instructs the clinician to connect the sensors and attach Device B to the patient's wrist. When this is complete, Device B renders a final confirmatory screen **176**, which when checked finalizes the swapping process. At this point Device B is officially associated with the patient, renders a standard screen **178**, and commences measuring vital signs from the patient. These vital signs, along with those collected from Device A, are included in a contiguous data file characterizing the patient.

As an alternative to the 'bumping' process, Device B's barcode can be read and processed to facilitate swapping the transceivers. In this case, an icon on Device A, when tapped, renders a screen **164** indicating that Device A is ready to read the barcode printed on Device B. At this point, Device B's barcode is swiped across Device A's barcode reader, decoded, and wirelessly transmitted to the PDS as indicated in FIG. 7. The PDS uses this information to associate Device B with the patient as described above. Once this is complete, Device B uses the same screens used for the 'bumping' transfer process (screens **166**, **168**, **170**, **172**, **176**, **178**) to associate Device B with the patient. The 'bumping' process shown in FIG. 6 takes place along the long axes of Device A and Device B. Alternatively, it can take place along the short axes of these devices. Or the short axis of one device can be bumped against the long axis of the other device to initiate the process.

The 'bumping' process described above can also be used for other applications relating to the wrist-worn transceiver. It can be used, for example, to pair the transceiver with an RVD, such as a display located at the patient's bedside, or at a central nursing station. In this embodiment, indicated in FIGS. 9 and 10, a clinician selects a transceiver **72** from the battery charger and brings it near an RVD **62**. Before attaching the transceiver **72** to the patient, the clinician 'bumps' it against a hard surface proximal to the RVD **62** (or against the RVD itself) to generate a sharp spike in the ACC waveform **133**. The waveform **133** is similar in shape to that generated when two transceivers are swapped with the bumping process, as described above. The RVD's location needs to be determined in order to pair it with the transceiver **72**. To do this, at a pre-determined time period (e.g. every few minutes) all neighboring wireless access points **56A**, **56B**, **56C** transmit a 'location beacon' **59A**, **59B**, **59C** to the transceiver, which is received and used to calculate a value for signal strength (typically characterized by an 'RSSI value') between the transceiver **72** and the respective access point **56A**, **56B**, **56C**. The transceiver concatenates values for RSSI and identifiers for the access points into a single 'location packet' **59D**, which it then transmits along with the ACC waveform **133** and an identifying code describing the transceiver (not shown in the figure) through a single access point **56B** to the PDS **60**. The PDS **60** receives the location packet **59D** and parses it to arrive at RSSI values for the three wireless access points **56A**, **56B**, **56C** within wireless range of the transceiver **72**. In other embodiments, the individual access points **56A**, **56B**, **56C** determine RSSI values characterizing the signal strength between them and the transceiver, and send these as individual packets to the PDS. Software on the PDS then concatenates these packets to determine signals similar to those included in the location packet.

Referring to FIG. 10, location-determining software operating on the PDS triangulates the signals, along with known

locations of each wireless access point **56A**, **56B**, **56C**, to determine an approximate location **71** of the transceiver **72**. The known locations of the access points are stored within a map grid **73** in a computer memory associated with the location-determining software. The transceiver's approximate location typically has an accuracy of 1-3 m. Using the map grid **73**, the software then processes the approximate location **71** and a known location of any RVD **62** lying within a pre-determined radius **75**. Typically the pre-determined radius is 3-5 m. If the location of the RVD **62** lies within the pre-determined radius **75**, the RVD **62** is automatically 'paired' with the transceiver **72**. Once paired, the RVD **62** then displays any follow-on waveform, motion, and vital sign information sent by the transceiver.

In related embodiments, the location-determining software described above uses triangulation algorithms to determine the patient's current and historical location. Such a process can be used to monitor and locate a patient in distress, and is described, for example, in the following issued patent, the contents of which are incorporated herein by reference: WIRELESS, INTERNET-BASED, MEDICAL DIAGNOSTIC SYSTEM (U.S. Pat. No. 7,396,330). If triangulation is not possible, the location-determining software may simply use proximity to a wireless access point (as determined from the strength of an RSSI value) to estimate the patient's location. Such a situation would occur if signals from at least three wireless access points were not available. In this case, the location of the patient is estimated with an accuracy of about 5-10 m. In embodiments, the RVD may be a central nursing station that displays vital sign, motion-related properties (e.g. posture and activity level) and location information from a group of patients. Such embodiments are described in the following co-pending patent application, the contents of which are fully incorporated herein by reference: BODY-WORN VITAL SIGN MONITOR (U.S. Ser. No. 12/560,077, filed Sep. 15, 2009). In other embodiments, the location-determining software determines the location of a patient-worn transceiver, and automatically pairs it to a RVD located nearby (e.g. within a pre-determined radius, such as that shown in FIG. 10). In this way, the patient's information can be displayed on different RVDs as they roam throughout the hospital.

In embodiments, the patient's location can be analyzed relative to a set of pre-determined boundaries (e.g. a 'geofence') to determine if they have wandered into a restricted area. Or their speed can be determined from their time-dependent location, and then analyzed relative to a pre-determined parameter to determine if they are walking too fast. In general, any combination of location, motion-related properties, vital signs, and waveforms can be collectively analyzed with software operating on either the transceiver or PDS to monitor the patient. Patients can be monitored, for example, in a hospital, medical clinic, outpatient facility, or the patient's home.

In the embodiments described above, location of the transceiver can be determined using off-the-shelf software packages that operate on the PDS. Companies that provide such software include, for example, by Cisco Systems (170 West Tasman Drive, San Jose, Calif. 93134; www.cisco.com), Ekahau (12930 Saratoga Avenue, Suite B-8, Saratoga, Calif. 95070; www.ekahau.com), and others.

In still other embodiments, software operating on the transceiver puts it into a 'sleep mode' when it is not attached to the patient. This way the transceiver can determine and transmit a location packet even when it is not used for patient monitoring. Using the above-described location-determining software, this allows the transceiver's location to be determined

and then analyzed if it has been lost, misplaced, or stolen. For example, the transceiver's serial number can be entered into the software and then used to send a 'ping' the transceiver. The transceiver responds to the ping by collecting and transmitting a location packet as described above. Or the location of all unused transceivers can be automatically rendered on a separate interface. In still other embodiments, the location-determining software can transmit a packet to a specific transceiver (e.g. one that is stolen) to disable it from operating further.

In other embodiments, the 'bumping' process described above can be used for a variety of applications involving the body-worn monitor, wrist-worn transceiver, PDS, and RVD. In embodiments, for example, one or more 'bumps' of a transceiver can modulate the ACC waveform, which is then processed and analyzed to initiate a specific application. Applications include turning the transceiver on/off; attaching sensors to the transceiver; pairing the transceiver with a handheld device (e.g. a cellular phone or personal digital assistant) over a peer-to-peer connection (using, e.g., 802.11 or 802.15.4); pairing the transceiver with a printer connected to a hospital network to print data stored in its computer memory; associating the transceiver with a specific clinician; and initiating display of a particular GUI. In general, the 'bumping' process can be used to initiate any application that can also be initiated with icons on the GUI.

Annotating the Medical Record Using the Wrist-Worn Transceiver

FIGS. 11-13 show how the wrist-worn transceiver can be used to communicate audible information from both the patient and a clinician. Audible information from the clinician **140** can be used, for example, to annotate vital sign information collected with the body-worn monitor. Audible information from the patient **141** can be transmitted to a clinician (e.g. a nurse working at a central station) to alert the clinician of a problem. In both applications, the transceiver **72** is attached to the patient's wrist **66** as described above and used to measure vital signs and waveform information. Audible information is received by a microphone **101** mounted on a circuit board within the transceiver. A speaker **120** mounted to the same circuit board enunciates voice information to the patient. In these and other voice-related applications, voice information is digitized by an internal analog-to-digital converter within the transceiver, and then wirelessly transmitted through a hospital's wireless network using conventional VOIP protocols. Systems that operate these protocols are marketed, for example, by Cisco Systems (170 West Tasman Drive, San Jose, Calif. 93134; www.cisco.com), Skype (22/24 Boulevard Royal, 6e etage, L-2449, Luxembourg; www.skype.com), and others.

FIG. 12 describes the annotation process in more detail. In this case, the transceiver **72** within the body-worn monitor is attached to the patient's wrist **66** to measure the patient's vital signs (e.g. blood pressure). During the measurement process, the clinician uses the GUI **50F** to activate an 'annotation' function which enables the transceiver to receive audible signals **140** which are used, for example, to annotate different medications administered to the patient. After the annotation function is activated, the clinician orally describes the medications. The microphone **101** within the transceiver **72** detects the voice signals, digitizes them with associated hardware, and then sends them and an associated time/date sample using a VOIP protocol through an access point **56** and to a PDS located within the hospital network **60**. Vital signs are transmitted before and after the annotation function is activated, and are stored along with the annotation in a computer

memory associated with the PDS. Typically these data are stored within a hospital's EMR.

As shown in graph **141**, annotated vital sign data can be viewed afterwards to determine, for example, how a patient responds to specific medications. In this case, administration of a beta blocker as a means of lowering the patient's blood pressure is recorded on the graph by a written description of the annotation, along with an icon (a black triangle) indicating when it occurred in time. To generate the written description the PDS requires software that performs a speech-to-text conversion. Such software is available, for example, from Nuance Systems (1 Wayside Road, Burlington, Mass. 01803; www.nuance.com). Similarly, the graph **141** shows a second annotation indicating that the patient was hydrated with saline to increase their blood pressure.

FIG. **13** shows a series of screens within the GUI **50F** that are used to control the annotation process. As described above, to annotate medical information the clinician taps an icon located in the upper right-hand portion of screen **180**. This action readies the voice recording features within the transceiver. Tapping the annotation icon drives the transceiver to render a second screen **182** that includes the type of annotation, e.g. audible content relating to medication, a specific intervention or procedure, a medical assessment, or another subject. Typically annotations are delivered as audible speech, in which case the 'Speech' button is tapped, as shown in screen **184**. Alternatively the annotation can be text or numerical; these can be typed in, e.g., using a 'soft' keyboard on the transceiver, or scanned in using the transceiver's barcode scanner. The annotations can also be associated with an alarm condition, such as those shown on screens **181**, **183**, **185**, **187**, **189**, **191**. Prior to recording an annotation, the GUI renders a screen **188** that, once tapped, initiates the recording. The recording can also be paused using screen **186**. After it is complete, the clinician taps the 'checkbox' on the screen **188**, thus saving the recording. It is then sent to the PDS as shown in FIG. **12**, and used to annotate the patient's medical information.

Other forms of annotation are also possible with the transceiver. For example, it can include a small CCD camera that allows images of the patient or their body (e.g. a wound) to be captured and used to annotate the medical information. In other applications, a barcode printed on medication administered to the patient can be scanned by the transceiver's barcode scanner, and the information encoded therein can be used to annotate vital sign information. In other embodiments, the transceiver can integrate with other equipment in the hospital room (e.g. an infusion pump, ventilator, or patient-controlled anesthesia pump) through a wired or wireless connection, and information from this equipment can be collected and transmitted to the PDS in order to annotate the vital sign information. In other embodiments, text annotations can be stored on the PDS, and then edited afterwards by the clinician.

Other GUI Applications

As shown in FIGS. **11** and **14**, the speaker **120** and microphone **101** within the transceiver **72**, combined with VOIP software operating on the hospital network, can also function as a nurse call system that communicates both distress signals and voice information. Here, the transceiver enables two-way communication between the patient and a remote clinician. During this application, the transceiver typically operates the 'patient GUI', shown schematically in FIG. **3** and in more detail in FIG. **14**. Here, the GUI shows a single screen **192** that indicates a nurse call function with an icon showing a telephone. When the patient taps on the telephone the transceiver initiates a call to a pre-programmed IP address, correspond-

ing, e.g., to a computer at a central nursing station or a VOIP-enabled phone. Alternatively the transceiver can call a pre-programmed phone number corresponding to a telephone. While the call is being place the GUI renders a screen **194** that shows the telephone's receiver being off the hook. A third screen **196** indicates that the patient is connected to the clinician. The call is terminated when the patient finishes talking to the clinician and taps the screen. Alternatively, the transceiver can include software that detects that no further voice communications are taking place, and then uses this information to terminate the call. In embodiments, the entire call can be stored in a computer memory on either the transceiver or the PDS.

The GUI operating on the wrist-worn transceiver's touch-panel display can render several other interfaces that facilitate patient monitoring in the hospital. For example, referring to FIG. **15**, the GUI can be used to monitor the patient's pain level, a parameter often considered by clinicians to be as important as vital signs for characterizing a patient. The GUI **200** shown in the figure features a simple series of icons that provide a relative indication of the patient's pain level. An index value of 0 (corresponding to a 'happy' face) indicates a low level of pain; an index value of 10 (corresponding to a 'sad' face) indicates a high level of pain. During a measurement, the patient simply touches the icon that best characterizes their pain level. The numerical value corresponding to this level is then wirelessly transmitted back to the PDS and stored in the patient's EMR. The GUI, for example, may be automatically rendered periodically (e.g. every hour) on the transceiver to continuously monitor the patient's pain level. In other embodiments, the GUI could render a graphical display that provides a more sophisticated metric for determining the patient's pain, such as the McGill Pain Questionnaire. This system described in the following journal article, the contents of which are incorporated herein by reference: 'The McGill Pain Questionnaire: Major Properties and Scoring Methods', Melzak, *Pain*, 1:277-299 (1975).

In a similar manner, the GUI can be used to gauge the patient's level of mentation, i.e. mental activity. Mentation has been consistently shown to be a valuable tool for diagnosing a patient, but is typically determined empirically by a clinician during a check-up or hospital visit. Such a diagnosis is somewhat arbitrary and requires the clinician to meet face-to-face with the patient, which is often impractical. But with the wrist-worn transceiver, diagnosis of mentation can be made automatically at the patient's bedside without a clinician needing to be present. FIG. **16**, for example, shows a GUI **202** that provides a simple 'mentation test' for the patient to complete. In this case, the mentation test involves a graphical representation of a series of non-sequential numbers. The patient completes this test by tapping on the numbers rendered by the touchpanel display in their numerical order. An algorithm then 'scores' the test based on accuracy and the time required to complete it. Once determined, the score is wirelessly transmitted back to the PDS, and then stored in the patient's EMR. Other simple tests with varying complexity can be used in place of that shown in FIG. **16**. The tests can vary depending on the specific mentation function to be tested. For example, unique tests can be generated for patients with head injuries, cardiac patients, patients in severe pain, Alzheimer's patients, etc. In all cases, the tests are designed to make a quantitative assessment of the patient's mental status; the transceiver sends a numerical value representing this parameter and an identifier for the test back to the EMR for analysis. The transceiver can be programmed so that the GUI **202** for the mentation test, like the GUI **200** for pain level shown in FIG. **15**, is automatically rendered at basically any

time interval on the touchpanel display. This time interval can be periodic and on an hourly basis, once/day, etc.

As shown in FIG. 17, the transceiver can include a GUI 204 that displays a photograph or video of the patient. The photograph could be taken by a digital camera within the transceiver, or with an external camera and then transferred to the transceiver through a variety of means, e.g. the hospital's wireless network, a peer-to-peer wireless connection, using a non-volatile memory such as an SD card, or even using a data-transfer process initiated by the 'bump' methodology described above. In general, the same means used to port a photograph from a standard digital camera to a personal computer or other device can be used in this application. Once the photograph is received, software on the transceiver displays it in either a default screen (e.g., in place of the 'beating heart' shown in FIGS. 1 and 3), or when the GUI 204 is activated through a tap of a corresponding icon. Displaying the patient's photograph in this manner provides a visual indicator which the clinician can use to correctly identify the patient. In other embodiments, a photograph of someone associated with the patient (e.g. a relative) can also be displayed on the GUI 204. Such an embodiment may be particularly useful for neo-natal hospitals wards, wherein one or more photographs of an infant's parents could be displayed on a transceiver attached to the infant. This way a clinician could check the photograph to ensure that visitors to the neo-natal hospital ward are, in fact, the infant's parents.

FIGS. 18 and 19 show other GUIs 206, 208 that can be rendered on the wrist-worn transceiver's display to carry out basic features in the hospital, such as meal ordering (FIG. 18), and changing the channel on a television or computer (FIG. 19). In these cases, the PDS associated with the transceiver receives a packet describing the function at hand (e.g., the meal that has been ordered, or the channel that is desired), and communicates with another software application in the hospital to complete the transaction. This communication, for example, can take place using a XML-based Web Services operation, such as that described in the following patent application, the contents of which are incorporated herein by reference: CUFFLESS BLOOD PRESSURE MONITOR AND ACCOMPANYING WEB SERVICES INTERFACE (U.S. Ser. No. 10/810,237, Filed Mar. 26, 2004). In related embodiments, a GUI similar to that shown in FIG. 19 can be used to order movies, video games, television programs stored on a digital video recorder, books, and music. Content corresponding to these components is typically stored on a remote server, and then accessed using an XML-based operation, as described above.

In yet another application, as shown in FIG. 20, the wrist-worn transceiver 72 and its associated barcode scanner 102 can be used to check medication before it is administered to the patient. In this embodiment, barcodes associated with the patient 63, clinician 65, and the medication 67 are read by the barcode scanner 102 within the transceiver 72. The transceiver then wirelessly transmits decoded barcode information through a local access point 56 and to the PDS connected to the hospital network 60. Using a software program, the PDS analyzes these data and communicates with the patient's record in the hospital EMR 58 to determine if the medication is appropriate for the patient. For example, the software program may check to see if the patient is allergic to the medication, if the dosage is correct, or if the patient has previously exhibited any detrimental side effects that may affect the dosage. In related embodiments, the transceiver may also include a GUI wherein the clinician enters ancillary information, such as the dosage of the medication or demographic information describing the patient, using a 'soft' keypad. Or

the GUI may include a simple questionnaire that guides the clinician through the process of checking the medication, and then administering it. In still other embodiments, the infusion pump that delivers the medication may include a wireless connection through the access point 56 to the PDS 60 or to the transceiver 72 to automatically supply information related to the medication to the software program.

Once the software program determines that it is safe to administer the medication, it sends a packet from the PDS 60, through the access point 56, and back to the transceiver 72, which then renders a GUI instructing the clinician to proceed. In other embodiments, the PDS 60 sends the packet through the access point 56 to either a remote computer 62 (e.g. a tablet computer) or a portable device 64 (e.g. a cellular telephone or personal digital assistant).

Form Factor of the Body-Worn Monitor

FIGS. 21A and 21B show how the body-worn monitor 100 described above attaches to a patient 70 to measure RR, SpO₂, cNIBP, and other vital signs. These figures show two configurations of the system: FIG. 21A shows the system used during the indexing portion of the Composite Technique, and includes a pneumatic, cuff-based system 85, while FIG. 21B shows the system used for subsequent measurements. The indexing measurement typically takes about 60 seconds, and is typically performed once every 4-8 hours. Once the indexing measurement is complete the cuff-based system 85 is typically removed from the patient. The remainder of the time the monitor 100 performs the RR, HR, SpO₂ and cNIBP measurements.

The body-worn monitor 100 features a wrist-worn transceiver 72, described in more detail in FIGS. 22A and 22B, featuring a touch panel interface 73 that displays the various GUIs described above and in FIG. 24. A wrist strap 90 affixes the transceiver 72 to the patient's wrist like a conventional wristwatch. A flexible cable 92 connects the transceiver 72 to an optical sensor 94 that wraps around the base of the patient's thumb. During a measurement, the optical sensor 94 generates a time-dependent PPG waveform which is processed along with an ECG to measure cNIBP, SpO₂, and, in some applications, RR. To determine ACC waveforms the body-worn monitor 100 features three separate accelerometers located at different portions on the patient's arm and chest. The first accelerometer is surface-mounted on a circuit board in the wrist-worn transceiver 72 and measures signals associated with movement of the patient's wrist. As described above, this motion can also be indicative of that originating from the patient's fingers, which will affect the SpO₂ measurement. The second accelerometer is included in a small bulkhead portion 96 included along the span of the cable 82. During a measurement, a small piece of disposable tape, similar in size to a conventional bandaid, affixes the bulkhead portion 96 to the patient's arm. In this way the bulkhead portion 96 serves two purposes: 1) it measures a time-dependent ACC waveform from the mid-portion of the patient's arm, thereby allowing their posture and arm height to be determined as described in detail above; and 2) it secures the cable 82 to the patient's arm to increase comfort and performance of the body-worn monitor 100, particularly when the patient is ambulatory. The third accelerometer is mounted in the sensor module 74 that connects through cables 80a-c to ECG electrodes 78a-c. Signals from these sensors are then digitized, transmitted through the cable 82 to the wrist-worn transceiver 72, where they are processed with an algorithm as described above to determine RR.

The cuff-based module 85 features a pneumatic system 76 that includes a pump, valve, pressure fittings, pressure sensor, manifold, analog-to-digital converter, microcontroller, and

rechargeable Li-ion battery. During an indexing measurement, the pneumatic system **76** inflates a disposable cuff **84** and performs two measurements according to the Composite Technique: 1) it performs an inflation-based measurement of oscillometry and measurement of a corresponding OSC waveform to determine values for SYS, DIA, and MAP; and 2) it determines a patient-specific relationship between PTT and MAP. These measurements are described in detail in the co-pending patent application entitled: 'VITAL SIGN MONITOR FOR MEASURING BLOOD PRESSURE USING OPTICAL, ELECTRICAL, AND PRESSURE WAVEFORMS' (U.S. Ser. No. 12/138,194; filed Jun. 12, 2008), the contents of which are incorporated herein by reference.

The cuff **84** within the cuff-based pneumatic system **85** is typically disposable and features an internal, airtight bladder that wraps around the patient's bicep to deliver a uniform pressure field. During the indexing measurement, pressure values are digitized by the internal analog-to-digital converter, and sent through a cable **86** according to a CAN protocol, along with SYS, DIA, and MAP blood pressures, to the wrist-worn transceiver **72** for processing as described above. Once the cuff-based measurement is complete, the cuff-based module **85** is removed from the patient's arm and the cable **86** is disconnected from the wrist-worn transceiver **72**. cNIBP is then determined using PTT, as described in detail above.

To determine an ECG, the body-worn monitor **100** features a small-scale, three-lead ECG circuit integrated directly into the sensor module **74** that terminates an ECG cable **82**. The ECG circuit features an integrated circuit that collects electrical signals from three chest-worn ECG electrodes **78a-c** connected through cables **80a-c**. As described above, the ECG electrodes **78a-c** are typically disposed in a conventional Einthoven's Triangle configuration, which is a triangle-like orientation of the electrodes **78a-c** on the patient's chest that features three unique ECG vectors. From these electrical signals the ECG circuit determines up at least three ECG waveforms, each corresponding to a unique ECG vector, which are digitized using an analog-to-digital converter mounted proximal to the ECG circuit and sent through the cable **82** to the wrist-worn transceiver **72** according to the CAN protocol. There, the ECG and PPG waveforms are processed to determine the patient's blood pressure. HR and RR are determined directly from the ECG waveform using known algorithms, such as those described above. More sophisticated ECG circuits (e.g. five and twelve-lead systems) can plug into the wrist-worn transceiver to replace the three-lead system shown in FIGS. **21A** and **21B**.

FIGS. **22A**, **22B** show three-dimensional views of the wrist-worn transceiver **72** before and after receiving cables **82**, **86**, **89** from sensors worn on the patient's upper arm and torso, as well as the cable **92** that connects to the optical sensor. The transceiver **72** is sealed in a waterproof plastic casing **117** featuring electrical interconnects (not shown in the figure) on its bottom surface that interface to the terminal ends **111**, **119a-c** of cables **82**, **86**, **89**, **92** leading to the monitor's various sensors. The electrical interconnects support serial communication through the CAN protocol, described in detail herein, particularly with reference to FIG. **25**. During operation, the transceiver's plastic casing **117** snaps into a plastic housing **106**, which features an opening **109** on one side to receive the terminal end **111** of the cable **92** connected to the optical sensor. On the opposing side the plastic housing **106** features three identical openings **104a-c** that receive the terminal ends **119a-c** of cables **82**, **86**, **89** connected to the ECG and accelerometer systems (cable **82**),

the pneumatic cuff-based system (cable **86**), and ancillary systems (cable **89**) described above. In addition to being waterproof, this design facilitates activities such as cleaning and sterilization, as the transceiver contains no openings for fluids common in the hospital, such as water and blood, to flow inside. During a cleaning process the transceiver **72** is simply detached from the plastic housing **106** and then cleaned.

The transceiver **72** attaches to the patient's wrist using a flexible strap **90** which threads through two D-ring openings in the plastic housing **106**. The strap **90** features mated Velcro patches on each side that secure it to the patient's wrist during operation. A touchpanel display **50** renders the various GUIs described above.

The electrical interconnects on the transceiver's bottom side line up with the openings **104a-c**, and each supports the CAN protocol to relay a digitized data stream to the transceiver's internal CPU, as described in detail with reference to FIG. **25**. This allows the CPU to easily interpret signals that arrive from the monitor's body-worn sensors, and means that these connectors are not associated with a specific cable. Any cable connecting to the transceiver **72** can be plugged into any opening **104a-c**. As shown in FIG. **22A**, the first opening **104a** receives the cable **82** that transports digitized ECG waveforms determined from the ECG circuit and electrodes, and digitized ACC waveforms measured by accelerometers in the cable bulkhead and the bulkhead portion associated with the ECG cable **82**.

The second opening **104b** receives the cable **86** that connects to the pneumatic cuff-based system used for the pressure-dependent indexing measurement. This connector receives a time-dependent pressure waveform delivered by the pneumatic system to the patient's arm, along with values for SYS, DIA, and MAP determined during the indexing measurement. The cable **86** unplugs from the opening **104b** once the indexing measurement is complete, and is plugged back in after approximately 4-8 hours for another indexing measurement.

The final opening **104c** can be used for an auxiliary device, e.g. a glucometer, infusion pump, body-worn insulin pump, ventilator, or end-tidal CO₂ monitoring system. As described with reference to FIG. **25**, digital information generated by these systems will include a header that indicates their origin so that the CPU can process them accordingly.

Measuring and Displaying Time-Dependent Physiological Signals

FIGS. **23A** and **23B** show how a network of sensors **78a-c**, **83**, **84**, **87**, **94** within the body-worn monitor **100** connect to a patient **70** to measure time-dependent ECG **261**, PPG **262**, OSC **263**, ACC **264**, and RR **265** waveforms. These, in turn, yield the patient's vital signs and motion parameters. Each waveform **261-265** relates to a unique physiological characteristic of the patient **70**. For example, each of the patient's heartbeats generates electrical impulses that pass through the body near the speed of light, along with a pressure wave that propagates through the patient's vasculature at a significantly slower speed. Immediately after the heartbeat, the pressure wave leaves the heart **148** and aorta **149**, passes through the subclavian artery **150** to the brachial artery **144**, and from there through the radial and ulnar arteries **145** to smaller arteries in the patient's fingers. Three disposable electrodes **78a-c** attached to the patient's chest measure unique electrical signals which pass to a single-chip ECG circuit **83** that terminates a distal end of the ECG cable. Typically, these electrodes attach to the patient's chest in a conventional 'Einthoven's triangle' configuration featuring three unique 'vectors', each corresponding to a different lead (e.g. LEAD

1, II, II). Related configurations can also be used when five and twelve-lead ECG systems are used in place of the three-lead system, as described above with reference to FIGS. 21A, 21B. Within the ECG circuit 83 signals are processed using an amplifier/filter circuit and analog-to-digital converter to generate a digital ECG waveform 261 corresponding to each lead. The ECG waveform 261 features a sharp, well-defined QRS complex corresponding to each heartbeat; this marks the initiation of the patient's cardiac cycle. Heart rate is determined directly from the ECG waveform 261 using known algorithms, such as those described in the following journal article, the contents of which are incorporated herein by reference: 'ECG Beat Detection Using Filter Banks', Afonso et al., *IEEE Trans. Biomed Eng.*, 46:192-202 (1999).

To generate an IP waveform 265, one of the ECG electrodes in the circuit 78a is a 'driven lead' that injects a small amount of modulated current into the patient's torso. A second, non-driven electrode 78c, typically located on the opposite side of the torso, detects the current, which is further modulated by capacitance changes in the patient's chest cavity resulting from breathing. Further processing and filtering of the IP waveforms 265 yields respiratory rate. Respiration can also be determined using an adaptive filtering approach that processes both the IP waveform and ACC waveform 264, as described in more detail in the following co-pending patent application, the contents of which are incorporated herein by reference: BODY-WORN MONITOR FOR MEASURING RESPIRATION RATE (U.S. Ser. No. 12/559,419, Filed Sep. 14, 2009).

The optical sensor 94 features two LEDs and a single photodetector that collectively measure a time-dependent PPG waveform 262 corresponding to each of the LEDs. The sensor and algorithms for processing the PPG waveforms are described in detail in the following co-pending patent application, the contents of which have been previously incorporated herein by reference: BODY-WORN PULSE OXIMETER (U.S. Ser. No. 12/559,379; filed Sep. 14, 2009). The waveform 262 represents a time-dependent volumetric change in vasculature (e.g. arteries and capillaries) that is irradiated with the sensor's optical components. Volumetric changes are induced by a pressure pulse launched by each heartbeat that travels from the heart 148 to arteries and capillaries in the thumb according to the above-described arterial pathway. Pressure from the pressure pulse forces a bolus of blood into this vasculature, causing it to expand and increase the amount of radiation absorbed, and decrease the transmitted radiation at the photodetector. The pulse shown in the PPG waveform 262 therefore represents the inverse of the actual radiation detected at the photodetector. It follows the QRS complex in the ECG waveform 261, typically by about one to two hundred milliseconds. The temporal difference between the peak of the QRS complex and the foot of the pulse in the PPG waveform 262 is the PTT, which as described in detail below is used to determine blood pressure according to the Composite Technique. PTT-based measurements made from the thumb yield excellent correlation to blood pressure measured with a femoral arterial line. This provides an accurate representation of blood pressure in the central regions of the patient's body.

Each accelerometer generates three time-dependent ACC waveforms 264, corresponding to the x, y, and z-axes, which collectively indicate the patient's motion, posture, and activity level. The body-worn monitor, as described above, features three accelerometers that attach to the patient: one in the wrist-worn transceiver 72, one in the ECG circuit 83, and one

the shape of the ACC waveform 264 indicate the type of motion that the patient is undergoing. For example, the waveform 264 can feature a relatively time-invariant component indicating a period of time when the patient is relatively still, and a time-variant component when the patient's activity level increases. Magnitudes of both components will depend on the relationship between the accelerometer and a gravity vector, and can therefore be processed to determine time-invariant features, such as posture and arm height. A frequency-dependent analysis of the time-variant components yields the type and degree of patient motion. Analysis of ACC waveforms 264 is described in detail in the above-mentioned patent applications, the contents of which have been fully incorporated herein by reference.

The OSC waveform 263 is generated from the patient's brachial artery 144 with the pneumatic system and a cuff-based sensor 84 during the pressure-dependent portion of the Composite Technique. It represents a time-dependent pressure which is applied to the brachial artery during inflation and measured by a digital pressure sensor within the pneumatic system. The waveform 263 is similar to waveforms measured during deflation by conventional oscillometric blood pressure monitors. During a measurement, the pressure waveform 263 increases in a mostly linear fashion as pressure applied by the cuff 84 to the brachial artery 144 increases. When it reaches a pressure slightly below the patient's diastolic pressure, the brachial artery 144 begins to compress, resulting in a series time-dependent pulsations caused by each heartbeat that couple into the cuff 84. The pulsations modulate the OSC waveform 263 with an amplitude that varies in a Gaussian-like distribution, with maximum modulation occurring when the applied pressure is equivalent to the patient's MAP. The pulsations can be filtered out and processed using digital filtering techniques, such as a digital bandpass filter that passes frequencies ranging from 0.5-20 Hz. The resulting waveform can be processed to determine SYS, DIA, and MAP, as is described in detail in the above-referenced patent applications, the contents of which have been previously incorporated herein by reference. The cuff 84 and pneumatic system are removed from the patient's bicep once the pressure-dependent component of the Composite Technique is complete.

The high-frequency component of the OSC waveform 263 (i.e. the pulses) can be filtered out to estimate the exact pressure applied to the patient's brachial artery during oscillometry. According to the Composite Technique, PTT measured while pressure is applied will gradually increase as the brachial artery is occluded and blood flow is gradually reduced. The pressure-dependent increase in PTT can be fit with a model to estimate the patient-specific relationship between PTT and blood pressure. This relationship, along with SYS, MAP, and DIA determined from the OSC waveform during inflation-based oscillometry, is used during the Composite Technique's pressure-free measurements to determine blood pressure directly from PTT.

There are several advantages to making the indexing measurement during inflation, as opposed to deflation. Measurements made during inflation are relatively fast and comfortable compared to those made during deflation. Inflation-based measurements are possible because of the Composite Technique's relatively slow inflation speed (typically 5-10 mmHg/second) and the high sensitivity of the pressure sensor used within the body sensor. Such a slow inflation speed can be accomplished with a small pump that is relatively lightweight and power efficient. Moreover, measurements made during inflation can be immediately terminated once systolic blood pressure is calculated. This tends to be more comfort-

able than conventional cuff-based measurements made during deflation. In this case, the cuff typically applies a pressure that far exceeds the patient's systolic blood pressure; pressure within the cuff then slowly bleeds down below the diastolic pressure to complete the measurement.

A digital temperature sensor proximal to the ECG circuit **83** measures the patient's skin temperature at their torso. This temperature is an approximation of the patient's core temperature, and is used mostly for purposes related to trending and alarms/alerts.

FIG. **24** shows how the above-described ECG, PPG, and IP waveforms, along with vital signs calculated from them, are rendered using different screens **300**, **304**, **306**, **308** within a GUI. In all cases, the waveforms are displayed with a rolling graphical technique, along with a moving bar that indicates the most current point in time. As per the AAMI/ANSI EC-13 reference standard, the ECG waveforms are displayed alongside a bar that indicates a signal intensity of 1 mV. Screen **308** shows different ECG vectors (corresponding to, e.g., Lead I, II, III, aVR, aVF) that are rendered when the clinician taps the ECG waveform on screen **300**, and then the corresponding lead on screen **308**. Waveforms for a particular vital sign (e.g. a PPG waveform for the SpO₂ measurement; an IP waveform for the RR measurement) are rendered when the clinician taps on the value of the corresponding vital sign. During a measurement both waveforms and the vital signs calculated from them are wirelessly transmitted to the PDS, as described above.

Communicating with Multiple Systems Using the CAN Protocol

As described above, the ECG, ACC, and pneumatic systems within the body-worn system send digitized information to the wrist-worn transceiver through the CAN protocol. FIG. **25** shows a schematic drawing indicating how CAN packets **201a-d**, **212a-e** transmitted between these systems facilitate communication. Specifically, each of the ACC **215**, ECG **216**, pneumatic **220**, and auxiliary **245** systems include a separate analog-to-digital converter, microcontroller, frequency-generating crystal oscillator (typically operating at 100 kHz), and real-time clock divider that collectively generate and transmit digital data packets **201a-d** according to the CAN protocol to the wrist-worn transceiver **72**. Each crystal uses the internal real-time clock on the internal microprocessor within the respective system. This allows the microcontroller within each system to be placed in a low-power state in which its real-time operating system (RTOS) dispatch system indicates that it is not ready to run a task. The real-time clock divider is programmed to create an interrupt which wakes up the microcontroller every 2 milliseconds.

The wrist-worn transceiver **72** features a 'master clock' that generates real-time clock 'ticks' at the sampling rate (typically 500 Hz, or 2 ms between samples). Each tick represents an incremented sequence number. Every second, the wrist-worn transceiver **72** transmits a packet **212e** over the CAN bus that digitally encodes the sequence number. One of the criteria for accurate timing is that the time delay between the interrupt and the transmission of the synchronizing packet **212e**, along with the time period associated with the CAN interrupt service routine, is predictable and stable. During initialization, the remote CAN buses do not sleep; they stay active to listen for the synchronization packet **212e**. The interrupt service routine for the synchronization packet **212e** then establishes the interval for the next 2 millisecond interrupt from its on-board, real-time crystal to be synchronized with the timing on the wrist-worn transceiver **72**. Offsets for the packet transmission and interrupt service delays are factored into the setting for the real-time oscillator to interrupt syn-

chronously with the microprocessor on the wrist-worn transceiver **72**. The magnitude of the correction factor to the real-time counter is limited to 25% of the 2 millisecond interval to ensure stability of this system, which represents a digital phase-locked loop.

When receipt of the synchronization packet **212e** results in a timing correction offset of either a 0, +1, or -1 count on the remote system's oscillator divider, software running on the internal microcontroller declares that the system is phase-locked and synchronized. At this point, it begins its power-down operation and enables measurement of data as described above.

Each remote system is driven with a 100 kHz clock, and a single count of the divider corresponds to 20 microseconds. This is because the clock divider divides the real-time clock frequency by a factor of 2. This is inherent in the microcontroller to ensure that the clock has a 50% duty cycle, and means the clock can drift +/-20 microseconds before the actual divider chain count will disagree by one count, at which time the software corrects the count to maintain a phase-locked state. There is thus a maximum of 40 microseconds of timing error between data transmitted from the remote systems over the CAN bus. Blood pressure is the one vital sign measured with the body-worn monitor that is calculated from time-dependent waveforms measured from different systems (e.g. PPG and ECG waveforms). For this measurement, the maximum 40-microsecond timing error corresponds to an error of +/-0.04 mmHg, which is well within the error (typically +/-5 mmHg) of the measurement.

In order to minimize power consumption, the wrist-worn transceiver **72** and remote systems **215**, **216**, **220**, **245** power down their respective CAN bus transceivers between data transfers. During a data transfer, each system generates a sequence number based that is included in the synchronization packet **212e**. The sequence number represents the interval between data transfers in intervals of 2 milliseconds. It is a factor of 500 (e.g. 2, 4, 5, 10) that is the number of 2 millisecond intervals between transfers on the CAN bus. Each remote system enables its CAN bus during the appropriate intervals and sends its data. When it has finished sending its data, it transmits a 'transmit complete' packet indicating that the transmission is complete. When a device has received the 'transmit complete' packet it can disable its CAN transceiver to further reduce power consumption.

Software in each of the ACC **215**, ECG **216**, pneumatic **220**, and auxiliary **245** systems receive the sequence packet **212e** and the corresponding sequence number, and set their clocks accordingly. There is typically some inherent error in this process due to small frequency differences in the crystals (from the ideal frequency of 100 kHz) associated with each system. Typically this error is on the order of microseconds, and has only a small impact on time-dependent measurements, such as PTT, which are typically several hundred milliseconds.

Once timing on the CAN bus is established using the above-described procedure, each of the ACC **215**, ECG **216**, and pneumatic **220** systems generate time-dependent waveforms that are transmitted in packets **201a-d**, each representing an individual sample. Each packet **201a-d** features a header portion which includes the sequence number **212a-d** and an initial value **210a-d** indicating the type of packet that is transmitted. For example, accelerometers used in the body-worn system are typically three-axis digital accelerometers, and generate waveforms along the x, y, and z-axes. In this case, the initial value **210a** encodes numerical values that indicate: 1) that the packet contains ACC data; and 2) the axis (x, y, or z) over which these data are generated. Similarly, the

ECG system **216** can generate a time-dependent ECG waveform corresponding to Lead I, II, or III, each of which represents a different vector measured along the patient's torso. Additionally, the ECG system **216** can generate processed numerical data, such as heart rate (measured from time increments separating neighboring QRS complexes), respiratory rate (from an internal impedance pneumography component), as well as alarms calculated from the ECG waveform that indicate problematic cardiovascular states such as VTAC, VFIB, and PVCs. Additionally, the ECG system can generate error codes indicating, for example, that one of the ECG leads has fallen off. The ECG system typically generates an alarm/alert, as described above, corresponding to both the error codes and potentially problematic cardiovascular states. In this case, the initial value **210b** encodes numerical values that indicate: 1) that the packet contains ECG data; 2) the vector (Lead I, II, or III) corresponding to the ECG data; and 3) an indication if a cardiovascular state such as VTAC, VFIB, or PVCs was detected.

The pneumatic system **220** is similar to the ECG system in that it generates both time-dependent waveforms (i.e. a pressure waveform, measured during oscillometry, characterizing the pressure applied to the arm and subsequent pulsations measured during an oscillometric measurement) and calculated vital signs (SYS, DIA, and MAP measured during oscillometry). In some cases errors are encountered during the oscillometric blood pressure measurement. These include, for example, situations where blood pressure is not accurately determined, an improper OSC waveform, over-inflation of the cuff, or a measurement that is terminated before completion. In these cases the pneumatic system **220** generates a corresponding error code. For the pneumatic system **220** the initial value **210c** encodes numerical values that indicate: 1) that the packet contains blood pressure data; 2) an indication that the packet includes an error code.

In addition to the initial values **210a-d**, each packet **201a-d** includes a data field **214a-d** that encodes the actual data payload. Examples of data included in the data fields **214a-d** are: 1) sampled values of ACC, ECG, and pressure waveforms; 2) calculated heart rate and blood pressure values; and 3) specific error codes corresponding to the ACC **215**, ECG **216**, pneumatic **220**, and auxiliary **225** systems.

Upon completion of the measurement, the wrist-worn transceiver **72** receives all the CAN packets **201a-d**, and synchronizes them in time according to the sequence number **212a-d** and identifier **210a-d** in the initial portions **216** of each packet. Every second, the CPU updates the time-dependent waveforms and calculates the patient's vital signs and motion-related properties, as described above. Typically these values are calculated as a 'rolling average' with an averaging window ranging from 10-20 seconds. The rolling average is typically updated every second, resulting in a new value that is displayed on the wrist-worn transceiver **72**. Each packet received by the transceiver **72** is also wirelessly retransmitted as a new packet **201b'** through a wireless access point **56** and to both an PDS and RVD within a hospital network **60**. The new packet **201b'** includes the same header **210b'**, **212b'** and data field information **214b'** as the CAN packets transmitted between systems within the body-worn monitor. Also transmitted are additional packets encoding the cNIBP, SpO2, and processed motion states (e.g. posture, activity level, degree of motion), which unlike heart rate and SYS, DIA, and MAP are calculated by the CPU in the wrist-worn transceiver. Upon receipt of the packet **201b'**, the RVD displays vital signs, waveforms, motion information, and alarms/alerts, typically with a large monitor that is easily viewed by a clinician. Additionally the PDS can send infor-

mation through the hospital network (e.g. in the case of an alarm/alert), store information in an internal database, and transfer it to a hospital EMR.

Alternate IT Configurations

FIG. **26** shows an alternate configuration of the invention wherein the transceiver **72** transmits both voice and data information through a wireless access point **56A** and to the Internet **55**, and from there to the hospital network and PDS **60**. Such a configuration would be used, for example, when the patient is located outside of the hospital (e.g. at home). It allows clinicians to monitor and care for a patient as if they were located in the hospital. Once information arrives at the PDS **60**, it can be transferred to the hospital EMR system **58**, or through a wireless access point **56B** within the hospital to an external computer **62** or a portable device **64**.

In an alternate embodiment, as shown in FIG. **27**, the first wireless access point **56A** shown in FIG. **26** is replaced by a wireless modem **64A**, such as a cellular telephone or personal digital assistant. Here, the wireless modem **64A** receives voice and data information from the transceiver through a peer-to-peer wireless interface (e.g. an interface based on 802.11b/g or 802.15.4). The wireless modem **64A** then transmits the voice and data information to the Internet **55** using, e.g., a cellular connection, such as one based on GSM or CDMA. In yet another embodiment, as shown in FIG. **28**, the transceiver **72** includes an internal long-range wireless transmitter based on a cellular protocol (e.g. GSM or CDMA), allowing it to transmit voice and data information directly to the Internet **55**. In the embodiments shown in both FIGS. **27** and **28**, information sent through the Internet is ultimately received by the PDS **60**, and is sent from there through a wireless access point **56** to either the remote computer **62** or portable device **64**.

In embodiments, the transceiver **72** features multiple wireless transmitters, and can operate in multiple modes, such as each of those shown in FIGS. **26-28**. In this case the wireless protocol (based on, e.g. 802.11 or cellular) is manually selected using the GUI, or automatically selected based on the strength of the ambient wireless signal.

Other Embodiments of the Invention

In addition to those methods described above, the body-worn monitor can use a number of additional methods to calculate blood pressure and other properties from the optical and electrical waveforms. These are described in the following co-pending patent applications, the contents of which are incorporated herein by reference: 1) CUFFLESS BLOOD-PRESSURE MONITOR AND ACCOMPANYING WIRELESS, INTERNET-BASED SYSTEM (U.S. Ser. No. 10/709,015; filed Apr. 7, 2004); 2) CUFFLESS SYSTEM FOR MEASURING BLOOD PRESSURE (U.S. Ser. No. 10/709,014; filed Apr. 7, 2004); 3) CUFFLESS BLOOD PRESSURE MONITOR AND ACCOMPANYING WEB SERVICES INTERFACE (U.S. Ser. No. 10/810,237; filed Mar. 26, 2004); 4) BILATERAL DEVICE, SYSTEM AND METHOD FOR MONITORING VITAL SIGNS (U.S. Ser. No. 11/420,774; filed May 27, 2006); 5) CUFFLESS BLOOD PRESSURE MONITOR AND ACCOMPANYING WIRELESS MOBILE DEVICE (U.S. Ser. No. 10/967,511; filed Oct. 18, 2004); 6) BLOOD PRESSURE MONITORING DEVICE FEATURING A CALIBRATION-BASED ANALYSIS (U.S. Ser. No. 10/967,610; filed Oct. 18, 2004); 7) PERSONAL COMPUTER-BASED VITAL SIGN MONITOR (U.S. Ser. No. 10/906,342; filed Feb. 15, 2005); 8) PATCH SENSOR FOR MEASURING BLOOD PRESSURE WITHOUT A CUFF (U.S. Ser. No. 10/906,315; filed Feb. 14, 2005); 9)

PATCH SENSOR FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/160,957; filed Jul. 18, 2005); 10) WIRELESS, INTERNET-BASED SYSTEM FOR MEASURING VITAL SIGNS FROM A PLURALITY OF PATIENTS IN A HOSPITAL OR MEDICAL CLINIC (U.S. Ser. No. 11/162,719; filed Sep. 9, 2005); 11) HAND-HELD MONITOR FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/162,742; filed Sep. 21, 2005); 12) CHEST STRAP FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/306,243; filed Dec. 20, 2005); 13) SYSTEM FOR MEASURING VITAL SIGNS USING AN OPTICAL MODULE FEATURING A GREEN LIGHT SOURCE (U.S. Ser. No. 11/307,375; filed Feb. 3, 2006); 14) BILATERAL DEVICE, SYSTEM AND METHOD FOR MONITORING VITAL SIGNS (U.S. Ser. No. 11/420,281; filed May 25, 2006); 15) SYSTEM FOR MEASURING VITAL SIGNS USING BILATERAL PULSE TRANSIT TIME (U.S. Ser. No. 11/420,652; filed May 26, 2006); 16) BLOOD PRESSURE MONITOR (U.S. Ser. No. 11/530,076; filed Sep. 8, 2006); 17) TWO-PART PATCH SENSOR FOR MONITORING VITAL SIGNS (U.S. Ser. No. 11/558,538; filed Nov. 10, 2006); and, 18) MONITOR FOR MEASURING VITAL SIGNS AND RENDERING VIDEO IMAGES (U.S. Ser. No. 11/682,177; filed Mar. 5, 2007).

Other embodiments are also within the scope of the invention. For example, other measurement techniques, such as conventional oscillometry measured during deflation, can be used to determine SYS, DIA, and MAP for the above-described algorithms. Additionally, processing units and probes for measuring pulse oximetry similar to those described above can be modified and worn on other portions of the patient's body. For example, optical sensors with finger-ring configurations can be worn on fingers other than the thumb. Or they can be modified to attach to other conventional sites for measuring SpO₂, such as the ear, forehead, and bridge of the nose. In these embodiments the processing unit can be worn in places other than the wrist, such as around the neck (and supported, e.g., by a lanyard) or on the patient's waist (supported, e.g., by a clip that attaches to the patient's belt). In still other embodiments the probe and processing unit are integrated into a single unit.

In embodiments, the interface rendered on the display at the central nursing station features a field that displays a map corresponding to an area with multiple sections. Each section corresponds to the location of the patient and includes, e.g., the patient's vital signs, motion parameter, and alarm parameter. For example, the field can display a map corresponding to an area of a hospital (e.g. a hospital bay or emergency room), with each section corresponding to a specific bed, chair, or general location in the area.

Further embodiments of the invention are within the scope of the following claims:

What is claimed is:

1. A system for monitoring a patient, comprising: at least one vital sign sensor configured to measure and transmit a time-dependent vital sign waveform; a vital sign monitor configured to be worn on the patient's body and comprising: a housing, a display device on a surface of the housing comprising a touch panel display configured to receive data input and to render a plurality of interfaces, a processing component comprising a microprocessor and a first interface to the vital sign sensor configured to receive the vital sign waveform, and a second interface to the display device configured to receive the data input and to cause the display device to render the plurality of interfaces, the processing component

configured to (i) determine from the vital sign waveform a vital sign value, (ii) cause the display device to render a first graphical user interface in a default mode, (iii) cause the display device to render a second graphical user interface which displays the vital sign value in a password-protected mode, (iv) cause the display device to render a third graphical user interface which collects data input characterizing the patient's cognitive ability from the display device, and (v) analyze the data input characterizing the patient's cognitive ability to determine a mentation parameter indicating the level of cognitive ability of the patient, and a wireless transmitter configured to receive the vital sign value and the mentation parameter from the processing component and transmit them over a wireless interface; and a remote computer configured to receive the vital sign value and the mentation parameter from the vital sign monitor.

2. The system of claim 1, wherein the third graphical user interface comprises the set of input fields configured to collect data input characterizing the patient's cognitive ability.

3. The system of claim 2, wherein the touchpanel display comprises a touch-sensitive area associated with each input field.

4. The system of claim 3, wherein the touch-sensitive area is configured to generate a digital signal after being contacted.

5. The system of claim 3, wherein each input field comprises a unique graphical icon.

6. The system of claim 2, wherein the set of input fields are organized to form a cognitive ability test.

7. The system of claim 6, wherein each input field comprises a numerical value, and the mentation test comprises an algorithm for determining if the input fields are contacted according to a numerical order corresponding to their numerical values.

8. The system of claim 1, wherein the vital sign monitor further comprises a communication system configured to receive audio information from the patient.

9. The system of claim 8, wherein the communication system comprises a microphone.

10. The system of claim 9, wherein the communication system comprises a speech-to-text converter configured to convert audio information received by the microphone into digital information.

11. The system of claim 10, wherein the microprocessor comprises a third interface to the communication system, the third interface configured to receive the digital information.

12. The system of claim 11, wherein the wireless transmitter is further configured to transmit the digital information to the remote computer.

13. The system of claim 1, wherein the remote computer comprises a display device.

14. The system of claim 13, wherein the remote computer is further configured to render both the vital sign value and a parameter determined from the mentation parameter on the display device.

15. The system of claim 1, wherein the remote computer comprises an alarming system.

16. The system of claim 15, wherein the alarming system is configured to generate an alarm if the mentation parameter or a parameter calculated therefrom exceeds a pre-determined threshold.

17. The system of claim 1, further comprising a flexible strap configured to attach the vital sign monitor the patient's arm.

UNITED STATES PATENT AND TRADEMARK OFFICE
CERTIFICATE OF CORRECTION

PATENT NO. : 8,727,977 B2
APPLICATION NO. : 12/762810
DATED : May 20, 2014
INVENTOR(S) : Matt Banet et al.


Page 1 of 1

It is certified that error appears in the above-identified patent and that said Letters Patent is hereby corrected as shown below:

On the Title page of the patent, inventor listed as “Gunner Trommer” should read
--Gunnar Trommer--.

On the Title page of the patent, inventor Jim Moon’s city address listed as “Porland” should read
--Portland--.

Signed and Sealed this
Tenth Day of February, 2015

A handwritten signature in black ink, reading "Michelle K. Lee". The signature is written in a cursive style with a large, stylized "M" and "L".

Michelle K. Lee
Deputy Director of the United States Patent and Trademark Office

专利名称(译)	体佩的生命体征监视器		
公开(公告)号	US8727977	公开(公告)日	2014-05-20
申请号	US12/762810	申请日	2010-04-19
[标]申请(专利权)人(译)	索泰拉无线公司		
申请(专利权)人(译)	SOTERA WIRELESS , INC.		
当前申请(专利权)人(译)	SOTERA WIRELESS , INC.		
[标]发明人	BANET MATT MOON JIM TROMMER GUNNER DIBATTISTA CAROLINE		
发明人	BANET, MATT MOON, JIM TROMMER, GUNNER DIBATTISTA, CAROLINE		
IPC分类号	A61B5/00		
CPC分类号	A61B5/00 A61B5/6824 A61B5/411 H04W4/206 A61B5/746 A61B5/0002 H04W4/02		
代理人(译)	WHITTAKER , MICHAEL A.		
优先权	61/312624 2010-03-10 US		
其他公开文献	US20110224557A1		
外部链接	Espacenet USPTO		

摘要(译)

本发明提供一种体戴式生命体征监测器，其测量患者的生命体征（例如血压，SpO2，心率，呼吸速率和温度），同时表征其活动状态（例如，休息，行走，抽搐，下降）和姿势。（直立，仰卧）。监视器处理此信息以最小化生命体征的损坏以及与运动相关的伪像的相关警报/警报。它还具有在触摸屏显示器上呈现的图形用户界面（GUI），有助于简化和改善医院和家庭中的患者监控和安全性的许多功能。

