



US009770183B2

(12) **United States Patent**
Sullivan

(10) **Patent No.:** **US 9,770,183 B2**
(45) **Date of Patent:** **Sep. 26, 2017**

(54) **ANALYZING ECG DATA IN DECIDING ON
PATIENT CHEST COMPRESSION
TREATMENT**

(75) Inventor: **Joseph L. Sullivan**, Kirkland, WA (US)

(73) Assignee: **Physio-Control, Inc.**, Redmond, WA
(US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 894 days.

(21) Appl. No.: **12/910,677**

(22) Filed: **Oct. 22, 2010**

(65) **Prior Publication Data**

US 2011/0082379 A1 Apr. 7, 2011

Related U.S. Application Data

(63) Continuation-in-part of application No. 12/572,691,
filed on Oct. 2, 2009, now Pat. No. 9,539,434.

(51) **Int. Cl.**
A61B 5/00 (2006.01)
A61B 5/0452 (2006.01)
A61N 1/39 (2006.01)

(52) **U.S. Cl.**
CPC *A61B 5/0452* (2013.01); *A61N 1/3925*
(2013.01)

(58) **Field of Classification Search**
CPC . A61N 2001/36; A61N 1/3925; A61B 5/0452
USPC 607/3-6; 600/510
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

2004/0267325 A1* 12/2004 Geheb et al. 607/5
2005/0137628 A1* 6/2005 Young et al. 607/5

* cited by examiner


Primary Examiner — Nadia A Mahmood

(74) *Attorney, Agent, or Firm* — Lane Powell PC

(57) **ABSTRACT**

Medical devices, software and methods are provided, for making a decision as to whether to pause patient chest compression treatment in connection with administering electric shock therapy to the patient. The decision is made depending whether signal spikes identified in the ECG data are determined to be QRS complexes, or merely likely impulsive artifact caused by the chest compressions.

20 Claims, 11 Drawing Sheets



DEFIBRILLATION SCENE





FIG. 1

DEFIBRILLATION SCENE

TYPE OF EXTERNAL DEFIBRILLATOR	INTENDED TO BE USED BY PERSONS:	
	IN THE MEDICAL PROFESSIONS	NOT IN THE MEDICAL PROFESSIONS
DEFIBRILLATOR – MONITOR	√	
AED	√	√


FIG. 2

TWO MAIN TYPES OF EXTERNAL DEFIBRILLATORS




COMPONENTS OF EXTERNAL DEFIBRILLATOR

FIG. 3




SAMPLE PATIENT ECG DATA

FIG. 4




SAMPLE PATIENT ECG DATA

FIG. 5

**FIG. 6**

SAMPLE ECG SIGNAL HAVING
IMPULSIVE ARTIFACTS AND NO
QRS COMPLEXES

**FIG. 7**

SAMPLE ECG SIGNAL HAVING
QRS COMPLEXES AND NO
IMPULSIVE ARTIFACTS

<u>800</u>	<u>891</u>	<u>892</u>	<u>893</u>	<u>894</u>	<u>895</u>	<u>896</u>
<u>810</u>	RECEIVED ECG WAVEFORM	PROPER PATIENT RHYTHM CLASSIFICATION	DESIRED OUTCOME OF SHOCK ADVISORY ALGORITHM	DETECTION OF SIGNAL SPIKES IN WAVEFORM (Y/N)	IF YES, ALGORITHM'S INTERPRETATION OF DETECTED SPIKES	OUTCOME OF SHOCK ADVISORY ALGORITHM DUE TO INTERPRETATION
<u>820</u>	QRSs & NON-IMPULSIVE ARTIFACT	NON-SHOCKABLE DUE TO QRSs	"NO SHOCK"	YES (FROM QRSs)	"QRSs"	"NO SHOCK ADVISED"
<u>830</u>	NO QRSs, NO IMPULSIVE ARTIFACT	VF, OR HIGH-RATE VT	"SHOCK"	NO (NO QRSs, NO IMPULSIVE ARTIFACT)	[N/A]	"SHOCK ADVISED"
<u>840</u>	NO QRSs, NO IMPULSIVE ARTIFACT	ASYSTOLE, ETC.	"NO SHOCK"	NO (NO QRSs, NO IMPULSIVE ARTIFACT)	[N/A]	"NO SHOCK ADVISED"
<u>850</u>	DOMINANT IMPULSIVE ARTIFACT	VF, OR HIGH-RATE VT	"SHOCK"	YES (FROM IMPULSIVE ARTIFACT)	"QRSs"	"NO SHOCK ADVISED"
<u>860</u>	DOMINANT IMPULSIVE ARTIFACT	ASYSTOLE, ETC.	"NO SHOCK"	YES (FROM IMPULSIVE ARTIFACT)	"QRSs"	"NO SHOCK ADVISED"






FIG. 8



METHODS

FIG. 9

1040METHODS**FIG. 10**




FIG. 11

SAMPLE POSITIONS OF SIGNAL SPIKES
AND SETTING OF WINDOW WITHIN CYCLE
INTERVAL FOR ECG DATA HAVING QRS
COMPLEXES




FIG. 12

SAMPLE POSITIONS OF SIGNAL SPIKES
AND SETTING OF WINDOW WITHIN CYCLE
INTERVAL FOR ECG DATA HAVING
IMPULSIVE ARTIFACTS AND NO QRS
COMPLEXES

<u>1300</u>	<u>1391</u>	<u>1392</u>	<u>1393</u>	<u>1394</u>	<u>1395</u>	<u>1396</u>
<u>1310</u>	RECEIVED ECG WAVEFORM	PROPER PATIENT RHYTHM CLASSIFICATION	DESIRED OUTCOME OF SHOCK ADVISORY ALGORITHM	DETECTION OF SIGNAL SPIKES IN WAVEFORM (Y/N)	IF YES, ALGORITHM'S INTERPRETATION OF DETECTED SPIKES	OUTCOME OF SHOCK ADVISORY ALGORITHM DUE TO INTERPRETATION
<u>1320</u>	QRSs & NON-IMPULSIVE ARTIFACT	NON-SHOCKABLE DUE TO QRSs	"NO SHOCK"	YES (FROM QRSs)	"QRSs"	"NO SHOCK ADVISED"
<u>1330</u>	NO QRSs, NO IMPULSIVE ARTIFACT	VF, OR HIGH-RATE VT	"SHOCK"	NO (NO QRSs, NO IMPULSIVE ARTIFACT)	[N/A]	"SHOCK ADVISED"
<u>1340</u>	NO QRSs, NO IMPULSIVE ARTIFACT	ASYSTOLE, ETC.	"NO SHOCK"	NO (NO QRSs, NO IMPULSIVE ARTIFACT)	[N/A]	"NO SHOCK ADVISED"
<u>1350</u>	DOMINANT IMPULSIVE ARTIFACT	VF, OR HIGH-RATE VT	"SHOCK"	YES (FROM IMPULSIVE ARTIFACT)	"IMPULSIVE ARTIFACT"	"PAUSE CPR FOR CLEAN ANALYSIS"
<u>1360</u>	DOMINANT IMPULSIVE ARTIFACT	ASYSTOLE, ETC.	"NO SHOCK"	YES (FROM IMPULSIVE ARTIFACT)	"IMPULSIVE ARTIFACT"	"PAUSE CPR FOR CLEAN ANALYSIS"

13581368

FIG. 13

1

ANALYZING ECG DATA IN DECIDING ON PATIENT CHEST COMPRESSION TREATMENT

CROSS-REFERENCE TO RELATED APPLICATION

The present application is a continuation-in-part of U.S. patent application Ser. No. 12/572,691, filed Oct. 2, 2009, which is hereby incorporated by reference in its entirety.

FIELD

This invention generally relates to deciding on patient chest compression treatment in connection with patient electric shock therapy.

BACKGROUND

In humans, the heart beats to sustain life. In normal operation, it pumps blood through the various parts of the body. More particularly, the various chamber of the heart contract and expand in a periodic and coordinated fashion, which causes the blood to be pumped regularly. More specifically, the right atrium sends deoxygenated blood into the right ventricle. The right ventricle pumps the blood to the lungs, where it becomes oxygenated, and from where it returns to the left atrium. The left atrium pumps the oxygenated blood to the left ventricle. The left ventricle, then, expels the blood, forcing it to circulate to the various parts of the body.

The heart chambers pump because of the heart's electrical control system. More particularly, the sinoatrial (SA) node generates an electrical impulse, which generates further electrical signals. These further signals cause the above-described contractions of the various chambers in the heart, in the right sequence. The electrical pattern created by the sinoatrial (SA) node is called a sinus rhythm.

Sometimes, however, the electrical control system of the heart malfunctions, which can cause the heart to beat irregularly, or not at all. The cardiac rhythm is then generally called an arrhythmia, and some of it may be caused by electrical activity from locations in the heart other than the SA node. Some types of arrhythmia may result in inadequate blood flow, thus reducing the amount of blood pumped to the various parts of the body. Some arrhythmias may even result in a Sudden Cardiac Arrest (SCA). In a SCA, the heart fails to pump blood effectively, and death can occur. In fact, it is estimated that SCA results in more than 250,000 deaths per year in the United States alone. Further, a SCA may result from a condition other than an arrhythmia.

One type of arrhythmia associated with SCA is known as Ventricular Fibrillation (VF). VF is a type of malfunction where the ventricles make rapid, uncoordinated movements, instead of the normal contractions. When that happens, the heart does not pump enough blood. The person's condition will deteriorate rapidly and, if not reversed in time, they will die soon, e.g. within ten minutes.

Ventricular Fibrillation can often be reversed using a life-saving device called a defibrillator. A defibrillator, if applied properly, can administer an electrical shock to the heart. The shock may terminate the VF, thus giving the heart the opportunity to resume pumping blood. If VF is not terminated, the shock may be repeated, often at escalating energies.

A challenge with defibrillation is that the electrical shock must be administered very soon after the onset of VF. There

2

is not much time: the survival rate of persons suffering from VF decreases by about 10% for each minute the administration of a defibrillation shock is delayed. After about 10 minutes the rate of survival for SCA victims averages less than 2%.

The challenge of defibrillating early after the onset of VF is being met in a number of ways. First, for some people who are considered to be at a higher risk of VF, an Implantable Cardioverter Defibrillator (ICD) can be implanted surgically. An ICD can monitor the person's heart, and administer an electrical shock as needed. As such, an ICD reduces the need to have the higher-risk person be monitored constantly by medical personnel.

Regardless, VF can occur unpredictably, even to a person who is not considered at risk. As such, VF can be experienced by many people who lack the benefit of ICD therapy. When VF occurs to a person who does not have an ICD, they collapse, because blood flow has stopped. They should receive therapy quickly.

For a VF victim without an ICD, a different type of defibrillator can be used, which is called an external defibrillator. External defibrillators have been made portable, so they can be brought to a potential VF victim quickly enough to revive them.

During VF, the person's condition deteriorates, because the blood is not flowing to the brain, heart, lungs, and other organs. Blood flow must be restored, if resuscitation attempts are to be successful.

Cardiopulmonary Resuscitation (CPR) is one method of forcing blood flow in a person experiencing cardiac arrest. In addition, CPR is the primary recommended treatment for some patients with some kinds of non-VF cardiac arrest, such as asystole and pulseless electrical activity (PEA). CPR is a combination of techniques that include chest compressions to force blood circulation, and rescue breathing to force respiration.

Properly administered CPR provides oxygenated blood to critical organs of a person in cardiac arrest, thereby minimizing the deterioration that would otherwise occur. As such, CPR can be beneficial for persons experiencing VF, because it slows the deterioration that would otherwise occur while a defibrillator is being retrieved. Indeed, for patients with an extended down-time, survival rates are higher if CPR is administered prior to defibrillation.

It is desired to improve patient outcomes, by making improved decisions of when to administer therapy, such as electrical shocks, CPR, pharmaceuticals, etc. Patient outcomes are sometimes analyzed in post-event review.

BRIEF SUMMARY

The present description gives instances of medical devices, software and methods, the use of which may help overcome problems and limitations of the prior art.

In some embodiments, a decision as to whether to pause chest compressions being administered to a patient is made differently, depending on whether signal spikes identified in the ECG data are determined to be QRS complexes or merely likely impulsive artifact caused by the chest compressions.

Embodiments include medical devices such as defibrillators and pacers, and also processors, computers, software, and methods of making the decision.

An advantage over the prior art is that patient outcomes can be improved, especially since impulsive artifacts resembling signal spikes are not misdiagnosed as QRS complexes.

These and other features and advantages of this description will become more readily apparent from the following Detailed Description, which proceeds with reference to the drawings, in which:

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 is a diagram of a scene where an external defibrillator is used to save the life of a person according to embodiments.

FIG. 2 is a table listing two main types of the external defibrillator shown in FIG. 1, and who they might be used by.

FIG. 3 is a diagram showing components of an external defibrillator, such as the one shown in FIG. 1, which is made according to embodiments.

FIGS. 4 and 5 are time diagrams of patient ECG data in the form of signals.

FIG. 6 is a time diagram of an ECG signal having impulsive artifacts and no QRS complexes.

FIG. 7 is a time diagram of an ECG signal having QRS complexes and no impulsive artifacts.

FIG. 8 is a table listing various decisions based on analysis of a detected ECG signal having QRS complexes, impulsive artifacts, or both, and further showing how a device in the prior art might misdiagnose an impulsive artifact as a QRS complex, and not administer a shock where otherwise a shock was needed.

FIGS. 9 and 10 are flowcharts for illustrating methods according to embodiments.

FIGS. 11 and 12 are diagrams for showing the arranging of signal peaks and setting of a window within a cycle interval.

FIG. 13 is a table listing various decisions based on analysis of a detected ECG signal having QRS complexes, impulsive artifacts, or both according to embodiments.

DETAILED DESCRIPTION

As has been mentioned, the present description is about making a decision of whether electric therapy should be administered or not. Embodiments include medical devices that can administer electrical therapy, such as defibrillators, pacers, etc. Examples are now described.

FIG. 1 is a diagram of a defibrillation scene. A person 82 is lying on their back. Person 82 could be a patient in a hospital, or someone found unconscious, and then turned to be on their back. Person 82 is experiencing a condition in their heart 85, which could be Ventricular Fibrillation (VF).

A portable external defibrillator 100 has been brought close to person 82. At least two defibrillation electrodes 104, 108 are usually provided with external defibrillator 100, and are sometimes called electrodes 104, 108. Electrodes 104, 108 are coupled with external defibrillator 100 via respective electrode leads 105, 109. A rescuer (not shown) has attached electrodes 104, 108 to the skin of person 82. Defibrillator 100 is administering, via electrodes 104, 108, a brief, strong electric pulse 111 through the body of person 82. Pulse 111, also known as a defibrillation shock, goes also through heart 85, in an attempt to restart it, for saving the life of person 82.

Defibrillator 100 can be one of different types, each with different sets of features and capabilities. The set of capabilities of defibrillator 100 is determined by planning who would use it, and what training they would be likely to have. Examples are now described.

FIG. 2 is a table listing two main types of external defibrillators, and who they are primarily intended to be used

by. A first type of defibrillator 100 is generally called a defibrillator-monitor, because it is typically formed as a unit with a patient monitor. A defibrillator-monitor is intended to be used by persons in the medical professions, such as doctors, nurses, paramedics, emergency medical technicians, etc. Such a defibrillator-monitor is intended to be used in a pre-hospital or hospital scenario.

As a defibrillator, the device can be one of different varieties, or even versatile enough to be able to switch among different modes that individually correspond to the varieties. One variety is that of an automated defibrillator, which can determine whether a shock is needed and, if so, charge to a predetermined energy level and instruct the user to administer the shock. Another variety is that of a manual defibrillator, where the user determines the need and controls administering the shock.

As a patient monitor, the device has features additional to what is minimally needed for mere operation as a defibrillator. These features can be for monitoring physiological signals of a person in an emergency scenario. For example, these signals can include a person's full ECG (electrocardiogram) signals. Additionally, these signals can be about the person's temperature, non-invasive blood pressure (NIBP), arterial oxygen saturation/pulse oximetry (SpO2), the concentration or partial pressure of carbon dioxide in the respiratory gases, which is also known as capnography, and so on.

A second type of external defibrillator 100 is generally called an AED, which stands for "Automated External Defibrillator". An AED typically makes the shock/no shock determination by itself, automatically. Indeed, it can sense enough physiological conditions of the person 82 via only the shown defibrillation electrodes 104, 108 of FIG. 1. In its present embodiments, an AED can either administer the shock automatically, or instruct the user to do so, e.g. by pushing a button. Being of a much simpler construction, an AED typically costs much less than a defibrillator-monitor. As such, it makes sense for a hospital, for example, to deploy AEDs at its various floors, in case the more expensive defibrillator-monitor is at an Intensive Care Unit, and so on.

AEDs, however, can also be used by people who are not in the medical profession. More particularly, an AED can be used by many professional first responders, such as policemen, firemen, etc. Even a person with only first-aid training can use one. And AEDs increasingly can supply instructions to whoever is using them.

AEDs are thus particularly useful, because it is so critical to respond quickly, when a person suffers from VF. Indeed, the people who will first reach the VF sufferer may not be in the medical professions.

Increasing awareness has resulted in AEDs being deployed in public or semi-public spaces, so that even a member of the public can use one, if they have obtained first aid and CPR/AED training on their own initiative. This way, defibrillation can be administered soon enough after the onset of VF, to hopefully be effective in rescuing the person.

There are additional types of external defibrillators, which are not listed in FIG. 2. For example, a hybrid defibrillator can have aspects of an AED, and also of a defibrillator-monitor. A usual such aspect is additional ECG monitoring capability.

FIG. 3 is a diagram showing components of an external defibrillator 300 made according to embodiments. These components can be, for example, in external defibrillator 100 of FIG. 1.

External defibrillator 300 is intended for use by a user 380, who would be the rescuer. Defibrillator 300 typically

includes a defibrillation port **310**, such as a socket. Defibrillation port **310** includes nodes **314**, **318**. Defibrillation electrodes **304**, **308**, which can be similar to electrodes **104**, **108**, can be plugged in defibrillation port **310**, so as to make electrical contact with nodes **314**, **318**, respectively. It is also possible that electrodes can be connected continuously to defibrillation port **310**, etc. Either way, defibrillation port **310** can be used for guiding via electrodes to person **82** an electrical charge that has been stored in defibrillator **300**, as will be seen later in this document.

If defibrillator **300** is actually a defibrillator-monitor, as was described with reference to FIG. 2, then it will typically also have an ECG port **319**, for plugging in ECG leads **309**. ECG leads **309** can sense a full ECG signal. Moreover, a defibrillator-monitor could have additional ports (not shown), and an other component **325** for the above described additional features.

Defibrillator **300** also includes a measurement circuit **320**. Measurement circuit **320** receives physiological signals from ECG port **319**, and also from other ports, if provided. These physiological signals are sensed, and information about them is rendered by circuit **320** as data, or other signals, etc.

If defibrillator **300** is actually an AED, it may lack ECG port **319**. Measurement circuit **320** can obtain physiological signals through nodes **314**, **318** instead, when defibrillation electrodes **304**, **308** are attached to person **82**. In these cases, a person's ECG signal can be sensed as a voltage difference between electrodes **304**, **308**. Plus, impedance between electrodes **304**, **308** can be sensed for detecting, among other things, whether these electrodes **304**, **308** have been inadvertently disconnected from the person.

Defibrillator **300** also includes a processor **330**. Processor **330** may be implemented in any number of ways. Such ways include, by way of example and not of limitation, digital and/or analog processors such as microprocessors and digital-signal processors (DSPs); controllers such as microcontrollers; software running in a machine; programmable circuits such as Field Programmable Gate Arrays (FPGAs), Field-Programmable Analog Arrays (FPAAs), Programmable Logic Devices (PLDs), Application Specific Integrated Circuits (ASICs), any combination of one or more of these, and so on.

Processor **330** can be considered to have a number of modules. One such module can be a detection module **332**, which senses outputs of measurement circuit **320**. Detection module **332** can include a VF detector. Thus, the person's sensed ECG can be used to determine whether the person is experiencing VF.

Another such module in processor **330** can be an advice module **334**, which arrives at advice based on outputs of detection module **332**. Advice module **334** can include a Shock Advisory Algorithm, implement decision rules, and so on. The advice can be to shock, to not shock, to administer other forms of therapy, and so on. If the advice is to shock, some external defibrillator embodiments merely report that to the user, and prompt them to do it. Other embodiments further execute the advice, by administering the shock. If the advice is to administer CPR, defibrillator **300** may further issue prompts for it, and so on.

Processor **330** can include additional modules, such as module **336**, for other functions. In addition, if other component **325** is indeed provided, it may be operated in part by processor **330**, etc.

Defibrillator **300** optionally further includes a memory **338**, which can work together with processor **330**. Memory **338** may be implemented in any number of ways. Such ways

include, by way of example and not of limitation, nonvolatile memories (NVM), read-only memories (ROM), random access memories (RAM), any combination of these, and so on. Memory **338**, if provided, can include programs for processor **330**, and so on. The programs can be operational for the inherent needs of processor **330**, and can also include protocols and ways that decisions can be made by advice module **334**. In addition, memory **338** can store prompts for user **380**, etc.

Defibrillator **300** may also include a power source **340**. To enable portability of defibrillator **300**, power source **340** typically includes a battery. Such a battery is typically implemented as a battery pack, which can be rechargeable or not. Sometimes, a combination is used, of rechargeable and non-rechargeable battery packs. Other embodiments of power source **340** can include AC power override, for where AC power will be available, and so on. In some embodiments, power source **340** is controlled by processor **330**.

Defibrillator **300** additionally includes an energy storage module **350**. Module **350** is where some electrical energy is stored, when preparing it for sudden discharge to administer a shock. Module **350** can be charged from power source **340** to the right amount of energy, as controlled by processor **330**. In typical implementations, module **350** includes one or more capacitors **352**, and so on.

Defibrillator **300** moreover includes a discharge circuit **355**. Circuit **355** can be controlled to permit the energy stored in module **350** to be discharged to nodes **314**, **318**, and thus also to defibrillation electrodes **304**, **308**. Circuit **355** can include one or more switches **357**. Those can be made in a number of ways, such as by an H-bridge, and so on.

Defibrillator **300** further includes a user interface **370** for user **380**. User interface **370** can be made in any number of ways. For example, interface **370** may include a screen, to display what is detected and measured, provide visual feedback to the rescuer for their resuscitation attempts, and so on. Interface **370** may also include a speaker, to issue voice prompts, etc. Interface **370** may additionally include various controls, such as pushbuttons, keyboards, and so on. In addition, discharge circuit **355** can be controlled by processor **330**, or directly by user **380** via user interface **370**, and so on.

Defibrillator **300** can optionally include other components. For example, a communication module **390** may be provided for communicating with other machines. Such communication can be performed wirelessly, or via wire, or by infrared communication, and so on. This way, data can be communicated, such as patient data, incident information, therapy attempted, CPR performance, and so on.

An additional feature of a defibrillator can be CPR-prompting. Prompts are issued to the user, visual or by sound, so that the user can administer CPR. Examples are taught in U.S. Pat. No. 6,334,070 and No. 6,356,785.

FIGS. 4 and 5 are time diagrams of patient ECG data in the form of signals. The ECG data exhibits an impulsive waveform having signal spikes or peaks that include both positive peaks **405** and negative peaks **410**. The time diagrams of FIGS. 4 and 5 also display cycle intervals **415** and **515**, respectively, that each correspond to chest compressions that are delivered to the patient,

The cycle intervals **415** illustrated in FIG. 4 are positioned such that each is substantially centered around an identified chest compression. The boundaries of each cycle interval **415** in FIG. 4 are thus each situated substantially midway

between the identified chest compression and the next closest chest compression on either side of the identified chest compression.

FIG. 5 shows the same ECG data presented in FIG. 4, including positive peaks 405 and negative peaks 410, but with different cycle intervals 515 that are positioned such that each is situated substantially between two successive identified chest compressions. Thus, the boundaries of each cycle interval 515 in FIG. 5 are each situated near or at one of the two successive identified chest compressions. Because the underlying ECG signal is the same in both FIGS. 4 and 5, the cycle intervals 415 and 515 in FIGS. 4 and 5, respectively, differ only in time relative to the ECG signal. The effects of such a time shift are illustrated below.

FIG. 6 is a time diagram of an ECG signal illustrating example signal spikes that represent impulsive signal artifacts and not QRS complexes. For example, the patient receiving chest compressions may be in a state of asystole such that the signal spikes represent only impulsive artifacts resulting from physical delivery of the chest compressions to the patient. The impulsive artifacts in the ECG signal may have both positive peaks and negative peaks.

FIG. 7 is a time diagram of an ECG signal illustrating example QRS complexes and no impulsive artifacts. The QRS complexes generally include both positive peaks and negative peaks.

As can be readily discerned by considering the two figures together, there are a number of similarities between the ECG signals of FIGS. 6 and 7. For example, the positive peaks are relatively closer in time to the corresponding negative peaks than to another positive peak. Also, the pattern of recurrence is similar. Consequently, current systems often have trouble discerning ECG signals that have impulsive artifacts and no QRS complexes from ECG signals that have QRS complexes and no impulsive artifacts. It follows that such current systems may generate a non-optimal decision regarding patient treatment based on mis-identification, as discussed below.

FIG. 8 is a table 800 listing various decisions based on analysis of a detected ECG signal having QRS complexes or impulsive artifacts, and further showing how a device in the prior art might misdiagnose an impulsive artifact as a QRS complex and, as a result, not advise a rescuer to administer a shock in situations where a shock is needed, for example. Impulsive artifacts have been particularly problematic for prior art devices because such devices are unable to distinguish between ECG signals having QRS complexes and VF signals having impulsive artifacts, for example. In addition, an ECG waveform jitter of only a few samples is often enough to result in a prior art device rendering a “no shock” decision in situations where an electric shock should actually be delivered for best patient treatment, as discussed below.

Row 810 of the table 800 provides header information corresponding to each of columns 891-896.

Row 820 pertains to situations in which a received ECG waveform contains QRS signals and non-impulsive artifacts, as indicated at column 891. Column 892 indicates that, in such situations, a proper patient rhythm classification would have the patient be deemed “not shockable” and, therefore, a “no shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at column 893.

Current systems will identify signal spikes in the received ECG waveform, i.e., due to the presence of the QRS complexes, and properly interpret the signal spikes as corresponding to the QRS complexes, as indicated at columns 894 and 895, respectively. In current systems, the outcome

of such a shock advisory algorithm due to the interpretation would be the issuing of a “no shock” prompt to the rescuer, as indicated at column 896.

Rows 830 and 840 pertain to situations in which a received ECG waveform contains neither QRS signals nor impulsive artifacts, as indicated at column 891. In such situations, a proper patient rhythm classification would provide that the patient is experiencing ventricular fibrillation (VF) or high-rate ventricular tachycardia (VT), as indicated at row 830, column 892, or asystole, as indicated at row 840, column 892. If the patient is experiencing VF or high-rate VT, a “shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at row 830, column 893. If the patient is experiencing asystole, however, a “no shock” prompt would be the desired outcome of such a shock advisory algorithm, as indicated at row 840, column 893.

Current systems will identify no signal spikes in the received ECG waveform, i.e., due to the absence of QRS complexes and impulsive artifacts, as indicated at column 894 for both rows 830 and 840. In current systems, the outcome of such a shock advisory algorithm due to the interpretation would be the issuing of a “shock advised” prompt if the patient is experiencing VF or high-rate VT, as indicated at row 830, column 896, or a “no shock advised” prompt if the patient is experiencing asystole, as indicated at row 840, column 896.

Rows 850 and 860 pertain to situations in which a received ECG waveform contains a dominant impulsive artifact, i.e., resulting from administration of chest compressions to the patient, as indicated at column 891. In such situations, a proper patient rhythm classification would provide that the patient is experiencing VF or high-rate VT, as indicated at row 850, column 892, or asystole, as indicated at row 860, column 892. If the patient is experiencing VF or high-rate VT, a “shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at row 850, column 893. If the patient is experiencing asystole, however, a “no shock” prompt would be the desired outcome of such a shock advisory algorithm, as indicated at row 860, column 893.

Current systems will identify signal spikes in the received ECG waveform, i.e., due to the presence of the dominant impulsive artifact, as indicated at column 894 for both rows 850 and 860. In situations where the patient is experiencing VF or high-rate VT, a shock advisory algorithm of current systems will incorrectly interpret the detected signal spikes as QRS complexes and, therefore, incorrectly issue a “no shock advised” prompt, as indicated at row 850, column 895, and row 860, column 896, respectively. Consequently, patients in these situations will not receive the desired outcome, i.e., administration of an electric shock, due to the erroneous interpretation of the signal spikes.

In situations where the patient is experiencing asystole, such a shock advisory algorithm will still incorrectly interpret the detected signal spikes as QRS complexes and thus issue a “no shock advised” prompt, as indicated at row 860, column 895, and row 860, column 896, respectively. While consistent with the desired outcome at row 860, column 893, the “no shock advised” prompt issued at row 860, column 896 is based on the incorrect interpretation of the detected signal spikes as corresponding to QRS complexes when, in actuality, there are no QRS complexes in the received ECG waveform because the patient is experiencing asystole.

The functions of this description may be implemented by one or more devices that include logic circuitry. The device performs functions and/or methods as are described in this

document. The logic circuitry may include a processor that may be programmable for a general purpose, or dedicated, such as microcontroller, a microprocessor, a Digital Signal Processor (DSP), etc. For example, the device may be a digital computer like device, such as a general-purpose computer selectively activated or reconfigured by a computer program stored in the computer. Alternately, the device may be implemented by an Application Specific Integrated Circuit (ASIC), etc.

Moreover, methods are described below. The methods and algorithms presented herein are not necessarily inherently associated with any particular computer or other apparatus. Rather, various general-purpose machines may be used with programs in accordance with the teachings herein, or it may prove more convenient to construct more specialized apparatus to perform the required method steps. The required structure for a variety of these machines will become apparent from this description.

In all cases there should be borne in mind the distinction between methods in this description, and the method of operating a computing machine. This description relates both to methods in general, and also to steps for operating a computer and for processing electrical or other physical signals to generate other desired physical signals.

Programs are additionally included in this description, as are methods of operation of the programs. A program is generally defined as a group of steps leading to a desired result, due to their nature and their sequence. A program is usually advantageously implemented as a program for a computing machine, such as a general-purpose computer, a special purpose computer, a microprocessor, etc.

Storage media are additionally included in this description. Such media, individually or in combination with others, have stored thereon instructions of a program made according to the invention. A storage medium according to the invention is a computer-readable medium, such as a memory, and is read by the computing machine mentioned above.

Performing the steps or instructions of a program requires physical manipulations of physical quantities. Usually, though not necessarily, these quantities may be transferred, combined, compared, and otherwise manipulated or processed according to the instructions, and they may also be stored in a computer-readable medium. These quantities include, for example electrical, magnetic, and electromagnetic signals, and also states of matter that can be queried by such signals. It is convenient at times, principally for reasons of common usage, to refer to these quantities as bits, data bits, samples, values, symbols, characters, images, terms, numbers, or the like. It should be borne in mind, however, that all of these and similar terms are associated with the appropriate physical quantities, and that these terms are merely convenient labels applied to these physical quantities, individually or in groups.

This detailed description is presented largely in terms of flowcharts, display images, algorithms, and symbolic representations of operations of data bits within at least one computer readable medium, such as a memory. Indeed, such descriptions and representations are the type of convenient labels used by those skilled in programming and/or the data processing arts to effectively convey the substance of their work to others skilled in the art. A person skilled in the art of programming may use these descriptions to readily generate specific instructions for implementing a program according to the present invention.

Often, for the sake of convenience only, it is preferred to implement and describe a program as various interconnected

distinct software modules or features, individually and collectively also known as software. This is not necessary, however, and there may be cases where modules are equivalently aggregated into a single program with unclear boundaries. In any event, the software modules or features of this description may be implemented by themselves, or in combination with others. Even though it is said that the program may be stored in a computer-readable medium, it should be clear to a person skilled in the art that it need not be a single memory, or even a single machine. Various portions, modules or features of it may reside in separate memories, or even separate machines. The separate machines may be connected directly, or through a network, such as a local access network (LAN), or a global network, such as the Internet.

It will be appreciated that some of these methods may include software steps which may be performed by different modules of an overall software architecture. For example, data forwarding in a router may be performed in a data plane, which consults a local routing table. Collection of performance data may also be performed in a data plane. The performance data may be processed in a control plane, which accordingly may update the local routing table, in addition to neighboring ones. A person skilled in the art will discern which step is best performed in which plane.

An economy is achieved in the present document in that a single set of flowcharts is used to describe both programs, and also methods. So, while flowcharts are described in terms of boxes, they can mean both method and programs.

For this description, the methods may be implemented by machine operations. In other words, embodiments of programs are made such that they perform methods of the invention that are described in this document. These may be optionally performed in conjunction with one or more human operators performing some, but not all of them. As per the above, the users need not be collocated with each other, but each only with a machine that houses a portion of the program. Alternately, some of these machines may operate automatically, without users and/or independently from each other.

Methods are now described.

FIGS. 9 and 10 show a flowchart 900 for describing methods according to embodiments. The methods of flowchart 900 may be practiced by systems, devices, and software according to embodiments. For example, the methods illustrated by flowchart 900 can be performed by the processor 330 illustrated in FIG. 3.

Chest compressions may be delivered to a patient by one or both of a rescuer and a chest compression machine controlled by the rescuer. A detection module, such as the detection module 332 illustrated in FIG. 3, may identify the occurrence of chest compressions, as shown at 910. For example, the occurrence of the chest compressions may be identified by one or both of an impedance signal of the patient and an input signal of a chest compression machine. Chest compression markers may be used in connection with a patient's ECG data to identify each chest compression being administered to the patient over a certain period of time. The detection module may also receive ECG data for the patient receiving the chest compressions, as shown at 920.

A module such as the detection module may identify signal spikes in the ECG data, as shown at 930. The module may then extract a grouping statistic about the timing of the signal spikes relative to the occurrence of the identified chest compressions, as shown at 940 and further illustrated by flowchart 1040 of FIG. 10. The grouping statistic is an effort

to identify whether the signal spikes are correlated or aligned with each other and/or with the compression cycle. The grouping statistic can be extracted in any number of ways, for example measuring distances, squares of distances, and so on. In certain embodiments, the grouping statistic may include a number of measured distances between an identified chest compression and certain identified signal spikes. For example, the module may generate a sorted list that ranks or orders the signal spikes by the corresponding measured distances. Alternatively, or in addition, the grouping statistic may include information pertaining to measured distances between an identified signal spike and some or all of the other identified signal spikes.

The detection module or another module, such as the advice module 334 or other module 336 illustrated in FIG. 3, may then compare the grouping statistic to a threshold. For example, the advice module 334 may determine whether the grouping statistic is less than a particular threshold, as shown at 950.

Responsive to a first outcome yielded by the comparing, such as a determination at 950 that the grouping statistic is less than the threshold, a module such as the advice module may cause a prompt to be issued for advising the rescuer to either administer or not administer a shock, based on the underlying algorithm, as shown at 960. For example, if the identified signal spikes indicate a condition of VF or high-rate VT, the module may cause a "shock" prompt to be issued for advising the rescuer to administer an electric shock to the patient due to the VF or VT. Alternatively, if the identified signal spikes indicate a condition of asystole, the module may cause a "no shock" prompt to be issued for advising the rescuer to not administer an electric shock to the patient due to the asystole.

Responsive to a second outcome yielded by the comparing, such as a determination at 950 that the grouping statistic is not less than the threshold, the module may cause a prompt to be issued for advising the rescuer to pause the chest compressions and wait for a clean signal, as shown at 970, while returning to 920, regardless of whether the patient is currently experiencing VF, VT, or asystole.

The flowchart 1040 of FIG. 10 describes that a cycle interval is established for the chest compressions being administered to the patient, as shown at 1042. The cycle interval may be established by centering an identified chest compression within the cycle interval. In these embodiments, each boundary of the cycle interval is thus situated approximately substantially halfway between the centered chest compression and the next closest chest compression on either side of the centered chest compression. Alternatively, the module may determine the cycle interval by determining a time distance between two successive identified chest compressions. In these embodiments, the boundaries of the cycle interval will generally correspond to successive identified chest compressions.

A window may be set within a portion of the cycle interval, as shown at 1044. In certain embodiments, the window may be based on a percentage of a width of the cycle interval. In situations where the cycle interval is determined based on a time distance between two successive identified chest compressions, the window may include two non-adjacent sub-windows within the cycle interval. The window may be static or, in certain embodiments, dynamically adjustable. Further, the window may be checked and adjusted after a specified period of time.

Respective positions of at least some of the signal spikes identified within the cycle interval may be determined, as shown at 1046 in FIG. 10 and further illustrated in FIGS. 11

and 12, discussed below. The position of each of the identified signal spikes within the cycle interval may be based on a distance in time between the corresponding signal spike and an identified chest compression. The position of each of the identified signal spikes within the cycle interval may also be determined by determining an absolute peak value of the corresponding signal spike. In certain embodiments, the window may be adjusted after the positioning of each identified signal peak or after a certain number of signal peaks have been positioned within the cycle interval. For example, the width of the window may be dependent upon the number of signal peaks positioned within the cycle interval.

A ratio indicating a percentage of the signal spikes that have a position within the window set at 1044 is determined, as shown at 1048 in FIG. 10 and further illustrated in FIGS. 11 and 12, discussed below. In determining the ratio, a numerical determination is made as to how many of the signal spikes within the cycle interval are further positioned within the window. For example, if seventy-five percent of the signal spikes have a position within the window, the determined ratio would indicate such a percentage. This ratio, or grouping statistic, may then be compared to a threshold, as shown at 950 of FIG. 9, discussed above. The threshold may be an arbitrary value, such as a fixed percentage of the cycle interval, or it may be calculated in any of a number of different ways. Further, the threshold may be adjusted over time, either manually, automatically, or both.

In certain embodiments, the cycle interval may be scaled, normalized, or both before, during, or after either of establishing the cycle interval at 1042 and positioning the signal spikes within the cycle interval at 1046.

FIGS. 11 and 12 are diagrams for showing the arranging of signal peaks and setting of a window within a cycle interval. The cycle intervals shown in FIGS. 11 and 12 each correspond to a mapping of ECG slope peak amplitude in the y-axis to time in the x-axis. In other words, for each signal spike in ECG data received from a patient receiving chest compressions, a point is positioned within the cycle interval that corresponds to the value of the signal peak in the y-axis and corresponds to the time of occurrence in the x-axis. In FIGS. 11 and 12, the value of each signal peak is an absolute value, though other embodiments may involve the actual value of the signal peaks. In certain embodiments, the time of occurrence for a point may be relative to a corresponding chest compression, such as the chest compression closest in time to the point. For example, in situations where signal spikes in the ECG signal generally correspond to chest compressions, the corresponding points will generally be positioned in relatively close proximity to each other along the x-axis, and possibly along the y-axis as well, particularly if the rescuer applies chest compressions to the patient in a generally uniform manner.

FIG. 11 illustrates a situation in which the patient ECG data has QRS complexes and no impulsive artifacts. Accordingly, the number of signal spike points positioned outside of the cycle interval window 1105 is significant and may even exceed the number of signal spike points positioned within the cycle interval window 1105. As a result, the ratio of the number of signal spike points positioned within the cycle interval window 1105 to the number of signal spike points not positioned within the cycle interval window 1105 is relatively low, e.g., below a certain threshold. This may happen in such situations because cardiac activity is typically asynchronous to chest compressions. Therefore, signal

spike points corresponding to the QRS complexes will generally be positioned at relatively random locations within the cycle interval.

FIG. 12 illustrates a situation in which the patient ECG data has impulsive artifacts and no QRS complexes. Accordingly, the number of signal spike points positioned within the cycle interval window 1205 significantly exceeds the number of signal spike points positioned outside of the cycle interval window 1105, if any. Consequently, the ratio of the number of signal spike peaks positioned within the cycle interval window 1205 to the number of signal spike peaks not positioned within the cycle interval window 1205 is relatively high, e.g., above the threshold. In the example illustrated in FIG. 12, there are in fact no signal spike peaks that are not within the cycle interval window 1205. This may happen in situations where the patient is experiencing asystole during the ECG analysis.

The cycle interval windows 1105 and 1205 illustrated in FIGS. 11 and 12, respectively, may be adjusted. For example, a window may be established before or as soon as one or more signal spike points are positioned within a cycle interval and, once established, the window may be re-sized, re-positioned, etc. Once a certain number of signal spike points have been positioned within the cycle interval, the window may be adjusted accordingly. For example, if a high percentage of the signal spike points are positioned in relatively close proximity to each other, at least in the x-axis, the width of the window may be decreased proportionately. Windows having a smaller width will generally provide greater accuracy than windows having a greater width because the determinations based thereon will have a correspondingly smaller margin of error.

It should be noted that FIGS. 11 and 12 are provided to illustrate visual representations of cycle intervals and cycle interval windows as used and applied herein. In certain embodiments, such data manipulations might not be performed in such a manner that a visual representation thereof is made available to the rescuer or anyone else. Alternatively, certain embodiments may include providing a visual representation of such information to the rescuer at the time of applying chest compressions to the patient or to others at a later time.

FIG. 13 is a table 1300 listing various decisions based on analysis of a detected ECG signal having QRS complexes, impulsive artifacts, or both according to embodiments.

Row 1310 of the table 1300 provides header information corresponding to each of columns 1391-1396.

Row 1320 pertains to situations in which a received ECG waveform contains QRS signals and non-impulsive artifacts, as indicated at column 1391. Column 1392 indicates that, in such situations, a proper patient rhythm classification would have the patient be deemed “not shockable” because of the presence of QRS complexes and, therefore, a “no shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at column 1393.

According to embodiments, systems will identify signal spikes in the received ECG waveform, i.e., due to the presence of the QRS complexes, and properly interpret the signal spikes as corresponding to the QRS complexes for row 1320, as indicated at columns 1394 and 1395, respectively. The outcome of such a shock advisory algorithm due to the interpretation would be the proper issuing of a “no shock” prompt to the rescuer, as indicated at column 1396.

Rows 1330 and 1340 pertain to situations in which a received ECG waveform contains neither QRS signals nor impulsive artifacts, as indicated at column 1391. In such situations, a proper patient rhythm classification would

provide that the patient is experiencing ventricular fibrillation (VF) or high-rate ventricular tachycardia (VT), as indicated at row 1330, column 1392, or asystole, as indicated at row 1340, column 1392. If the patient is experiencing VF or high-rate VT, a “shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at row 1330, column 1393. If the patient is experiencing asystole, however, a “no shock” prompt would be the desired outcome of such a shock advisory algorithm, as indicated at row 1340, column 1393.

According to embodiments, systems will identify no signal spikes in the received ECG waveform, i.e., due to the absence of QRS complexes and impulsive artifacts, as indicated at column 1394 for both rows 1330 and 1340. The outcome of such a shock advisory algorithm due to the interpretation would be the proper issuing of a “shock advised” prompt if the patient is experiencing VF or high-rate VT, as indicated at row 1330, column 1396, or a “no shock advised” prompt if the patient is experiencing asystole, as indicated at row 1340, column 1396.

Rows 1350 and 1360 pertain to situations in which a received ECG waveform contains a dominant impulsive artifact, i.e., resulting from administration of chest compressions to the patient, as indicated at column 1391. In such situations, a proper patient rhythm classification would provide that the patient is experiencing VF or high-rate VT, as indicated at row 1350, column 1392, or asystole, as indicated at row 1360, column 1392. If the patient is experiencing VF or high-rate VT, a “shock” prompt would be the desired outcome of a shock advisory algorithm, as indicated at row 1350, column 1393. If the patient is experiencing asystole, however, a “no shock” prompt would be the desired outcome of such a shock advisory algorithm, as indicated at row 1360, column 1393.

According to embodiments, systems will identify signal spikes in the received ECG waveform, i.e., due to the presence of the dominant impulsive artifact, as indicated at column 1394 for both rows 1350 and 1360. In situations where the patient is experiencing VF or high-rate VT, a shock advisory algorithm according to embodiments will correctly interpret the detected signal spikes as impulsive artifacts and, therefore, properly issue a “pause CPR for clean analysis” prompt, as indicated at row 1350, column 1395, and row 1360, column 1396, respectively. Consequently, patients in these situations will receive the desired outcome, i.e., the rescuer will be directed to pause CPR and to wait for a clean signal. In contrast, there is nothing in the table 800 of FIG. 8 that describes any type of indication provided to a rescuer that the rescuer should wait for a clean signal in such situations.

In situations where the patient is experiencing asystole, a shock advisory algorithm according to embodiments will correctly interpret the detected signal spikes as impulsive artifacts and thus properly issue a “pause CPR for clean analysis” prompt, as indicated at row 1360, column 1395, and row 1360, column 1396, respectively. Consequently, patients in these situations will also receive the desired outcome, i.e., the rescuer will be directed to pause CPR and to wait for a clean signal. In contrast, there is nothing in the table 800 of FIG. 8 that describes any type of indication provided to a rescuer that the rescuer should wait for a clean signal in such situations.

In this description, numerous details have been set forth in order to provide a thorough understanding. In other instances, well-known features have not been described in detail in order to not obscure unnecessarily the description.

15

A person skilled in the art will be able to practice the present invention in view of this description, which is to be taken as a whole. The specific embodiments as disclosed and illustrated herein are not to be considered in a limiting sense. Indeed, it should be readily apparent to those skilled in the art that what is described herein may be modified in numerous ways. Such ways can include equivalents to what is described herein. In addition, the invention may be practiced in combination with other systems.

The following claims define certain combinations and sub-combinations of elements, features, steps, and/or functions, which are regarded as novel and non-obvious. Additional claims for other combinations and sub-combinations may be presented in this or a related document.

What is claimed is:

1. A medical device, comprising:

an energy storage module configured to store an electrical charge; and

a processor for determining whether the stored electrical charge should be guided to a patient receiving chest compressions, the processor configured to:

identify the occurrence of the chest compressions;

receive ECG data of the patient;

identify signal spikes in the ECG data;

extract a grouping statistic about the timing of the signal spikes relative to the occurrence of the identified chest compressions,

compare the grouping statistic to a threshold, and

when the comparing yields a first outcome, determine according to an underlying algorithm whether the stored electrical charge should be guided to the patient, and

when the comparing yields a second outcome, cause a prompt to be issued for advising a rescuer to cause the chest compressions to stop being received by the patient for subsequent ECG data to be received.

2. The medical device of claim 1, in which

the first outcome occurs when the grouping statistic is less than the threshold.

3. The medical device of claim 1, in which

the processor is configured to extract the grouping statistic by:

establishing a cycle interval for the identified chest compressions,

setting a window within a portion the cycle interval, determining respective positions of at least some of the signal spikes within the cycle interval, and

in which the grouping statistic is determined from a ratio indicating a ratio of the respective positions within the window.

4. The medical device of claim 3, in which

the cycle interval is established from a time distance between two successive identified chest compressions.

5. The medical device of claim 3, in which

the window is set as a percentage of a width of the cycle interval.

6. The medical device of claim 3, in which

the window comprises two non-adjacent sub-windows within the cycle interval.

7. The medical device of claim 3, in which

one of the respective positions is determined by a distance in time between its corresponding signal spike and a corresponding detected chest compression.

8. The medical device of claim 7, in which

the determined respective position is measured from an absolute peak value of the corresponding signal spike and the corresponding detected chest compression.

16

9. A method, comprising:

identifying the occurrence of chest compressions received by a patient;

receiving ECG data of the patient;

identifying signal spikes in the ECG data;

extracting a grouping statistic about the timing of the signal spikes relative to the occurrence of the identified chest compressions, and

comparing the grouping statistic to a threshold, and

when the comparing yields a first outcome, determining according to an underlying algorithm whether the stored electrical charge should be guided to the patient, and

when the comparing yields a second outcome, causing a prompt to be issued for advising a rescuer to cause the chest compressions to stop being received by the patient for subsequent ECG data to be received.

10. The method of claim 9, in which

the first outcome occurs when the grouping statistic is less than the threshold.

11. The method of claim 9, in which the extracting comprises:

establishing a cycle interval for the identified chest compressions,

setting a window within a portion the cycle interval,

determining respective positions of at least some of the signal spikes within the cycle interval, and

in which the grouping statistic is determined from a ratio indicating a ratio of the respective positions within the window.

12. The method of claim 11, in which

the cycle interval is established from a time distance between two successive identified chest compressions.

13. The method of claim 11, in which

the window is set as a percentage of a width of the cycle interval.

14. The method of claim 11, in which

the window comprises two non-adjacent sub-windows within the cycle interval.

15. The method of claim 11, in which

one of the respective positions is determined by a distance in time between its corresponding signal spike and a corresponding detected chest compression.

16. The method of claim 15, in which

the determined respective position is measured from an absolute peak value of the corresponding signal spike and the corresponding detected chest compression.

17. A method, comprising:

identifying the occurrence of chest compressions being received by the patient during CPR;

receiving ECG data of the patient;

identifying signal spikes in the ECG data;

determining a grouping statistic based on the relative timing of the signal spikes and the occurrence of the chest compressions,

performing a shock advisory algorithm in response to the grouping statistic meeting a first criterion, the shock advisory algorithm determining whether electrical charge should be guided to the patient, and

causing a prompt to be issued in response to the grouping statistic meeting a second criterion, the prompt directing a rescuer to cause pausing of the CPR.

18. The method of claim 17, further comprising:

wherein pausing the CPR results in subsequent ECG data that is of the patient is received that is free of impulsive artifacts caused by CPR.

17

19. The method of claim **17**, in which
the occurrence of chest compressions being received by
the patient during CPR is identified by an impedance
signal of the patient.

20. The method of claim **17**, in which
the occurrence of chest compressions being received by
the patient during CPR is identified by one or both of
an impedance signal of the patient and an input signal
of a chest compression machine.

* * * * *

10

18

专利名称(译)	分析心电图数据决定患者胸部按压治疗		
公开(公告)号	US9770183	公开(公告)日	2017-09-26
申请号	US12/910677	申请日	2010-10-22
申请(专利权)人(译)	生理控制, INC.		
当前申请(专利权)人(译)	生理控制, INC.		
[标]发明人	SULLIVAN JOSEPH L		
发明人	SULLIVAN, JOSEPH L.		
IPC分类号	A61B5/00 A61B5/0452 A61N1/39		
CPC分类号	A61B5/0452 A61N1/3925		
其他公开文献	US20110082379A1		
外部链接	Espacenet USPTO		

摘要(译)

提供了医疗装置, 软件和方法, 用于决定是否暂停与对患者施用电击疗法相关的患者胸部按压治疗。根据ECG数据中识别的信号尖峰是否被确定为QRS复合波, 或者仅仅是由胸部按压引起的冲动伪影来做出决定。

