



(19) **United States**

(12) **Patent Application Publication**
Al-Ali

(10) **Pub. No.: US 2003/0181798 A1**

(43) **Pub. Date: Sep. 25, 2003**

(54) **PHYSIOLOGICAL MEASUREMENT COMMUNICATIONS ADAPTER**

(52) **U.S. Cl. 600/324**

(76) **Inventor: Ammar Al-Ali, Tustin, CA (US)**

(57) **ABSTRACT**

Correspondence Address:
KNOBBE MARTENS OLSON & BEAR LLP
2040 MAIN STREET
FOURTEENTH FLOOR
IRVINE, CA 92614 (US)

A sensor interface is configured to receive a sensor signal. A transmitter modulates a first baseband signal responsive to the sensor signal so as to generate a transmit signal. A receiver demodulates a receive signal corresponding to the transmit signal so as to generate a second baseband signal corresponding to the first baseband signal. Further, a monitor interface is configured to communicate a waveform responsive to the second baseband signal to a sensor port of a monitor. The waveform is adapted to the monitor so that measurements derived by the monitor from the waveform are generally equivalent to measurements derivable from the sensor signal. The communications adapter may further comprise a signal processor having an input in communications with the sensor interface, where the signal processor is operable to derive a parameter responsive to the sensor signal and where the first baseband signal is responsive to the parameter. The parameter may correspond to at least one of a measured oxygen saturation and a pulse rate.

(21) **Appl. No.: 10/377,933**


(22) **Filed: Feb. 28, 2003**

Related U.S. Application Data

(60) **Provisional application No. 60/367,428, filed on Mar. 25, 2002.**

Publication Classification

(51) **Int. Cl.⁷ A61B 5/00**






FIG. 1 (Prior Art)




FIG. 2 (Prior Art)




FIG. 3




FIG. 4A




FIG. 4B




FIG. 5A




FIG. 5B




FIG. 5C




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10




FIG. 11




FIG. 12




FIG. 13




FIG. 14

PHYSIOLOGICAL MEASUREMENT COMMUNICATIONS ADAPTER

REFERENCE TO RELATED APPLICATION

[0001] The present application claims priority benefit under 35 U.S.C. §119(e) from U.S. Provisional Application No. 60/367,428, filed Mar. 25, 2002, entitled "*Physiological Measurement Communications Adapter*," which is incorporated herein by reference.

BACKGROUND OF THE INVENTION

[0002] Patient vital sign monitoring may include measurements of blood oxygen, blood pressure, respiratory gas, and EKG among other parameters. Each of these physiological parameters typically require a sensor in contact with a patient and a cable connecting the sensor to a monitoring device. For example, FIGS. 1-2 illustrate a conventional pulse oximetry system 100 used for the measurement of blood oxygen. As shown in FIG. 1, a pulse oximetry system has a sensor 110, a patient cable 140 and a monitor 160. The sensor 110 is typically attached to a finger 10 as shown. The sensor 110 has a plug 118 that inserts into a patient cable socket 142. The monitor 160 has a socket 162 that accepts a patient cable plug 144. The patient cable 140 transmits an LED drive signal 252 (FIG. 2) from the monitor 160 to the sensor 110 and a resulting detector signal 254 (FIG. 2) from the sensor 110 to the monitor 160. The monitor 160 processes the detector signal 254 (FIG. 2) to provide, typically, a numerical readout of the patient's oxygen saturation, a numerical readout of pulse rate, and an audible indicator or "beep" that occurs in response to each arterial pulse.

[0003] As shown in FIG. 2, the sensor 110 has both red and infrared LED emitters 212 and a photodiode detector 214. The monitor 160 has a sensor interface 271, a signal processor 273, a controller 275, output drivers 276, a display and audible indicator 278, and a keypad 279. The monitor 160 determines oxygen saturation by computing the differential absorption by arterial blood of the two wavelengths emitted by the sensor emitters 212, as is well-known in the art. The sensor interface 271 provides LED drive current 252 which alternately activates the red and IR LED emitters 212. The photodiode detector 214 generates a signal 254 corresponding to the red and infrared light energy attenuated from transmission through the patient finger 10 (FIG. 1). The sensor interface 271 also has input circuitry for amplification, filtering and digitization of the detector signal 254. The signal processor 273 calculates a ratio of detected red and infrared intensities, and an arterial oxygen saturation value is empirically determined based on that ratio. The controller 275 provides hardware and software interfaces for managing the display and audible indicator 278 and keypad 279. The display and audible indicator 278 shows the computed oxygen status, as described above, and provides the pulse beep as well as alarms indicating oxygen desaturation events. The keypad 279 provides a user interface for setting alarm thresholds, alarm enablement, and display options, to name a few.

SUMMARY OF THE INVENTION

[0004] Conventional physiological measurement systems are limited by the patient cable connection between sensor and monitor. A patient must be located in the immediate

vicinity of the monitor. Also, patient relocation requires either disconnection of monitoring equipment and a corresponding loss of measurements or an awkward simultaneous movement of patient equipment and cables. Various devices have been proposed or implemented to provide wireless communication links between sensors and monitors, freeing patients from the patient cable tether. These devices, however, are incapable of working with the large installed base of existing monitors and sensors, requiring caregivers and medical institutions to suffer expensive wireless upgrades. It is desirable, therefore, to provide a communications adapter that is plug-compatible both with existing sensors and monitors and that implements a wireless link replacement for the patient cable.

[0005] An aspect of a physiological measurement communications adapter comprises a sensor interface configured to receive a sensor signal. A transmitter modulates a first baseband signal responsive to the sensor signal so as to generate a transmit signal. A receiver demodulates a receive signal corresponding to the transmit signal so as to generate a second baseband signal corresponding to the first baseband signal. Further, a monitor interface is configured to communicate a waveform responsive to the second baseband signal to a sensor port of a monitor. The waveform is adapted to the monitor so that measurements derived by the monitor from the waveform are generally equivalent to measurements derivable from the sensor signal. The communications adapter may further comprise a signal processor having an input in communications with the sensor interface, where the signal processor is operable to derive a parameter responsive to the sensor signal and where the first baseband signal is responsive to the parameter. The parameter may correspond to at least one of a measured oxygen saturation and a pulse rate.

[0006] One embodiment may further comprise a waveform generator that synthesizes the waveform from a predetermined shape. The waveform generator synthesizes the waveform at a frequency adjusted to be generally equivalent to the pulse rate. The waveform may have a first amplitude and a second amplitude, and the waveform generator may be configured to adjusted the amplitudes so that measurements derived by the monitor are generally equivalent to a measured oxygen saturation.

[0007] In another embodiment, the sensor interface is operable on the sensor signal to provide a plethysmograph signal output, where the first baseband signal is responsive to the plethysmograph signal. This embodiment may further comprise a waveform modulator that modifies a decoded signal responsive to the second baseband signal to provide the waveform. The waveform modulator may comprise a demodulator that separates a first signal and a second signal from the decoded signal, an amplifier that adjusts amplitudes of the first and second signals to generate a first adjusted signal and a second adjusted signal, and a modulator that combines the first and second adjusted signals into the waveform. The amplitudes of the first and second signals may be responsive to predetermined calibration data for the sensor and the monitor.

[0008] An aspect of a physiological measurement communications adapter method comprises the steps of inputting a sensor signal at a patient location, communicating patient data derived from the sensor signal between the patient

location and a monitor location, constructing a waveform at the monitor location responsive to the sensor signal, and providing the waveform to a monitor via a sensor port. The waveform is constructed so that the monitor calculates a parameter generally equivalent to a measurement derivable from the sensor signal.

[0009] In one embodiment, the communicating step may comprise the substeps of deriving a conditioned signal from the sensor signal, calculating a parameter signal from the conditioned signal, and transmitting the parameter signal from the patient location to the monitor location. The constructing step may comprise the substep of synthesizing the waveform from the parameter signal. In an alternative embodiment, the communicating step may comprise the substeps of deriving a conditioned signal from said sensor signal and transmitting the conditioned signal from the patient location to the monitor location. The constructing step may comprise the substeps of demodulating the conditioned signal and re-modulating the conditioned signal to generate the waveform. The providing step may comprise the substeps of inputting a monitor signal from an LED drive output of the sensor port, modulating the waveform in response to the monitor signal, and outputting the waveform on a detector input of the sensor port.

[0010] Another aspect of a physiological measurement communications adapter comprises a sensor interface means for inputting a sensor signal and outputting a conditioned signal, a transmitter means for sending data responsive to the sensor signal, and a receiver means for receiving the data. The communications adapter further comprises a waveform processor means for constructing a waveform from the data so that measurements derived by a monitor from the waveform are generally equivalent to measurements derivable from the sensor signal, and a monitor interface means for communicating the waveform to a sensor port of the monitor. The communications adapter may further comprise a signal processor means for deriving a parameter signal from the conditioned signal, where the data comprises the parameter signal. The waveform processor means may comprise a means for synthesizing the waveform from the parameter signal. The data may comprise the conditioned signal, and the waveform processor means may comprise a means for modulating the conditioned signal in response to the monitor.

BRIEF DESCRIPTION OF THE DRAWINGS

[0011] FIG. 1 is an illustration of a prior art pulse oximetry system;

[0012] FIG. 2 is a functional block diagram of a prior art pulse oximetry system;

[0013] FIG. 3 is an illustration of a physiological measurement communications adapter;

[0014] FIGS. 4A-B are illustrations of communications adapter sensor modules;

[0015] FIGS. 5A-C are illustrations of communications adapter monitor modules;

[0016] FIG. 6 is a functional block diagram of a communications adapter sensor module;

[0017] FIG. 7 is a functional block diagram of a communications adapter monitor module;

[0018] FIG. 8 is a functional block diagram of a sensor module configured to transmit measured pulse oximeter parameters;

[0019] FIG. 9 is a functional block diagram of a monitor module configured to received measured pulse oximeter parameters;

[0020] FIG. 10 is a functional block diagram of a sensor module configured to transmit a plethysmograph;

[0021] FIG. 11 is a functional block diagram of a monitor module configured to receive a plethysmograph;

[0022] FIG. 12 is a functional block diagram of a waveform modulator;

[0023] FIG. 13 is a functional block diagram of a sensor module configured for multiple sensors; and

[0024] FIG. 14 is a functional block diagram of a monitor module configured for multiple sensors.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENT Overview

[0025] FIG. 3 illustrates one embodiment of a communications adapter. FIGS. 4-5 illustrate physical configurations for a communications adapter. In particular, FIGS. 4A-B illustrate sensor module configurations and FIGS. 5A-C illustrate monitor module configurations. FIGS. 6-14 illustrate communications adapter functions. In particular, FIGS. 6-7 illustrate general functions for a sensor module and a monitor module, respectively. FIGS. 8-9 functionally illustrate a communications adapter where derived pulse oximetry parameters, such as saturation and pulse rate are transmitted between a sensor module and a monitor module. Also, FIGS. 10-12 functionally illustrate a communications adapter where a plethysmograph is transmitted between a sensor module and a monitor module. FIGS. 13-14 functionally illustrate a multiple-parameter communications adapter.

[0026] FIG. 3 illustrates a communications adapter 300 having a sensor module 400 and a monitor module 500. The communications adapter 300 communicates patient data derived from a sensor 310 between the sensor module 400, which is located proximate a patient 20 and the monitor module 500, which is located proximate a monitor 360. A wireless link 340 is provided between the sensor module 400 and the monitor module 500, replacing the conventional patient cable, such as a pulse oximetry patient cable 140 (FIG. 1). Advantageously, the sensor module 400 is plug-compatible with a conventional sensor 310. In particular, the sensor connector 318 connects to the sensor module 400 in a similar manner as to a patient cable. Further, the sensor module 400 outputs a drive signal to the sensor 310 and inputs a sensor signal from the sensor 310 in an equivalent manner as a conventional monitor 360. The sensor module 400 may be battery powered or externally powered. External power may be for recharging internal batteries or for powering the sensor module during operation or both.

[0027] As shown in FIG. 3, the monitor module 500 is advantageously plug-compatible with a conventional monitor 360. In particular, the monitor's sensor port 362 connects to the monitor module 500 in a similar manner as to a patient cable, such as a pulse oximetry patient cable 140 (FIG. 1). Further, the monitor module 500 inputs a drive signal from

the monitor **360** and outputs a corresponding sensor signal to the monitor **360** in an equivalent manner as a conventional sensor **310**. As such, the combination sensor module **400** and monitor module **500** provide a plug-compatible wireless replacement for a patient cable, adapting an existing wired physiological measurement system into a wireless physiological measurement system. The monitor module **500** may be battery powered, powered from the monitor, such as by tapping current from a monitor's LED drive, or externally powered from an independent AC or DC power source.

[0028] Although a communications adapter **300** is described herein with respect to a pulse oximetry sensor and monitor, one of ordinary skill in the art will recognize that a communications adapter may provide a plug-compatible wireless replace for a patient cable that connects any physiological sensor and corresponding monitor. For example, a communications adapter **300** may be applied to a biopotential sensor, a non-invasive blood pressure (NIBP) sensor, a respiratory rate sensor, a glucose sensor and the corresponding monitors, to name a few.

[0029] Sensor Module Physical Configurations

[0030] FIGS. 4A-B illustrate physical embodiments of a sensor module **400**. FIG. 4A illustrates a wrist-mounted module **410** having a wrist strap **411**, a case **412** and an auxiliary cable **420**. The case **412** contains the sensor module electronics, which are functionally described with respect to FIG. 6, below. The case **412** is mounted to the wrist strap **411**, which attaches the wrist-mounted module **410** to a patient **20**. The auxiliary cable **420** mates to a sensor connector **318** and a module connector **414**, providing a wired link between a conventional sensor **310** and the wrist-mounted module **410**. Alternatively, the auxiliary cable **420** is directly wired to the sensor module **400**. The wrist-mounted module **410** may have a display **415** that shows sensor measurements, module status and other visual indicators, such as monitor status. The wrist-mounted module **410** may also have keys (not shown) or other input mechanisms to control its operational mode and characteristics. In an alternative embodiment, the sensor **310** may have a tail (not shown) that connects directly to the wrist-mounted module **410**, eliminating the auxiliary cable **420**.

[0031] FIG. 4B illustrates a clip-on module **460** having a clip **461**, a case **462** and an auxiliary cable **470**. The clip **461** attaches the clip-on module **460** to patient clothing or objects near a patient **20**, such as a bed frame. The auxiliary cable **470** mates to the sensor connector **318** and functions as for the auxiliary cable **420** (FIG. 4A) of the wrist-mounted module **410** (FIG. 4A), described above. The clip-on module **460** may have a display **463** and keys **464** as for the wrist-mounted module **410** (FIG. 4A). Either the wrist-mounted module **410** or the clip-on module **460** may have other input or output ports (not shown) that download software, configure the module, or provide a wired connection to other measurement instruments or computing devices, to name a few examples.

[0032] Monitor Module Physical Configurations

[0033] FIGS. 5A-C illustrate physical embodiments of a monitor module **500**. FIG. 5A illustrates a direct-connect module **510** having a case **512** and an integrated monitor connector **514**. The case **512** contains the monitor module electronics, which are functionally described with respect to

FIG. 7, below. The monitor connector **514** mimics that of the monitor end of a patient cable, such as a pulse oximetry patient cable **140** (FIG. 1), and electrically and mechanically connects the monitor module **510** to the monitor **360** via the monitor's sensor port **362**.

[0034] FIG. 5B illustrates a cable-connect module **540** having a case **542** and an auxiliary cable **550**. The case **542** functions as for the direct-connect module **510** (FIG. 5A), described above. Instead of directly plugging into the monitor **360**, the cable-connect module **540** utilizes the auxiliary cable **550**, which mimics the monitor end of a patient cable, such as a pulse oximetry patient cable **140** (FIG. 1), and electrically connects the cable-connect module **540** to the monitor sensor port **362**.

[0035] FIG. 5C illustrates a plug-in module **570** having a plug-in case **572** and an auxiliary cable **580**. The plug-in case **572** is mechanically compatible with the plug-in chassis of a multiparameter monitor **370** and may or may not electrically connect to the chassis backplane. The auxiliary cable **580** mimics a patient cable and electrically connects the plug-in module **570** to the sensor port **372** of another plug-in device. A direct-connect module **510** (FIG. 5A) or a cable-connect module **540** (FIG. 5B) may also be used with a multiparameter monitor **370**.

[0036] In a multiparameter embodiment, such as described with respect to FIGS. 13-14, below, a monitor module **500** may connect to multiple plug-in devices of a multiparameter monitor **370**. For example, a cable-connect module **540** (FIG. 5B) may have multiple auxiliary cables **550** (FIG. 5B) that connect to multiple plug-in devices installed within a multiparameter monitor chassis. Similarly, a plug-in module **570** may have one or more auxiliary cables **580** with multiple connectors for attaching to the sensor ports **372** of multiple plug-in devices.

[0037] Communications Adapter Functions

[0038] FIGS. 6-7 illustrate functional embodiments of a communications adapter. FIG. 6 illustrates a sensor module **400** having a sensor interface **610**, a signal processor **630**, an encoder **640**, a transmitter **650** and a transmitting antenna **670**. A physiological sensor **310** provides an input sensor signal **612** at the sensor connector **318**. Depending on the sensor **310**, the sensor module **400** may provide one or more drive signals **618** to the sensor **310**. The sensor interface **610** inputs the sensor signal **612** and outputs a conditioned signal **614**. The conditioned signal **614** may be coupled to the transmitter **650** or further processed by a signal processor **630**. If the sensor module configuration utilizes a signal processor **630**, it derives a parameter signal **632** responsive to the sensor signal **612**, which is then coupled to the transmitter **650**. Regardless, the transmitter **650** inputs a baseband signal **642** that is responsive to the sensor signal **612**. The transmitter **650** modulates the baseband signal **642** with a carrier to generate a transmit signal **654**. The transmit signal **654** may be derived by various amplitude, frequency or phase modulation schemes, as is well known in the art. The transmit signal **654** is coupled to the transmit antenna **670**, which provides wireless communications to a corresponding receive antenna **770** (FIG. 7), as described below.

[0039] As shown in FIG. 6, the sensor interface **610** conditions and digitizes the sensor signal **612** to generate the conditioned signal **614**. Sensor signal conditioning may be

performed in the analog domain or digital domain or both and may include amplification and filtering in the analog domain and filtering, buffering and data rate modification in the digital domain, to name a few. The resulting conditioned signal 614 is responsive to the sensor signal 612 and may be used to calculate or derive a parameter signal 632.

[0040] Further shown in FIG. 6, the signal processor 630 performs signal processing on the conditioned signal 614 to generate the parameter signal 632. The signal processing may include buffering, digital filtering, smoothing, averaging, adaptive filtering and frequency transforms to name a few. The resulting parameter signal 632 may be a measurement calculated or derived from the conditioned signal, such as oxygen saturation, pulse rate, blood glucose, blood pressure and EKG to name a few. Also, the parameter signal 632 may be an intermediate result from which the above-stated measurements may be calculated or derived.

[0041] As described above, the sensor interface 610 performs mixed analog and digital pre-processing of an analog sensor signal and provides a digital output signal to the signal processor 630. The signal processor 630 then performs digital post-processing of the front-end processor output. In alternative embodiments, the input sensor signal 612 and the output conditioned signal 614 may be either analog or digital, the front-end processing may be purely analog or purely digital, and the back-end processing may be purely analog or mixed analog or digital.

[0042] In addition, FIG. 6 shows an encoder 640, which translates a digital word or serial bit stream, for example, into the baseband signal 642, as is well-known in the art. The baseband signal 642 comprises the symbol stream that drives the transmit signal 654 modulation, and may be a single signal or multiple related signal components, such as in-phase and quadrature signals. The encoder 640 may include data compression and redundancy, also well-known in the art.

[0043] FIG. 7 illustrates a monitor module 500 having a receive antenna 770, a receiver 710, a decoder 720, a waveform processor 730 and a monitor interface 750. A receive signal 712 is coupled from the receive antenna 770, which provides wireless communications to a corresponding transmit antenna 670 (FIG. 6), as described above. The receiver 710 inputs the receive signal 712, which corresponds to the transmit signal 654 (FIG. 6). The receiver 710 demodulates the receive signal to generate a baseband signal 714. The decoder 720 translates the symbols of the demodulated baseband signal 714 into a decoded signal 724, such as a digital word stream or bit stream. The waveform processor 730 inputs the decoded signal 724 and generates a constructed signal 732. The monitor interface 750 is configured to communicate the constructed signal 732 to a sensor port 362 of a monitor 360. The monitor 360 may output a sensor drive signal 754, which the monitor interface 750 inputs to the waveform processor 730 as a monitor drive signal 734. The waveform processor 730 may utilize the monitor drive signal 734 to generate the constructed signal 732. The monitor interface 750 may also provide characterization information 758 to the waveform processor 730, relating to the monitor 360, the sensor 310 or both, that the waveform processor 730 utilizes to generate the constructed signal 732.

[0044] The constructed signal 732 is adapted to the monitor 360 so that measurements derived by the monitor 360

from the constructed signal 732 are generally equivalent to measurements derivable from the sensor signal 612 (FIG. 6). Note that the sensor 310 (FIG. 6) may or may not be directly compatible with the monitor 360. If the sensor 310 (FIG. 6) is compatible with the monitor 360, the constructed signal 732 is generated so that measurements derived by the monitor 360 from the constructed signal 732 are generally equivalent (within clinical significance) with those derivable directly from the sensor signal 612 (FIG. 6). If the sensor 310 (FIG. 6) is not compatible with the monitor 360, the constructed signal 732 is generated so that measurements derived by the monitor 360 from the constructed signal 732 are generally equivalent to those derivable directly from the sensor signal 612 (FIG. 6) using a compatible monitor.

[0045] Wireless Pulse Oximetry

[0046] FIGS. 8-11 illustrate pulse oximeter embodiments of a communications adapter. FIGS. 8-9 illustrate a sensor module and a monitor module, respectively, configured to communicate measured pulse oximeter parameters. FIG. 10-11 illustrate a sensor module and a monitor module, respectively, configured to communicate a plethysmograph signal.

[0047] Parameter Transmission

[0048] FIG. 8 illustrates a pulse oximetry sensor module 800 having a sensor interface 810, signal processor 830, encoder 840, transmitter 850, transmitting antenna 870 and controller 890. The sensor interface 810, signal processor 830 and controller 890 function as described with respect to FIG. 2, above. The sensor interface 810 communicates with a standard pulse oximetry sensor 310, providing an LED drive signal 818 to the LED emitters 312 and receiving a sensor signal 812 from the detector 314 in response. The sensor interface 810 provides front-end processing of the sensor signal 812, also described above, providing a plethysmograph signal 814 to the signal processor 830. The signal processor 830 then derives a parameter signal 832 that comprises a real time measurement of oxygen saturation and pulse rate. The parameter signal 832 may include other parameters, such as measurements of perfusion index and signal quality. In one embodiment, the signal processor is an MS-5 or MS-7 board available from Masimo Corporation, Irvine, Calif.

[0049] As shown in FIG. 8, the encoder 840, the transmitter 850 and the transmitting antenna 870 function as described with respect to FIG. 6, above. For example, the parameter signal 832 may be a digital word stream that is serialized into a bit stream and encoded into a baseband signal 842. The baseband signal 842 may be, for example, two bit symbols that drive a quadrature phase shift keyed (QPSK) modulator in the transmitter 850. Other encodings and modulations are also applicable, as described above. The transmitter 850 inputs the baseband signal 842 and generates a transmit signal 854 that is a modulated carrier having a frequency suitable for short-range transmission, such as within a hospital room, doctor's office, emergency vehicle or critical care ward, to name a few. The transmit signal 854 is coupled to the transmit antenna 870, which provides wireless communications to a corresponding receive antenna 970 (FIG. 9), as described below.

[0050] FIG. 9 illustrates a monitor module 900 having a receive antenna 970, a receiver 910, a decoder 920, a

waveform generator **930** and an interface cable **950**. The receive antenna **970**, receiver **910** and decoder **920** function as described with respect to **FIG. 7**, above. In particular, the receive signal **912** is coupled from the receive antenna **970**, which provides wireless communications to a corresponding transmit antenna **870** (**FIG. 8**). The receiver **910** inputs the receive signal **912**, which corresponds to the transmit signal **854** (**FIG. 8**). The receiver **810** demodulates the receive signal **912** to generate a baseband signal **914**. Not accounting for transmission errors, the baseband signal **914** corresponds to the sensor module baseband signal **842** (**FIG. 8**), for example a symbol stream of two bits each. The decoder **920** assembles the baseband signal **914** into a parameter signal **924**, which, for example, may be a sequence of digital words corresponding to oxygen saturation and pulse rate. Again, not accounting for transmission errors, the monitor module parameter signal **924** corresponds to the sensor module parameter signal **832** (**FIG. 8**), derived by the signal processor **830** (**FIG. 8**).

[0051] Also shown in **FIG. 9**, the waveform generator **930** is a particular embodiment of the waveform processor **730** (**FIG. 7**) described above. The waveform generator **930** generates a synthesized waveform **932** that the pulse oximeter monitor **360** can process to calculate SpO_2 and pulse rate values or exception messages. In the present embodiment, the waveform generator output does not reflect a physiological waveform. In particular, the synthesized waveform is not physiological data from the sensor module **800**, but is a waveform synthesized from predetermined stored waveform data to cause the monitor **360** to calculate oxygen saturation and pulse rate equivalent to or generally equivalent (within clinical significance) to that calculated by the signal processor **830** (**FIG. 8**). The actual intensity signal from the patient received by the detector **314** (**FIG. 8**) is not provided to the monitor **360** in the present embodiment. Indeed, the waveform provided to the monitor **360** will usually not resemble a plethysmographic waveform or other physiological data from the patient to whom the sensor module **800** (**FIG. 8**) is attached.

[0052] The synthesized waveform **932** is modulated according to the drive signal input **934**. That is, the pulse oximeter monitor **360** expects to receive a red and IR modulated intensity signal originating from a detector, as described with respect to **FIGS. 1-2**, above. The waveform generator **930** generates the synthesized waveform **932** with a predetermined shape, such as a triangular or sawtooth waveform stored in waveform generator memory or derived by a waveform generator algorithm. The waveform is modulated synchronously with the drive input **934** with first and second amplitudes that are processed in the monitor **360** as red and IR portions of a sensor signal. The frequency and the first and second amplitudes are adjusted so that pulse rate and oxygen saturation measurements derived by the pulse oximeter monitor **360** are generally equivalent to the parameter measurements derived by the signal processor **830** (**FIG. 8**), as described above. One embodiment of a waveform generator **930** is described in US Patent Application No. **60/117,097** entitled "Universal/Upgrading Pulse Oximeter," assigned to Masimo Corporation, Irvine, Calif. and incorporated by reference herein. Although the waveform generator **930** is described above as synthesizing a waveform that does not resemble a physiological signal, one of ordinary skill will recognize that another embodiment of

the waveform generator **930** could incorporate, for example, a plethysmograph simulator or other physiological signal simulator.

[0053] Further shown in **FIG. 9**, the interface cable **950** functions in a manner similar to the monitor interface **750** (**FIG. 7**) described above. The interface cable **950** is configured to communicate the synthesized waveform **932** to the monitor **360** sensor port and to communicate the sensor drive signal **934** to the waveform generator **930**. The interface cable **950** may include a ROM **960** that contains monitor and sensor characterization data. The ROM **960** is read by the waveform generator **930** so that the synthesized waveform **932** is adapted to a particular monitor **360**. For example, the ROM **960** may contain calibration data of red/IR versus oxygen saturation, waveform amplitude and waveform shape information. An interface cable is described in U.S. Patent Application No. **60/117,092**, referenced above. Monitor-specific SatShare™ brand interface cables are available from Masimo Corporation, Irvine, Calif. In an alternative embodiment, such as a direct connect monitor module as illustrated in **FIG. 5A**, an interface cable **950** is not used and the ROM **960** may be incorporated within the monitor module **900** itself.

[0054] Plethysmograph Transmission

[0055] **FIG. 10** illustrates another pulse oximetry sensor module **1000** having a sensor interface **1010**, encoder **1040**, transmitter **1050**, transmitting antenna **1070** and controller **1090**, which have the corresponding functions as those described with respect to **FIG. 8**, above. The encoder **1040**, however, inputs a plethysmograph signal **1014** rather than oxygen saturation and pulse rate measurements **832** (**FIG. 8**). Thus, the sensor module **1000** according to this embodiment encodes and transmits a plethysmograph signal **1014** to a corresponding monitor module **1100** (**FIG. 11**) in contrast to derived physiological parameters, such as oxygen saturation and pulse rate. The plethysmograph signal **1014** is illustrated in **FIG. 10** as being a direct output from the sensor interface **1010**. In another embodiment, the sensor module **1000** incorporates a decimation processor, not shown, after the sensor interface **1010** so as to provide a plethysmograph signal **1014** having a reduced sample rate.

[0056] **FIG. 11** illustrates another pulse oximetry monitor module **1100** having a receive antenna **1170**, a receiver **1110**, a decoder **1120** and an interface cable **1150**, which have the corresponding functions as those described with respect to **FIG. 9**, above. This monitor module embodiment **1100**, however, has a waveform modulator **1200** rather than a waveform generator **930** (**FIG. 9**), as described above. The waveform modulator **1200** inputs a plethysmograph signal from the decoder **1120** rather than oxygen saturation and pulse rate measurements, as described with respect to **FIG. 9**, above. Further, the waveform modulator **1200** provides a modulated waveform **1132** to the pulse oximeter monitor **360** rather than a synthesized waveform, as described with respect to **FIG. 9**. The modulated waveform **1132** is a plethysmographic waveform modulated according to the monitor drive signal input **1134**. That is, the waveform modulator **1200** does not synthesize a waveform, but rather modifies the received plethysmograph signal **1124** to cause the monitor **360** to calculate oxygen saturation and pulse rate generally equivalent (within clinical significance) to that derivable by a compatible, calibrated pulse oximeter directly

from the sensor signal 1012 (FIG. 10). The waveform modulator 1200 is described in further detail with respect to FIG. 12, below.

[0057] FIG. 12 shows a waveform modulator 1200 having a demodulator 1210, a red digital-to-analog converter (DAC) 1220, an IR DAC 1230, a red amplifier 1240, an IR amplifier 1250, a modulator 1260, a modulator control 1270, a look-up table (LUT) 1280 and a ratio calculator 1290. The waveform modulator 1200 demodulates red and IR plethysmographs ("pleths") from the decoder output 1124 into a separate red pleth 1222 and IR pleth 1232. The waveform modulator 1200 also adjusts the amplitudes of the pleths 1222, 1232 according to stored calibration curves for the sensor 310 (FIG. 10) and the monitor 360 (FIG. 11). Further, the waveform modulator 1200 re-modulates the adjusted red pleth 1242 and adjusted IR pleth 1252, generating a modulated waveform 1132 to the monitor 360 (FIG. 11).

[0058] As shown in FIG. 12, the demodulator 1210 performs the demodulation function described above, generating digital red and IR pleth signals 1212, 1214. The DACs 1220, 1230 convert the digital pleth signals 1212, 1214 to corresponding analog pleth signals 1222, 1232. The amplifiers 1240, 1250 have variable gain control inputs 1262, 1264 and perform the amplitude adjustment function described above, generating adjusted red and IR pleth signals 1242, 1252. The modulator 1260 performs the re-modulation function described above, combining the adjusted red and IR pleth signals 1242, 1252 according to a control signal 1272. The modulator control 1270 generates the control signal 1272 synchronously with the LED drive signal(s) 1134 from the monitor 360. 100531 Also shown in FIG. 12, the ratio calculator 1290 derives a red/IR ratio from the demodulator outputs 1212, 1214. The LUT 1280 stores empirical calibration data for the sensor 310 (FIG. 10). The LUT 1280 also downloads monitor-specific calibration data from the ROM 1160 (FIG. 11) via the ROM output 1158. From this calibration data, the LUT 1280 determines a desired red/IR ratio for the modulated waveform 1132 and generates red and IR gain outputs 1262, 1264 to the corresponding amplifiers 1240, 1250, accordingly. A desired red/IR ratio is one that allows the monitor 360 (FIG. 11) to derive oxygen saturation measurements from the modulated waveform 1132 that are generally equivalent to that derivable directly from the sensor signal 1012 (FIG. 10).

[0059] One of ordinary skill in the art will recognize that some of the signal processing functions described with respect to FIGS. 8-11 may be performed either within a sensor module or within a monitor module. Signal processing functions performed within a sensor module may advantageously reduce the transmission bandwidth to a monitor module at a cost of increased sensor module size and power consumption. Likewise, signal processing functions performed within a monitor module may reduce sensor module size and power consumption at a cost of increase transmission bandwidth.

[0060] For example, a monitor module embodiment 900 (FIG. 9) described above receives measured pulse oximeter parameters, such as oxygen saturation and pulse rate, and generates a corresponding synthesized waveform. In that embodiment, the oxygen saturation and pulse rate computations are performed within a sensor module 800 (FIG. 8).

Another monitor module embodiment 1100 (FIG. 11), also described above, receives a plethysmograph waveform and generates a remodulated waveform. In that embodiment, minimal signal processing is performed within a sensor module 1000 (FIG. 10). In yet another embodiment, not shown, a sensor module transmits a plethysmograph waveform or a decimated plethysmograph waveform having a reduced sample rate. A corresponding monitor module has a signal processor, such as described with respect to FIG. 8, in addition to a waveform generator, as described with respect to FIG. 9. The signal processor computes pulse oximeter parameters and the waveform generator generates a corresponding synthesized waveform, as described above. In this embodiment, minimal signal processing is performed within the sensor module, and the monitor module functions are performed on the pulse oximeter parameters computed within the monitor module.

[0061] Wireless Multiple Parameter Measurements

[0062] FIGS. 13-14 illustrate a multiple parameter communications adapter. FIG. 13 illustrates a multiple parameter sensor module 1300 having sensor interfaces 1310, one or more signal processors 1330, a multiplexer and encoder 1340, a transmitter 1350, a transmitting antenna 1370 and a controller 1390. One or more physiological sensors 1301 provide input sensor signals 1312 to the sensor module 1300. Depending on the particular sensors 1301, the sensor module 1300 may provide one or more drive signals 1312 to the sensors 1301 as determined by the controller 1390. The sensor interfaces 1310 input the sensor signals 1312 and output one or more conditioned signals 1314. The conditioned signals 1314 may be coupled to the transmitter 1350 or further processed by the signal processors 1330. If the sensor module configuration utilizes signal processors 1330, it derives multiple parameter signals 1332 responsive to the sensor signals 1312, which are then coupled to the transmitter 1350. Regardless, the transmitter 1350 inputs a baseband signal 1342 that is responsive to the sensor signals 1312. The transmitter 1350 modulates the baseband signal 1342 with a carrier to generate a transmit signal 1354, which is coupled to the transmit antenna 1370 and communicated to a corresponding receive antenna 1470 (FIG. 14), as described with respect to FIG. 6, above. Alternatively, there may be multiple baseband signals 1342, and the transmitter 1350 may transmit on multiple frequency channels, where each channel conveys data responsive to one or more of the sensor signals 1314.

[0063] As shown in FIG. 13, the sensor interface 1310 conditions and digitizes the sensor signals 1312 as described for a single sensor with respect to FIG. 6, above. The resulting conditioned signals 1314 are responsive to the sensor signals 1312. The signal processors 1330 perform signal processing on the conditioned signals 1314 to derive parameter signals 1332, as described for a single conditioned signal with respect to FIG. 6, above. The parameter signals 1332 may be physiological measurements such as oxygen saturation, pulse rate, blood glucose, blood pressure, EKG, respiration rate and body temperature to name a few, or may be intermediate results from which the above-stated measurements may be calculated or derived. The multiplexer and encoder 1340 combines multiple digital word or serial bit streams into a single digital word or bit stream. The multiplexer and encoder also encodes the digital word or bit

stream to generate the baseband signal 1342, as described with respect to FIG. 6, above.

[0064] FIG. 14 illustrates a multiple parameter monitor module 1400 having a receive antenna 1470, a receiver 1410, a demultiplexer and decoder 1420, one or more waveform processors 1430 and a monitor interface 1450. The receiver 1410 inputs and demodulates the receive signal 1412 corresponding to the transmit signal 1354 (FIG. 13) to generate a baseband signal 1414 as described with respect to FIG. 7, above. The demultiplexer and decoder 1420 separates the symbol streams corresponding to the multiple conditioned signals 1314 (FIG. 13) and/or parameter signals 1332 (FIG. 13) and translates these symbol streams into multiple decoded signals 1422, as described for a single symbol stream with respect to FIG. 7, above. Alternatively, multiple frequency channels are received to generate multiple baseband signals, each of which are decoded to yield multiple decoded signals 1422. The waveform processors 1430 input the decoded signals 1422 and generate multiple constructed signals 1432, as described for a single decoded signal with respect to FIGS. 7-12, above. The monitor interface 1450 is configured to communicate the constructed signals 1432 to the sensor ports of a multiple parameter monitor 1401 or multiple single parameter monitors, in a manner similar to that for a single constructed signal, as described with respect to FIGS. 7-12, above. In particular, the constructed signals 1432 are adapted to the monitor 1401 so that measurements derived by the monitor 1401 from the constructed signals 1432 are generally equivalent to measurements derivable directly from the sensor signals 1312 (FIG. 13).

[0065] A physiological measurement communications adapter is described above with respect to wireless communications and, in particular, radio frequency communications. A sensor module and monitor module, however, may also communicate via wired communications, such as telephone, Internet or fiberoptic cable to name a few. Further, wireless communications can also utilize light frequencies, such as IR or laser to name a few.

[0066] A physiological measurement communications adapter has been disclosed in detail in connection with various embodiments. These embodiments are disclosed by way of examples only. One of ordinary skill in the art will appreciate many variations and modifications of a physiological measurement communications adapter within the scope of the claims that follow.

What is claimed is:

1. A communications adapter comprising:
 - a sensor interface configured to receive a sensor signal;
 - a transmitter that modulates a first baseband signal responsive to said sensor signal so as to generate a transmit signal;
 - a receiver that demodulates a receive signal corresponding to said transmit signal so as to generate a second baseband signal corresponding to said first baseband signal; and
 - a monitor interface configured to communicate a waveform responsive to said second baseband signal to a sensor port of a monitor, said waveform adapted to said monitor so that measurements derived by said monitor

from said waveform are generally equivalent to measurements derivable from said sensor signal.

2. The communications adapter according to claim 1 further comprising a signal processor having an input in communications with said sensor interface, said signal processor operable to derive a parameter responsive to said sensor signal, said first baseband signal responsive to said parameter.

3. The communications adapter according to claim 2 wherein said parameter corresponds to at least one of a measured oxygen saturation and a pulse rate.

4. The communications adapter according to claim 3 further comprising a waveform generator that synthesizes said waveform from a predetermined shape.

5. The communications adapter according to claim 4 wherein said waveform generator synthesizes said waveform at a frequency adjusted to be generally equivalent to said pulse rate.

6. The communications adapter according to claim 4 wherein said waveform has a first amplitude and a second amplitude, said waveform generator configured to adjusted said amplitudes so that measurements derived by said monitor are generally equivalent to said measured oxygen saturation.

7. The communications adapter according to claim 1 wherein said sensor interface is operable on said sensor signal to provide a plethysmograph signal output, said first baseband signal responsive to said plethysmograph signal.

8. The communications adapter according to claim 7 further comprising a waveform modulator that modifies a decoded signal responsive to said second baseband signal to provide said waveform.

9. The communications adapter according to claim 8 wherein said waveform modulator comprises:

- a demodulator that separates a first signal and a second signal from said decoded signal;

- an amplifier that adjusts amplitudes of said first and second signals to generate a first adjusted signal and a second adjusted signal; and

- a modulator that combines said first and second adjusted signals into said waveform.

10. The communications adapter according to claim 9 wherein amplitudes of said first and second signals are responsive to predetermined calibration data for said sensor and said monitor.

11. A communications adapter method comprising the steps of:

- inputting a sensor signal at a patient location;

- communicating patient data derived from said sensor signal between said patient location and a monitor location;

- constructing a waveform at said monitor location responsive to said sensor signal; and

- providing said waveform to a monitor via a sensor port, said waveform constructed so that said monitor calculates a parameter generally equivalent to a measurement derivable from said sensor signal.

12. The communications adapter method according to claim 11 wherein said communicating step comprises the substeps of:

deriving a conditioned signal from said sensor signal;
calculating a parameter signal from said conditioned
signal; and

transmitting said parameter signal from said patient loca-
tion to said monitor location.

13. The communications adapter method according to
claim 12 wherein said constructing step comprises the
substep of synthesizing said waveform from said parameter
signal.

14. The communications adapter method according to
claim 11 wherein said communicating step comprises the
substeps of:

deriving a conditioned signal from said sensor signal;
transmitting said conditioned signal from said patient
location to said monitor location.

15. The communications adapter method according to
claim 14 wherein said constructing step comprises the
substeps of:

demodulating said conditioned signal; and
re-modulating said conditioned signal to generate said
waveform.

16. The communications adapter method according to
claim 11 wherein said providing step comprises the substeps
of:

inputting a monitor signal from an LED drive output of
said sensor port;

modulating said waveform in response to said monitor
signal; and

outputting said waveform on a detector input of said
sensor port.

17. A communications adapter comprising:

a sensor interface means for inputting a sensor signal and
outputting a conditioned signal;

a transmitter means for sending data responsive to said
sensor signal;

a receiver means for receiving said data;

a waveform processor means for constructing a waveform
from said data so that measurements derived by a
monitor from said waveform are generally equivalent
to measurements derivable from said sensor signal; and

a monitor interface means for communicating said wave-
form to a sensor port of said monitor.

18. The communications adapter according to claim 17
further comprising a signal processor means for deriving a
parameter signal from said conditioned signal, said data
comprising said parameter signal.

19. The communications adapter according to claim 18
wherein said waveform processor means comprises a means
for synthesizing said waveform from said parameter signal.

20. The communications adapter according to claim 17
wherein said data comprises said conditioned signal and said
waveform processor means comprises a means for modu-
lating said conditioned signal in response to said monitor.

* * * * *

专利名称(译)	生理测量通信适配器		
公开(公告)号	US20030181798A1	公开(公告)日	2003-09-25
申请号	US10/377933	申请日	2003-02-28
[标]申请(专利权)人(译)	AL ALI AMMAR		
申请(专利权)人(译)	AL-ALI AMMAR		
当前申请(专利权)人(译)	AL-ALI AMMAR		
[标]发明人	AL ALI AMMAR		
发明人	AL-ALI, AMMAR		
IPC分类号	A61B5/00 A61B5/024		
CPC分类号	A61B5/02438 A61B5/1455 A61B5/6826 A61B5/6838 A61B5/0026 A61B5/002 A61B5/02427 A61B5/14552 Y10S128/903		
优先权	60/367428 2002-03-25 US		
其他公开文献	US6850788		
外部链接	Espacenet	USPTO	

摘要(译)

传感器接口配置为接收传感器信号。发射器响应于传感器信号调制第一基带信号，以便产生发射信号。接收器解调对应于发送信号的接收信号，以便产生对应于第一基带信号的第二基带信号。此外，监视器接口被配置为响应于第二基带信号将波形传送到监视器的传感器端口。波形适用于监视器，因此监视器从波形得到的测量值通常等于可从传感器信号导出的测量值。通信适配器还可以包括信号处理器，其具有与传感器接口通信的输入，其中信号处理器可操作以响应于传感器信号导出参数，并且第一基带信号响应于参数。该参数可以对应于测量的氧饱和度和脉搏率中的至少一个。

