




US 20130310699A1

(19) **United States**(12) **Patent Application Publication**
Hart et al.(10) **Pub. No.: US 2013/0310699 A1**(43) **Pub. Date: Nov. 21, 2013**(54) **PATIENT MONITORING METHOD AND
MONITORING DEVICE****Publication Classification**(71) Applicants: **King's College London, of The Strand;
Guy's and St. Thomas' NHS
Foundation Trust, of The Counting
House, Guy's Hospital, London (GB)**(51) **Int. Cl.**
A61B 5/0205 (2006.01)
A61B 5/00 (2006.01)
A61B 5/0472 (2006.01)
A61B 5/0488 (2006.01)
A61B 5/04 (2006.01)(72) Inventors: **Nicholas Hart, London (GB); John
Moxham, London (GB); Fiammetta
Fedele, London (GB)**(52) **U.S. Cl.**
CPC *A61B 5/0205* (2013.01); *A61B 5/0488*
(2013.01); *A61B 5/04012* (2013.01); *A61B*
5/0472 (2013.01); *A61B 5/7203* (2013.01)
USPC **600/484; 600/529**(73) Assignees: **King's College London, of The Strand;
Guy's and St. Thomas' NHS
Foundation Trust, of The Counting
House, Guy's Hospital, London (GB)**(21) Appl. No.: **13/627,383**(22) Filed: **Sep. 26, 2012****Related U.S. Application Data**(60) Provisional application No. 61/541,708, filed on Sep.
30, 2011.(30) **Foreign Application Priority Data**

Sep. 30, 2011 (GB) 1116860.6

(57) **ABSTRACT**

A method of monitoring a patient includes measuring neural respiratory drive using a monitoring device (10), repeating the measurement either continuously or at regular time intervals, and comparing the measurements obtained in order to predict treatment failure and/or clinical deterioration and/or re-admission. In embodiments of the invention, the neural respiratory drive is measured by obtaining a measure of the second intercostal space parasternal electromyogram. A monitoring device (10) includes a signal input (20), a processing unit (30), and an output unit (50), and is arranged to measure the neural respiratory drive, store the measured value and compare it to a previously measured value for the neural respiratory drive.



Electrode placement for parasternal EMG




Fig. 1. Electrode placement for parasternal EMG




Fig. 2. Example of parasternal EMG trace




Fig. 3




Fig. 4




Fig.5




Fig. 6




Fig. 7




Figure 8

Patient ID

8888

Signal Quality

BAD

Date

20102010

Artifacts removal

Sniff

☒ Sniff acquired

Sniff Quality

GOOD

(a)

Sniff

Avg HR [bpm]

67

Avg rms [mV]

35

Avg RespR [brpm]

29

Avg rms % Sniff

29

Avg rms % Sniff x RespR

955

Avg rms area % Sniff

78

Avg rms area [mVs]

20

Avg rms area % Sniff

2302


(c)

START

Comments


Write here your comments

Acquired signal




(b)


% Trends



% Rms Sniff



% Rms Sniff x RespR



Save data location:

Exit

(a) Record details

- Signal quality assurance
- electronic patient record

(b) Electromyogram raw signal with trend monitor of neural respiratory drive

(c) Neutral respiratory drive parameters

(d) Quality assurance for Sniff electromyogram signal

Figure 9





Fig.11




Fig. 12A




Fig. 12B




Fig. 13A




Fig. 13B




Fig. 13C




Fig. 13D




Fig. 14A




Fig. 14B




Fig. 15A




Fig. 15B




Fig. 16




Fig. 17





Fig. 18



PATIENT MONITORING METHOD AND MONITORING DEVICE

CROSS-REFERENCE

[0001] This application claims the benefit of U.S. Provisional Patent Application No. 61/541,708, filed Sep. 30, 2011, and GB 1116860.6, filed Sep. 30, 2011, which applications are incorporated herein by reference in their entirety.

BACKGROUND

[0002] Various abbreviations are used throughout this application and for ease of reference those most frequently used are set out below:

- [0003] EMG—electromyogram
- [0004] ECG—electrocardiogram
- [0005] COPD—chronic obstructive pulmonary disease
- [0006] AECOPD—acute exacerbation of COPD
- [0007] HR—Heart Rate
- [0008] NRD—neural respiratory drive
- [0009] NRDI—neural respiratory drive index
- [0010] NRDT—neural respiratory drive time product
- [0011] NRDTI—neural respiratory drive time index
- [0012] RR—respiratory rate
- [0013] RMS—root mean square

[0014] The use of physiological biomarkers to monitor and track clinical change in patients in both acute care organisations and in the community is an area of increasing interest to health care providers. Early detection of clinical deterioration and assessing the response to treatment correspond to improved clinical outcomes in patients. Despite the rationale for this approach, there is little published evidence to support the use of the current basic physiological monitoring systems available for detection of deterioration in patients are sufficiently sensitive or specificity for early detection (Hillman et al. (2005) *Lancet* 365, 2091-7).

[0015] Whilst heart rate (HR) and respiratory rate (RR) are commonly used clinical physiological variables in acute care, measurements of neural respiratory drive (NRD) and neural respiratory drive index (NRDI) are advanced physiological biomarkers that have been shown to have greater sensitivity and specificity in assessing the intensity, timing and duration of respiratory effort, which is defined by the balance between respiratory muscle load and capacity (Duiverman et al. (2004) *J. Appl. Physiol.* 96, 1723-9; Jolley et al. (2009) *Eur. Respir. J.* 33, 289-97; Steier et al. (2009) *Thorax* 64; 719-25).

[0016] Whilst this technique is an established research technique (Jolley et al. (2009); Steier et al. (2011) *Thorax* 66, 609-614; Murphy et al. (2011) *Thorax* 66, 602-8, incorporated herein by reference in its entirety), the limitations set out above have prevented it from becoming useful as a clinical tool. Indeed, the time involved in undertaking the analysis has limited its applicability to research situations. Whilst generic devices exist to perform the first steps of the process there is no known system available in clinical use to display NRD and NRDI as a physiological biomarker. One system utilises surface electrodes to measure EMG activity (Duiverman et al. (2004)), however, the commercial device has not been robust or reliable enough to move from the research to the clinical arena. Furthermore, the processing of the signal displays a log ratio of the EMG activity, which has no clinical advantage. In addition, this method has not focussed on an approach of normalisation of the EMG_{para} signal to the maximal EMG_{para} signal manoeuvre. These factors have prevented it from becoming a useful clinical tool.

SUMMARY

[0017] The present disclosure relates to a patient monitoring method and device. In particular, it relates to a method of, and device for, monitoring a patient with severe respiratory disease, and predicting the likelihood of both clinical deterioration and hospital re-admission.

[0018] According to an aspect of the present disclosure, there is provided a method of monitoring a patient, including measuring neural respiratory drive, repeating the measurement continuously and/or at regular time intervals, and comparing to the initial values measured in order to predict treatment failure and/or clinical deterioration and/or re-admission.

[0019] The neural respiratory drive may be measured by obtaining a measure of the second intercostal space parasternal electromyogram.

[0020] The method may include:

[0021] 1) obtaining a value for a normalised neural respiratory drive (which may be calculated as $EMG_{para}/EMG_{para\max}$ and termed $EMG_{para\%max}$);

[0022] 2) obtaining the respiratory rate of the patient, expressed as breaths per minute (RR/min);

[0023] 3) obtaining a value for the neural respiratory drive index by multiplying the value of the neural respiratory drive obtained in Step 1 by the respiratory rate ($EMG_{para\%max} \times RR$);

[0024] 4) repeating Steps 1 to 3 continuously and/or after a first given period of time and comparing the two neural respiratory drive index values obtained.

[0025] In an example, Steps 1 to 3 may be carried out upon admission into hospital and repeated just prior to discharge from hospital.

[0026] The method may further include obtaining a value for the neural respiratory drive time index by multiplying the value of the neural respiratory drive time product by the respiratory rate.

[0027] In an embodiment, the method includes:

[0028] 1) obtaining a value for a normalised neural respiratory drive by:

[0029] a) carrying out parasternal electromyography to obtain a raw signal during normal breathing;

[0030] identifying and obtaining the root mean square of the raw signal to obtain a rectified trace;

[0031] identifying and obtaining the peak magnitude of the rectified trace for each inspiration;

[0032] calculating the mean of the peak magnitudes identified ($EMG_{para\%peak}$);

[0033] b) carrying out parasternal electromyography to obtain a raw signal during at least two (or at least three) sniff maximal manoeuvres;

[0034] obtaining the root mean square of the raw signal to obtain a rectified trace;

[0035] identifying and selecting the peak magnitude of the rectified trace ($EMG_{paramax\%peak}$);

[0036] c) expressing the mean of the peak calculated in Step 1a as a percentage of the peak magnitude selected in Step 1b ($EMG_{para\%max\%peak}$);

[0037] 2) obtaining the respiratory rate of the patient, expressed as breaths per minute;

[0038] 3) obtaining a value for the neural respiratory drive index by multiplying the value of the $EMG_{para\%max\%peak}$ (NRD) obtained in Step 1c by the respiratory rate;

[0039] 4) repeating Steps 1 to 3 continuously and/or after a first given period of time and comparing the two neural respiratory drive index values obtained.

[0040] In one embodiment, the first period of time is selected from the range of 8 to 24 hours. For example, an initial measurement may be taken upon hospital admission, with a second measurement being taken 8 to 24 hours after admission. Optionally, the measurement and comparison is performed continuously or repeatedly at an approximate frequency of the first period of time.

[0041] Additionally or alternatively, the method may include: obtaining a value for the neural respiratory drive time product by carrying out parasternal electromyography to obtain a raw signal during normal breathing; obtaining the root mean square of the raw signal to obtain a rectified trace;

[0042] measuring the area under the rectified trace;

[0043] obtaining a value for the neural respiratory drive time product index by multiplying the neural respiratory drive time product by the respiratory rate; and

[0044] comparing the two neural respiratory drive time index values obtained.

[0045] The method may include identifying and obtaining the area under the curve of the rectified root mean square of the raw signal; calculating the mean of the area under the curves identified ($EMG_{paraAUC}$); identifying and selecting the area under the curve of the rectified trace ($EMG_{paramaxAUC}$); expressing the mean area under the curve as a percentage of the maximum area under the curve ($EMG_{para\%maxAUC}$) and obtaining neural respiratory drive time index by multiplying the value of the $EMG_{para\%maxAUC}$ (NRDTP) by the respiratory rate.

[0046] In embodiments, the patient has respiratory disease, (for example, cardiorespiratory disease). The respiratory disease may be (1) an acute exacerbation of chronic obstructive pulmonary disease; (2) an acute exacerbation of chronic respiratory disease; (3) acute respiratory failure; (4) chronic respiratory disease; (5) chronic respiratory failure; (6) acute exacerbation of chronic heart failure (7) acute heart failure and (8) chronic heart failure, for example.

[0047] The method may be used to predict clinical deterioration and/or the likelihood of hospital readmission within 28 days of discharge.

[0048] In an embodiment, the peak magnitude is obtained for each inspiration over a time period of approximately 30 seconds to 3 minutes.

[0049] In some cases, the neural drive index values and/or the neural respiratory drive time product index values are obtained upon admission to hospital, intermittently or continuously throughout admission until the day of discharge from hospital. The NRD, NRDI, NRDTP and NRDTI may alternatively be obtained in the community or home setting either intermittently at various time points (for example, each day after hospital discharge) or continuously.

[0050] In an exemplary embodiment, the method includes obtaining an electrocardiography signal, and removing artifacts from the electrocardiography signal from the raw parasternal electromyography trace.

[0051] As an example, the monitoring can be carried out in real time.

[0052] According to another aspect of the present disclosure, there is provided a monitoring device including a signal input, a processing unit, and an output unit, the monitoring device being arranged to:

[0053] (1) receive a first raw parasternal electromyography signal at the signal input;

[0054] (2) determine the root mean square of the raw parasternal signal to obtain a rectified trace;

[0055] (3) identify the peak magnitude of the rectified parasternal trace for each inspiration over a second given period of time;

[0056] (4) calculate the mean of the peak magnitudes identified;

[0057] (5) receive further and determine raw parasternal electromyography signals during at least two (or at least three) sniff manoeuvres;

[0058] (6) determine the root mean square of the raw signal to obtain a further rectified trace;

[0059] (7) determine the peak magnitude of the further rectified trace;

[0060] (8) express the mean of the first peak magnitude as a percentage of the peak magnitude of the further rectified trace obtained during the sniff manoeuvre;

[0061] (9) receive data on the respiratory rate of the patient, expressed as breaths per minute;

[0062] (10) determine a value for the neural respiratory drive index by multiplying the value of the neural respiratory drive by the respiratory rate;

[0063] (11) store the measured/determined values of the neural respiratory (for example, NRD and NRDI, and optionally NRDTP and NRDTI) in the data repository.

[0064] In some embodiments, the monitoring device is arranged to determine a subsequent neural respiratory drive index and compare with the stored value.

[0065] In some embodiments, the monitoring device is also operable to identify the area under the curve of the rectified root mean square of the raw parasternal signal for each inspiration over a second given period of time; to calculate the mean of the area under the curves identified; and to determine the values for NRDTP and NRDTI from the area under the curve of the rectified trace.

[0066] In some cases, the monitoring device is operable to display the measured/determined neural respiratory drive values (for example, NRD and NRDI, and optionally NRDTP and NRDTI) in real time.

[0067] Embodiments of the present disclosure provide for substantially real time processing of biological signals received from surface electrodes and convert them into a clinically useful physiological biomarker. The signals from these electrodes are processed in one embodiment to obtain and display heart rate (HR), respiratory rate (RR), neural respiratory drive (NRD), neural respiratory drive index (NRDI), and optionally neural respiratory drive time product (NRDTP) and neural respiratory drive time index (NRDTI). Whilst HR and RR are commonly used clinical physiological variables in acute care, the NRD and NRDI have been shown to be more sensitive markers of neural respiratory drive and respiratory effort and have previously been shown to correspond to the balance between the respiratory muscle load and the respiratory muscle capacity. Additionally or alternatively NRDTP and NRDTI may be used.

[0068] NRD can be determined from the electromyogram of the 2nd intercostal parasternal muscles (EMG_{para}). These are acquired using electrodes and amplifiers and the signals are processed using analogue to digital conversion followed by digital filtering and arithmetic conversion of the signal.

[0069] In one embodiment, signals are recorded during resting breathing to provide an indication of the patient's current respiratory effort determined by objectively measuring the EMG_{para} activity. At the end of recording, the patient is asked to perform repeated maximum sniff manoeuvres in

order to allow the signal to be normalised for an individual patient maximum effort ($EMG_{para\%max}$).

[0070] In one embodiment, the process of acquiring these biological signals and producing clinical useful data is divided into a four step process, with the final step of integrating and processing the parasternal electromyogram. The steps involved are detailed below:

[0071] 1. Patients at potential risk of deterioration are identified (for example, upon acute admission to hospital) by clinical staff and have surface electrodes placed over the parasternal muscles of the second intercostal space along with a reference electrode over the electrically neutral clavicle.

[0072] 2. Electrical signals from the two recording electrodes are amplified. The signal is amplified to a factor of 1000 and analogue filtered at 10 Hz and 2000 Hz to remove contributions from other muscle activity and maximise signal from the parasternal muscles.

[0073] 3. The amplified signal is passed to an analogue to digital converter to allow for further computer processing.

[0074] 4. The final step of the process is to integrate the signals to produce the essential biological variables for clinical interpretation. The initial signal is assessed for signal quality and advises if repositioning of the electrodes is required. If signal quality is adequate the patient performs a series of maximum sniff manoeuvres in order to normalise the signal for individual variations in subcutaneous fat distribution that can alter signal strength. Once the system has detected and analysed the maximum manoeuvres the patient undergoes a period of testing that comprises of relaxed breathing at which time continuous measures of HR, RR and NRD, NRDI, and optionally NRDTTP and NRDTI are displayed. The EMG_{para} signal is processed to remove ECG artefact by a simple 20 Hz digital filter. As the signal comprises of positive and negative deflections further processing occurs to produce a representative value by converting the EMG_{para} to a root mean squared (RMS) with a moving window of 40 ms. As well as displaying the previously used marker of NRD (EMG_{para}), NRDI is displayed that incorporates respiratory rate to provide a measure of neurological drive to breathe over a minute rather than the per breath measure provided by EMG_{para} .

[0075] In selected embodiments, some or all of the following features may apply or be included:

[0076] Signal quality assessment

[0077] ECG (QRS) detection and HR calculation

[0078] ECG artefact removal from EMG_{para} signal

[0079] Calculation of the RMS

[0080] EMG_{para} analysis including peak-peak analysis to calculate respiratory rate, peak value to calculate $EMG_{para\%maxpeak}$ and multiplication by respiratory rate to calculate NRDI

[0081] EMG_{para} analysis to calculate area under the curve of the signal to calculate $EMG_{para\%maxAUC}$ by normalizing for the $EMG_{paramaxAUC}$ and multiplying by respiratory rate to calculate NRDTI

[0082] Other potential uses in both hospital and home setting include:

[0083] Home:

[0084] Objective measure of breathlessness in patients with chronic respiratory disease, for example COPD

[0085] Identification of patients undergoing clinical deterioration, for example, an exacerbation of COPD

[0086] Facilitation of the out of hospital set up of patients with chronic respiratory failure (for example, COPD, neuromuscular disease and obesity hypoventilation syndrome) requiring domiciliary non-invasive ventilation allowing improved set up without hospital admission, of particular benefit in conditions associated with a poor prognosis. The proposed approach is to optimise patient-ventilator interaction using NRD, NRDI, and optionally NRDTTP and NRDTI as an objective physiological biomarker.

[0087] Hospital:

[0088] Monitoring respiratory deterioration in acute critical illness

[0089] Facilitate setup of domiciliary non-invasive ventilation in patients with chronic respiratory failure

[0090] Facilitate setup for acute non-invasive ventilation in patients with acute and acute on chronic respiratory failure

[0091] Stratification of patients that are high risk of re-admission

[0092] Potential benefits of selected embodiments include:

[0093] Real time display of important clinical parameters

[0094] Data log for trend display of changes in parameters

[0095] Single tool for six clinical parameters (HR, RR, NRD, NRDI, NRDTTP and NRDTI)

[0096] Signal analysis and processing to allow removal of ECG artefact and computing neural respiratory drive

[0097] Use of neural respiratory drive as a clinical respiratory biomarker

BRIEF DESCRIPTION OF THE DRAWINGS

[0098] Exemplary embodiments of the present disclosure are described below with reference to the accompanying drawings, in which:

[0099] FIG. 1 illustrates exemplary electrode placement locations;

[0100] FIG. 2 illustrates a sample signal in both raw and partially processed states;

[0101] FIG. 3 is a schematic diagram of a monitoring device according to an embodiment of the present disclosure;

[0102] FIGS. 4 to 7 are illustrations of signals during processing of an embodiment of the present disclosure;

[0103] FIG. 8 is a flow diagram illustrating aspects of filtering and subsequent steps of a method according to an embodiment of the present disclosure;

[0104] FIGS. 9A-D are screen shots of an output produced in an embodiment of the present disclosure;

[0105] FIGS. 10A and 10B show Bland-Altman analysis of EMG_{para} measured in healthy volunteers on two separate test days (A) with EMG_{para} shown in μV with 95% upper and lower limits of agreement indicated and $EMG_{para\%max}$ (B) measured in stable COPD patients attending pulmonary rehabilitation on two separate occasions with $EMG_{para\%max}$ shown with 95% upper and lower limits of agreement;

[0106] FIG. 11 shows daily changes in (A) $EMG_{para\%max}$ and (B) NRDI during the course of admission between patients designated as 'improvers' or 'deteriorators' during the first 24 hours of study participation (plotted as mean $\pm$ standard error of the mean);

[0107] FIGS. 12A and 12 B show changes in NRD during hospital admission;

[0108] FIGS. 13A-D compare ABorg score with ΔEMG_{para} , $\Delta\text{EMG}_{para\%max}$, ΔNRDI and ΔFEV_1 ; and

[0109] FIGS. 14A, 14B, 15A, and 15B show receiver operating characteristics plots.

[0110] FIGS. 16, 17, and 18 depict measurements taken and expected outcomes for patients 1, 2, and 3, respectively.

DETAILED DESCRIPTION

[0111] NRDI is calculated from the electromyogram of the 2nd intercostal parasternal muscles (EMG_{para}). These signals can be acquired using conventional electrodes and amplifiers. The signals may be processed using basic analogue to digital conversion followed by digital filtering and arithmetic conversion of the signal. The subsequent digitised and converted signal is hand analysed discounting sections of the trace that have interference from the electrical signals of the heart muscle, or electrocardiogram (ECG).

[0112] This technique requires the signals to be recorded during resting breathing to provide an indication of the patient's current respiratory effort based on their EMG_{para} activity. At the end of recording, the patient is asked to perform repeated maximum sniff manoeuvres in order to allow the signal to be normalised for an individual patient maximum effort ($\text{EMG}_{para\%max}$). The data are then manually analysed off-line at a later time with each breath being manually marked, measured and breaths contaminated by ECG artefact being removed.

[0113] The device 10 includes a signal input 20, a processing unit 30, a data repository 40 and an output unit 50.

[0114] The device 10 receives a signal at the signal input 20 that has been measured from the parasternal electromyogram (EMG).

[0115] The signal is passed to the processing unit 30 where the raw, biphasic parasternal EMG signal is rectified using a root mean square operation as shown in FIG. 4. Preferably, a window is used to segment the signal and simplify RMS calculation.

$$\text{rms}(n) = \sqrt{\frac{\sum_{i=n-M}^{n+M} \text{resp}^2(i)}{2M}} \quad \text{with } M = 25 \text{ ms} \times Fs = 50 \text{ samples.}$$

[0116] The peak magnitude ($\text{EMG}_{para\text{peak}}$) of the rectified rmsEMG_{para} trace is then identified by the processing unit 30 for each inspiration over 30 seconds to 3 minutes as shown in FIG. 5.

[0117] The mean of the $\text{EMG}_{para\text{peak}}$ values is then calculated for all inspirations over a 30-second to 3 minute time period. This is termed mean ($\text{EMG}_{para\text{peak}}$).

[0118] The magnitude of the rmsEMG_{para} trace during a maximal inspiratory manoeuvre is recorded as shown in FIG. 6.

[0119] In practice, the patient performs several sniff manoeuvres. The manoeuvre that results in the greatest magnitude of rmsEMG_{para} is selected. This value is termed $\text{EMG}_{para\text{max}}$.

[0120] The neural respiratory drive is represented by the quantity $\text{EMG}_{para\%max}$. This is derived by expressing the mean ($\text{EMG}_{para\text{peak}}$) as a percentage of the $\text{EMG}_{para\text{max}}$:

$$\text{EMG}_{para\%max} = [\text{mean}(\text{mean}(\text{EMG}_{para\text{peak}}) / \text{EMG}_{para\text{max}}) \times 100\%] \quad (1)$$

[0121] The Neural Respiratory Drive Index (NRDI) is then calculated by the processing unit 30.

[0122] NRDI is the product of $\text{EMG}_{para\%max}$ and the respiratory rate (RR, breaths per minute)

$$\text{NRDI} = \text{EMG}_{para\%max} \times \text{RR} \quad (2)$$

[0123] The area under the curve of the rmsEMG_{para} trace, termed neural respiratory drive time product; NRDTP, is then calculated as shown in FIG. 7. This approximation is used to enable substantially real time processing, although it will be appreciated that more accurate measurements could be made. NRDTP indicates of the total electrical activity of the muscle during inspiration, a non-invasive surrogate for the work of breathing.

[0124] The neural respiratory drive time index (NRDTI) is the product of NRDTP and respiratory rate normalised for the $\text{EMG}_{para\text{maxAUC}}$:

$$\text{NRDTI} = \text{NRDTP} \times \text{RR} \quad (3)$$

[0125] NRDTI more accurately reflects the total neural respiratory drive compared to NRDI, as the latter takes into account only the peak value of the parasternal EMG signal during each inspiration. By contrast, the area under the rmsEMG_{para} curve represents the electrical activity of the parasternal muscles during the whole of inspiration.

[0126] The device 10 automatically calculates the NRDTI and stores it in the data repository 40. Subsequent measurements and calculations are stored in the data repository and compared to earlier values of NRDTI. A reduction in value is indicative of reduced likelihood of re-admission. A consistent value or an increase is indicative of higher likelihood of re-admission.

[0127] The presence of ECG (cardiac) signal artefact in the parasternal EMG trace significantly affects accurate analysis of neural respiratory drive.

[0128] In exemplary embodiments of the present disclosure, the processing unit 30 applies a filtering algorithm to remove the ECG artefact from the raw parasternal EMG trace. Aspects of the algorithm are shown in FIG. 8. A high pass filter is applied to the raw signal to remove baseline noise.

[0129] Additional band pass filtering (BPF) between 12-20 Hz is performed to condition the signal spectral content to remove respiration artefacts prior to ECG detection when high levels of neural respiratory drive are present. This additional filtering is based on the principle that the electromyogram is a broad-band signal, whilst the QRS complex of the ECG main spectral content is below 20 Hz and the empirical observation (from previous patient data) that patient with low NRDI have EMG_{para} signals with a spectral content above 0.4 Hz less than 60%.

[0130] ECG detection can be carried out by comparison of a non-linear derivation of the first derivative of the filtered EMG_{para} signal with a threshold. The algorithm is an implementation of a QRS complex detection algorithm which is particularly suitable for signals (such as EMG_{para}) affected by motion artefacts. The specific threshold ($\text{Speak} = 0.0001$) used by the algorithm was chosen empirically, as best performing (90% Sensitivity) in the range 10^{-5} - 10^{-2} when testing the algorithm on 52 different blocks of data of 20-30 s length.

[0131] In some patients residual influence of QRS peaks would lead to outlier peaks in the rmsEMG_{para} , which can be removed by the activation of an optional routine.

[0132] The routine examines the distribution of the peaks in a window of 30 s of rmsEMG_{para} signal, calculates their 99.3 percentile (P993, a value above 99.3% of all peaks) and compares it with the maximum in the distribution (rmsEMG_{para}).

$paramax$). If $P993/rmsEMG_{paramax} < 0.7$ then all the peaks above P993 are considered outliers and eliminated.

[0133] Knowledge of the location of the QRS peaks is used to eliminate the contribution of a patient's ECG to a EMG_{para} signal. In particular, following the example of previous studies only a portion of signal between two R peaks is kept. The portion of rectified signal between 30%-75% of each RR interval is considered as respiration only and is kept unmodified. For example, if one R peak is at 10 s and the following one is at 10.8 s (R-R interval=800 ms), then the portion of signal between 10.24 s (10 s+30% of 800 ms) and 10.6 s (10 s+75% of 800 ms) it is left unchanged. A larger interval than previous studies was used to avoid loss of data. The portions of signals corresponding to QRS complexes are replaced by linear interpolations connecting adjacent unmodified data segments. In the designing phase zero padding, zero order interpolation and the use of the root means square of two adjacent segments were also attempted, but discarded. This is because these options were affecting the signal morphology to a point where each respiration peak would be split in two and the software would return a doubled respiration rate.

[0134] It will be appreciated that the use of a real time measure of NRD could be used not only on other patients with acute respiratory illness but also in the long term management of chronic disease. The technique is simple and quick to perform and is totally painless for the patient and could be applied in clinic patients for the monitoring of chronic respiratory disorders such as COPD, obesity related respiratory failure (for example, obesity hypoventilation syndrome and hypercapnic obstructive sleep apnoea), asthma, bronchiectasis, neuromuscular disease and interstitial lung disease in order to more sensitively track clinical change. The technique also offers the opportunity to closely monitor and optimise patient-ventilator interaction in patients receiving both acute and domiciliary non-invasive ventilation (NIV). NIV confers significant clinical outcome benefits to patients with acute and chronic respiratory failure but its use can be limited by poor patient tolerance as a consequence of poor adherence to the ventilator prescription resulting from poor patient-ventilator interaction. The ability to match patient and ventilator effort would offer the opportunity of improving patient comfort and thus improve the adherence to therapy. The potential uses in the home setting are:

[0135] Objective measure of breathlessness in patients with chronic respiratory disease for example, COPD

[0136] Identification of patients undergoing clinical deterioration for example, an exacerbation of COPD

[0137] Facilitation of the out of hospital set up of patients requiring domiciliary NIV allowing improved set up without hospital admission, of particular benefit in conditions associated with a poor prognosis.

[0138] In one embodiment, a monitoring device may be connected to a patient in the home and measurement communicated (for example via a mobile data connection, WIFI etc.) back to a server at the hospital for monitoring, analysis which would be supported by a clinical team utilising a clinical decisions algorithm leading to clinical intervention. For example, the patient can be monitored remotely and if the patient's clinical condition were to deteriorate this would be identified by a change in RR, HR, NRD, NRDI, and/or NRDT, NRDTI as described above with an alert communicated to the clinical team. The clinical team would alert the patient and/or carer either through the device or directly to advise the patient and/or carer to seek medical attention.

Alternatively, or in addition, an automated call could be made to a designated phone number to provide an alert. Appropriate medical intervention can then be carried out.

[0139] The potential uses are in the hospital setting are:

[0140] Monitoring respiratory deterioration in acute critical illness

[0141] Facilitate setup of NIV for domiciliary NIV

[0142] Facilitate setup for acute NIV

[0143] Stratification of patients that are high risk of re-admission

[0144] Where a patient is identified as being at high risk of hospital re-admission, changes can be made to their medical care so that it can be optimised; hospital discharge could be delayed; and/or after outreach support may be intensified (for example a carer may visit twice daily instead of only once).

[0145] The monitor display produced, as shown in FIG. 9, allows the monitoring and recording of 'resting' breathing and the 'Sniff manoeuvre'.

[0146] The patient is identified by a unique 'Patient ID' integer number (FIG. 2-(a)), that the operator can choose; the monitoring 'Date' is automatically set-up by the program.

[0147] The monitoring device is set to collect 5 s of data a time at a frequency of 2 kHz such as by a National Instruments USB DAQ device (NI BNC 6221) and return an immediate feedback on 'Signal Quality', based on a comparison of detectable respiratory activity and background noise level.

[0148] Once at least 30 s of resting data have been acquired the monitor will display the following signal features:

[0149] Avg HR [beats per minute]: average heart rate in beats per minutes in the last 30 s to 3 minute sampling period

[0150] Avg RR [breaths per minute]: average respiration rate in breath per minute in the last 30 s to 3 minute sampling period

[0151] Avg rms [mV]: average peak root mean square in mVolts in the last 30 s to 3 minute sampling period

[0152] In case of data of poor quality with low NRD, only the HR might be extractable, in such cases a 'NaN' (Not a Number) will be returned for the other parameters. Whenever possible, it is advised not to proceed with data of 'BAD' quality and to check the electrode contact to assess if it possible to gain data of 'GOOD' quality.

[0153] The quality of a resting segment is considered 'GOOD' if the EMG_{para} signal is easily distinguishable from noise. The algorithm calculates a threshold equal to 40% of a smoothed EMG_{para} signal and determines if it can detect EMG_{para} peaks above this threshold. If EMG_{para} peaks are detectable, the signal is considered of 'GOOD' quality and an average RR is estimated from the intervals between the detected peaks. Otherwise the signal is considered of 'BAD' quality and the RR is set equal to 'Not a number' (NaN).

[0154] As signal quality is updated every 5 s, one can expect samples with no respiration activity, therefore flagged as 'BAD', during a continuous acquisition. Whenever the operator is flagged constantly 'BAD' quality the acquisition should be interrupted and electrode contact with the patients and the rest of the acquisition variables should be checked to improve signal quality.

[0155] Once the operator (for example, nurse, doctor, technician) is satisfied that the data are 'GOOD' quality, the patient will be asked to perform a 'sniff manoeuvre', and 10 s to 20 s of 'sniff manoeuvre' data should be collected pressing the 'START SNIFF' button in the 'Sniff' sub-panel (FIG. 2-(d); button not shown in figure).

[0156] The same features as above are returned in a 'sniff manoeuvre' session if the sniff signal is of 'GOOD' quality, however, in case of the 'sniff manoeuvre', peak values and area under curve rather than average peak values are considered.

[0157] Once 'GOOD' quality sniff data have been acquired, the operator can proceed with monitoring of resting traces and receive the additional feedback on NRD, NRD_I, NRD_TP and NRD_TI for the last recorded 30 s:

[0158] FIG. 9 is a screen shot of a possible user interface produced in one embodiment of the present disclosure. The interface is organised into different blocks.

[0159] Block (a) reports the treatment details (Patient ID and Date), the signal quality, and a 'checkbox' that enables further removal of ECG artifacts from NRD signals.

[0160] Block (b) shows the last 30 s of signal acquired on the left, and the trends of the two important NRD features on the right: '% rms sniff' (30 s average $EMG_{peak}/EMG_{paramax}=NRD$) and '% rms sniff $\times$ RR (30 s average $EMG_{peakpara}/EMG_{paramax}\times RR=NRD_I$) since monitoring started. These trends are only shown after a 'sniff' EMG_{para} signal has been acquired.

[0161] Block (c) reports values averaged on 30 s of all the features extracted: Avg HR (average heart-rate), Avg rms (average peak rms), Avg rms area (average area under the rms curve), Avg RespR (average respiratory rate), Avg rms % sniff (average rms normalized to the 'sniff' peak rms value), Avg rms area % sniff (average area normalized to the 'sniff' peak rms value), Avg rms % sniff $\times$ RespR (Average rms % sniff multiplied by the respiratory rate), Avg rms area % sniff $\times$ RespR (% Average rms area % sniff multiplied by the respiratory rate). The last four parameters are available only if a sniff signal has been acquired.

[0162] Block (d) shows the information about the sniff signal acquired. Prior to 'sniff' acquisition it also contains a 'START Sniff' button that allows the acquisition of a 10 s 'sniff' trace. The interface also has a 'comment' box and one showing where the acquired data are saved. A 'START' button allows the beginning of a monitoring phase.

[0163] The applicant has shown that NRD_I is a feasible clinical physiological biomarker in patients with an acute exacerbation of chronic obstructive pulmonary disease (COPD), which can provide useful information on treatment response and risk of hospital readmission (Murphy et al. (2011)). It has shown that the technique of measuring NRD_I to monitor patients with COPD has potential across acute care services, critical care services and community services. In patients with COPD, the signals have been shown to be wholly reproducible and more sensitive and specific than standard clinical parameters at monitoring clinical change, and more importantly clinical deterioration. NRD_I was also shown to be useful in risk stratifying readmission in COPD patients following discharge from hospital following an acute exacerbation.

[0164] The use of the above-described monitoring device and method has a number of advantages: (1) it is less time consuming; (2) data points are separated from the ECG signal and are therefore not lost; (3) signal analysis and clinical interpretation are performed in real time; and (4) the opportunity to modify the clinical management if the NRD signal does not change in response to treatment is provided.

[0165] At least a portion of the embodiments described herein may be performed by a process defined by code executing within a processor of a computer, state machine of suitably configured hardware based processor such as a field programmable gate array (FPGA). The code can comprise software instructions that are provided to a physical memory that is accessible to the processor of the computer. The code can be arranged as firmware or software, and can be organized as a set of modules such as discrete code modules, function calls, procedure calls or objects in an object-oriented programming environment. If implemented using modules, the embodiment can comprise a single module or a plurality of modules that operate in cooperation with one another.

[0166] It will also be appreciated that embodiments of the present disclosure are possible in which processing may be done locally (for example at a hospital bedside, in a patient's home using portable measuring equipment) or remotely (in which case the monitoring device may simply capture raw data and transmit back to a remote server such as via ftp, a secure Web interface or some other communication means and then processed at the server) In one embodiment, a centralized user interface, for example in the form of an on-screen dashboard, can provide filtered up-to-date results to medical staff for a number of patients, both local and/or remote so that semi-automated tracking of status can be done and fed to a remote station for review.

[0167] Optional embodiments of the disclosure can be understood as including the parts, elements and features referred to or indicated herein, individually or collectively, in any or all combinations of two or more of the parts, elements or features, and wherein specific integers are mentioned herein which have known equivalents in the art to which the invention relates, such known equivalents are deemed to be incorporated herein as if individually set forth.

[0168] Although illustrated embodiments of the present invention have been described, it should be understood that various changes, substitutions, and alterations can be made by one of ordinary skill in the art without departing from the present invention which is defined by the recitations in the claims below and equivalents thereof.

EXAMPLES

Example 1

Validation of Reproducibility of EMG_{para}

[0169] Methods

[0170] Subjects

[0171] EMG_{para} measurements in healthy volunteers were taken one week apart. EMG_{para} measurements from stable COPD patients were recorded at consecutively attended classes, 3 to 4 days apart.

[0172] EMG_{para} Recording and Data Processing

[0173] The second intercostal space was identified using bony landmarks and skin was prepared with EMG contact gel (Nuprep, DO Weaver and Co, USA). Wet gel electrodes (Neuroline 720, Ambu, Denmark) were placed adjacent to the sternal edge in the second intercostal spaces. The signal was amplified and processed using a high differential amplifier with band pass filters set at 10 Hz and 2000 Hz (1902, Cambridge Electronic Design, Cambridge, UK). Additional analogue 50 Hz notch filter and AC coupling were used. Amplified signals were passed to an analogue to digital convertor (Powerlab, ADInstruments, Chalgrove, UK) and passed to a

personal computer. Further digital filtering occurred at 20 Hz after data acquisition (LabChart v7.1, ADInstruments, Chalgrove, UK). EMG_{para} recordings were performed with the patient relaxed in a chair or semi-recumbent in bed with arms supported. EMG_{para} signals were acquired during resting breathing for at least 5 minutes and until more than 2 minutes of stable breathing were recorded. Repeat sniff manoeuvres were then performed with verbal encouragement until a consistent EMG_{para} signal was recorded which was subsequently used as the maximum EMG_{para} measurement. EMG_{para} signals were analysed using the root mean squared (RMS) of the raw EMG_{para} signal with a 40 ms moving window and normalised to the maximum RMS EMG_{para} value (EMG_{para%max}) analogous to the algorithm previously described for the analysis of EMG_{di} (Jolley et al. (2009)). Whilst EMG_{para%max} reflects neural drive per breath and has been used in stable patients, here a neural respiratory drive index (NRDI, arbitrary units—AU) that incorporated respiratory rate to develop a measurement of neural drive to the respiratory muscles per unit time was used.

[0174] Results

[0175] EMG_{para} Reproducibility in Healthy Subjects

[0176] Ten healthy subjects had a mean age of 28 ± 5.0 years (range 20–36 years), 60% male. The mean EMG_{para} on visit 1 was 4.43 ± 2.09 μ V and on visit 2 was 4.15 ± 2.01 μ V with a co-efficient of variation (Cv) between visits of 0.10 ± 0.08 . The agreement between the first and second visit EMG_{para} data, assessed using a Bland and Altman plot, is shown in FIG. 10A. The mean EMG_{para} difference was -0.28 ± 0.51 μ V with limits of agreement of -1.28 and 0.73 μ V. The mean EMG_{para%max} value on visit 1 was $2.74 \pm 2.53\%$ and on visit 2 was $2.92 \pm 2.98\%$ with a Cv of 0.37 ± 0.27 . The bias in EMG_{para%max} measurements from visit 1 to visit 2 was $0.18 \pm 1.68\%$ within limits of agreement— 3.12 to 3.47% . Pearson correlation coefficient showed strong correlation with both EMG_{para} ($r=0.97$; $p<0.001$) and EMG_{para%max} ($r=0.83$; $p=0.003$).

[0177] EMG_{para} Reproducibility in Stable COPD Patients

[0178] Ten patients with stable COPD were studied with a mean age 75 ± 6.9 years (range 66–85 years), 50% male. Mean FEV₁ was 0.97 ± 0.41 L. The mean EMG_{para} on visit 1 was 8.83 ± 5.29 μ V and on visit 2 was 10.09 ± 6.72 μ V with a Cv of 0.19 ± 0.13 . The mean EMG_{para} difference was 1.26 ± 2.45 μ V with limits of agreement of -3.54 and 6.06 mV. The mean EMG_{para%max} value on visit 1 was $11.59 \pm 5.70\%$ and on visit 2 was $11.96 \pm 5.11\%$ with a Cv of 0.15 ± 0.09 . The mean EMG_{para%max} difference was $-0.37 \pm 2.93\%$ with limits of agreement of -6.12 and 5.38% . A Bland-Altman plot of the first and second visit EMG_{para%max} data is shown in FIG. 10B. Regression analysis between visit 1 and visit 2 demonstrated a strong correlation with both EMG_{para} ($r=0.94$; $p<0.001$) and EMG_{para%max} ($r=0.89$; $p=0.002$). No significant relationships could be identified between forced expiratory volume in 1 second (FEV₁) and either EMG_{para} ($p=0.78$) or EMG_{para%max} ($p=0.46$) or the dyspnoea domain of the CRDQ and EMG_{para} ($p=0.17$) or EMG_{para%max} ($p=0.37$).

[0179] Reproducibility of EMG_{para}

[0180] EMG_{para} data is reported as root mean squared (RMS; μ V). A representative trace of a patient with stable COPD during the sniff manoeuvres is shown in FIG. 11. Reproducibility of EMG_{para} was confirmed in 10 healthy volunteers and 10 patients with stable COPD.

Example 2

EMG_{para} as a Physiological Biomarker to Monitor Change in AECOPD (to Predict Deterioration and Re-Admission)

[0181] Methods

[0182] Subjects

[0183] Patients with AECOPD were recruited. AECOPD was defined based on clinical features and basic investigations. Initial patient management was according to standard local guidelines with oral corticosteroids, antibiotics and a combination metered-dose inhalers and nebulised bronchodilators. The first EMG_{para} measurement recorded within 24 hours of hospital arrival. Repeat EMG_{para} measurements and the clinical dataset were recorded daily until the patient was reported as stable and suitable for hospital discharge.

[0184] Baseline Data

[0185] Demographic and anthropometric data were collected. Borg (Borg (1982) *Med. Sci. Sports Exerc.* 14, 377–81) and MRC dyspnoea score (Mahler et al. (1987) *Am. Rev. Respir. Dis.* 135, 1229–33; Celli et al. (2004) *N. Engl. J. Med.* 350, 1005–12) were used to assess subjective breathlessness. HRQL data was obtained using the Chronic Respiratory Disease Questionnaire (CRDQ). (Guyatt et al. (1987) *Thorax* 42, 773–8). Spirometry was performed with a handheld device (EasyOne Diagnostic Spirometer, ndd Medical Technologies, Switzerland) according to standard guidelines (Quanjer et al. (1993) *Eur. Respir. J. Suppl.* 16, 5–40; Statement of the American Thoracic Society (1987) *Am. Rev. Respir. Dis.* 136; 1285–98). Repeat measurements were taken during admission. The patient was seated and rested for at least 5 minutes; bronchodilator therapy was withheld for the previous 4 hours. Heart rate (HR), oxygen saturations (S_pO_2) and respiratory rate (RR) were measured over one minute. Clinical data (HR, S_pO_2 , RR, temperature (T), blood pressure (BP) and medical early warning score (MEWS) (Subbe et al. (2001) *QJM.* 94, 521–26) and the supervising senior clinician's summary opinion on clinical course were recorded from the medical notes and observation charts. A patient was defined as a clinical 'deteriorator' or 'improver' based on the summary opinion of the senior attending respiratory physician (respiratory specialist registrar or consultant) and the requirement for increased treatment. The respiratory clinicians were blinded to the EMG_{para} measurement, which was analysed off line following patient discharge. EMG_{para} signals were acquired either in a chair or semi-recumbent in bed. Oxygen therapy was only used when the S_pO_2 was 88%.

[0186] EMG_{para} Measurement

[0187] The second intercostal space was identified using surface bony landmarks and the skin was prepared prior to placement of electrodes. EMG_{para} signal acquisition and processing was analogous to the method described previously for EMG_{di} as described above for with respect to Example 1. The resting signal was normalised to the maximum value obtained from a reproducible maximum sniff manoeuvre to produce the EMG_{para%max}. To reflect changes in respiratory pattern, the product of EMG_{para%max} and respiratory rate was calculated to produce the neural respiratory drive index (NRDI; arbitrary units AU). Nasal cannulae connected to a differential pressure transducer (Validyne DP45, Validyne, Northridge, Calif., US) identified inspiratory and expiratory phases of breathing.

Data Analysis and Statistics

[0188] Reproducibility was assessed using co-efficient of variability and Bland-Altman analysis (Bland & Altman (1986) *Lancet* 327, 307–10). Relationships between EMG_{para}, EMG_{para%max} and NRDI and lung function parameters

and HRQL data were analysed using regression analysis. Data were analysed using independent or paired t-test where appropriate. Data that were not normally distributed, as defined by the Kolmogorov-Smirnov test, were transformed and then analysed as parametric data or if the logarithm of the data remained non-normal then a non-parametric equivalent was used. Data analysis was conducted using SPSS software (SPSS, Chicago, Ill., USA). All data are presented as mean $\pm$ SD, unless otherwise stated with a p value <0.05 considered as statistically significant.

[0189] Results

[0190] Change in EMG_{para} in Patients with AECOPD

[0191] 30 patients were recruited with a mean age of 72 $\pm$ 10 years (47% male). On admission, the median MRC dyspnoea score was 5 (2-5). The median previous admission frequency and length of stay was 3 admissions (0-13) and 6 days (2-34), respectively. Baseline data is provided in Table 1 and full details can be found in Table 2.

TABLE 1

Standard Clinical Parameters and Indices of Neural Respiratory Drive on Admission		
	Emergency Department	Baseline Measurements
MEWS	3 (0-7)	2 (0-4)
FEV ₁ (L)	—	0.60 $\pm$ 1.65

TABLE 1-continued

	Emergency Department	Baseline Measurements
FVC (L)	—	1.53 $\pm$ 0.82
P _a O ₂ (kPa)	10.0 $\pm$ 3.5	—
P _a CO ₂ (kPa)	6.3 $\pm$ 1.4	—
EMG _{para%max} (%)	—	20.3 $\pm$ 9.9
NRDI (AU)	—	455 $\pm$ 263

Data presented as median (range) and mean $\pm$ standard deviation; MEWS = medical early warning score; FEV₁ = forced expiratory volume in 1 second; P_aO₂ = arterial partial pressure of oxygen; P_aCO₂ = arterial partial pressure of carbon dioxide;

[0192] Changes in EMG_{sc} in Patients with AECOPD

[0193] Baseline EMG_{sc%max} was 13.4 $\pm$ 8.5% with no statistically significant changes occurring during the first 24 hours of admission in either improvers or deteriorators. There were no significant relationships between EMG_{sc%max} and other markers of NRD, measures of dyspnoea, spirometric measures or in the standard clinical variables.

[0194] Changes in NRD During Hospital Admission

[0195] Both 'improvers' and 'deteriorators' had similar levels of NRD at initial reading that were significantly different at the follow up reading 24 hours later (mean difference EMG_{para%max}=8.1, 95% CI 0.2-16.0, p=0.046; mean difference NRDI=335, 95% CI 163-507, p<0.001). Differences in NRD did not persist in subsequent measurements (FIG. 12).

TABLE 2

Baseline data of patients with AECOPD									
	Age (years)	Sex (M/F)	EMG _{para} (μ V)	EMG _{para%max} (%)	NRDI (AU)	MEWS	Borg	RR (bpm)	FEV ₁ (L)
1	64	F	12.6	9.1	238	3	4	26	UTP
2	73	F	8.4	12.9	271	2	1	21	0.32
3	57	M	40.0	25.0	799	3	4	32	UTP
4	70	F	43.1	52.1	1512	4	7	29	0.52
5	72	M	7.5	11.0	221	1	3	20	1.64
6	72	F	5.8	21.0	543	3	6	26	UTP
7	77	M	3.7	11.5	298	2	7	26	0.49
8	81	F	14.0	21.2	509	2	6	24	0.46
9	68	F	22.4	20.3	406	1	4	20	UTP
10	64	M	16.8	25.3	506	1	7	20	0.69
11	80	F	6.5	13.9	291	4	3	21	0.58
12	74	F	9.2	14.8	355	2	2	24	0.45
13	69	F	18.6	16.0	383	2	8	24	0.49
14	89	F	21.8	36.7	808	2	1	22	0.24
15	90	M	8.3	14.0	252	1	0.5	18	0.69
16	79	M	6.4	16.9	422	3	5	25	0.48
17	85	F	7.5	16.2	390	3	7	24	0.53
18	72	F	16.4	10.4	249	2	4	24	UTP
19	72	F	9.5	12.2	231	1	5	19	0.62
20	75	F	19.8	15.7	378	4	8	24	0.44
21	63	F	18.4	28.1	534	1	8	19	0.6
22	72	M	20.9	32.3	550	1	3	17	0.67
23	75	F	16.3	36.7	770	3	4	21	0.34
24	83	M	15.7	17.9	287	0	0	16	1.58
25	43	M	13.3	21.0	440	2	3	21	0.5
26	64	M	17.4	25.1	527	2	9	21	0.45
27	62	M	16.7	29.7	594	2	3	20	0.58
28	85	M	6.2	13.2	304	2	3	23	0.82
29	63	M	3.0	8.0	144	1	5	18	2.11
30	80	M	10.5	21.3	426	1	9	20	0.96
Mean $\pm$ SD	72 $\pm$ 10		14.6 $\pm$ 9.3	20.32 $\pm$ 9.85	455 $\pm$ 263	2 (0-4)*	4 (0-9)*	22 $\pm$ 4	0.60 $\pm$ 1.65

*median (range)

Abbreviations:

UTP = patient unable to perform;

MEWS = medical early warning score;

FEV₁ = forced expiratory volume in 1 second

[0196] Three patients received non-invasive ventilation with all cases initiated in the first 4 hours of admission in the emergency department. Nine patients were discharged with home oxygen, all were previously prescribed long term oxygen therapy.

[0197] Twenty-four patients had recordings on two occasions, five patients had recordings on three occasions and one patient had recordings on four occasions, producing 37 data pairs. Δ Borg score had a significant relationship with Δ EMG_{para} ($r=+0.50$; $p=0.001$), Δ EMG_{para%max} ($r=+0.57$; $p<0.001$) and Δ NRDI ($r=+0.60$; $p<0.001$) as shown in FIG. 13 and Δ FEV₁ ($r=-0.58$; $p=0.002$). There was no relationship observed with Δ S_pO₂ ($p=0.16$) or Δ RR ($p=0.08$).

[0198] There were significant differences observed in mean change between 'improvers' and 'deteriorators' in all three EMG_{para} indices. However, there were no significant between group differences in changes in RR, HR, S_pO₂ or FEV₁ (Table 3).

TABLE 3

Difference between consecutive recordings of measured physiological variables in 30 patients from day of baseline measurement to repeat reading							
	Δ MEWS*	Δ RR	Δ S _p O ₂	Δ FEV ₁ [†]	Δ EMG _{para}	Δ EMG _{para%max}	Δ NRDI
'Deteriorators'	0.50 (0-1)	4.5 $\pm$ 6.0	1.2 $\pm$ 2.4	0.03 $\pm$ 0.18	7.8 $\pm$ 4.9	6.2 $\pm$ 4.3	226 $\pm$ 58
'Improvers'	0 (-2-1)	-1.8 $\pm$ 3.8	0.9 $\pm$ 2.7	0.06 $\pm$ 0.14	-1.7 $\pm$ 5.5	-3.5 $\pm$ 8.1	-113 $\pm$ 221
Mean difference		6.3	-0.6	0.03	9.6	9.6	339
(95% CI)		(-0.1-12.6)	(-2.0-3.1)	(-0.42-0.36)	(4.4-14.8)	(4.5-14.8)	(234-444)
P value	0.02	0.05	0.6	0.7	0.003	0.001	<0.001

Data presented as mean $\pm$ standard deviation or *median (range);

[†]7 patients were unable to perform spirometry on 1 or more occasions, therefore analysis of FEV₁ was performed on 'improvers' n = 20 and 'deteriorators' n = 3.

MEWS = medical early warning score;

FEV₁ = forced expiratory volume in 1 second;

95% CI = 95% confidence intervals

[0199] A significant ($p=0.02$), but clinically small ($+0.5$), difference was observed in MEWS between 'improvers' and 'deteriorators'. Patients who improved had statistically significant reduction in dyspnoea (Δ Borg -1.5; 95% CI -0.7--2.3), respiratory rate (Δ RR -1.8 bpm; 95% CI -0.2--3.3) and increase in FVC (Δ FVC 0.22 L; 95% CI -0.05-0.40), with no statistically significant differences demonstrable in physiological variables in the 'deteriorators'.

[0200] Receiver operating characteristics (ROC) plots (FIG. 14A) with change 'cut offs' $>+6.6$ for EMG_{para%max} and $>+160$ AU for NRDI had sensitivities of 83% (95% CI 54-100%) and 100% (95% CI 100-100%) and specificities of 96% (95% CI 88-100%) and 100% (95% CI 100-100%) for both EMG_{para%max} and NRDI, respectively. ROC plots of the

standard clinical variables either did not statistically differ from the null hypothesis or could not produce a 'cut off' providing sensitivity $>80\%$ without reducing specificity to $<90\%$ (FIG. 14B).

[0201] Change in EMG Para Between Admission and Discharge to Predict Readmission

[0202] A significant difference in Δ EMG_{para} and Δ NRDI between admission (1st measurement within 24 hours of admission) and discharge (final measurement within 24 hours of clinical stability) was demonstrated between those patients readmitted within 14 days as a consequence of a respiratory deterioration and those patients who remained at home. However, Δ MEWS, Δ FEV₁ and number of previous admissions did not differ between patients who were and were not readmitted (Table 4).

TABLE 4

Difference between admission and discharge of measured physiological variables in 30 patients either readmitted (n = 9) or not readmitted (n = 21) within 14 days of hospital discharge					
	Δ MEWS*	Δ FEV ₁ [†]	Previous admissions*	Δ EMG _{para%max}	Δ NRDI
Readmitted	0 (-1-2)	0.09 $\pm$ 0.15	4 (0-14)	1.98 $\pm$ 4.36	76 $\pm$ 134
Not readmitted	0 (-3-2)	0.08 $\pm$ 0.10	3 (0-10)	-4.05 $\pm$ 10.30	-127 $\pm$ 305

TABLE 4-continued

Difference between admission and discharge of measured physiological variables in 30 patients either readmitted (n = 9) or not readmitted (n = 21) within 14 days of hospital discharge					
	ΔMEWS*	ΔFEV ₁ †	Previous admissions*	ΔEMG _{para%max}	ΔNRDI
Mean difference (95% CI)		0.1 (0.14-0.11)		6.03 (11.5-0.54)	203 (39-366)
P value	0.5	0.8	0.1	0.03	0.02

Data presented as mean $\pm$ standard deviation or

*median (range);

†7 patients were unable to perform spirometry on 2 occasions, therefore analysis of FEV₁ was performed on readmitted n = 6, not readmitted n = 17.

MEWS = medical early warning score; FEV₁ = forced expiratory volume in 1 second; 95% CI = 95% confidence intervals

[0203] ROC plots (FIG. 15A) were calculated with 'cut offs' of a change in EMG_{para%max} >0% and NRDI >50AU during admission producing sensitivities of 67% (95% CI 36-97%) and 67% (95% CI 36-97%) and specificities of 62% (95% CI 41-83%) and 71% (95% CI 52-91%) for EMG_{para%max} and NRDI, respectively. None of the ROC plots for routine clinical variables differed significantly from the null hypothesis (FIG. 15B).

[0204] Discussion

[0205] Examples 1 and 2 demonstrate that 2nd intercostal space (ICS) parasternal NRDI, calculated as a product of EMG_{para} and RR normalised for maximum EMG_{para}, is a reproducible physiological biomarker in stable COPD patients and has greater sensitivity and specificity than standard clinical physiological parameters to identify AECOPD patients failing to respond to treatment. Furthermore, the failure of NRDI to fall during AECOPD requiring hospitalisation identifies patients who are more likely to be readmitted with a further respiratory deterioration.

[0206] Patient Selection

[0207] The AECOPD patients recruited were not consecutive admissions and therefore subject to selection bias. Despite this limitation, demographics and severity of patients were similar to previously reported data (Roberts et al. (2011) *Thorax* 66, 43-8). Furthermore, the goal of this study was to provide pilot data to demonstrate the feasibility and clinical usefulness of using non-invasive EMG monitoring as a physiological biomarker in the acute setting.

[0208] Surface EMG Para Measurement

[0209] Although the issues of surface EMG recording are well described (Luo et al. (2008) *Clin. Sci. (Lond)*. 115, 233-44), contamination from other chest wall muscles cannot be excluded. Patient and electrode position during data acquisition were carefully observed to maximise the contribution of 2nd ICS parasternal muscle to the inspiratory EMG_{para} signal and minimising the non-respiratory muscle activity of other muscles. Needle electrode technique could be used to isolate parasternal muscle activity, but similar to oesophageal measurement of diaphragm electrical activity, this invasive technique is not suitable for the acute setting.

[0210] Validity and Reproducibility of EMG_{para}

[0211] EMG_{para} as a measure of NRD, was shown to have satisfactory inter-occasion reproducibility in both healthy subjects and patients with stable COPD. Although the degree of variability with EMG_{para} in COPD patients was greater in this study than that shown previously using EMG_{di} (Jolley et al. (2009)) and EMG_{para} in patients with cystic fibrosis (Reilly et al. (2011) *Thorax* 66, 240-46), the inter-occasion

correlation for both healthy subjects and stable COPD was >0.80, which is a level that has previously been used to indicate acceptable inter-test agreement for surface electromyogram (Duiveman et al. (2004)).

[0212] Definition of Clinical Change

[0213] All patients met an event based criteria for a severe exacerbation of COPD requiring hospital admission (Anthonisen et al. (1987) *Ann. Intern. Med.* 106, 196-204; Trappenburg et al. (2011) *Eur. Respir. J.* 37, 1260-8). There is no 'gold standard' to predict or measure acute clinical progress in AECOPD (Trappenburg et al. (2011)), and therefore EMG_{para} was compared to standard clinical parameters and the summary opinion of the supervising senior physician. Whilst this is a broad definition it is widely used in both research and clinical practice and is the benchmark by which other novel influential assessment tools have been judged (Leidy et al. (2011) *Am. J. Respir. Crit. Care Med.* 183, 323-9). Despite the limitations inherent with this choice of outcome it allows the data to be easily interpreted. Clinical gestalt is the interpretation and analysis by the physician of the patient's report of their clinical state as well as the findings from the physical examination incorporating standard physiological variables and clinical parameters, which are subsequently processed as part of learnt complex clinical algorithm to determine the clinical state of the patient and the response to treatment. In order to have measured the performance of this novel technique against a more definable objective marker the study population may have to be limited to those patients in respiratory failure. However, this would have limited the applicability of the study and would have been difficult to demonstrate that measures of NRD added to already established and widely available techniques to measure clinical progress in this group.

[0214] Significance of Findings

[0215] Parasternal Muscle Activity

[0216] Chest wall respiratory muscles have increased importance in patients with advanced COPD as progressive hyperinflation impacts adversely on diaphragm positioning and efficiency (Sharp et al. (1977) *Am. Rev. Respir. Dis.* 115, 47-56; Kyroussis et al. (2000) *Eur. Respir. J.* 15, 649-55), which results in a compensatory increase in chest wall and accessory respiratory muscle activity (Man et al. (2004) *Thorax* 59, 471-76). In particular, the uppermost parasternal intercostal muscles have been shown to be important inspiratory muscles (Legrand et al. (1996) *J. Appl. Physiol.* 80, 2097-101; De Troyer et al. (1996) *J. Appl. Physiol.* 80, 1490-4; De Troyer & Leduc (2006) *J. Appl. Physiol.* 101, 169-75). Furthermore, these parasternals have minimal post-inspira-

tory activity (Easton et al. (1999) *J. Appl. Physiol.* 87, 1097-101) with the 2nd ICS parasternal muscle demonstrating similar activity to the diaphragm (Gandevia et al. (2006) *J. Physiol.* 573; 263-75). However, during increasing hyperinflation, as observed during an AECOPD, the resting length of the parasternals is less affected than the diaphragm, such that the parasternals make a greater contribution to inspiratory pressure generation (Martinez et al. (1991) *Am. Rev. Respir. Dis.* 143; 476-480). This increase in parasternal activity is also associated with higher levels of dyspnoea (Ward et al. (1988) *J. Appl. Physiol.* 65, 2181-9). These data provide the scientific rationale to develop EMG_{para} as a physiological biomarker to track changes in clinical state in patients with AECOPD. Patients failing to respond to therapy have persistent hyperinflation with sustained elevation in EMG_{para} activity compared to those responding to therapy and associated respiratory muscle unloading which have a decline in EMG_{para} activity as the lung volumes and diaphragm and parasternal activity return to baseline. Finally, NRD was expressed as a product of EMG_{para} and RR normalised for maximum EMG_{para} to produce NRDI, which incorporates the peak RMS inspiratory parasternal muscle activity per unit time as a ratio of maximum NRD.

[0217] Dyspnoea

[0218] Dyspnoea provides a significant symptom burden in COPD. An objective method of assessing the severity of breathlessness has previously been lacking, with clinicians using subjective assessment tools. Physiological indicators of disease severity in COPD, such as FEV_1 , are acknowledged to be poorly predictive of dyspnoea (Mahler (1992) *Chest* 101, 242S-7S). In contrast, changes in NRD have been shown to explain variance in exercise induced dyspnea (Marin et al. (1999) *Chest* 115, 1293-300) The applicant observed a similar relationship between change in Borg score and change in EMG_{para} . Furthermore, as the initial measurements were recorded following commencement of emergency therapy, in some patients there were relatively small changes in 2nd ICS parasternal muscle electrical activity and breathlessness, indicating that that this technique is sensitive enough to monitor relatively modest changes even after treatment initiation. These data, therefore, support the use of non-invasive 2nd ICS parasternal electromyography as a physiological biomarker of NRD that reflects perception of dyspnoea severity during AECOPD. Furthermore, as FEV_1 has a weak relationship with dyspnoea, there is potential for EMG_{para} to be applied to patients in the stable state to monitor progression of disease and detect exacerbation onset, although more work is required to fully elucidate this relationship.

[0219] Monitoring Response to Treatment

[0220] NRD was shown to monitor response to treatment in patients admitted with AECOPD when calculated as $EMG_{para\%max}$ and NRDI. The reproducibility data indicate that the 'cut off' chosen for maximum sensitivity and specificity of detection for clinical change ($EMG_{para\%max} > 6.6\%$) represents a genuine and detectable change in NRD as it is above the 95% upper limit of agreement on the Bland-Altman plot. In this population of AECOPD patients, this would have correctly tracked deterioration in 5 out of 6 occasions. This high sensitivity and specificity was further improved with the addition of respiratory rate to produce the NRDI, which correctly identified all episodes of deterioration in this sample set. This demonstrates the potential clinical utility of the test with the integrated physiological signal accurately reflecting

the summary opinion of the senior attending respiratory physician in a way unable to be replicated by any of the standard clinical variables assessed.

[0221] Re-Admission

[0222] In addition to the ability to track changes in NRD during AECOPD using this technique, the applicant has also shown that failure of NRDI to fall in response to therapy identifies those patients who are likely to be readmitted within 14 days of discharge with a further respiratory deterioration. There has not previously been reported any clinically useful biomarker that can predict readmission in such patients (Cao et al. (2006) *Respirology* 11, 188-95). Previous data in COPD patients with severe disease, as indicated by an $FEV_1 < 1$ L at discharge or > 2 previous admissions in the preceding 12 months, reported that these patients were more likely to be readmitted following an exacerbation of COPD (Garcia-Aymerich et al. (2003) *Thorax* 58, 100-105). The specificity of these particular predictors in the current cohort of patients was poor at < 0.5 and therefore these are not clinically useful. Failure of NRDI to fall in response to treatment provides an easy to apply novel physiological biomarker to predict readmission in high risk patients. Data from the ECLIPSE study has suggested the 'frequent exacerbator' is a distinct phenotype in COPD (Hurst et al. (2010) *N. Engl. J. Med.* 363, 1128-38). The measurement of NRDI in this context is not simply acting as a measure of disease severity or to identify the frequent exacerbator phenotype, as if the analysis were limited to those patients with ≥ 2 previous admissions ($n=22$), the sensitivity and specificity to predict readmission at 14 days remained similar to the whole cohort at 63% and 64%, respectively. The ability of this physiological tool to maintain its sensitivity and specificity in the higher risk group of patients increases the clinical utility with the ability to further risk stratify the most high risk patients. With the increasing role of early discharge and COPD outreach schemes to support patients in the community, this technique could facilitate clinical selection to identify those patients that require greater community support or further hospital treatment prior to discharge. This approach has increasing importance as the rising incidence of failed hospital discharge have been highlighted by the UK government as an area for improved performance with potential financial penalties for hospitals.

[0223] The Examples set out below illustrate various situations, measurements taken and expected outcomes.

Example 3

[0224] 4.1 Patient 1

[0225] Scenario: Neural respiratory drive falls between hospital admission and discharge

[0226] Outcome: The patient did not require re-admission to hospital within 28 days.

TABLE 5

Trace	RR	Mean(peak- $EMG_{para\ 30}$)	$EMG_{para\%max}$	NRDI	$maxEMG_{para}$
Admission	21	31.41	23.10	485.13	135.95
Discharge	20	9.01	6.98	139.67	129.01

[0227] Data are depicted in FIG. 16.

Example 4

[0228] Scenario: Neural respiratory drive rises between hospital admission and discharge

[0229] Outcome: The patient required re-admission to hospital within 28 days

TABLE 6

Trace	RR	Mean(peak- EMG _{para} 30)	EMG _{para} %max	NRDI	maxEMG _{para}
Admission	23	21.00	28.01	644.24	74.98
Discharge	28	25.10	38.94	1090.37	64.45

[0230] Data are depicted in FIG. 17.

Example 5

[0231] Scenario: Neural respiratory drive rises significantly during hospital admission as clinical respiratory status declines due to treatment failure, i.e., monitoring acute deterioration

[0232] Outcome: Escalation in medical treatment above standard initial therapy

TABLE 7

Trace	RR	Mean(peak- EMG _{para} 30)	EMG _{para} %max	NRDI	maxEMG _{para}
1	16	11.81	16.24	259.90	72.72
2	22	12.83	21.76	478.76	58.95
3	25	14.28	16.72	418.12	85.38
4	18	9.02	8.51	153.23	105.98

[0233] Data are depicted in FIG. 18.

[0234] The skilled person will appreciate that the above description and Examples are exemplary only and modifications may be made thereto.

What is claimed is:

1. A method of monitoring a patient, including: measuring neural respiratory drive using a monitoring device; repeating the measurement continuously and/or at regular time intervals; and comparing the measurements obtained in order to predict treatment failure and/or clinical deterioration and/or admission from home or re-admission to hospital.
2. A method as claimed in claim 1, wherein the neural respiratory drive is measured by obtaining a measure of the second intercostal space parasternal electromyogram.
3. A method as claimed in claim 1, including obtaining a value for the neural respiratory drive time index by multiplying the value of the neural respiratory drive time product by the respiratory rate.
4. A method as claimed in claim 1, including:
 - 1) obtaining a value for a normalized neural respiratory drive (which may be calculated as $EMG_{para}/EMG_{para\%max}$ and termed $EMG_{para\%max}$);
 - 2) obtaining the respiratory rate of the patient, expressed as breaths per minute (RR/min);
 - 3) obtaining a value for the neural respiratory drive index by multiplying the value of the neural respiratory drive obtained in Step 1 by the respiratory rate ($EMG_{para\%max} \cdot RR$);

4) repeating Steps 1 to 3 continuously and/or after a first given period of time and comparing the two neural respiratory drive index values obtained.

5. A method as claimed in claim 4, wherein Steps 1 to 3 are carried out upon admission into hospital and repeated just prior to discharge from hospital.

6. A method as claimed in claim 1, including:

1) obtaining a value for a normalized neural respiratory drive by:

a) carrying out parasternal electromyography to obtain a raw signal during normal breathing;

identifying and obtaining the root mean square of the raw signal to obtain a rectified trace;

identifying and obtaining the peak magnitude of the rectified trace for each inspiration;

calculating the mean of the peak magnitudes identified ($EMG_{para\%peak}$);

b) carrying out parasternal electromyography to obtain a raw signal during at least two (or preferably at least three) sniff maximal maneuvers;

obtaining the root mean square of the raw signal to obtain a rectified trace;

identifying and selecting the peak magnitude of the rectified trace ($EMG_{paramax\%peak}$);

c) expressing the mean of the peak calculated in Step 1a as a percentage of the peak magnitude selected in Step 1b ($EMG_{para\%max\%peak}$);

2) obtaining the respiratory rate of the patient, expressed as breaths per minute;

3) obtaining a value for the neural respiratory drive index by multiplying the value of the $EMG_{para\%max\%peak}$ (NRD) obtained in Step 1c by the respiratory rate;

4) repeating Steps 1 to 3 continuously and/or after a first given period of time and comparing the two neural respiratory drive index values obtained.

7. A method as claimed in claim 1, including:

obtaining a value for the neural respiratory drive time product by carrying out parasternal electromyography to obtain a raw signal during normal breathing;

obtaining the root mean square of the raw signal to obtain a rectified trace;

measuring the area under the rectified trace;

obtaining a value for the neural respiratory drive time product index by multiplying the neural respiratory drive time product by the respiratory rate; and

comparing the two neural respiratory drive time index values obtained.

8. A method as claimed in claim 7, including:

identifying and obtaining the area under the curve of the rectified root mean square of the raw signal;

calculating the mean of the area under the curves identified ($EMG_{paraAUC}$);

identifying and selecting the area under the curve of the rectified trace ($EMG_{paramaxAUC}$);

expressing the mean area under the curve as a percentage of the maximum area under the curve ($EMG_{para\%maxAUC}$); and

obtaining neural respiratory drive time index by multiplying the value of the $EMG_{para\%maxAUC}$ (NRDTP) by the respiratory rate.

9. A method as claimed in claim 1, wherein the patient has respiratory disease, for example, wherein the respiratory disease is:

- (1) an acute exacerbation of chronic obstructive pulmonary disease;
- (2) an acute exacerbation of chronic respiratory disease;
- (3) acute respiratory failure;
- (4) chronic respiratory disease;
- (5) chronic respiratory failure;
- (6) acute exacerbation of chronic heart failure;
- (7) acute heart failure; or
- (8) chronic heart failure.

10. A method as claimed in claim 6, wherein the peak magnitude is obtained for each inspiration over a time period of approximately 30 seconds to 3 minutes.

11. A monitoring device including: a signal input, and a processing unit, the monitoring device being arranged to:

- (1) receive a first raw parasternal electromyography signal at the signal input;
execute computer program code in the processing unit to:
- (2) determine the root mean square of the raw parasternal signal to obtain a rectified trace;
- (3) identify the peak magnitude of the rectified parasternal trace for each inspiration over a second given period of time;
- (4) calculate the mean of the peak magnitudes identified;
- (5) receive further and determine raw parasternal electromyography signals during at least two (preferably at least three) sniff manoeuvres;
- (6) determine the root mean square of the raw signal to obtain a further rectified trace;
- (7) determine the peak magnitude of the further rectified trace;
- (8) express the mean of the first peak magnitude as a percentage of the peak magnitude of the further rectified trace obtained during the sniff manoeuvre;
- (9) receive data on the respiratory rate of the patient, expressed as breaths per minute;
- (10) determine a value for the neural respiratory drive index by multiplying the value of the neural respiratory drive by the respiratory rate; and
- (11) store the measured/determined values of the neural respiratory drive a data repository.

12. A monitoring device as claimed in claim 11, being arranged to determine a subsequent neural respiratory drive index and compare with the stored value.

13. A monitoring device as claimed in claim 11, operable to identify the area under the curve of the rectified root mean square of the raw parasternal signal for each inspiration over a second given period of time to display the measured/determined neural respiratory drive values in real time; to calculate the mean of the area under the curves identified; and to determine the values for NRDTp and NRDTI from the area under the curve of the rectified trace.

14. A method as claimed in claim 1, including carrying out substantially real time processing of biological signals received from surface electrodes, and converting them into a

physiological biomarker by processing the signals from the electrodes to obtain and display heart rate (HR), respiratory rate (RR), neural respiratory drive (NRD), neural respiratory drive index (NRDI).

15. A method as claimed in claim 14, including obtaining and displaying neural respiratory drive time product (NRDTP) and neural respiratory drive time index (NRDTI).

16. A method as claimed in claim 14, wherein neural respiratory drive (NRD) is determined from the electromyogram of the 2nd intercostal parasternal muscles (EMG_{para}) by using electrodes and amplifiers, and processing the para, signals using analog to digital conversion followed by digital filtering and arithmetic conversion of the signal.

17. A method as claimed in claim 1, wherein signals are recorded during resting breathing to provide an indication of a patient's current respiratory effort determined by objectively measuring EMG_{para} activity, and wherein the patient performs repeated maximum sniff manoeuvres in order to allow the signal to be normalised for an individual patient maximum effort ($EMG_{para\%max}$).

18. A method as claimed in claim 1, including the following steps:

- 1) identifying patients at potential risk of deterioration;
- 2) placing surface electrodes over the parasternal muscles of the second intercostal space along with a reference electrode over the electrically neutral clavicle;
- 3) amplifying electrical signals from the two recording electrodes;
- 4) passing the amplified signal to an analog to digital converter to allow for further computer processing; and
- 5) integrating the signals to produce the essential biological variables for clinical interpretation.

19. A method as claimed in claim 14, wherein the signal is amplified to a factor of 1000 and analog filtered at 10 Hz and 2000 Hz to remove contributions from other muscle activity and maximize signal from the parasternal muscles.

20. A method as claimed in claim 1, including at least one of the following features:

- Signal quality assessment (including initial signal)
- ECG (QRS) detection and HR calculation
- ECG artifact accommodation or removal from EMG_{para} signal
- Calculation of RMS
- EMG_{para} analysis including peak-peak analysis to calculate respiratory rate, peak value to calculate $EMG_{para\%maxpeak}$ and multiplication by respiratory rate to calculate NRDI
- EMG_{para} analysis to calculate area under the curve of the signal to calculate $EMG_{para\%maxAUC}$ by normalizing for the $EMG_{paramaxAUC}$ and multiplying by respiratory rate to calculate NRDTI.

* * * * *

专利名称(译)	病人监护方法和监测装置		
公开(公告)号	US20130310699A1	公开(公告)日	2013-11-21
申请号	US13/627383	申请日	2012-09-26
[标]申请(专利权)人(译)	伦敦国王学院 GUYS & 圣托马斯NHS中发现的帐房里GUYS医院的信赖		
申请(专利权)人(译)	伦敦大学国王学院，钢绞线 盖伊和圣托马斯的NHS信托基金会，计数室，盖伊的医院		
当前申请(专利权)人(译)	伦敦大学国王学院，钢绞线 盖伊和圣托马斯的NHS信托基金会，计数室，盖伊的医院		
[标]发明人	HART NICHOLAS MOXHAM JOHN FEDELE FIAMMETTA		
发明人	HART, NICHOLAS MOXHAM, JOHN FEDELE, FIAMMETTA		
IPC分类号	A61B5/0205 A61B5/00 A61B5/0472 A61B5/0488 A61B5/04		
CPC分类号	A61B5/0205 A61B5/0488 A61B5/04012 A61B5/0472 A61B5/7203 A61B5/0245 A61B5/04017 A61B5/0452 A61B5/08		
优先权	2011016860 2011-09-30 GB 61/541708 2011-09-30 US		
外部链接	Espacenet USPTO		

摘要(译)

监测患者的方法包括使用监测装置 (10) 测量神经呼吸驱动，连续地或以规则的时间间隔重复测量，并比较所获得的测量值以预测治疗失败和/或临床恶化和/或重新测量入场。在本发明的实施例中，通过获得第二肋间隙胸骨旁肌电图的测量来测量神经呼吸驱动。监测装置 (10) 包括信号输入 (20)，处理单元 (30) 和输出单元 (50)，并且布置成测量神经呼吸驱动，存储测量值并将其与先前测量的值进行比较神经呼吸驱动的价值。

