



US010321877B2

(12) **United States Patent**
Volpe et al.

(10) **Patent No.:** **US 10,321,877 B2**
(45) **Date of Patent:** **Jun. 18, 2019**

(54) **MEDICAL DEVICE WITH ACOUSTIC SENSOR**

(71) Applicant: **ZOLL Medical Corporation**,
Chelmsford, MA (US)

(72) Inventors: **Shane S. Volpe**, Saltsburg, PA (US);
Gregory R. Frank, Mt. Lebanon, PA (US); **Thomas E. Kaib**, North
Huntingdon, PA (US); **Steven J. Szymkiewicz**, Bethel Park, PA (US);
Gary A. Freeman, Waltham, MA (US)

(73) Assignee: **ZOLL Medical Corporation**,
Chelmsford, MA (US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 293 days.

(21) Appl. No.: **15/074,020**

(22) Filed: **Mar. 18, 2016**

(65) **Prior Publication Data**

US 2016/0270738 A1 Sep. 22, 2016

Related U.S. Application Data

(60) Provisional application No. 62/134,881, filed on Mar.
18, 2015.

(51) **Int. Cl.**
A61B 7/04 (2006.01)
A61B 5/0205 (2006.01)
(Continued)

(52) **U.S. Cl.**
CPC **A61B 5/7278** (2013.01); **A61B 5/0408**
(2013.01); **A61B 5/08** (2013.01);
(Continued)

(58) **Field of Classification Search**
CPC A61B 5/11; A61B 5/1118; A61B 5/4815;
A61B 5/681; A61B 7/04; A61B 5/00;
(Continued)

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,094,310 A 6/1978 McEachern et al.
4,632,122 A 12/1986 Johansson et al.
(Continued)

FOREIGN PATENT DOCUMENTS

DE 2644236 C3 4/1981
EP 0295497 B1 9/1993
(Continued)

OTHER PUBLICATIONS

PCT Search Report and Written Opinion for PCT Application No.
PCT/US2016/023068, dated Jun. 3, 2016, 14 pages.

(Continued)

Primary Examiner — Carl H Layno


Assistant Examiner — Jon Eric C Morales

(74) *Attorney, Agent, or Firm* — Finch & Maloney PLLC

(57) **ABSTRACT**

In at least one example, a medical device is provided. The medical device includes at least one therapy electrode, at least one electrocardiogram (ECG) electrode, at least one acoustic sensor, and at least one processor coupled with the at least one acoustic sensor, the at least one ECG electrode, and the at least one therapy electrode. The at least one processor can receive at least one acoustic signal from the at least one acoustic sensor, receive at least one electrode signal from the ECG electrode, detect at least one unverified cardiopulmonary anomaly using the at least one electrode signal, and verify the at least one unverified cardiopulmonary anomaly with reference to data descriptive of the at least one acoustic signal.

25 Claims, 6 Drawing Sheets



- (51) **Int. Cl.**
A61B 5/00 (2006.01)
A61B 5/0408 (2006.01)
A61B 7/00 (2006.01)
A61N 1/04 (2006.01)
A61B 5/08 (2006.01)
A61N 1/372 (2006.01)
A61N 1/39 (2006.01)
A61N 1/365 (2006.01)
A61B 5/0464 (2006.01)
- (52) **U.S. Cl.**
 CPC *A61B 7/003* (2013.01); *A61B 7/04* (2013.01); *A61N 1/046* (2013.01); *A61N 1/0484* (2013.01); *A61N 1/3655* (2013.01); *A61N 1/36542* (2013.01); *A61N 1/37247* (2013.01); *A61N 1/37258* (2013.01); *A61N 1/37282* (2013.01); *A61N 1/39* (2013.01); *A61N 1/3993* (2013.01); *A61B 5/0464* (2013.01); *A61B 5/4818* (2013.01); *A61B 5/6805* (2013.01); *A61B 5/6823* (2013.01); *A61B 5/6844* (2013.01); *A61B 5/721* (2013.01); *A61B 2562/0219* (2013.01); *A61B 2562/17* (2017.08)
- (58) **Field of Classification Search**
 CPC ... *A61B 5/0205*; *A61B 5/0402*; *A61B 5/6833*; *A61B 5/04085*; *A61N 1/39*; *G10L 25/48*
 USPC 600/587, 301, 509, 528, 586; 607/6; 704/270
 See application file for complete search history.
- (56) **References Cited**
 U.S. PATENT DOCUMENTS
- | | | | |
|--------------|----|---------|---------------------|
| 4,928,690 | A | 5/1990 | Heilman et al. |
| 4,978,926 | A | 12/1990 | Zerod et al. |
| 4,991,217 | A | 2/1991 | Garrett et al. |
| 5,062,834 | A | 11/1991 | Gross et al. |
| 5,078,134 | A | 1/1992 | Heilman et al. |
| 5,348,008 | A | 9/1994 | Bornn et al. |
| 5,365,932 | A | 11/1994 | Greenhut |
| 5,381,798 | A | 1/1995 | Burrows |
| 5,472,453 | A | 12/1995 | Alt |
| 5,662,689 | A | 9/1997 | Elsberry et al. |
| 5,718,242 | A | 2/1998 | McClure et al. |
| 5,738,102 | A | 4/1998 | Lemelson |
| 5,741,306 | A | 4/1998 | Glegyak et al. |
| 5,758,443 | A | 6/1998 | Pedrazzini |
| 5,792,190 | A | 8/1998 | Olson et al. |
| 5,827,196 | A | 10/1998 | Yeo et al. |
| 5,887,978 | A | 3/1999 | Lunghofer et al. |
| 5,929,601 | A | 7/1999 | Kaib et al. |
| 5,944,669 | A | 8/1999 | Kaib et al. |
| 6,016,445 | A | 1/2000 | Baura |
| 6,045,503 | A | 4/2000 | Grabner et al. |
| 6,047,203 | A | 4/2000 | Sackner et al. |
| 6,065,154 | A | 5/2000 | Hulings et al. |
| 6,097,982 | A | 8/2000 | Glegyak et al. |
| 6,097,987 | A | 8/2000 | Milani |
| 6,148,233 | A | 11/2000 | Owen et al. |
| 6,169,397 | B1 | 1/2001 | Steinbach et al. |
| 6,253,099 | B1 | 6/2001 | Oskin et al. |
| 6,280,461 | B1 | 8/2001 | Glegyak et al. |
| 6,304,783 | B1 | 10/2001 | Lyster et al. |
| 6,390,996 | B1 | 5/2002 | Halperin et al. |
| 6,406,426 | B1 | 6/2002 | Reuss et al. |
| 6,681,003 | B2 | 1/2004 | Linder et al. |
| 6,690,969 | B2 | 2/2004 | Bystrom et al. |
| 6,804,554 | B2 | 10/2004 | Ujhelyi et al. |
| 6,827,695 | B2 | 12/2004 | Palazzolo et al. |
| 6,878,121 | B2 | 4/2005 | Krausman et al. |
| 6,908,437 | B2 | 6/2005 | Bardy |
| 6,947,565 | B2 | 9/2005 | Halleck et al. |
| 6,961,612 | B2 | 11/2005 | Elghazzawi et al. |
| 6,990,373 | B2 | 1/2006 | Jayne et al. |
| 7,149,579 | B1 | 12/2006 | Koh et al. |
| 7,220,235 | B2 | 5/2007 | Geheb et al. |
| 7,340,296 | B2 | 3/2008 | Stahmann et al. |
| 7,453,354 | B2 | 11/2008 | Reiter et al. |
| 7,488,293 | B2 | 2/2009 | Marcovecchio et al. |
| 7,712,373 | B2 | 5/2010 | Nagle et al. |
| 7,831,303 | B2 | 11/2010 | Rueter et al. |
| 7,974,689 | B2 | 7/2011 | Volpe et al. |
| 7,991,460 | B2 | 8/2011 | Fischell et al. |
| 8,121,683 | B2 | 2/2012 | Bucher et al. |
| 8,140,154 | B2 | 3/2012 | Donnelly et al. |
| 8,271,082 | B2 | 9/2012 | Donnelly et al. |
| 8,412,323 | B2 | 4/2013 | Bauer |
| 2003/0004547 | A1 | 1/2003 | Owen et al. |
| 2003/0060723 | A1 | 3/2003 | Joo et al. |
| 2003/0095648 | A1 | 5/2003 | Kaib et al. |
| 2003/0109904 | A1 | 6/2003 | Silver et al. |
| 2003/0149462 | A1 | 8/2003 | White et al. |
| 2003/0158593 | A1 | 8/2003 | Heilman et al. |
| 2003/0174049 | A1 | 9/2003 | Beigel et al. |
| 2003/0195567 | A1 | 10/2003 | Jayne et al. |
| 2003/0212311 | A1 | 11/2003 | Nova et al. |
| 2004/0049233 | A1 | 3/2004 | Edwards |
| 2004/0162510 | A1 | 8/2004 | Jayne et al. |
| 2004/0249419 | A1 | 12/2004 | Chapman et al. |
| 2005/0049515 | A1 | 3/2005 | Misczynski et al. |
| 2005/0131465 | A1 | 6/2005 | Freeman et al. |
| 2005/0246199 | A1 | 11/2005 | Futch |
| 2005/0283198 | A1 | 12/2005 | Haubrich et al. |
| 2006/0036292 | A1 | 2/2006 | Smith et al. |
| 2006/0085049 | A1 | 4/2006 | Cory et al. |
| 2006/0178706 | A1 | 8/2006 | Lisogurski et al. |
| 2006/0211934 | A1 | 9/2006 | Hassonjee et al. |
| 2006/0270952 | A1 | 11/2006 | Freeman et al. |
| 2006/0293714 | A1 | 12/2006 | Salo et al. |
| 2007/0118056 | A1 | 5/2007 | Wang et al. |
| 2007/0143864 | A1 | 6/2007 | Cabana et al. |
| 2007/0161913 | A1 | 7/2007 | Farrell et al. |
| 2007/0169364 | A1 | 7/2007 | Townsend et al. |
| 2007/0197878 | A1 | 8/2007 | Shklarski |
| 2007/0239214 | A1 | 10/2007 | Cinbis |
| 2007/0239220 | A1 | 10/2007 | Greenhut et al. |
| 2007/0265671 | A1 | 11/2007 | Roberts et al. |
| 2008/0004536 | A1 | 1/2008 | Baxi et al. |
| 2008/0021532 | A1 | 1/2008 | Kveen et al. |
| 2008/0033495 | A1 | 2/2008 | Kumar |
| 2008/0045815 | A1 | 2/2008 | Derchak et al. |
| 2008/0046015 | A1 | 2/2008 | Freeman et al. |
| 2008/0058884 | A1 | 3/2008 | Matos |
| 2008/0103402 | A1 | 5/2008 | Stickney et al. |
| 2008/0167535 | A1 | 7/2008 | Stivoric et al. |
| 2008/0177341 | A1 | 7/2008 | Bowers |
| 2008/0249591 | A1 | 10/2008 | Gaw et al. |
| 2008/0287749 | A1 | 11/2008 | Reuter |
| 2008/0294019 | A1 | 11/2008 | Tran |
| 2008/0306560 | A1 | 12/2008 | Macho et al. |
| 2008/0312520 | A1 | 12/2008 | Rowlandson et al. |
| 2008/0312522 | A1 | 12/2008 | Rowlandson et al. |
| 2008/0312709 | A1 | 12/2008 | Vople et al. |
| 2009/0018428 | A1 | 1/2009 | Dias et al. |
| 2009/0073991 | A1 | 3/2009 | Landrum et al. |
| 2009/0076336 | A1 | 3/2009 | Mazar et al. |
| 2009/0076340 | A1 | 3/2009 | Libbus et al. |
| 2009/0076341 | A1 | 3/2009 | James et al. |
| 2009/0076342 | A1 | 3/2009 | Amurthur et al. |
| 2009/0076344 | A1 | 3/2009 | Libbus et al. |
| 2009/0076345 | A1 | 3/2009 | Manicka et al. |
| 2009/0076346 | A1 | 3/2009 | James et al. |
| 2009/0076348 | A1 | 3/2009 | Manicka et al. |
| 2009/0076349 | A1 | 3/2009 | Libbus et al. |
| 2009/0076350 | A1 | 3/2009 | Bly et al. |
| 2009/0076363 | A1 | 3/2009 | Bly et al. |
| 2009/0076364 | A1 | 3/2009 | Libbus et al. |
| 2009/0076397 | A1 | 3/2009 | Libbus et al. |
| 2009/0076405 | A1 | 3/2009 | Amurthur et al. |
| 2009/0076410 | A1 | 3/2009 | Libbus et al. |

(56)

References Cited

U.S. PATENT DOCUMENTS

2009/0076559 A1 3/2009 Libbus et al.
 2009/0093687 A1 4/2009 Telfort et al.
 2009/0138059 A1 5/2009 Ouwerkerk
 2009/0146822 A1 6/2009 Soliman
 2009/0234410 A1 9/2009 Libbus et al.
 2009/0264792 A1 10/2009 Mazar
 2009/0275848 A1 11/2009 Brockway et al.
 2009/0281394 A1 11/2009 Russell et al.
 2009/0287120 A1 11/2009 Ferren et al.
 2009/0292194 A1 11/2009 Libbus et al.
 2009/0295326 A1 12/2009 Daynes et al.
 2009/0307266 A1 12/2009 Fleizach et al.
 2009/0318779 A1 12/2009 Tran
 2010/0052892 A1 3/2010 Allen et al.
 2010/0052897 A1 3/2010 Allen et al.
 2010/0056881 A1 3/2010 Libbus et al.
 2010/0069735 A1 3/2010 Berkner
 2010/0076513 A1 3/2010 Warren et al.
 2010/0081962 A1 4/2010 Hamaguchi et al.
 2010/0114243 A1 5/2010 Nowak et al.
 2010/0234716 A1 9/2010 Engel
 2010/0241181 A1 9/2010 Savage et al.
 2010/0295674 A1 11/2010 Hsieh et al.
 2010/0298899 A1 11/2010 Donnelly et al.
 2011/0015533 A1 1/2011 Cox et al.
 2011/0093840 A1 4/2011 Pynenburg et al.
 2011/0098765 A1 4/2011 Patel
 2011/0288604 A1 11/2011 Kaib et al.
 2011/0288605 A1 11/2011 Kaib et al.
 2012/0011382 A1 1/2012 Volpe et al.
 2012/0053479 A1 3/2012 Hopenfeld
 2012/0112903 A1 5/2012 Kaib et al.
 2012/0146797 A1 6/2012 Oskin et al.
 2012/0150008 A1 6/2012 Kaib et al.
 2012/0158075 A1 6/2012 Kaib et al.
 2013/0060149 A1 3/2013 Song et al.

2013/0325078 A1 12/2013 Whiting et al.
 2014/0277243 A1 9/2014 Maskara et al.
 2015/0005588 A1 1/2015 Herken et al.

FOREIGN PATENT DOCUMENTS

EP 0335356 B1 3/1996
 EP 1455640 B1 1/2008
 EP 1720446 B1 7/2010
 JP 5115450 A 5/1993
 JP 2002200059 A 7/2002
 WO 200002484 A1 1/2000
 WO 2004054656 A1 7/2004
 WO 2004067083 A2 8/2004
 WO 2004078259 A1 9/2004
 WO 2005082454 A1 9/2005
 WO 2006050325 A2 5/2006
 WO 2007019325 A2 2/2007
 WO 2009034506 A1 3/2009
 WO 2010025432 A1 3/2010
 WO 2010077997 A2 7/2010

OTHER PUBLICATIONS

De Bock et al., "Captopril Treatment of Chronic Heart Failure in the Very Old", *Journal of Gerontology: Medical Sciences*, 1994, pp. M148-M152, vol. 49, No. 3.
<http://web.archive.org/web/20030427001846/http://www.lifecor.com/imagelib/imageproduct.asp>; Published by LifeCor, Inc., 2002, on webpage owned by LifeCor, Inc.
 "ATS Statement: Guidelines for the Six-Minute Walk Test", *American Journal of Respiratory and Critical Care Medicine*, 2002, pp. 111-117, vol. 166, American Thoracic Society, , available at <http://ajrccm.atsjournals.org/cgi/content/full/166/1/111>.
 O'Keeffe et al., "Reproductivity and Responsiveness of Quality of Life Assessment and Six Minute Walk Test in Elderly Heart Failure Patients", *Heart*, 1998, 80: 377-382.




FIG. 1




FIG. 2




FIG. 3B




FIG. 3A




FIG. 4




FIG. 5




FIG. 6

**MEDICAL DEVICE WITH ACOUSTIC
SENSOR****RELATED APPLICATIONS**

This application claims benefit under 35 U.S.C. § 119(e) of U.S. Provisional Application Ser. No. 62/134,881, titled "MEDICAL DEVICE WITH ACOUSTIC SENSOR," filed Mar. 18, 2015, which is hereby incorporated herein by reference in its entirety.

BACKGROUND**Technical Field**

This disclosure relates to medical devices, and more particularly to medical devices that monitor the cardiopulmonary system using one or more acoustic sensors.

Discussion

Some conventional medical devices that monitor the cardiopulmonary system obtain a subject's electrocardiogram (ECG) signal from body surface electrodes. Determining the true characteristics of an ambulatory subject's cardiac cycle based on an ECG signal in this manner can be difficult. Electrical noise and electrode fall-off frequently degrade the quality of the ECG signal. In addition, the characteristics of ECG signals vary from subject to subject due to factors such as the subject's state of health and individual physiology.

Known ambulatory wearable defibrillators, such as the LifeVest® Wearable Cardioverter Defibrillator available from ZOLL Medical Corporation of Chelmsford, Mass., use four ECG sensing electrodes in a dual-channel configuration. That is, an electrical signal provided by one of the four ECG sensing electrodes is paired with the electrical signal provided by another of the four ECG sensing electrodes to form a channel. This arrangement of ECG sensing electrodes is usually suitable because in most cases it is rare that noise or electrode movement affects the entire body circumference. The dual-channel configuration provides redundancy and allows the system to operate on a single channel if necessary. Because signal quality also varies from subject to subject, having two channels provides the opportunity to have improved signal pickup, since the ECG sensing electrodes are located in different body positions.

SUMMARY

Some aspects and examples disclosed herein manifest an appreciation for the sporadic inability of medical devices relying solely on conventional ECG sensing electrodes to determine the functional effect of a cardiopulmonary anomaly. For example, medical devices in accord with at least one example disclosed herein utilize a combination of electrode and acoustic sensor data to detect cardiopulmonary anomalies with increased accuracy and precision relative to conventional ECG sensing electrode based systems. With this enhanced cardiopulmonary data, various medical devices disclosed herein are better able to discriminate between detected cardiopulmonary anomalies that substantially impair cardiac or pulmonary function and those that do not. Using the enhanced cardiopulmonary data, these medical devices may modify the manner in which treatment is provided to a subject. For example, the medical devices may treat those anomalies that substantially impair cardiopulmonary function and may defer treatment where cardiopulmonary function is not substantially impaired.

In at least one example, a medical device is provided. The medical device includes at least one therapy electrode, at least one electrocardiogram (ECG) electrode, at least one acoustic sensor, and at least one processor coupled with the at least one acoustic sensor, the at least one ECG electrode, and the at least one therapy electrode. The at least one processor can receive at least one acoustic signal from the at least one acoustic sensor, receive at least one electrode signal from the ECG electrode, detect at least one unverified cardiopulmonary anomaly using the at least one electrode signal, and verify the at least one unverified cardiopulmonary anomaly with reference to data descriptive of the at least one acoustic signal.

In the medical device, the data descriptive of the acoustic signal includes at least one of S1, S2, S3, S4, EMAT, % EMAT, SDI, and LVST. The unverified cardiopulmonary anomaly includes at least one of supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure. The at least one processor can verify disconnection of the at least one ECG electrode with reference to the data descriptive of the at least one acoustic signal.

The medical device may further include at least one accelerometer coupled with the at least one processor, wherein the at least one processor can further verify the at least one unverified cardiopulmonary anomaly with reference to data descriptive of at least one motion signal from the at least one accelerometer. In the medical device, the at least one processor can acquire and record the at least one acoustic signal during a configurable period of time. The at least one processor can also execute a sleep test in response to receiving input requesting the sleep test.

In another example, a medical device is provided. The medical device includes at least one electrocardiogram (ECG) electrode, at least one acoustic sensor, and at least one processor coupled with the at least one acoustic sensor and the at least one ECG electrode and configured to receive at least one acoustic signal from the at least one acoustic sensor, to receive at least one electrical signal from the at least one ECG electrode, and to detect at least one cardiopulmonary anomaly using the at least one electrical signal and the at least one acoustic signal. The medical device may include an ambulatory medical device.

In the medical device, the at least one processor may be configured to detect the at least one cardiopulmonary anomaly at least in part by identifying the at least one cardiopulmonary anomaly using the at least one electrical signal and verifying the at least one cardiopulmonary anomaly using the at least one acoustic signal. The medical device may further include a garment. The at least one ECG electrode and the at least one acoustic sensor may be integrated with the garment. The medical device may further include at least one therapy electrode coupled with the at least one processor and integrated with the garment. The at least one acoustic signal may include at least one of S1, S2, S3, and S4. The at least one processor may be configured to detect the at least one cardiopulmonary anomaly at least in part by calculating, based on the at least one acoustic signal, at least one of electromechanical activation time (EMAT), EMAT as a percentage of cardiac cycle time, systolic dysfunction index, and left ventricular systolic time. The at least one cardiopulmonary anomaly may include at least one of supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure.

In the medical device, the at least one processor may be configured to detect disconnection of the at least one ECG

electrode based on the at least one acoustic signal. The medical device may further include at least one motion sensor coupled with the at least one processor. The at least one processor may be configured to receive at least one motion signal from the at least one motion sensor and to detect the at least one cardiopulmonary anomaly using the at least one electrical signal, the at least one acoustic signal, and the at least one motion signal.

In the medical device, the at least one acoustic sensor may include at least one motion sensor. The at least one acoustic sensor may include at least one accelerometer. The at least one processor may be configured to receive at least one signal from the at least one accelerometer and to partition the at least one signal into the at least one acoustic signal and at least one motion signal. The at least one motion signal may include frequencies less than approximately 10 hertz. The at least one acoustic signal may include frequencies greater than approximately 10 hertz. The at least one processor may be configured to partition the at least one signal into a first frequency band for subject motion, a second frequency band for chest compression motion and/or sounds, a third frequency band for heart sounds, and a fourth frequency band for breath sounds. The first frequency band may include frequencies less than approximately 6 hertz, the second frequency band may include frequencies between approximately 6 and 20 hertz, the third frequency band may include frequencies between approximately 20 hertz and 150 hertz, and the fourth frequency band may include frequencies between approximately 300 to 1200 hertz.

In another example, a method of monitoring a subject using a wearable defibrillator is provided. The wearable defibrillator includes at least one electrocardiogram (ECG) electrode and at least one acoustic sensor. The method comprises acts of receiving at least one acoustic signal from the at least one acoustic sensor, receiving at least one electrical signal from the at least one ECG electrode, and detecting at least one cardiopulmonary anomaly using the at least one electrical signal and the at least one acoustic signal.

In the method, the act of receiving the at least one acoustic signal may include an act of receiving at least one of S1, S2, S3, and S4. The act of detecting the at least one cardiopulmonary anomaly may include an act of calculating, based on the at least one acoustic signal, at least one of electromechanical activation time (EMAT), EMAT as a percentage of cardiac cycle time, systolic dysfunction index, and left ventricular systolic time. The act of detecting the at least one cardiopulmonary anomaly may include an act of detecting at least one of supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure.

The medical device may further include at least one motion sensor. The method may further include acts of receiving at least one motion signal from the at least one motion sensor and detecting the at least one cardiopulmonary anomaly using the at least one electrical signal, the at least one acoustic signal, and the at least one motion signal.

In another example, a medical device is provided. The medical device includes a memory, at least one accelerometer, and at least one processor coupled with the at least one accelerometer and the memory. The at least one processor is configured to receive at least one acoustic signal from the at least one accelerometer, to receive at least one motion signal from the at least one accelerometer, and to execute a sleep test configured to store data descriptive of the at least one acoustic signal and the at least one motion signal in the memory.

The medical device may further include comprising at least one ECG electrode coupled with the at least one processor. The at least one processor may be configured to receive at least one electrical signal from the at least one ECG electrode and the sleep test is configured to store data descriptive of the at least one electrical signal in the memory. The medical device may further include a garment. The at least one ECG electrode and the at least one accelerometer may be integrated with the garment. The medical device may further include at least one therapy electrode coupled with the at least one processor and integrated with the garment. The at least one processor may be configured to execute the sleep test at a configurable time. The medical device may further include at least one motion sensor distinct from the accelerometer and coupled with the at least one processor. The at least one motion sensor may be positioned on a wrist of a subject. The at least one processor may be configured to receive one or more motion signals from the at least one motion sensor. The sleep test may be configured to store data descriptive of the one or more motion signals in the memory.

Still other aspects and advantages of the examples disclosed herein are discussed in detail below. Moreover, it is to be understood that both the foregoing information and the following detailed description are merely illustrative examples of various aspects, and are intended to provide an overview or framework for understanding the nature and character of the claimed subject matter. Any example disclosed herein may be combined with any other example. References to "an example," "some examples," "an alternate example," "various examples," "one example," "at least one example," "this and other examples" or the like are not necessarily mutually exclusive and are intended to indicate that a particular feature, structure, or characteristic described in connection with the example may be included in at least one example. The appearances of such terms herein are not necessarily all referring to the same example.

Furthermore, in the event of inconsistent usages of terms between this document and documents incorporated herein by reference, the term usage in the incorporated references is supplementary to that of this document; for irreconcilable inconsistencies, the term usage in this document controls. In addition, the accompanying drawings are included to provide illustration and a further understanding of the various aspects and examples, and are incorporated in and constitute a part of this specification. The drawings, together with the remainder of the specification, serve to explain principles and operations of the described and claimed aspects and examples.

BRIEF DESCRIPTION OF DRAWINGS

The accompanying drawings are not intended to be drawn to scale. In the drawings, components that are identical or nearly identical may be represented by a like numeral. For purposes of clarity, not every component is labeled in every drawing. In the drawings:

FIG. 1 is a functional schematic one example of a pacing device;

FIG. 2 is an illustration of one example of an ambulatory medical device;

FIGS. 3A-B are illustrations of one example of a medical device controller for an ambulatory medical device;

FIG. 4 is an illustration of one example of an external medical device;

FIG. 5 is a flow diagram of one example of a process analyzing enhanced cardiopulmonary data; and

FIG. 6 is an illustration of potential locations for accelerometers/acoustic sensors on the body of a subject.

DETAILED DESCRIPTION

Medical devices in accord with various examples disclosed herein utilize enhanced cardiopulmonary data to implement a wide variety of functionality. For instance, according to some examples, a medical device includes a cardiopulmonary function analyzer configured to process enhanced cardiopulmonary data, which may include acoustic and electrode signals among other signals, to identify cardiopulmonary anomalies that are difficult to detect using conventional ECG sensing electrodes. Examples of these cardiopulmonary anomalies include super ventricular tachycardia (SVT) as distinct from ventricular tachycardia (VT), asystole, ECG sensing electrode falloff, pulseless electrical activity (PEA), asystole, pseudo-PEA, bradycardia, heart murmurs, sleep apnea, and respiratory arrest.

In some examples, the cardiopulmonary function analyzer also determines how to address the identified cardiopulmonary anomaly. For example, where the medical device is a monitor, the cardiopulmonary function analyzer may issue an alarm describing the anomaly to an external entity, such as a user or a computer system distinct from the medical device. Where the medical device is a defibrillator, the cardiopulmonary function analyzer may address the anomaly by delivering a defibrillating shock to a subject. Where the medical device is a pacing device, the cardiopulmonary function analyzer may address the anomaly by delivering one or more pacing pulses.

Any of the medical devices disclosed herein may be non-invasive, bodily-attached, or ambulatory. As used herein, the term non-invasive means that the device does not penetrate the body of a subject. This is in contrast to invasive devices, such as implantable medical devices, in which at least a portion of the device is disposed subcutaneously. The term bodily-attached means that at least a portion of the device (other than its electrodes in the case of a defibrillator, cardioverter, or pacer) is removably attached to the body of a subject, such as by mechanical coupling (e.g., by a wrist strap, cervical collar, bicep ring), adhesion (e.g., by an adhesive gel intermediary), suction, magnetism, fabric or other flexible material (e.g., by straps or integration into a garment) or other body mounting features not limited by the aforementioned examples. These coupling elements hold the device in a substantially fixed position with respect to the body of the subject. The term ambulatory means that the device is capable of and designed for moving with the subject as the subject goes about their daily routine.

The examples of the methods and apparatuses discussed herein are not limited in application to the details of construction and the arrangement of components set forth in the following description or illustrated in the accompanying drawings. The methods and apparatuses are capable of implementation in other examples and of being practiced or of being carried out in various ways. Examples of specific implementations are provided herein for illustrative purposes only and are not intended to be limiting. In particular, acts, elements and features discussed in connection with any one or more examples are not intended to be excluded from a similar role in any other examples.

Also, the phraseology and terminology used herein is for the purpose of description and should not be regarded as limiting. Any references to examples or elements or acts of the systems and methods herein referred to in the singular may also embrace examples including a plurality of these

elements, and any references in plural to any example or element or act herein may also embrace examples including only a single element. References in the singular or plural form are not intended to limit the presently disclosed systems or methods, their components, acts, or elements. The use herein of “including,” “comprising,” “having,” “containing,” “involving,” and variations thereof is meant to encompass the items listed thereafter and equivalents thereof as well as additional items. References to “or” may be construed as inclusive so that any terms described using “or” may indicate any of a single, more than one, and all of the described terms.

Medical Device

Various examples disclosed herein utilize enhanced cardiopulmonary data (e.g., acoustic data and electrode data) to detect and treat cardiopulmonary anomalies, such as SVT as distinct from VT, asystole, ECG sensing electrode falloff, PEA, pseudo-PEA, bradycardia, heart murmurs, sleep apnea, and respiratory arrest. FIG. 1 illustrates a medical device 100 in accord with some examples. As shown in FIG. 1, the medical device 100 includes a medical device controller 102, one or more accelerometers/acoustic sensors 122, and one or more electrodes 120. As illustrated in FIG. 1, the medical device controller 102 includes a processor 118, a sensor interface 114, a cardiopulmonary function analyzer 112, a therapy delivery interface 116, data storage 104, a network interface 106, a user interface 108, and a battery 110. The processor 118 can be implemented using a variety of commercially available processors. Specific examples of the processor 118 are described further below. The data storage 104 includes cardiopulmonary data 117. The sensor interface 114 can include an acoustic signal processing component 124, an electrode signal processing component 126, and motion signal processing component 128. In some implementations, components such as cardiopulmonary function analyzer 112 and sensor interface 114 can be implemented within processor 118. In some implementations, the components can be implemented in circuitry that is separate from processor 118.

The medical device 100 may be any of a variety of medical devices including defibrillators, monitors, CPR systems, pacing devices, and other medical devices. More specific examples of medical devices in accord with the medical device 100 are described further below with reference to FIGS. 2-4.

As shown in FIG. 1, the acoustic signal processing component 124 is coupled to and receives acoustic signals from the accelerometer/acoustic sensor 122. Similarly, the electrode signal processing component 126 is coupled to and receives electrode signals from the electrode 120. The electrode 120 may comprise any of a variety of commercially available electrodes, some examples of which are described further below.

Likewise, the motion signal processing component 128 is coupled to and receives motion signals from the accelerometer/acoustic sensor 122. As illustrated in FIG. 1, in some examples, the cardiopulmonary function analyzer 112 can be coupled with (via the processor 118) and receive processed acoustic data from the acoustic signal processing component 124, processed electrode data from the electrode signal processing component 126, and processed motion data from the motion signal processing component 128. Examples of processes executed by the acoustic signal processing component 124, the electrode signal processing component 126, and the motion signal processing component 128 are described further below with reference to FIG. 5.

According to one example illustrated by FIG. 1, the cardiopulmonary function analyzer 112 is configured to detect heart beats, breath sounds, and cardiopulmonary anomalies and determine whether the detected anomalies substantially impair cardiac or pulmonary function. For example, the cardiopulmonary function analyzer 112 can be implemented using a variety of hardware or software components. When executing according to this configuration, in some examples, the cardiopulmonary function analyzer 112 detects cardiopulmonary anomalies by scanning processed acoustic data and processed electrode data for patterns indicative of cardiopulmonary anomalies. Responsive to identifying a data pattern indicative of a cardiopulmonary anomaly, the cardiopulmonary function analyzer 112 identifies a routine to address the cardiopulmonary anomaly based on the identity of the anomaly and a confidence that the anomaly actually exists. Next, the cardiopulmonary function analyzer 112 initiates the identified routine. The data patterns scanned for by the cardiopulmonary function analyzer 112 are indicative of a wide variety of cardiopulmonary anomalies. Examples of these anomalies include cardiac arrhythmias (e.g., bradycardia, tachycardia, SVT, an irregular cardiac rhythm, PEA, and asystole), murmurs, sleep apnea, and respiratory arrest. The data patterns may also indicate problems with the medical device itself such as faulty or disconnected sensors. Specific examples of the data patterns scanned for by the cardiopulmonary function analyzer 112 and actions taken responsive to detection of the data patterns are described further below with reference to FIG. 5.

In some examples, the cardiopulmonary function analyzer 112 is configured to leverage the differing originating modalities of the processed acoustic data and electrode data to advantageous effect. For example, data patterns that indicate VT are sometimes present in processed electrode data where the subject is actually experiencing SVT. Accordingly, in an implementation, the cardiopulmonary function analyzer 112 can analyze the processed acoustic data to determine whether VT indicated by the processed electrode data actually exists. If the cardiopulmonary function analyzer 112 determines based on the processed acoustic data that no VT condition exists, the cardiopulmonary function analyzer 112 may not initiate treatment as doing so may convert SVT into VT, which poses a greater risk to the health of the subject than SVT. In some examples, as described further below with reference to FIG. 5, the cardiopulmonary function analyzer 112 verifies that no VT condition exists by comparing the intensity and variability of S1 to predefined thresholds. S1 is a heart sound described further below.

Similarly, in some examples, the cardiopulmonary function analyzer 112 can analyze processed acoustic data to determine whether the heart of a subject is mechanically pumping blood where the processed electrode data indicates asystole or PEA. If sufficient blood is being pumped (e.g., if the ejection fraction exceeds a predefined threshold), the cardiopulmonary function analyzer 112 may defer treatment. In some examples, as described further below with reference to FIG. 5, the cardiopulmonary function analyzer 112 determines whether the ejection fraction exceeds the predefined threshold by comparing EMAT to a predefined threshold. EMAT is a metric calculated based on heart sounds and is described further below.

In some examples, the cardiopulmonary function analyzer 112 can analyze processed acoustic data that covers periods of time where electrode data is not available (e.g., due to an enforced blanking interval, temporary electrode saturation,

or electrode fall off) and takes appropriate action based on the condition of a subject as indicated by the processed acoustic data.

In some implementations, the cardiopulmonary function analyzer 112 can analyze processed acoustic data and processed motion data received via multiple signals generated by the accelerometer/acoustic sensor 122. In this regard, the sensor 122 can provide the signals over one or more channels to maintain signal separation. In some implementations, the multiple signals can be combined or multiplexed within a signal channel. For example, the cardiopulmonary function analyzer 112 may analyze data representative of a subject's respiration encoded from a first channel, data representative of a subject's heart sounds encoded from a second channel, and data representative of a subject's position encoded from a third channel.

One example of a cardiopulmonary anomaly detection process that the cardiopulmonary function analyzer 112 is configured to execute is described further below with reference to FIG. 5.

In some implementations, the cardiopulmonary function analyzer 112 can be configured to execute an at-home sleep test. When executing according to this configuration, the cardiopulmonary function analyzer 112 may receive information indicating that a user has requested execution of the sleep test via a user interface of the medical device, such as the user interface 108. Responsive to receipt of this request information, the cardiopulmonary function analyzer 112 can record and store enhanced cardiopulmonary data for a configurable period of time (e.g., 6-10 hours) specified in the request information or within a discrete configurable parameter of the medical device. This period of time may, for example, represent time when the subject is expected to be asleep. Enhanced cardiopulmonary data can include information based on one or more of the processed signals from the sensor interface 114 and can include motion data, acoustic data, electrode data, or aggregated data generated by processing other data. In some examples, the configurable period of time can include a future time period (e.g., 2 days from the current day) or a reoccurring period (e.g., every other Tuesday for the next month). Upon completion of each sleep test recording, the cardiopulmonary function analyzer 112 may transmit the recorded data to a remote system via, for example, a network interface, such as the network interface 106.

In some examples, the medical device 100 can generate a variety of reports and metrics based on the data analyzed by the cardiopulmonary function analyzer 112. For example, the medical device 100 can cause a report to be presented to a user via the user interface 108. In some cases, the medical device 100 can cause a report (and any associated analytic metrics) to be transmitted to a network output device (e.g., a network printer on a same or different network as the medical device 100). In some instances, the report can be transmitted to a server on same or different network (e.g. at a remote location) for further processing and storage.

In an example, the cardiopulmonary function analyzer 112 may generate both summary and full disclosure reports in portable document format for each recorded period. The cardiopulmonary function analyzer 112 may save or print these reports in response to a request for the same. The full disclosure report that depicts all enhanced cardiopulmonary data (e.g., electrode data and acoustic data) for each recorded period. The summary report summarizes the information presented in the full disclosure report.

In some examples, the variety of reports include one or more reports that present information descriptive of the

hemodynamic performance of the heart and respiratory effort of a subject. For instance, a report may provide information descriptive of chest rise and fall during respiration as determined from acoustic vibration of the chest wall during respiration. These reports may also display information descriptive of the restlessness of a subject, as determined from processed motion data. This processed motion data may be generated by one or more accelerometers positioned on the torso and wrist of a subject, such as within a pocket, belt, or watch worn by the subject and in communication with the cardiopulmonary function analyzer 112 via a BAN. For example, such data may provide insights into a subject's body position during sleep or limb movements and position. The processed motion data may also be generated by an internal 3-axis gyroscope included in some examples.

In some examples, the cardiopulmonary function analyzer 112 reads values of one or more configurable parameters that specify targeted operational characteristics of the cardiopulmonary function analyzer 112. These operational parameters may include upload filter criteria that specifies the data type and frequency with which the cardiopulmonary function analyzer 112 transmits enhanced cardiopulmonary data to a remote computer, such as the centralized server described further below with reference to FIG. 2 via the network interface 106.

In some examples illustrated by FIG. 1, the acoustic signal processing component 124 is configured to detect and record a variety of sounds related to cardiopulmonary function. To process analog and digital acoustic signals received from the accelerometer/acoustic sensor 122, the acoustic signal processing component 124 may include various circuitry, such as amplifiers, filters, transducers, analog to digital converters, analog signal processors, and digital signal processors. In at least one example, the acoustic signal processing component 124 processes signals received from one or more acoustic channels. Further according to some examples, the acoustic signal processing component 124 transmits processed acoustic data descriptive of the acoustic signals to the cardiopulmonary function analyzer 112 for subsequent analysis. In some examples, the processor 118 exchanges data with the sensor interface 114 and executes the cardiopulmonary function analyzer 112. Additional aspects and functions of the processor 118 are described further below.

In healthy adults, there are at least two normal heart sounds, commonly referred to as S1 and S2. A third heart sound, commonly referred to as S3 (also called a protodiastolic gallop or ventricular gallop), may be indicative of a problem with a subject's heart when present. For example, in subjects over 40 years old, S3 has been associated with an abnormal diastolic filling pattern. The presence of S3 may signal cardiac problems like a failing left ventricle as in dilated congestive heart failure. A fourth heart sound, commonly referred to as S4 (also called a presystolic gallop or atrial gallop), is indicative of a problem with a subject's heart when present. For example, S4 is often associated with an increased left ventricular stiffness. Heart murmurs may also be present in some subjects and may indicate cardiac problems.

In some examples, the acoustic signal processing component 124 is configured to detect and record heart sound values including any one or all of S1, S2, S3, and S4. Other heart sound values which may be monitored and recorded by the acoustic signal processing component 124 may include any one or more of electromechanical activation time

(EMAT), percentage of EMAT (% EMAT), systolic dysfunction index (SDI), and left ventricular systolic time (LVST).

EMAT is generally measured from the onset of the Qwave on the ECG to the closure of the mitral valve within the S1 heart sound. Prolonged EMAT has been associated with reduced left ventricular ejection fraction (LV EF, being a measure of how much blood is being pumped out of the left ventricle of the heart with each contraction).

% EMAT is EMAT divided by a dominant RR interval. In this regard, % EMAT is related to an efficiency of the pump function of the heart.

SDI is a multiplicative combination of ECG and heart sound values (EMA, S3, QRS duration and QR interval). SDI can be used to predict left ventricular systolic dysfunction.

LVST is the systolic portion of the cardiac cycle and is defined as the time interval between the S1 and the S2 heart sounds. LVST may have some heart rate dependence, and can be approximately 40% (range 30-50%) of the cardiac cycle. LVST can be affected by disease that produces poor contractility or a low ejection fraction.

In some examples, during an initial fitting of the medical device, the medical device may record enhanced cardiopulmonary data to use as a baseline for an individual subject. For example, a baseline EMAT may be recorded and later used to analyze degree of degradation occurring within the cardiopulmonary function of the subject. In one example, the baseline EMAT and baseline ejection fraction can be recorded and subsequent EMAT values can be used to approximate subsequent ejection fractions.

In some examples, the recorded heart sound values described above may be stored in non-volatile data storage. For example, in one example these values are stored as cardiopulmonary data 117 within the data storage 104. In at least one example, the cardiopulmonary function analyzer 112 periodically monitors the recorded heart sound values for events of interest. In this example, where the recorded heart sound values indicate a potential health risk, the cardiopulmonary function analyzer 112 may notify an external entity of the risk. For example, where an EMAT value degrades gradually over a specified period of time, the cardiopulmonary function analyzer 112 may transmit a communication to a healthcare provider indicating this degradation.

In other examples, when executing according to its configuration, the acoustic signal processing component 108 identifies and records breathing sounds such as those that accompany normal respiration, snoring, episodes of sleep apnea, and respiratory arrest.

In some examples, the acoustic signal processing component 124 reads values of one or more configurable parameters that specify targeted operational characteristics of the acoustic signal processing component 124 or the accelerometer/acoustic sensor 122. These operational parameters may specify the sampling rate, filter coefficients, recording duration and interval, and noise thresholds, used to process acoustic data. The value of the recording duration and interval operational parameter specifies the length of time of each acoustic recording and the interval of time between recordings. In one example, the value of the recording duration and interval operational parameter specifies a duration of 5 seconds to be recorded every 10 minutes. In some examples, the acoustic signal processing component 124 and the accelerometer/acoustic sensor 122 continuously record, process, and store processed acoustic data.

In one example illustrated by FIG. 1, the electrode signal processing component 126 is configured to detect and record

cardiac activity of a subject. For example, when executing according to this configuration, the electrode signal processing component **126** may detect and record ECG signals. Further according to this example, the electrode signal processing component **126** transmits information descriptive of the ECG signals to the cardiopulmonary function analyzer **112** for subsequent analysis. As described above, the processed data from the electrode signal processing component **126** can be monitored for cardiopulmonary anomalies either alone or in combination with processed data from the acoustic signal processing component **124** or the motion signal processing component **128**.

In one example illustrated by FIG. 1, the motion signal processing component **128** is configured to detect and record motion and physical movement of a subject. For example, when executing according to this configuration, the motion signal processing component **128** may detect and record physical movement such as walking, falling, breath motion (e.g., chest rising and falling), and the like. Further according to this example, the motion signal processing component **128** transmits information descriptive of the accelerometer signals to the cardiopulmonary function analyzer **112** for subsequent analysis.

In one example illustrated by FIG. 1, the accelerometer/acoustic sensor **122** may comprise any device that may detect sounds from a subject's cardiopulmonary system and provide an output signal responsive to the detected heart and breath sounds. In some examples the accelerometer/acoustic sensor **122** comprises a microphone or an accelerometer. In some examples, the accelerometer comprises a microelectromechanical system (MEMS) accelerometer or a multi-channel accelerometer, for example, a three channel accelerometer. The acoustic sensor may comprise a three channel accelerometer configured to sense movement in each of three orthogonal axes. An example accelerometer which may be utilized in some examples is a LIS344ALH accelerometer, available from STMicroelectronics.

The accelerometer/acoustic sensor **122** and associated electronics may be configured to monitor any one or more of a subject's respiration, a subject's heart sounds, a subject's position, and an activity level of a subject. The accelerometer/acoustic sensor **122** and associated electronics may additionally or alternatively be configured to monitor other sounds which may be indicative of a state of health of a subject, for example, gastrointestinal sounds or the sounds of snoring or the absence of such sounds to, for example, provide an indication of the subject experiencing sleep apnea or respiratory arrest. The accelerometer/acoustic sensor **122** may provide signals indicative of the subject's heart sounds on a first channel, signals indicative of the subject's position on a second channel, and signals indicative of the subject's activity level on a third channel. In some examples, the different channels may be utilized to provide signals indicative of more than one physiological characteristic or other characteristics associated with the state of the subject. For example, the accelerometer/acoustic sensor **122** may provide signals indicative of the subject's heart sounds on a first channel, signals indicative of the subject's respiration on a second channel, and signals indicative of the subject's body position on any or all of the first, second, and a third channel. It should be appreciated that dependent on the underlying characteristic that is being monitored, multiple signals related to the characteristic being monitored may be received over a single channel or a number of different channels.

In one example illustrated by FIG. 1, the electrode **120** may include any type of sensing electrode, such as one or

more ECG sensing electrodes as described further below with reference to FIGS. 2 and 4.

In some examples in accord with FIG. 1, the battery **110** is a rechargeable 3 cell 2200 mAh lithium ion battery pack that provides electrical power to the other device components with a minimum 24 hour runtime between charges. It is appreciated that the battery capacity, runtime, and type (e.g., lithium ion, nickel-cadmium, or nickel-metal hydride) may be changed to best fit the specific application of the medical device controller **102**.

According to the example illustrated in FIG. 1, the processor **118** is coupled to the sensor interface **114**, the therapy delivery interface **116**, the data storage **104**, the network interface **106**, and the user interface **108**. The processor **118** performs a series of instructions that result in manipulated data which are stored in and retrieved from the data storage **104**. According to a variety of examples, the processor **118** is a commercially available processor such as a processor manufactured by Texas Instruments, Intel, AMD, Sun, IBM, Motorola, Freescale, and ARM Holdings. However, the processor **118** may be any type of processor, multiprocessor or controller, whether commercially available or specially manufactured. For instance, according to one example, the processor **118** may include a power conserving processor arrangement as described in U.S. Pat. No. 8,904,214, titled SYSTEM AND METHOD FOR CONSERVING POWER IN A MEDICAL DEVICE, issued Dec. 2, 2014, which is hereby incorporated herein by reference in its entirety. In one example, the processor **118** is an Intel® PXA270.

In addition, in several examples the processor **118** is configured to execute a conventional real-time operating system (RTOS), such as RTLinux. In these examples, the RTOS may provide platform services to application software, such as some examples of the cardiopulmonary function analyzer **112**. These platform services may include inter-process and network communication, file system management and standard database manipulation. One or more of many operating systems may be used, and examples are not limited to any particular operating system or operating system characteristic. For instance, in some examples, the processor **118** may be configured to execute a non-real time operating system, such as BSD or GNU/Linux.

As illustrated in FIG. 1, the cardiopulmonary function analyzer **112**, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** may be implemented using hardware or a combination of hardware and software. For instance, in one example, the cardiopulmonary function analyzer **112**, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** are implemented as software components that are stored within the data storage **104** and executed by the processor **118**. In this example, the instructions included in the cardiopulmonary function analyzer **112**, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** program the processor **118** to analyze the cardiopulmonary function of a subject. In some examples, cardiopulmonary function analyzer **112**, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** may be application-specific integrated circuits (ASICs) that are coupled to the processor **118** and tailored to analyze the cardiopulmonary function of a subject. Thus, examples of the cardiopulmonary function analyzer **112**, the acoustic signal pro-

cessing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** are not limited to a particular hardware or software implementation.

In some examples, the components disclosed herein, such as the cardiopulmonary function analyzer **112**, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** may read parameters that affect the functions performed by the components. These parameters may be physically stored in any form of suitable memory including volatile memory, such as RAM, or nonvolatile memory, such as a flash memory or magnetic hard drive. In addition, the parameters may be logically stored in a propriety data structure, such as a database or file defined by a user mode application, or in a commonly shared data structure, such as an application registry that is defined by an operating system. In addition, some examples provide for both system and user interfaces, as may be implemented using the user interface **108**, that allow external entities to modify the parameters and thereby configure the behavior of the components.

The data storage **104** includes a computer readable and writeable nonvolatile data storage medium configured to store non-transitory instructions and data. In addition, the data storage **104** includes processor memory that stores data during operation of the processor **118**. In some examples, the processor memory includes a relatively high performance, volatile, random access memory such as dynamic random access memory (DRAM), static memory (SRAM) or synchronous DRAM. However, the processor memory may include any device for storing data, such as a non-volatile memory, with sufficient throughput and storage capacity to support the functions described herein. According to several examples, the processor **118** causes data to be read from the nonvolatile data storage medium into the processor memory prior to processing the data. In these examples, the processor **118** copies the data from the processor memory to the non-volatile storage medium after processing is complete. A variety of components may manage data movement between the non-volatile storage medium and the processor memory and examples are not limited to particular data management components. Further, examples are not limited to a particular memory, memory system, or data storage system.

The instructions stored on the data storage **104** may include executable programs or other code that can be executed by the processor **118**. The instructions may be persistently stored as encoded signals, and the instructions may cause the processor **118** to perform the functions described herein. The data storage **104** also may include information that is recorded, on or in, the medium, and this information may be processed by the processor **118** during execution of instructions. The medium may, for example, be optical disk, magnetic disk or flash memory, among others, and may be permanently affixed to, or removable from, the medical device controller **102**.

In some examples, the cardiopulmonary data **117** includes cardiopulmonary data detected, identified, and stored by the cardiopulmonary function analyzer **112**. More particularly, according to the illustrated example, the cardiopulmonary data **117** includes information descriptive of cardiac function and respiratory function. For example, the cardiopulmonary data **117** may include data such as ECG signal data, interpretations of the ECG signal data (e.g., heartbeats), analog heart sounds, analog breath sounds, analog motion data,

acoustic signals, electrode signals, motion signals, processed motion data, processed acoustic data, and processed electrode data.

As illustrated in FIG. 1, the cardiopulmonary function analyzer **112** and the cardiopulmonary data **117** are separate components. However, in some examples, the cardiopulmonary function analyzer **112** and the cardiopulmonary data **117** may be combined into a single component or re-organized so that a portion of the data included in the cardiopulmonary function analyzer **112**, such as executable code that causes the processor **118** to analyze enhanced cardiopulmonary data, resides in the cardiopulmonary data **117**, or vice versa. Such variations in these and the other components illustrated in FIG. 1 are intended to be within the scope of the examples disclosed herein.

The cardiopulmonary data **117** may be stored in any logical construction capable of storing information on a computer readable medium including, among other structures, flat files, indexed files, hierarchical databases, relational databases or object oriented databases. These data structures may be specifically configured to conserve storage space or increase data exchange performance. In addition, various examples organize the cardiopulmonary data **117** into particularized and, in some cases, unique structures to perform the functions disclosed herein. In these examples, the data structures are sized and arranged to store values for particular types of data, such as integers, floating point numbers, character strings, arrays, linked lists, and the like.

As shown in FIG. 1, the medical device controller **102** includes several system interface components **116**, **106**, and **114**. Each of these system interface components is configured to exchange, i.e. send or receive, data with one or more specialized devices that may be located within the housing of the medical device controller **102** or elsewhere. The components used by the interfaces **116**, **106**, and **114** may include hardware components, software components, or a combination of both. Within each interface, these components physically and logically couple the medical device controller **102** to the specialized devices. This physical and logical coupling enables the medical device controller **102** to communicate with and, in some examples, power or control the operation of the specialized devices. These specialized devices may include physiological sensors, therapy delivery devices, and computer networking devices.

According to various examples, the hardware and software components of the interfaces **116**, **106**, and **114** implement a variety of coupling and communication techniques. In some examples, the interfaces **116**, **106**, and **114** use leads, cables or other wired connectors as conduits to exchange data between the medical device controller **102** and specialized devices. In some examples, the interfaces **116**, **106**, and **114** communicate with specialized devices using wireless technologies such as radio frequency, infrared technology, and body area network (BAN) technology. The software components included in the interfaces **116**, **106**, and **114** enable the processor **118** to communicate with specialized devices. These software components may include elements such as objects, executable code, and populated data structures. Together, these software components provide software interfaces through which the processor **118** can exchange information with specialized devices. Moreover, in at least some examples where one or more specialized devices communicate using analog signals, the interfaces **116**, **106**, and **114** further include components configured to convert analog information into digital information, and vice versa, to enable the processor **118** to communicate with specialized devices.

15

As discussed above, the system interface components **116**, **106**, and **114** shown in FIG. 1 support different types of specialized devices. For instance, the components of the sensor interface **114** couple the processor **118** to one or more physiological sensors such as a body temperature sensors, respiration monitors, and ECG sensing electrodes, one or more environmental sensors such as atmospheric thermometers, airflow sensors, video sensors, audio sensors, accelerometers, GPS locators, and hygrometers. In these examples, the sensors may include sensors with a relatively low sampling rate, such as wireless sensors. Additionally, in at least one example, the acoustic signal processing component **124**, the electrode signal processing component **126**, and the motion signal processing component **128** described above with reference to FIG. 1 are integrated into the sensor interface **114**.

The components of the therapy delivery interface **116** couple one or more therapy delivery devices, such as capacitors, defibrillator electrode assemblies, pacing electrode assemblies, or mechanical chest compression devices, to the processor **118**. It is appreciated that the functionality of the therapy delivery interface **116** may be incorporated into the sensor interface **114** to form a single interface coupled to the processor **118**. In addition, the components of the network interface **106** couple the processor **118** to a computer network via a networking device, such as a bridge, router or hub. According to a variety of examples, the network interface **106** supports a variety of standards and protocols, examples of which include USB (via, for example, a dongle to a computer), TCP/IP, Ethernet, Wireless Ethernet, Bluetooth®, ZigBee, M-Bus, CAN-bus, IP, IPV6, UDP, DTN, HTTP, FTP, SNMP, CDMA, NMEA and GSM. It is appreciated that the network interface **106** of medical device controller **102** may enable communication between other medical device controllers within a certain range.

To ensure data transfer is secure, in some examples, the medical device controller **102** can transmit data via the network interface **106** using a variety of security measures including, for example, TLS, SSL, or VPN. In some examples, the network interface **106** includes both a physical interface configured for wireless communication and a physical interface configured for wired communication. According to various examples, the network interface **106** enables communication between the medical device controller **102** and a variety of personal electronic devices including, for example, computer enabled glasses, watches, and earpieces. In these examples, the network interface **106** may connect to and communicate through a body area network.

Thus, the various system interfaces incorporated in the medical device controller **102** allow the device to interoperate with a wide variety of devices in various contexts. For instance, some examples of the medical device controller **102** are configured to perform a process of sending critical events and data to a centralized server via the network interface **106**. An illustration of a process in accord with these examples is disclosed in U.S. Pat. No. 6,681,003, titled "DATA COLLECTION AND SYSTEM MANAGEMENT FOR SUBJECT-WORN MEDICAL DEVICES," issued on Jan. 20, 2004, which is hereby incorporated by reference in its entirety.

As illustrated in FIG. 1, the therapy delivery interface **116** and the network interface **106** are optional and may not be included in every example. For instance, a heart rate monitor may employ the medical device controller **102** to issue alarms but may not include a therapy delivery interface **116** to treat cardiopulmonary abnormalities. Similarly, an ambu-

16

latory defibrillator may include the medical device controller **102** to provide defibrillation functionality but may not include a network interface **106** where, for example, the ambulatory defibrillator is designed to rely on the user interface **108** to announce alarms.

The user interface **108** shown in FIG. 1 includes a combination of hardware and software components that allow the medical device controller **102** to communicate with an external entity, such as a subject or other user. These components may be configured to receive information from actions such as physical movement, verbal intonation or thought processes. In addition, the components of the user interface **108** can provide information to external entities. Examples of the components that may be employed within the user interface **108** include keyboards, mouse devices, trackballs, microphones, electrodes, touch screens, printing devices, display screens, and speakers. In some examples, the electrodes include an illuminating element, such as an LED. In some examples, the printing devices include printers capable of rendering visual or tactile (Braille) output.

Other examples may include a variety of features not shown in FIG. 1. For example, although the accelerometer/acoustic sensor **122** and the electrode **120** are shown in FIG. 1 as discrete sensors, some examples may integrate the accelerometer/acoustic sensor **122** and the electrode **120** into a single assembly. In other examples, the accelerometer/acoustic sensor **122** is integrated within a therapy electrode assembly. One such arrangement is described further in U.S. Patent Application Publication No. 2015/0005588, titled "THERAPEUTIC DEVICE INCLUDING ACOUSTIC SENSOR," published Jan. 1, 2015, which is hereby incorporated herein by reference in its entirety. In other examples, the accelerometer/acoustic sensor **122** is integrated within a garment such as the garment described further below with reference to FIG. 2. For instance, the accelerometer/acoustic sensor **122** may be integrated within a belt, vest, or harness, such as the harness **210** or the belt **250**. Thus the examples disclosed herein are not limited to a particular number or arrangement of accelerometer/acoustic sensors or electrodes.

Example Ambulatory Medical Device

In some examples, the medical device **100** described above with reference to FIG. 1 is a wearable defibrillator that includes a garment (e.g., a vest or belt) that is worn by the subject. FIG. 2 illustrates a wearable defibrillator **200** in accord with these examples. In at least one example, the wearable defibrillator **200** may be a LifeVest® wearable cardioverter defibrillator available from ZOLL Medical Corporation of Chelmsford, Mass. The wearable defibrillator **200** monitors the subject's ECG with sensing electrodes, monitors the subject heart sounds with acoustic sensors, detects life-threatening arrhythmias, records events of interest, and delivers therapy in the form of one or more pacing pulses or a defibrillating shock through the therapy electrodes if treatment is necessary. As shown in FIG. 2, the wearable defibrillator **200** includes a harness **210** having a pair of shoulder straps and a belt that is worn about the torso of a subject. The wearable defibrillator **200** includes a plurality of ECG sensing electrodes **120** that are attached to the harness **210** at various positions about the subject's body and electrically coupled to the sensor interface **114** of the medical device controller **102** via a connection pod **220**. The plurality of ECG sensing electrodes **120** are coupled to the medical device controller **102** to monitor the cardiac function of the subject and generally include an anterior/posterior pair of ECG sensing electrodes and a side/side pair of ECG sensing electrodes. The plurality of ECG sensing

electrodes **120** may incorporate any electrode system, including conventional stick-on adhesive electrodes, dry-sensing capacitive ECG electrodes, radio transparent electrodes, segmented electrodes, or one or more long term wear electrodes that are configured to be continuously worn by a subject for extended periods (e.g., 3 or more days). One example of such a long term wear electrode is described in U.S. Patent Application Publication No. 2013/0325096, titled "LONG TERM WEAR MULTIFUNCTION BIO-MEDICAL ELECTRODE," published Dec. 5, 2013, which is hereby incorporated herein by reference in its entirety. Additional ECG sensing electrodes may be provided, and the plurality of ECG sensing electrodes **120** may be disposed at various locations about the subject's body.

The wearable defibrillator **200** also includes one or more accelerometer/acoustic sensors **122** that are attached to a belt **250** of the harness **210** at various positions about the subject's body and electrically coupled to the sensor interface **114** of the medical device controller **102** via the connection pod **220**. The one or more accelerometer/acoustic sensors **122** are coupled to the medical device controller **102** to monitor the cardiopulmonary function of the subject and generally positioned on the surface of a subject's body in the precordial area.

FIG. 6 illustrates several potential example locations, labeled P, A, B, and T where accelerometer/acoustic sensors **122** can be placed. In one example, the accelerometer/acoustic sensors **122** are positioned near the third intercostal parasternal area. Additionally, table 1 lists primary surface anatomy locations where the accelerometer/acoustic sensors **122** are positioned to detect various cardiac features and anomalies according to some examples.

TABLE 1

Primary Auscultation Location	Anomaly or Feature
Second intercostal space at right sternal border	Aortic Value
Second intercostal space at left sternal border	Pulmonary Value
Third intercostal space at left sternal border	Erb's Point
Fourth or fifth intercostal spaces at the left sternal border	Tricuspid Value
Fifth intercostal space at midclavicular line	Mitral Value

In some examples, the accelerometer/acoustic sensors **122** are positioned at secondary posterior locations that correspond to the anterior positions described above. For instance, the accelerometer/acoustic sensors **122** may be positioned at locations near the scapula such as the posterior auxiliary line (V7), midscapular location (V8), or paraspinal location (V9). In one example, the accelerometer/acoustic sensors **122** are positioned at a secondary surface location proximal to the atria.

Although not shown is FIG. 2, the wearable defibrillator **200** may include additional sensors, other than the plurality of ECG sensing electrodes **120**, capable of monitoring the physiological condition or activity of the subject. For example, sensors capable of measuring blood pressure (via, for example, video blood pressure detection), heart rate, heart sounds, thoracic impedance, pulse oxygen level (via, for example, reflectance-based pulse oximetry to determine oxygen concentration), respiration rate, and the activity level of the subject may also be provided.

The wearable defibrillator **200** also includes a plurality of therapy electrodes **214** that are electrically coupled to the medical device controller **102** via the connection pod **220** and which are configured to deliver one or more therapeutic defibrillating shocks to the body of the subject, if it is

determined that such treatment is warranted. Each therapy electrode of the plurality of therapy electrodes may be housed in a therapy electrode assembly that further includes conductive gel disposed within one or more reservoirs. Prior to delivering therapy, the therapy electrode assembly may dispense the conductive gel to improve conductivity between the therapy electrode and the body of the subject. The connection pod **220** electrically couples the plurality of ECG sensing electrodes **120** and the plurality of therapy electrodes **214** to the therapy delivery interface **116** of the medical device controller **102**, and may include electronic circuitry configured for this purpose. The connection pod **220** may also include other electronic circuitry, such as a motion sensor or accelerometer through which subject activity may be monitored.

As shown in FIG. 2, the wearable defibrillator **200** also includes a user interface pod **240** that is electrically coupled to, or integrated in with, the user interface **108** of the medical device controller **102**. The user interface pod **240** can be attached to the subject's clothing or to the harness **210**, for example, via a clip (not shown) that is attached to a portion of the interface pod **240**. In some examples, the user interface pod **240** may simply be held in a person's hand. In some examples, the user interface pod **240** may communicate wirelessly with the user interface **108** of the medical device controller **102**, for example, using a Bluetooth®, Wireless USB, ZigBee, Wireless Ethernet, GSM, or other type of communication interface.

The user interface pod **240** includes a number of buttons by which the subject, or a bystander can communicate with the medical device controller **102**, and a speaker by which the medical device controller **102** may communicate with the subject or the bystander. For example, where the medical device controller **102** determines that the subject is experiencing a cardiac arrhythmia, the medical device controller **102** may issue an audible alarm via a speaker on the medical device controller **102** or the user interface pod **240** alerting the subject and any bystanders to the subject's medical condition. The medical device controller **102** may also instruct the subject to press and hold one or more buttons on the user interface **108** of the medical device controller **102** or on the user interface pod **240** to indicate that the subject is conscious, thereby instructing the medical device controller **102** to withhold the delivery of one or more therapeutic defibrillating shocks. If the subject does not respond, the device may infer that the subject is unconscious, and proceed with the treatment sequence, culminating in the delivery of one or more defibrillating shocks to the body of the subject.

In one example, the functionality of the user interface pod **240** is integrated into the housing of the medical device controller **102**. FIGS. 3A-B illustrate such an example of the medical device controller **102**. The medical device controller **102** includes two response buttons **310** on opposing sides of the housing of the medical device controller **102**. As shown in FIGS. 3A-B, the response buttons **310** are recessed to reduce the likelihood of accidental activation (e.g., a subject falling on the response button). The medical device controller **102** also includes, in this example, a display screen **320** and a speaker to enable the communication of audible and visual stimuli to the subject. It is appreciated that the response buttons **310** do not have to be placed on opposing sides of the housing as illustrated in FIGS. 3A-B. The response buttons, for example, may be located adjacent to each other in the housing the medical device controller **102**. The adjacent placement of the response buttons may

make it easier for individuals with smaller hands or less dexterity to engage the response buttons.

Example Automated Medical Device

In some examples, the medical device **100** described above with reference to FIG. **1** is an automated external medical device (AED). AEDs are small portable defibrillators that are capable of monitoring cardiac rhythms, determining when a defibrillating shock is necessary, and administering the defibrillating shock either automatically, or under the control of a trained rescuer (e.g., an EMT or other medically training personnel). The AED, in addition, may be configured to provide counseling to an operator as to how to perform cardiac resuscitation (CPR). FIG. **4** illustrates an AED **400**. The AED **400** may be, for example, an AED Plus® automated external defibrillator available from ZOLL Medical Corporation of Chelmsford, Mass. As shown, the AED **400** includes a medical device controller **102** and an electrode assembly **402**.

The electrode assembly **402** includes one or more sensing electrodes **120** (e.g., ECG sensors), one or more acoustic sensors **122**, one or more therapy electrodes **404** (e.g., defibrillation pads), a connector **406**, wiring **408** electrically coupling the connector **406** to the one or more sensing electrodes **120**, the one or more acoustic sensors, and the one or more therapy electrodes **404**. As shown in FIG. **4**, the connector is configured to couple the electrode assembly **402** to the medical device controller **102** and, more specifically, the one or more sensing electrodes **120** and the one or more acoustic sensors **122** to the sensor interface **114** and the one or more therapy electrodes to the therapy delivery interface **116**.

The medical device controller **102** of the AED **400** is configured to detect the cardiac rhythm of the subject using ECG and heart sounds data and provide pacing and defibrillating shocks to the subject as appropriate. This process is similar to the process described with regard to medical device controller **102** of the ambulatory medical device **200**. The user interface **108** of the AED **400** may include a variety of components configured to communicate with the operator including, but not limited to, a display screen, a speaker, and one or more buttons. In this example, the AED **400** includes a display screen to display notifications to an operator. The notifications may provide instructions to the operator regarding the proper administration of CPR to the subject. The notifications on the display may be accompanied by audible alarms from the speaker to further assist the operator in administering CPR to the subject.

In another example, the medical device controller **102** (and more particularly, the cardiopulmonary function analyzer **112** of FIG. **1**) of the AED is configured to utilize processed acoustic data to guide an operator through a CPR procedure. In this example, the cardiopulmonary function analyzer **112** issues step by step instructions to the operator and analyzes processed electrode data, processed acoustic data, and processed motion data representative of heart sounds, breath sounds, and physical movement (e.g., chest compressions) to validate performance of each step by the operator. Further, where performance of a given step is inadequate (e.g., chest compressions are not deep enough), the cardiopulmonary function analyzer **112** may issue additional instructions to help the operator adequately perform the given step.

Anomaly Detection Processes

As described above, various examples implement processes through which a medical device identifies and addresses cardiopulmonary anomalies using enhanced cardiopulmonary data. FIG. **5** illustrates one such identification

process **500** that utilizes enhanced cardiopulmonary data to identify cardiopulmonary anomaly.

In act **502**, a medical device (e.g., the medical device **100** of FIG. **1**) receives acoustic, electrode, and motion signals generated from detectable characteristics of physical activity involving the subject, such as motion of the subject's body and/or the subject's cardiopulmonary function. These signals may be received via one or more electrodes (e.g., the electrode **120** of FIG. **1**), one or more acoustic sensors, (e.g., the accelerometer/acoustic sensor **122** of FIG. **1**), or one or more motion sensors (e.g., the accelerometer/acoustic sensor **122** of FIG. **1**). In act **504**, the medical device analyzes the received signals using an acoustic signal processing component (e.g., the acoustic signal processing component **124** of FIG. **1**), an electrode signal processing component (e.g., the electrode signal processing component **126** of FIG. **1**), a motion signal processing component (e.g., the motion signal processing component **128**) and a cardiopulmonary function analyzer (e.g., the cardiopulmonary function analyzer **112** of FIG. **1**).

For instance, in some examples of the act **504**, the acoustic signal processing component and/or the motion signal processing component receive one or more signals from the one or more acoustic sensors, partition the one or more signals into one or more frequency bands, and provide the partitioned signals to the cardiopulmonary function analyzer for further processing. In some examples, the frequency bands correspond to signals generated by certain subject body movements and/or motion such as chest movements during patient breathing (e.g., frequencies less than 3 hertz), chest compressions being performed on the subject (e.g., frequencies between 3 and 20 hertz), heart sounds of the subject (e.g., frequencies between 20 hertz and 150 hertz), and breath sounds of the subject (e.g., frequencies between 300 to 1200 hertz).

The frequency bands recited are approximate and may vary as needed to supply the cardiopulmonary function analyzer with frequency domain information required to infer information regarding subject activity. For example, in one implementation, an estimate of motion may be separated from an estimate of acoustic sounds by partitioning the frequency band into low frequency (e.g., less than 10 hertz may be classified as indicative of patient motion such as breathing movements and/or chest compressions) and high frequency (e.g., greater than 10 hertz, may be classified as indicative as patient acoustic sounds).

In act **506**, the cardiopulmonary function analyzer determines whether the subject's cardiopulmonary activity is normal. In some examples, the cardiopulmonary function analyzer determines that the subject's cardiopulmonary activity is normal by analyzing processed electrode data and processed acoustic data covering the same cardiac cycle. More specifically, in some examples, the cardiopulmonary function analyzer compares processed electrode data to preconfigured benchmark data for the subject. This preconfigured benchmark data may include, for example, metrics such as heart rate, QRS duration, QRS axis, PR interval, QT interval, etc. In these examples, the cardiopulmonary function analyzer also, within the act **506**, compares processed acoustic data to preconfigured benchmark data that may include, for example, EMAT, LVST, LDPT, and % EMAT. Where the processed electrode data and processed acoustic data substantially match the preconfigured benchmark data, the cardiopulmonary function analyzer determines that the subject's cardiopulmonary activity is normal.

If the subject's cardiopulmonary activity is normal, the process **500** ends. In the event of a detected anomaly, the

cardiopulmonary function analyzer attempts to identify and verify an anomaly in act 508. For instance, in some examples of the act 508, the cardiopulmonary function analyzer attempts to identify and verify VT (as distinct from SVT) as follows. First, the cardiopulmonary function analyzer detects a pattern within processed electrode data that indicates whether a subject is experiencing VT. For example, VT can manifest as high frequency ventricular pulses (e.g., more than 100 bpm, or in a range of 110-250 bpm). The processed electrode data may indicate one or more of a non-sustained VT (e.g., a VT that is less than 30 seconds duration), a sustained VT, a monomorphic VT (ventricular beats having similar or same configuration), a polymorphic VT (ventricular beats having changing configurations), or a biphasic VT (a VT having a QRS complex that changes from beat to beat). In some cases, VT can originate from an area that suffered a previous injury (e.g., an earlier myocardial infarction site). As such, the cardiopulmonary function analyzer can consider a source of the VT and correlate the information with the subject's history. To verify a VT episode detected in the processed electrode data, the cardiopulmonary function analyzer can analyze the processed acoustic data covering the same cardiac cycle as the processed electrode data. For example, VT can have a lower S1 intensity or higher S1 variability than supraventricular rhythm (2.63 ± 1.78 mV vs. 4.70 ± 5.03 mV and 0.45 ± 0.24 vs. 0.21 ± 0.11 , respectively). In this regard, the cardiopulmonary function analyzer can determine whether S1 intensity is below a value of a first configurable parameter. In some examples, the cardiopulmonary analyzer can also determine whether S1 beat-to-beat variability is above a value of a second configurable parameter. Where the S1 intensity is below the value of the first configurable parameter or the S1 variability is above the value of the second configurable parameter, the cardiopulmonary function analyzer may conclude that the detected VT episode is actual and verifies the VT episode. Otherwise, the cardiopulmonary function analyzer may conclude the detected VT episode is unverified (e.g., the anomaly may be SVT). For example, the first configuration parameter for the S1 intensity threshold can be configured to be in a range defined by 4.70 mV ± 5.03 mV. For example, the second configurable parameter for the S1 variability threshold can be in a range defined by 0.21 ± 0.11 . It should be understood that other values outside these ranges are possible, and are not to be limited to the ranges described here.

In some examples of the act 508, the cardiopulmonary function analyzer can identify PEA as follows. For example, the cardiopulmonary function analyzer analyzes processed electrode data and detects a pattern within the processed electrode data that indicates a subject is experiencing PEA. For example, the cardiopulmonary function analyzer may detect that the QRS duration of the subject is greater than the value of a configurable parameter (e.g., 180 milliseconds). To verify this detected PEA, the cardiopulmonary function analyzer can analyze processed acoustic data covering the same cardiac cycle as the processed electrode data. More specifically, the cardiopulmonary function analyzer can determine, for example, that the EMAT within the cardiac cycle is greater than the value of a configurable parameter (e.g., 150 milliseconds), or that the processed acoustic data includes no heart sounds. Where the cardiopulmonary function analyzer identifies either of these patterns in the processed acoustic data, the cardiopulmonary function analyzer can confirm the detected PEA diagnosis and verifies the PEA episode. Otherwise, the cardiopulmonary function analyzer can conclude that the detected PEA is unverified.

In some examples of the act 508, the cardiopulmonary function analyzer can identify asystole as follows. For example, the cardiopulmonary function analyzer analyzes processed electrode data and detects a pattern within the processed electrode data that indicates a subject is experiencing asystole. For example, asystole can manifest as a flatline in ECG data. To verify this detected asystole, the cardiopulmonary function analyzer can analyze processed acoustic data covering the same cardiac cycle as the processed electrode data. For example, the cardiopulmonary function analyzer can determine whether the processed acoustic data includes no heart sounds. If no heart sounds are detected, the cardiopulmonary function analyzer can confirm that the detected asystole episode is a correct diagnosis and verifies the asystole episode. Otherwise, the cardiopulmonary function analyzer can identify the detected asystole as unverified. In some implementations, the cardiopulmonary function analyzer can use other information such as, motion data (e.g., chest rise and fall), respiratory movements, or respiratory sounds to confirm the asystole diagnosis.

In some examples of the act 508, the cardiopulmonary function analyzer can identify pseudo-PEA as follows. For example, the cardiopulmonary function analyzer can analyze processed electrode data and detect a pattern within processed electrode data that indicates a subject is experiencing pseudo-PEA. For example, the cardiopulmonary function analyzer may detect that the QRS duration of the subject is within a range associated with pseudo-PEA (e.g., 120-180 milliseconds). This range may be specified by one or more configurable parameters. To verify this detected pseudo-PEA, the cardiopulmonary function analyzer can analyze processed acoustic data covering the same cardiac cycle as the processed electrode data. More specifically, the cardiopulmonary function analyzer can determine, for example, that the EMAT within the cardiac cycle is within a range specified by one or more configurable parameters (e.g., 90-150 milliseconds). Where the cardiopulmonary function analyzer identifies this pattern in the processed acoustic data, the cardiopulmonary function analyzer can confirm the detected pseudo-PEA diagnosis and verify the pseudo-PEA episode. Otherwise, the cardiopulmonary function analyzer can conclude that the detected pseudo-PEA is unverified.

In some examples of the act 508, the cardiopulmonary function analyzer can identify degenerative bradycardia/pseudo-PEA/PEA where the cardiopulmonary function analyzer detects a gradually degrading EMAT within processed acoustic data over a period of time having a duration specified by a configurable parameter. In at least one example, the period of time is a configurable parameter with a value set between 10 minutes and 1 hour. Where the EMAT degrades at a rate specified by a configurable parameter, the cardiopulmonary function analyzer verifies the degenerative bradycardia/pseudo-PEA/PEA episode.

In some examples of the act 508, the cardiopulmonary function analyzer can identify respiratory arrest as follows. For example, the cardiopulmonary function analyzer can analyze EMAT data within processed acoustic data and detect a gradually degrading EMAT over a period of time. For example, such a period of time can have a duration specified by a configurable parameter. Next the cardiopulmonary function analyzer can determine whether, within the period of time with a gradually degrading EMAT, a gradually increasing time span between the apex of chest rise and the nadir of chest fall is being exhibited by the subject. In some examples, the cardiopulmonary function analyzer

determines the time span between the apex of the chest rise and the nadir of chest fall by analyzing low frequency (e.g., 0.1 Hz) and high frequency breath sounds. Where the time span transgresses a threshold value defined by a configurable parameter (and thus indicates labored breathing), the cardiopulmonary function analyzer can confirm the detected respiratory arrest diagnosis, and thereby verify the detected respiratory arrest.

In some examples of the act **508**, the cardiopulmonary function analyzer can identify sleep apnea by analyzing the strength of the S3 sound of a subject. Where the cardiopulmonary analyzer determines that the S3 sound has strength greater than a value specified by a configurable parameter (e.g., 5), the cardiopulmonary function analyzer can confirm that sleep apnea diagnosis and an episode of sleep apnea is verified.

In some examples of the act **508**, the cardiopulmonary function analyzer can identify ECG sensing electrode falloff in some situations (e.g., when the processed electrode data indicates that the subject is undergoing asystole). For example, consider a scenario in which the cardiopulmonary function analyzer analyzes processed electrode data and detects a pattern within the processed electrode data that indicates a subject is experiencing asystole. To verify this detected asystole episode, the cardiopulmonary function analyzer can analyze the processed acoustic data from the same time period as the time period in which the cardiopulmonary function analyzer analyzes the processed electrode data. If the processed acoustic data includes normal heart sounds, the cardiopulmonary function analyzer can conclude that the detected asystole episode is unverified. The cardiopulmonary function analyzer can then alert the subject, a support person, or other external entity, about a potential electrode falloff, rather than initiate a routine relating to treating the asystole condition.

In some examples, the cardiopulmonary function analyzer can further analyze processed motion data from a same time period as a time period in which the cardiopulmonary function analyzer analyzes the processed electrode data. Where the processed motion data indicates a subject is displaying a level of activity greater than a value of a configurable parameter, the cardiopulmonary function analyzer can conclude that the detected asystole episode is actually a verified electrode falloff.

If, within the act **508**, the cardiopulmonary function analyzer identifies and verifies an anomaly, the cardiopulmonary function analyzer proceeds to act **510**. Otherwise, the cardiopulmonary function analyzer proceeds to the act **514**. In the act **514** the cardiopulmonary function analyzer stores a record of the detected abnormal cardiopulmonary activity, issues a notification indicating the abnormal activity to an external entity (e.g., user or remote system), and returns to the act **502**.

In act **510**, the cardiopulmonary function analyzer identifies a routine associated with the identified anomaly by, for example, referring to a configurable cross-reference parameter stored in a data storage (e.g., the data storage **104** of FIG. **1**) that associates each detectable anomaly with a routine to address the anomaly. Next, within act **512**, the cardiopulmonary function analyzer addresses the anomaly by executing the identified routine. These routines may address anomalies by executing a variety of processes. In one example of the act **512**, the cardiopulmonary function analyzer addresses verified VT, verified asystole, and verified PEA by executing a defibrillation sequence, culminating in the delivery of one or more defibrillating shocks to the body of the subject. In some examples of the act **512**, the

cardiopulmonary function analyzer addresses ECG sensor falloff by communicating an alarm to an external entity, such as a subject, a caregiver, a technician, or a remote computer system. In at least one example of the act **512**, the cardiopulmonary function analyzer addresses verified bradycardia and verified pseudo-PEA by executing a pacing routine. In some examples of the act **512**, the cardiopulmonary function analyzer addresses verified respiratory arrest by communicating an alarm to the subject or a caregiver. In some examples the alarm may include a mild shock or high volume audio signal to stir the subject. In some examples of the act **512**, the cardiopulmonary function analyzer addresses verified or unverified SVT by communicating an alarm to a caregiver. Examples of the types of alarms communicated in the act **512** are described in U.S. Patent Application Publication No. 2012/0293323, titled "SYSTEM AND METHOD FOR ADAPTING ALARMS IN A WEARABLE MEDICAL DEVICE," filed Mar. 23, 2012, which is hereby incorporated herein by reference in its entirety.

After execution of the act **512**, the process **500** ends. The process **500** may execute repeatedly during operation of the medical device to monitor and potentially treat the subject as needed.

Process **500** depicts one particular sequence of acts in a particular example. The acts included in this process may be performed by, or using, one or more computer systems specially configured as discussed herein. Some acts are optional and, as such, may be omitted in accord with one or more examples. Additionally, the order of acts can be altered, or other acts can be added, without departing from the scope of the systems and methods discussed herein. Furthermore, as discussed above, in at least one example, the acts are performed on a particular, specially configured machine, namely a medical device configured according to the examples disclosed herein.

Having thus described several aspects of at least one example of this disclosure, it is to be appreciated various alterations, modifications, and improvements will readily occur to those skilled in the art. Such alterations, modifications, and improvements are intended to be part of this disclosure, and are intended to be within the scope of the disclosure. Accordingly, the foregoing description and drawings are by way of example only.

The invention claimed is:

1. A medical device comprising:

at least one electrocardiogram (ECG) electrode;

at least one acoustic sensor comprising at least one accelerometer; and

at least one processor coupled with the at least one acoustic sensor and the at least one ECG electrode and configured

to receive at least one signal from the at least one accelerometer,

to partition the at least one signal into at least one acoustic signal comprising frequencies greater than approximately 10 hertz and at least one motion signal comprising frequencies less than approximately 10 hertz,

to receive at least one electrical signal from the at least one ECG electrode, and

to detect at least one cardiopulmonary anomaly using the at least one electrical signal and the at least one acoustic signal.

2. The medical device of claim 1, wherein the at least one processor is configured to detect the at least one cardiopulmonary anomaly at least in part by identifying the at least

25

one cardiopulmonary anomaly using the at least one electrical signal and verifying the at least one cardiopulmonary anomaly using the at least one acoustic signal.

3. The medical device of claim 1, further comprising a garment, wherein the at least one ECG electrode and the at least one acoustic sensor are integrated with the garment.

4. The medical device of claim 3, further comprising at least one therapy electrode coupled with the at least one processor, the at least one therapy electrode integrated with the garment.

5. The medical device of claim 1, wherein the at least one acoustic signal comprises at least one of S1, S2, S3, and S4 signals.

6. The medical device of claim 1, wherein the at least one processor is configured to detect the at least one cardiopulmonary anomaly at least in part by calculating, based on the at least one acoustic signal, at least one of electromechanical activation time (EMAT), EMAT as a percentage of cardiac cycle time, systolic dysfunction index, and left ventricular systolic time.

7. The medical device of claim 1, wherein the at least one cardiopulmonary anomaly comprises at least one of supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure.

8. The medical device of claim 1, wherein the at least one processor is configured to detect disconnection of the at least one ECG electrode from a patient's skin based on the at least one acoustic signal.

9. The medical device of claim 1, wherein the at least one processor is configured to partition the at least one signal into a first frequency band for subject motion, a second frequency band for chest compression motion, a third frequency band for heart sounds, and a fourth frequency band for breath sounds.

10. The medical device of claim 9, wherein the first frequency band comprises frequencies less than approximately 3 hertz, the second frequency band comprises frequencies between approximately 3 and 20 hertz, the third frequency band comprises frequencies between approximately 20 hertz and 150 hertz, and the fourth frequency band comprises frequencies between approximately 300 to 1200 hertz.

11. The medical device of claim 1, wherein the medical device comprises an ambulatory medical device.

12. A method of monitoring a subject using a wearable defibrillator comprising at least one electrocardiogram (ECG) electrode and at least one acoustic sensor comprising at least one accelerometer, the method comprising:

receiving at least one signal from the at least one accelerometer;

partitioning the at least one signal into at least one acoustic signal comprising frequencies greater than approximately 10 hertz and at least one motion signal comprising frequencies less than approximately 10 hertz;

receiving at least one electrical signal from the at least one ECG electrode; and

detecting at least one cardiopulmonary anomaly using the at least one electrical signal and the at least one acoustic signal.

13. The method of claim 12, wherein partitioning the at least one signal comprises partitioning the at least one signal into at least one of S1, S2, S3, and S4 signals.

14. The method of claim 12, wherein detecting the at least one cardiopulmonary anomaly comprises calculating, based on the at least one acoustic signal, at least one of electro-

26

mechanical activation time (EMAT), EMAT as a percentage of cardiac cycle time, systolic dysfunction index, and left ventricular systolic time.

15. The method of claim 12, wherein detecting the at least one cardiopulmonary anomaly comprises detecting at least one of supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure.

16. The method of claim 12, further comprising detecting the at least one cardiopulmonary anomaly using the at least one electrical signal, the at least one acoustic signal, and the at least one motion signal.

17. The method of claim 12, wherein detecting the at least one cardiopulmonary anomaly comprises identifying the at least one cardiopulmonary anomaly using the at least one electrical signal and verifying the at least one cardiopulmonary anomaly using the at least one acoustic signal.

18. A medical device comprising:

at least one electrocardiogram (ECG) electrode;

at least one acoustic sensor comprising at least one accelerometer; and

at least one processor coupled with the at least one acoustic sensor and the at least one ECG electrode and configured

to receive at least one signal from the at least one accelerometer,

to partition the at least one signal into at least one acoustic signal and at least one motion signal,

wherein to partition the at least one signal into the at least one acoustic signal and the at least one motion signal comprises to partition the at least one signal into a first frequency band comprising frequencies less than approximately 3 hertz for subject motion, a second frequency band comprising frequencies between approximately 3 hertz and 20 hertz for chest compression motion, a third frequency band comprising frequencies between approximately 20 hertz and 150 hertz for heart sounds, and a fourth frequency band comprising frequencies between approximately 300 hertz to 1200 hertz for breath sounds,

receive at least one electrical signal from the at least one ECG electrode, and

detect at least one cardiopulmonary anomaly using the at least one electrical signal and the at least one signal.

19. The medical device of claim 18, wherein the at least one processor is configured to detect the at least one cardiopulmonary anomaly at least in part by identifying the at least one cardiopulmonary anomaly using the at least one electrical signal and verifying the at least one cardiopulmonary anomaly using the at least one acoustic signal.

20. The medical device of claim 18, further comprising a garment, wherein the at least one ECG electrode and the at least one acoustic sensor are integrated with the garment.

21. The medical device of claim 20, further comprising at least one therapy electrode coupled with the at least one processor, the at least one therapy electrode integrated with the garment.

22. The medical device of claim 18, wherein the at least one acoustic signal comprises at least one of S1, S2, S3, and S4 signals.

23. The medical device of claim 22, wherein the at least one processor is configured to detect the at least one cardiopulmonary anomaly at least in part by calculating, based on the at least one acoustic signal, at least one of electro-

mechanical activation time (EMAT), EMAT as a percentage of cardiac cycle time, systolic dysfunction index, and left ventricular systolic time.

24. The medical device of claim **18**, wherein the at least one cardiopulmonary anomaly comprises at least one of 5 supraventricular tachycardia, ventricular tachycardia, pulseless electrical activity, asystole, a heart murmur, sleep apnea, and respiratory failure.

25. The medical device of claim **18**, wherein the at least one processor is configured to detect disconnection of the at 10 least one ECG electrode from a patient's skin based on the at least one acoustic signal.

* * * * *

专利名称(译)	带声学传感器的医疗设备		
公开(公告)号	US10321877	公开(公告)日	2019-06-18
申请号	US15/074020	申请日	2016-03-18
[标]申请(专利权)人(译)	卓尔医学产品公司		
申请(专利权)人(译)	ZOLL医疗公司		
当前申请(专利权)人(译)	ZOLL医疗公司		
[标]发明人	VOLPE SHANE S FRANK GREGORY R KAIB THOMAS E SZYMKIEWICZ STEVEN J FREEMAN GARY A		
发明人	VOLPE, SHANE S. FRANK, GREGORY R. KAIB, THOMAS E. SZYMKIEWICZ, STEVEN J. FREEMAN, GARY A.		
IPC分类号	A61B7/04 A61N1/39 A61B5/0205 A61B5/00 A61B5/0408 A61B7/00 A61N1/04 A61B5/08 A61N1/365 A61N1/372 A61B5/0464		
CPC分类号	A61B5/7278 A61B5/0408 A61B5/08 A61B7/003 A61B7/04 A61N1/046 A61N1/0484 A61N1/3655 A61N1/36542 A61N1/37247 A61N1/37258 A61N1/37282 A61N1/39 A61N1/3993 A61B5/0464 A61B5/4818 A61B5/6805 A61B5/6823 A61B5/6844 A61B5/721 A61B2562/0219 A61B2562/17 A61N1/3904 A61N1/39044		
优先权	62/134881 2015-03-18 US		
其他公开文献	US20160270738A1		
外部链接	Espacenet		

摘要(译)

在至少一个示例中，提供了一种医疗设备。所述医疗装置包括至少一个治疗电极，至少一个心电图（ECG）电极，至少一个声传感器，以及与所述至少一个声传感器，所述至少一个ECG电极耦合的至少一个处理器，以及至少一个治疗电极。所述至少一个处理器可以从所述至少一个声传感器接收至少一个声信号，从所述ECG电极接收至少一个电极信号，使用所述至少一个电极信号检测至少一个未验证的心肺异常，并验证所述至少一个。至少一个未验证的心肺异常，参考描述至少一个声信号的数据。

