



US010448851B2

(12) **United States Patent**
Addison et al.

(10) **Patent No.: US 10,448,851 B2**
(45) **Date of Patent: Oct. 22, 2019**

(54) **SYSTEM AND METHOD FOR
DETERMINING STROKE VOLUME OF A
PATIENT**

(71) Applicant: **COVIDIEN LP**, Mansfield, MA (US)

(72) Inventors: **Paul Stanley Addison**, Edinburgh
(GB); **James Nicholas Watson**,
Dunfermline (GB)

(73) Assignee: **Covidien LP**, Mansfield, MA (US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 253 days.

(21) Appl. No.: **14/977,301**

(22) Filed: **Dec. 21, 2015**

(65) **Prior Publication Data**

US 2016/0106322 A1 Apr. 21, 2016

Related U.S. Application Data

(62) Division of application No. 13/609,566, filed on Sep.
11, 2012, now Pat. No. 9,241,646.

(51) **Int. Cl.**

A61B 5/029 (2006.01)
A61B 5/024 (2006.01)
A61B 5/026 (2006.01)
A61B 5/00 (2006.01)
A61B 8/02 (2006.01)
A61B 5/1455 (2006.01)

(52) **U.S. Cl.**

CPC **A61B 5/029** (2013.01); **A61B 5/024**
(2013.01); **A61B 5/0261** (2013.01); **A61B**
5/02427 (2013.01); **A61B 5/742** (2013.01);
A61B 5/02416 (2013.01); **A61B 5/14551**
(2013.01); **A61B 8/02** (2013.01)

(58) **Field of Classification Search**

CPC **A61B 5/024**
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,094,308 A	6/1978	Cormier
4,282,655 A	8/1981	Tinman
4,289,141 A	9/1981	Cormier
4,450,527 A	5/1984	Sramek

(Continued)

FOREIGN PATENT DOCUMENTS

EP	0537383	4/1993
EP	0841034	5/1998

(Continued)

OTHER PUBLICATIONS

Moody et al., Derivation of Respiratory Signals from Multi-lead
ECGs (1985).

(Continued)

Primary Examiner — Carl H Layno

Assistant Examiner — Jennifer L Ghand


(74) *Attorney, Agent, or Firm* — Shumaker & Sieffert,
P.A.

(57)

ABSTRACT

A PPG system for determining a stroke volume of a patient includes a PPG sensor configured to be secured to an anatomical portion of the patient. The PPG sensor is configured to sense a physiological characteristic of the patient. The PPG system may include a monitor operatively connected to the PPG sensor. The monitor receives a PPG signal from the PPG sensor. The monitor includes a pulse trending module determining a slope transit time of an upslope of a primary peak of the PPG signal. The pulse trending module determines a stroke volume of the patient as a function of the slope transit time.

18 Claims, 6 Drawing Sheets



(56)

References Cited

U.S. PATENT DOCUMENTS

5,092,339 A	3/1992	Geddes	2005/0143665 A1	6/2005	Huiku et al.
5,178,151 A	1/1993	Sackner	2005/0177046 A1	8/2005	Mills
5,275,159 A	1/1994	Griebel	2005/0192488 A1	9/2005	Brynton et al.
5,331,960 A	7/1994	Krenzke	2005/0240037 A1	10/2005	Keenan
5,408,327 A	4/1995	Geiler	2006/0167363 A1	7/2006	Osyepka et al.
5,595,182 A	1/1997	Kriviski	2006/0184051 A1	8/2006	Hempstead
5,743,268 A	4/1998	Kabal	2006/0224053 A1	10/2006	Black
5,817,010 A	10/1998	Hibl	2006/0253007 A1	11/2006	Cheng et al.
5,833,618 A	11/1998	Caro	2006/0258921 A1	11/2006	Addison et al.
5,913,826 A	6/1999	Blank	2007/0093702 A1	4/2007	Yu
5,935,066 A	8/1999	Harris	2007/0213619 A1	9/2007	Linder
6,004,272 A	12/1999	Barry	2007/0213625 A1	9/2007	Nayak
6,045,509 A	4/2000	Caro	2007/0249949 A1	10/2007	Hadley
6,155,984 A	12/2000	Krivitski	2008/0082004 A1	4/2008	Banet
6,292,686 B1	9/2001	Chaiken	2008/0119329 A1	5/2008	Punkka
6,371,921 B1	4/2002	Caro	2008/0139958 A1	6/2008	Uemura
6,389,306 B1	5/2002	Chaiken	2008/0167540 A1	7/2008	Korhonen et al.
6,503,206 B1	1/2003	Li	2008/0183232 A1	7/2008	Voss
6,561,986 B2	5/2003	Baura et al.	2008/0255433 A1	10/2008	Prough et al.
6,616,613 B1	9/2003	Goodman	2008/0287815 A1	11/2008	Chon
6,719,705 B2	4/2004	Mills	2009/0099459 A1	4/2009	Svanberg
6,740,072 B2	5/2004	Starkweather	2009/0105556 A1	4/2009	Fricke et al.
6,754,523 B2	6/2004	Toole	2009/0149762 A1	6/2009	Yang
6,758,822 B2	7/2004	Romano	2009/0163821 A1	6/2009	Sola I Caros et al.
6,760,608 B2	7/2004	Lynn	2009/0177110 A1	7/2009	Lyden
6,776,764 B2	8/2004	Pinsky	2009/0198140 A1	8/2009	Aboy
6,816,266 B2	11/2004	Varshneya	2009/0204012 A1	8/2009	Joeken
6,819,950 B2	11/2004	Mills	2009/0240119 A1	9/2009	Schwaibold
6,869,402 B2	3/2005	Arnold	2009/0326353 A1	12/2009	Watson
6,875,176 B2	4/2005	Mourad	2009/0326388 A1	12/2009	Watson
7,022,077 B2	4/2006	Mourad	2009/0326395 A1	12/2009	Watson
7,033,320 B2	4/2006	Von Behren	2010/0016739 A1	1/2010	Shelley
7,035,679 B2	4/2006	Addison et al.	2010/0049007 A1	2/2010	Sterling
7,056,292 B2	6/2006	Hutchinson	2010/0049071 A1	2/2010	Goor
7,171,271 B2	1/2007	Koh	2010/0069761 A1	3/2010	Karst et al.
7,220,230 B2	5/2007	Roteliuk	2010/0081895 A1	4/2010	Zand
7,344,497 B2	3/2008	Kline	2010/0152547 A1	6/2010	Sterling
7,367,949 B2	5/2008	Korhonen et al.	2010/0152591 A1	6/2010	Yu
7,367,954 B2	5/2008	Starr et al.	2010/0160794 A1	6/2010	Banet
7,407,486 B2	8/2008	Chon et al.	2010/0191128 A1	7/2010	Shelley
7,452,333 B2	11/2008	Roteliuk	2010/0210924 A1	8/2010	Parthasarathy
7,462,152 B2	12/2008	Kolluri	2010/0228102 A1	9/2010	Addison et al.
7,615,011 B2	11/2009	Sugo	2010/0249542 A1	9/2010	Thijs
7,674,231 B2	3/2010	McCombie	2010/0249559 A1	9/2010	Lovejoy
7,704,209 B2	4/2010	Bennett	2010/0249612 A1	9/2010	Cohen
7,747,301 B2	6/2010	Cheng	2010/0268090 A1	10/2010	Rubinstein
7,785,263 B2	8/2010	Roteliuk	2010/0268101 A1	10/2010	Sugo
7,806,830 B2	10/2010	Bernstein	2010/0268518 A1	10/2010	Sugo
7,850,617 B2	12/2010	Goedje	2010/0298689 A1	11/2010	Wang
7,881,762 B2	2/2011	Kling	2010/0324388 A1	12/2010	Moon
7,887,502 B2	2/2011	Ross et al.	2010/0324431 A1	12/2010	Addison
7,894,869 B2	2/2011	Hoaran	2010/0324827 A1	12/2010	Addison
7,899,510 B2	3/2011	Hoaran	2011/0009754 A1	1/2011	Wenzel
7,976,472 B2	7/2011	Kiani	2011/0009755 A1	1/2011	Wenzel
8,073,516 B2	12/2011	Scharf	2011/0026784 A1	2/2011	Van Slyke
8,073,518 B2	12/2011	Chin	2011/0034813 A1	2/2011	Cohen
8,092,386 B1	1/2012	Wenzel et al.	2011/0040345 A1	2/2011	Wenzel
8,187,197 B2	5/2012	Shapira	2011/0060234 A1	3/2011	Zhou
8,211,031 B2	7/2012	Poupko	2011/0060531 A1	3/2011	Sugo
8,216,136 B2	7/2012	Addison et al.	2011/0077532 A1	3/2011	Kim
8,551,006 B2	10/2013	Karst et al.	2011/0087115 A1	4/2011	Sackner
8,740,806 B2	6/2014	Parfenova et al.	2011/0098112 A1	4/2011	LeBoeuf
8,768,438 B2	7/2014	Mestha et al.	2011/0098546 A1	4/2011	Farazi
8,814,800 B2	8/2014	Fortin et al.	2011/0105918 A1	5/2011	Fortin
8,977,348 B2	3/2015	Su et al.	2011/0144711 A1	6/2011	Bornzin et al.
9,060,745 B2	6/2015	Su et al.	2011/0172504 A1	7/2011	Wegerich
2002/0022785 A1	2/2002	Romano	2011/0196245 A1	8/2011	Poupko et al.
2002/0038078 A1	3/2002	Ito	2011/0209915 A1	9/2011	Telfort
2002/0082485 A1	6/2002	Faithfull	2011/0224564 A1	9/2011	Moon
2003/0167012 A1	9/2003	Friedman	2011/0270097 A1	11/2011	Aboy
2005/0010116 A1	1/2005	Korhonen et al.	2011/0301436 A1	12/2011	Teizeira
2005/0070774 A1	3/2005	Addison et al.	2012/0022350 A1	1/2012	Teizeira
2005/0080345 A1	4/2005	Finburgh	2012/0029320 A1	2/2012	Watson
2005/0085707 A1	4/2005	Korsten	2012/0029361 A1	2/2012	Addison
2005/0124903 A1	6/2005	Roteliuk	2012/0029363 A1	2/2012	Lund
			2012/0046557 A1	2/2012	Kiani
			2012/0053433 A1	3/2012	Chamoun
			2012/0053469 A1	3/2012	Melker
			2012/0065485 A1	3/2012	Benni

(56) References Cited

U.S. PATENT DOCUMENTS

2012/0065527	A1	3/2012	Gill
2012/0065528	A1	3/2012	Gill
2012/0078069	A1	3/2012	Melker
2012/0109018	A1	5/2012	Gertner
2012/0136261	A1	5/2012	Sethi
2012/0172723	A1	7/2012	Gertner
2012/0172732	A1	7/2012	Gertner
2012/0259235	A1	10/2012	Addison et al.

FOREIGN PATENT DOCUMENTS

EP	1443856	2/2006
EP	1769737	4/2007
EP	1884189	2/2008
EP	2281508	2/2011
EP	2047794	2/2012
EP	2217140	2/2012
WO	91/13589	9/1991
WO	94/14372	7/1994
WO	97/47236	12/1997
WO	98/41279	9/1998
WO	98/51212	11/1998
WO	99/53834	10/1999
WO	02/03076	1/2002
WO	03/000125	1/2003
WO	03/039326	5/2003
WO	03/082099	10/2003
WO	2004/071292	8/2004
WO	2004/075746	9/2004
WO	2005/055825	6/2005
WO	2006/100676	9/2006
WO	2007/109065	9/2007
WO	2008/094598	8/2008
WO	2008/144404	11/2008
WO	2008/144525	11/2008
WO	2009/009761	1/2009
WO	2009/014420	1/2009
WO	2009/101140	8/2009
WO	2010/001231	1/2010
WO	2010/045556	4/2010
WO	2010/096475	8/2010
WO	2010/111073	9/2010
WO	2010/124034	10/2010
WO	2010/146326	12/2010
WO	2010/146327	12/2010
WO	2011/047211	4/2011
WO	2011/050066	4/2011
WO	2011/051822	5/2011
WO	2011/060220	5/2011
WO	2011/077294	6/2011
WO	2011/080190	7/2011
WO	2011/080194	7/2011
WO	2011/087927	7/2011
WO	2011/089488	7/2011
WO	2012/009350	1/2012
WO	2012/014065	2/2012
WO	2012/015426	2/2012

WO	2012/027613	3/2012
WO	2012/032413	3/2012
WO	2012/032536	3/2012
WO	2012/052926	4/2012
WO	2012/075322	6/2012
WO	2012/076957	6/2012

OTHER PUBLICATIONS

Allen, "Photoplethysmography and its application in clinical physiological measurement," *Physiol. Meas.* 28 (2007).

Zaja, "Venus Oximetry," *Signa* 2007.

Hoffman, "Near-Infrared Spectrometry (NIRS) and Venous-side Monitoring of the Circulation".

Elgendi, "On the Analysis of Fingertip Photoplethysmogram Signals," *Current Cardiology Reviews*, 2012.

Gonzales et al., "A Computer Based Photoplethysmographic Vascular Analyzer Through Derivatives," *Computers in Cardiology* (2008).

Millasseau et al., "Noninvasive Assessment of the Digital Volume Pulse: Comparison with the Peripheral Pressure Pulse," *Journal of the American Heart Association* (2000).

Poon, "Non-Invasive Estimation of Cardiac Output from Finger Photoplethysmogram Based of Windkessel Model," *Bulletin of Advance Technology Research*, vol. 4, No. 6 (2010).

Yoon et al., "Non-constrained Blood Pressure Monitoring Using ECG and PPG for Personal Healthcare," (2008).

Vogel et al., "How to measure heat rate?" *Eur. J. Clin. Pharmacol* (2004) 60, 461-466.

Fox et al., "Resting Heart Rate in Cardiovascular Disease," *Journal of the American College of Cardiology* vol. 50, No. 9 (2007).

Nauman, "Why measure resting heat rate?" (2012).

Marcinkevics, et al., "The shape and dimensions of photoplethysmographic pulse waves; a measurement repeatability study," *Acta Universitatis Latviensis*, vol. 753, Biology, pp. 99-106 (2009).

Selvavaj, et al., "Monitoring of Reactive Hyperemia Using Photoplethysmographic Pulse Amplitude and Transit Time," *Journal of Clinical Monitoring and Computing* 23:315-322 (2009).

Sheinfeld, et al., "Photoacoustic thermal diffusion flowmetry," *Biomedical Optics Express* vol. 3, No. 4 (2012).

Sheinfeld, et al., "Flow dependent photothermal modulation of the photoacoustic response," *Photonics Plus Ultrasound: Imaging and Sensing* (2012).

Cannesson, et al., "Relation between respiratory variations in pulse oximetry plethysmographic waveform amplitude and arterial pulse pressure in ventilated patients," *Critical Care* (2005).

Chandler, et al., "Pulse oximeter plethysmograph variation and its relationship to the arterial waveform in mechanically ventilated childer," *J. Clin. Monit. Comput.* (2012).

Natalani, et al., "Variations in Arterial Blood Pressure and Photoplethysmography During Mechanical Ventilation," *Technology, Computing and Simulation*, vol. 103, No. 5 (2006).

Alty et al., "Predicting Arterial Stiffness from the Digital Volume Pulse Waveform," *IEEE Transactions on Biomedical Engineering*, vol. 54, No. 12 (2007).




FIG. 1A




FIG. 1B

FIG. 1C

FIG. 1D





FIG. 2



30




FIG. 4




FIG. 5




FIG. 6




FIG. 7




FIG. 8




FIG. 9

1

SYSTEM AND METHOD FOR DETERMINING STROKE VOLUME OF A PATIENT

CROSS-REFERENCE TO RELATED APPLICATIONS

This application is a divisional of U.S. application Ser. No. 13/609,566, filed Sep. 11, 2012, and issued as U.S. Pat. No. 9,241,646, entitled "System and Method for Determining Stroke Volume of a Patient."

BACKGROUND

Embodiments of the present disclosure generally relate to physiological signal processing and, more particularly, to processing physiological signals to determine a stroke volume or cardiac output of a patient.

In cardiovascular physiology, stroke volume (SV) is the volume of blood pumped from one ventricle of the heart with each beat. SV is calculated using measurements of ventricle volumes from an echocardiogram (ECG) and subtracting the volume of the blood in the ventricle at the end of a beat (end-systolic volume) from the volume of blood just prior to the beat (called end-diastolic volume). Stroke volume is an important determinant of cardiac output, which is the product of stroke volume and heart rate. Because stroke volume decreases in certain conditions and disease states, stroke volume itself correlates with cardiac function.

Systems and methods have been used to measure stroke volume. For example, photoplethysmogram (PPG) systems have been used in connection with ECG systems to aid in determining SV by measuring pulse transit times from the pulse measurement by the ECG system and the pulse measurement by the PPG system. The PPG system performs a non-invasive, optical measurement that may be used to detect changes in blood volume within tissue, such as skin, of an individual.

SUMMARY

Certain embodiments provide a PPG system for determining a stroke volume of a patient. The PPG system may include a PPG sensor configured to be secured to an anatomical portion of the patient. The PPG sensor is configured to sense a physiological characteristic of the patient. The PPG system may include a monitor operatively connected to the PPG sensor. The monitor receives a PPG signal from the PPG sensor. The monitor includes a pulse trending module configured to determine a slope transit time of an upslope of a primary peak of the PPG signal. The pulse trending module configured to determine a stroke volume of the patient as a function of the slope transit time. The PPG system may include a monitor that displays an output of the stroke volume on a display.

Optionally, the slope transit time may be calculated based upon a unit amplitude of the pulse. The slope transit time may be calculated by the pulse trending module as a temporal gradient of the PPG signal. The slope transit time may be calculated by the pulse trending module as an inverse of a gradient of the upslope of the primary peak of the PPG signal.

The PPG signal may form a PPG waveform. The pulse trending module may analyze a contour of the PPG waveform along the upslope of the primary peak to identify a gradient of the contour per unit amplitude. The pulse trending module may calculate the slope transit time as the

2

inverse of the gradient. The pulse trending module may analyze a contour of the PPG waveform along the upslope of the primary peak to identify a temporal gradient of the contour per unit amplitude. The pulse trending module may calculate the slope transit time as a function of the temporal gradient.

The pulse trending module may analyze a contour of the PPG waveform along the upslope of the primary peak to identify at least one waveform characteristic thereof. The waveform characteristic may include at least one of an amplitude (A) of the contour, a gradient (m) of the contour and a temporal gradient (k) of the contour. The pulse trending module may calculate the stroke volume based upon at least one of the ratios of $m=A/STT$, $k=1/m$ and $k=STT$.

The pulse trending module may calculate stroke volume based upon empirically determined constants. The pulse trending module may use a calibration constant to determine the stroke volume.

Certain embodiments provide a PPG system for determining a stroke volume of a patient. The PPG system may include a PPG sensor configured to be secured to an anatomical portion of the patient to sense a physiological characteristic of the patient and generate a PPG signal. The PPG signal may have a primary peak, a trailing peak and a dichrotic notch for each pulse. The PPG system may include a monitor operatively connected to the PPG sensor. The monitor receives the PPG signal from the PPG sensor. The monitor includes a pulse trending module configured to determine waveform characteristics of an upslope of the primary peak of the PPG signal. The waveform characteristics may include an amplitude, a gradient and a slope transit time of the PPG signal. The pulse trending module may be configured to determine a stroke volume as a function of the waveform characteristics of the upslope of the primary peak of the PPG signal.

Certain embodiments provide a method of determining a stroke volume of a patient from a PPG system. The method may include securing a PPG sensor to an anatomical portion of the patient, sensing a physiological characteristic of the patient with the PPG sensor and receiving a PPG signal from the sensor at a monitor. The monitor includes a pulse trending module. The method includes analyzing the PPG signal at the pulse trending module to determine a slope transit time of an upslope of a primary peak of the PPG signal. The method includes calculating a stroke volume at the pulse trending module of the patient based on the slope transit time of the upslope of the primary peak of the PPG signal.

The analyzing operation may include correlating the slope transit time of the upslope of the primary peak of the PPG signal to a slope transit time of an initial pressure wave. The analyzing operation may include analyzing a contour of the PPG waveform along the upslope of the primary peak to identify at least one waveform characteristic thereof, the waveform characteristic including at least one of an amplitude (A) of the contour, a gradient (m) of the contour and a temporal gradient (k) of the contour. The pulse trending module may calculate the slope transit time based upon at least one of the ratios of $m=A/STT$, $k=1/m$ and $k=STT$.

The correlating operation may include calculating the slope transit time of the initial pressure wave as a function of effects of reflected waves of previous pulses based on at least one of wave periods, pulse periods, pulse transit times, and shapes of the previous waveforms.

The calculating operation may include calculating the stroke volume based upon the slope transit time and empirically determined constants.

Certain embodiments provide a tangible and non-transitory computer readable medium that includes one or more sets of instructions configured to direct a computer to receive a physiological signal from a sensor secured to an anatomical portion of a patient, determine a slope transit time of an upslope of a primary peak of the PPG signal, and calculate a stroke volume of the patient based on the slope transit time of the upslope of the primary peak of the PPG signal.

Embodiments of the present disclosure allow for quick and simple determination of stroke volume through analysis of a PPG signal. The PPG signal may be obtained from a single pleth-only system. In contrast to previous systems and methods, embodiments may not require multiple PPG sensors at multiple locations on the patient to determine pulse transit times for calculating stroke volume. In contrast to previous systems and methods, embodiments may not require another type of monitoring system such as an ECG system for measuring physiological conditions of the patient for use in calculations for the stroke volume. Moreover, embodiments may be used to measure temporal components of the PPG waveform that are independent of pulse transit times, respiratory effects and the like.

Certain embodiments may include some, all, or none of the above advantages. One or more other technical advantages may be readily apparent to those skilled in the art from the figures, descriptions, and claims included herein. Moreover, while specific advantages have been enumerated above, various embodiments may include all, some, or none of the enumerated advantages.

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1A illustrates a simplified block diagram of a system configured to determine a physiological parameter of a patient, according to an embodiment.

FIG. 1B illustrates an electrocardiogram (ECG) waveform of the patient, according to an embodiment.

FIG. 1C illustrates a phonocardiogram (PCG) waveform of the patient, according to an embodiment.

FIG. 1D illustrates a photoplethysmogram (PPG) waveform of the patient, according to an embodiment.

FIG. 2 illustrates an isometric view of a PPG system, according to an embodiment.

FIG. 3 illustrates a simplified block diagram of a PPG system, according to an embodiment.

FIG. 4 is an illustrative processing system in accordance with an embodiment.

FIG. 5 illustrates a PPG signal over time, according to an embodiment.

FIG. 6 illustrates a pulse waveform of a PPG signal, according to an embodiment.

FIG. 7 illustrates a pulse waveform of a PPG signal, according to an embodiment.

FIG. 8 illustrates a flow chart of a method of determining a stroke volume of a patient, according to an embodiment.

FIG. 9 illustrates a flow chart of a method of operating a PPG system, according to an embodiment.

DETAILED DESCRIPTION

FIG. 1A illustrates a simplified block diagram of a system configured to determine a physiological parameter of a patient 102. The system 100 is configured to acquire physi-

ological signals (or biosignals) from the patient 102 and analyze the physiological signals to determine a physiological parameter, such as a stroke volume (SV) and/or a cardiac output (CO) of the patient 102. The SV is the volume of blood leaving the heart in a given contraction. The CO is the volume of blood being pumped by the heart, and is a function of SV and a heart rate (HR) of the patient 102. The physiological signals are indicative of phenomena occurring in the patient. For example, the physiological signals may describe cardiac activity in which the heart undergoes a number of cardiac cycles (e.g., heart beats). The physiological signals may be electrical, optical, and/or acoustical signals.

The system 100 may include a sensor 104 that is configured to detect one or more types of physiological signals. By way of example, the system 100 may include an electrocardiogram (ECG) system that detects electrical signals corresponding to muscle excitation of the heart. In such cases, the sensor 104 may include a plurality of electrodes that are coupled to different anatomical locations of the patient 102 (e.g., chest, wrists, and/or ankles). FIG. 1B illustrates, according to an embodiment, a representative ECG waveform 110A based on the ECG signals acquired by the electrode-sensor 104.

As another example, the system 100 may include a phonocardiogram (PCG) system that detects sounds that may be caused by the closing of heart valves. In such cases, the sensor 104 may include one or more microphones that are coupled to the patient 102. FIG. 1C illustrates, according to an embodiment, a representative PCG waveform 110B based on the PCG signals acquired by the microphone-sensor 104. In alternative embodiments, the system 100 includes an ultrasound system configured to detect heart beats from the patient 102.

In certain embodiments, the system 100 includes a photoplethysmogram (PPG) system, which can measure changes in blood volume through an anatomical portion or location (e.g., a finger). A typical example of a PPG system is a pulse oximetry system although other PPG systems exist and may be used with embodiments described herein. The PPG-sensor 104 may include a probe having one or more light sources and one or more light detectors that are coupled to the patient 102. The light source(s) provide an incident light that is scattered, absorbed, reflected, and/or transmitted by the blood. The light detector(s) detect an amount of light that may correspond to blood volume. For example, as the volume of blood increases at the anatomical location, the light is attenuated more and, as such, a smaller amount of light is detected. FIG. 1C illustrates, according to an embodiment, a representative PPG waveform 110C based on the PPG signals acquired by the PPG-sensor 104.

As shown in FIG. 1A, the system 100 may include a monitor (or computing system) 106 that includes one or more components for analyzing and/or processing the physiological signals. For example, the monitor 106 may include a pre-processing module 112, a validation module 113, a rate-determining module 114, an analysis module 115, and a graphical user interface (GUI) module 116. As used herein, a "module" may include hardware components (e.g., processor, controller), software components, or a combination thereof including any associated circuitry.

The pre-processing module 112 is configured to remove unwanted signal data (e.g., noise) from raw physiological signal data obtained from the individual 102. For example, raw PPG signals may include artifacts caused by motion of the patient relative to the light detector, instrumentation bias (e.g., bias by amplifiers used in the PPG system), powerline

interference, low amplitude PPG signals, etc. Raw physiological signals from other types of monitoring systems, such as ECG and PCG systems, may also include unwanted noise. The pre-processing module **112** is configured to remove the noise to provide clearer and/or cleaner physiological signals to the other components of the system **100**.

The validation module **113** is configured to analyze the physiological signals to identify valid heart beats and waveforms from the physiological signals. In some embodiments, the validation module **113** is part of the pre-processing module **112** or another module. The validation module **113** may analyze the physiological signals after the physiological signals have been processed. In some embodiments, the validation module **113** examines the physiological signals to identify one or more reference features in the physiological signals. For instance, a series of data points over time may provide waveforms, such as the waveforms **110A-110C**. A reference feature may be an identifiable point, segment, or characteristic of the waveform (e.g., peak, trough (or foot), notch, slope of a designated segment, threshold, etc.) that may be relied upon in analysis of the physiological signals. In many cases, a reference feature of a waveform corresponds to a known physiological activity (e.g., excitation of heart muscles, closure or opening of valves, maximum volume of blood at an anatomical location, etc.). The validation module **113** may examine the data points, or a select number of data points (e.g., a segment of the waveform), to confirm that the data points are caused by a designated event of a cardiac cycle and are not a result of noise or other unwanted event, such as when the sensor **104** is being adjusted. The data points associated with valid heart beats may then be used by a rate-determining module **114** to determine a heart rate signal. In some embodiments, the data points that are not identified as corresponding to heart beats may not be considered in subsequent analysis.

The rate-determining module **114** is configured to analyze the heart beats or, more specifically, the data points corresponding to the valid heart beats identified by the validation module **113** and determine a HR of the individual at a designated moment of time. For example, the HR may be calculated by analyzing time intervals between two or more heart beats or by analyzing portions of a waveform that corresponds to a single heart beat. By way of example only, when analyzing the physiological signals, the rate-determining module **114** may identify one or more reference features (e.g., points, segments, and/or characteristics that correspond to a waveform) that may be used to calculate HR. For example, in the ECG waveform **110A**, the rate-determining module **114** may identify an R-wave peak **118** in each heart beat. A time interval **120** between the R-wave peaks **118A**, **118B** may be determined and divided by a unit of time to calculate the HR. For example, if the time interval is 0.90 seconds between the two R-wave peaks **118A**, **118B**, then the HR is 67 beats/minute.

Corresponding to each heart beat, the PPG waveform **110C** may include a systolic peak **122**, a diastolic peak **124**, and a dichrotic notch **126** that exists therebetween. In some cases, the diastolic peak is not a peak but instead a change in slope. To determine HR, the rate-determining module **114** may identify for each heart beat a reference point that exists at a foot **128** of the wave before the systolic peak **122**. The HR may be determined in a similar manner as described above with respect to the ECG waveform by identifying a time interval **129** between the foot **128A** and the foot **128B**.

However, it should be noted that the above description is just exemplary and that many reference points and/or waveform segments may be analyzed and used in calculating a

HR of an individual. Furthermore, the physiological signals may be processed in various manners to determine a HR. For example, a first derivative or second derivative of the PPG waveform may be used to locate certain reference data points in the PPG waveform. Such reference data points may be used for determining the heart rate or for determining other physiological parameters.

As will be described in greater detail below, the analysis module **115** is configured to identify data points from the physiological signals. The signal data points may be a limited number of data points from a series of data points. For example, the signal data points may correspond to a peak data point, a trough data point, a segment of data points that correspond to a slope of the waveform, and the like. To calculate a physiological parameter, such as the SV and/or CO, the analysis module **115** may use one or more of the data points to calculate the physiological parameter.

The system **100** may also include a user interface **130** that includes a display **132**. The user interface **130** may include hardware, firmware, software, or a combination thereof that enables a user to directly or indirectly control operation of the system **100** and the various components thereof. The display **132** is configured to display one or more images, such as one or more of the waveforms **110A-110C**. The display **132** may also be configured to show a representation of the physiological parameter, for example, a number representing the SV or the CO. In some embodiments, the user interface **130** may also include one or more input devices (not shown), such as a physical keyboard, mouse, touchpad, and/or touch-sensitive display. The user interface **130** may be operatively connected to the GUI module **116** and receive instructions from the GUI module **116** to display designated images on the display **132**. The user interface **130** may also include a printer or other device for providing (e.g. printing) a report. The user interface **130** may also include an alarm or alert system.

There are many medical conditions in which SV and CO are relevant. For example, uncontrolled atrial fibrillation is a condition characterized by an abnormally rapid heart rate caused by unregulated firing of electrical pulses within the heart muscles. This rapid firing induces an elevated heart rate (known as tachycardia) while simultaneously preventing the ventricles from filling completely with blood before the next contraction. In this condition, the HR increases while SV decreases. As a result, the CO (total volume of blood pumped to the body from the heart) can decrease during atrial fibrillation, leading to a decrease in blood pressure (BP).

Detecting a change in the SV, CO, HR and/or BP can alert medical providers to potentially dangerous patient conditions. Analyzing and/or processing the physiological signals to provide a representation of SV and/or CO on a display for a care provider may be more meaningful than merely monitoring BP and HR readings. A monitoring system that tracks and provides information relating to SV and/or CO for a care provider and indicates a patient status in response to calculated SV and/or CO provides a tool in patient diagnosis and treatment. The present disclosure relates to systems and methods for detecting SV and/or CO to determine patient status, and more particularly, relates to analyzing a trending nature of a waveform to determine SV and/or CO to alert a care provider to a patient condition. For example, the present disclosure relates to systems and methods that analyze a slope transit time of a PPG signal for determining SV and/or CO.

FIG. 2 illustrates an isometric view of a PPG system **210**, according to an embodiment. The PPG system **210** may be

used as part of the system **100** (shown in FIG. 1). While the system **210** is shown and described as a PPG system **210**, the system may be various other types of physiological detection systems, such as an electrocardiogram system, a phonocardiogram system, and the like. The PPG system **210** may be a pulse oximetry system, for example. The system **210** may include a PPG sensor **212** and a PPG monitor **214**. The PPG sensor **212** may include an emitter **216** configured to emit light into tissue of a patient. For example, the emitter **216** may be configured to emit light at two or more wavelengths into the tissue of the patient. The PPG sensor **212** may also include a detector **218** that is configured to detect the emitted light from the emitter **216** that emanates from the tissue after passing through the tissue.

The system **210** may include a plurality of sensors forming a sensor array in place of the PPG sensor **212**. Each of the sensors of the sensor array may be a complementary metal oxide semiconductor (CMOS) sensor, for example. Alternatively, each sensor of the array may be a charged coupled device (CCD) sensor. In another embodiment, the sensor array may include a combination of CMOS and CCD sensors. The CCD sensor may include a photoactive region and a transmission region configured to receive and transmit, while the CMOS sensor may include an integrated circuit having an array of pixel sensors. Each pixel may include a photodetector and an active amplifier.

The emitter **216** and the detector **218** may be configured to be located at opposite sides of a digit, such as a finger or toe, in which case the light that is emanating from the tissue passes completely through the digit. The emitter **216** and the detector **218** may be arranged so that light from the emitter **216** penetrates the tissue and is reflected by the tissue into the detector **218**, such as a sensor designed to obtain pulse oximetry data.

The sensor **212** or sensor array may be operatively connected to and draw power from the monitor **214**. Optionally, the sensor **212** may be wirelessly connected to the monitor **214** and include a battery or similar power supply (not shown). The monitor **214** may be configured to calculate physiological parameters based at least in part on data received from the sensor **212** relating to light emission and detection. Alternatively, the calculations may be performed by and within the sensor **212** and the result of the oximetry reading may be passed to the monitor **214**. Additionally, the monitor **214** may include a display **220** configured to display the physiological parameters or other information about the system **210**. The monitor **214** may also include a speaker **222** configured to provide an audible sound that may be used in various other embodiments, such as for example, sounding an audible alarm in the event that physiological parameters are outside a predefined normal range.

The sensor **212**, or the sensor array, may be communicatively coupled to the monitor **214** via a cable **224**. Alternatively, a wireless transmission device (not shown) or the like may be used instead of, or in addition to, the cable **224**.

The system **210** may also include a multi-parameter workstation **226** operatively connected to the monitor **214**. The workstation **226** may be or include a computing sub-system **230**, such as standard computer hardware. The computing sub-system **230** may include one or more modules and control units, such as processing devices that may include one or more microprocessors, microcontrollers, integrated circuits, memory, such as read-only and/or random access memory, and the like. The workstation **226** may include a display **228**, such as a cathode ray tube display, a flat panel display, such as a liquid crystal display (LCD), light-emitting diode (LED) display, a plasma display, or any

other type of monitor. The computing sub-system **230** of the workstation **226** may be configured to calculate physiological parameters and to show information from the monitor **214** and from other medical monitoring devices or systems (not shown) on the display **228**. For example, the workstation **226** may be configured to display SV information, CO information, an estimate of a patient's blood oxygen saturation generated by the monitor **214** (referred to as an SpO₂ measurement), pulse rate information from the monitor **214** and blood pressure from a blood pressure monitor (not shown) on the display **228**.

The monitor **214** may be communicatively coupled to the workstation **226** via a cable **232** and/or **234** that is coupled to a sensor input port or a digital communications port, respectively and/or may communicate wirelessly with the workstation **226**. Alternatively, the monitor **214** and the workstation **226** may be integrated as part of a common device. Additionally, the monitor **214** and/or workstation **226** may be coupled to a network to enable the sharing of information with servers or other workstations. The monitor **214** may be powered by a battery or by a conventional power source such as a wall outlet.

FIG. 3 illustrates a simplified block diagram of the PPG system **210**, according to an embodiment. When the PPG system **210** is a pulse oximetry system, the emitter **216** may be configured to emit at least two wavelengths of light (for example, red and infrared) into tissue **240** of a patient. Accordingly, the emitter **216** may include a red light-emitting light source such as a red light-emitting diode (LED) **244** and an infrared light-emitting light source such as an infrared LED **246** for emitting light into the tissue **240** at the wavelengths used to calculate the patient's physiological parameters. For example, the red wavelength may be between about 600 nm and about 700 nm, and the infrared wavelength may be between about 800 nm and about 1000 nm. In embodiments where a sensor array is used in place of single sensor, each sensor may be configured to emit a single wavelength. For example, a first sensor may emit a red light while a second sensor may emit an infrared light.

As discussed above, the PPG system **210** is described in terms of a pulse oximetry system. However, the PPG system **210** may be various other types of systems. For example, the PPG system **210** may be configured to emit more or less than two wavelengths of light into the tissue **240** of the patient. Further, the PPG system **210** may be configured to emit wavelengths of light other than red and infrared into the tissue **240**. As used herein, the term "light" may refer to energy produced by radiative sources and may include one or more of ultrasound, radio, microwave, millimeter wave, infrared, visible, ultraviolet, gamma ray or X-ray electromagnetic radiation. The light may also include any wavelength within the radio, microwave, infrared, visible, ultraviolet, or X-ray spectra, and that any suitable wavelength of electromagnetic radiation may be used with the system **210**. The detector **218** may be configured to be specifically sensitive to the chosen targeted energy spectrum of the emitter **216**.

The detector **218** may be configured to detect the intensity of light at the red and infrared wavelengths. Alternatively, each sensor in the array may be configured to detect an intensity of a single wavelength. In operation, light may enter the detector **218** after passing through the tissue **240**. The detector **218** may convert the intensity of the received light into an electrical signal. The light intensity may be directly related to the absorbance and/or reflectance of light in the tissue **240**. For example, when more light at a certain wavelength is absorbed or reflected, less light of that wave-

length is received from the tissue by the detector **218**. After converting the received light to an electrical signal, the detector **218** may send the signal to the monitor **214**, which calculates physiological parameters based on the absorption of the red and infrared wavelengths in the tissue **240**.

In an embodiment, an encoder **242** may store information about the sensor **212**, such as sensor type (for example, whether the sensor is intended for placement on a forehead or digit) and the wavelengths of light emitted by the emitter **216**. The stored information may be used by the monitor **214** to select appropriate algorithms, lookup tables and/or calibration coefficients stored in the monitor **214** for calculating physiological parameters of a patient. The encoder **242** may store or otherwise contain information specific to a patient, such as, for example, the patient's age, weight, and diagnosis. The Information may allow the monitor **214** to determine, for example, patient-specific threshold ranges related to the patient's physiological parameter measurements, and to enable or disable additional physiological parameter algorithms. The encoder **242** may, for instance, be a coded resistor that stores values corresponding to the type of sensor **212** or the types of each sensor in the sensor array, the wavelengths of light emitted by emitter **216** on each sensor of the sensor array, and/or the patient's characteristics. Optionally, the encoder **242** may include a memory in which one or more of the following may be stored for communication to the monitor **214**: the type of the sensor **212**, the wavelengths of light emitted by emitter **216**, the particular wavelength each sensor in the sensor array is monitoring, a signal threshold for each sensor in the sensor array, any other suitable information, or any combination thereof.

Signals from the detector **218** and the encoder **242** may be transmitted to the monitor **214**. The monitor **214** may include a general-purpose control unit, such as a microprocessor **248** connected to an internal bus **250**. The microprocessor **248** may be configured to execute software, which may include an operating system and one or more applications, as part of performing the functions described herein. A read-only memory (ROM) **252**, a random access memory (RAM) **254**, user inputs **256**, the display **220**, and the speaker **222** may also be operatively connected to the bus **250**. The control unit and/or the microprocessor **248** may include a pulse trending module **249** that is configured to determine a trending nature, index or value of the PPG signals or waveform. In an embodiment, the pulse trending module **249** analyzes the PPG signal to determine a slope transit time of a segment of the PPG signal as a basis for determining the stroke volume of the patient. The pulse trending module **249** is configured to determine a stroke volume based on calculations, measurements or other information, data or signals received from the PPG sensor **212**.

The RAM **254** and the ROM **252** are illustrated by way of example, and not limitation. Any suitable computer-readable media may be used in the system for data storage. Computer-readable media are configured to store information that may be interpreted by the microprocessor **248**. The information may be data or may take the form of computer-executable instructions, such as software applications, that cause the microprocessor to perform certain functions and/or computer-implemented methods. The computer-readable media may include computer storage media and communication media. The computer storage media may include volatile and non-volatile media, removable and non-removable media implemented in any method or technology for storage of Information such as computer-readable instructions, data structures, program modules or other data. The computer storage media may include, but are not limited to, RAM,

ROM, EPROM, EEPROM, flash memory or other solid state memory technology, CD-ROM, DVD, or other optical storage, magnetic cassettes, magnetic tape, magnetic disk storage or other magnetic storage devices, or any other medium which may be used to store desired information and that may be accessed by components of the system.

The monitor **214** may also include a time processing unit (TPU) **258** configured to provide timing control signals to a light drive circuitry **260**, which may control when the emitter **216** is illuminated and multiplexed timing for the red LED **244** and the infrared LED **246**. The TPU **258** may also control the gating-in of signals from the detector **218** through an amplifier **262** and a switching circuit **264**. The signals are sampled at the proper time, depending upon which light source is illuminated. The received signal from the detector **218** may be passed through an amplifier **266**, a low pass filter **268**, and an analog-to-digital converter **270**. The digital data may then be stored in a queued serial module (QSM) **272** (or buffer) for later downloading to RAM **254** as QSM **272** fills up. In an embodiment, there may be multiple separate parallel paths having amplifier **266**, filter **268**, and A/D converter **270** for multiple light wavelengths or spectra received.

The microprocessor **248** may be configured to determine the patient's physiological parameters, such as SV, CO, SpO₂, pulse rate, and the like, using various algorithms and/or look-up tables based on the value(s) of the received signals and/or data corresponding to the light received by the detector **218**. The signals corresponding to information about a patient, and regarding the intensity of light emanating from the tissue **240** over time, may be transmitted from the encoder **242** to a decoder **274**. The transmitted signals may include, for example, encoded information relating to patient characteristics. The decoder **274** may translate the signals to enable the microprocessor **248** to determine the thresholds based on algorithms or look-up tables stored in the ROM **252**. The user inputs **256** may be used to enter information about the patient, such as age, weight, height, diagnosis, medications, treatments, and so forth. The display **220** may show a list of values that may generally apply to the patient, such as, for example, age ranges or medication families, which the user may select using the user inputs **256**.

As noted, the PPG system **210** may be a pulse oximetry system. A pulse oximeter is a medical device that may determine oxygen saturation of blood. The pulse oximeter may indirectly measure the oxygen saturation of a patient's blood (as opposed to measuring oxygen saturation directly by analyzing a blood sample taken from the patient) and changes in blood volume in the skin. Ancillary to the blood oxygen saturation measurement, pulse oximeters may also be used to measure the pulse rate of a patient. Pulse oximeters typically measure and display various blood flow characteristics including, but not limited to, the oxygen saturation of hemoglobin in arterial blood.

A pulse oximeter may include a light sensor, similar to the sensor **212**, that is placed at a site on a patient, typically a fingertip, toe, forehead or earlobe, or in the case of a neonate, across a foot. The pulse oximeter may pass light using a light source through blood perfused tissue and photoelectrically sense the absorption of light in the tissue. For example, the pulse oximeter may measure the intensity of light that is received at the light sensor as a function of time. A signal representing light intensity versus time or a mathematical manipulation of this signal (for example, a scaled version thereof, a log taken thereof, a scaled version of a log taken thereof, and/or the like) may be referred to as

the photoplethysmogram (PPG) signal. In addition, the term “PPG signal,” as used herein, may also refer to an absorption signal (for example, representing the amount of light absorbed by the tissue) or any suitable mathematical manipulation thereof. The light intensity or the amount of light absorbed may then be used to calculate the slope transit time of the pulse, the amount of the blood constituent (for example, oxyhemoglobin) being measured as well as the pulse rate and when each individual pulse occurs.

The light passed through the tissue is selected to be of one or more wavelengths that are absorbed by the blood in an amount representative of the amount of the blood constituent present in the blood. The amount of light passed through the tissue varies in accordance with the changing amount of blood constituent in the tissue and the related light absorption. Red and infrared wavelengths may be used because it has been observed that highly oxygenated blood will absorb relatively less red light and more infrared light than blood with lower oxygen saturation. By comparing the intensities of two wavelengths at different points in the pulse cycle, it is possible to estimate the blood oxygen saturation of hemoglobin in arterial blood.

Stroke volume may correlate with cardiac health and other physiological parameters important in patient care, such as fluid responsiveness. Fluid responsiveness relates to the volume of fluid, such as blood, in the arteries, veins, and vasculature of an individual. In general, fluid responsiveness may include a measurement of the response of stroke volume, the volume of blood passing out of the heart with each heartbeat, to venous return, the volume of blood entering the heart with each heartbeat, caused by the clinical administration of fluid into the vasculature, such as through an intravenous injection. With each heartbeat, a certain amount of blood is pumped out of the heart. The more blood that fills the heart, the more blood the heart can pump out with each heartbeat. If blood volume is too low, the heart may not fully fill with blood. Therefore, the heart may not pump out as much blood with each heartbeat. Consequently, low blood volume may lead to low blood pressure, and organs and tissues may not receive enough blood to optimally and/or properly function. Monitoring stroke volume and fluid responsiveness allows a physician to determine whether additional fluid should be provided to a patient, such as through an intravenous fluid injection. In short, fluid responsiveness represents a prediction of whether or not additional intravenous fluid may improve blood flow within a patient. Fluid responsiveness may be viewed as a response of a heart in relation to overall fluid within a patient.

Stroke volume may be monitored in, for example, critically-ill patients because fluid administration plays an important role in optimizing cardiac output and oxygen delivery to organs and tissues. However, clinicians need to balance central blood volume depletion and volume overloading. Critically-ill patients are generally at greater risk for volume depletion and severe hypotension is a common life-threatening condition in critically-ill patients. Conversely, administering too much fluid may induce life-threatening adverse effects, such as volume overload, systemic and pulmonary edema, and Increased tissue hypoxia. Therefore, obtaining reliable information and parameters that aid clinicians in fluid management decisions may help improve patient outcomes.

FIG. 4 is an illustrative processing system 400 in accordance with an embodiment. In an embodiment, an input signal generator 410 generates an input signal 416. The input signal generator 410 includes a pre-processor 420 coupled to a sensing device 418. It will be understood that the input

signal generator 410 may include any suitable signal source, signal generating data, signal generating equipment, or any combination thereof to produce the signal 416. The signal 416 may be a single signal, or may be multiple signals transmitted over a single pathway or multiple pathways.

The pre-processor 420 may apply one or more signal processing techniques to the signal generated by the sensing device 418. For example, the pre-processor 420 may apply a pre-determined transformation to the signal provided by the sensing device 418 to produce an input signal 416 that can be appropriately interpreted by the processor 412. The pre-processor 420 may also perform any of the following operations to the signal provided by the sensing device 418: reshaping the signal for transmission; multiplexing the signal; modulating the signal onto carrier signals; compressing the signal; encoding the signal; and filtering the signal.

In the embodiment of FIG. 4, the signal 416 is coupled to the processor 412. The processor 412 may be any suitable software, firmware, and/or hardware, and/or combinations thereof for processing the signal 416. For example, the processor 412 may include one or more hardware processors (e.g., integrated circuits), one or more software modules, computer-readable media such as memory, firmware, or any combination thereof. The processor 412 may, for example, be a computer or may be one or more chips (i.e., integrated circuits). The processor 412 may, for example, be configured of analog electronic components. The processor 412 may perform some or all of the calculations associated with the monitoring methods of the present disclosure. For example, the processor 412 may analyze the physiological signals, waveforms, and the like and compute pulse trending characteristics thereof to determine a stroke volume and/or cardiac output, as discussed further below. The processor 412 may also perform any suitable signal processing to filter the signal 416, such as any suitable band-pass filtering, adaptive filtering, dosed-loop filtering, and/or any other suitable filtering, and/or any combination thereof. The processor 412 may also receive input signals from additional sources (not shown). For example, the processor 412 may receive an input signal containing information about the patient or treatments provided to the patient. These additional input signals may be used by the processor 412 in any of the calculations or operations it performs in accordance with the processing system 400.

The processor 412 may be coupled to one or more memory devices (not shown) or incorporate one or more memory devices such as any suitable volatile memory device (e.g., RAM, registers, etc.), non-volatile memory device (e.g., ROM, EPROM, magnetic storage device, optical storage device, flash memory, etc.), or both. In an embodiment, the processor 412 may store physiological measurements or previously received data from the signal 416 in a memory device for later retrieval. The processor 412 may be coupled to a calibration device (not shown) that may generate or receive as input reference measurements for use in calibrating calculations.

The processor 412 is coupled to an output 414 through a patient status indicator signal 419, and may be coupled through additional signal pathways not shown. The output 414 may be any suitable output device such as, for example, one or more medical devices (e.g., a medical monitor that displays various physiological parameters, a medical alarm, or any other suitable medical device that either displays physiological parameters or uses the output of the processor 412 as an input), one or more display devices (e.g., monitor, PDA, mobile phone, any other suitable display device, or any combination thereof), one or more audio devices, one or

more memory devices (e.g., hard disk drive, flash memory, RAM, optical disk, any other suitable memory device, or any combination thereof), one or more printing devices, any other suitable output device, or any combination thereof. In an embodiment, the patient status indicator signal **419** includes at least one of an identification of a medical condition of the patient; an alert, a current stroke volume measurement; a current cardiac output measurement; a current HR measurement; a current BP measurement; another current physiological measurement; an estimated patient status; and an estimated patient outcome. In some embodiments, the patient status indicator signal **419** will be stored in a memory device or recorded in another physical form for future, further analysis.

It will be understood that the system **400** may be incorporated into the system **100** (shown in FIG. 1) and/or the system **210** (shown in FIGS. 2 and 3) in which, for example, the input signal generator **410** may be implemented as parts of the sensor **212** and/or the monitor **214** and the processor **412** may be implemented as part of the monitor **214**. In some embodiments, portions of the system **400** may be configured to be portable. For example, all or a part of the system **400** may be embedded in a small, compact object carried with or attached to the patient (e.g., a watch, other piece of jewelry, or cellular telephone). In such embodiments, a wireless transceiver (not shown) may also be included in the system **400** to enable wireless communication with other components of system **210**. As such, the system **210** may be part of a fully portable and continuous monitoring solution.

FIG. 5 illustrates a PPG signal **500** over time, according to an embodiment. The PPG signal **500** is an example of a physiological signal. However, embodiments may be used in relation to various other physiological signals, such as electrocardiogram signals, phonocardiogram signals, ultrasound signals, and the like. The PPG signal **500** may be determined, formed, and displayed as a waveform by the monitor **214** (shown in FIG. 2) that receives signal data from the PPG sensor **212** (shown in FIG. 2). For example, the monitor **214** may receive signals from the PPG sensor **212** positioned on a finger of a patient. The monitor **214** processes the received signals, and displays the resulting PPG signal **500** on the display **228** (shown in FIG. 2).

The PPG signal **500** may include a plurality of pulses **502a-502n** over a predetermined time period. The time period may be a fixed time period, or the time period may be variable. Moreover, the time period may be a rolling time period, such as a 5 second rolling timeframe.

Each pulse **502a-502n** may represent a single heartbeat and may include a pulse-transmitted or primary peak **504** separated from a pulse-reflected or trailing peak **506** by a dichrotic notch **508**. The primary peak **504** represents a pressure wave generated from the heart to the point of detection, such as in a finger where the PPG sensor **212** is positioned. The trailing peak **506** represents a pressure wave that is reflected from the location proximate where the PPG sensor **212** is positioned back toward the heart. Characteristics of the primary peak **504** and/or the trailing peak **506** may be analyzed by the pulse trending module **249** (shown in FIG. 3) to calculate the stroke volume or other physiological parameters of the patient.

As shown in FIG. 5, each pulse **502a-502n** has a particular amplitude. For example, the pulse **502a** has an amplitude **A**. The amplitudes **A** may differ with respect to one another. In general, the overall amplitude of the PPG signal **500** over time **t** may modulate. The pulse trending module **249** (shown in FIG. 3) of the monitor **214** may track and store the magnitude of the amplitude modulation of the PPG signal

500 over time **t**. Optionally, the pulse trending module **249** of the monitor **214** may track and store the magnitude of the amplitude of any number of the pulses **502a-502n** for use in determining a stroke volume, cardiac output or other physiological parameter of the patient. For example, the pulse trending module **249** of the monitor **214** may use a single pulse **502a**, analyze the single pulse to determine a trending nature of the waveform thereof, and calculate the stroke volume, cardiac output or other physiological parameter of the patient based on the single pulse **502a**. Alternatively, the pulse trending module **249** may use multiple pulses **502a-502n**, analyze the trending nature of the waveforms thereof, and calculate the stroke volume, cardiac output or other physiological parameter of the patient based upon a comparison of the amplitudes or other aspects of the waveforms of the pulses **502a-502n** to calculate the stroke volume, cardiac output or other physiological parameter of the patient. Optionally, the pulse trending module **249** of the monitor **214** may determine an average modulation of the pulses **502a-502n** over a time period **t** and use the average modulation to calculate the stroke volume, cardiac output or other physiological parameter of the patient.

As shown in FIG. 5, the frequency of the pulses **502a-502n** may vary. For example, the frequency **f₁** of the pulses over a first period of time may vary from a frequency **f₂** over a later period of time. The monitor **214** (shown in FIG. 2) may monitor and determine the frequencies **f₁** and **f₂**. The frequency variation may be based upon respiration, blood pressure, heart rate, or other factors. The pulse trending module **249** of the monitor **214** may detect a magnitude of frequency modulation over a time period **t**. The pulse trending module **249** of the monitor **214** may use the frequency of the pulses, or any other temporal element of the pulses, to analyze the trending nature of the waveforms thereof, and calculate the stroke volume, cardiac output or other physiological parameter of the patient.

FIG. 6 illustrates a pulse waveform **600** of a PPG signal, according to an embodiment. While the pulse waveform **600** is described as a component of a PPG signal, the pulse waveform **600** may be a component of various other physiological signals. The pulse waveform **600** represents a single heartbeat and may include a primary peak **602** separated from a trailing peak **604** by a dichrotic notch **606**.

The pulse waveform **600** is defined by a plurality of internal pulse components **608** of a given pulse. The internal pulse components **608** aggregate together, and may aggregate with internal pulse components of previous pulses) to define the pulse waveform **600**. The internal pulse components **608** generally include an initial pressure wave **610** generated from the heart to the point of detection, such as in a finger where the PPG sensor **212** (shown in FIG. 2) is positioned, and a series of reflected pressure waves **612**, **614** (any number of reflected waves may occur, while only two are illustrated in FIG. 6 as additional reflected waves have diminishing effects on the pulse waveform **600**). The primary peak **602** is dominated by the initial pressure wave **610** (but may be affected by reflected waves of previous pulses as evidenced by the slight increase in amplitude of the primary peak **602** as compared to the initial pressure wave **610**). The trailing peak **604** is defined mostly, if not entirely, by the reflected pressure waves **612**, **614** that are reflected from the location proximate where the PPG sensor **212** is positioned back toward the heart.

Various waveform characteristics may be measured and/or calculated from the pulse waveform **600**. The waveform characteristics may be used by the pulse trending module **249** (shown in FIG. 3) to calculate the stroke volume,

cardiac output or other physiological parameter of the patient. For example, as described in further detail below, the pulse trending module 249 may utilize a wave period, a pulse period, a transit time, an amplitude, a peak, a temporal element, a gradient, a change in any of the waveform characteristics, and the like to calculate the stroke volume, cardiac output or other physiological parameter of the patient.

An internal wave period 616 is a measure of a temporal element of the initial pressure wave 610. For example, the internal wave period 616 may be indicative of a transit time of a given pulse. Optionally, the reflected pressure waves 612, 614 may have the same wave periods as the initial wave period. The wave period 616 may be affected by other physiological conditions of the patient, such as blood pressure, heart rate, respiration, and the like. The wave period 616 is variable depending on the physiological status of the patient.

A pulse transit time (PTT) 618 is a measure of a temporal element of the pulse. For example, the PTT 618 may be a transit time of a given pulse from the heart to the location proximate where the PPG sensor 212 is positioned. The PTT 618 may be calculated by using an ECG system to detect the pulse at the heart and a PPG system to detect the pulse at the finger, and the time difference between the detection at the heart and the detection at the finger corresponds to the PPT 618. The PTT 618 may be calculated by using a dual-pleth system where two PPG sensors are attached at two different locations of the patient and measuring a differential time between detection of the pulse at the first PPG sensor and detection of the pulse at the second PPG sensor (e.g. at a finger and at an ear). The time difference between the pulse detections correspond to the PPT 618. Other methods of detecting and/or calculating the PPT 618 may be used in other embodiments. The PPT 618 may be affected by other physiological conditions of the patient, such as blood pressure, heart rate, respiration, and the like. The PTT 618 is variable depending on the physiological status of the patient.

A second pulse waveform 620 is illustrated in FIG. 6 and is represented by a dashed line along the waveform. Internal pulse components 622 of the second pulse are also illustrated by dashed lines of corresponding internal pulse components. A pulse period 624, defined by the heart rate (HR) of the patient, may be calculated by measuring the time difference between the pulses. For example, the pulse period 624 may be a measurement of the time difference from the initiation of the pulse waveform 600 to the initiation of the second pulse waveform 620. Alternatively, the pulse period 624 may be a measurement of the time difference from a peak 626 of the pulse waveform 600 to a peak 628 of the second pulse waveform 620. The pulse period 624 may be affected by other physiological conditions of the patient, such as blood pressure, heart rate, respiration, and the like. The pulse period 624 is variable depending on the physiological status of the patient.

In an embodiment, characteristics of the pulse waveform 600 may be useful in determining stroke volume, cardiac output or other physiological parameters of the patient. For example, the pulse trending module 249 may use a slope transit time 630 associated with the primary peak 602 to calculate stroke volume, cardiac output or other physiological parameters of the patient. The slope transit time 630 of the upslope of the primary peak 602 of the pulse waveform 600 corresponds to a gradient 632 of the upslope for a given amplitude 634 of the pulse waveform 600. As shown in FIG. 6, because the initial pressure wave 610 dominates the shape of the primary peak 602, particularly the upslope of the

primary peak 602, the measured gradient 632, amplitude 634 and slope transit time 630 of the primary peak 602 can be assumed to be equivalent to a gradient (m), amplitude (A) and a slope transit time (STT) of the initial pressure wave 610.

The gradient m provides a measure of the upslope of the initial pressure wave 610. The gradient m is an amplitude change per unit time and may generally be represented by:

$$m=A/STT \quad \text{Equation (1)}$$

where m is the gradient of the initial pressure wave 610, A is the amplitude of the initial pressure wave 610, and STT is the slope transit time of the initial pressure wave 610.

The inverse of the gradient m is the time change per unit amplitude and may generally be represented by:

$$k=1/m \quad \text{Equation (2)}$$

where k is the temporal gradient of the initial pressure wave 610. The temporal gradient k is proportional to STT and is not dominated by HR effects, as described in further detail below. Optionally, the temporal gradient $k=STT$. The temporal gradient k may be used by the pulse trending module 249 to calculate stroke volume. For example, SV may be a function of k (e.g., f(k)). The pulse trending module 249 may calculate SV according to the following equations:

$$SV=a(k)+b \quad \text{Equation (3)}$$

$$SV=K(a(k)+b) \quad \text{Equation (4)}$$

$$SV=a(\ln [k])+b \quad \text{Equation (5)}$$

where a and b are empirically-determined constants that may be determined through clinical examinations of patients, K is a calibration constant, and ln is the natural logarithm. The constants a, b and K may be dependent upon the nature of the subject and/or the nature of the signal detecting devices. The constants a, b and K may be computed from relationships derived from observed historical data (e.g., relationships with patient demographic data such as body mass index (BMI), height, weight, and the like) and/or measured signal characteristics (e.g., heart rate, PTT, differential PTT, and the like). The calibration constant K may be derived in whole or in part by calibration through dilution methods for obtaining stroke volume or cardiac output or other methods, after which the SV or CO may be calculated continuously using the calibrated equation of SV with respect to k.

In some embodiments, the PPG signal may be corrected or normalized to account for changes in vascular tone and/or motion artifacts through analysis of the PPG signal. Normalizing may be performed prior to calculating k.

In some embodiments, the pulse trending module 249 may calculate the gradient m, amplitude A and/or slope transit time STT of the pressure wave 610 as a function of the measured gradient 632 and slope transit time 630, such as by predicting the effects of reflected waves of previous pulses based on wave periods, pulse periods, pulse transit times, shapes of previous waveforms, and the like.

FIG. 7 illustrates a pulse waveform 650 of a PPG signal, according to an embodiment. The pulse waveform 650 is similar to the pulse waveform 600 shown in FIG. 6, however the pulse waveform 650 is a pulse waveform taken during inhalation whereas the pulse waveform 600 is a pulse waveform taken during exhalation. In FIG. 7, like reference numerals from FIG. 6 are used to identify like components.

Comparison of the waveform characteristics between FIG. 6 and FIG. 7 illustrate how the pulse waveforms have

two competing temporal elements that affect the pulse waveforms. For example, during exhalation (FIG. 6), the blood pressure is increasing but the heart rate is decreasing. During inhalation (FIG. 7), the opposite occurs where the blood pressure decreases and the heart rate increases. During exhalation (FIG. 6), as the blood pressure increases, the internal pulse waves travel faster, thus reducing PTT 618 and having shorter wave periods 616. However, the pulse period 624 lengthens as the heart rate decreases during the exhalation phase of the respiratory cycle. During inhalation (FIG. 7), as the blood pressure decreases, the internal pulse waves travel slower, thus increasing PTT 618 and having longer wave periods 616. However, the pulse period 624 shortens due to the increasing heart rate. The opposing temporal effects make analysis of the pulse waveforms difficult as the time measured between two separate fiducial points on a single pleth or pulse waveform may be pulled in competing directions.

As shown in FIGS. 6 and 7, the gradient 632 of the upslope of the primary peak 602 corresponds to the PTT 618 and wave period 616. As the PTT 618 and wave period 616 increase (e.g. during inhalation), the trend of the gradient is to decrease, whereas as the PTT 618 and wave period 616 decrease (e.g. during exhalation), the trend of the gradient is to increase. However, during both exhalation and inhalation, the gradients 632 of the upslopes associated with the primary peaks 602 closely follow the actual gradients m of the initial pressure waves 610. The measured gradients 632 are a close approximation of the actual gradients m of the pressure wave 610. Such correlation between the measured gradient 632 and the actual gradient m occurs because the initial pressure wave 610 dominates such temporal region of the resulting (aggregate) pleth cardiac pulse waveform. Measurements of the gradients 632 and STTs 630 over a fixed amplitude 634 (e.g. height change of the waveform) at a highly localized region along the upslopes of the primary peaks 602 eliminates the effect of the heart rate on the pulse waveforms 600, 650 that may influence a transit time calculated from two separated fiducial points on the waveform. The pulse trending module 249 (shown in FIG. 3) calculates the time period for the pulse waveform 600, 650 to increase a fixed amplitude, and uses such pulse trend to calculate SV, such as according to equation 3.

FIG. 8 illustrates a flow chart of a method of determining a stroke volume of a patient, according to an embodiment. The method may be performed by various systems, such as the system 100 (shown in FIG. 1), the system 210 (shown in FIGS. 2 and 3), or other capable systems. The method may include acquiring at 800 PPG signals. The PPG signals may be acquired by securing a PPG sensor to an anatomical portion of the patient and sensing a physiological characteristic of the patient with the PPG sensor. While the embodiment of the method described herein references acquiring PPG signals, such as using the PPG sensor 212 (shown in FIG. 2), the method may include acquiring other types of physiological signals, such as ECG signals, PCG signals, and/or ultrasound signals that characterize or describe cardiac activity. The physiological signals may be obtained from the individual for at least a designated period of time.

In an embodiment, the PPG signal is analyzed by the pulse trending module 249 (shown in FIG. 3). The pulse trending module 249 analyzes the PPG signal to determine waveform characteristics of the PPG signal, such as an amplitude of the PPG signal, an amplitude trending component of the PPG signal (e.g., a change in amplitude per unit time, a rate of increase of the amplitude, a rate of

decrease of the amplitude, and the like), a temporal component of the PPG signal, a temporal trending component of the PPG signal (e.g., a change in time per unit amplitude), and the like.

At 802, the PPG signal is analyzed to determine a gradient of the upslope of the primary peak of the PPG signal. The gradient may correspond to a fixed height change or amplitude of the PPG signal. Then, at 804, the system correlates the gradient of the primary peak to a gradient of the pressure wave. Because the PPG signal is an aggregation of pressure waves and reflected pressure waves, the system may determine or calculate the actual gradient of the pressure wave from the measured PPG signal. In an embodiment, because the initial pressure wave dominates the measured PPG signal at the upslope of the primary peak, the measured gradient may be assumed to be equal to the gradient of the pressure wave. In such embodiment, the system uses the measured gradient as the actual gradient. Alternatively, the system may calculate the actual gradient from the measured gradient by using a scaling factor, an algorithm, a look-up table, or by other methods. Optionally, the system may calculate the gradient as a function of the measured gradient by predicting the effects of reflected waves of previous pulses based on wave periods, pulse periods, pulse transit times, shapes of previous waveforms, and the like.

Similarly, at 806, the PPG signal is analyzed to determine a slope transit time (STT) of the upslope of the primary peak of the PPG signal. The STT may correspond to a fixed height change or amplitude of the PPG signal. Then, at 808, the system correlates the STT of the primary peak to a STT of the pressure wave. Optionally, the system may determine the STT as an inverse of the gradient. For example, because the gradient $m=A/STT$, for a unit amplitude the $STT=k=1/m$.

At 810, a stroke volume is calculated based on the STT of the pressure wave. The stroke volume is calculated as a function of STT. The stroke volume may be calculated according to any of equations 3-5. The stroke volume may be calculated with a pleth-only system. For example, the system may be operated without the need for an invasive monitoring system, an ECG or any other monitoring system. The system may calculate the stroke volume with the use of a single PPG sensor. Further, the system, at 812, may calculate the cardiac output based on the stroke volume and the heart rate. For example, the $CO=SV*HR$.

FIG. 9 illustrates a flow chart of a method of operating a PPG system, according to an embodiment. The method may be performed by various systems, such as the system 100 (shown in FIG. 1), the system 210 (shown in FIGS. 2 and 3), or other capable systems. The method may include acquiring at 820 PPG signals. In an embodiment, the PPG signal is analyzed by the pulse trending module 249 (shown in FIG. 3). The pulse trending module 249 analyzes the PPG signal to determine waveform characteristics of the PPG signal.

At 822, the PPG signal is analyzed to determine an amplitude, a gradient and/or a slope transit time of the upslope of the primary peak of the pulse. Then, at 824, a stroke volume is calculated based on the STT of the pulse. The stroke volume is calculated as a function of STT. The stroke volume may be calculated according to any of equations 3-5.

At 826, the system displays the stroke volume on a monitor, such as the monitor 214 (shown in FIG. 2). The stroke volume may be displayed as a number, a grade, a graphical representation, and the like.

At 828, the system determines if the stroke volume exceeds a threshold. The threshold may be based on physiological conditions of the patient, such as age, weight,

height, diagnosis, medications, treatments, and so forth. If the stroke volume exceeds the threshold, the system at 830 provides an alarm condition. The alarm may be a visual alarm, an audible alarm, or another type of alarm. The alarm may be triggered on the monitor 214 and/or may be transmitted to another location, such as a central monitoring station to alert medical professionals.

Various embodiments described herein provide a tangible and non-transitory (for example, not an electric signal) machine-readable medium or media having instructions recorded thereon for a processor or computer to operate a system to perform one or more embodiments of methods described herein. The medium or media may be any type of CD-ROM, DVD, floppy disk, hard disk, optical disk, flash RAM drive, or other type of computer-readable medium or a combination thereof.

The various embodiments and/or components, for example, the control units, modules, or components and controllers therein, also may be implemented as part of one or more computers or processors. The computer or processor may include a computing device, an input device, a display unit and an interface, for example, for accessing the Internet. The computer or processor may include a microprocessor. The microprocessor may be connected to a communication bus. The computer or processor may also include a memory. The memory may include Random Access Memory (RAM) and Read Only Memory (ROM). The computer or processor further may include a storage device, which may be a hard disk drive or a removable storage drive such as a floppy disk drive, optical disk drive, and the like. The storage device may also be other similar means for loading computer programs or other instructions into the computer or processor.

As used herein, the term “computer,” “computing system,” or “module” may include any processor-based or microprocessor-based system including systems using microcontrollers, reduced instruction set computers (RISC), application specific integrated circuits (ASICs), logic circuits, and any other circuit or processor capable of executing the functions described herein. The above examples are exemplary only, and are thus not intended to limit in any way the definition and/or meaning of the term “computer” or “computing system.”

The computer, computing system, or processor executes a set of instructions that are stored in one or more storage elements, in order to process input data. The storage elements may also store data or other information as desired or needed. The storage element may be in the form of an information source or a physical memory element within a processing machine.

The set of instructions may include various commands that instruct the computer or processor as a processing machine to perform specific operations such as the methods and processes of the various embodiments of the subject matter described herein. The set of instructions may be in the form of a software program. The software may be in various forms such as system software or application software. Further, the software may be in the form of a collection of separate programs or modules, a program module within a larger program or a portion of a program module. The software also may include modular programming in the form of object-oriented programming. The processing of input data by the processing machine may be in response to user commands, or in response to results of previous processing, or in response to a request made by another processing machine.

As used herein, the terms “software” and “firmware” are interchangeable, and include any computer program stored in memory for execution by a computer, including RAM memory, ROM memory, EPROM memory, EEPROM memory, and non-volatile RAM (NVRAM) memory. The above memory types are exemplary only, and are thus not limiting as to the types of memory usable for storage of a computer program.

As discussed above, embodiments may provide a system and method of determining a stroke volume of a patient through analysis of physiological signals, such as PPG signals, by analyzing waveform characteristics of the PPG signal and calculating a slope transit time of the PPG signal. The slope transit time is less influenced by momentary rises in the HR and is a good measurement tool for analyzing the PPG signal to determine stroke volume.

Certain embodiments of the present disclosure may include some, all, or none of the above advantages. One or more other technical advantages may be readily apparent to those skilled in the art from the figures, descriptions, and claims included herein. Moreover, while specific advantages have been enumerated above, various embodiments may include all, some, or none of the enumerated advantages.

It is to be understood that the above description is intended to be illustrative, and not restrictive. For example, the above-described embodiments (and/or aspects thereof) may be used in combination with each other. In addition, many modifications may be made to adapt a particular situation or material to the teachings without departing from its scope. While the dimensions, types of materials, and the like described herein are intended to define the parameters of the disclosure, they are by no means limiting and are exemplary embodiments. Many other embodiments will be apparent to those of skill in the art upon reviewing the above description. The scope of the disclosure should, therefore, be determined with reference to the appended claims, along with the full scope of equivalents to which such claims are entitled. In the appended claims, the terms “including” and “in which” are used as the plain-English equivalents of the respective terms “comprising” and “wherein.” Moreover, in the following claims, the terms “first,” “second,” and “third,” etc. are used merely as labels, and are not intended to impose numerical requirements on their objects. Further, the limitations of the following claims are not written in means-plus-function format and are not intended to be interpreted based on 35 U.S.C. § 112, sixth paragraph, unless and until such claim limitations expressly use the phrase “means for” followed by a statement of function void of further structure.

What is claimed is:

1. A method, comprising:

receiving a photoplethysmogram (PPG) signal from a single PPG sensor at a PPG monitor comprising a processor, wherein the PPG sensor is configured to be secured to a skin surface of a patient, and wherein the PPG sensor comprises one or more light sources configured to emit light into a tissue and one or more detectors configured to detect an amount of the light from the tissue, and wherein the PPG signal comprises a primary peak separated from a trailing peak by a dichrotic notch;

analyzing the PPG signal via the processor to determine an inverse gradient of an upslope of the primary peak of the PPG signal; and

calculating a stroke volume via the processor based on the inverse gradient of the upslope of the primary peak of the PPG signal.

21

2. The method of claim 1, wherein the PPG signal comprises a PPG waveform, and wherein analyzing the PPG signal comprises analyzing a contour of the PPG waveform along the upslope of the primary peak to identify the inverse gradient.

3. The method of claim 1, wherein analyzing the PPG signal comprises correlating the inverse gradient of the upslope of the primary peak of the PPG signal to a slope transit time of an initial pressure wave.

4. The method of claim 3, wherein correlating the inverse gradient of the upslope to a slope transit time comprises calculating the inverse gradient of the PPG signal as a function of effects of reflected waves of previous pulses.

5. The method of claim 1, wherein calculating the stroke volume comprises calculating the stroke volume based upon one or more empirically determined constants.

6. The method of claim 5, wherein the one or more empirically determined constants are determined through clinical examinations of the patient.

7. The method of claim 5, wherein the one or more empirically determined constants are derived from historical data related to the patient, wherein the historical data comprises patient demographic data, a body mass index, a height, a weight, or a combination thereof.

8. The method of claim 5, wherein the one or more empirically determined constants are derived from measured characteristics of the PPG signal, wherein the measured characteristics comprise a heart rate, a pulse transit time, a differential pulse transit time, or a combination thereof.

9. The method of claim 1, comprising calculating a heart rate based on a frequency of or a time interval between one or more primary peaks in the PPG signal.

10. The method of claim 9, comprising calculating a cardiac output based on the calculated stroke volume and the heart rate.

11. The method of claim 10, comprising displaying the calculated stroke volume, the heart rate, or the cardiac output on a display of the PPG monitor.

12. The method of claim 1, comprising triggering an alarm when the stroke volume exceeds a threshold, wherein the threshold is based on physiological conditions of the patient, and wherein the physiological conditions comprise an age, a weight, a height, a diagnosis, medications, treatments, or a combination thereof.

22

13. The method of claim 1, comprising:

measuring a gradient of the upslope of the primary peak, wherein the gradient is a fixed height change or an amplitude of the PPG signal; and

determining an actual gradient of the upslope of the primary peak from the measured gradient by using a scaling factor, an algorithm, a lookup table, or a combination thereof.

14. The method of claim 13, wherein the inverse gradient is the inverse of the measured gradient or the actual gradient.

15. A method, comprising:

receiving a photoplethysmogram (PPG) signal from a single PPG sensor at a PPG monitor comprising a processor, wherein the PPG sensor is configured to be secured to a skin surface of a patient, and wherein the PPG sensor comprises one or more light sources configured to emit light into a tissue and one or more detectors configured to detect an amount of the light from the tissue, and wherein the PPG signal comprises one or more primary peaks;

calculating a heart rate based on a frequency of or a time interval between the one or more primary peaks in the PPG signal;

determining an inverse gradient of an upstroke of a primary peak of the PPG signal;

calculating a stroke volume based at least in part on the inverse gradient;

calculating a cardiac output based at least in part on the stroke volume and the heart rate; and

displaying the cardiac output, the stroke volume, the heart rate, or a combination thereof, on a display.

16. The method of claim 15, wherein the PPG signal comprises a PPG waveform, and wherein analyzing the PPG signal comprises analyzing a contour of the PPG waveform along the upslope of the primary peak to identify the inverse gradient.

17. The method of claim 15, comprising calculating the stroke volume based at least in part on the inverse gradient of the PPG signal and one or more empirical constants.

18. The method of claim 17, wherein the one or more empirically determined constants comprise a calibration constant or the one or more empirically determined constants are derived from clinical examinations of the patient.

* * * * *

专利名称(译)	用于确定患者的中风量的系统和方法		
公开(公告)号	US10448851	公开(公告)日	2019-10-22
申请号	US14/977301	申请日	2015-12-21
[标]申请(专利权)人(译)	柯惠有限合伙公司		
申请(专利权)人(译)	COVIDIEN LP		
当前申请(专利权)人(译)	COVIDIEN LP		
[标]发明人	ADDISON PAUL STANLEY WATSON JAMES NICHOLAS		
发明人	ADDISON, PAUL STANLEY WATSON, JAMES NICHOLAS		
IPC分类号	A61B5/029 A61B5/024 A61B5/026 A61B5/00 A61B8/02 A61B5/1455		
CPC分类号	A61B5/029 A61B5/0261 A61B5/742 A61B5/02427 A61B5/024 A61B8/02 A61B5/02416 A61B5/14551		
其他公开文献	US20160106322A1		
外部链接	Espacenet		

摘要(译)

用于确定患者的中风量的PPG系统包括被配置为固定至患者的解剖部位的PPG传感器。PPG传感器被配置为感测患者的生理特征。PPG系统可以包括可操作地连接到PPG传感器的监视器。监视器从PPG传感器接收PPG信号。监视器包括脉冲趋势模块，该脉冲趋势模块确定PPG信号的主峰的上坡的斜率渡越时间。脉冲趋势模块根据斜率通过时间确定患者的中风量。

