



US009247991B2

(12) **United States Patent**
Norris et al.

(10) **Patent No.:** **US 9,247,991 B2**
(45) **Date of Patent:** **Feb. 2, 2016**

(54) **TEMPERATURE PROFILE MAPPING AND
GUIDED THERMOTHERAPY**

(71) Applicant: **Tomophase Corporation**, Burlington,
MA (US)

(72) Inventors: **Peter E. Norris**, Cambridge, MA (US);
Xiao-Li Li, Lexington, MA (US);
Andrey Vertikov, Westwood, MA (US)

(73) Assignee: **TOMOPHASE, INC.**, Burlington, MA
(US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 0 days.

(21) Appl. No.: **13/903,983**

(22) Filed: **May 28, 2013**

(65) **Prior Publication Data**

US 2013/0261613 A1 Oct. 3, 2013

Related U.S. Application Data

(62) Division of application No. 12/920,309, filed as
application No. PCT/US2009/035773 on Mar. 2,
2009, now Pat. No. 8,452,383.

(60) Provisional application No. 61/032,853, filed on Feb.
29, 2008.

(51) **Int. Cl.**

A61B 6/00 (2006.01)

A61B 18/18 (2006.01)

A61B 1/00 (2006.01)

A61B 5/00 (2006.01)

A61B 5/01 (2006.01)

A61B 17/32 (2006.01)

A61B 18/20 (2006.01)

(52) **U.S. Cl.**

CPC **A61B 18/18** (2013.01); **A61B 1/00096**
(2013.01); **A61B 1/00172** (2013.01); **A61B**
5/0066 (2013.01); **A61B 5/0084** (2013.01);
A61B 5/015 (2013.01); **A61B 5/6852**
(2013.01); **A61B 17/320068** (2013.01); **A61B**
18/20 (2013.01)

(58) **Field of Classification Search**

CPC A61B 18/24; A61B 18/12; A61B 18/14;
A61B 2018/00005; A61B 17/320068; A61B
18/18; A61B 18/20; A61B 1/00096; A61B
1/00172; A61B 5/0066; A61B 5/0084; A61B
5/015; A61B 5/6852; A61M 2025/0026

USPC 600/473–480, 549; 606/1–19, 32, 34,
606/35, 41, 45, 49–50; 607/98–101
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,140,364	A *	2/1979	Yamashita		A61B 1/00165
					359/367
5,150,254	A *	9/1992	Saitou		A61B 1/018
					348/65
6,626,902	B1 *	9/2003	Kucharczyk		A61B 5/14503
					606/34
7,004,911	B1 *	2/2006	Tu		A61B 5/015
					600/549
7,153,299	B1 *	12/2006	Tu		A61N 5/0601
					128/898
7,344,528	B1 *	3/2008	Tu		A61B 18/245
					606/1
7,976,537	B2 *	7/2011	Lieber		A61B 18/1492
					606/13
8,467,858	B2 *	6/2013	Vertikov		A61B 5/14532
					600/471
2002/0038074	A1 *	3/2002	Hakamata		A61B 1/00186
					600/178
2005/0075704	A1 *	4/2005	Tu		A61B 18/245
					607/88
2005/0143664	A1 *	6/2005	Chen		A61B 5/6852
					600/478
2005/0234347	A1 *	10/2005	Yamataka		A61B 1/0008
					600/476
2005/0288582	A1 *	12/2005	Yu		A61B 5/0066
					600/437
2009/0005771	A1 *	1/2009	Lieber		A61B 18/1492
					606/34
2009/0137997	A1 *	5/2009	Zhang		A61B 18/24
					606/11
2011/0029049	A1 *	2/2011	Vertikov		A61B 5/14532
					607/104
2013/0261613	A1 *	10/2013	Norris		A61B 1/00096
					606/14
2013/0282083	A1 *	10/2013	Vertikov		A61B 5/14532
					607/100

* cited by examiner

Primary Examiner — Sanjay Cattungal

(57) **ABSTRACT**

Techniques, apparatus and systems that use an optical probe
head to deliver light to a target and to collect light from the
target for imaging and monitoring a target while a separate
radiation is applied to treat the target.

26 Claims, 10 Drawing Sheets




FIG. 1A




FIG. 1B




FIG. 2





FIG. 3



FIG. 4

FIG. 5






FIG. 7





FIG. 8



FIG. 9

FIG. 10

TEMPERATURE PROFILE MAPPING AND GUIDED THERMOTHERAPY

CROSS REFERENCE TO RELATED APPLICATIONS

This patent document is a divisional of U.S. patent application Ser. No. 12/920,309, filed on Nov. 29, 2010, which is a 371 of International Patent Application No. PCT/US2009/035773, filed on Mar. 2, 2009, which claims the benefit of priority of U.S. Provisional Patent Application No. 61/032,853, filed on Feb. 29, 2008. The entire contents of the before-mentioned patent applications are incorporated by reference as part of the disclosure of this document.

BACKGROUND

This application relates to techniques, apparatus and systems that use optical waveguides to deliver light to a targeted area for optical detection of tissues, organs and other objects in medical, biological and other applications.

Light can be guided through a light pipe or optical waveguide such as optic fiber to a target to obtain optical images, optical measurements and other operations of the target. The optical waveguide such as optic fiber can be used to reach the target at a location that is otherwise difficult to reach or requires some preparatory procedures to make the target more accessible. For example, the tissue of an internal organ of a patient may be made available for a medical examination or therapy procedure through a natural orifice or an incision to expose the internal organ. Such a procedure may be performed by delivering probe light to the tissue via an endoscope instrument or catheter to reduce or minimize the degree of invasiveness of a diagnostic or therapeutic procedure. At the distal end of the instrument, light is pointed to certain direction or steered to interact with an area or a slice of tissue of interest.

Various devices and techniques based on optical coherence domain reflectometry (OCDR) may be used for non-invasive optical probing of various substances, including but not limited to skins, body tissues and organs of humans and animals, to provide tomographic measurements of these substances. In many OCDR systems, the light from a light source is split into a sampling beam and a reference beam which propagate in two separate optical paths, respectively. The light source may be partially coherent source. The sampling beam is directed along its own optical path to impinge on the substances under study, or sample, while the reference beam is directed in a separate path towards a reference surface. The beams reflected from the sample and from the reference surface are then brought to overlap with each other to optically interfere. Because of the wavelength-dependent phase delay, the optical interference results in no observable interference fringes unless the two optical path lengths of the sampling and reference beams are similar. This provides a physical mechanism for ranging. A beam splitter may be used to split the light from the light source and to combine the reflected sampling beam and the reflected reference beam for detection at an optical detector. The application of OCDR in medical diagnoses in certain optical configurations has come to be known as "optical coherence tomography" (OCT).

SUMMARY

This application includes implementations and examples of techniques, apparatus and systems that use an optical probe head in an endoscope device to optically measure a target for

various applications, including techniques, apparatus and systems that use an optical probe head to deliver light to a target and to collect light from the target for imaging and monitoring a target while a separate radiation is applied to treat the target.

In one aspect, an implementation of an endoscope device includes an optical catheter comprising an optical fiber to guide an optical imaging beam and an optical probe head, located at a distal end of the fiber to direct the optical imaging beam from the fiber to a target and to receive light returned from the target under illumination of the optical imaging beam; a detection module to process the light returned from the target under illumination of the optical imaging beam to measure a temperature at a location illuminated by the optical imaging beam; an energy source that produces energy that is applied to the target to raise a temperature at a location of the target where the energy is applied; and a control mechanism that controls the energy source to set a power level of the energy produced by the energy source based on the measured temperature from the detection module.

In another aspect, an implementation of a method for guiding the application of a thermotherapeutic radiation to a target tissue uses an optical catheter comprising an optical fiber to guide an optical imaging beam to direct the optical imaging beam from the fiber to a target and to receive light returned from the target under illumination of the optical imaging beam, processes the light returned from the target tissue under illumination of the optical imaging beam to measure a temperature at a location illuminated by the optical imaging beam; applies a thermotherapeutic energy to the target tissue to raise a temperature at a location of the target tissue where the energy is applied; and controls an amount of the thermotherapeutic energy applied to the target tissue, based on the measured temperature by using the optical imaging beam, to control the temperature at the target tissue between a low limit above which a thermotherapeutic effect is present and a high limit above which a damage to the target tissue occurs.

In yet another aspect, an implementation of an endoscope device for providing guided thermotherapy includes an endoscope tube comprising a hollow working channel and an optical catheter comprising an optical fiber located inside the hollow working channel to guide an optical probe beam and, an optical probe head, located at a distal end of the fiber, to reflect a first portion of the optical probe beam back to the fiber and to direct a section portion of the optical probe beam to a target tissue as an optical imaging beam. The optical probe head receives light returned from the target tissue under illumination of the optical imaging beam to overlap the light returned from the target tissue with the first portion to co-propagate in the fiber away from the optical probe head. This device includes an optical delay device coupled to the fiber to receive the first portion and the light returned from the target tissue to produce a variable relative phase delay between the first portion and the light returned from the target tissue; an optical detector that detects the light of the first portion and the light returned from the target tissue from the optical delay device; a processing unit to receive output from the optical detector and to extract a temperature at a location illuminated by the optical imaging beam from information of the target carried by the light returned from the target tissue; an RF applicator engaged to the endoscope tubing and near the optical probe head to apply RF energy to the target tissue to raise a temperature at a location of the target where the RF energy is applied; and a control mechanism that controls an amount of the RF energy to be applied by the RF applicator to the target tissue based on the measured temperature.

These and other aspects of various techniques, apparatus and systems are described in detail in the drawings, the description and the claims.

BRIEF DESCRIPTION OF TEE DRAWINGS

FIGS. 1A and 1B show an example of an endoscope for applying an optical imaging beam and RF radiation to a target in various applications, including using as a bronchoscope bronchial thermoplasty.

FIG. 2 shows an example of an endoscope for applying an optical imaging beam and RF radiation to a target with a cooling mechanism.

FIG. 3 shows one operation of the endoscope in FIGS. 1A and 1B or in FIG. 2.

FIGS. 4-8 show examples of an optical catheter for directing the optical imaging beam to the target in the endoscope in FIGS. 1A and 1B or in FIG. 2 and associated optical detection and processing, where the optical probe head reflects a first portion of an optical probe beam back to the fiber and to direct a section portion of the optical probe beam to a target tissue as the optical imaging beam and further receives light returned from the target tissue under illumination of the optical imaging beam to overlap the light returned from the target tissue with the first portion to co-propagate in the fiber away from the optical probe head.

FIG. 9 shows measurements of water density and water temperature.

FIG. 10 shows measurements of refractive index of water and its temperature.

DETAILED DESCRIPTION

This application describes examples of techniques, apparatus and systems that use optical imaging to achieve temperature profile mapping in various applications, including a thermotherapy process. One example of thermotherapy processes is radio frequency (RF) Ablation (RFA) that uses RF energy to destroy malignant cells via thermal treatment and heating. Thermotherapy depends on achieving a proper temperature range to achieve desired cell necrosis efficacy. If a malignancy is being treated using RFA such as treating liver cancer, RF power levels sufficient to produce complete cell necrosis must be employed. If the power level is too low, incomplete necrosis will occur increasing the likelihood of recurrence. If power levels are too high, extensive charring of the tissues result, impairing recovery. Mapping of the tissue temperature profile during thermotherapy is therefore very important to achieve optimum clinical outcomes.

Endoscope devices described here use light-based tomographic imaging and temperature-induced optical property changes to provide temperature profile mapping of a target, such as tissue, an organ or other object.

For temperatures above normal body temperature, the mass density of water (in grams/cc) decreases as temperature increases. Although the density change is not large, sufficiently sensitive instruments and sufficiently sophisticated interpretative techniques can detect and interpret mass density changes in the water content of soft tissues. When heat is then applied to the tumor to elevate the local tissue temperature to approach 42° C., the process of cell death begins. As the tissue temperature increases the time required to assure cell death decreases. At about 58° C. proteins begin to denature and at higher temperatures tissue coagulates. The denaturation and coagulation of tissue contained within approximately the same volume, produces density changes. As heat is transported from the central coagulation zone and conducted

to surrounding tissues, an expanding coagulation zone develops and increases as a function of thermal power delivered over time. The primary treatment for both malignant and benign tumors is surgery. While surgery has proven to be effective in cancer treatment, it is expensive and invasive, frequently requiring lengthy hospital stays for patient recovery. A critical measure of surgical success is the complete removal of the tumor with surgical margins testing negative to tumor cells upon pathology review.

With the number of cancer cases steadily increasing throughout the world, a less invasive, less costly way of treating primary and metastatic tumors is desired. Modern diagnostic imaging modalities and new interventional methods have set the stage for bringing less invasive methods to the field of oncology treatment much as imaging has enabled the diagnosis and minimally invasive treatment of coronary arteries with balloon catheters and drug eluting stents.

The devices and techniques described in this document combines real time imaging with interventional access to solid tumors offers the possibility to non surgically treat, or ablate, the tumor mass. Improved minimally invasive methods are being developed to treat both primary and metastatic solid tumors. The present devices and techniques can be used to create thermal injury to tissue when used in conjunction with a percutaneous approach to the lesion. The energy sources used of creating the desired thermal injury can vary depending on implementations and may use RF (radiofrequency), laser, microwave energy or high intensity focused ultrasound.

While thermal ablative techniques such as radiofrequency ablation or laser ablation are rapidly gaining acceptance in the treatment of inoperable tumors, incomplete treatments are common since there is no reliable method to monitor the treatment zone during the ablation. A treatment that does not encompass the entire tumor will result in recurrent growth of the tumor, usually within one or two years. Ablative techniques will be held back from full adoption as long as the treatment zone cannot be monitored during the ablation.

The present optical imaging devices and techniques provide optical monitoring and feedback control for thermal therapy. Various aspects of the such devices and techniques can be found in PCT Application No. PCT/US2005/37730 (PCT Publication No. WO 2006/045013) entitled "Integrated Disease Diagnosis and Treatment System" which is incorporated by reference as part of the disclosure of this document. Such optical imaging systems can be used for mapping of spatially non-uniform thermal distributions and can have special utility in thermotherapies where microstructural imaging plays an important role. As an example, bronchial wall microstructure is very important in a new, non-drug asthma thermotherapy called Bronchial Thermoplasty while, in various CT-guided RFA therapy, low resolution CT imaging is performed at a low rate of CT imaging and thus is inadequate for real time monitoring of RFA therapy. The present optical imaging devices provide fast imaging of the target tissue under treatment, e.g., 30 images/sec, and thus can be used to measure the thermal effect caused by the RF radiation during the RF treatment.

Bronchial Thermoplasty [BT] is a non-drug asthma therapy in clinical trials and has been developed over the past seven years as a means of reducing the severity of asthmatic exacerbations in patients whose asthma is not well controlled by cortico-steroid or anti-inflammatory inhalers. BT uses RF energy, applied to the airway smooth muscle [ASM], which surrounds the bronchus in a somewhat random pattern, to reduce the ability of the ASM to contract and obstruct airflow.

The RF catheter is inserted into the working channel of a bronchoscope and positioned in a secondary/tertiary bronchus, but since the ASM is not visible from within the bronchus, there is no way to position the RF electrode in close proximity to the ASM. As a consequence, RF energy is applied periodically, as the catheter is withdrawn from the bronchus, and, on the average, damages some of the ASM to some degree. An interesting statistic is that ASM covers, on the average, only 2.5-4.5% of the surface of a major airway. The procedure does provide relief, but in all likelihood, could be more effective if microstructural cross-sectional imaging allowed the physician to accurately position the RF electrode directly above the ASM. In addition, some patients experience adverse events [including hospitalization] after BT. Since the same treatment parameters are used for each patient, it seems reasonable to conclude that differences in bronchial wall microanatomy may be responsible for these adverse events.

The optical imaging systems described in this document can be used to play the role of a companion diagnostic for BT as well as providing procedural imaging for guidance and control.

FIGS. 1A and 1B below show an example endoscope apparatus that combines the optical imaging for thermal mapping and RFA mechanism. This endoscope device includes an optical catheter 4 comprising an optical fiber 18 to guide an optical imaging beam 14 and an optical probe head 1, located at a distal end of the fiber 18 to direct the optical imaging beam 14 from the fiber 18 to a target and to receive light returned from the target under illumination of the optical imaging beam 14. The optical probe head 1 as illustrated in this example includes a prism reflector 12 and the distal end facet of the fiber 18 and may be implemented in other configurations. The optical probe head 1 is fixed to the distal end of the fiber 18. The fiber 18 is attached to a torque cable that can rotate the fiber 18 and thus the optical probe head 1 to allow the optical imaging beam 14 from the optical probe head 1 to scan around at different positions. In addition, the torque cable can be used to push or pull the optical probe head 1 to different positions along the longitudinal direction of the fiber 18. Therefore, under this design, the optical probe head 1 can be used to sense images at different locations along the longitudinal direction of the fiber 18 and, at each location, the optical probe head 1 can be rotated to sense the surrounding areas at different directions.

In this device, a detection module is provided to process the light returned from the target under illumination of the optical imaging beam 14 to measure a temperature at a location illuminated by the optical imaging beam 14. An energy source is provided to produce energy that is applied to the target to raise a temperature at a location of the target where the energy is applied. In this example, the energy source is an RF source and an RF applicator 6 is provided to direct the RF energy to the target. A control mechanism is also provided to control the energy source to set a power level of the energy produced by the energy source based on the measured temperature from the detection module.

An endoscope tube or sheath 2 with a hollow working channel is provided to house the fiber 18 and to hold the RF applicator 6 and an RF transmission path 8 that connects the RF applicator 6 and the RF source. The RF transmission path 8 can be an electrically conductive tube in some implementations. When implemented as a bronchoscope, the endoscope tube 2 may include other components such as an imaging camera and a biopsy module. The RF applicator 6 can be an RF conductive wires in form of a basket and is attached to an outer sheath 10 outside the optical fiber 18 in the optical

catheter 14 in a way that allows the basket to expand when the outer sheath 10 is pulled in the endoscope tube or sheath 2. It is a sliding fit so that the optical imaging catheter 4 can rotate within the endoscope sheath 2 to form the image. As illustrated, an applicator mount 7 is provided to engage the RF applicator 6 to the outer sheath 10. The optical catheter 4 also may be moved in a longitudinal manner, to image proximally, distally or exactly at the point where the basket struts contact the airway wall tissue. The configuration is such that the coaxial imaging catheter 4 can rotate inside the sheath 10 which is fixed to the basket applicator 6 at the tip. The sheath/imaging catheter unit is initially extended to allow the collapsed wire RF applicator basket 6 to fit through the working channel orifice as used in bronchoscopes and other endoscopes. The size may not be so critical in this particular example as the RF Bronchial Thermoplasty therapy is used in the larger airways.

FIG. 2 shows another example of an endoscope where the optical probe head 1 of the optical catheter 4 includes a prism reflector 12 and a focusing lens 16 that provide optical focusing between the prism reflector 12 and the distal end of the fiber 18. The focusing lens 16 or its interface with the distal end of the fiber 18 can be used by the optical probe head 1 to reflect a first portion of an optical probe beam guided by the fiber 18 towards the optical probe head back to the fiber 18 and to direct a section portion of the optical probe beam to a target tissue as the optical imaging beam 14. The prism reflector 12, the focusing lens 16 and the distal end of the fiber 18 are fixed in position relative to each other and rotate or move together as the fiber 18 is rotated or moved by a torque cable 20 attached to the fiber 18. The optical probe head 1 receives light returned from the target tissue under illumination of the optical imaging beam 14 to overlap the light returned from the target tissue with the first portion that does not reach the target tissue to co-propagate in the fiber 18 away from the optical probe head for optical detection and processing.

In many cases, for example in the case of ASM, the therapeutic area of interest [ASM] occurs below the surface of a target tissue under treatment. In this case, if RF power is applied such that the temperature of the ASM is optimized for cytolysis [time/temperature], the airway surface tissue reaches a temperature higher than that of the ASM which may lead to damage and an Adverse Event [AE]. Significant severe AEs are reported after BT therapy [10%]. The endoscope design in FIG. 2 includes a cooling mechanism that cools the surface of the treated target issue to be lower than a temperature underneath the surface. In this regard, a coolant, e.g., a cooling gas or liquid, can be direct to pass near the tissue surface to cool the surface. As a specific example, the conductive struts of the RF basket applicator can be made from flexible hollow tubing which carries the cooling gas or liquid. When a liquid coolant is used, a loop for carrying the liquid coolant is need to direct the coolant out of the endoscope. When a gaseous coolant is used, the gaseous coolant may be directed to the target surface without being collected. In FIG. 2, two hollow basket struts 22 and 23 are used to form two cooling loops near the tip of the endoscope which a coolant is passed. An applicator mount 7 is provided to hold the hollow basket struts 22 and 23 relative to the sheath 10. This design can reduce the surface temperature of the contact region of the target tissue and the peak temperature of the treated location on the target issue occurs at some distance below the surface, depending on the coolant used, and its temperature and flow rate. This places the peak temperature in closer proximity to the target ASM, resulting in improved therapeutic efficacy.

When the above described endoscope is used as a bronchoscope for BT treatment, an image can be acquired from each

bronchial location, e.g., over a 1-second imaging time. During BT local tissue temperatures rise and when approaching approximately 40° C. the process of cell death begins. As the tissue temperature increases the time required to assure cell death decreases. Between about 55-60° C. proteins begin to lose water and at even higher temperatures tissue chars, producing density changes. During the RF power application interval, heat is conducted to surrounding tissues, expanding the tissue damage zone as a function of RF energy [RF power×time]. Periodic tomographic images may be taken at frequent intervals (e.g., few seconds) to monitor process temperature and/or ASM condition.

After suitable pilot studies have been performed to establish the correlation between adverse events and unique microstructural features [e.g. location/extent of ASM, submucosal position/thickness] it will be possible to make therapeutic decisions and set therapeutic parameters based on pretherapeutic imaging results. This should allow physicians to utilize optimum RF power settings to achieve efficacy while minimizing the number and extent of adverse events. Since BT changes the water content and hence the optical properties of ASM, BT induced ASM damage can be ascertained by comparing pre- and post-therapy microstructure images. A growing experience base will allow physicians to personalize therapeutic parameters to obtain the best clinical outcome for the patient. This ability includes locating the ASM as well as temperature/tissue monitoring during RF power application at each site.

The optical catheter **4** in the endoscope in FIGS. **1A** and **1B** or FIG. **2** can be used to obtain both (1) imaging of the target tissue and (2) thermal profile of the target tissue. The imaging information allows the physician to inspect the images of a treat area and the thermal profile of the target tissue allows proper control of the delivery of the RF power for thermotherapy to achieve effect treatment without damaging the tissue. FIG. **3** shows one exemplary process for operating the endoscope in FIGS. **1A** and **1B** or FIG. **2**. First, an optical imaging beam is directed from the endoscope head to the target tissue to obtain reflection of the optical imaging beam from the target tissue and the therapeutic energy is directed to the target tissue being imaged to administer thermal therapy (step **31**). Next, the light returned by the target under illumination of the optical imaging beam is collected and processed to obtain (1) imaging of the target tissue and (2) thermal profile of the target tissue (step **32**). The thermal information of the target tissue obtained from the thermal profile of the target tissue is used as a feedback from the target tissue to control an amount of the therapeutic energy locally applied to different locations of the target tissue to keep a temperature at each location treated by the therapeutic energy between a low threshold and a high threshold for the thermal thermotherapy (step **33**).

In some implementations, the optical probe head **1** can be used to do an imaging scan of the entire treatment area, e.g., performing imaging the surrounding area at each location of the optical probe head **1** along the longitudinal direction of the fiber **18** and repeating this imaging operation at different locations in the longitudinal direction of the fiber **18**. This pre-treatment imaging scan allows construction of an imaging map of the entire area to be treated and identification of selected locations in the entire area to be treated. This pre-treatment imaging scan can be used as a map to control and guide the application of the thermotherapy radiation to the selected locations. During treatment at each of the selected locations, the optical probe head **1** is used to monitoring the thermal profile at each selected location and the monitoring result is used to control the power level of the thermotherapy

radiation being applied. Alternatively, in other implementations, the optical probe head **1** can be directed to each location of the entire treatment area to perform an imaging scan by rotating the optical probe head **1** to detect whether there is a target area to be treated and then perform the radiation treatment if a target area is detected. Then the optical probe head **1** is moved along the longitudinal direction of the fiber **18** to a different location to repeat the optical imaging for detecting one or more targets and then radiation treatment of each detected target. During the radiation treatment, the optical probe head **1** is used to monitoring the thermal profile at each selected location and the monitoring result is used to control the power level of the thermotherapy radiation being applied.

FIGS. **4-8** show examples of an optical catheter **4** for directing the optical imaging beam **14** to the target in the endoscope in FIGS. **1A** and **1B** or in FIG. **2** and associated optical detection and processing. The optical probe head reflects a first portion of an optical probe beam back to the fiber and to direct a section portion of the optical probe beam to a target tissue as the optical imaging beam and further receives light returned from the target tissue under illumination of the optical imaging beam to overlap the light returned from the target tissue with the first portion to co-propagate in the fiber away from the optical probe head. Such design allows for superposition and interplay of different optical waves and modes propagating along substantially the same optical path provided by the fiber **18**. When one of the optical waves or modes interacts with the target, its superposition with another wave or mode can be used for acquiring information about the optical properties of the substance. This use of a common optical path for different optical waves which may be in the same mode or different modes avoids separation of the reference light beam from the sample light beam in various optical coherence domain reflectometry (OCDR) systems and associated technical issues caused by the separation of optical paths such as uncontrolled fluctuations in the relative optical phase or differential delay between the two beams that may adversely affect the measurements. The use of the common optical path for different optical waves in the same or different modes may be advantageously used to stabilize the relative phase among different radiation waves and modes in the presence of environmental fluctuations in the system such as variations in temperatures, physical movements of the system especially of the waveguides, and vibrations and acoustic impacts to the waveguides and system. In this context, such systems have a “built-in” stability of the differential optical path by virtue of their optical designs and are beneficial for some phase-sensitive measurement, such as the determination of the absolute reflection phase and birefringence.

FIG. **4** shows one example of a sensing device according to one implementation. This device directs light in two propagation modes along the same waveguide to an optical probe head near a sample **205** for acquiring information of optical inhomogeneity in the sample. A sample holder may be used to support the sample **205** in some applications. Light radiation from a broadband light source **201** is coupled into the first dual-mode waveguide **271** to excite two orthogonal propagation modes, **001** and **002**. A light director **210** is used to direct the two modes to the second dual-mode waveguide **272** that is terminated by a probe head **220**. The probe head **220** may be configured to perform at least the following functions. The first function of the probe head **220** is to reverse the propagation direction of a portion of light in the waveguide **272** in the mode **001**; the second function of the probe head **220** is to reshape and deliver the remaining portion of the light in mode **002** to the sample **205**; and the third function of the probe head **220** is to collect the light reflected from the sample **205**

back to the second dual-mode waveguide **272**. The back traveling light in both modes 001 and 002 is then directed by light director **210** to the third waveguide **273** and further propagates towards a differential delay modulator **250**. The differential delay modulator **250** is capable of varying the relative optical path length and optical phase between the two modes 001 and 002. A detection subsystem **260** is used to superpose the two propagation modes 001 and 002 to form two new modes, mutually orthogonal, to be received by photo-detectors. Each new mode is a mixture of the modes 001 and 002.

The superposition of the two modes 001 and 002 in the detection subsystem **260** allows for a range detection. The light entering the detection subsystem **260** in the mode 002 is reflected by the sample, bearing information about the optical inhomogeneity of the sample **205**, while the other mode, 001, bypassing the sample **205** inside probe head **220**. So long as these two modes 001 and 002 remain independent through the waveguides their superposition in the detection subsystem **260** may be used to obtain information about the sample **205** without the separate optical paths used in some conventional Michelson interferometer systems.

For the simplicity of the analysis, consider a thin slice of the source spectrum by assuming that the amplitude of the mode 001 is E_{001} in a first linear polarization and that of the mode 002 is E_{002} in a second, orthogonal linear polarization in the first waveguide **271**. The sample **205** can be characterized by an effective reflection coefficient r that is complex in nature; the differential delay modulator **250** can be characterized by a pure phase shift Γ exerted on the mode 001. Let us now superpose the two modes 001 and 002 by projecting them onto a pair of new modes, E_A and E_B , by a relative 45-degree rotation in the vector space. The new modes, E_A and E_B , may be expressed as following:

$$\begin{cases} E_A = \frac{1}{\sqrt{2}}(e^{j\Gamma}E_{001} + rE_{002}); \\ E_B = \frac{1}{\sqrt{2}}(e^{j\Gamma}E_{001} - rE_{002}) \end{cases} \quad (1)$$

It is assumed that all components in the system, except for the sample **205**, are lossless. The resultant intensities of the two superposed modes are

$$\begin{cases} I_A = \frac{1}{2}[E_{001}^2 + E_{002}^2 + |r|E_{001}E_{002}\cos(\Gamma - \varphi)]; \\ I_B = \frac{1}{2}[E_{001}^2 + E_{002}^2 - |r|E_{001}E_{002}\cos(\Gamma - \varphi)] \end{cases} \quad (2)$$

where ϕ is the phase delay associated with the reflection from the sample. A convenient way to characterize the reflection coefficient r is to measure the difference of the above two intensities, i.e.

$$I_A - I_B = |r|E_{001}E_{002}\cos(\Gamma - \phi). \quad (3)$$

If Γ is modulated by the differential delay modulator **250**, the measured signal, Eq. (3), is modulated accordingly. For either a periodic or a time-linear variation of Γ , the measured signal responds with a periodic oscillation and its peak-to-peak value is proportional to the absolute value of r .

For a broadband light source **201** in FIG. 2, consider the two phases, Γ and ϕ to be dependent on wavelength. If the two modes 001 and 002 experience significantly different path lengths when they reach the detection system **260**, the overall phase angle, $\Gamma - \phi$, should be significantly wavelength dependent as well.

Consequently the measured signal, being an integration of Eq. (3) over the source spectrum, yields a smooth function even though Γ is being varied. The condition for a significant oscillation to occur in the measured signal is when the two modes 001 and 002 experience similar path lengths at the location of their superposition. In this case the overall phase angle, $\Gamma - \phi$, becomes wavelength independent or nearly wavelength independent. In other words, for a given relative path length set by the modulator **250**, an oscillation in the measured signal indicates a reflection, in the other mode, from a distance that equalizes the optical path lengths traveled by the two modes 001 and 002. Therefore the system depicted in FIG. 2 can be utilized for ranging reflection sources.

Due to the stability of the relative phase between the two modes, 001 and 002, phase-sensitive measurements can be performed with the system in FIG. 2 with relative ease. The following describes an exemplary method based on the system in FIG. 2 for the determination of the absolute phase associated with the radiation reflected from the sample **205**.

In this method, a sinusoidal modulation is applied to the differential phase by the differential delay modulator **250**, with a modulation magnitude of M and a modulation frequency of Ω . The difference in intensity of the two new modes is the measured and can be expressed as follows:

$$I_A - I_B = |r|E_{001}E_{002}\cos[M\sin(\Omega t) - \phi]. \quad (4)$$

It is clear from Eq. (4) that the measured exhibits an oscillation at a base frequency of Ω and oscillations at harmonic frequencies of the base frequency Ω . The amplitudes of the base frequency and each of the harmonics are related to ϕ and $|r|$. The relationships between r and the harmonics can be derived. For instance, the amplitude of the base-frequency oscillation and the second harmonic can be found from Eq. (4) to be:

$$A_{\Omega} = E_{001}E_{002}J_1(M)|r|\sin\phi; \quad (5a)$$

$$A_{2\Omega} = E_{001}E_{002}J_2(M)|r|\cos\phi; \quad (5b)$$

where J_1 and J_2 are Bessel functions of the first and second order, respectively. Eq. (5a) and (5b) can be used to solve for $|r|$ and ϕ , i.e. the complete characterization of r . We can therefore completely characterize the complex reflection coefficient r by analyzing the harmonic content of various orders in the measured signal. In particular, the presence of the base-frequency component in the measured is due to the presence of ϕ .

FIG. 5 shows an exemplary implementation of the system depicted in FIG. 4. The spectrum of source **201** may be chosen to satisfy the desired ranging resolution. The broader the spectrum is the better the ranging resolution. Various light sources may be used as the source **201**. For example, some semiconductor superluminescent light emitting diodes (SLED) and amplified spontaneous emission (ASE) sources may possess the appropriate spectral properties for the purpose. In this particular example, a polarization controller **302** may be used to control the state of polarization in order to proportion the magnitudes of the two modes, 001 and 002, in the input waveguide **371**. The waveguide **371** and other waveguides **372** and **373** may be dual-mode waveguides and are capable of supporting two independent polarization modes which are mutually orthogonal. One kind of practical and commercially available waveguide is the polarization maintaining (PM) optical fiber. A polarization maintaining fiber can carry two independent polarization modes, namely, the s-wave polarized along its slow axis and the p-wave polarized along its fast axis. In good quality polarization maintaining fibers these two modes can have virtually no energy

exchange, or coupling, for substantial distances. Polarization preserving circulator **310** directs the flow of optical waves according to the following scheme: the two incoming polarization modes from fiber **371** are directed into the fiber **372**; the two incoming polarization modes from fiber **372** are directed to the fiber **373**. A polarization-preserving circulator **310** may be used to maintain the separation of the two independent polarization modes. For instance, the s-wave in the fiber **371** should be directed to the fiber **372** as s-wave or p-wave only. Certain commercially available polarization-preserving circulators are adequate for the purpose.

The system in FIG. **5** implements an optical probe head **320** coupled to the waveguide **372** for optically probing the sample **205**. The probe head **320** delivers a portion of light received from the waveguide **372**, the light in one mode (e.g., 002) of the two modes 001 and 002, to the sample **205** and collects reflected and back-scattered light in the same mode 002 from the sample **205**. The returned light in the mode 002 collected from the sample **205** carries information of the sample **205** and is processed to extract the information of the sample **205**. The light in the other mode 001 in the waveguide **372** propagating towards the probe head **320** is reflected back by the probe head **320**. Both the returned light in the mode 002 and the reflected light in the mode 001 are directed back by the probe head **320** into the waveguide **372** and to the differential delay modulator **250** and the detection system **260** through the circulator **310** and the waveguide **373**.

In the illustrated implementation, the probe head **320** includes a lens system **321** and a polarization-selective reflector (PSR) **322**. The lens system **321** is to concentrate the light energy into a small area, facilitating spatially resolved studies of the sample in a lateral direction. The polarization-selective reflector **322** reflects the mode 001 back and transmits the mode 002. Hence, the light in the mode 002 transmits through the probe head **320** to impinge on the sample **205**. Back reflected or scattered the light from the sample **205** is collected by the lens system **321** to propagate towards the circulator **310** along with the light in the mode 001 reflected by PSR **322** in the waveguide **372**.

In the above described examples, the optical probe head sends out light in two different propagation modes where light in one of the two modes carries the information from the sample. Alternatively, light in a single propagation mode may be used as the input light to the optical probe head and as output light from the optical probe head. Hence, devices based on this design not only use a common optical path to direct light to and from the probe head and sample but also control the light in a single mode. In comparison with above examples where two different modes are used for light coming out of the probe heads, this single-mode design further eliminates or reduces any differences between different modes that propagate in the same optical path.

FIG. **6** shows one exemplary system for acquiring information of optical inhomogeneity and other properties in substances with only one propagation mode inside waveguides. A broadband or low-coherence light from Broadband Light Source **201** is directed to a probe head **2110** by means of polarization-maintaining waveguides **271** and **272**. A partial reflector inside the probe head **2110** reverses the direction of a small portion of the input light to create a radiation wave **1** while transmitting the remainder of the input light to the sample **205**. Backscattered or reflected light from the sample **205** becomes a second radiation wave **2** and is collected by the probe head **2110**. The probe head **2110** combines and couples both the radiation waves **1** and **2** back into the waveguide **272**. The radiation waves **1** and **2** travel in the waveguide **272** towards Light the light director **210** which directs radiation

waves **1** and **2** through the waveguide **273** towards the detection module **2101**. Notably, the radiation waves **1** and **2** output from the probe head **2110** are in the same mode as the input light to the probe head **2110**. the probe head **2110** does not change the mode of light when directing the radiation waves **1** and **2** to the waveguide **272**.

The detection module **2101** includes a beam Splitter **2120**, two optical paths **2121** and **2122**, an optical variable delay element **2123** in the path **2122**, a beam combiner **2130**, and two optical detectors **2141** and **2142**. The beam splitter **2120** splits the light in the waveguide **273**, which includes the radiation waves **1** and **2** in the same mode, into two parts that respectively propagate in the two optical paths **2121** and **2122**. Notably, each of the two parts includes light from both the radiation waves **1** and **2**. The variable delay element or delay line **2123** in the optical path **2122** is controlled by a control signal to adjust the relative optical delay between the two optical paths **2121** and **2122** and may be implemented by, e.g., the exemplary delay elements described in this application and other delay designs. The beam combiner **2130** combines the signals of the two optical paths to overlap with each other and to output two optical signals for optical detectors **2141** and **2142**, respectively. The beam combiner may be a polarization beam splitter which splits the combined light into two parts, orthogonal in polarization to one another.

The probe head **2110** may include a partial reflector to produce the radiation wave **1** which does not reach the sample **205**. Assuming the single propagation mode for the light to the probe head **2110** and the light out of the probe head **2110** is a polarization mode, the light reflected from the partial reflector in the probe head **2110**, i.e., the radiation wave **1**, has the same polarization as the light collected from the sample, the radiation wave **2**. Therefore, both Radiation **1** and **2** travel in the same propagation mode in the waveguides, **272** and **273**. Because the radiation waves **1** and **2** are reflected from different locations, they experience different optical path lengths when reaching the beam splitter **2120**. The effect of variable delay element **2123** is to add an adjustable amount of the delay in the light in the path **2122** relative to the light in the path **2121**.

In operation, the variable delay element **2123** can be adjusted so that the partial radiation **1** reaching the polarization beam splitter **2130** through the path **2122** can be made to experience a similar optical path length as the partial radiation **2** reaching the beam splitter **2130** via the other path **2121**. The superposition of the two beams at the photo detectors **2141** and **2142** causes a measurable intensity variation as their relative path length is being varied by the variable delay element **2123**. This variation can be utilized to retrieve information on the inhomogeneity and other properties of the sample **205**.

FIG. **7** shows an exemplary implementation of the system in FIG. **6** using polarization maintaining optical fibers. A polarization controller **202** may be placed at the output of the light source **201** to control the polarization of the input light in one polarization mode. The optical head **2110** is shown to include a lens system **2111** and a partial reflector **2112**. Two mirrors **1** and **2** are used to construct the two optical paths between the beam splitters **2120** and **2130**. The optical radiation reflected from the partial reflector **2122** and from the sample **205** travel in the polarization-maintaining (PM) fiber **272** in the same mode. The main portions of the radiation waves **1** and **2** are deflected to the mirror **1** while the remaining portions are directed to the mirror **2** by the beam splitter **2120**.

The incident plane of the polarizing beam splitter **2130** can be made to have a finite angle with respect to the polarization

directions of light from both the Mirror **2** in one optical path and the variable delay element **2123** from the other optical path. In this configuration, light energies received by both detectors **2141** and **2142** are the superposition of the two radiations, i.e., Radiation **1** and Radiation **2**. It should be appreciated that the linkage between the beam splitters **2120** and **2130** can be made by means of optical fibers or other optical waveguides to eliminate the free space paths and the two mirrors **1** and **2**.

In some implementations, the probe head may be designed to cause a first portion of the first mode to reverse its propagation direction while directing the remaining portion, or a second portion, to reach the sample. The reflection or back scattered light of the second portion from the sample is collected by the probe head and is controlled in the second propagation mode different from the first mode to produce a reflected second portion. Both the reflected first portion in the first propagation mode and the reflected second portion in the second propagation mode are directed by the probe head through a common waveguide into the detection module for processing. In comparison with the implementations that use light in two modes throughout the system, this alternative design further improves the stability of the relative phase delay between the two modes at the detection module and provides additional implementation benefits.

FIG. **8** illustrate one exemplary design of the optical layout of the optical sensing system and its system implementation with an electronic controller. An input waveguide **871** is provided to direct light in a first propagation mode, e.g., the mode 001, from the broadband light source **201** to a light director **810**. The waveguide **871** may be a mode maintaining waveguide designed to support at least one propagation mode such as the mode 001 or 002. When light is coupled into the waveguide **871** in a particular mode such as the mode 001, the waveguide **871** essentially maintains the light in the mode 001. A polarization maintaining fiber supporting two orthogonal linear polarization modes, for example, may be used as the waveguide **871**. Dual-mode waveguides **272** and **273** are used to direct the light. A light director **510** is used to couple the waveguides **871**, **272**, and **273**, to convey the mode 001 from the input waveguide **871** to one of the two modes (e.g., modes 001 and 002) supported by the dual-mode waveguide **272**, and to direct light in two modes from the waveguide **272** to the dual-mode waveguide **273**. In the example illustrated in FIG. **8**, the light director **810** couples the light in the mode 001 from the waveguide **871** into the same mode 001 in the waveguide **272**. Alternatively, the light director **810** may couple the light in the mode 001 from the waveguide **871** into the different mode 002 in the waveguide **272**. The dual-mode waveguide **271** is terminated at the other end by a probe head **820** which couples a portion of light to the sample **205** for sensing.

The probe head **820** is designed differently from the probe head **320** in that the probe head **830** converts part of light in the mode 001 into the other different mode 002 when the light is reflected or scattered back from the sample **205**. Alternatively, if the light in the waveguide **272** that is coupled from the waveguide **871** is in the mode 002, the probe head **820** converts that part of light in the mode 002 into the other different mode 001 when the light is reflected or scattered back from the sample **205**. In the illustrated example, the probe head **820** performs these functions: a) to reverse the propagation direction of a small portion of the incoming radiation in mode 001; b) to reshape the remaining radiation and transmit it to the sample **205**; and c) to convert the radiation reflected from the sample **205** to an independent mode 002 supported by the dual-mode waveguide **272**. Since the

probe head **820** only converts part of the light into the other mode supported by the waveguide **272**, the probe head **820** is a partial mode converter in this regard. Due to the operations of the probe head **820**, there are two modes propagating away from the probe head **820**, the mode 001 that bypasses the sample **205** and the mode 002 for light that originates from sample reflection or back scattering.

In the examples shown in FIGS. **4-8**, an optical delay is introduced at the output of the common optical waveguide to perform a time-domain imaging in extracting images from the returned light by the optical probe head. The depth scan is obtained based on a Fourier-transform from the measured signals at different phase delays at different times. Alternatively, the same designs for the optical probe head and the common optical waveguide can be used to extract images from the returned light by the optical probe head based on optical frequency domain imaging (OFDI) under a swept source optical coherence tomography configuration. Under OFDI, a wavelength-swept light source is used to produce a probe beam to the optical probe head for probing the amplitude, phase, polarization and spectral properties of back scattering light from a target, e.g., a tissue. The depth scan is obtained based on a Fourier-transform from the acquired spectra. As such, the optical delay used in the time-domain technique can be eliminated. OFDI offers intrinsic signal-to-noise ratio (SNR) advantage over the time domain techniques because the interference signal can be effectively integrated through a Fourier transform enabling significant improvements in imaging speed, sensitivity and ranging depth required for in-vivo tissue imaging.

The optical imaging information obtained from the devices in FIGS. **4-8** based on either the time-domain imaging and frequency-domain imaging can be processed to map the thermal field or profile of a target. Changes in density, due to temperature changes, affect the optical properties [such as refractive index] of water and tissue. FIGS. **9** and **10** show measurements in water to illustrate the correlations between the temperature, density and the refractive index. FIG. **10** shows the refractive index of water as a function of temperature at standard density and 589 nm wavelength. A linear fit shows that the curve can be parameterized by $n=1.341-2.262 \times 10^{-5}T$.

In some implementations, the thermal treatment radiation exposure is not applied to a tissue occurs during a monitoring period when the optical imaging beam is directed to the target to measure the target. In other implementations, both the optical imaging beam and the thermal treatment radiation can be applied to the target at the same time. The result of an optical index change can cause a phase shift which can lead to a full fringe for a small [1 deg. C] temperature change allowing a two dimensional representation of the tissue density change to be assembled. The phase changes on each scanning of the relative phase between the optical imaging beam and the reflected beam that does not reach the target are translated into contrast changes. Scanning the optical imaging beam in a circumferential manner, while rapidly performing a series of imaging scans allows a raster scan image to be constructed where the thermally-induced density changes show up as contrast variations. The contrast variations reflect the combined effect of temperature changes and tissue denaturing [water loss], two important parameters for BT. By comparing the pre- and intra-BT images, the temperature increase due to thermotherapy can be extracted.

Since cytotoxicity requires only a few minutes exposure to temperatures between 50-60 deg C., and denaturation requires temperatures in excess of 60 deg C., thermally-induced image contrast is almost completely due to the change in tempera-

15

ture rather than water loss. However, above 60 deg C., water loss [denaturation] becomes increasingly important as an additional contribution to image contrast. The post-BT image [after cooling] contrast will reflect changes in ASM density due to BT. Notably, the optical techniques described in this document can be used to obtain imaging of the, microstructure, spectral properties and temperature profile of a target tissue, organ or object. Consequently, if we obtain a tomographic tissue image and apply an energy source [light, RF] so as to subsequently generate a temperature change [non-uniform in many cases], we can map that temperature gradient by the image contrast changes it produces. The overall functions of the present techniques include 1) monitoring therapeutic temperature [‘process control’] to maintain it in the correct ‘thermotherapeutic window’ and 2) pre- and post-tissue imaging to assess the extent of tissue ‘damage’ as a result of thermotherapy.

While this specification contains many specifics, these should not be construed as limitations on the scope of an invention that is claimed or of what may be claimed, but rather as descriptions of features specific to particular embodiments. Certain features that are described in this specification in the context of separate embodiments can also be implemented in combination in a single embodiment. Conversely, various features that are described in the context of a single embodiment can also be implemented in multiple embodiments separately or in any suitable sub-combination. Moreover, although features may be described above as acting in certain combinations and even initially claimed as such, one or more features from a claimed combination can in some cases be excised from the combination, and the claimed combination may be directed to a sub-combination or a variation of a sub-combination

Only a few examples and implementations are described. One of ordinary skill in the art can readily recognize that variations, modifications and enhancements to the described examples may be made.

What is claimed is:

1. An endoscope device, comprising:
a light source that provides an optical imaging beam;
an optical catheter comprising an optical fiber to guide an optical imaging beam and an optical probe head, located at a distal end of the fiber to direct the optical imaging beam from the fiber to a target and to receive light returned from the target under illumination of the optical imaging beam, the optical probe head of the optical catheter comprising a prism that reflects the optical imaging beam from the fiber to a location of the target and collects the light returned from the target into the fiber under the illumination of the optical imaging beam;
a detection module to process the light returned from the target under illumination of the optical imaging beam to measure a temperature at a location illuminated by the optical imaging beam;
an energy source that produces energy that is applied to the target to raise a temperature at a location of the target where the energy is applied; and
a control mechanism that controls the energy source to set a power level of the energy produced by the energy source based on the measured temperature from the detection module.

2. The device of claim 1, wherein: the energy source is an RF source, and the device comprises an RF applicator engaged to the sheath to receive RF energy from the RF source and to apply the received RF energy to the target.

16

3. The device of claim 2, wherein:

the RF applicator includes a hollow tube that both conducts the RF energy for application of the RF energy to the target and carries a coolant inside the hollow tube to cool a surface of the target to allow a temperature at a location underneath the surface to be higher than a temperature of the surface under application of the RF energy.

4. The device of claim 2, wherein:

the RF applicator surrounds the optical catheter and allows the optical catheter to rotate within the sheath.

5. The device of claim 4, comprising:

a housing comprising a hollow working channel, in which the optical catheter is placed to rotate within the hollow working channel, and

an RF transmission path connecting the RF applicator and the RF source RF applicator and located inside the hollow working channel.

6. The device of claim 5, wherein:

the RF applicator comprises an expandable basket that can carry RF energy, the expandable basket structured to expand when the expandable basket is located outside a distal opening of the hollow working channel and to collapse when the expandable basket is pulled into the distal opening of the hollow working channel.

7. The device of claim 1, wherein:

the optical probe head of the optical catheter splits the optical imaging beam from the fiber into a first portion that is directed to the target and a second portion that does not reach the target and is reflected back to the fiber to co-propagate with the light returned from the target that is collected by the optical probe head along the fiber, and

the detection module receives both the second portion and the light returned from the target via the fiber and processes the received second portion and the light returned from the target to extract the measured temperature.

8. The device as in claim 7, wherein:

the detection module comprises an optical differential delay unit to produce and control a relative phase delay between the second portion and the light returned from the target via the fiber and performs a time-domain Fourier transform of the obtained signals at different relative phase delays.

9. The device as in claim 7, wherein:

the detection module performs a frequency-domain Fourier transform of the spectra of the light from the optical probe head.

10. The device of claim 1, comprising:

a beam scanning mechanism to scan the optical imaging beam on the target to direct the optical imaging beam to different positions on the target, wherein the detection module changes a phase delay between two portions of received light of the second portion and the light returned from the target via the fiber at each location of the optical imaging beam on the target to obtain an image of the each location of the target.

11. The device of claim 1, wherein:

the detection module processes received light of the second portion and the light returned from the target via the fiber to measure a change of a density of the target caused by application of the energy from the energy source.

12. The device of claim 1, comprising:

a mechanism to cool a surface of the target to allow a temperature at a location underneath the surface to be higher than a temperature of the surface under application of the energy from the energy source.

17

13. The device of claim 12, wherein:
the mechanism comprises a tube that carries a cooling flow
to pass a location near the distal end of the optical cath-
eter.
14. The device of claim 1, wherein:
the energy source produces a laser beam, separate from the
optical imaging beam, as the energy applied to the target.
15. The device of claim 1, wherein:
the energy source generates a microwave energy signal as
the energy applied to the target.
16. The device of claim 1, wherein:
the energy source generates an ultrasound signal as the
energy applied to the target.
17. The device of claim 2, further comprising:
two hollow basket struts that are used to form two cooling
loops near a tip of the endoscope which a coolant is
passed; and an applicator mount that is provided to hold
the hollow basket struts relative to the sheath.
18. A method for guiding the application of a thermothera-
peutic radiation to a target tissue, comprising:
using an optical catheter comprising
an optical fiber to guide an optical imaging beam to
direct the optical imaging beam from the fiber to a
target, and
an optical probe head comprising a prism to reflect the
optical imaging beam from the fiber to a location of
the target and to collect the light returned from the
target into the fiber under illumination of the optical
imaging beam;
processing the light returned from the target tissue under
illumination of the optical imaging beam to measure a
temperature at a location illuminated by the optical
imaging beam;
applying a thermotherapeutic energy to the target tissue to
raise a temperature at a location of the target tissue
where the energy is applied; and
controlling an amount of the thermotherapeutic energy
applied to the target tissue, based on the measured tem-
perature by using the optical imaging beam, to control
the temperature at the target tissue between a low limit
above which a thermotherapeutic effect is present and a
high limit above which a damage to the target tissue
occurs.
19. The method as in claim 18, comprising:
using the optical imaging beam to obtain an image of a
location of the target tissue illuminated by the optical
imaging beam, in addition to obtaining the measured
temperature; and
using both the measured temperature and the obtained
image from the optical imaging beam to monitor and
control application of the thermotherapeutic energy to
the target tissue.
20. The method as in claim 18, comprising: using the
optical catheter to split the optical imaging beam from the
fiber into a first portion that is directed to the target and a

18

second portion that does not reach the target and is reflected
back to the fiber to co-propagate with the light returned from
the target tissue.

21. The method as in claim 20, comprising:

- producing a relative phase delay between the second por-
tion and the light returned from the target tissue via the
fiber to cause an optical interference between the second
portion and the light returned from the target tissue;
detecting a change to the optical interference caused by the
application of the thermotherapeutic energy to the target
tissue to measure a change in density of the target tissue;
and

using the measured change in density of the target tissue to
obtain the measured temperature.

22. The method as in claim 20, comprising:

- performing a frequency-domain Fourier transform of spec-
tra of light in the fiber that combines the light returned
from the target tissue and the second portion that does
not reach the target to extract a change to the optical
interference caused by the application of the thermo-
therapeutic energy to the target tissue and to measure a
change in density of the target tissue; and

using the measured change in density of the target tissue to
obtain the measured temperature.

23. The method as in claim 18, comprising:

- scanning the optical imaging beam on the target tissue to
obtain a two-dimensional map of temperatures at loca-
tions of the target tissue; and

using the two-dimensional map of temperatures at loca-
tions of the target tissue to control the application of the
thermotherapeutic energy to the target tissue.

24. The method as in claim 18, wherein:

- the processing of the light returned from the target tissue
under illumination of the optical imaging beam com-
prises:

obtaining image contrasts from the light returned from the
target tissue before and after the thermotherapeutic
energy;

comparing the obtained image contrasts from the light
returned from the target tissue before and after the ther-
motherapeutic energy to obtain a contrast variation
caused by application of the thermotherapeutic energy;
and

using the contrast variation to obtain the measured tem-
perature.

25. The method as in claim 18, comprising:

- cooling a surface of the target tissue to allow a temperature
at a location underneath the surface to be higher than a
temperature of the surface under application of the ther-
motherapeutic energy.

26. The method as in claim 25, wherein:

- using a tube that is located near the distal end of the optical
catheter to carry a cooling flow to providing the cooling
of the surface of the target tissue.

* * * * *

专利名称(译)	温度曲线图和引导热疗法		
公开(公告)号	US9247991	公开(公告)日	2016-02-02
申请号	US13/903983	申请日	2013-05-28
[标]申请(专利权)人(译)	突慧公司		
申请(专利权)人(译)	TOMOPHASE CORPORATION		
当前申请(专利权)人(译)	TOMOPHASE INC.		
[标]发明人	NORRIS PETER E LI XIAO LI VERTIKOV ANDREY		
发明人	NORRIS, PETER E. LI, XIAO-LI VERTIKOV, ANDREY		
IPC分类号	A61B6/00 A61B18/18 A61B1/00 A61B5/00 A61B5/01 A61B17/32 A61B18/20		
CPC分类号	A61B1/00096 A61B1/00172 A61B5/0066 A61B5/0084 A61B5/015 A61B5/6852 A61B18/18 A61B17/320068 A61B18/20 A61B2017/320069		
优先权	PCT/US2009/035773 2009-03-02 WO 61/032853 2008-02-29 US		
其他公开文献	US20130261613A1		
外部链接	Espacenet USPTO		

摘要(译)

使用光学探头将光传递到目标并收集来自目标的光以用于成像和监测目标同时施加单独的辐射以治疗目标的技术，装置和系统。

