




US 20150216612A1

(19) **United States**(12) **Patent Application Publication****Leo et al.**(10) **Pub. No.: US 2015/0216612 A1**(43) **Pub. Date: Aug. 6, 2015**(54) **MEDICAL APPARATUS SYSTEM HAVING
OPTICAL FIBER SENSING CAPABILITY****Publication Classification**(71) Applicant: **St. Jude Medical Luxembourg Holding
S.à.r.l.**, Luxembourg (LU)(72) Inventors: **Giovanni Leo**, Cologny (CH); **Nicolas
Aeby**, Geneva (CH); **Daniele Inaudi**,
Lugano (CH)(51) **Int. Cl.****A61B 19/00** (2006.01)**A61B 5/00** (2006.01)(52) **U.S. Cl.**CPC **A61B 19/46** (2013.01); **A61B 5/6852**
(2013.01); **A61B 18/1492** (2013.01)(21) Appl. No.: **14/573,666**(22) Filed: **Dec. 17, 2014****Related U.S. Application Data**(63) Continuation of application No. 13/096,647, filed on
Apr. 28, 2011, now Pat. No. 8,932,288, which is a
continuation of application No. 11/436,926, filed on
May 15, 2006, now Pat. No. 8,075,498, which is a
continuation-in-part of application No. 11/237,053,
filed on Sep. 28, 2005, now Pat. No. 8,182,433.(60) Provisional application No. 60/704,825, filed on Aug.
1, 2005.(30) **Foreign Application Priority Data**

Mar. 4, 2005 (EP) 05004852.9

(57) **ABSTRACT**

An apparatus and method for diagnosis or treatment of a vessel or organ. The apparatus includes a deformable body such as a catheter having a tissue ablation end effector and an irrigation channel in fluid communication therewith. At least two sensors are disposed within a distal extremity of the deformable body, the sensors being responsive to a wave in a specified range of frequency to detect deformations resulting from a contact force applied to the distal extremity. A microprocessor can be operatively coupled with the sensors to receive outputs therefrom, the microprocessor being configured to resolve a multi-dimensional force vector corresponding to the contact force. In one embodiment, the sensors are fiber Bragg grating sensors, and the wave is injected into the fiber Bragg grating strain sensors from a laser diode.







FIG. 2




FIG. 3




FIG. 4




FIG. 5




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10







FIG. 14






FIG. 15




FIG. 16




FIG. 17

FIG. 18




FIG. 19




FIG. 20





FIG. 21



MEDICAL APPARATUS SYSTEM HAVING OPTICAL FIBER SENSING CAPABILITY

RELATED APPLICATIONS

[0001] The present application is a continuation of U.S. patent application Ser. No. 11/11/436,926, filed May 15, 2006, which is a continuation-in-part of U.S. patent application Ser. No. 11/237,053, filed Sep. 28, 2005, which claims the benefit of U.S. Provisional Application No. 60/704,825, filed Aug. 1, 2005 and which also claims priority from European Patent Application No. EP 05004852.9 filed Mar. 4, 2005, all of which are hereby incorporated by reference herein.

FIELD OF THE INVENTION

[0002] The present invention relates to apparatus for exploring and treating an organ that permits computation of a multi-dimensional force vector resulting from contact between the distal extremity of the apparatus and the tissue of the wall of the organ.

BACKGROUND OF THE INVENTION

[0003] For many years, exploration and treatment of various organs or vessels has been possible using catheter-based diagnostic and treatment systems. Such catheters are introduced through a vessel leading to the cavity of the organ to be explored or treated or alternatively may be introduced directly through an incision made in the wall of the organ. In this manner, the patient avoids the trauma and extended recuperation times typically associated with open surgical procedures.

[0004] To provide effective diagnosis or therapy, it is frequently necessary to first map the zone to be treated with great precision. Such mapping may be performed, for example, when it is desired to selectively ablate current pathways within a heart to treat atrial fibrillation. Often, the mapping procedure is complicated by difficulties in locating the zone (s) to be treated due to periodic movement of the heart throughout the cardiac cycle.

[0005] Previously-known systems for mapping the interior of a vessel or organ are described, for example, in U.S. Pat. Nos. 6,546,271 and 6,226,542. The catheters described in those patents employ electromagnetic, magnetic or acoustic sensors to map the position of a distal end of the catheter in space and then construct a three-dimensional visualization of the vessel or organ interior.

[0006] One drawback of such previously known mapping systems is that they rely on manual feedback of the catheter and/or impedance measurements to determine when the catheter is properly positioned in the vessel or organ. Those systems do not measure contact forces with the vessel or organ wall or detect contact forces applied by the catheter against the organ or vessel wall that may modify the true wall location. Instead, previously known mapping methods are time-consuming, dependent upon the skill of the clinician, and cannot compensate for artifacts created by excessive contact forces.

[0007] It therefore would be desirable to provide apparatus and methods for detecting and monitoring contact forces between a mapping catheter and the wall of the organ or vessel to permit faster and more accurate mapping. It also would be desirable to provide apparatus and methods that permit the process to be automated, thereby improving reg-

istration of measured electro-physiologic values and spatial coordinates, for example, by recording such values only where the contact forces fall within a predetermined range.

[0008] Once the topography of the vessel or organ is mapped, either the same or a different catheter may be employed to effect treatment. Depending upon the specific treatment to be applied to the vessel or organ, the catheter may comprise any of a number of end effectors, such as RF ablation electrodes, a rotary cutting head, laser ablation system, injection needle or cryogenic fluid delivery system. Exemplary systems are described, for example, in U.S. Pat. Nos. 6,120,520, 6,102,926, 5,575,787, 5,409,000 and 5,423,807.

[0009] Because the effectiveness of such end effectors often depends having the end effector in contact with the tissue of the wall of the organ or vessel, many previously-known treatment systems include expandable baskets or hooks that stabilize the distal extremity of the catheter in contact with the tissue. Such arrangements, however, may be inherently imprecise due to the motion of the organ or vessel. Moreover, the previously-known systems do not provide the ability of sense the load applied to the distal extremity of the catheter by movement of the tissue wall.

[0010] For example, in the case of a cardiac ablation system, at one extreme the creation of a gap between the end effector of the treatment system and the tissue wall may render the treatment ineffective, and inadequately ablate the tissue zone. At the other extreme, if the end effector of the catheter contacts the tissue wall with excessive force, it may inadvertently puncture the tissue, resulting in cardiac tamponade.

[0011] In view of the foregoing, it would be desirable to provide a catheter-based diagnostic or treatment system that permits sensing of the load applied to the distal extremity of the catheter, including periodic loads arising from movement of the organ or tissue. It further would be desirable to have a load sensing system coupled to control operation of the end effector, so that the end effector is operated, either manually or automatically, only when the contact force is detected to fall within a predetermined range.

[0012] U.S. Pat. No. 6,695,808 proposes several solutions to measure the force vector arising from contact with the tissue surface, including mechanical, capacitive, inductive and resistive pressure sensing devices. One drawback of such devices, however, is that they are relatively complex and must be sealed to prevent blood or other liquids from disturbing the measurements. In addition, such load sensing devices may result in an increase in the insertion profile of the distal extremity of the catheter. Still further, sensors of the types described in that patent may be subject to electromagnetic interference.

[0013] One previously-known solution for dealing with potential electromagnetic interference in the medical environment is to use light-based systems rather than electrical measurement systems, such as described in U.S. Pat. No. 6,470,205 to Bosselman. That patent describes a robotic system for performing surgery comprising a series of rigid links coupled by articulated joints. A plurality of Bragg gratings are disposed at the articulated joints so that the bend angle of each joint may be determined optically, for example, by measuring the change in the wavelength of light reflected by the Bragg gratings using an interferometer. Calculation of the bend angles does not require knowledge of the characteristics of the rigid links.

[0014] International Publication No. WO 01/33165 to Bucholtz describes an alternative spatial orientation system wherein wavelength changes measured in a triad of optical fiber strain sensors are used to compute the spatial orientation of a catheter or other medical instrument. Although the publication discloses that the strain sensors may be encased within a deformable sheath, as in Bosselman, calculation of the bend angles is not described as requiring characterization of the material properties of the deformable sheath.

[0015] Accordingly, it would be desirable to provide diagnostic and treatment apparatus, such as a catheter or guide wire, that permits sensing of loads applied to a distal extremity of the apparatus, but which do not substantially increase the insertion profile of the apparatus.

[0016] It further would be desirable to provide diagnostic and treatment apparatus, such as a catheter and guide wire, that permits computation of forces applied to a distal extremity of the apparatus, and which are substantially immune to electromagnetic interference.

[0017] It still further would be desirable to provide a diagnostic and treatment apparatus, such as catheter system, that permits computation of forces applied to a distal extremity of the catheter that is substantially immune to environmental conditions encountered during use of the catheter, such as exposure to body fluids and the presence of room-to-body temperature gradients.

SUMMARY OF THE INVENTION

[0018] In view of the foregoing, it is object of the present invention to provide diagnostic or treatment apparatus that permits sensing of the load applied to a distal extremity of apparatus, including periodic loads arising from movement of the organ or tissue.

[0019] It is another object of this invention to provide apparatus and methods for detecting and monitoring contact forces between an interventional apparatus, such as a mapping catheter or guide wire, and the wall of the organ or vessel to facilitate the speed and accuracy of such mapping.

[0020] It is a further object of the present invention to provide apparatus and methods that enable a mapping or treatment process to be automated, thereby improving registration of measured electro-physiologic values and spatial coordinates, for example, by recording such values only where the contact forces fall within a predetermined range.

[0021] It is also an object of this invention to provide apparatus wherein a load sensing system is coupled to control operation of an end effector of a diagnostic or treatment apparatus, so that the end effector is operated, either manually or automatically, only when the contact force is detected to fall within a predetermined range.

[0022] It is another object of this invention to provide diagnostic and treatment apparatus, that permit sensing of loads applied to a distal extremity of the apparatus, but which do not substantially increase the insertion profile of the apparatus.

[0023] It is a further object of this invention to provide diagnostic and treatment apparatus that permit computation of forces applied to a distal extremity of the apparatus, and which are substantially immune to electromagnetic interference.

[0024] It is still another object of this invention to provide a diagnostic and treatment apparatus that permits computation of forces applied to a distal extremity of the catheter, but which is substantially immune to environmental conditions

encountered during use of the catheter, such as exposure to body fluids and the presence of room-to-body temperature gradients.

[0025] It is also an object of this invention to provide apparatus for use in a hollow-body organ, such as the heart, that permits sensing of loads applied to a distal extremity of the apparatus during movement of the organ, so as to optimize operation of an end effector disposed within the distal extremity.

[0026] These and other objects of the invention are accomplished by providing medical apparatus, such as catheter, having at least two optical fiber sensors disposed in a distal extremity configured to deform responsive to contact forces, and processing logic programmed to compute at least a two-dimensional force vector responsive to detected changes in the optical characteristics of the optical fiber sensors. The apparatus of the present invention may be configured as a catheter or guide wire, or may be employed in other medical apparatus where knowledge of tissue contact forces is desired.

[0027] More preferably, the apparatus of the present invention comprises three optical fiber sensors disposed within the distal extremity so that they are not co-planar. For example, the three optical fiber sensors may be arranged at the apices of an equilateral triangle centered on the geometric axis of the apparatus, although other configurations also may be employed. Use of three such optical fiber sensors advantageously permits the computation of a three-dimensional force vector. The optical fiber sensors preferably are chosen from among a Fiber Bragg Grating (FBG), an Intrinsic Fabry-Perot Interferometer (IFPI), an Extrinsic Fabry-Perot Interferometer (EFPI), a Long Period Grating (LPG), a two, three or four arm Michelson interferometer (MI), a Brillouin scattering strain sensor, or intensity-based fiber optic strain sensor.

[0028] Further in accordance with the principles of the present invention, the apparatus includes processing logic, such as programmed general purpose microprocessor or application specific integrated circuit, operatively coupled to receive an output signal from the optical fiber sensors, and to compute a two- or three-dimensional force vector from that output signal, depending upon the number of optical fiber sensors employed. The processing logic may be programmed with a matrix of values associated with physical properties of an individual device, and applies those values to the detected changes in wavelength to compute the external forces applied to the distal extremity. More preferably, a force-strain conversion matrix specific for each device is determined during manufacture, and that force-strain conversion is associated with the device via an appropriate memory device, label or tag.

[0029] In accordance with the one aspect of the present invention, two optical fiber sensors may be used provided that the neutral axis of the distal extremity of the apparatus is well characterized. More preferably, three optical fiber sensors are disposed within the distal extremity to allow deformations (elongation or contraction) imposed on the deformable body to be measured at three or more non-planar points.

[0030] The extremely small dimensions of the optical fiber sensors provide ample space in the distal extremity of the apparatus to house for other diagnostic or treatment devices. When configured as a catheter or guide wire, the device has a substantially reduced insertion profile relative to previously-known systems having force-sensing capability. In addition, the optical nature of the sensors ensures that the possible

presence of liquids does not disturb the measurements, and ensures a high degree of immunity from electromagnetic interference.

[0031] The apparatus of the present invention optionally may include any of a number of previously-known end effectors disposed in the distal extremity for treating a vessel or organ, for example, an electrode to measure an electric potential (e.g., to perform an endocavity electrocardiogram), an electrode configured to ablate tissue by deposition of radiofrequency energy, an irrigation channel, and/or a three-dimensional positioning sensor.

[0032] Advantageously, the load sensing system of the present invention may be employed to continuously monitor deflection of a distal extremity. For example, the signal output by the load sensing system may be used to guide or control the use and operation of an end effector of a catheter either manually or automatically. Illustratively, when employed as part of an electrophysiology mapping catheter, the present invention permits electrical potentials of the tissue to be measured only at contact positions where the contact force applied to the distal extremity of the catheter by the tissue wall falls within a predetermined range. Such an arrangement not only offers to improve spatial registration between the mapped values and tissue location, but also makes possible the use of robotic systems capable of automating the mapping process. As a further example, the output of the load sensing system may be used to control operation of a treatment end effector, for example, to position the end effector in contact with the organ wall and to energize the ablation electrode only when the contact force is detected to fall within a predetermined range.

[0033] In addition, the distal part of at least one of the optical fibers, or an additional optical fiber, extends beyond the others and is equipped with an additional FBG, LPG, IFPI, EFPI or Brillouin scattering type sensor to permit the temperature of the distal extremity to be monitored.

[0034] Alternatively, or in addition, a temperature sensor may be disposed in the distal extremity in close proximity to the optical fiber sensors. Temperatures measured by the temperature sensor may be used to compensate for deformations of the deformable body arising from temperature variations, which might otherwise erroneously be interpreted as force-related deformations. The temperature sensor may comprise any of a number of temperature sensors. More specifically, the temperature sensor comprises an additional optic fiber that is not constrained to deform in unison with the distal extremity, but instead is free to expand due to temperature variations. In a preferred embodiment, the temperature sensor comprises an additional FBG, LPG, IFPI, EFPI or Brillouin scattering type optical fiber sensor.

[0035] The additional optical fiber also could extend beyond the other optical fibers and include an additional FBG, LPG, IFPI, EFPI or Brillouin scattering type sensor to measure the temperature of the distal extremity. Alternatively, the distal part of the additional fiber extends beyond the other optical fibers in the distal extremity and includes a temperature sensor comprising a Michelson interferometer sensor or an intensity sensor.

[0036] In accordance with a preferred alternative embodiment, the apparatus may comprise an electrophysiology catheter comprising an elongated portion, a distal extremity, and a proximal end. An irrigation tube extends from the proximal end to the distal extremity and has a plurality of optical fibers arranged symmetrically around its circumference. The opti-

cal fibers include sensors, such as Bragg Gratings, disposed near the distal extremity. In accordance with one aspect of the invention, the irrigation tube in the vicinity of the distal extremity comprises a flexible tube having a low thermal expansion coefficient which reduces sensor artifacts introduced by environmental effects, such as a temperature fluctuations.

BRIEF DESCRIPTION OF THE DRAWINGS

[0037] Further features of the invention, its nature and various advantages will be more apparent from the accompanying drawings and the following detailed description of the preferred embodiments, in which:

[0038] FIG. 1 is a schematic view of apparatus according to the invention;

[0039] FIG. 2 is a schematic plan view of the distal extremity of FIG. 1;

[0040] FIG. 3 is a section according to III-III of FIG. 2;

[0041] FIG. 4 is a schematic view of the side of the distal extremity showing the disposition of the Fiber Bragg Grating (FBG) or Long Period Grating (LPG) sensors;

[0042] FIG. 5 is a schematic view of the side of the distal extremity showing the disposition of the Intrinsic Fabry-Perot Interferometer (IFPI) sensors;

[0043] FIG. 6 is a schematic view of the side of the distal extremity showing the disposition of the Extrinsic Fabry-Perot Interferometer (EFPI) sensors;

[0044] FIG. 7 is a schematic view of the side of the distal extremity showing the disposition of the Michelson interferometer sensors;

[0045] FIG. 8 is a schematic view of the side of the distal extremity showing the disposition of the High Resolution Brillouin sensors;

[0046] FIG. 9 is a schematic view of the side of the distal extremity showing the disposition of the reflection intensity sensors;

[0047] FIG. 10 is a schematic view of the side of the distal extremity showing the disposition of the microbending intensity sensors;

[0048] FIG. 11 is a perspective view of three optical fibers in contact with each other;

[0049] FIG. 12 is a perspective view of three optical fibers in contact with each other and forming an integral part;

[0050] FIG. 13 is a schematic plan view of the distal extremity with the optical fibers of FIG. 6 forming an integral part of the distal extremity;

[0051] FIG. 14 is an exploded perspective view of the distal extremity of an exemplary catheter constructed in accordance with the present invention;

[0052] FIG. 15 is a schematic plan view of the distal extremity including a fourth optical fiber;

[0053] FIG. 16 is a schematic view of apparatus of the present invention wherein the output of the load sensing system is utilized to control automated operation of the apparatus;

[0054] FIG. 17 is a schematic view of an alternative embodiment of apparatus of the present application;

[0055] FIG. 18 is a perspective view of a distal subassembly of the apparatus of FIG. 17;

[0056] FIG. 19 is a perspective view of the distal subassembly of FIG. 18 including a protective housing, which is partially cut-away;

[0057] FIG. 20 is a cross-sectional view of the distal subassembly of FIG. 19 taken along line 20-20; and

[0058] FIG. 21 is a perspective view of an exemplary deflectable catheter shaft for use with the distal subassembly of FIG. 19.

DETAILED DESCRIPTION OF THE DRAWINGS

[0059] The present invention is directed to medical apparatus and methods for use with diagnostic and treatment systems wherein it is desired to measure contact forces between a distal extremity of the apparatus and a tissue wall of an organ or vessel. The load sensing capability of the present invention may be used intermittently to measure the contact forces at discrete points, or alternatively, used to continuously monitor contact forces to assist in manipulation and operation of the apparatus.

[0060] Medical apparatus incorporating the present invention illustratively may be configured as catheters or guide wires to be manually manipulated by a clinician, with the clinician using a visual or audio cue output by the load sensing system to determine, for example, optimum position for measuring an electrophysiologic value or performing treatment. Alternatively, the medical apparatus may be robotically controlled, with the load sensing system of the present invention providing a feedback and control system.

[0061] Advantageously, medical apparatus equipped with the load sensing system of the present invention are expected to permit faster, more accurate diagnosis or treatment of a vessel of organ, with improved registration between measured values and spatial locations. For example, a catheter with the inventive load sensing system would permit mapping of cardiac electrical potentials by providing reproducible contact forces between the distal extremity of the catheter and the tissue wall, thereby making the results of the mapping process less dependent on the skill of the individual clinician and facilitating automated procedures.

[0062] Referring now to FIGS. 1 and 2, exemplary apparatus constructed in accordance with the principles of the present invention comprises catheter 1 having proximal end 2 coupled to console 3 via cable 4. As described in detail below, catheter 1 includes distal extremity 5 that illustratively carries any one or more of a number of end effectors known in the art for diagnosing or treating a vessel or organ. While the present invention is described in the context of a catheter system for cardiac mapping and ablation, it will be understood that medical apparatus constructed in accordance with the present invention advantageously may be used for other purposes, such as delivering drugs or bioactive agents to a vessel or organ wall or performing transmyocardial revascularization or cryo-ablation, such as described in the above-referenced patents.

[0063] Proximal end 2 preferably includes storage device 2a, such as a memory chip, RFID tag or bar code label, which stores data that may be used in computing a multi-dimensional force vector, as described herein after. Alternatively, storage device 2a need not be affixed to proximal end 2, but instead could be a separate item, e.g., packaging, individually associated with each catheter. Proximal end 2 may be manipulated manually or automatically to cause a desired amount of articulation or flexion of distal extremity 5 using mechanisms which are per se known in the art, such as pull wires or suitably configured electroactive polymers. Catheter 1 also may be advanced, retracted and turned manually or automatically.

[0064] Distal extremity 5 of catheter 1 comprises a deformable body having at least two optical fiber sensors that extend

proximally and are coupled to console 3 via proximal end 2 and cable 4. More preferably, catheter 1 includes three optical fiber sensors disposed therein. In addition, control signals to and from the end effector(s) in distal extremity 5 are transmitted via suitable components of cable 4 to console 3, to a tactile component of proximal end 2. As will be apparent, the nature of cable 4 depends on the nature of the end effectors disposed in distal extremity 5 of catheter 1.

[0065] Console 3 comprises electronic and optical components to drive the optical fiber sensors and to interpret the output signals therefrom. Console 3 further includes processing logic 6, such as a programmed general purpose microprocessor or application-specific integrated circuit, which receives an output signal corresponding to wavelength changes manifested in the optical fiber sensors due to forces applied to the distal extremity of the deformable body. Processing logic 6 computes a multi-dimensional force vector based upon that output signal and a matrix of physical characteristics of the individual deformable body, as described in detail below. Console 3 preferably also includes means to manifest an output from the load sensing system, such as a visual display or an auditory device. Alternatively, console 3 may output a signal for display on a separate monitor.

[0066] Referring now to FIGS. 2 and 3, catheter 1 preferably has at least two optical fiber sensors 7 disposed within it, so that deformation of distal extremity 5 is transferred to the sensors 7. Two optical fiber sensors may be employed so long as the location of the neutral axis of the distal extremity is known or determined during manufacture. More preferably, distal extremity 1 includes at least three optical fiber sensors, and comprises a molded, machined or extruded material, such as typically are used in making guide wires or catheters. To ensure that the optical fibers form an integral part of catheter 1, the optical fibers may be affixed within the distal extremity using adhesive or other means as, for example, overmolding or co-extrusion. In FIG. 3, optical fibers 7 are glued into distal extremity 5 using adhesive 8.

[0067] Preferably, catheter 1 comprises a liquid crystal polymer ("LCP") that has a small positive or even negative coefficient of thermal expansion in the direction of extrusion. A variety of liquid crystal polymers are known in the art and such materials may be coated with parylene or a metallic coating to enhance resistance to fluid absorption.

[0068] Where three optical fiber sensors are employed, optical fibers 7 are disposed in distal extremity 5 so that the optical fiber sensors are not co-planar, i.e., are not situated in a single plane. Illustratively, the optical fibers are disposed at the apices of an equilateral triangle centered on the longitudinal axis of the catheter. Other configurations are possible, so long as optical fibers experience different degrees of bending and elongation during deformation of distal extremity 5. Optical fiber sensors 7 may be chosen from among a Fiber Bragg Grating (FBG), a Long Period Grating (LPG), an Intrinsic Fabry-Perot Interferometer (IFPI), an Extrinsic Fabry-Perot Interferometer (EFPI), a two, three or four arm Michelson interferometer (MI), a Brillouin scattering strain sensor, or intensity-based fiber optic strain sensor.

[0069] Referring now to FIG. 4, catheter 1 is depicted housing three optical fibers 7 having FBG or LPG strain sensors 9 disposed in distal extremity 5. An FBG sensor is an interferometer in which a stable Bragg grating is permanently impressed (e.g., photo-etched) into the core of the fiber. The region of periodic variation in the index of refraction of the fiber core acts as a very narrowband reflection filter that

reflects light having a predetermined Bragg wavelength. Light therefore is reflected from the FBG in a narrow spike with a center wavelength that is linearly dependent on the Bragg wavelength and the mean index of refraction of the core. Consequently, deformations that alter the grating characteristics result in a shift in the reflected Bragg wavelength.

[0070] An LPG is similar in construction to an FBG, and comprises a single mode fiber having periodic index modulation of the refractive index of the fiber core with a much longer period than an FBG. Use and operation of a catheter employing LPGs rather than FBGs is similar to that described below.

[0071] During use of the apparatus, the distal extremity of catheter 1 is compressed and bent due to loads imposed by contacting the tissue of the organ. The portions of optical fibers 7 that are situated in the distal extremity also are deformed but in a varying degrees according to their respective positions in the distal extremity. In addition, the distal extremity may be deflected by deflecting a more proximal portion of the catheter using any of a variety of previously-known catheter deflection mechanisms, such as described in U.S. Pat. No. 4,960,134 to Webster, which is incorporated herein by reference. In this case, the apparatus will compute the force with which the distal extremity contacts the tissue of the organ or vessel.

[0072] The initial calibration of the FBG sensors, i.e., the average wavelength reflected from the Bragg grating in the absence of any applied forces (referred to as the “Bragg wavelength”) is determined from grating characteristics impressed during manufacture of the optical fiber. Any deviations from the Bragg wavelength are proportionately related to an exact parameter, such as strain. In the embodiment of FIG. 4, the Bragg grating allows the deformation (elongation or contraction) of each of optical fibers 7 to be quantified by measuring the change in wavelength of the light reflected by the Bragg grating.

[0073] The foregoing information, together with known physical properties of the distal extremity of the catheter, enable processing logic 6 of console 3 to calculate the components of a multidimensional force vector with appropriate algorithms. The force vector then may be displayed or otherwise manifested, for example, as a graphic on a display screen or by varying the pitch emitted from an auditory device housed in or associated with console 3.

[0074] Still referring to FIG. 4, one of optical fibers 7 preferably extends beyond the others and includes second FBG (or LPG) 10 for measuring the temperature of the front end of the distal extremity. Temperature changes at the front end of the distal extremity may arise, e.g., due to operation of an ablation electrode, and will cause a change in the associated Bragg wavelength. By knowing the physical properties of the fiber and measuring the wavelength of the light reflected by the grating, processing logic 6 may compute the temperature at the level of the distal extremity, for example, to monitor tissue ablation progress.

[0075] Referring again to FIG. 1, console 3 comprises a laser, preferably a tunable laser diode, arranged to inject a beam of light into the optical fibers through cable 4, and a photodetector that detects variations in the characteristics of the reflected light beam due to deformations imposed on the strain sensors and distal extremity 5. Preferably, console 3 includes a Fiber Bragg Grating Demodulator.

[0076] In such a system, each of the optical fiber sensors has a Bragg grating with a different wavelength, and which

therefore responds in a specified range of frequency. A tunable laser is coupled to all of the optical fiber sensors and scans a certain frequency several times per second. A photodiode records the wavelength change for each Bragg grating when the frequency of the laser centers on the grating frequency. In this manner, each of the optical fiber sensors may be interrogated as the tunable laser scans through the grating frequencies of the sensors.

[0077] Further in accordance with the principles of the present invention, processing logic 6 is programmed to compute a two- or three-dimensional force vector from the output of the Fiber Bragg Grating Demodulator. The theory underlying these computations is now described.

[0078] For apparatus having three fiber optic Bragg strain sensors embedded within the distal extremity of the catheter, the total strain may be computed using:

$$\begin{bmatrix} \varepsilon_{1,t} \\ \varepsilon_{2,t} \\ \varepsilon_{3,t} \\ \Delta T_t \end{bmatrix} = \begin{bmatrix} C_\varepsilon & 0 & 0 & C_{\varepsilon T} \\ 0 & C_\varepsilon & 0 & C_{\varepsilon T} \\ 0 & 0 & C_\varepsilon & C_{\varepsilon T} \\ 0 & 0 & 0 & C_T \end{bmatrix} \cdot \begin{bmatrix} \lambda_{1,t} \\ \lambda_{2,t} \\ \lambda_{3,t} \\ \lambda_{4,t} \end{bmatrix} - \begin{bmatrix} \lambda_{1,r} \\ \lambda_{2,r} \\ \lambda_{3,r} \\ \lambda_{4,r} \end{bmatrix} \quad (1.1)$$

$$\varepsilon_t = C \cdot (\lambda_t - \lambda_r) \quad (1.1a)$$

[0079] Where: r—time when reference (zero) measurement is set

[0080] t—time relative to reference time

[0081] $\lambda_{i,r}$, i=1,4—reference wavelengths of Bragg-gratings

[0082] $\lambda_{i,t}$, i=1,4—wavelengths of Bragg-gratings at time t

[0083] $\varepsilon_{i,t}$, i=1,3—total strain values at time t

[0084] ΔT_t —Temperature change at time t

[0085] C_ε —coefficient of linearity between the wavelength and strain

[0086] $C_{\varepsilon T}$ —coefficient of temperature compensation of the Bragg-grating

[0087] C_T —coefficient of linearity between the wavelength and temperature

[0088] λ_r —Matrix (vector) of Bragg-gratings reference wavelengths

[0089] λ_t —Matrix (vector) of Bragg-gratings wavelengths at time t

[0090] ε_t —Matrix (vector) of total strain and temperature changes

[0091] C—Strain transducer and compensation matrix

[0092] The total strain includes a component due to thermal expansion of the distal extremity arising from the difference between the measured temperature of the distal extremity and a predetermined reference temperature. The elastic strain, which is a function of the applied force, therefore may be calculated using:

$$\begin{bmatrix} \varepsilon_{el1,t} \\ \varepsilon_{el2,t} \\ \varepsilon_{el3,t} \end{bmatrix} = \begin{bmatrix} 1 & 0 & 0 & -\alpha_{Tc} \\ 0 & 1 & 0 & -\alpha_{Tc} \\ 0 & 0 & 1 & -\alpha_{Tc} \end{bmatrix} \cdot \begin{bmatrix} \varepsilon_{1,t} \\ \varepsilon_{2,t} \\ \varepsilon_{3,t} \\ \Delta T_t \end{bmatrix} \quad (1.2)$$

$$\varepsilon_{el,t} = \alpha_T \cdot \varepsilon_t \quad (1.2a)$$

[0093] Where: $\epsilon_{eli,t}$, $i=1,3$ —elastic strain values at time t

[0094] α_T —Thermal expansion coefficient of catheter material (PEEK)

[0095] $\epsilon_{el,t}$ —Matrix (vector) of elastic strain at time t

[0096] α_T —Temperature reduction matrix

$$(1.1a) \wedge (1.2a) \Rightarrow \epsilon_{el,t} = \alpha_T \cdot C \cdot (\lambda_r - \lambda_r) \quad (1.3)$$

[0097] The elastic strains are related to the internal forces experienced by the optical fiber sensors as a function of both the physical dimensions of, and the material properties of, the distal extremity:

$$\begin{bmatrix} \epsilon_{el1,t} \\ \epsilon_{el2,t} \\ \epsilon_{el3,t} \end{bmatrix} = \begin{bmatrix} 1 & y_1 & -x_1 \\ 1 & y_2 & -x_2 \\ 1 & y_3 & -x_3 \end{bmatrix} \cdot \begin{bmatrix} \frac{1}{E_{ten} \cdot A} & 0 & 0 \\ 0 & \frac{1}{E_{flex} \cdot I_x} & 0 \\ 0 & 0 & \frac{1}{E_{flex} \cdot I_y} \end{bmatrix} \cdot \begin{bmatrix} N_{z,t} \\ M_{x,t} \\ M_{y,t} \end{bmatrix} \quad (2.1)$$

$$\epsilon_{el,t} = G \cdot \delta \cdot I_{F,t} \quad (2.1a)$$

[0098] Where: x_i and y_i , $i=1,3$ —coordinates of Bragg gratings with respect to center of gravity of the catheter cross-section

[0099] E_{ten} —Equivalent tension/compression Young modulus of catheter

[0100] E_{flex} —Equivalent flexural Young modulus of catheter

[0101] I_x —Moment of inertia with respect to x axis

[0102] I_y —Moment of inertia with respect to y axis

[0103] $N_{z,t}$ —Normal force in direction of z axis at time t

[0104] $M_{x,t}$ —Bending moment with respect to x axis at time t

[0105] $M_{y,t}$ —Bending moment with respect to y axis at time t

[0106] G —Geometry matrix

[0107] δ —Matrix of flexibility

[0108] $I_{F,t}$ —Matrix (vector) of internal forces at time t

[0109] Equation (2.1) may be rearranged to solve for the internal forces as a function of the elastic strain. The elastic strain from equation (1.3) may then be substituted into the rearranged matrix system to compute the internal forces as a function of the elastic strain, as shown in Equation (2.3) below:

$$(2.1) \Rightarrow \begin{bmatrix} N_{z,t} \\ M_{x,t} \\ M_{y,t} \end{bmatrix} = \quad (2.2)$$

$$\begin{bmatrix} E_{ten} \cdot A & 0 & 0 \\ 0 & E_{flex} \cdot I_y & 0 \\ 0 & 0 & E_{flex} \cdot I_x \end{bmatrix} \cdot \begin{bmatrix} 1 & y_1 & -x_1 \\ 1 & y_2 & -x_2 \\ 1 & y_3 & -x_3 \end{bmatrix}^{-1} \cdot \begin{bmatrix} \epsilon_{el1,t} \\ \epsilon_{el2,t} \\ \epsilon_{el3,t} \end{bmatrix}$$

$$(2.1a) \Rightarrow I_{F,t} = S \cdot G^{-1} \cdot \epsilon_{el,t}$$

[0110] Where: $S = \delta^{-1}$ —Stiffness matrix

$$(1.3) \wedge (2.1a) \Rightarrow I_{F,t} = S \cdot G^{-1} \cdot \alpha_T \cdot C \cdot (\lambda_r - \lambda_r) \quad (2.3)$$

[0111] It remains only to relate the internal forces experienced by the optical fiber sensors to the external contact

forces actually exerted on the distal extremity by the tissue wall. These forces are computed based on the positions of the optical fiber sensors from the exterior wall of the distal extremity, assuming the catheter material is substantially incompressible:

$$\begin{bmatrix} F_{x,t} \\ F_{y,t} \\ F_{z,t} \end{bmatrix} = \begin{bmatrix} 0 & 0 & -\frac{1}{d} \\ 0 & \frac{1}{d} & 0 \\ -1 & 0 & 0 \end{bmatrix} \cdot \begin{bmatrix} N_{z,t} \\ M_{x,t} \\ M_{y,t} \end{bmatrix} \quad (3.1)$$

$$F_t = d \cdot I_{F,t} \quad (3.1a)$$

[0112] Where: $F_{x,t}$ —Touching external transversal force at time t , in direction of x axis (with opposite sense)

[0113] $F_{y,t}$ —Touching external transversal force at time t , in direction of y axis (with opposite sense)

[0114] $F_{z,t}$ —Touching external normal force at time t , in direction of z axis (with opposite sense, compression is positive)

[0115] d —distance between the touching point of lateral forces and the cross-section with sensors (along z axis)

[0116] F_t —Matrix of touching external forces at time t

[0117] d —Matrix of conversion

$$(2.3) \wedge (3.1a) \Rightarrow F_t = d \cdot S \cdot G^{-1} \cdot \alpha_T \cdot C \cdot (\lambda_r - \lambda_r) \quad (3.2)$$

$$F_t = K_\lambda \cdot (\lambda_r - \lambda_r) = K_\lambda \cdot \lambda_r - F_r \quad (3.3)$$

[0118] Where: K_λ —Force transducer matrix,

$$K_\lambda = d \cdot S \cdot G^{-1} \cdot \alpha_T \cdot C \quad (3.4)$$

[0119] F_r —Reference force matrix (vector),

$$F_r = K_\lambda \cdot \lambda_r \quad (3.5)$$

[0120] Solution of equations (3.1) to (3.5) provides the normal and transverse forces applied to the external surface of the deformable body, i.e., $F_{norm,t} = F_{z,t}$ and $F_{trans,t}$ —square root ($F_{x,t}^2 + F_{y,t}^2$). The angle γ_t of application of the transverse force may be computed from Table I:

TABLE I

$F_{x,t}$	$F_{y,t}$	γ_t
≥ 0	≥ 0	$\arcsin(F_{y,t}/F_{trans,t})$
< 0	≥ 0	$\Pi - \arcsin(F_{y,t}/F_{trans,t})$
< 0	< 0	$\Pi - \arcsin(F_{y,t}/F_{trans,t})$
≥ 0	< 0	$2 * \Pi + \arcsin(F_{y,t}/F_{trans,t})$

[0121] Many of the values employed in equations (1.1) to (3.5) are related to the material properties of the distal extremity or optical fiber sensors, such as the Bragg wavelengths, thermal expansion coefficients and elastic moduli. Other values, such as the distances between the optical fiber sensors and the external surface of the distal extremity may be subject to variations as a consequence of the manufacturing process employed.

[0122] To ensure the accuracy of the computed force vector, specific information for each catheter may be stored in storage device 2a. Generally, the information make take the form of a data file that is input to console 3 prior to use of the catheter. For example, storage device 2a may comprise a

memory chip associated with cable 4 in which such information is stored, or a bar code or a RFID tag located on proximal end 2 of the catheter or the packaging for the catheter. Alternatively, data specific to an individual catheter may be uploaded to console 3 from an item of removable storage (e.g., CD) or via secure download from the manufacturer's website.

[0123] The information specific to each catheter may be obtained during a calibration step, conducted during manufacture of the catheter, by subjecting the distal extremity of the catheter to a series of known forces. In this case, the foregoing equations may be collapsed so the normal and transverse forces may be computed directly from a force-to-wavelength conversion matrix:

$$F(t)=K(\lambda(t)-\lambda_0) \quad (4.0)$$

where:

[0124] $F(t)$ is the vector of forces $[F_{x,t}, F_{y,t}, F_{z,t}]$,

[0125] $\lambda(t)$ is the vector of wavelengths $[\lambda_{1,t}, \lambda_{2,t}, \lambda_{3,t}]$ measured for the individual sensors,

[0126] λ_0 is the vector of wavelengths $[\lambda_1^0, \lambda_2^0, \lambda_3^0]$ measured for the individual sensors with zero applied force, and

[0127] K is a matrix computed when the deformable body is subjected to the series of known forces.

[0128] During the calibration step of manufacture, the catheter is subjected to the following forces in series: (1) a purely axial force of known magnitude F' ; (2) a lateral force of known magnitude F'' ; and (3) a lateral force of known magnitude F''' applied 90 degrees to the orientation of force F'' . When all of the forces F' , F'' , F''' and wavelengths are known, the force-to-strain conversion matrix K may be computed as:

$$K=F(\lambda(t)-\lambda_0)^{-1} \quad (5.0)$$

or:

$$\begin{bmatrix} F_x & F'_x & F''_x \\ F_y & F'_y & F''_y \\ F_z & F'_z & F''_z \end{bmatrix} \begin{bmatrix} (\lambda_1 - \lambda_1^0) & (\lambda'_1 - \lambda_1^0) & (\lambda''_1 - \lambda_1^0) \\ (\lambda_2 - \lambda_2^0) & (\lambda'_2 - \lambda_2^0) & (\lambda''_2 - \lambda_2^0) \\ (\lambda_3 - \lambda_3^0) & (\lambda'_3 - \lambda_3^0) & (\lambda''_3 - \lambda_3^0) \end{bmatrix}^{-1} = \begin{bmatrix} a_{11} & a_{12} & a_{13} \\ a_{21} & a_{22} & a_{23} \\ a_{31} & a_{32} & a_{33} \end{bmatrix} \quad (5.1)$$

[0129] Force-to-strain conversion matrix K then may be stored in storage device 2a associated with the corresponding device, as described herein above. The values of the force-to-conversion matrix then may be input to console 3 when the catheter is coupled to the console using a bar code reader, input pad or direct electrical connection through cable 4. Once matrix K is provided for a given distal extremity, the normal force, transverse force and angle of application of the transverse force may be computed as described above and using Table I.

[0130] The values for the normal force, transverse force and angle of application of the transverse force, computed as described above, may be output as numerical values to a display monitor that forms part of console 3 or which is associated with console 3. In addition, a graphic including a variable size or colored arrow may be displayed pointing at a position on the circumference of a circle to visualize the magnitude and direction of the transverse force applied to the distal extremity. By monitoring this display, the operator may

continuously obtain feedback concerning the contact forces applied to the distal extremity of the catheter.

[0131] Referring now to FIG. 5, an alternative embodiment is described in which optical fiber strain sensors 7 comprise Intrinsic Fabry-Perot Interferometers (IFPI). One of the optical fibers is extended and comprises a second IFPI sensor 13 for measuring the temperature of the front end of the distal extremity.

[0132] An IFPI comprises a single mode optical fiber having segment having reflectors 12 disposed at either end to define optical cavity 11. The reflectors may comprise semi-reflective mirror surfaces formed in the fiber, or alternatively may comprise two FBGs. Light emitted from a laser diode disposed in console 3 impinges upon the proximal reflector and is partially reflected back at specific wavelengths 14. Light passing through the proximal reflector and impinging upon the distal reflector is also reflected back. The two reflected beams result in constructive and destructive interferences that are detected by a photodetector disposed in console 3.

[0133] A variation in strain or temperature changes the optical length of optical cavity 11 and sensor 13, and influences the reflection characteristics from which relative deflections of the optical fibers may be computed. This information in turn permits computation of the force vector imposed upon distal extremity 5 due to contact with the tissue of the wall of the organ or vessel.

[0134] FIG. 6 illustrates a further alternative embodiment of the distal extremity of catheter 1 and contains three Extrinsic Fabry-Perot interferometer (EFPI) sensors. One of the optical fibers extends beyond the others and comprises a second EFPI sensor 17 to measure the temperature of the front end of the distal extremity. An EFPI sensor comprises optical cavity 11 formed by hollow capillary tube 15 and cut ends 16 of the optical fiber. The hollow capillary tube contains air. Operation of the EPFI is similar to that described above for the IFPI, except that the cut ends of the fiber act as the reflectors to reflect specific wavelengths 18. Light reflected from cut ends 16 result in two beams that constructively and destructively interfere. A variation in strain or temperature changes the length of the optical cavity and influences the reflection characteristics.

[0135] FIG. 7 illustrates a further alternative embodiment of the distal extremity of catheter 1, wherein the distal extremity contains three optical fibers 7 that form a Michelson interferometer. Each optical fiber 7 includes reflector 19 at its distal extremity; the fibers are coupled at their proximal ends by optical coupler 20. A wave is injected into fiber 21 from a laser diode disposed in console 3 and is separated by coupler 20 into each of the optical fibers ("arms") of the interferometer. The coupler 20 combines the back reflected light from each arm. Using a coherent or low coherence interferometer, variations in the relative phases of the reflected light from the different fibers are measured to compute the strain experienced by the distal extremity of catheter 1. Based upon the computed strain, the contact force between the distal extremity and the tissue of the organ or vessel wall may be determined.

[0136] Referring now to FIG. 8, an embodiment wherein the optical fibers comprise high resolution Brillouin sensors is described. Brillouin sensors use the principle of scattering 22 that is an intrinsic phenomenon of optical fiber. This phenomenon results from the interaction between the light and the phonons (pressure wave) present in the fiber. Wave 23 is

backscattered with a shift in optical frequency relative to the injected wave. One of the optical fibers 7 extends beyond the others and comprises a second Brillouin scattering sensor 24 to measure the temperature at the front end of the distal extremity. A variation in strain or temperature changes the shift in optical frequency. Using impulsion, phase modulation or other techniques, it is possible to select different locations 26 along the fiber and to measure the state of strain at these locations.

[0137] Referring to FIGS. 9 and 10, further embodiments of the present invention are described that employ intensity-type optical fiber sensors. More specifically, FIG. 9 illustrates use of reflection intensity sensors while FIG. 10 illustrates use of microbending intensity sensors.

[0138] In FIG. 9, reflection intensity sensors comprise connection zones 25 within optical fibers 7. Under the effect of a strain caused by deformation of the distal extremity, or a temperature variation, connection zones 25 modulate the amplitude of the optical wave 26 that is transmitted and/or reflected. The variation in intensity of the reflected light is measured by apparatus, which is per se known. An additional optical fiber also may be provided to perform temperature measurement.

[0139] In FIG. 10, microbending intensity sensors comprise connection zones 27 disposed along the length of optical fibers 7. Connection zones 27 may be obtained by introducing microbendings in the fibers. Under the effect of a strain caused by deformation of the distal extremity, or a temperature variation, connection zones 27 modulate the amplitude of the optical wave 28 that is transmitted and/or reflected. The variation in intensity of the reflected light is measured by apparatus, which is per se known.

[0140] According to a preferred embodiment, the three optical fibers may be assembled with each other to form an integral part, as depicted in FIG. 11, or embedded with an adhesive or other suitable deformable material to form cylindrical element 29, as depicted in FIG. 12. This arrangement provides a very small solid assembly that may in turn be affixed within a lumen of a catheter of otherwise conventional construction, as depicted in FIG. 13, while also protecting the optical fibers from breakage. In accordance with the principles of the present invention, bundling the fibers as shown in FIGS. 11-13 ensures that all three of the optical fibers are not co-planar.

[0141] Referring now to FIGS. 4 and 14, the distal extremity of an exemplary ablation catheter utilizing the load sensing capability of the present invention is described. Catheter 1 includes electrodes 30, 31 and 32 and is coupled to front end 33 having irrigation ports 34. Electrodes 30, 31, 32, 33 are provided according to the function of the specific application for the catheter, for example, endocavity electrocardiogram, radiofrequency ablation, etc. Front end 33 also may be an electrode. Sensor 35 also may be provided that provides three-dimensional positioning of the distal extremity of the catheter, with sensor 35 being based upon electromagnetic, magnetic, electric, ultrasound principles.

[0142] The distal extremity of catheter 1 includes at least three fiber optic sensors 9 configured as described hereinabove. One of the optical fibers extends beyond the others and includes, for example, second Bragg grating 10 that serves as a temperature sensor. Bragg grating 10 is received within front end 33 and may be used to compute temperature changes in front end 33 resulting from operation of the electrode. Irrigation ports 34 communicate with one or more

channels situated inside the catheter and may be used to deliver a cooling solution, e.g., saline, to the distal extremity of the catheter during operation of the front end electrode to lower the temperature of the front end and control the ablation of tissue.

[0143] Although front end 33 is illustratively described as configured for performing radiofrequency ablation, other tissue ablation or treatment end effectors could be used, such as laser, ultrasound, radiation, microwave and others. Furthermore, other therapeutic means such as the injector of medication, stem or other types of cells may also be situated in the head of the catheter.

[0144] With respect to FIG. 15, a further alternative embodiment is described wherein a fourth optical fiber is used to measure the temperature of the distal extremity in the vicinity of the other optical fiber strain sensors. Because the material of the distal extremity of catheter 1 may be sensitive to temperature variations, a change of temperature of the distal extremity may result in expansion or contraction of the distal extremity and the embedded optical fibers. This effect may result in computation of a false force vector. Accordingly, fourth optical fiber 7 is slidably disposed in distal extremity 1 so that it is not affected by temperature induced expansion or contraction of the distal extremity of the catheter, and thus provides a reference measurement. If the temperature of the sensor body is known, however, such as by using a fourth optical fiber, thermal expansion or compression of the distal extremity may be compensated in the computation of the force vector.

[0145] Referring now to FIG. 16, an alternative embodiment of apparatus utilizing the load sensing system of the present invention is described. Apparatus 40 includes catheter 41 having distal extremity 42 and proximal end 43 coupled to console 45 via cable 44. Construction and operation of components 41-45 is similar to that described above for the embodiment of FIG. 1.

[0146] In accordance with one aspect of the present invention, apparatus 40 of FIG. 16 further includes robotic control system comprising controller 46, input and display device 47 and actuator 48. Actuator 48 is coupled to catheter 41 to manipulate the catheter responsive to commands generated by programmed microprocessor 46. Controller 46 is programmed via instructions input via input and display device 47, and the operation of the actuator 48 may be monitored via a display portion that device 47. Controller 46 is coupled to console 45 to receive the output of the load sensing system of the present invention, and to use that information to control manipulation of catheter 41 and actuator 48. Console 45 also may receive an input from controller 46 that is used to determine when the end effector of catheter 41 is operated.

[0147] For example, catheter 41 may comprise an electrophysiology catheter designed to map electrical potentials within a patient's heart. In this case, distal extremity 42 may include a series of mapping and ablation electrodes as described herein above with respect to FIG. 14. As described above, previously known methods of mapping electrical potentials within a patient's heart is a time consuming activity, because the clinician determines engagement of with the tissue wall by tactile feedback through the catheter shaft or using impedance measurements.

[0148] In accordance with the principles of the present invention, actuator 48 comprises a multi-axis tool capable of advancing and rotating the catheter within the patient's heart. Controller 46 may be programmed to manipulate the catheter

until the contact force encountered by distal extremity 42 falls within a predetermined range, as determined via monitoring by console 45. Once the contact force is determined to fall within the predetermined range, the electrical potential may be measured and recorded. Controller 46 then may reposition the catheter as required to map other desired portions of the patient's heart.

[0149] Advantageously, because the contact forces applied by the distal extremity can be controlled within desired ranges, the risk of deforming the tissue wall is reduced. Accordingly, if a three dimensional locator system also is provided in the catheter, such as described above, accurate registration of the measured values and the spatial locations of the measurement points may be obtained. The load sensing system of the present invention similarly may be integrated into a treatment system, for example, including the ablation electrode described above with respect to FIG. 14, in which the ablation electrode may be energized to ablate tissue only when the contact force between the distal extremity and the tissue wall exceeds a predetermined minimum value or falls within a predetermined range.

[0150] In addition, where distal extremity 42 of catheter 41 is articulable, controller 46 also may provide a signal to console 45 that adjusts the articulation of the distal extremity. In this manner, the load sensing system of the present invention may be configured not only to serve as part of a feedback loop to an external controller, but may itself accept an external control signal that controls operation of an end effector of the catheter.

[0151] Referring now to FIG. 17, a further alternative embodiment of an ablation catheter utilizing the load sensing features of the present invention is described. Applicant has observed that some polymers routinely employed in catheter construction, such as polyethylene have a relatively high coefficient of thermal expansion, and a tendency to absorb moisture when exposed to bodily fluids.

[0152] Although the dimensional changes resulting from moisture absorption and temperature fluctuations may be small, these environmental factors may introduce artifacts into the forces computed by the apparatus. Moreover, the environmental effects may not be entirely removed by use of an additional optical fiber sensor, such as described with respect to FIG. 15. To address the temperature fluctuation issue, a tube having a low thermal expansion coefficient is disposed in the distal extremity of the catheter in the vicinity of the sensor portions of the optical fibers.

[0153] Referring again to FIG. 17, apparatus 50 comprises catheter 51 having proximal end 52 coupled via cable 53 to console 54 having processor 55. Apparatus 50 further comprises distal extremity 56 attached to the distal end of catheter 51 and includes electrode 57 having irrigation ports 58 for cooling the tissue during an RF ablation procedure. Proximal end 52, cable 53, console 54, and processor 55 are similar in design and construction to proximal end 2, cable 4, console 3 and processor 6 of the embodiment of FIG. 1, respectively, which are described in detail above. Apparatus 50 differs mainly in the construction of distal extremity 56, as described below.

[0154] Referring now to FIGS. 18 and 19, subassembly 60 disposed within distal extremity 56 of apparatus 50 is described. Subassembly 60 comprises irrigation tube 61 coupled at proximal end 52 to an infusion port (not shown) and at distal end 62 to irrigation ports 58 of front end 63.

[0155] Front end 63 preferably is metallic and acts as an ablation electrode, and includes irrigation ports 58 in fluid communication with the interior of irrigation tube 61, so that fluid injected via the infusion port exits through irrigation ports 58.

[0156] In FIG. 19, subassembly 60 is disposed within polymeric housing 64, shown partially cut-away for ease of understanding. Optical fiber sensors 65 are arranged around the circumference of irrigation tube 61, preferably spaced 120 degrees apart. Sensors 65 are similar in design and construction to optical fiber sensors 7 of the preceding embodiments, and may be configured to measure strain in any appropriate manner, such as described above and depicted in FIGS. 4-10. Preferably, sensors 65 are Bragg Gratings.

[0157] In accordance with one aspect of the invention, irrigation tube 61 preferably comprises proximal portion 66 and distal portion 67. Proximal portion 66 preferably comprises a polymer and more preferably comprises a thin polyimide tube, such as made from Kapton, available from DuPont, and extends from proximal end 52 to within about 1 cm of distal end 62.

[0158] Distal portion 67 couples proximal portion 66 to front end 63. Distal portion 67 preferably is electrically conductive, so as to conduct electrical current to front end 63, for example, by wire 59 coupled to the proximal end of proximal portion 66. Preferably, distal portion 67 is formed of a material having a relatively low coefficient of thermal expansion compared to the rest of catheter 51. Distal portion 67 preferably also has a Young's modulus of elasticity such that, when configured as a thin tube, its axial deformation under an applied load is sufficient to obtain a force resolution with the optical fiber sensors 65 of 1 gram. In a preferred embodiment, distal portion 67 comprises titanium and has a length of approximately 1 cm, whereas the length of the measurement regions of optical fibers 65 is about 4 mm.

[0159] Referring now to FIGS. 19 and 20, housing 64 is described in greater detail. Housing 64 preferably comprises a polymer and extends over distal portion 67 of irrigation tube 61 to enclose and protect the measurement regions of optical fiber sensors 65. Housing 64 is bonded to distal portion 67, e.g., with glue or other known attachment means, so that the distal end of the housing does not contact front end 63, but instead forms gap 68.

[0160] Housing 64 includes central channel 69 configured to receive distal portion 67 of subassembly 60, and may include grooves 70 on the exterior surface of the housing 64 to accept wires that electrodes on the exterior of housing 64 to proximal end 52. Housing 64 also includes ribs 71 that prevent the housing from directly contacting optical fiber sensors 66. Housing 64 further includes stepped diameter region 72 that facilitates joining the housing to the proximal portion of catheter 51.

[0161] As described above, apparatus 50 may be configured to include the capability to deflect the distal extremity of catheter 51 using any of variety of well-known mechanisms, such as pull-wires. More particularly, referring to FIG. 21, an illustrative embodiment of a deflectable catheter shaft suitable for use with subassembly 60 of FIGS. 18 and 19 is described.

[0162] Catheter shaft 80 includes handle 81, elongated shaft 82 and deflectable region 83. Shaft 82 preferably comprises braided wire tube 84 embedded within biocompatible polymer 85. Deflectable region 83 preferably comprises flexible catheter material 86 having wire coil 87 embedded with

it. Pull wire **88** is coupled to anchor ring **89** disposed at distal end **90** of deflectable region **83**, and extends through coil spring **91** to handle **81**. Electrical wires **92**, irrigation tube **93** (corresponding to irrigation tube **61** in FIG. **18**) and the optical fibers (not shown) extend from handle **81** through anchor ring **89** to the housing of the distal extremity.

[**0163**] Stepped diameter region **72** of housing **64** engages the distal end of catheter shaft **80**, so that housing **64** and electrode **57** are disposed distal to anchor ring **89**. In this manner, deflection of deflectable region **83** does not impact the strains computed by the optical fiber sensors used to compute contact forces between the distal extremity of the catheter and the wall of the vessel, tissue or organ.

[**0164**] It will be appreciated that other embodiments of an ablation catheter may employ other features discussed elsewhere in this application. For example, an additional sensor may be added to apparatus **50** for measuring temperature using the above-described principles.

[**0165**] In summary, use of optical fiber strain sensors permits computation of a multi-dimensional force vector that arises during contact of the distal extremity of the catheter with the wall of the tissue, organ or vessel. When such information is combined with a 3D positioning sensor, precise mapping may be obtained to permit diagnosis or treatment of tissue at an optimal applied force. The small size of the optical fiber strain sensors and high resolution of measurements obtained by these devices allows highly precise measurements to be obtained even in environments that are humid and subject to electromagnetic interference.

[**0166**] While preferred illustrative embodiments of the invention are described above, it will be apparent to one skilled in the art that various changes and modifications may be made therein without departing from the invention. The appended claims are intended to cover all such changes and modifications that fall within the true spirit and scope of the invention.

1-15. (canceled)

16. A method to determine a force, comprising:
receiving a plurality of signals from a plurality of sensors located at a distal end of a catheter;
processing the plurality of signals to determine a force vector applied to the distal end of the catheter;
outputting a signal indicative of the force vector.

17. The method according to claim **16**, wherein the force vector is a three-dimensional force vector.

18. The method according to claim **16**, wherein the plurality of signals comprises three signals and wherein the plurality of sensors comprises three sensors.

19. The method according to claim **16**, further comprising receiving specific information pertaining to the catheter.

20. The method according to claim **19**, wherein the specific information comprises calibration information acquired during a calibration step and stored in a memory of the catheter.

21. The method according to claim **20**, wherein the calibration information comprises a force conversion matrix.

22. The method according to claim **16**, wherein the force vector comprises a normal force, a transverse force, and an angle of application of the transverse force.

23. The method according to claim **16**, wherein outputting a signal comprises outputting a graphic including a representation of a magnitude and a direction of a transverse force applied to the distal end of the catheter.

24. The method according to claim **16**, further comprising determining a temperature of the distal end of the catheter by changes in the plurality of signals received from the plurality of sensors.

25. The method according to claim **16**, wherein processing the plurality of signals further comprises accounting for changes in temperature at the distal end of the catheter.

26. The method according to claim **16**, wherein processing the plurality of signals comprises a continuous process and outputting a signal comprises a continuous process such that the outputted signal of the force vector applied to the distal end of the catheter can be used to guide or control a use of the catheter.

27. A system for determining a force, comprising:
a console comprising a processing logic,

wherein the console is configured to couple to a catheter and to receive an output of at least two sensors, wherein the processing logic is configured to determine a force vector from the output of the at least two sensors, and wherein the console is configured to output a signal indicative of the force vector.

28. The system according to claim **27**, wherein the processing logic is configured to determine a three-dimensional force vector.

29. The system according to claim **27**, wherein the at least two sensors comprises three sensors.

30. The system according to claim **27**, wherein the console is further configured to receive specific information pertaining to the catheter.

31. The system according to claim **30**, wherein the specific information comprises calibration information acquired during a calibration step and stored in a memory of the catheter.

32. The system according to claim **27**, wherein the force vector comprises a normal force, a transverse force, and an angle of application of the transverse force.

33. The system according to claim **27**, wherein the outputted signal comprises a graphic including a representation of a magnitude and a direction of a transverse force applied to the distal end of the catheter.

34. The system according to claim **27**, wherein the processing logic is further configured to determine a temperature of the distal end of the catheter by tracking changes in the output received from the at least two sensors.

35. The system according to claim **27**, wherein the processing logic is further configured to account for changes in temperature at the distal end of the catheter by tracking changes in the output received from the at least two sensors.

* * * * *

专利名称(译)	具有光纤传感能力的医疗设备系统		
公开(公告)号	US20150216612A1	公开(公告)日	2015-08-06
申请号	US14/573666	申请日	2014-12-17
[标]申请(专利权)人(译)	ST Jude Medical公司卢森堡控股S A R L		
申请(专利权)人(译)	ST.犹达医疗LUXEMBOURG HOLDING S.A.R.L.		
[标]发明人	LEO GIOVANNI AEBY NICOLAS INAUDI DANIELE		
发明人	LEO, GIOVANNI AEBY, NICOLAS INAUDI, DANIELE		
IPC分类号	A61B19/00 A61B5/00		
CPC分类号	A61B5/0084 A61B5/6852 A61B18/1492 A61B19/2203 A61B19/46 A61B2017/00039 A61B2017/00053 A61B2017/00084 A61B2019/2211 A61B2019/442 A61B2019/448 A61B2019/465 A61B2019/547 A61B2562/0266 A61B5/6885 A61B90/98 A61B5/0066 A61B34/30 A61B90/06 A61B90/96 A61B2034 /301 A61B2090/065 A61B2090/397		
优先权	60/704825 2005-08-01 US 2005004852 2005-03-04 EP		
其他公开文献	US9907618		
外部链接	Espacenet USPTO		

摘要(译)

用于诊断或治疗血管或器官的装置和方法。该装置包括可变形体，例如导管，其具有组织消融末端执行器和与其流体连通的冲洗通道。至少两个传感器设置在可变形体的远端末端内，传感器响应于指定频率范围内的波，以检测由施加到远端末端的接触力引起的变形。微处理器可以与传感器可操作地耦合以从其接收输出，微处理器被配置为解析对应于接触力的多维力矢量。在一个实施例中，传感器是光纤布拉格光栅传感器，并且波从激光二极管注入到光纤布拉格光栅应变传感器中。

