



- (51) **International Patent Classification:**
A61B 8/06 (2006.01)
- (21) **International Application Number:**
PCT/IB2011/055703
- (22) **International Filing Date:**
15 December 2011 (15.12.2011)
- (25) **Filing Language:** English
- (26) **Publication Language:** English
- (30) **Priority Data:**
61/426,669 23 December 2010 (23.12.2010) US
61/466,048 22 March 2011 (22.03.2011) US
- (71) **Applicant (for all designated States except US):**
KONINKLIJKE PHILIPS ELECTRONICS N.V.
[NL/NL]; Groenewoudseweg 1, NL-5621 BA Eindhoven (NL).
- (72) **Inventors; and**
- (75) **Inventors/Applicants (for US only):** **WEL, Qifeng** [US/US]; c/o High Tech Campus Building 44, NL-5656 AE Eindhoven (NL). **THIELE, Karl E.** [US/US]; c/o High Tech Campus Building 44, NL-5656 AE Eindhoven (NL). **YOGANATHAN, Ajit, P.** [US/US]; c/o High Tech Campus Building 44, NL-5656 AE Eindhoven (NL). **YAP, Choon-Hwai** [MY/US]; c/o High Tech Campus Building 44, NL-5656AE Eindhoven (NL).
- (74) **Agents:** **VAN VELZEN, Maaïke, M.** et al.; High Tech Campus 44, NL-5656 AE Eindhoven (NL).
- (81) **Designated States (unless otherwise indicated, for every kind of national protection available):** AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IS, JP, KE, KG, KM, KN, KP, KR, KZ, LA, LC, LK, LR, LS, LT, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SC, SD, SE, SG, SK, SL, SM, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, ZA, ZM, ZW.
- (84) **Designated States (unless otherwise indicated, for every kind of regional protection available):** ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

Published:

— with international search report (Art. 21(3))

(54) **Title:** ANALYSIS OF MITRAL REGURGITATION FROM SLIT ORIFICES BY ULTRASONIC IMAGING




FIG. 5

(57) **Abstract:** An ultrasonic diagnostic imaging system is described which quantifies regurgitant flow through a plurality of pinhole leaks or a slit leak of a mitral valve. A plurality of orifice locations of a leaking valve are identified and Doppler values obtained from a flow velocity field proximal each orifice. The Doppler values of each flow velocity field vectorially relating to the orifice location are processed to produce a measure of flow through the orifice. The flow measurements for a plurality of such orifices are summed to produce a quantified measure of regurgitant flow through a plurality of pinhole leaks or along a slit leak.



ANALYSIS OF MITRAL REGURGITATION
FROM SLIT ORIFICES BY ULTRASONIC IMAGING

5 This application claims the benefit of U.S.
provisional patent application serial number
61/426,669, filed December 23, 2010.

10 This invention relates to medical diagnostic
ultrasound systems and, in particular, to the use of
diagnostic ultrasonic imaging systems for the
analysis of mitral regurgitation flow from a slit
opening in a mitral valve.

Regurgitant flow is a serious medical condition
which required analysis and appropriate treatment.
Just prior to the contraction of the left ventricle
15 to pump blood into the body, the mitral valve must
close completely so that the contraction will eject
all of the blood flow into the aorta. If the valve
does not close completely, some of the blood in the
left ventricle will be ejected back into the left
20 atria through the opening in the incompletely sealed
valve. This backflow of blood, typically a small,
momentary jet of blood flow squirting back through
the incompletely closed valve leaflets, reduces the
outflow of blood from the heart and hence the
25 efficiency of each heart contraction. The heart must
then pump more rapidly in order to supply the body
with its necessary supply of nourishing blood flow.
The heart is overworked due to its inefficiency,
leading to heart failure.

30 Clinicians have used ultrasonic imaging for many
years to try to detect regurgitant blood flow.
Ultrasonic detection of valvular regurgitation was
initially done by looking for the above-mentioned jet
of blood in an ultrasound image of the left side of
35 the heart. During the past twenty years observation

of the jet has been facilitated by two dimensional (2D) colorflow Doppler imaging, in which the high speed and turbulence of the small jet of blood is detected by careful search for these abnormal local flow velocities near the leaking heart valve. But acquisition of the image plane in which the jet is most prevalent, coupled with heart and valve motion and blood flow turbulence in the vicinity of the mitral valve, as well as the momentary occurrence of the jet, pose challenges to this subjective approach. In recent years, in cases where the location of the jet can be observed ultrasonically, clinicians have used a technique called PISA, an acronym for Proximal Iso-velocity Surface Area, to try to quantify the regurgitant blood flow. In this method the suspect valve and the region inside the LV heart chamber and proximal to the valve are imaged by colorflow Doppler imaging. At the time of occurrence of the jet a flow convergence region (FCR) is formed in the proximal region as blood flow velocities in the region instantaneously accelerate toward the regurgitant orifice. This flow pattern results in aliasing in the colorflow image as the flow velocities momentarily exceed the velocity range used for the colorflow image. A colorflow image at this moment is captured and frozen on the display screen. A measurement is then made of the velocity v at the first aliasing line of the FCR, and a measurement is made of the distance r from the aliasing line to the presumed center of the valve orifice. These two measurements are then used to compute the flow rate through the orifice using the expression $Q_t = 2\pi r^2 v$.

Several difficulties arise when conducting this procedure. One is that the greatest accuracy is obtained when the jet is captured in the colorflow image at its

very peak. The duration of the jet during a heart cycle can be only 300-450 milliseconds, however, while a typical colorflow frame rate may be in the range of 10-20 frames per second. Thus it is probable that the time of acquisition of one of the colorflow image frames will not be the same as the exact moment when the jet is at its peak. The clinician can repeat the colorflow acquisition sequence for additional cardiac cycles, or can settle for the inaccuracy caused by making the measurements at other than the exact peak of the jet.

Another problem is that the center of the valve orifice is not easy to define in the colorflow image. The valve tissue produces large reflections of ultrasound and is moving rapidly as scanning takes place, and can appear as a bulky, blurred or indistinct mass in the image. Thus it is possible that the accuracy of the measurement r will be compromised by the inability to estimate the exact location of the orifice.

Yet a third problem is that the basic PISA technique is only a single, one-dimensional measurement. Only one velocity measurement is made and only a single radius r to the orifice is used in the computation. The method assumes that the rest of the blood flow in the FCR is behaving the same as that of the single measurement. Obviously, any inaccuracy in making the single measurement will yield an inaccurate result.

An extension of the basic PISA technique which attempts to cure these inaccuracies is to make multiple velocity measurements around the arc delineating the outer boundary of the FCR in the two dimensional image. The distance r from each velocity measurement point to the orifice is measured, and the multiple measurements are used to calculate the flow rate Q_t . While the multiple measurements may prevent

reliance on a single inaccurate measurement, another problem arises. The single measurement (1D) technique generally is done by aiming a center beam of the ultrasound probe through the apex of the heart and directly through the presumed regurgitant orifice, and making the velocity v and distance r measurements along this beam line. With the beam line thus aligned with this vector of the regurgitant flow, the velocity measured will be accurate and not affected by a non-zero Doppler angle. As is well known, ultrasound Doppler velocity measurements are affected by the angle between the flow direction and the ultrasound beam direction. Flow which is directly in line with the beam direction will be accurately measured, whereas flow which is at a non-zero angle to the beam direction will be reduced by the cosine of the angle. Flow which is orthogonal (90°) to the beam direction will produce no Doppler response. Hence, a Doppler velocity measured by ultrasound must be corrected as a function of the angle between the flow vector and the beam direction in order to produce a more accurate velocity measurement. In the case of the 2D PISA technique, the additional velocity measurements along the FCR arc will be inaccurate due to variation of the angles between their flow vectors and the beam direction at each point where a velocity measurement is made. Thus, whereas multiple measurements can compensate for an error made in a single velocity measurement, the aggregate measurements will understate the flow rate due to the variation in the Doppler angles at each measurement point.

Yet a further problem inherent in the PISA technique is inaccuracy in knowing the precise location of the regurgitant orifice of the valve. As

mentioned above, aliasing in the colorflow image is prevalent in the vicinity of the valve leak due to the sudden acceleration of the blood flow toward and through the orifice. The colorflow image is thus cluttered with the flash of aliasing colors in the Doppler image in the vicinity of the orifice. Additionally, the mitral valve plane is in motion during the heart contraction. As the accuracy of the PISA technique is reliant upon knowing the orifice location in order to measure the distance r , these impediments to precisely knowing the orifice location can result in an inaccurate r measurement and hence an inaccurate calculation of flow rate. Experiments have demonstrated that an error of as little as one millimeter in the orifice location can lead to a significant error in calculating the flow rate through the leak. Furthermore, it is known that many regurgitant valves do not have a single pinhole leak, but leak along a slit of imprecise valve closure. The assumption of a leak as being caused by a single pinhole through the valve may thus not be valid.

Accordingly it is desirable to provide a method and apparatus for quantifying the flow rate and volume flow of mitral valve regurgitation which overcomes the above limitations of the PISA technique. It is an object of the present invention to accurately identify the precise location of the orifice of the valve regurgitation. It is a further object of the present invention to make a range of measurements which are unaffected by the Doppler angle at each measurement location. It is a further object of the present invention to be able to identify and quantify mitral valve regurgitation due not only to a single pinhole leak, but also due to a slit along the closure of the valve.

In accordance with the principles of the present invention, an ultrasonic diagnostic imaging system and method are described for quantifying regurgitant blood flow. An arcuate (two dimensional) or
5 hemispherical (three dimensional) region of blood flow is delineated proximal to a presumed location of a mitral valve leak in an ultrasound image of the regurgitant valve. The inner boundary of the region is at or just beyond the aliasing region adjacent to
10 the orifice of the leak, and the outer boundary of the region is outward from and concentric with the inner boundary. Initial assumptions are made as to the regurgitant flow, including the location of the regurgitant orifice. These assumptions are used to
15 compute a model of the velocity field of the regurgitant flow. The parameters of the model are adjusted in accordance with the physics of ultrasound and/or settings of the ultrasound system to determine expected velocity vectors to be observed by the
20 ultrasound system, V_{OBS} , in the delineated region. Velocity measurements are made by the ultrasound system within the delineated region, and the measured velocities are compared with the expected velocities. Differences between the measured and expected values
25 are computed and, through non-linear curve-fitting, adjustments are made to one or more of the model values. The process is iteratively repeated until the expected and measured values acceptably converge. The iteratively adjusted model parameters can be used
30 to produce a quantified measurement of the flow rate or the location of the leaking orifice in the valve.

In accordance with a further aspect of the present invention, an initially assumed location of the regurgitant orifice in the ultrasound image is
35 automatically updated by the foregoing process to

indicate the actual location of the orifice in the image. The clinician is thus shown a precise location of the regurgitant orifice despite the presence of image clutter in the vicinity of the mitral valve.

In accordance with yet another aspect of the present invention, the foregoing technique is repeated at multiple spatial locations along a mitral valve. The aggregate regurgitant flow measured at the spatial locations provides a measure of regurgitant flow due to a regurgitant leak which is a slit in the mitral valve closure rather than a single leaking pinhole orifice.

In accordance with a further aspect of the present invention, several wall filter characteristics are described which are preferred for an implementation of regurgitant flow measurement in accordance with the present invention, including a wall filter with a characteristic that peaks at intermediate flow velocities which are often found in a flow region proximal a regurgitant orifice.

In the drawings:

FIGURE 1 illustrates in block diagram form an ultrasonic diagnostic imaging system constructed in accordance with the principles of the present invention.

FIGURE 2 is a block diagram illustrating the functioning of the flow quantification processor of FIGURE 1.

FIGURE 3 illustrates an ultrasound Doppler image of a regurgitant jet.

FIGURE 4 illustrates an ultrasound image with an initially assumed orifice location and an automatically indicated orifice location in accordance with the present invention.

FIGURE 5 illustrates a technique for quantifying the regurgitant flow of a slit along a valve in accordance with the present invention.

FIGURE 6 illustrates the response characteristic of a wall filter suitable for use in an ultrasound system of the present invention.

Referring first to FIGURE 1, an ultrasonic diagnostic imaging system constructed in accordance with the principles of the present invention is shown in block diagram form. In FIGURE 1 a transducer array 10' is provided in an ultrasound probe 10 for transmitting ultrasonic waves and receiving echo information. The transducer array 10' is preferably a two dimensional array of transducer elements capable of scanning in three dimensions, for instance, in both elevation and azimuth about the location of the mitral valve, for 3D imaging. The transducer array is coupled to a microbeamformer 12 in the probe which controls transmission and reception of signals by the array elements. Microbeamformers are capable of at least partial beamforming of the signals received by groups or "patches" of transducer elements as described in US Pats. 5,997,479 (Savord et al.), 6,013,032 (Savord), and 6,623,432 (Powers et al.) The microbeamformer is coupled by the probe cable to a transmit/receive (T/R) switch 16 which switches between transmission and reception and protects the main beamformer 20 from high energy transmit signals. The transmission of ultrasonic beams from the transducer array 10 under control of the microbeamformer 12 is directed by the transmit controller 18 coupled to the T/R switch and the beamformer 20, which receives input from the user's operation of the user interface or control panel 38. One of the functions controlled by

the transmit controller is the direction in which beams are steered. Beams may be steered straight ahead from (orthogonal to) the transducer array, or at different angles for a wider field of view.

5 The partially beamformed signals produced by the microbeamformer 12 are coupled to a main beamformer 20 where partially beamformed signals from the individual patches of elements are combined into a fully beamformed signal. For example, the main
10 beamformer 20 may have 128 channels, each of which receives a partially beamformed signal from a patch of 12 transducer elements. In this way the signals received by over 1500 transducer elements of a two dimensional array can contribute efficiently to a
15 single beamformed signal.

 The beamformed signals are coupled to a signal processor 22. The signal processor 22 can process the received echo signals in various ways, such as bandpass filtering, decimation, I and Q component
20 separation, and harmonic signal separation which acts to separate linear and nonlinear signals so as to enable the identification of nonlinear echo signals returned from tissue and microbubbles. The signal processor may also perform additional signal
25 enhancement such as speckle reduction, signal compounding, and noise elimination.

 The processed signals are coupled to a B mode processor 26 and a Doppler processor 28. The B mode processor 26 employs amplitude detection for the
30 imaging of structures in the body such as the tissue of the heart wall, the mitral valve, and blood cells. B mode images of structure of the body may be formed in either the harmonic mode or the fundamental mode or a combination of both as described in US Pat.
35 6,283,919 (Roundhill et al.) and US Pat. 6,458,083

(Jago et al.) The Doppler processor 28 processes temporally distinct signals from tissue and blood flow for the detection of the motion of substances such as the flow of blood cells in the image field.

5 The Doppler processor typically includes a wall filter with parameters which may be set to pass and/or reject echoes returned from selected types of materials in the body. For instance, the wall filter can be set to have a passband characteristic which

10 passes signal of relatively low amplitude from higher velocity materials while rejecting relatively strong signals from lower or zero velocity material. This passband characteristic will pass signals from flowing blood while rejecting signals from nearby

15 stationary or slowly moving objects such as the wall of the heart. An inverse characteristic would pass signals from moving tissue of the heart while rejecting blood flow signals for what is referred to as tissue Doppler imaging, detecting and depicting

20 the motion of tissue. The Doppler processor receives and processes a sequence of temporally discrete echo signals from different points in an image field, the sequence of echoes from a particular point referred to as an ensemble. An ensemble of echoes received in

25 rapid succession over a relatively short interval can be used to estimate the Doppler shift frequency of flowing blood, with the correspondence of the Doppler frequency to velocity indicating the blood flow velocity. An ensemble of echoes received over a

30 longer period of time is used to estimate the velocity of slower flowing blood or slowly moving tissue. For mitral regurgitation assessment of a rapidly occurring jet, short ensemble lengths (fewer samples) are generally employed so that a high

35 acquisition frame rate can be realized. The Doppler

shift Δf may be estimated by an equation of the form
$$\Delta f = \frac{2vf_0 \cos\theta}{c}$$
 where f_0 is the transmit frequency, c is
the ultrasound propagation speed, v is velocity, and
 θ is the angle between the beam direction and the
5 direction of the blood flow.

The structural and motion signals produced by
the B mode and Doppler processors are coupled to a
scan converter 32 and a multiplanar reformatter 44.
The scan converter arranges the echo signals in the
10 spatial relationship from which they were received in
a desired image format. For instance, the scan
converter may arrange the echo signal into a two
dimensional (2D) sector-shaped format, or a pyramidal
three dimensional (3D) image. The scan converter can
15 overlay a B mode structural image with colors
corresponding to motion at points in the image field
corresponding with their Doppler-estimated velocities
to produce a color Doppler image which depicts the
motion of tissue and blood flow in the image field.
20 The multiplanar reformatter will convert echoes which
are received from points in a common plane in a
volumetric region of the body into an ultrasonic
image of that plane, as described in US Pat.
6,443,896 (Detmer). A volume renderer 42 converts
25 the echo signals of a 3D data set into a projected 3D
image as viewed from a given reference point as
described in US Pat. 6,530,885 (Entrekin et al.) The
2D or 3D images are coupled from the scan converter
32, multiplanar reformatter 44, and volume renderer
30 42 to an image processor 30 for further enhancement,
buffering and temporary storage for display on an
image display 40.

In accordance with the principles of the present
invention, blood flow velocity values produced by the

Doppler processor 28 are coupled to a flow quantification processor 34. The flow quantification processor operates as described below to produce a measure of the flow rate through a regurgitant orifice, the volume flow through the orifice, and the spatial location of the orifice. The flow quantification processor may receive input from the user control panel 38, such as an initial estimate of the location of the orifice as described below.

Output data from the flow quantification processor is coupled to a graphics processor 36 for the reproduction of output data from the processor with the image on the display 40. The graphics processor 36 can also generate graphic overlays for display with the ultrasound images. These graphic overlays can contain standard identifying information such as patient name, date and time of the image, imaging parameters, and the like. For these purposes the graphics processor receives input from the user interface 38, such as a typed patient name. The user interface is also coupled to the transmit controller 18 to control the generation of ultrasound signals from the transducer array 10' and hence the images produced by the transducer array and the ultrasound system. The user interface is also coupled to the multiplanar reformatter 44 for selection and control of a display of multiple multiplanar reformatted (MPR) images which may be used to quantify regurgitant flow in the MPR images in accordance with the present invention as described below.

FIGURE 2 describes the operation of the flow quantification processor of FIGURE 1. The processor is based upon a mathematical model of the flow velocity field at sample points within an inclusion zone as described below. In a 2D image embodiment

the inclusion zone is preferably a curved strip delineated by two arcs centered on the regurgitant orifice location, $\{x_o, y_o, z_o\}$. In a 3D image embodiment the curved strip is of the form of a hemispherical shell, as discussed in conjunction with FIGURE 3 below. In a preferred embodiment the model is a vector velocity model of the flow velocity field of the form

$$\vec{V}_{Model}(r) = \vec{n}_{rel} |V(r)|$$

or

$$\vec{V}_{Model}(r) = \frac{-\vec{r}}{\|\vec{r}\|} \circ |V(r)|$$

where $|V(r)| = \vec{F} / 2\pi r^2$ for the flow vector $\vec{F}$, and r is the distance from a point at (x, y, z) in three dimensional coordinates in the inclusion zone to the regurgitant orifice, and the orifice is located at coordinates $\{x_o, y_o, z_o\}$ in three dimensional space. This means that the distance r , when expressed vectorially to indicate the direction to the orifice, is of the form

$$\vec{r} = \{x - x_o, y - y_o, z - z_o\}$$

In box 50 of FIGURE 2, the flow velocity of every point in the inclusion zone is modeled in this way. Initially the model can start with assumed or estimated values for the unknown parameters, including the flow $\vec{F}$ and the $\{x_o, y_o, z_o\}$ location of the regurgitant orifice. For instance, the model can start with parameters which are nominally characteristic of regurgitant flow. The user can enter parameters such as by indicating the presumed location of the regurgitant orifice in the ultrasound image. Or a known technique such as PISA can be used

to calculate values used as the starting parameters for the flow velocity field.

The flow velocity field model approximates velocity vectors which would be accurate if the true physiologic velocity vectors were known. The velocity vectors approximated by the model are indicated at the output of box 50 as $\sim V_{\text{TRUE}}$. Box 52 then imposes some limits and adjustments to $\sim V_{\text{TRUE}}$ due to practical factors such as the physics of ultrasound and operating parameters of the ultrasound system being used. The flow velocity field model is then adjusted or scaled to take these practical factors into account in consideration of the actual velocity values which would be observed by an ultrasound system. One of the practical factors for which adjustment can be made is the Doppler angle. As mentioned above, Doppler measurements as performed by an ultrasound system are precisely accurate only when the direction of flow is in line with the direction of the ultrasound beam, a Doppler angle of zero. For all other angles between the beam direction and the flow direction the velocity is understated. In the Doppler shift equation presented above, the Doppler angle is weighed in the result by the $\cos\theta$ term, where θ is the Doppler angle. Another practical factor of an ultrasound system is wall filter bias. A Doppler wall filter will typically exhibit a nonlinear characteristic which, for blood flow detection, will have a zero response at DC (no motion) and rise to a maximum response at a selected high frequency of $\pm f$. Alternatively, the wall filter can exhibit a maximum response at a frequency less than that dictated by the Nyquist limit of the ensemble sampling rate, as discussed below. A sample volume in the body at which the Doppler shift is

measured will not be a single point in the body but will have a finite size, resulting in the return of Doppler signals indicative of a range of velocities. The non-uniform response of the wall filter can cause

5 a wide spread of velocities to experience greater gain at different locations of the response characteristic, e.g., higher velocities are more greatly emphasized than lower velocities. This non-uniform response can produce a shift in the perceived

10 center of the spread of velocities referred to as wall filter bias. The wall filter bias effect can also be taken into account by an adjustment to the model. Another related factor which can be taken into account is spectral spread, the Doppler spectral

15 broadening effect resulting from the different paths and angles from the sample volume to each receiving element of the active aperture of an array transducer. See US Pat. 5,606,972 (Routh). Yet another factor which can be taken into account is

20 aliasing effects, the mis-reporting of Doppler frequencies and velocities when the motion of blood flow is at a velocity in excess of that which can be detected unambiguously by the Nyquist limit of the sampling rate of the echo ensemble. In box 52,

25 factors such as these are taken into account by a dot product adjustment to the flow velocity field model which, for the models previous described, can be of the form

$$\vec{V}_{OBS}(r) = |V(r)| \left[\frac{-\vec{r}}{\|\vec{r}\|} \circ \vec{n}_{scan} \right]$$

where $\vec{n}_{scan}$ is a vectorial representation of the

30 physical and ultrasound system factors for which adjustment is being made. The result as shown in FIGURE 2 is V_{OBS} , the model as adjusted by expected effects and what is to be expected in the signals

measured by the ultrasound system.

A comparator 54 compares the expected velocity values from the flow velocity field model, V_{OBS} , with actual velocity measurements from points (sample
5 volumes) in the field, V_{MEAS} , produced by the Doppler processor 28. In box 56 the differences between the expected and received values are squared to produce an error term for each point. The error terms are integrated over the full inclusion zone which may be
10 a one dimensional (e.g., 1D line), two dimensional (e.g., 2D line arcuate area), or three dimensional (e.g., 3D hemispherical shell) inclusion zone to produce a mean squared error term from the full zone. The error term is then used to adjust the parameters
15 of the field model such as r and the flow rate to cause a reduction in a subsequently measured error term. The preferred adjustment technique is to use a non-linear curve fit to adapt the model toward error reduction. One such non-linear curve fitting
20 technique which may be used is the Levenberg-Marquardt algorithm, which refines the coordinates of the regurgitant orifice location and the flow ($\vec{F}$) or flow rate (Q_t) of the field toward or through the orifice.

25 The loop of FIGURE 2 is iteratively repeated to reduce the error term. With each pass through the loop a more precisely adapted field model is adjusted and compared with ultrasound velocity measurements from the acceptance zone and the error term is
30 iteratively reduced. When the error term has been reduced to an acceptable low value, the model exhibits the desired actual measurements of the orifice location and the blood flow through the orifice. These terms, and others as desired, are
35 produced as outputs by the flow quantification

processor 34 and are presented on the display to the user.

FIGURE 3 is a diagrammatic 2D ultrasound image illustrating practice of the present invention to
5 measure mitral valve regurgitation. The line 100 represents the plane of the mitral valve on which a regurgitant leak exists through an orifice O. When the left ventricle contracts, a jet of blood 102 escapes back into the right atria. In the PISA
10 method this jet would be interrogated by a Doppler beam 110 extending through the orifice O. US Pat. 6,719,697 (Li) presents an improvement to the PISA technique in which a color M mode display is produced over the the heart cycle from the location of Doppler
15 beam 110. The color Doppler display processor is set for the display limits indicated by the color bar 120 on the right side of the image, extending from a zero velocity center reference (BK=black) to +V and -V maximum velocities in opposite directions. A range
20 of colors (Y=yellow; DR-dark red; etc.) is displayed in accordance with this range of velocities. Aliasing occurs when a velocity being measured exceeds the Nyquist limits of +V and -V. Immediately adjacent the orifice is the flow convergent region
25 (FCR) 104. Beyond the flow convergent region is the acceptance zone 112 (S) used by an implementation of the present invention.

Both the PISA technique and the present invention are premised on the assumption that the
30 regurgitant flow in the left ventricle near the orifice is converging and flowing toward the orifice location O. This is indicated by the flow vectors $\vec{V}_1$, $\vec{V}_2$, and $\vec{V}_3$ in the acceptance zone S. But as the direction of the Doppler beam 110 illustrates,
35 virtually all of the flow vectors will not be aligned

with the beam direction, even when the varying beam angles of a phased array sector probe are used. Consequently there will be Doppler angles of different magnitudes for the different flow vectors, which is taken into account by the Doppler angle adjustments to the flow velocity field model in box 52 of FIGURE 2 as described above.

The color Doppler image of the FCR 104 will be chaotic and erratic. This is because the spatial orientation of blood cells and blood flow velocities are changing very rapidly in this region as blood flow changes direction and accelerates momentarily toward the orifice O during systolic contraction. The ensemble samples acquired during this interval are often uncorrelated with one another, which defeats the correlation within an ensemble upon which the usual Doppler estimation processors are reliant. As a result, even though the flow within the FCR may be generally laminar, the colorflow display can be that of highly turbulent flow and aliasing. The inner boundary 108 of the acceptance zone S is preferably set beyond the flow convergence region 104 to avoid the use of erratic velocity estimates from the FCR. One approach to setting the boundary 108 is to set it at or beyond the velocity shear boundary. This can be done visually with reference to the colorflow display, or automatically with reference to a velocity shear threshold.

The outer boundary 106 of the acceptance zone may be set in relation to measurable velocities. While the velocities of regurgitant flow are relatively high near the orifice, they become progressively lower at increasing distances from the orifice. The outer boundary 106 can be set at a distance from the orifice O at which low Doppler

velocities can still be reliably measured. This may be determined in relation to a percentage of the Nyquist limit or in relation to a low flow velocity such as 5 mm/sec. The outer boundary 106 can thus be
5 set at a distance at which acceptable sensitivity to low flows can still be realized by the Doppler processor.

The acceptance zone 112 in the example of FIGURE 3 is seen to be a two dimensional arcuate area S
10 having a center of rotation at the orifice O. The present invention may be used in one, two or three dimensions. A one dimensional implementation and model may just consider the segment of the beam line 110 which is between arcs 108 and 106, for instance.
15 A two dimensional implementation would be one which samples an acceptance zone which is planar such as acceptance zone 112 in FIGURE 3. A three dimensional implementation would consider an arcuate acceptance such as 112 but in a full hemisphere centered on the
20 orifice. Greater accuracy would be anticipated with the two and three dimensional implementations. A preferred implementation would use a 3D imaging probe with a two dimensional array transducer as shown in FIGURE 1, with acquisition and modeling done in a
25 hemispherical or semi-hemispherical volume shell about the orifice. Higher frame rates can be realized with two dimensional imaging, in which case an MPR frame through the orifice and jet can be selected from the volume scanned with the 3D probe by
30 use of the multiplanar reformatter 44. The desired plane can be repeatedly scanned at a high acquisition frame rate and the velocity measurements and flow calculations done with a two dimensional acceptance zone shown in the 2D MPR image.

35 While acceptance zones from which flow converges

in a circular pattern toward the orifice can exist as shown in FIGURE 3, other non-circular patterns have been found to be present in some cases. Stated differently, the flow vectors pointing toward the orifice may be arranged in a pattern other than an arc of a pure circle. Instead of the circular shape of FIGURE 3, an acceptance zone may have a parabolic or flattened circular shape. In three dimensional implementations the acceptance zone may be a paraboloid or oblate spheroid shell. The shape of the acceptance zone, that is, the area or volume in which velocities are measured in comparison to the model, may be dynamically changed during iterative passages through the processing loop of FIGURE 2 so that the process will adapt and converge to an acceptance zone with an altered shape from that used in the initial model.

As previously mentioned, it is frequently difficult for a clinician to accurately pinpoint the location of a tiny regurgitant orifice on an ultrasound image. When the heart is not contracting, the orifice is not present, and when it is present, its location is marred by imaging artifacts of the resultant turbulence such as those in the vicinity of the flow convergence region. In accordance with a further aspect of the present invention, an implementation of the present invention can automatically indicate the location of a leaking orifice on an ultrasound image. FIGURE 4 illustrates an ultrasound image in which elements previously described in conjunction with FIGURE 3 bear the same reference numerals. This image also includes a small box 130 which has been placed over the image by a clinician at the start of a diagnosis to indicate the presumed location of the regurgitant orifice. A

control of the control panel 38 such as the trackball is manipulated by the clinician to place the icon 130 at the location in the image where the clinician believes the orifice is located. The coordinates of the indicated orifice icon 130 are used to seed the processing of the flow quantification processor shown in FIGURE 2, whereby the coordinates are used as the initial coordinates of the orifice in the mathematical model of box 50. The processor 34 iteratively refines the modeled values of the flow velocity vectors toward an orifice location in response to the receipt of the measured velocity values V_{MEAS} . The vertical placement of the orifice location has been found to have the greatest effect on the convergence of the modeled velocity vectors with the measured velocity values. While the heart is contracting during systole, the movement of contraction causes the mitral valve to move vertically in the image toward the center of the left ventricle. Thus, the natural contractive movement of the heart can be the source of greatest error in regurgitant flow assessment. The processor 34 iteratively refines the location of the orifice to reduce the error disparity between the measured and estimated velocity values in the flow velocity field (acceptance zone S) between boundaries 106 and 108. As the flow quantification processor iterates to converge on the true coordinates of the orifice, the finally determined coordinates are used by the flow quantification processor 34 and the graphics processor 36 to automatically move orifice icon 130 to the calculated coordinates in the image. Alternatively the graphics processor will place another (calculated) orifice icon 132 on the display at the true coordinate location determined by the

iterative model adjustment. As mentioned previously, misplacement of the orifice by as little as 1mm can significantly affect the orifice flow rate and volume flow calculations, volume flow being the integral of the flow rate over the time during which the heart is contract (about 1/3 of a heart cycle). The flow volume will peak at approximately the middle of the systolic interval. In FIGURE 4 it is seen that the system has placed the calculated orifice icon 132 at its true location in the image field, which is different from the clinician's initial estimate of the orifice location. When the ultrasound study is performed with 2D imaging the orifice coordinates will generally be in (x,y,z) Cartesian coordinates. When 3D imaging is used a spherical (r,θ,φ) coordinate system will generally be employed. In addition, as the FIGURE 4 ultrasound image illustrates, the acceptance zone S is graphically delineated in the displayed ultrasound image and any modification of the shape of the acceptance zone as discussed above is incorporated in the displayed shape and graphical delineation of the acceptance zone on the display.

As mentioned above, regurgitant orifices do not always exist as single pinhole leaks in a closed valve. The leaks may be produced by slit openings or multiple small leaks. FIGURE 5 is an illustration of the use of an implementation of the present invention to asses the flow rate and volume flow of slits and multiple leaks in valve closure. As FIGURE 5 illustrates, the process of the present invention is performed for a series of orifice locations arrayed along the leaking mitral valve 100'. These discrete pinhole orifice locations can be used to model complex slit-like orifices or multiple orifices in

the closed valve. FIGURE 5 shows an example of three such determinations, each with its own acceptance zone, indicated by outer zone boundaries 106, 106' and 106'', and flow vectors (indicated by the small arrows) directed to converge at a different orifice location along the valve 100'. The vector velocity field associated with each pinhole orifice is combined vectorially with the other orifices, such that a single combined acceptance zone is defined along with a single combined vector velocity field. For multiple orifices, the combined vector velocity field now approximates the true physiologic velocity vectors ($\sim V_{true}$, output of box 50 in FIGURE 2). The processing of the combined vector velocity field determines the flow rate or volume flow of all the pinhole orifices corresponding to the entire slit along the mitral valve closure. While FIGURE 5 gives an example of a vector velocity field generated from three pinhole orifices, any number of orifices may be used to adequately model the slit. The acceptance zones used for each measurement may be combined as shown in FIGURE 5, as the flow that is determined for the combined orifices is vectorial flow directed toward a unique orifice or position along a slit. In a three dimension implementation the line of measurements is not constrained to a straight line in a plane, but may follow a non-linear path of closure of the mitral valve leaflets.

FIGURE 6 illustrates the response characteristics of two Doppler wall filters which may be employed in an implementation of the present invention. The coordinates of the abscissa of the plot of FIGURE 6 are in units of the Nyquist limits of a sampled data wall filter, where the limits of +1 and -1 are the normalized Nyquist limits of the

filter corresponding to the blood flow velocities. A zero (DC, or no-flow) condition is in the center. The ordinate coordinates indicate the relative amplitude response of the filter characteristic. The response curve 140 is a typical wall filter response curve for measuring blood flow and may be used in an implementation of the present invention. This characteristic has a response of zero at the center, resulting in no response to stationary objects such as stationary vessel walls. The response is seen to progressively increase as the curve extends out from zero, with a maximum response at the Nyquist limits for the highest velocity flows detectable without aliasing at the chosen sampling interval. As a result, this response characteristic is most sensitive to flow at the highest flow velocities.

The response curve 150 is one which is often preferred for the wall filter of an implementation of the present invention. The response of this curve 150 is seen to peak at a relatively high but intermediate sampling rate just above ± 0.5 Nyquist, preferably in the range of $\frac{1}{2}$ to $\frac{2}{3}$ of Nyquist, and drops to a response of zero at the Nyquist limits. This filter characteristic is designed to be more sensitive to lower flow velocities, which can be expected around the outer boundary 106 of the acceptance zone S. This is helpful to offset the low sensitivity to low velocity flow which results from the short ensembles typically used for colorflow imaging of regurgitant flow. This can be seen by the greater response of the curve 150 for lower velocities near the center of the plot. While the curve 150 is seen to drop to zero at the Nyquist limits, this loss of high velocity sensitivity is generally acceptable when traded off for greater

sensitivity to low velocity flows.

WHAT IS CLAIMED IS:

1. A method for measuring regurgitant flow from multiple or slit orifices comprising:
 - 5 identifying a first orifice location;
 - acquiring ultrasonic echo signals in the vicinity of the first orifice location;
 - processing the echo signals to produce a quantification of flow through the first orifice
 - 10 location;
 - identifying a second orifice location;
 - acquiring ultrasonic echo signals in the vicinity of the second orifice location;
 - processing the echo signals to produce a
 - 15 quantification of flow through the second orifice location; and
 - combining the quantifications of flow through the two orifice locations.
- 20 2. The method of Claim 1, wherein the first processing step further comprises Doppler processing the echo signals to identify flow vectors toward the first orifice location; and
- wherein the second processing step further
- 25 comprises Doppler processing the echo signals to identify flow vectors toward the second orifice location.
3. The method of Claim 1, wherein the first
- 30 acquiring step further comprises acquiring ultrasound echo signals from a two dimensional acceptance zone proximal the first orifice location; and
- wherein the second acquiring step further
- comprises acquiring ultrasound echo signals from a
- 35 two dimensional acceptance zone proximal the second

orifice location.

4. The method of Claim 3, wherein each acceptance zone is arcuate in shape having a center of curvature substantially at the first or second orifice location.

5. The method of Claim 3, wherein the two acceptance zones are spatially overlapping.

6. The method of Claim 1, wherein the first acquiring step further comprises acquiring ultrasound echo signals from a three dimensional acceptance zone proximal the first orifice location; and

wherein the second acquiring step further comprises acquiring ultrasound echo signals from a three dimensional acceptance zone proximal the second orifice location.

7. The method of Claim 6, wherein each acceptance zone is hemispherical in shape having a center of curvature substantially at the first or second orifice location.

8. The method of Claim 6, wherein the two acceptance zones are spatially overlapping.

9. The method of Claim 1, wherein the first and second orifice locations each comprise the location of a pinhole leak of a closed mitral valve.

10. The method of Claim 1, wherein the first and second orifice locations each comprise a location on a slit leak of a closed mitral valve.

35

11. The method of Claim 1, wherein combining
further comprises summing a plurality of flow
measurements for a plurality of orifices and
displaying a quantification of total flow rate or
5 volume flow for multiple or slit orifices.

12. A diagnostic ultrasound system for
measuring regurgitant flow from multiple or slit
orifices comprising:
10 an ultrasound probe having a transducer array
for transmitting ultrasonic energy to and receiving
ultrasonic echoes from a plurality of locations of
regurgitant flow in a body;
an image processor, responsive to received
15 echoes, for producing an ultrasound image of the
locations of regurgitant flow;
a Doppler processor responsive to received
echoes from the vicinity of each location for
producing Doppler ultrasound measurements of blood
20 flow velocity in the vicinity of the locations;
a flow quantification processor configured to
produce a measurement of a flow velocity field in the
vicinity of each of the locations and to sum the
measurements for the locations; and
25 a display device, coupled to the image processor
and the flow quantification processor, for displaying
the ultrasound image of the locations of regurgitant
flow and the summed measurements.

30 13. The diagnostic ultrasound system of Claim
12, wherein each flow velocity field contains flow
vectors relating to a specific orifice location.

35 14. The diagnostic ultrasound system of Claim
12, wherein each flow velocity field is two

dimensional.

15. The diagnostic ultrasound system of Claim
12, wherein each flow velocity field is three
5 dimensional.

10




FIG. 1








FIG. 2





INTERNATIONAL SEARCH REPORT

International application No
PCT/IB2011/055703

A. CLASSIFICATION OF SUBJECT MATTER
INV. A61B8/06
ADD.

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

EP0-Internal

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y	US 2002/151794 A1 (LI XIANG-NING [US]) 17 October 2002 (2002-10-17) cited in the application the whole document ----- -/--	1-15

☒ Further documents are listed in the continuation of Box C.

☒ See patent family annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier document but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

"&" document member of the same patent family

Date of the actual completion of the international search

30 March 2012

Date of mailing of the international search report

05/04/2012

Name and mailing address of the ISA/

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040,
Fax: (+31-70) 340-3016

Authorized officer

Artikis, T

INTERNATIONAL SEARCH REPORT

International application No
PCT/IB2011/055703

C(Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y	CHANDRA S ET AL: "Volumetric measurement of the anatomic regurgitant orifice area in mitral regurgitation: Comparison with two-dimensional flow convergence analysis", COMPUTERS IN CARDIOLOGY, 2009, IEEE, PISCATAWAY, NJ, USA, 13 September 2009 (2009-09-13), pages 769-772, XP031656187, ISBN: 978-1-4244-7281-9 abstract page 769, left-hand column, paragraph 2 - page 772, left-hand column, paragraph 1 figure 4	1-15
A	----- LIN B A ET AL: "Color Doppler Jet Area Overestimates Regurgitant Volume when Multiple Jets are Present", JOURNAL OF THE AMERICAN SOCIETY OF ECHOCARDIOGRAPHY, MOSBY-YEAR BOOK, INC. ST. LOUIS, MO, US, vol. 23, no. 9, 1 September 2010 (2010-09-01), pages 993-1000, XP027224814, ISSN: 0894-7317 [retrieved on 2010-08-08] the whole document	1-15
A	----- US 2009/043208 A1 (HERGUM TORBJORN [NO] ET AL) 12 February 2009 (2009-02-12) abstract page 994, left-hand column, paragraph METHODS - page 998, right-hand column, paragraph 4 figures 1,4 -----	1,15

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No

PCT/IB2011/055703

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 2002151794 A1	17-10-2002	NONE	

US 2009043208 A1	12-02-2009	US 2009043208 A1	12-02-2009
		WO 2009031034 A2	12-03-2009

专利名称(译)	超声成像分析狭缝孔二尖瓣关闭不全		
公开(公告)号	EP2654571A1	公开(公告)日	2013-10-30
申请号	EP2011811570	申请日	2011-12-15
[标]申请(专利权)人(译)	皇家飞利浦电子股份有限公司		
申请(专利权)人(译)	皇家飞利浦N.V.		
当前申请(专利权)人(译)	皇家飞利浦N.V.		
[标]发明人	WEI QIFENG THIELE KARL E YOGANATHAN AJIT P YAP CHOON HWAI		
发明人	WEI, QIFENG THIELE, KARL E. YOGANATHAN, AJIT, P. YAP, CHOON-HWAI		
IPC分类号	A61B8/06		
CPC分类号	A61B8/06 A61B8/145 A61B8/4444 A61B8/4488 A61B8/463 A61B8/483 A61B8/488 A61B8/5207 A61B8/5223		
代理机构(译)	STEFFEN , THOMAS		
优先权	61/426669 2010-12-23 US 61/466048 2011-03-22 US		
其他公开文献	EP2654571B1		
外部链接	Espacenet		

摘要(译)

描述了一种超声诊断成像系统，其量化通过多个针孔泄漏或二尖瓣的狭缝泄漏的反流。识别泄漏阀的多个孔位置，并从靠近每个孔的流速场获得多普勒值。对与孔口位置矢量相关的每个流速场的多普勒值进行处理，以产生通过孔口的流量测量值。将多个这样的孔的流量测量值相加以产生通过多个针孔泄漏或沿着狭缝泄漏的反流流量的量化测量。