



(11) **EP 2 831 573 B1**

(12) **EUROPEAN PATENT SPECIFICATION**

(45) Date of publication and mention
of the grant of the patent:
08.11.2017 Bulletin 2017/45

(51) Int Cl.:
G01N 29/44 (2006.01) **A61B 8/00** (2006.01)
G01N 29/265 (2006.01) **G06T 17/00** (2006.01)
G01N 29/22 (2006.01) **A61B 8/08** (2006.01)

(21) Application number: **13767924.7**

(86) International application number:
PCT/CA2013/000302

(22) Date of filing: **28.03.2013**

(87) International publication number:
WO 2013/142978 (03.10.2013 Gazette 2013/40)

(54) **ACTUATOR FOR MOVING AN ULTRASOUND PROBE**

AKTUATOR ZUR BEWEGUNG EINER ULTRASCHALLSONDE

ACTIONNEUR POUR LE DÉPLACEMENT D'UNE SONDE À ULTRASONS

(84) Designated Contracting States:
**AL AT BE BG CH CY CZ DE DK EE ES FI FR GB
GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO
PL PT RO RS SE SI SK SM TR**

- **NESHAT, Hamid**
London, Ontario N6H 2M3 (CA)
- **KAKANI, Nirmal**
London, Ontario N6A 5A5 (CA)

(30) Priority: **28.03.2012 US 201261617031 P**

(74) Representative: **Faulkner, Thomas John**
Cleveland
10 Fetter Lane
London EC4A 1BR (GB)

(43) Date of publication of application:
04.02.2015 Bulletin 2015/06

(73) Proprietor: **The University Of Western Ontario**
London, Ontario N6G 4X8 (CA)

(56) References cited:
WO-A1-96/00402 WO-A2-02/30287
CA-A1- 2 749 465 CN-A- 102 497 131
GB-A- 2 027 197 US-A- 4 290 309
US-A- 5 445 154 US-A- 5 456 256
US-A- 5 487 388 US-A- 5 562 095
US-A1- 2009 326 554 US-A1- 2010 030 078
US-A1- 2011 152 690 US-A1- 2011 224 551

(72) Inventors:
• **BARKER, Kevin**
London, Ontario N6G 5M9 (CA)
• **FENSTER, Aaron**
London, Ontario N6G 4N9 (CA)

Note: Within nine months of the publication of the mention of the grant of the European patent in the European Patent Bulletin, any person may give notice to the European Patent Office of opposition to that patent, in accordance with the Implementing Regulations. Notice of opposition shall not be deemed to have been filed until the opposition fee has been paid. (Art. 99(1) European Patent Convention).

Description**Field of the Invention**

5 **[0001]** The present invention relates to ultrasound imaging and in particular, to an actuator for moving an ultrasound probe and to an ultrasound imaging system employing the same.

Background of the Invention

10 **[0002]** A wide variety of approaches have been developed to produce three-dimensional (3D) ultrasound images using both one-dimensional (1D) ultrasound arrays and two-dimensional (2D) ultrasound arrays. The use of 1D ultrasound arrays to produce 3D ultrasound images requires methods to determine the position and orientation of acquired 2D ultrasound images within the 3D ultrasound image volume. The use of 2D ultrasound arrays to produce 3D ultrasound images requires a 3D scan-converter to build the 3D ultrasound images from the sequence of transmit acoustic signals.

15 **[0003]** U.S. Patent No. 5,562,095 to Downey et al. discloses a 3D ultrasound imaging system for the eye, prostate and other organs, comprising an assembly onto which an ultrasound probe may be mounted, a motor and drive for either rotating or scanning the ultrasound probe relative to the human organ under investigation, and a computer for executing software for controlling movement of the assembly to rotate or scan the ultrasound probe. Ultrasound signals from the probe are processed via a clinical ultrasound machine for generating multiple 2D ultrasound images of the organ. The software executed on the computer collects the 2D ultrasound images of the clinical ultrasound machine and reconstructs the 2D ultrasound images to form a 3D ultrasound image which can be viewed and manipulated in real time, or stored for later retrieval.

The document WO02/30287 A2 discloses an ultrasound probe having sensors coupled to the probe to determine the probe's location. The probe may be moved across cellular tissue at a uniform rate, which is synchronized with the image capture rate of an ultrasound scanner.

25 The document US 2010/030078 A1 discloses apparatus for imaging mammary gland and breast whilst compressing the breast with a compression plate.

The document US 5487388A, which represents the closest prior art to the subject-matter of independent claim 1, discloses a scanning device for scanning the breast intercostally to provide a 3D image format.

30 **[0004]** Improvements in 3D ultrasound imaging are generally desired. It is therefore an object at least to provide a novel actuator for moving an ultrasound probe and a novel ultrasound imaging system employing the same.

Summary of the Invention

35 **[0005]** The invention is defined by the independent claim 1. Accordingly, in one aspect there is provided an actuator for an ultrasound probe comprising a housing assembly configured to accommodate an ultrasound probe and to be positioned adjacent a target region of interest to be examined, and a motor assembly coupled to the housing assembly, the motor assembly configured to move the housing assembly in a manner to allow an ultrasound probe when accommodated by the housing assembly to perform a compound scan of the target region of interest.

40 **[0006]** In one embodiment, during the compound scan, the motor assembly is configured to move the housing assembly along a linear path and to tilt the housing assembly during its movement along the linear path. The motor assembly comprises a first motor system and a second motor system, the first motor system being configured to move the housing assembly along the linear path and the second motor system being configured to tilt the housing assembly. The first motor system comprises a motor coupled to a linear slide for moving the housing assembly along the linear path. The second motor system comprises a motor coupled to a plurality of linkage arms for tilting the housing assembly between first and second boundaries.

45 **[0007]** In one embodiment, the motor assembly is further configured to move the housing assembly along the linear path to allow the ultrasound probe when accommodated by the housing assembly to perform a linear scan of the target region of interest or to tilt the housing assembly to allow the ultrasound probe when accommodated by the housing assembly to perform a tilt scan of the target region.

50 **[0008]** In one embodiment, the housing assembly comprises a harness to retain the ultrasound probe therein. A casing encloses at least the motor assembly and has user actuatable controls thereon. Formations are provided in the casing to accommodate a user's fingers and the user actuatable controls are positioned approximate the formations.

55 **[0009]** According to another aspect there is provided an ultrasound imaging system comprising the above actuator, an ultrasound probe accommodated by the housing assembly of the actuator, and processing structure configured to process ultrasound signals output by the ultrasound probe and to control the motor assembly of the actuator.

[0010] According to another aspect there is provided a method of reconstructing a three-dimensional ultrasound image of a target region of interest, the method comprising: receiving a plurality of 2D ultrasound images captured during a

compound scan of the target region of interest; and mapping pixels of each of the received 2D ultrasound images to a corresponding voxel within a reconstruction matrix.

[0011] In one embodiment, the 2D ultrasound images are received in succession. Each pixel of each 2D ultrasound image is mapped to a corresponding voxel within the reconstruction matrix. The reconstruction matrix is displayed on a display device. A mapping matrix is calculated to map pixels of each received 2D ultrasound image to the corresponding voxel within the reconstruction matrix.

[0012] According to yet another aspect there is provided a method of reconstructing a three-dimensional ultrasound image of a target region of interest, the method comprising: calculating a mapping matrix to map pixels within an acquisition matrix to a corresponding voxel within a reconstruction matrix; receiving a plurality of 2D ultrasound images captured during a compound scan of the target region of interest and storing pixels of each of the plurality of 2D ultrasound images within the acquisition matrix; mapping each pixel within the acquisition matrix with a corresponding voxel within the reconstruction matrix using the mapping matrix; and storing a pixel value associated with each pixel within the acquisition matrix at the corresponding voxel within the reconstruction matrix.

Brief Description of the Drawings

[0013] Embodiments will now be described more fully with reference to the accompanying drawings in which:

Figure 1 is a schematic block diagram of a 3D ultrasound imaging system;

Figure 2a is an isometric view of a motor assembly and a housing assembly forming part of an actuator of the 3D ultrasound imaging system of Figure 1;

Figure 2b is a top plan view of the motor assembly and the housing assembly of Figure 2a;

Figure 2c is a front elevational view of the motor assembly and the housing assembly of Figure 2a;

Figure 2d is a side elevational view of the motor assembly and the housing assembly of Figure 2a;

Figure 2e is a top plan view of the motor assembly disassembled;

Figure 3a is an isometric view of the motor assembly, the housing assembly, an ultrasound probe and a casing;

Figure 3b is a top plan view of the motor assembly, the housing assembly, the ultrasound probe and the casing;

Figure 3c is a front elevational view of the motor assembly, the housing assembly, the ultrasound probe and the casing;

Figure 3d is a side elevational view of the motor assembly, the housing assembly, the ultrasound probe and the casing;

Figure 4 is a flowchart showing a 3D ultrasound imaging system operation method;

Figures 5a to 5c are side elevational views of the 3D ultrasound imaging system of Figure 1 during operation in a linear scanning mode;

Figures 6a to 6c are side elevational views of the 3D ultrasound imaging system of Figure 1 during operation in a tilt scanning mode;

Figures 7a to 7c are side elevational views of the 3D ultrasound imaging system of Figure 1 during operation in a compound scanning mode;

Figure 8 is a flowchart showing a method for reconstructing a 3D ultrasound image during the compound scanning mode; and

Figure 9 is an exemplary ultrasound image reconstructed according to the method of Figure 8.

Detailed Description of the Embodiments

[0014] In the following, an actuator for an ultrasound probe and an ultrasound imaging system employing the same are described. The actuator comprises a housing assembly configured to accommodate an ultrasound probe and to be positioned adjacent to a target region of interest to be examined. The region of interest may be a human organ or tissue such as for example a liver, a kidney, a breast, etc. A motor assembly is coupled to the housing assembly and is configured to move the housing assembly and hence the ultrasound probe, in a manner to allow the ultrasound probe when accommodated by the housing assembly to perform a compound scan of the target region of interest.

[0015] Turning now to Figure 1, a schematic block diagram of a 3D ultrasound imaging system 100 is shown. In this embodiment, the 3D ultrasound imaging system 100 is operable in a plurality of modes including a linear scanning mode, a tilt scanning mode, and a compound scanning mode, as will be described. The 3D ultrasound imaging system 100 comprises a general purpose computing device 200 that is communicatively coupled to an actuator 250. The actuator 250 comprises a motor assembly 300 and a housing assembly 400. The motor assembly 300 receives control commands from the general purpose computing device 200 and in response, moves the housing assembly 400. The housing assembly 400 accommodates an ultrasound probe 502 (see Figures 3a to 3d) that communicates with a clinical ultrasound machine 500. The clinical ultrasound machine 500 processes ultrasound signals received from the ultrasound probe 502 during scanning of a target region of interest to generate 2D ultrasound images of the region of interest. The clinical ultrasound machine 500 is communicatively coupled to the general purpose computing device 200 and sends generated

2D ultrasound images thereto. The general purpose computing device 200 in turn processes the received 2D ultrasound images to reconstruct a 3D ultrasound image of the target region of interest, which is communicated to a user interface 600 for display thereon.

[0016] The general purpose computing device 200 in this embodiment is a personal computer or other suitable processing device comprising, for example, a processing unit, system memory (volatile and/or non-volatile memory), other non-removable or removable memory (e.g., a hard disk drive, RAM, ROM, EEPROM, CD-ROM, DVD, flash memory, etc.) and a system bus coupling the various computing device components to the processing unit. The general purpose computing device 200 may also comprise networking capability using Ethernet, WiFi, and/or other network format, to access shared or remote drives, one or more networked computers, or other networked devices.

[0017] Figures 2a to 2e illustrate the motor assembly 300 and the housing assembly 400 and Figures 3a and 3d illustrate the motor assembly 300, the housing assembly 400, the ultrasound probe 502, and a casing 700 enclosing the motor assembly 300 and a portion of the housing assembly 400.

[0018] In this embodiment, motor assembly 300 comprises a linear motor system 302 and a tilt motor system 304. The linear motor system 302 comprises a motor 306 coupled to a linear slide carriage 310 via a lead screw 312. Each side of the linear carriage 310 is connected to a linear slide 308. Each linear slide 308 is slidably mounted to a respective stationary rail which is connected to an interior surface of the casing 700, as will be described. As the motor 306 rotates the lead screw 312, the linear slide carriage 310 moves along a linear path. The movement of the linear slide carriage 310 causes movement of each linear slide 308 along its respective stationary rail. As such, the linear motor system 302 moves the housing assembly 400 along the linear path. In this embodiment, the motor 306 has a gear ratio of about 112:1 and a maximum no load output speed of about 59 revolutions-per-minute (RPM). The lead screw 312 has a pitch of about 0.94 cm (1/2.7 inches). The linear motor system 302 is configured to move the linear slides 308 a total of about 3cm along the linear path at a speed of about 0.75cm/s.

[0019] The tilt motor system 304 comprises a motor 314 and a plurality of linkage arms 316, 318 and 320 formed of aluminum or other suitable material. The tilt motor system 304 is shown disassembled in Figure 2e together with linear slide carriage 310. The linkage arms 316, 318 and 320 and linear slide carriage 310 are connected to one another, as will be described, via a polyether ether ketone (PEEK) bushing on a stainless steel shaft. A first connection point 316-1 of linkage arm 316 is connected to a first connection point 318-1 of linkage arm 318. A second connection point 316-2 of linkage arm 316 is connected to the linear slide carriage 310 at connection point 310-1. A front end defined by connection points 320-1 and 320-2 of linkage arm 320 is connected to the linear slide carriage 310 at connection points 310-2 and 310-3, respectively. A back end of linkage arm 320 defined by connection points 320-3 and 320-4 is connected to housing assembly 400. A connection point 318-2 of linkage arm 318 is connected to a driving shaft (not shown) of motor 314. Linkage arm 318 is connected to a lower portion of housing assembly 400 at connection points 318-3 and 318-4. As will be appreciated, motor 314, when actuated, imparts rotation of linkage arm 318. As a result of the rotation, linkage arms 316 and 320 rotate and thus, the tilt motor system 304 provides tilt motion to housing assembly 400 between a range of about -30° and 30°. In this embodiment, the motor 314 has a gear ratio of about 207:1 and a maximum no load output speed of about 32 RPM. The tilt motor system 304 is configured to complete a 60° sweep in about 4 seconds.

[0020] Housing assembly 400 comprises a linkage arm 402 connected to the connection points 320-3 and 320-4 of linkage arm 320. A bottom portion 404 of the housing assembly 400 is connected to connection points 318-3 and 318-4 of linkage arm 318. The housing assembly 400 has a housing port 406 dimensioned to securely hold a harness 410 that is configured to retain the ultrasound probe 502. A handle 408 is tangentially connected to a side of the housing assembly 400. The top of the handle 408 comprises a locking button for locking the harness 410 therein. The bottom of the handle 408 comprises an unlocking button for unlocking the harness 410 therefrom. In this embodiment, the housing assembly 400, including the linkage arm 402 and handle 408, is made of a polyether ether ketone (PEEK) material. Harness 410 is positioned around a top portion of the ultrasound probe 502 and retains the ultrasound probe 502. The harness 410 is custom made for different types of ultrasound probes. The harness 410 is opened by loosening a screw associated therewith and when opened, the harness 410 is able to receive the ultrasound probe 502.

[0021] The casing 700 in this embodiment is made of a molded plastic material such as for example ABS-M30 manufactured by Stratasy®. The interior surface of the casing 700 comprises a pair of stationary rails (not shown). Each stationary rail extends along a respective side of the interior surface of the casing 700 and is configured to support a respective linear slide 308. Each side of the exterior surface of the casing 700 has a pair of grooves 702 formed therein, each of which is dimensioned to receive a finger of a user. A plurality of operational buttons is positioned within the grooves 702. The buttons can be user actuated to allow the user to control and operate the 3D ultrasound imaging system 100. Exemplary functions associated with the operational buttons include start scan, stop scan, jog, etc.

[0022] The user interface 600 in this embodiment comprises a display monitor such as for example a liquid crystal display (LCD) panel or other suitable display panel or device connected to the general purpose computing device 200 via a cable such as for example a high-definition multimedia interface (HDMI) cable or a video graphics array (VGA) cable or other suitable wired or wireless communications link. The user interface 600 displays the graphical user interface of an application program executed by the general purpose computing device 200. The user interface 600 further

comprises a keyboard and a mouse and/or other suitable user input devices, which are connected to the general purpose computing device 200 via a universal serial bus (USB) cable or other suitable wired or wireless connection.

[0023] Turning now to Figure 4, a method of operating the 3D ultrasound imaging system 100 is shown and is generally identified by reference numeral 800. During operation, when the general purpose computing device 200 receives a mode selection from the user interface 600 (step 802), a check is performed to determine which mode has been selected (step 804). If the linear scanning mode has been selected, when a start scan command has been input via actuation of the associated operational button, the linear motor system 302 is actuated by the general purpose computing device 200 (step 806) and thus, a linear scan of the target region of interest is performed by the 3D ultrasound imaging system 100, and the method ends. If the tilt scanning mode has been selected, when a start scan command has been input via actuation of the associated operational button, the tilt motor system 304 is actuated by the general purpose computing device 200 (step 808) and thus, a tilt scan of the target region of interest is performed by the 3D ultrasound imaging system 100, and the method ends. If the compound scanning mode has been selected, when a start scan command has been input via actuation of the associated operational button, both the linear motor system 302 and the tilt motor system 304 are actuated by the general purpose computing device 200 (step 810) and thus a compound scan of the target region of interest is performed by the 3D ultrasound imaging system 100, and the method ends.

[0024] Turning to Figure 5a, the start position of the actuator 250 in the linear scanning mode is shown. As can be seen, ultrasound probe 502 is positioned such that its longitudinal axis is generally normal to the surface of the target region of interest to be scanned. Each linear slide 308 is at its rearward most position on the respective stationary rail. As will be appreciated, each linear slide 308 is contained within casing 700, and thus is not visible in Figure 5a.

[0025] Figure 5b shows the finish position of the actuator 250 in the linear scanning mode. As can be seen, ultrasound probe 502 remains positioned such that its longitudinal axis is generally normal to the surface of the target region of interest to be scanned. Each linear slide 308 is at its forward most position on the respective stationary rail. As such, the linear slide carriage 310 extends out of casing 700. Again, as each linear slide 308 is contained within casing 700, it is not visible in Figure 5b.

[0026] Figure 5c shows both the start position and the finish position of the actuator 250 in the linear scanning mode. As can be seen, the ultrasound probe 502 moves from the start position to the finish position along a linear path indicated by arrow LP. The linear path LP is defined by movement of the linear slide carriage 310 and the resultant movement of each linear slide 308 along its respective stationary rail. As such, the ultrasound probe 502 performs a linear scan of the target region of interest when moved along linear path LP. It will be appreciated that in the linear scanning mode, there is no tilt movement of the ultrasound probe 502.

[0027] During operation in the linear scanning mode, the ultrasound probe 502 scans the target region of interest as it moves along the linear path LP and outputs ultrasound signals to the clinical ultrasound machine 500. The ultrasound signals from the ultrasound probe 502 are processed via the clinical ultrasound machine 500 to generate a plurality of 2D ultrasound images of the target region of interest at the desired frame rate (typically 15 or 30 frames per second). The 2D ultrasound images are then communicated to the general purpose computing device 200 for processing. The general purpose computing device 200 processes the 2D ultrasound images to reconstruct a 3D ultrasound image covering a rectangular volume of the scanned target region of interest according to a known method such as that described in U.S. Patent No. 5,562,095 to Downey et al.

[0028] Turning to Figure 6a, the start position of the actuator 250 in the tilt scanning mode is shown. As can be seen, ultrasound probe 502 is positioned such that its longitudinal axis is approximately 30° to the normal of the surface of the target region of interest to be scanned. Each linear slide 308 is at its rearward most position on the respective stationary rail. As will be appreciated, each linear slide 308 is contained within casing 700, and thus is not visible in Figure 6a.

[0029] Figure 6b shows the finish position of the actuator 250 in the tilt scanning mode. As can be seen, ultrasound probe 502 is positioned such that its longitudinal axis is approximately -30° to the normal of the surface of the target region of interest to be scanned. Each linear slide 308 is at its rearward most position on the respective stationary rail. As will be appreciated, each linear slide 308 is contained within casing 700, and thus is not visible in Figure 6b.

[0030] Figure 6c shows both the start position and the finish position of the actuator 250 in the tilt scanning mode. As can be seen, the ultrasound probe 502 moves from the start position to the finish position along a path indicated by arrow TP. As such, the ultrasound probe 502 performs a tilt scan of the target region of interest when moved along path TP. It will be appreciated that in the tilt scanning mode, there is no linear movement of the ultrasound probe 502.

[0031] During operation in the tilt scanning mode, the ultrasound probe 502 scans the target region of interest as it moves along path TP and outputs ultrasound signals to the clinical ultrasound machine 500. The ultrasound signals from the ultrasound probe 502 are processed via the clinical ultrasound machine 500 to generate a plurality of 2D ultrasound images of the target region of interest at the desired frame rate (typically 15 or 30 frames per second). The 2D ultrasound images are communicated to the general purpose computing device 200 for processing. The general purpose computing device 200 in turn processes the 2D ultrasound images to reconstruct a 3D ultrasound image covering a pie-shaped volume of the scanned target region of interest according to a known method such as that described in U.S. Patent No. 5,562,095 to Downey et al.

[0032] Turning to Figure 7a, the start position of the actuator 250 in the compound scanning mode is shown. As can be seen, ultrasound probe 502 is positioned such that its longitudinal axis is approximately -30° to the normal of the surface of the target region of interest to be scanned. Each linear slide 308 is at its rearward most position on the respective stationary rail. As will be appreciated, each linear slide 308 is contained within casing 700, and thus is not visible in Figure 7a.

[0033] Figure 7b shows the finish position of the actuator 250 in the tilt scanning mode. As can be seen, ultrasound probe 502 is positioned such that its longitudinal axis is approximately 30° to the normal of the surface of the target region of interest to be scanned. Each linear slide 308 is also at its forward most position on the respective stationary rail. Again, as each linear slide 308 is contained within casing 700, it is not visible in Figure 7b.

[0034] Figure 7c shows both the start position and the finish position of the actuator 250 in the compound scanning mode. As can be seen, the ultrasound probe 502 moves from the start position to the finish position along a path indicated by arrow CP. As such, the ultrasound probe 502 performs a compound scan of the target region of interest when moved along path CP. Comparing the scan path of the ultrasound probe 502 during operation in the compound scanning mode (Figure 7c) to the scan paths of the ultrasound probe 502 during operation in the linear scanning mode (Figure 5c) and tilt scanning mode (Figure 6c), a larger volume is covered by the ultrasound probe 502 during operation in the compound scanning mode.

[0035] During operation in the compound scanning mode, the ultrasound probe 502 scans the target region of interest as it moves along path CP. The ultrasound signals output by the ultrasound probe 502 are processed via the clinical ultrasound machine 500 to generate a plurality of 2D ultrasound images of the target region of interest at the desired frame rate (typically 15 or 30 frames per second). The 2D ultrasound images are communicated to the general purpose computing device 200 for processing. The general purpose computing device 200 in turn processes the 2D ultrasound images to reconstruct a 3D ultrasound image covering a frusto-conical shaped volume of the scanned target region of interest as will now be described below.

[0036] As will be appreciated, in the compound scanning mode, reconstructing a 3D ultrasound image from the plurality of 2D ultrasound images must take into account both the linear and tilt movements of the ultrasound probe 502. For this purpose, two 3D matrices are created within memory associated with the general purpose computing device 200 and mapped to one another.

[0037] The first 3D matrix, hereinafter referred to as an acquisition matrix, holds pixel values of each acquired 2D ultrasound image. The size of the acquisition matrix is expressed as:

$$h \times w \times n \quad (1)$$

where h is the number of pixels in each 2D ultrasound image along the Y-direction, w is the number of pixels in each 2D ultrasound image along the X-direction, and n is the number of 2D ultrasound images to be captured during the compound scan. It should be noted that the number of pixels in each 2D ultrasound image along the Y-direction h and X-direction w are fixed for the compound scan. The number of 2D ultrasound images to be captured n is determined by dividing the maximum sweeping angle of the ultrasound probe 502 by a desired angle between two consecutive 2D ultrasound images. The desired angle between two consecutive 2D ultrasound images is set by the user prior to ultrasound image acquisition via the user interface 600. As will be appreciated, the grey-value of each pixel may be one or two bytes depending on the number of bits selected to represent the grey-values. Prior to scanning, each pixel within the acquisition matrix is set to a value of zero, and thus is set to the color black.

[0038] The second 3D matrix, hereinafter referred to as a reconstruction matrix, holds grey-values of each voxel of the reconstructed 3D ultrasound image. The size of the reconstruction matrix is determined by the acquired image parameters, scanning mode, and maximum sweep either set as default values within the 3D ultrasound imaging system 100 or set by the user via the user interface 600. Specifically, the size of the reconstruction matrix is expressed as:

$$h \times w \times z \quad (2)$$

where h is the number of pixels in each 2D ultrasound image along the Y-direction and w is the number of pixels in each 2D ultrasound image along the X-direction. The value of z is calculated as:

$$z = (n - 1) \left(1 + \frac{2h}{L} \sin(\theta) \right) \quad (3)$$

where n is the number of 2D ultrasound images to be captured, h is the number of pixels in each 2D ultrasound image along the Y-direction, L is the extent of the linear motion during the compound scan, and θ is the maximum angle of the tilt motion during the compound scan.

[0039] To determine an association between a voxel in the reconstruction matrix and a corresponding pixel in the acquisition matrix, a mapping matrix is used. In this embodiment, the mapping matrix is calculated using a closest point method, wherein a particular voxel within the reconstruction matrix is mapped to a pixel within the acquisition matrix that has coordinates closest thereto. The mapping matrix is calculated before the compound scan is performed and is used to fill each voxel within the reconstruction matrix as corresponding pixels within the acquisition matrix are filled based on the captured 2D ultrasound images. The mapping matrix is stored within memory associated with the general purpose computing device 200 and may be used with future compound scans comprising the same parameters (i.e. angular step angle, linear step distance, linear extent of scan L , maximum angle θ , size of image in X-direction, number of pixels in X-direction, and number of pixels in Y-direction). During the compound scan, the reconstruction matrix is displayed by the user interface 600 such that a user can follow the scanning and reconstructing process.

[0040] Turning to Figure 8, a flowchart showing a method for reconstructing a 3D ultrasound image from a plurality of 2D ultrasound images obtained during a compound scan of the target region of interest is shown and is generally identified by reference numeral 900. Method 900 begins by calculating the mapping matrix (step 902). The mapping matrix is calculated by first transferring each pixel in the acquisition matrix according to corresponding 3D scanning geometry, and then using each transferred pixel to find a reverse map according to the closest point method. Each pixel ($X_{im(i)}$, $Y_{im(i)}$) in acquired 2D ultrasound image i in the acquisition matrix (containing a total number of acquired 2D ultrasound images n) is transferred to 3D geometry using the following equations:

$$X = X_{im(i)} \quad (4)$$

$$Y = Y_{im(i)} dx \cos(\theta_i) \quad (5)$$

$$Z = i \times dL + Y_{im(i)} dy \sin \theta_i \quad (6)$$

where the transformed 3D coordinates $[X, Y, Z]$ represent the corresponding location in the 3D ultrasound image, θ_i is the angle of the ultrasound probe 502 at which 2D ultrasound image i is acquired, and dL is the linear scanning distance between two consecutive captured 2D ultrasound images. It is noted that the transformed 3D coordinates $[X, Y, Z]$ do not necessarily contain integer values and do not directly correspond to a location within the reconstruction matrix, since coordinates within the reconstruction matrix are integer values. As will be appreciated, if only the integer part of the transformed 3D coordinates $[X, Y, Z]$ is used, depending on the linear scanning distance between two consecutive captured 2D ultrasound images, there may be gaps between the planes in the reconstruction matrix that must be filled. Further, there may be cases wherein two or more pixels in the acquired 2D ultrasound images contain the same integer values.

[0041] Once the 3D transformation for each pixel ($X_{im(i)}$, $Y_{im(i)}$) in the acquisition matrix has been calculated according to Equations (4), (5) and (6), each voxel within the reconstruction matrix that falls within the boundaries of the scanned target region of interest is evaluated. In this embodiment, any voxel within the reconstruction matrix that is outside of the boundaries of the scanned target region of interest will have a grey-value that results in a black colored voxel. For each of these voxels, a pixel within the acquisition matrix that has coordinates closest to each particular pixel within the reconstruction matrix is identified, and the transformed 3D coordinates $[X, Y, Z]$ of that pixel are recorded within the mapping matrix.

[0042] To minimize processing, given the geometry of the scanned volumes, once a single slice perpendicular to the compound scan direction within the mapping matrix is calculated, the remaining slices within the mapping matrix are determined by changing the value of X . To further minimize processing, given the symmetry about the mid-line of the compound scan, that is, when the ultrasound probe 502 is positioned such that its longitudinal axis is generally perpendicular to the surface of the target region of interest to be scanned, once half of the pixels in a slice have been calculated for the mapping matrix, the second half of the pixels can be readily determined by mirroring the Y value of the calculated pixels. As will be appreciated, each element within the mapping matrix has three components represented by $X_{im(i)}$, $Y_{im(i)}$ and the acquired 2D ultrasound image number i .

[0043] Once the mapping matrix has been calculated, the compound scan begins (step 904). A 2D ultrasound image of the target region of interest is acquired (step 906) and stored in the acquisition matrix (step 908). The pixels of the acquired 2D ultrasound image are mapped to corresponding voxels within the reconstruction matrix using the mapping

matrix, and the resulting reconstruction matrix is displayed on the user interface 600 (step 910). A check is performed to determine if the compound scan is complete (step 912). If the compound scan is not complete, the method returns to step 906 to acquire another 2D ultrasound image of the region of interest. Once all of the 2D ultrasound images of the compound scan have been acquired and processed as described above, the method ends.

[0044] A side view of an exemplary ultrasound image 1000 reconstructed according to method 900 is shown in Figure 9. Reconstructed image 1000 was acquired using a linear scanning distance L of 25cm and a maximum angle θ of 27° . As can be seen, reconstructed image 1000 is symmetric about center line 1002, which is shown for illustrative purposes. Each line 1004 represents the side view of a single 2D ultrasound image stored within the reconstruction matrix.

[0045] Although the mapping matrix is calculated above using a closest point method, in another embodiment the mapping matrix may be calculated using an interpolation method, wherein the grey-value of a particular voxel within the reconstruction matrix is determined using the grey-value of a combination of pixels within the acquisition matrix that have coordinates closest to the particular voxel within the reconstruction matrix. In this embodiment, once the 3D transformation for each pixel ($X_{im(i)}$, $Y_{im(i)}$) in the acquisition matrix has been calculated according to Equations (4), (5) and (6), a linear combination of 2, 4 or 6 pixels within the acquisition matrix that have coordinates closest to each particular voxel within the reconstruction matrix are identified, the grey-value of each of these pixels is used to calculate the grey-value of the corresponding voxel in the reconstruction matrix, and the transformed 3D coordinates $[X, Y, Z]$ of each of the 2, 4 or 6 pixels are stored in the mapping matrix. As will be appreciated, if 2 pixels are used, each pixel within the mapping matrix will contain 8 components represented by $X_{im(i)}$, $Y_{im(i)}$ and the acquired 2D ultrasound image number i for each of the 2 pixels identified in the acquisition matrix, as well as 2 values associated with the weighting factors of the 2 pixels. If 4 pixels are used, each pixel within the mapping matrix will contain 16 components represented by $X_{im(i)}$, $Y_{im(i)}$ and the acquired 2D ultrasound image number i for each of the 4 pixels identified in the acquisition matrix, as well as 4 values associated with the weighting factors of the 4 pixels. Similarly, if 6 pixels are used, each pixel within the mapping matrix will contain 24 components represented by $X_{im(i)}$, $Y_{im(i)}$ and the acquired 2D ultrasound image number i for each of the 6 pixels identified in the acquisition matrix, as well as 6 values associated with the weighting factors of the 6 pixels.

[0046] Those skilled in the art will appreciate that virtually any type of ultrasound probe may be used and virtually any type of clinical ultrasound machine that is compatible with the selected ultrasound probe may be used. For example, clinical ultrasound machines utilizing B-mode ultrasound, Doppler imaging ultrasound, or power Doppler imaging ultrasound may be used.

[0047] Although the linkage arms are described above as being made of aluminum, those skilled in the art will appreciate that other types of material may be used such as for example stainless steel, plastic materials, etc.

[0048] Although the linear motor system is described above as comprising a linear slide to provide movement along a single axis, those skilled in the art will appreciate that other arrangements may be used. For example, a box slide or dovetail may be used. Similarly, the linear movement may be achieved through use of rods or round rails. Also, although the linear motor system is described above as comprising a motor driven lead screw to provide movement along a single axis, those skilled in the art will appreciate that other arrangements may be used. For example, devices such as a belt, chain, rack and pinion gearing may be used.

[0049] Although the tilt motor system is described above as comprising a motor directly connected to one or more linkage arms, the tilt motor system may alternatively use one or more gears to adjust the speed of the tilt motor system.

[0050] Although the linear motor system and tilt motor system are described above as comprising separate motors, those skilled in the art will appreciate that one motor may be used. For example, a single motor may be used in combination with a specific orientation of linkage arms to operate within tilt and linear motion parameters, thereby achieving proper motion for a compound scan. It will also be appreciated that different types of motors may be used such as for example brushed DC motors.

[0051] Although the ultrasound probe is described above as being retained within the housing assembly via a harness, those skilled in the art will appreciate that other types of mechanical retaining devices may be used. For example, a quick release mechanism or spring-loaded quick release mechanism may be used to retain the ultrasound probe within the housing assembly.

[0052] Although voxels within the reconstruction matrix are considered to be outside of the scanned target region of interest if they are black in color, those skilled in the art will appreciate that the voxels to be evaluated may be selectable by a user via user interface 600.

[0053] In embodiments described above, particular motor gear ratios and output speeds, lead screw pitches, path distances, sweeps and speeds as well as particular component materials have been specified. Those of skill in the art will appreciate that these specifics are exemplary and that alternatives are available.

[0054] Although embodiments have been described with reference to the drawings, those of skill in the art will appreciate that variations and modifications may be made without departing from the scope thereof as defined by the appended claims.

Claims

1. An actuator (250) for an ultrasound probe comprising:

a housing assembly (400) configured to accommodate the ultrasound probe (502) and to be positioned adjacent to a target region of interest to be examined;
a motor assembly (300) coupled to the housing assembly, the motor assembly configured to move the housing assembly in a manner to allow the ultrasound probe (502) when accommodated by the housing assembly to perform a compound scan of the target region of interest; and a casing (700) enclosing at least the motor assembly (300); **characterised in that** the actuator (250) further comprises user actuatable controls on the casing (700).

2. The actuator of claim 1 wherein during the compound scan, the motor assembly (300) is configured to move the housing assembly (400) along a linear path and to tilt the housing assembly during its movement along the linear path.

3. The actuator of claim 2 wherein the motor assembly (300) comprises a first motor system (302) and a second motor system (304), the first motor system being configured to move the housing assembly (400) along the linear path and the second motor system (304) being configured to tilt the housing assembly.

4. The actuator of claim 3 wherein the first motor system (302) comprises a motor (306) coupled to a linear slide carriage (310) for moving the housing assembly (400) along the linear path.

5. The actuator of claim 3 or 4 wherein the second motor system (304) comprises a motor (314) coupled to a plurality of linkage arms (316, 318, 320) for tilting the housing assembly (400) between first and second boundaries.

6. The actuator of claim 5 wherein the first and second boundaries are defined by first and second angles measured with respect to the normal to a surface of the target region of interest.

7. The actuator of claim 6 wherein the first and second angles are complimentary.

8. The actuator of claim 6 or 7 wherein the first angle is equal to approximately 30° and the second angle is equal to approximately -30°.

9. The actuator of any one of claims 2 to 8 wherein the motor assembly (300) is further configured to move the housing assembly (400) along the linear path to allow the ultrasound probe (502) when accommodated by the housing assembly to perform a linear scan of the target region of interest or to tilt the housing assembly (400) to allow the ultrasound probe when accommodated by the housing assembly (400) to perform a tilt scan of the target region of interest.

10. The actuator of any one of claims 1 to 9 wherein the housing assembly (400) comprises a harness (410) to retain the ultrasound probe therein.

11. The actuator of any one of claims 1 to 10 wherein the casing (700) has formations (702) therein shaped to accommodate a user's fingers and wherein the user actuatable controls are positioned proximate the formations.

12. The actuator of claim 11 wherein said formations (702) comprise a pair of grooves formed in each exterior side surface of said casing.

13. An ultrasound imaging system (100) comprising:

an actuator (250) according to any one of claims 1 to 12; the ultrasound probe (502) accommodated by the housing assembly (400) of the actuator; and
processing structure (200) configured to process ultrasound signals output by the ultrasound probe (502) and to control the motor assembly (300) of the actuator (250).

14. The ultrasound imaging system of claim 13 wherein the processing structure (200) comprises a clinical ultrasound machine (500) to which the ultrasound probe is connected, the clinical ultrasound machine being configured to generate a succession of two-dimensional ultrasound images during the compound scan, and a processing device

configured to reconstruct the two-dimensional ultrasound images to form a three-dimensional ultrasound image of the target region of interest.

15. The ultrasound imaging system of claim 14 further comprising a user interface (600) communicating with said processing structure (200) via a suitable wired or wireless communication link.

Patentansprüche

1. Ein Aktor (250) für eine Ultraschallsonde, umfassend:

einen Gehäuseaufbau (400), der eingerichtet ist, um die Ultraschallsonde (502) aufzunehmen und um an einem zu untersuchenden Zielbildbereich angrenzend positioniert zu werden;
einen Motoraufbau (300), der an den Gehäuseaufbau gekoppelt ist, wobei der Motoraufbau eingerichtet ist, um den Gehäuseaufbau auf eine Weise zu bewegen, um der Ultraschallsonde (502), wenn diese von dem Gehäuseaufbau aufgenommen ist, eine zusammengesetzte Abtastung ("compound scan") des Zielbildbereichs zu gestatten; und
ein Gehäuse (700), das wenigstens den Motoraufbau (300) umschließt;
dadurch gekennzeichnet, dass der Aktor (250) ferner benutzerbedienbare Bedienelemente auf dem Gehäuse (700) umfasst.

2. Der Aktor nach Anspruch 1, wobei während der zusammengesetzten Abtastung der Motoraufbau (300) eingerichtet ist, um den Gehäuseaufbau (400) entlang eines linearen Pfads zu bewegen und den Gehäuseaufbau während seiner Bewegung entlang des linearen Pfads zu kippen.

3. Der Aktor nach Anspruch 2, wobei der Motoraufbau (300) ein erstes Motorsystem (302) und ein zweites Motorsystem (304) umfasst, wobei das erste Motorsystem eingerichtet ist, den Gehäuseaufbau (400) entlang des linearen Pfads zu bewegen, und das zweite Motorsystem (304) eingerichtet ist, um den Gehäuseaufbau zu kippen.

4. Der Aktor nach Anspruch 3, wobei das erste Motorsystem (302) einen Motor (306) umfasst, der an einen linearen Gleitschlitten (310) gekoppelt ist, zum Bewegen des Gehäuseaufbaus (400) entlang des linearen Pfads.

5. Der Aktor nach Anspruch 3 oder 4, wobei das zweite Motorsystem (304) einen Motor (314) umfasst, der an eine Vielzahl an Verbindungsarmen (316, 318, 320) gekoppelt ist, zum Kippen des Gehäuseaufbaus (400) zwischen ersten und zweiten Begrenzungen.

6. Der Aktor nach Anspruch 5, wobei die ersten und zweiten Grenzen durch erste und zweite Winkel definiert sind, die bezüglich der Normalen auf eine Oberfläche des Zielbildbereichs gemessen werden.

7. Der Aktor nach Anspruch 6, wobei die ersten und zweiten Winkel komplementär sind.

8. Der Aktor nach Anspruch 6 oder 7, wobei der erste Winkel ungefähr gleich 30° ist und der zweite Winkel ungefähr gleich -30° ist.

9. Der Aktor nach einem der Ansprüche 2 bis 8, wobei der Motoraufbau (300) ferner eingerichtet ist, um den Gehäuseaufbau (400) entlang des linearen Pfads zu bewegen, um der Ultraschallsonde (502), wenn diese von dem Gehäuseaufbau aufgenommen ist, zu ermöglichen, eine lineare Abtastung des Zielbildbereichs durchzuführen, oder um den Gehäuseaufbau (400) zu kippen, um der Ultraschallsonde, wenn diese von dem Gehäuseaufbau (400) aufgenommen ist, zu ermöglichen, eine Kippabtastung des Zielbildbereichs durchzuführen.

10. Der Aktor nach einem der Ansprüche 1 bis 9, wobei der Gehäuseaufbau (400) eine Kopfbefestigung (410) umfasst, um die Ultraschallsonde darin zu sichern.

11. Der Aktor nach einem der Ansprüche 1 bis 10, wobei das Gehäuse (700) darin Strukturen (702) aufweist, die geformt sind, um Finger eines Benutzers aufzunehmen und wobei die benutzerbedienbaren Bedienelemente nahe den Strukturen angeordnet sind.

12. Der Aktor nach Anspruch 11, wobei die Strukturen (702) ein Paar Vertiefungen umfasst, die in jeder Außenseiten-

fläche besagten Gehäuses geformt sind.

13. Ein Ultraschallbildgebungssystem (100), umfassend:

einen Aktor (250) nach einem der Ansprüche 1 bis 12;
die Ultraschallsonde (502), die von dem Gehäuseaufbau (400) des Aktors aufgenommen ist; und
eine Verarbeitungsstruktur (200), die eingerichtet ist, um Ultraschallsignale zu verarbeiten, die von der Ultraschallsonde (502) ausgegeben werden, und um den Motoraufbau (300) des Aktors (250) zu steuern.

14. Das Ultraschallbildgebungssystem nach Anspruch 13, wobei die Verarbeitungsstruktur (200) eine klinische Ultraschallmaschine (500) umfasst, mit welcher die Ultraschallsonde verbunden ist, wobei die klinische Ultraschallmaschine eingerichtet ist, um eine Folge von zweidimensionalen Ultraschallbildern während der zusammengesetzten Abtastung zu erzeugen, und ein Verarbeitungsgerät, das eingerichtet ist, um die zweidimensionalen Ultraschallbilder zu rekonstruieren, um ein dreidimensionales Ultraschallbild des Zielbildbereichs zu bilden.

15. Das Ultraschallbildgebungssystem nach Anspruch 14, ferner umfassend eine Benutzerschnittstelle (600), die mit besagter Verarbeitungsstruktur (200) über eine geeignete kabelgebundene oder kabellose Kommunikationsverbindung kommuniziert.

Revendications

1. Actionneur (250) pour une sonde à ultrasons comprenant :

un ensemble d'enceinte (400) configuré pour loger la sonde à ultrasons (502) et être positionné à proximité adjacente d'une région cible d'intérêt à examiner ;
un ensemble moteur (300) couplé à l'ensemble d'enceinte, l'ensemble moteur étant configuré pour déplacer l'ensemble d'enceinte de manière à permettre à la sonde à ultrasons (502) lorsqu'elle est logée par l'ensemble d'enceinte d'effectuer un balayage composé de la région cible d'intérêt ;
et un boîtier (700) entourant au moins l'ensemble moteur (300) ; **caractérisé en ce que** l'actionneur (250) comprend en outre des commandes actionnables par utilisateur sur le boîtier (700).

2. Actionneur selon la revendication 1, dans lequel, pendant le balayage composé, l'ensemble moteur (300) est configuré pour déplacer l'ensemble d'enceinte (400) le long d'un trajet linéaire et pour incliner l'ensemble d'enceinte pendant son déplacement le long du trajet linéaire.

3. Actionneur selon la revendication 2, dans lequel l'ensemble moteur (300) comprend un premier système de moteur (302) et un second système de moteur (304), le premier système de moteur étant configuré pour déplacer l'ensemble d'enceinte (400) le long du trajet linéaire et le second système de moteur (304) étant configuré pour incliner l'ensemble d'enceinte.

4. Actionneur selon la revendication 3, dans lequel le premier système de moteur (302) comprend un moteur (306) couplé à un chariot à glissement linéaire (310) pour déplacer l'ensemble d'enceinte (400) le long du trajet linéaire.

5. Actionneur selon la revendication 3 ou 4, dans lequel le second système de moteur (304) comprend un moteur (314) couplé à une pluralité de bras de liaison (316, 318, 320) pour incliner l'ensemble d'enceinte (400) entre des première et seconde limites.

6. Actionneur selon la revendication 5, dans lequel les première et seconde limites sont définies par des premier et second angles mesurés par rapport à la normale à une surface de la région cible d'intérêt.

7. Actionneur selon la revendication 6, dans lequel les premier et second angles sont complémentaires.

8. Actionneur selon la revendication 6 ou 7, dans lequel le premier angle est égal à approximativement 30 ° et le second angle est égal à approximativement -30 °.

9. Actionneur selon l'une quelconque des revendications 2 à 8, dans lequel l'ensemble moteur (300) est en outre configuré pour déplacer l'ensemble d'enceinte (400) le long du trajet linéaire pour permettre à la sonde à ultrasons

(502) lorsqu'elle est logée par l'ensemble d'enceinte d'effectuer un balayage linéaire de la région cible d'intérêt ou pour incliner l'ensemble d'enceinte (400) pour permettre à la sonde à ultrasons lorsqu'elle est logée par l'ensemble d'enceinte (400) d'effectuer un balayage incliné de la région cible d'intérêt.

- 5 10. Actionneur selon l'une quelconque des revendications 1 à 9, dans lequel l'ensemble d'enceinte (400) comprend un harnais (410) pour retenir la sonde à ultrasons à l'intérieur de celui-ci.
11. Actionneur selon l'une quelconque des revendications 1 à 10, dans lequel le boîtier (700) comporte à l'intérieur des formations (702) formées pour recevoir les doigts d'un utilisateur et dans lequel les commandes actionnables par
10 utilisateur sont positionnées à proximité des formations.
12. Actionneur selon la revendication 11, dans lequel lesdites formations (702) comprennent une paire de rainures formées dans chaque surface latérale extérieure dudit boîtier.
- 15 13. Système d'imagerie à ultrasons (100) comprenant :

un actionneur (250) selon l'une quelconque des revendications 1 à 12 ;
la sonde à ultrasons (502) logée par l'ensemble d'enceinte (400) de l'actionneur ; et
une structure de traitement (200) configurée pour traiter des signaux à ultrasons délivrés par la sonde à ultrasons
20 (502) et pour commander l'ensemble moteur (300) de l'actionneur (250).
14. Système d'imagerie à ultrasons selon la revendication 13 dans lequel la structure de traitement (200) comprend une machine à ultrasons clinique (500) à laquelle la sonde à ultrasons est connectée, la machine à ultrasons clinique étant configurée pour générer une succession d'images ultrasonores bidimensionnelles pendant le balayage composé, et un dispositif de traitement configuré pour reconstruire les images ultrasonores bidimensionnelles afin de
25 former une image ultrasonore tridimensionnelle de la région cible d'intérêt.
15. Système d'imagerie à ultrasons selon la revendication 14 comprenant en outre une interface utilisateur (600) communiquant avec ladite structure de traitement (200) via une liaison de communication câblée ou sans fil appropriée.
30




Figure 1




Figure 2a




Figure 2b




Figure 2d

Figure 2c




Figure 2e




Figure 3a

Figure 3b




Figure 3d




Figure 3c




Figure 4




Figure 5a




Figure 5b




Figure 5c




Figure 6a




Figure 6b




Figure 6c




Figure 7a




Figure 7b




Figure 7c




Figure 8




Figure 9

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- US 5562095 A, Downey [0003] [0027] [0031]
- WO 0230287 A2 [0003]
- US 2010030078 A1 [0003]
- US 5487388 A [0003]

专利名称(译)	用于移动超声探头的致动器		
公开(公告)号	EP2831573B1	公开(公告)日	2017-11-08
申请号	EP2013767924	申请日	2013-03-28
[标]申请(专利权)人(译)	西安大略大学		
申请(专利权)人(译)	加拿大西安大略大学		
当前申请(专利权)人(译)	加拿大西安大略大学		
[标]发明人	BARKER KEVIN FENSTER AARON NESHAT HAMID KAKANI NIRMAL		
发明人	BARKER, KEVIN FENSTER, AARON NESHAT, HAMID KAKANI, NIRMAL		
IPC分类号	G01N29/44 A61B8/00 G01N29/265 G06T17/00 G01N29/22 A61B8/08		
CPC分类号	A61B8/4461 A61B8/13 A61B8/4218 A61B8/466 A61B8/483 A61B8/5207 A61B8/5253 A61B8/54 G01N29/225 G01N29/265		
优先权	61/617031 2012-03-28 US		
其他公开文献	EP2831573A4 EP2831573A1		
外部链接	Espacenet		

摘要(译)

用于超声探头的致动器包括：壳体组件，被配置为容纳超声探头并且被定位在待检查的目标感兴趣区域附近；以及马达组件，被联接到壳体组件，马达组件被配置为移动壳体组件。以一种方式允许超声探头在被壳体组件容纳时执行目标感兴趣区域的复合扫描。

