



EUROPEAN PATENT APPLICATION
published in accordance with Art. 153(4) EPC

(43) Date of publication:
06.01.2016 Bulletin 2016/01

(51) Int Cl.:
A61N 7/00 (2006.01) **A61B 8/00** (2006.01)
A61B 8/14 (2006.01) **A61B 18/00** (2006.01)
G01S 15/89 (2006.01)

(21) Application number: **13876137.4**

(22) Date of filing: **28.02.2013**

(86) International application number:
PCT/KR2013/001669

(87) International publication number:
WO 2014/133208 (04.09.2014 Gazette 2014/36)

(84) Designated Contracting States:
AL AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO PL PT RO RS SE SI SK SM TR
Designated Extension States:
BA ME

- **KANG, Kookjin**
Yongin-si
Gyeonggi-do 448-536 (KR)
- **KIM, Dae Seung**
Seoul 157-882 (KR)
- **KIM, Myungdeok**
Seoul 152-020 (KR)
- **JUN, Sukhwan**
Incheon 402-200 (KR)

(71) Applicant: **Alpinion Medical Systems Co., Ltd.**
Hwaseong-si, Gyeonggi-do 445-380 (KR)

(72) Inventors:
• **SON, Keonho**
Seongnam-si
Gyeonggi-do 463-440 (KR)

(74) Representative: **Thorniley, Peter et al**
Kilburn & Strode LLP
20 Red Lion Street
London WC1R 4PJ (GB)

(54) **METHOD FOR FOCAL POINT COMPENSATION, AND ULTRASONIC MEDICAL APPARATUS THEREFOR**

(57) A method and an ultrasound medical apparatus for compensating for a focal point are disclosed. A method and an ultrasound medical apparatus are provided, for transmitting a high-intensity ultrasound having an intensity lower than a predetermined output value to a target object (an affected area of a patient) in a manner that

no errors are generated between a focal point on the target object and an actual ultrasound focusing location when transmitting a high-intensity focused ultrasound and compensating a distance of the focal point by using a reflected signal of the ultrasound.




FIG. 1

Description

[Technical Field]

[0001] The present disclosure relates to a method and an apparatus for compensating a focal point, and more particularly, to a method and an ultrasound medical apparatus for focusing a high-intensity ultrasound having an intensity lower than a predetermined output value on a target object (an affected area of a patient), and compensating for a distance of the focal point by a difference as determined to be present between the focal point on the target object and the actual ultrasound focusing location.

[Background]

[0002] The statements in this section merely provide background information related to the present disclosure and do not necessarily constitute prior art.

[0003] A high-intensity focused ultrasound (HIFU) is generally used for treating (processing) a biological tissue such as cancer, tumor, and lesion. The treatment method using a high-intensity ultrasound necrotizes a biological tissue by using a heat generated by transmitting the high-intensity ultrasound to a single spot in a focused manner. In this case, the high-intensity ultrasound needs to be controlled not to damage a healthy biological tissue, and using such a high-intensity focused ultrasound obviates an incision process from a surgical operation.

[0004] In the treatment method using the high-intensity ultrasound, an ultrasound for acquiring an image is transmitted to a biological tissue to be treated, an image is acquired by using an echo signal which is a reflected signal of the ultrasound, and then a high-intensity ultrasound is transmitted to the corresponding biological tissue. In this case, a possible difference between the focal point to be treated by the ultrasound medical apparatus and the actual location where the high-intensity ultrasound is focused causes damage to a biological tissue out of the focal point.

[Disclosure]

[Technical Problem]

[0005] The present disclosure has been made in an effort to effectively resolving at least a part of the above-mentioned problems, and it is an object of some embodiments of the present disclosure to provide a method and an ultrasound medical apparatus for focusing a high-intensity ultrasound having an intensity lower than a predetermined output value on a target object or an affected area of a patient, and when it is determined that there is a difference between the focal point on the target object and the actual ultrasound focusing location, compensating for a distance of the focal point by the difference.

[Summary]

[0006] According to some embodiments of the present disclosure, an ultrasound medical apparatus is provided, which includes an imaging transducer configured to transmit a first imaging ultrasound to a target object, to receive a first echo signal reflected from the target object, and to generate a first reception signal based on the first echo signal, a signal processing unit configured to generate a first image picture based on the first reception signal and to output the first image picture to a display unit, a user input unit configured to receive an input of focal-point information on the first image picture, a treatment transducer configured to transmit a high-intensity ultrasound having an intensity lower than a predetermined output value to a focal point corresponding to the focal-point information, a focusing-location extracting unit configured to extract actual focal-point information by using a change amount of a second image picture that is generated based on a second imaging ultrasound transmitted at a time when the high-intensity ultrasound reaches the focal point, and a compensating unit configured to generate a compensation value based on the focal-point information and the actual focal-point information and to apply the compensation value to the focal-point information.

[0007] According to another embodiment of the present disclosure, a method for compensating for a focal point in an ultrasound medical apparatus is provided, which includes:

a first reception signal generating step including transmitting a first imaging ultrasound to a target object, receiving a first echo signal reflected from the target object, and generating a first reception signal based on the first echo signal, a first image processing step including generating a first image picture based on the first reception signal, and outputting the first image picture to a display unit, a focal point inputting step including receiving an input of focal-point information on the first image picture, a high-intensity ultrasound transmitting step including transmitting a high-intensity ultrasound having an intensity

lower than a predetermined output value to a focal point corresponding to the focal-point information,
 an actual focal-point extracting step including extracting actual focal-point information by using a change amount
 of a second image picture that is generated based on a second imaging ultrasound transmitted at a time when the
 high-intensity ultrasound reaches the focal point, and
 a compensating step including generating a compensation value based on the focal-point information and the actual
 focal-point information, and applying the compensation value to the focal-point information.

[Advantageous Effects]

[0008] According to some embodiments of the present disclosure as described above, a pre-targeting can be achieved by focusing a high-intensity ultrasound having an intensity lower than a predetermined output value on a target object or an affected area of a patient and using a reflected signal of the ultrasound, and when an error occurs between the focal point on the target object and the actual ultrasound focusing location, a distance of the focal point can be compensated by as much as the difference. According to some embodiments, a pre-targeting is performed to avail of the correct focus by transmitting an undamaging level of high-intensity ultrasound to a human tissue at the focal point, and a possible mismatch between the focal point on the target object and the actual ultrasound focusing location can be compensated by using an electronic compensation value without a physical movement.

[0009] Further, according to some embodiments, off-focusing issues are addressed when the treatment transducer transmits the high-intensity ultrasound for treatment to the set focal point and if there is a mismatch between the focal-point information and the actual focusing location information risking damage to a biological tissue (target object) deviated from the treatment focal-point information, by compensating the difference between the focal-point information and the actual focusing location information, thereby treating the site to be treated in a precise manner.

[Brief Description of Drawings]

[0010]

FIG. 1 is a block diagram of an ultrasound medical apparatus according to some embodiments.

FIG. 2 is a schematic diagram of a high-intensity ultrasound having an intensity lower than a predetermined output value according to some embodiments.

FIG. 3 is a schematic diagram for illustrating focal-point information on an image picture according to some embodiments.

FIG. 4 is a flowchart of a method for compensating for a focal point according to some embodiments.

FIGs. 5A and 5B are schematic diagrams for illustrating a difference between the focal-point information and the actual focusing location information according to some embodiments.

FIG. 6 is a schematic diagram for illustrating a process of generating a three-dimensional image picture by an ultrasound medical apparatus according to some embodiments.

FIG. 7 is a schematic diagram of a treatment transducer including a plurality of groups according to some embodiments.

REFERENCE NUMERALS

[0011]

100:	Ultrasound Medical Apparatus	110:	Imaging Transducer
112:	Treatment Transducer	120:	Analog-to-digital Converter
130:	Image Processing Unit	132:	Beamformer
134:	Signal Processing Unit	136:	Scan Converter
138:	Focusing-location Extracting Unit	139:	Compensating Unit
140:	Display Unit	150:	User Input Unit
160:	Storage Unit		

[Detailed Description]

[0012] Hereinafter, at least one embodiment of the present disclosure will be described in detail with reference to the accompanying drawings.

[0013] A first imaging ultrasound described in some embodiments refers to an ultrasound that is transmitted to a target object to obtain an image by using an imaging transducer 110. On the other hand, a second imaging ultrasound described in some embodiments refers to an ultrasound that is transmitted to the target object by the imaging transducer 110 at a time when a high-intensity ultrasound is transmitted by a treatment transducer 112 and reaches a focal point. When a high-intensity ultrasound having an intensity lower than a predetermined output value is transmitted by the treatment transducer 112 based on a sort of synchronization between the imaging transducer 110 and the treatment transducer 112, the imaging transducer 110 can determine the time when the high-intensity ultrasound reaches the focal point and transmit the second imaging ultrasound to the target object according to the arrival time of the high-intensity ultrasound. Such a synchronization can be set based on a distance information between the imaging transducer 110 and the focal point, distance information between the treatment transducer 112 and the focal point, and traveling speed information of the first imaging ultrasound, the second imaging ultrasound, and the high-intensity ultrasound traveling in the medium of the target object.

[0014] A diagnostic image described in some embodiments includes a B-mode image, a C-mode image, and the like. The B-mode image is a grayscale image in an image mode for displaying a motion of a target object, and the C-mode image is an image in a color flow image mode. A BC-mode image is an image in an image mode for displaying a blood flow or a motion of the target object by using a Doppler effect, which simultaneously provides the B-mode image and the C-mode image to provide anatomic information together with the blood flow and the motion of the target object. The B-mode is a grayscale image mode for displaying the motion of the target object, and the C-mode is a color flow image mode for displaying the blood flow or the motion of the target object. An ultrasound medical apparatus 100 according to some embodiments is capable of simultaneously providing the B-mode image and the C-mode image that is a color flow image; however, for the purpose of illustration, descriptions are given with an assumption that a diagnostic image is a B-mode image that is provided by the ultrasound medical apparatus 100 in the present disclosure.

[0015] FIG. 1 is a block diagram of an ultrasound medical apparatus according to some embodiments.

[0016] An ultrasound medical apparatus 100 includes an imaging transducer 110, a treatment transducer 112, an analog-to-digital converter 120, an image processing unit 130, a display unit 140, a user input unit 150, and a storage unit 160. To illustrate an exemplary technical idea of some embodiments, the ultrasound medical apparatus 100 specifically includes the imaging transducer 110, treatment transducer 112, analog-to-digital converter 120, image processing unit 130, display unit 140, user input unit 150 and storage unit 160 in some embodiments, although one of ordinary skill in the pertinent art would appreciate that various modifications, additions and substitutions are possible in the constituent elements of the ultrasound medical apparatus 100 without departing from the idea and scope of the embodiments.

[0017] The imaging transducer 110 includes an imaging transducer array, transmits a first imaging ultrasound and a second imaging ultrasound to a target object, and receives a first echo signal and a second echo signal reflected from the target object. This transducer converts an electrical analog signal into an ultrasound, transmits the ultrasound to the target object, converts the echo signal reflected from the target object into an electrical analog signal, and includes a plurality of transducer elements coupled to each other. This imaging transducer 110 includes one-dimensional (1 D), 1.25D, 1.5D, 1.75D, or 2D transducer array. For example, when the imaging transducer 110 includes 1 D, 1.25D, 1.5D, or 1.75D transducer array, the 1 D, 1.25D, 1.5D, or 1.75D transducer array transmits the first imaging ultrasound or the second imaging ultrasound while rotating in a predetermined angle (0° to 360°), and forms a first reception signal or a second reception signal by receiving the first echo signal or the second echo signal from the target object. On the other hand, when the imaging transducer 110 includes 2D transducer array, the imaging transducer 110 transmits the first imaging ultrasound or the second imaging ultrasound via the 2D transducer array without a rotation, and forms the first reception signal or the second reception signal by receiving the first echo signal or the second echo signal from the target object. The imaging transducer 110 transmits an ultrasound beam, which is focused by appropriately delaying an input time of a pulse inputted to each transducer, to the target object along a transmission scanline. On the other hand, the first echo signal or the second echo signal reflected from the target object is inputted to each transducer with a different reception time, and each transducer outputs the first echo signal or the second echo signal to a beamformer 132.

[0018] The imaging transducer 110 transmits the first imaging ultrasound to the target object, receives the first echo signal reflected from the target object, and forms the first reception signal based on the first echo signal. In some embodiments, the imaging transducer 110 further transmits the second imaging ultrasound to the target object at a time when the high-intensity ultrasound reaches the focal point, receives the second echo signal reflected in response to the second imaging ultrasound, and forms the second reception signal based on the second echo signal. When a high-intensity ultrasound having an intensity lower than a predetermined output value reaches the focal point, although a human tissue corresponding to the focal point is not damaged, the corresponding human tissue is depressed or a density of the human tissue is changed by the high-intensity ultrasound having the intensity lower than the predetermined output

value. The second imaging ultrasound is transmitted to the target object at the time when the high-intensity ultrasound reaches the focal point, and then reflected second echo signal is received, which can detect the spot where the human tissue is depressed or the density of the human tissue is changed by the high-intensity ultrasound having the intensity lower than the predetermined output value. In other words, when the focal point is irradiated with an acoustic pressure having a high amplitude (high-intensity ultrasound), the medium is instantly depressed to cause a change of density and wave velocity, and the ultrasound is reflected due to a difference in the acoustic impedance. In order to determine the time when the high-intensity ultrasound reaches the focal point by the imaging transducer 110, when the high-intensity ultrasound having the intensity lower than the predetermined output value is transmitted by the treatment transducer 112 under a sort of synchronization between the imaging transducer 110 and the treatment transducer 112, the imaging transducer 110 transmits the second imaging ultrasound to the target object according to the arrival time of the high-intensity ultrasound. Such a synchronization can be set based on the distance information between the imaging transducer 110 and the focal point, the distance information between the treatment transducer 112 and the focal point, and the traveling speed information of the first imaging ultrasound, the second imaging ultrasound, and the high-intensity ultrasound traveling in the medium of the target object.

[0019] The treatment transducer 112 includes a high-intensity transducer array, and transmits a high-intensity ultrasound having an intensity lower than a predetermined output value or a high-intensity ultrasound for treatment (corresponding to the predetermined output value) to a specific region of the target object. The treatment transducer 112 transmits the high-intensity ultrasound having the intensity lower than the predetermined output value to a focal point corresponding to focal-point information. The treatment transducer 112 transmits a high-intensity ultrasound having a duty ratio lower than a predetermined ratio and a pulse amplitude exceeding a predetermined threshold value to the target object. For example, the treatment transducer 112 transmits the high-intensity ultrasound for treatment and the high-intensity ultrasound for pre-targeting. That is, the treatment transducer 112 transmits the high-intensity ultrasound for pre-targeting with a low intensity (with the duty ratio lower than the predetermined ratio, e.g., equal to or smaller than 5%) not to damage the human tissue (target object) or transmits the high-intensity ultrasound with a high intensity (with the duty ratio of the predetermined ratio of, e.g., 100%). At this time, the treatment transducer 112 repeatedly transmits the high-intensity ultrasound having the intensity lower than the predetermined output value for predetermined times. For example, the treatment transducer 112 makes a test pre-targeting by performing about two repeat transmissions of the high-intensity ultrasound having the intensity lower than the predetermined output value to the target object. In some embodiments, the two repeat transmissions are good for the high-intensity ultrasound having the intensity lower than the predetermined output value; however, the present disclosure is not limited thereto.

[0020] An exemplary operation of the treatment transducer 112 is described below. The treatment transducer 112 transmits the high-intensity ultrasound (one having the intensity lower than the predetermined output value or having the intensity corresponding to the predetermined output value) to a specific site determined via the user input unit 150. In this case, the user first transmits the first imaging ultrasound to the target object via the imaging transducer 110, receives the first echo signal reflected from the target object, and determines the specific site (focal point corresponding to the focal-point information) of the target object through a first image picture generated based on the first reception signal obtained upon receiving the first echo signal. In order for the user to determine the specific site (focal point corresponding to the focal-point information), the user inputs a location value corresponding to the specific site via the user input unit 150 or determines the corresponding site by using a direction key such as a joystick. With this operation, the high-intensity ultrasound can be transmitted to the specific site of the target object, such as cancer tissue, tumor tissue, lesion tissue, or the like. In some embodiments, the treatment transducer 112 is manufactured in a circular shape with the imaging transducer 110 arranged at the center; however, the present disclosure is not limited to this scheme.

[0021] In some embodiments, the treatment transducer 112 is divided into a plurality of groups. The treatment transducer 112 includes a high-intensity transducer array which is divided into a first transducer that transmits a first high-intensity ultrasound having an intensity lower than a predetermined output value based on a fundamental frequency to the specific site of the target object, a second transducer that transmits a second high-intensity ultrasound having an intensity lower than a predetermined output value based on a second frequency to the specific site of the target object, and an N-th transducer that transmits an N-th high-intensity ultrasound having an intensity lower than a predetermined output value based on an N-th frequency to the specific site of the target object. The treatment transducer 112 includes a first group including at least one first transducer, a second group including at least one second transducer, and an N-th group including at least one N-th transducer. Further, in some embodiments, the treatment transducer 112 includes the first transducer, the second transducer, and the N-th transducer arranged in a random manner. In this case, the treatment transducer 112 transmits a combination of a plurality of high-intensity ultrasounds with a ratio loaded to the specific region of the target object, or transmits the high-intensity ultrasound having the intensity lower than the predetermined output value by using each of the groups.

[0022] The analog-to-digital converter 120 converts the electrical analog signal obtained by the transducer of the imaging transducer 110 or the treatment transducer 112 into an electrical digital signal. FIG. 1 shows that the analog-to-digital converter 120 is arranged between the transducer (the imaging transducer 110 or the treatment transducer

112) and the beamformer 132, although the present invention is not limited to this scheme. One of ordinary skill in the pertinent art would appreciate that various modifications are possible in the arrangement of the analog-to-digital converter 120 between other modules in the ultrasound medical apparatus 100 without departing from the idea and scope of the embodiments.

5 **[0023]** The image processing unit 130 generates a first image picture by using the first reception signal generated based on the first echo signal corresponding to the first imaging ultrasound from the imaging transducer 110. Further, the image processing unit 130 generates a second image picture by using the second reception signal generated based on the second echo signal corresponding to the second imaging ultrasound from the imaging transducer 110.

10 **[0024]** Further, the image processing unit 130 is configured to set a region of interest (ROI) on the first image picture by an operation or an instruction from the user input unit 150. The image processing unit 130 receives the first echo signal from an image transducer that is rotated by a predetermined angle (0° to 360°) or from the imaging transducer 110 that includes a two-dimensional (2D) transducer array, and forms the first image picture as a three-dimensional image based on the received first echo signal. Further, the image processing unit 130 receives the second echo signal from the imaging transducer that is rotated by a predetermined angle (0° to 360°) or from the imaging transducer 110 that includes a 2D transducer array, and forms the second image picture as a three-dimensional image based on the received second echo signal.

15 **[0025]** The image processing unit 130 includes a beamformer 132, a signal processing unit 134, a scan converter 136, a focusing-location extracting unit 138, and a compensating unit 139. The beamformer 132 focuses the first echo signal and the second echo signal received by each transducer of the imaging transducer 110, and generates frame data as raw data. The beamformer 132 generates a receive focusing signal based on the electrical digital signal obtained by the analog-to-digital converter 120. Further, the beamformer 132 is configured to generate the receive focusing signal by adding an appropriate delay to each electrical digital signal and then summing the electrical digital signals, considering a time required to reach each transducer of the imaging transducer 110 and the treatment transducer 112 from the target object. The beamformer 132 focuses the ultrasound on a focal point (specific site) corresponding to the focal-point information by adjusting driving timing of each transducer in the imaging transducer 110 and the treatment transducer 112 when the imaging transducer 110 transmits the first imaging ultrasound or the second imaging ultrasound or the treatment transducer 112 transmits the high-intensity ultrasound. Further, the beamformer 132 focuses the first echo signal and the second echo signal by adding a time delay to each of the first imaging ultrasound signal and the second imaging ultrasound signal of the imaging transducer 110, considering the difference in time required for the first echo signal and the second echo signal reflected at the target subject to reach each transducer of the imaging transducer 110. The beamformer 132 in its implementation includes an imaging beamformer corresponding to the imaging transducer 110 and a treatment beamformer corresponding to the treatment transducer 112, and each of the imaging beamformer and the treatment beamformer includes a transmit beamformer and a receive beamformer.

20 **[0026]** The signal processing unit 134 generates the first image picture or the second image picture by performing a digital signal processing on the frame data signal generated by the beamformer 132. That is, the signal processing unit 134 generates the first image signal based on the first reception signal received from the imaging transducer 110, and outputs the first image picture to the display unit 140. Further, the signal processing unit 134 generates the second image signal based on the second reception signal received from the imaging transducer 110, and outputs the second image picture to the display unit 140. The scan converter 136 converts a data format of the first image picture or the second image picture into a data format that can be used in the display unit 140 of a predetermined scanline display format. That is, the scan converter 136 converts the first image picture signal or the second image picture signal into a signal of a data format that can be actually displayed on the display unit 140.

25 **[0027]** The focusing-location extracting unit 138 according to some embodiments extracts actual focusing-location information by using a change amount of the second image picture generated based on the second imaging ultrasound transmitted at the time when the high-intensity ultrasound reaches the focal point. That is, the imaging transducer 110 transmits the second imaging ultrasound at the time when the high-intensity ultrasound having the intensity lower than the predetermined output value transmitted by the treatment transducer 112 reaches the focal point, receives the second echo signal reflected in response to the second imaging ultrasound, and generates the second reception signal based on the second echo signal. Thereafter, the signal processing unit 134 generates the second image picture based on the second reception signal, and at this time, the focusing-location extracting unit 138 extracts the actual focusing-location information by using the change amount of the second image picture. In order to detect the change amount of the second image picture, the focusing-location extracting unit 138 extracts a location having the largest change amount of gray-scale of the second image picture, and determines the location having the largest change amount as the actual focusing-location information. The gray-scale is an index of brightness difference, which is a gradual step ranging from white to black. In a digital image, each pixel value indicates a single sample image, and the digital image delivers luminous intensity information only. When the second image picture is a two-dimensional image, the focusing-location extracting unit 138 determines two-dimensional coordinates as the actual focusing-location information (x_1, z_1) , and when the second image picture is a three-dimensional image, it determines three-dimensional coordinates as the actual focusing-

location information (x_1, y_1, z_1).

[0028] The compensating unit 139 according to some embodiments extracts a compensation value based on the focal-point information and the actual focusing-location information, and then applies the compensation value to the focal-point information. When the second image picture is a two-dimensional image, the compensating unit 139 generates the compensation value ($\Delta x, \Delta z$) based on a difference ($x - x_1, z - z_1$) between the focal-point information (x, z) on two-dimensional coordinates and the actual focal-point information (x_1, z_1) on the two-dimensional coordinates, and when the second image picture is a three-dimensional image, it generates the compensation value ($\Delta x, \Delta y, \Delta z$) based on a difference ($x - x_1, y - y_1, z - z_1$) between the focal-point information (x, y, z) on three-dimensional coordinates and the actual focal-point information (x_1, y_1, z_1) on the three-dimensional coordinates. Further, when the second image picture is a two-dimensional image, the compensating unit 139 generates compensation location information ($x + \Delta x, z + \Delta z$) by applying the compensation value ($\Delta x, \Delta z$) to the focal-point information (x, z) on the two-dimensional coordinates, and when the second image picture is a three-dimensional image, it generates compensation location information ($x + \Delta x, y + \Delta y, z + \Delta z$) by applying the compensation value ($\Delta x, \Delta y, \Delta z$) to the focal-point information (x, y, z) on the three-dimensional coordinates. That is, when the treatment transducer 112 transmits the high-intensity ultrasound for treatment with the focal-point information set by the user input unit 150, a possible difference between the focal-point information (x, y, z) and the actual focusing-location information (x_1, y_1, z_1) risks damage to a biological tissue (target object) deviated from the focal-point information to be treated may be damaged. Therefore, the compensating unit 139 according to some embodiments detects the difference between the focal-point information (x, y, z) and the actual focusing-location information (x_1, y_1, z_1), and compensates for the distance ($\Delta x, \Delta y, \Delta z$) by the difference, thus accurately treating the site at issue.

[0029] The display unit 140 outputs the first image picture or the second image picture received from the image processing unit 130 as a B-mode image or a C-mode image. The user input unit 150 receives an instruction by an operation or an input of a user. The user instruction includes a control instruction to control the ultrasound medical apparatus 100 and the like. Further, the user input unit 150 receives an input of the focal-point information on the first image picture. That is, when the first image picture is a two dimensional image, the user input unit 150 receives an input of the focal-point information of two-dimensional coordinates (x, z), and when the first image picture is a three-dimensional image, it receives an input of the focal-point information of three-dimensional coordinates (x, y, z). Moreover, the user input unit 150 is configured to set an ROI on the first image picture by an operation or an instruction from the user. The storage unit 160 is for storing various data for driving the ultrasound medical apparatus 100, which stores at least one of the first image picture or the second image picture. Further, the storage unit 160 stores the first reception signal or the second reception signal generated by using the imaging transducer 110.

[0030] FIG. 2 is a schematic diagram of a high-intensity ultrasound having an intensity lower than a predetermined output value according to some embodiments.

[0031] As shown in FIG. 2, the high-intensity ultrasound having the intensity lower than the predetermined output value according to some embodiments has a duty ratio lower than a predetermined ratio and a pulse amplitude exceeding a predetermined threshold value. That is, the treatment transducer 112 is configured to transmit a high-intensity ultrasound for treatment and another high-intensity ultrasound for pre-targeting. The treatment transducer 112 transmits the high-intensity ultrasound for pre-targeting with a low intensity (with the duty ratio lower than the predetermined ratio, e.g., equal to or smaller than 5%) not to damage the human tissue (target object). On the other hand, the treatment transducer 112 transmits the high-intensity ultrasound with a high intensity (with the duty ratio of the predetermined ratio of, e.g., 100%) to treat the human tissue. That is, in the example shown in FIG. 2, the treatment transducer 112 transmits the high-intensity ultrasound for pre-targeting with a low intensity (with the duty ratio lower than the predetermined ratio, e.g., equal to or smaller than 5%) not to damage the human tissue (target object). At this time, the treatment transducer 112 transmits the high-intensity ultrasound with a low intensity (with the duty ratio lower than the predetermined ratio, e.g., equal to or smaller than 5%) not to damage the human tissue (target object) in a repeated manner for a predetermined number of times to a specific site of the target object (focal point corresponding to the focal-point information). That is, the treatment transducer 112 performs a number of (about two) repeat transmissions of the high-intensity ultrasound having the intensity lower than the predetermined output value to the target object, thus performing the pre-targeting.

[0032] As shown in FIG. 2, in some embodiments, the high-intensity ultrasound that is set as low as not to damage the human tissue (target object) (to have a duty ratio lower than a predetermined ratio (e.g., equal to or smaller than about 5%)) has a pulse amplitude of equal to or larger than about 3 MPa; however, the present invention is not limited to this scheme. The unit MPa is a unit of acoustic pressure, and the acoustic pressure can be represented as an acoustic Intensity (unit: W/cm²). That is, the acoustic pressure of 3 MPa can be converted into the acoustic intensity of 580 W/cm². When treating the human tissue (target object), energy is equal to power x time, and hence in order to reduce the damage, the time should be decreased. To this end, the duty ratio is set to be small (e.g., equal to or smaller than about 5%).

[0033] FIG. 3 is a schematic diagram for illustrating focal-point information on an image picture according to some embodiments.

[0034] In some embodiments, the ultrasound medical apparatus 100 transmits the first imaging ultrasound or the second imaging ultrasound to the target object, receives the first echo signal or the second echo signal reflected from

the target object, generates the first reception signal or the second reception signal based on the first echo signal or the second echo signal, generates the first image picture or the second image picture based on the first reception signal or the second reception signal, and then outputs the first image picture or the second image picture to the display unit 140. In some embodiments, the first image picture or the second image picture generated by the ultrasound medical apparatus 100 is a three-dimensional image; however, the present invention is not limited to this scheme. When the imaging transducer 110 in the ultrasound medical apparatus 100 includes a 2D transducer array, the ultrasound medical apparatus 100 generates the first image picture as a three-dimensional image based on the first echo signal received corresponding to the first imaging ultrasound transmitted by the 2D transducer array. On the other hand, when the imaging transducer 110 in the ultrasound medical apparatus 100 includes a 2D transducer array, the ultrasound medical apparatus 100 generates the second image picture as a three-dimensional image based on the second echo signal received corresponding to the second imaging ultrasound transmitted by the 2D transducer array.

[0035] However, when the imaging transducer 110 in the ultrasound medical apparatus 100 includes a 1D, 1.25D, 1.5D and 1.75D transducer arrays, the ultrasound medical apparatus 100 transmits the first imaging ultrasound or the second imaging ultrasound while rotating the 1D, 1.25D, 1.5D and 1.75D transducer arrays by a predetermined angle (0° to 360°), receives the first echo signal or the second echo signal reflected from the target object, and then generates the first image picture or the second image picture based on the first echo signal or the second echo signal. As described above, when the first image picture or the second image picture is a three-dimensional image, as shown in FIG. 3, the coordinates can be represented as (x, y, z) . When the imaging transducer 110 includes a 1D, 1.25D, 1.5D and 1.75D transducer arrays, the first image picture or the second image can be generated as a two-dimensional image, and in this case, the coordinates can be represented as (x, z) .

[0036] FIG. 4 is a flowchart of a method of compensating for a focal point according to some embodiments.

[0037] The imaging transducer 110 of the ultrasound medical apparatus 100 transmits the first imaging ultrasound to the target object, receives the first echo signal reflected from the target object, and generates the first reception signal based on the first echo signal (step S410). The signal processing unit 134 of the ultrasound medical apparatus 100 generates the first image picture based on the first reception signal, and outputs the first image picture to the display unit 140 (step S420).

[0038] The user input unit 150 of the ultrasound medical apparatus 100 receives an input of the focal-point information on the first image picture (step S430). At step S430, when the first image picture is a two-dimensional image, the user input unit 150 receives the focal-point information of two-dimensional coordinates (x, z) , and when the first image picture is a three-dimensional image, it receives the focal-point information of three-dimensional coordinates (x, y, z) .

[0039] The treatment transducer 112 of the ultrasound medical apparatus 100 transmits the high-intensity ultrasound having the intensity lower than the predetermined output value to the focal point corresponding to the focal-point information (step S440). The high-intensity ultrasound having the predetermined output value has a duty ratio lower than a predetermined ratio and a pulse amplitude exceeding a predetermined threshold value. That is, the high-intensity ultrasound having the intensity lower than the predetermined output value means a high-intensity ultrasound having an intensity as low as not to damage the human tissue (target object) or having a duty ratio lower than a predetermined ratio, e.g., equal to or smaller about 5%.

[0040] The imaging transducer 110 of the ultrasound medical apparatus 100 transmits the second imaging ultrasound to the target object at the time when the high-intensity ultrasound from the treatment transducer 112 reaches the focal point, receives the second echo signal reflected in response to the second imaging ultrasound, and generates the second reception signal based on the second echo signal (step S450). The signal processing unit 134 of the ultrasound medical apparatus 100 generates the second image picture based on the second reception signal (step S460).

[0041] The focusing-location extracting unit 138 of the ultrasound medical apparatus 100 extracts the actual focusing-location information by using the change amount of the second image picture (step S470). At step S470, the focusing-location extracting unit 138 of the ultrasound medical apparatus 100 extracts a location having the largest change amount of the gray-scale of the second image picture, and determines the location having the largest change amount as the actual focusing-location information. At this time, when the second image picture is a two-dimensional image, the focusing-location extracting unit 138 of the ultrasound medical apparatus 100 determines the two-dimensional coordinates as the actual focusing-location information (x_1, z_1) , and when the second image picture is a three-dimensional image, it determines the three-dimensional coordinates as the actual focusing-location information (x_1, y_1, z_1) .

[0042] The compensating unit 139 of the ultrasound medical apparatus 100 extracts a compensation value based on the focal-point information and the actual focusing-location information, and then applies the compensation value to the focal-point information (step S480). At step S480, when the second image picture is a two-dimensional image, the compensating unit 139 of the ultrasound medical apparatus 100 generates the compensation value $(\Delta x, \Delta z)$ based on a difference $(x - x_1, z - z_1)$ between the focal-point information (x, z) on two-dimensional coordinates and the actual focal-point information (x_1, z_1) on the two-dimensional coordinates, and when the second image picture is a three-dimensional image, it generates the compensation value $(\Delta x, \Delta y, \Delta z)$ based on a difference $(x - x_1, y - y_1, z - z_1)$ between the focal-point information (x, y, z) on three-dimensional coordinates and the actual focal-point information (x_1, y_1, z_1) on the three-

dimensional coordinates. Thereafter, when the second image picture is a two-dimensional image, the compensating unit 139 of the ultrasound medical apparatus 100 generates compensation location information ($x + \Delta x$, $z + \Delta z$) by applying the compensation value (Δx , Δz) to the focal-point information (x , z) on the two-dimensional coordinates, and when the second image picture is the three-dimensional image, it generates compensation location information ($x + \Delta x$, $y + \Delta y$, $z + \Delta z$) by applying the compensation value (Δx , Δy , Δz) to the focal-point information (x , y , z) on the three-dimensional coordinates.

[0043] Steps S410 to S480 are described to be sequentially performed in FIG. 4 as a mere example for describing the technical idea of some embodiments, although one of ordinary skill in the pertinent art would appreciate that various modifications, additions and substitutions are possible by performing the sequences shown in FIG. 4 in a different order or at least one of Steps S410 to S480 in parallel without departing from the idea and scope of the embodiments, and hence the example shown in FIG. 4 is not limited to the chronological order.

[0044] The method of compensating the focal point shown in FIG. 4 can be implemented as a computer program, and can be recorded on a computer-readable medium. The computer-readable recording medium on which the method of compensating the focal point according to some embodiments is recordable includes any type of recording device on which data that can be read by a computer system are recordable. Examples of the computer-readable recording medium include a ROM, a RAM, a CD-ROM, a magnetic tape, a floppy disk, an optical data storage device, and the like, and also include one implemented in the form of carrier wave (e.g., transmission through the Internet). Further, the computer-readable recording medium can be distributed in computer systems connected via a network, and computer-readable codes can be stored and executed in a distributed mode. Moreover, functional programs, codes, and code segments for implementing some embodiments can be easily deduced by a programmer in technical fields to which some embodiments belong.

[0045] FIGs. 5A and 5B are schematic diagrams for illustrating a difference between the focal-point information and the actual focusing location information according to some embodiments.

[0046] As shown in FIGs. 5A and 5B, the user input unit 150 of the ultrasound medical apparatus 100 receives an input of the focal-point information on the first image picture. When the first image picture is a three-dimensional image, the user input unit 150 receives an input of the focal-point information of three-dimensional coordinates (x , y , z). As shown in FIGs. 5A and 5B, a location corresponding to the three-dimensional coordinates (x , y , z) is indicated by "+" on the first image picture. The treatment transducer 112 of the ultrasound medical apparatus 100 transmits the high-intensity ultrasound having the intensity lower than the predetermined output value to the focal point ("+" point shown in FIG. 5A) corresponding to the focal-point information of the target object. Here, the ultrasound medical apparatus 100 transmits the high-intensity ultrasound having the intensity lower than the predetermined output value to a subject to be treated (target object) by using the treatment transducer 112, to increase the reflected signal (second echo signal). That is, when the focal point is irradiated with an acoustic pressure with a large amplitude (high-intensity ultrasound), the medium is instantly depressed to cause changes of density and wave velocity, and hence the ultrasound is reflected due to a difference in the acoustic impedance.

[0047] Thereafter, the imaging transducer 110 of the ultrasound medical apparatus 100 transmits the second imaging ultrasound to the target object at the time when the high-intensity ultrasound having the intensity lower than the predetermined output value reaches the focal point. The imaging transducer 110 of the ultrasound medical apparatus 100 receives the second echo signal reflected in response to the second imaging ultrasound, and generates the second reception signal. The signal processing unit 134 of the ultrasound medical apparatus 100 generates the second image picture based on the second reception signal. The focusing-location extracting unit 138 of the ultrasound medical apparatus 100 extracts the actual focusing-location information (location indicated by "•" in FIG. 5A) by using a change amount of the second image picture.

[0048] The focusing-location extracting unit 138 of the ultrasound medical apparatus 100 extracts a location having the largest change amount of the gray-scale of the second image picture (location indicated by "•" in FIG. 5A), and determines the location having the largest change amount as the actual focusing-location information (x_1 , y_1 , z_1). The ultrasound medical apparatus 100 extracts the location having the largest change amount of the gray-scale (location indicated by "•" in FIGs. 5A and 5B) in the ROI set on the second image picture by using the focusing-location extracting unit 138, to determine the actual focusing-location information (x_1 , y_1 , z_1).

[0049] The compensating unit 139 of the ultrasound medical apparatus 100 extracts the compensation value based on the focal-point information (location indicated by "+" in FIG. 5A) and the actual focusing-location information (location indicated by "•" in FIG. 5A), and then applies the compensation value to the focal-point information. Thereafter, as shown in FIG. 5B, when the second image picture is a three-dimensional image, the compensating unit 139 of the ultrasound medical apparatus 100 generates the compensation value (Δx , Δy , Δz) based on a difference ($x - x_1$, $y - y_1$, $z - z_1$) between the focal-point information (x , y , z) on three-dimensional coordinates and the actual focal-point information (x_1 , y_1 , z_1) on the three-dimensional coordinates. As shown in FIG. 5B, when the second image picture is a three-dimensional image, the compensating unit 139 of the ultrasound medical apparatus 100 generates compensation location information ($x + \Delta x$, $y + \Delta y$, $z + \Delta z$) by applying the compensation value (Δx , Δy , Δz) to the focal-point information (x , y , z) on the

three-dimensional coordinates.

[0050] FIG. 6 is a schematic diagram for illustrating a process of generating a three-dimensional image picture by an ultrasound medical apparatus according to some embodiments.

[0051] In some embodiments, the ultrasound medical apparatus 100 transmits the first imaging ultrasound or the second imaging ultrasound to the target object, receives the first echo signal or the second echo signal reflected from the target object, generates the first reception signal or the second reception signal based on the first echo signal or the second echo signal, generates the first image picture or the second image picture based on the first reception signal or the second reception signal, and then outputs the first image picture or the second image picture to the display unit 140.

[0052] In some embodiments, the first image picture or the second image picture generated by the ultrasound medical apparatus 100 is a three-dimensional image; however, the present disclosure is not limited to this scheme. When the imaging transducer 110 of the ultrasound medical apparatus 100 includes a two-dimensional transducer array, the ultrasound medical apparatus 100 receives the first echo signal corresponding to the first imaging ultrasound transmitted by the 2D transducer array, and generates the first image picture as a three-dimensional image based on the first echo signal, or receives the second echo signal corresponding to the second imaging ultrasound, and generates the second image picture as a three-dimensional image based on the second echo signal. However, when the imaging transducer 110 of the ultrasound medical apparatus 100 includes a 1 D, 1.25D, 1.5D and 1.75D transducer arrays, the ultrasound medical apparatus 100 can generate a three-dimensional image in a manner shown in FIG. 3.

[0053] As shown in FIG. 6, the ultrasound medical apparatus 100 transmits the first imaging ultrasounds to the target object while rotating the 1 D, 1.25D, 1.5D and 1.75D transducer arrays by a predetermined angle (0° to 360°), receives the respective first echo signals reflected from the target object, and then generates the first image picture as a three-dimensional image based on the respective first echo signals, or transmits the second imaging ultrasounds to the target object, receives the respective second echo signals reflected from the target object, and then generates the second image picture as a three-dimensional image based on the second echo signal. For example, the first imaging ultrasounds or the second ultrasounds are transmitted to the target object by using the imaging transducer 110 in a horizontal direction and a vertical direction with respect to the treatment transducer 112, and based on the first echo signals or the second echo signals reflected and received from the target object, the first image picture or the second image picture can be generated as a three-dimensional image.

[0054] FIG. 7 is a schematic diagram of a treatment transducer including a plurality of groups according to some embodiments.

[0055] To take account of the nonuniform internal tissues of a human body (target object), which cause variation of the phase of each acoustic wave reaching the set focal point through different human tissues, the high-intensity transducer array of the treatment transducer 112 can be divided into a plurality (at least two or more) groups. The ultrasound medical apparatus 100 utilizes the treatment transducer 112 to transmit the high-intensity ultrasound having the intensity lower than the predetermined output value to the subject to be treated (target object).

[0056] A plurality of grouped divisions of the high-intensity transducer array can also be used successfully to transmit the high-intensity ultrasound having the intensity lower than the predetermined output value to the subject to be treated (target object) and then receive the second echo signals before harming the subject to be treated (human tissue). The ultrasound medical apparatus 100 compensates for the phase by applying an optimization algorithm to the plurality of groups. In this case, the applicable phase is one that occurs when the reflected signal is maximized out of the focal point information, and the applicable optimum solution may be one among a set of various solutions for use as the optimization algorithm.

[0057] The treatment transducer 112 is described below with reference to FIG. 7. As shown in FIG. 7, in the treatment transducer 112, the transducers are arranged to form a specific group depending on the first high-intensity ultrasound having an intensity lower than a predetermined output value, the second high-intensity ultrasound having an intensity lower than the predetermined output value, and the N-th high-intensity ultrasound having an intensity lower than the predetermined output value to be transmitted to the focal point corresponding to the focal-point information or arranged in a random manner. The treatment transducer 112 transmits a combination of a plurality of high-intensity ultrasounds with a ratio loaded to the specific region of the target object, or transmits the high-intensity ultrasound having the intensity lower than the predetermined output value by using each of the groups.

[0058] Further, the treatment transducer 112 transmits the high-intensity ultrasound to a specific site (the focal point corresponding to the focal-point information) adjusted via the user input unit 150. A user (i.e., an operator) first transmits the first imaging ultrasound to the target object by using the imaging transducer 110, and determines a specific region of the target object through the first image picture generated based on the first reception signal generated based on the first echo signal reflected from the target object. In order for the user (operator) to determine the specific region, the user inputs a location value corresponding to the specific region to the user input unit 150 or determines the corresponding location by adjusting a direction key such as a joystick. With this operation, the high-intensity ultrasound having the intensity lower than the predetermined output value can be transmitted to the specific region of the target object, such as cancer tissue, tumor tissue, or lesion tissue. As shown in FIG. 7, the treatment transducer 112 can be manufactured

in a circular shape with the imaging transducer 110 arranged at the center; however, the present invention is not limited to this scheme.

[0059] The treatment transducer 112 transmits a combination of a plurality of high-intensity ultrasounds with a ratio loaded to the specific site of the target object under a control of the control unit. As shown in FIG. 7, the treatment transducer 112 includes a first transducer that transmits a first high-intensity ultrasound having an intensity lower than a predetermined output value based on a fundamental frequency to the specific site of the target object, a second transducer that transmits a second high-intensity ultrasound having an intensity lower than a predetermined output value based on a second frequency to the specific site of the target object, and an N-th transducer that transmits an N-th high-intensity ultrasound having an intensity lower than a predetermined output value based on an N-th frequency to the specific site of the target object. The first transducer, the second transducer, and the N-th transducer are physically similar transducers; however, in some embodiments, they are distinguishably described as the first transducer, the second transducer, and the N-th transducer for the purpose of illustration.

[0060] More specifically, as shown in FIG. 7, the treatment transducer 112 includes a first group including the first transducers, a second group including the second transducers, and an N-th group including the N-th transducers. Further, in some embodiments, as shown in FIG. 7, the treatment transducer 112 includes the first transducers, the second transducers, and the N-th transducers arranged in a random manner.

[0061] Although exemplary embodiments of the present disclosure have been described for illustrative purposes, those skilled in the art will appreciate that various modifications, additions and substitutions are possible, without departing from the spirit and scope of the claimed invention. Therefore, exemplary embodiments of the present disclosure have been described for the sake of brevity and clarity. Accordingly, one of ordinary skill in the pertinent art would understand the scope of the claimed invention is not to be limited by the explicitly described above embodiments but by the claims and equivalents thereof.

CROSS-REFERENCE TO RELATED APPLICATION

[0062] If applicable, this application claims priority under 35 U.S.C §119(a) of Patent Application No. 10-2013-0021931, filed on February 28, 2013 in Korea, the entire content of which is incorporated herein by reference. In addition, this non-provisional application claims priority in countries, other than the U.S., with the same reason based on the Korean patent application, the entire content of which is hereby incorporated by reference.

Claims

1. An ultrasound medical apparatus, comprising:

an imaging transducer configured
to transmit a first imaging ultrasound to a target object,
to receive a first echo signal reflected from the target object, and
to generate a first reception signal based on the first echo signal;
a signal processing unit configured
to generate a first image picture based on the first reception signal, and
to output the first image picture to a display unit;
a user input unit configured to receive an input of focal-point information on the first image picture;
a treatment transducer configured to transmit a high-intensity ultrasound having an intensity lower than a predetermined output value to a focal point corresponding to the focal-point information;
a focusing-location extracting unit configured to extract actual focal-point information by using a change amount of a second image picture that is generated based on a second imaging ultrasound transmitted at a time when the high-intensity ultrasound reaches the focal point; and
a compensating unit configured
to generate a compensation value based on the focal-point information and the actual focal-point information, and
to apply the compensation value to the focal-point information.

2. The ultrasound medical apparatus according to claim 1, wherein

the imaging transducer is further configured
to transmit the second imaging ultrasound to the target object at the time when the high-intensity ultrasound reaches the focal point,
to receive a second echo signal reflected from the target object in response to the second imaging ultrasound, and
to generate a second reception signal based on the second echo signal, and

the signal processing unit generates the second image picture based on the second reception signal.

3. The ultrasound medical apparatus according to claim 1, wherein the treatment transducer is configured to transmit the high-intensity ultrasound having a duty ratio lower than a predetermined ratio and a pulse amplitude exceeding a predetermined threshold value.
4. The ultrasound medical apparatus according to claim 3, wherein the treatment transducer is configured to perform a predetermined repeat transmissions of the high-intensity ultrasound.
5. The ultrasound medical apparatus according to claim 1, wherein the signal processing unit is configured to receive the first echo signal from the imaging transducer that is rotated by a predetermined angle in a range from 0 degrees to 360 degrees, or to receive the first echo signal from the imaging transducer including a two-dimensional transducer array, and to generate the first image picture as a three-dimensional image based on the first echo signal.
6. The ultrasound medical apparatus according to claim 1, wherein the focusing-location extracting unit is configured to extract a location having a largest change amount of a gray-scale of the second image picture, and to determine the location having the largest change amount of the gray-scale of the second image picture as the actual focal-point information.
7. The ultrasound medical apparatus according to claim 6, wherein the focusing-location extracting unit is configured to determine, when the second image picture is a two-dimensional image, two-dimensional coordinates of the location having the largest change amount of the gray-scale of the second image picture as the actual focal-point information (x_1, z_1) , and to determine, when the second image picture is a three-dimensional image, three-dimensional coordinates of the location having the largest change amount of the gray-scale of the second image picture as the actual focal-point information (x_1, y_1, z_1) .
8. The ultrasound medical apparatus according to claim 1, wherein the compensating unit is configured to generate, when the second image picture is a two-dimensional image, the compensation value $(\Delta x, \Delta z)$ based on a difference $(x - x_1, z - z_1)$ between the focal-point information (x, z) on two-dimensional coordinates and the actual focal-point information (x_1, z_1) on the two-dimensional coordinates, and to generate, when the second image picture is a three-dimensional image, the compensation value $(\Delta x, \Delta y, \Delta z)$ based on a difference $(x - x_1, y - y_1, z - z_1)$ between the focal-point information (x, y, z) on three-dimensional coordinates and the actual focal-point information (x_1, y_1, z_1) on the three-dimensional coordinates.
9. The ultrasound medical apparatus according to claim 8, wherein the compensating unit is configured to generate, when the second image picture is the two-dimensional image, compensation location information $(x + \Delta x, z + \Delta z)$ by applying the compensation value $(\Delta x, \Delta z)$ to the focal-point information (x, z) on the two-dimensional coordinates, and to generate, when the second image picture is the three-dimensional image, compensation location information $(x + \Delta x, y + \Delta y, z + \Delta z)$ by applying the compensation value $(\Delta x, \Delta y, \Delta z)$ to the focal-point information (x, y, z) on the three-dimensional coordinates.
10. The ultrasound medical apparatus according to claim 1, wherein the user input unit is configured to receive, when the first image picture is a two-dimensional image, an input of the focal-point information of two-dimensional coordinates (x, z) , and to receive, when the first image picture is a three-dimensional image, an input of the focal-point information of three-dimensional coordinates (x, y, z) .
11. The ultrasound medical apparatus according to claim 1, wherein the treatment transducer includes:
 - at least one first transducer configured to transmit a first high-intensity ultrasound having an intensity lower than a predetermined output value with respect to a fundamental frequency to the focal point,
 - at least one second transducer configured to transmit a second high-intensity ultrasound having an intensity lower than a predetermined output value with respect to a secondary frequency to the focal point, and
 - at least one N-th transducer configured to transmit an N-th high-intensity ultrasound having an intensity lower than a predetermined output value with respect to an N-th frequency to the focal point.

12. The ultrasound medical apparatus according to claim 11, wherein the treatment transducer includes:

a first group consisting of the first transducers,
a second group consisting of the second transducers, and
an N-th group consisting of the N-th transducers.

13. The ultrasound medical apparatus according to claim 11, wherein the treatment transducer includes the first transducers, the second transducers, and the N-th transducers arranged in a random manner.

14. A method for compensating for a focal point in an ultrasound medical apparatus, the method comprising:

a first reception signal generating step including:

transmitting a first imaging ultrasound to a target object,
receiving a first echo signal reflected from the target object, and
generating a first reception signal based on the first echo signal;

a first image processing step including

generating a first image picture based on the first reception signal, and
outputting the first image picture to a display unit;

a focal point inputting step including receiving an input of focal-point information on the first image picture;

a high-intensity ultrasound transmitting step including transmitting a high-intensity ultrasound having an intensity lower than a predetermined output value to a focal point corresponding to the focal-point information;

an actual focal-point extracting step including extracting actual focal-point information by using a change amount of a second image picture that is generated based on a second imaging ultrasound transmitted at a time when the high-intensity ultrasound reaches the focal point; and

a compensating step including:

generating a compensation value based on the focal-point information and the actual focal-point information,
and
applying the compensation value to the focal-point information.

15. The method according to claim 14, wherein the actual focal-point extracting step further includes:

a second reception signal generating step including:

transmitting the second imaging ultrasound to the target object at the time when the high-intensity ultrasound reaches the focal point,
receiving a second echo signal reflected from the target object in response to the second imaging ultrasound,
and
generating a second reception signal based on the second echo signal, and

a second image processing step including generating the second image picture based on the second reception signal.

16. The method according to claim 14, wherein the high-intensity ultrasound transmitting step further includes transmitting the high-intensity ultrasound having a duty ratio lower than a predetermined ratio and a pulse amplitude exceeding a predetermined threshold value.

17. The method according to claim 14, wherein the actual focal-point extracting step further includes:

extracting a location having a largest change amount of a gray-scale of the second image picture, and
determining the location having the largest change amount of the gray-scale of the second image picture as the actual focal-point information.

18. The method according to claim 14, wherein the compensating step further includes:

generating, when the second image picture is a two-dimensional image, the compensation value (Δx , Δz) based

on a difference $(x - x_1, z - z_1)$ between the focal-point information (x, z) on two-dimensional coordinates and the actual focal-point information (x_1, z_1) on the two-dimensional coordinates, and
generating, when the second image picture is a three-dimensional image, the compensation value $(\Delta x, \Delta y, \Delta z)$
based on a difference $(x - x_1, y - y_1, z - z_1)$ between the focal-point information (x, y, z) on three-dimensional
coordinates and the actual focal-point information (x_1, y_1, z_1) on the three-dimensional coordinates.

19. The method according to claim 18, wherein the compensating step further includes:

generating, when the second image picture is the two-dimensional image, compensation location information
 $(x + \Delta x, z + \Delta z)$ by applying the compensation value $(\Delta x, \Delta z)$ to the focal-point information (x, z) on the two-
dimensional coordinates, and
generating, when the second image picture is the three-dimensional image, compensation location information
 $(x + \Delta x, y + \Delta y, z + \Delta z)$ by applying the compensation value $(\Delta x, \Delta y, \Delta z)$ to the focal-point information (x, y, z)
on the three-dimensional coordinates.




FIG. 1




FIG. 2





FIG. 3

**FIG. 4**




FIG. 5A




FIG. 5B




FIG. 6




FIG. 7

INTERNATIONAL SEARCH REPORT

International application No.

PCT/KR2013/001669

A. CLASSIFICATION OF SUBJECT MATTER

A61N 7/00(2006.01)i, A61B 8/00(2006.01)i, A61B 8/14(2006.01)i, A61B 18/00(2006.01)i, G01S 15/89(2006.01)i

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

A61N 7/00; A61B 8/12; G06T 19/20; G06T 5/50; A61B 8/08; A61B 8/00; A61N 7/02; A61B 8/14; A61B 18/00; G01S 15/89

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched
Korean Utility models and applications for Utility models: IPC as above
Japanese Utility models and applications for Utility models: IPC as aboveElectronic data base consulted during the international search (name of data base and, where practicable, search terms used)
eKOMPASS (KIPO internal) & Keywords: ultrasonic waves, transducer, focus, imaging, compensation, compensation

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	KR 10-1031504 B1 (SAMSUNG MEDISON CO.,LTD.) 29 April 2011 See abstract; claims 1-3 and 6	1-19
A	KR 10-2012-0010012 A (SOGANG UNIVERSITY RESEARCH FOUNDATION) 02 February 2012 See abstract; claims 1-4	1-19
A	KR 10-2010-0120091 A (SIEMENS MEDICAL SOLUTIONS USA, INC.) 12 November 2010 See abstract; claim 19; paragraph 62	1-19
A	JP 07-059777 A (ALOKA CO LTD.) 07 March 1995 See abstract; claim 1	1-19
A	KR 10-2011-0039506 A (SAMSUNG MEDISON CO.,LTD.) 19 April 2011 See abstract; claims 1 and 3	1-19

☐ Further documents are listed in the continuation of Box C.
 ☒ See patent family annex.

* Special categories of cited documents:

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier application or patent but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"I" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

"&" document member of the same patent family

Date of the actual completion of the international search

25 SEPTEMBER 2013 (25.09.2013)

Date of mailing of the international search report

25 SEPTEMBER 2013 (25.09.2013)

Name and mailing address of the ISA/KR


 Korean Intellectual Property Office
 Government Complex-Daejeon, 189 Seonsa-ro, Daejeon 302-701,
 Republic of Korea

Facsimile No. 82-42-472-7140

Authorized officer

Telephone No.

INTERNATIONAL SEARCH REPORT
Information on patent family members

International application No.

PCT/KR2013/001669

Patent document cited in search report	Publication date	Patent family member	Publication date
KR 10-1031504 B1	14/06/2011	NONE	
KR 10-2012-0010012 A	02/02/2012	NONE	
KR 10-2010-0120091 A	12/11/2010	CN 101879077 A	10/11/2010
		JP 2010-259806 A	18/11/2010
		US 2010-0280373 A1	04/11/2010
		US 2010-0317971 A1	16/12/2010
		US 8343050 B2	01/01/2013
JP 07-059777 A	07/03/1995	JP 3316268 B2	19/08/2002
KR 10-2011-0039506 A	19/04/2011	KR 10-1097539 B1	15/12/2011

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- KR 1020130021931 [0062]

专利名称(译)	焦点补偿方法及其超声波医疗设备		
公开(公告)号	EP2962728A4	公开(公告)日	2016-10-19
申请号	EP2013876137	申请日	2013-02-28
[标]申请(专利权)人(译)	爱飞纽医疗机械贸易有限公司		
申请(专利权)人(译)	ALPINION MEDICAL SYSTEMS CO. , LTD.		
当前申请(专利权)人(译)	ALPINION医疗系统CO.LTD.		
[标]发明人	SON KEONHO KANG KOOKJIN KIM DAE SEUNG KIM MYUNGDEOK JUN SUKHWAN		
发明人	SON, KEONHO KANG, KOOKJIN KIM, DAE SEUNG KIM, MYUNGDEOK JUN, SUKHWAN		
IPC分类号	A61N7/00 A61B8/00 A61B8/14 A61B18/00 G01S15/89 A61N7/02 A61B8/08		
CPC分类号	A61B8/58 A61B8/468 A61B8/5207 A61B8/54 A61N7/02 A61N2007/0052 A61N2007/0078 A61N2007/0086 A61N2007/0091 A61N2007/0095 G01S7/52049 G01S15/899		
优先权	1020130021931 2013-02-28 KR		
其他公开文献	EP2962728A1 EP2962728B1		
外部链接	Espacenet		

摘要(译)

公开了一种用于补偿焦点的方法和超声波医疗设备。提供了一种用于将具有低于预定输出值的强度的高强度超声波以如下方式发送到目标对象（患者的受影响区域）的方法和超声医疗装置：在焦点当传送高强度聚焦超声时，在所述目标对象上和实际超声聚焦位置处，并且通过使用所述超声的反射信号来补偿所述焦点的距离。