



(11)

EP 3 097 859 A1

(12)

EUROPEAN PATENT APPLICATION
published in accordance with Art. 153(4) EPC

(43) Date of publication:

30.11.2016 Bulletin 2016/48

(51) Int Cl.:

A61B 8/08 (2006.01)

A61B 8/12 (2006.01)

A61B 8/14 (2006.01)

(21) Application number: **14879286.4**

(86) International application number:

PCT/JP2014/078998

(22) Date of filing: **24.10.2014**

(87) International publication number:

WO 2015/111278 (30.07.2015 Gazette 2015/30)

(84) Designated Contracting States:

**AL AT BE BG CH CY CZ DE DK EE ES FI FR GB
GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO
PL PT RO RS SE SI SK SM TR**

Designated Extension States:

BA ME

(72) Inventor: **MIYAKI, Hironaka**

Hachioji-shi,

Tokyo 192-8507 (JP)

(74) Representative: **Gunzelmann, Rainer**

Wuesthoff & Wuesthoff

Patentanwälte PartG mbB

Schweigerstraße 2

81541 München (DE)

(30) Priority: **24.01.2014 JP 2014011840**


(71) Applicant: **Olympus Corporation**

Tokyo 192-8507 (JP)

(54) **ULTRASONIC OBSERVATION DEVICE, OPERATING METHOD FOR ULTRASONIC OBSERVATION DEVICE, AND OPERATING PROGRAM FOR ULTRASONIC OBSERVATION DEVICE**

(57) An ultrasound observation apparatus includes: a THI image data generation unit that generates image data of a grayscale THI image, based on first and second received signals which are obtained by transmitting first and second ultrasound signals having different phases or amplitudes to a specimen along the same line and receiving first and second ultrasound echo signals reflected from the specimen; a feature extraction unit that computes first and second features based on frequency spectra of the first and second ultrasound echo signals; a tissue characteristic image data generation unit that generates image data of a tissue characteristic image having a color display mode in an area whose value of a third feature based on the first and second features is greater than or equal to or less than or equal to a predetermined threshold; and an image composition unit that generates image data of a composite image where the THI image and the tissue characteristic image are combined.

FIG.15



EP 3 097 859 A1

Description

Field

5 **[0001]** The present invention relates to an ultrasound observation apparatus for observing tissue of a specimen using ultrasound waves, a method for operating the ultrasound observation apparatus, and an operating program for the ultrasound observation apparatus.

Background

10 **[0002]** In ultrasound observation techniques, there is known THI (tissue harmonic imaging) that performs image generation using the nonlinearity of body tissue (see, for example, Patent Literature 1). THI is a technique in which a plurality of ultrasound signals having different phases or amplitudes are sequentially transmitted, and received signals that represent ultrasound echoes of the ultrasound signals are subjected to addition, subtraction, or the like, by which a signal whose harmonic component is enhanced is obtained, and based on the signal an image is generated. According to THI, an ultrasound image with improved resolution and reduced artifacts can be obtained.

15 **[0003]** On the other hand, there is also known a technique in which a feature image that represents differences in tissue characteristics of body tissue is generated using frequency feature of ultrasound scattered in body tissue (see, for example, Patent Literature 2). In this technique, a frequency spectrum is computed by performing frequency analysis by performing fast Fourier transform (FFT) on a received signal representing an ultrasound echo, and a feature image is generated based on feature which is extracted by, for example, performing an approximation process on the frequency spectrum.

Citation List

25 Patent Literature

[0004]

30 Patent Literature 1: JP 2013-56033 A
Patent Literature 2: JP 5114609 B2

Summary

35 Technical Problem

[0005] However, the above-described feature image has a problem that since FFT computation is performed on a received signal included in a predetermined time range, it is difficult to obtain a high-resolution image like a THI image.

40 **[0006]** The present invention has been made in view of the foregoing, and an object of the invention is to provide an ultrasound observation apparatus, a method for operating the ultrasound observation apparatus, and an operating program for the ultrasound observation apparatus which are capable of generating an ultrasound image that makes it possible to identify tissue characteristics and that has excellent resolution.

Solution to Problem

45 **[0007]** In order to solve the above described problems and achieve the object, an ultrasound observation apparatus according to the invention includes: an ultrasound probe configured to transmit a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and to receive first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal; an image data generation unit configured to perform a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute a third received signal, and to generate image data of an ultrasound image having a predetermined display mode, using the third received signal; a feature extraction unit configured to extract at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and an image composition unit configured to generate image data of a composite image using the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first

and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

[0008] In the ultrasound observation apparatus, the feature extraction unit is configured to compute the first and second features, and to calculate a weighted sum of the first and second features using positive weights to extract a third feature, and the image composition unit is configured to generate the image data of the composite image having the display mode different from the predetermined display mode at the pixel position where the third feature is not less than or not more than the predetermined threshold.

[0009] The ultrasound observation apparatus further includes a second image data generation unit configured to generate image data of a second ultrasound image having a second display mode different from the predetermined display mode at the pixel position where the value based on at least one of the first and second features is not less than or not more than the predetermined threshold. The image composition unit is configured to combine the ultrasound image with the second ultrasound image to generate the image data of the composite image.

[0010] In the ultrasound observation apparatus, the predetermined display mode is grayscale display, and the second display mode is color display. The second image data generation unit is configured to: assign, to a first pixel whose value in the second ultrasound image based on at least one of the first and second features is not less than or not more than the predetermined threshold, a pixel value of a color according to the value; and assign a pixel value of zero to a second pixel other than the first pixel in the second ultrasound image.

[0011] In the ultrasound observation apparatus, the image composition unit is configured to: assign a pixel value obtained by a weighted sum of pixel values of pixels of the ultrasound image and the second ultrasound image at corresponding positions, to a pixel in the composite image at a corresponding position to the first pixel; and assign a pixel value of the second pixel to a pixel in the composite image at a corresponding position to the second pixel.

[0012] In the ultrasound observation apparatus, the predetermined display mode is grayscale display, and for a pixel in the ultrasound image corresponding to the pixel position where the value based on at least one of the first and second features is not less than or not more than the predetermined threshold, the image composition unit is configured to change a percentage of each color component in the grayscale display according to the value to generate the image data of the composite image.

[0013] The ultrasound observation apparatus further includes a display unit configured to display the composite image.

[0014] In the ultrasound observation apparatus, the display unit is configured to display the composite image and the ultrasound image side by side.

[0015] In the ultrasound observation apparatus, the second ultrasound signal has a same phase as that of the first ultrasound signal and has an amplitude that is $1/n$ of an amplitude of the first ultrasound signal ($n > 0$ and $n \neq 1$), and the image data generation unit is configured to compute the third received signal by multiplying the second received signal by a factor of n and subtracting the resulting second received signal from the first received signal.

[0016] In the ultrasound observation apparatus, the second ultrasound signal has a same amplitude as that of the first ultrasound signal and has a phase inverted relative to a phase of the first ultrasound signal, and the image data generation unit is configured to generate the third received signal by adding up the first and second received signals.

[0017] A method for operating an ultrasound observation apparatus according to the invention includes: an ultrasound transmitting and receiving step of, by an ultrasound probe, transmitting a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and receiving first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal; an image data generation step of, by an image data generation unit, performing a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute a third received signal, and generating image data of an ultrasound image having a predetermined display mode, using the third received signal; a feature extraction step of extracting, by a feature extraction unit, at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and an image composition step of generating, by an image composition unit, image data of a composite image using the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

[0018] An operating program for an ultrasound observation apparatus according to the invention instructs the ultrasound observation apparatus to execute: an ultrasound transmitting and receiving step of, by an ultrasound probe, transmitting a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and receiving first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal; an image data generation step of, by an image data generation unit, performing a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute

a third received signal, and generating image data of an ultrasound image having a predetermined display mode, using the third received signal; a feature extraction step of extracting, by a feature extraction unit, at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and an image composition step of generating, by an image composition unit, image data of a composite image using the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

Advantageous Effects of Invention

[0019] According to the present invention, an ultrasound image having excellent resolution is generated using a third received signal, and image data of a composite image is generated using the ultrasound image. The composite image has a display mode different from that of the ultrasound image at a pixel position where a feature extracted from frequency spectra of ultrasound echo signals satisfies a predetermined condition, and has the same display mode as that of the ultrasound image at other pixel positions. Thus, it is possible to generate the ultrasound image that allows identification of tissue characteristics and that has excellent resolution.

Brief Description of Drawings

[0020]

FIG. 1 is a block diagram illustrating an exemplary configuration of an ultrasound observation apparatus according to a first embodiment of the present invention.

FIG. 2 is a diagram illustrating a relationship between amplification factor and reception depth for an amplification process (STC correction) performed by a received signal processing unit illustrated in FIG. 1.

FIG. 3 is a diagram illustrating a relationship between reception depth and amplification factor for an amplification correction process performed by an amplification correction unit illustrated in FIG. 1.

FIG. 4 is a schematic diagram illustrating examples of waveforms of a first ultrasound signal and an ultrasound echo signal thereof.

FIG. 5 is a schematic diagram illustrating examples of waveforms of a second ultrasound signal and an ultrasound echo signal thereof.

FIG. 6 is a flowchart illustrating the operation of the ultrasound observation apparatus illustrated in FIG. 1.

FIG. 7 is a diagram illustrating exemplary display of a THI image.

FIG. 8 is a flowchart illustrating a frequency analysis process performed by a frequency analysis unit illustrated in FIG. 1.

FIG. 9 is a diagram schematically illustrating a data array of one sound ray.

FIG. 10 is a diagram illustrating an example of a frequency spectrum computed by the frequency analysis unit illustrated in FIG. 1.

FIG. 11 is a flowchart illustrating a tissue characteristic image data generating process performed by a tissue characteristic image data generation unit illustrated in FIG. 1.

FIG. 12 is a diagram illustrating an example of a tissue characteristic image.

FIG. 13 is a diagram schematically illustrating the image illustrated in FIG. 12 in black and white.

FIG. 14 is a diagram illustrating exemplary display of a composite image.

FIG. 15 is a diagram schematically illustrating the image illustrated in FIG. 14 in black and white.

FIG. 16 is a diagram illustrating another exemplary display of the composite image.

FIG. 17 is a schematic diagram for describing an attenuation correction method of a modification of the first embodiment of the present invention.

FIG. 18 is a schematic diagram illustrating examples of waveforms of a second ultrasound signal and an ultrasound echo signal thereof which are transmitted in a second embodiment of the present invention.

FIG. 19 is a flowchart illustrating the operation of an ultrasound observation apparatus according to a third embodiment of the present invention.

Description of Embodiments

[0021] Embodiments of an ultrasound observation apparatus, a method of operating an ultrasound observation apparatus, and an operating program for an ultrasound observation apparatus according to the present invention will be

described in detail below with reference to the drawings. Note that the present invention is not limited to these embodiments. The same reference signs are used to designate the same elements throughout the drawings.

(First Embodiment)

[0022] FIG. 1 is a block diagram illustrating an exemplary configuration of an ultrasound observation apparatus according to a first embodiment of the present invention. An ultrasound observation apparatus 1 illustrated in the drawing is an apparatus that observes a specimen using ultrasound.

[0023] The ultrasound observation apparatus 1 includes an ultrasound probe 2 that outputs ultrasound pulses externally and receives ultrasound echoes reflected externally; a transmitting and receiving unit 3 that performs transmission and reception of electrical signals with the ultrasound probe 2; a computation unit 4 that performs a predetermined calculation process on electrical echo signals into which ultrasound echoes are converted; an image processing unit 5 that performs generation of image data corresponding to the electrical echo signals into which the ultrasound echoes are converted; an input unit 6 that accepts input of various types of information to the ultrasound observation apparatus 1; a display unit 7 that is implemented using a display panel including liquid crystal, organic EL, or the like, and that displays various types of information including an image generated by the image processing unit 5; a storage unit 8 that stores parameters used in the calculation process and image processing for the echo signals and various types of information such as results of these processes; and a control unit 9 that performs operational control of the ultrasound observation apparatus 1.

[0024] The ultrasound probe 2 includes a signal converter 21 including a plurality of ultrasound transducers that convert an electrical pulse signal received from the transmitting and receiving unit 3 into ultrasound pulses (ultrasound signal) and transmit the ultrasound pulses, and receive an ultrasound echo reflected from a specimen and convert the ultrasound echo into an electrical echo signal. The type of the ultrasound probe 2 is not particularly limited and may be of an external type (a linear type, a sector type, a convex type, etc.) which is used outside the body such as the abdomen or the chest, or may be an ultrasound endoscope which is inserted into the body and used. In addition, a scanning method for the ultrasound probe 2 may be a mechanical scanning method that controls the direction of ultrasound transmission by mechanical control performed on the ultrasound transducers, or may be an electronic scanning method that controls the direction of ultrasound transmission by electronic control performed on the plurality of ultrasound transducers. Furthermore, a 2D array that can acquire three-dimensional image information may be used.

[0025] The transmitting and receiving unit 3 is electrically connected to the ultrasound probe 2, and transmits an electrical pulse signal to the ultrasound probe 2 and receives an electrical echo signal from the ultrasound probe 2. More specifically, the transmitting and receiving unit 3 includes a transmission signal generation unit 31 and a received signal processing unit 32. Note that when the ultrasound probe 2 employs the electronic scanning method, the transmitting and receiving unit 3 includes a multichannel circuit for combining beams which is provided for the plurality of ultrasound transducers.

[0026] The transmission signal generation unit 31 generates a pulse signal having a predetermined waveform under control of the control unit 9, and performs processes, such as amplification and transmission delay for transmitting ultrasound in a desired direction, on the pulse signal and transmits the processed pulse signal to the ultrasound probe 2.

[0027] The received signal processing unit 32 performs predetermined signal processing, such as amplification, filtering, A/D conversion, and delay addition, on an electrical echo signal received from the ultrasound probe 2, and thereby generates and outputs a digital received signal (digital RF signal) for each line corresponding to a direction of ultrasound transmission.

[0028] More specifically, the received signal processing unit 32 performs, as an amplification process, STC (Sensitivity Time Control) correction where an echo signal with a larger reception depth is amplified with a higher amplification factor. FIG. 2 is a diagram illustrating a relationship between reception depth and amplification factor for an amplification process performed by the received signal processing unit 32. A reception depth z illustrated in FIG. 2 is an amount computed based on the elapsed time from a time point when reception of ultrasound is started. As illustrated in FIG. 2, when the reception depth z is smaller than a threshold z_{th} , an amplification factor β (dB) increases linearly from β_0 to β_{th} ($> \beta_0$) along with an increase in the reception depth z . In addition, when the reception depth z is greater than or equal to the threshold z_{th} , the amplification factor β takes a constant value β_{th} . The value of the threshold z_{th} is a value at which almost all of an ultrasound signal received from a specimen is attenuated and thus noise is dominant. More generally, when the reception depth z is smaller than the threshold z_{th} , the amplification factor β monotonously increases along with an increase in the reception depth z .

[0029] In addition, a received digital signal for each line (sound ray) corresponding to a direction of ultrasound transmission, which is output from the received signal processing unit 32 is hereinafter also referred to as sound ray data.

[0030] The computation unit 4 includes an amplification correction unit 41 that performs amplification correction where the amplification factor is constant irrespective of the reception depth, on a received signal which is output from the transmitting and receiving unit 3; a frequency analysis unit 42 that performs frequency analysis of an echo signal by performing a fast Fourier transform (FFT) on the received signal having been subjected to the amplification correction;

and a feature extraction unit 43 that extracts feature from frequency spectra (power spectra) which are computed by the frequency analysis unit 42, so as to be associated with pixel positions in an image.

[0031] FIG. 3 is a diagram illustrating a relationship between reception depth and amplification factor for an amplification correction process performed by the amplification correction unit 41. As illustrated in FIG. 3, an amplification factor β (dB) for the amplification process performed by the amplification correction unit 41 takes a maximum value $\beta_{th} - \beta_0$ when a reception depth z is zero, and decreases linearly until the reception depth z reaches a threshold z_{th} from zero, and is zero when the reception depth z is greater than or equal to the threshold z_{th} . By the amplification correction unit 41 performing amplification correction on a received signal using the amplification factor as defined above, the influence of STC correction performed by the received signal processing unit 32 is cancelled out, thereby enabling to output a signal with a constant amplification factor β_{th} . Note that the relationship between reception depth z and the amplification factor P performed by the amplification correction unit 41, of course, varies depending on the relationship between reception depth and amplification factor for the received signal processing unit 32.

[0032] Here, the STC correction performed by the received signal processing unit 32 is correction to uniformly amplify the amplitude of an analog signal waveform over the entire frequency band. Hence, as will be described later, when a THI (tissue harmonic imaging) image that uses the amplitude of ultrasound is generated, sufficient effect can be obtained by performing STC correction, whereas when an ultrasound frequency spectrum is computed, the influence of attenuation associated with propagation of ultrasound cannot be accurately eliminated. To solve this problem, it is considered that when a THI image is generated, received signals having been subjected to STC correction are output; on the other hand, when an image based on frequency spectra is generated, another transmission different from transmission for creating a THI image is performed and received signals having not been subjected to STC correction are output. However, in this case, there is a problem that the frame rate of image data generated based on received signals decreases. Hence, by providing the amplification correction unit 41 at the previous stage to the frequency analysis unit 42, while the frame rate of image data to be generated is maintained, the influence of STC correction can be eliminated once for signals having been subjected to STC correction for a THI image.

[0033] The frequency analysis unit 42 performs, for received signals output from the transmitting and receiving unit 3, a fast Fourier transform on FFT data groups each having a predetermined amount of data, and thereby computes frequency spectra for a plurality of locations (data positions) on sound rays. In general, a frequency spectrum shows different tendencies depending on the tissue characteristic of a specimen. This is because the frequency spectrum has a correlation with the size, density, acoustic impedance, etc., of a specimen serving as a scatterer that scatters ultrasound. Note that in the first embodiment the "tissue characteristic" refers to a condition of tissue, e.g., cancer, an endocrine tumor, a mucinous tumor, normal tissue, or a vessel.

[0034] The feature extraction unit 43 includes an approximation unit 431 that performs an approximation process on the frequency spectra computed by the frequency analysis unit 42, and thereby computes pre-correction feature; an attenuation correction unit 432 that performs an attenuation correction process, which reduces contribution of attenuation occurring according to the reception depth and frequency of ultrasound when the ultrasound is propagated, on the pre-correction feature computed by the approximation unit 431; and a weighting and adding unit 433 that calculates a weighted sum of the features having been subjected to the attenuation correction process.

[0035] The approximation unit 431 approximates a frequency spectrum by regression analysis using a linear expression (regression line), and thereby extracts pre-correction feature that characterizes the approximated linear expression. Specifically, the approximation unit 431 extracts, as pre-correction feature, the slope a_0 and intercept b_0 of a linear expression by regression analysis. Note that the approximation unit 431 may compute the intensity (also referred to as Mid-band fit) $c_0 = a_0 f_M + b_0$ of a center frequency $f_M = (f_L + f_H)/2$ of a frequency band ($f_L < f < f_H$), as pre-correction feature other than the slope a_0 and the intercept b_0 .

[0036] Of the three pre-correction feature, the slope a_0 has a correlation with the size of an ultrasound scatterer, and in general, it is considered that the larger the scatterer the smaller the value of the slope. In addition, the intercept b_0 has a correlation with the size of the scatterer, the difference in acoustic impedance, the density (concentration) of the scatterer, etc. Specifically, it is considered that the larger the scatterer the larger the value of the intercept b_0 , the larger the difference in acoustic impedance the larger the value of the intercept b_0 , and the higher the density (concentration) of the scatterer the larger the value of the intercept b_0 . The intensity of the center frequency f_M (hereinafter, simply referred to as "intensity") c_0 is an indirect parameter derived from the slope a_0 and the intercept b_0 , and provides spectral intensity at the center of an effective frequency band. Hence, it is considered that the intensity c_0 has a certain degree of correlation with the luminance of a THI image or the like which is generated based on the amplitudes of received signals, in addition to the size of the scatterer, the difference in acoustic impedance, and the density of the scatterer. Note that an approximate polynomial computed by the approximation unit 431 is not limited to a linear expression, and it is also possible to use quadratic or higher-order approximate polynomials.

[0037] The attenuation correction unit 432 performs an attenuation correction process, which reduces contribution of attenuation occurring according to the reception depth and frequency of ultrasound when the ultrasound is propagated, on the pre-correction feature extracted by the approximation unit 431.

[0038] In general, the amount of attenuation of ultrasound $A(f, z)$ is represented by:

$$A(f, z) = 2\alpha z f \dots (1)$$

where α is the attenuation factor, z is the reception depth of ultrasound, and f is the frequency. As is also clear from equation (1), the amount of attenuation $A(f, z)$ is proportional to the frequency f . When an observation target is a living body, a specific value of the attenuation factor α is 0.0 to 1.0 (dB/cm/MHz), more preferably, 0.3 to 0.7 (dB/cm/MHz), and is determined according to the region of the living body. For example, when the observation target is the pancreas, the attenuation factor α may be determined such that $\alpha = 0.6$ (dB/cm/MHz). Note that in the first embodiment it may be configured such that the value of the attenuation factor α is settable or changeable by input from the input unit 6.

[0039] The attenuation correction unit 432 performs attenuation correction on pre-correction feature (slope a_0 , intercept b_0 , and intensity c_0) which is acquired by the approximation process, in the following manner and thereby extracts feature:

$$a = a_0 + 2\alpha z \dots (2)$$

$$b = b_0 \dots (3)$$

$$c = c_0 + 2\alpha z f_M (= a f_M + b) \dots (4)$$

As is also clear from (2) and (4), the attenuation correction unit 432 performs correction where the larger the reception depth z of ultrasound the larger the amount of correction. In addition, according to equation (3), correction for the intercept is identify transformation. This is because the intercept is a frequency component for a frequency of 0 (Hz) and thus is not subjected to the influence of attenuation.

[0040] A straight line corresponding to corrected feature is represented by the following equation:

$$I = a f + b = (a_0 + 2\alpha z) f + b_0 \dots (5)$$

As is also clear from this equation (5), the straight line corresponding to the corrected feature has a large slope and an identical intercept, compared to a straight line corresponding to pre-correction feature.

[0041] The weighting and adding unit 433 calculates the weighted sum of the features that are respectively extracted from sound ray data corresponding to two ultrasound echo signals which are generated by transmitting two ultrasound signals toward a specimen at different points in time along the same line and by reflection of the two ultrasound signals from the specimen.

[0042] The image processing unit 5 includes a THI image data generation unit (image data generation unit) 51 that generates image data of a THI image (ultrasound image), based on harmonic components of sound ray data; a tissue characteristic image data generation unit (second image data generation unit) 52 that generates image data of a tissue characteristic image (second ultrasound image) representing the tissue characteristics of body tissue, based on feature extracted by the feature extraction unit 43; and an image composition unit 53 that generates image data of a composite image where the THI image and the tissue characteristic image are combined.

[0043] The THI image data generation unit 51 extracts harmonic components from received signals corresponding to echoes of ultrasound signals (ultrasound echoes), respectively, which are sequentially transmitted to a specimen along the same line and which have the same waveform and different amplitudes or phases; and generates THI image data based on the extracted harmonic components.

[0044] Now, the principle of THI will be described with reference to FIGS. 4 and 5. FIGS. 4 and 5 are schematic diagrams illustrating examples of waveforms of transmitted ultrasound signals and ultrasound echo signals thereof. Here, the case of transmitting two ultrasound signals having the same waveform and the same phase and different amplitudes will be described.

[0045] It is known that, when ultrasound is propagated through body tissue, a nonlinear effect occurs more strongly for larger amplitude of ultrasound. For example, when a first ultrasound signal having a waveform illustrated in (a) of FIG. 4 is transmitted, due to the nonlinear effect of body tissue, as illustrated in (b) of FIG. 4, distortion occurs in the waveform of a first ultrasound echo signal generated by reflection of the first ultrasound signal by a specimen. The first ultrasound echo signal can be represented as a combined wave of a fundamental wave component illustrated in (c) of

FIG. 4, a second harmonic component illustrated in (d) of FIG. 4, and furthermore, a third or higher harmonic component.

[0046] On the other hand, when, as illustrated in (a) of FIG. 5, a second ultrasound signal having the same phase as the first ultrasound signal and a smaller amplitude than the first ultrasound signal is transmitted, the nonlinear effect of body tissue is smaller than that for the case of the first ultrasound signal. Hence, in a second ultrasound echo signal generated by reflection of the second ultrasound signal by the specimen, as illustrated in (b) of FIG. 5, waveform distortion is smaller compared to that in (b) of FIG. 4. The second ultrasound echo signal can be represented as a combined wave of a fundamental wave component illustrated in (c) of FIG. 5, a second harmonic component illustrated in (d) of FIG. 5, and furthermore, a third or higher harmonic component.

[0047] Here, the amplitude of the fundamental wave component has a value nearly proportional to the transmitted ultrasound signal. On the other hand, the amplitude of the harmonic component has a value nearly proportional to the square of the amplitude of the fundamental wave component. Hence, when the amplitude of the second ultrasound signal is a factor of $1/n$ smaller than that of the first ultrasound signal ($n > 0$ and $n \neq 1$, $n = 2$ in FIGS. 4 and 5), the amplitude of the second harmonic component of the second ultrasound echo signal (see (d) of FIG. 5) becomes $1/n^2$ ($1/4$ in FIGS. 4 and 5) of the amplitude of the second harmonic component of the first ultrasound echo signal (see (d) of FIG. 4). Hence, by multiplying the second ultrasound echo signal by a factor of n and subtracting the resulting second ultrasound echo signal from the first ultrasound echo signal, the fundamental wave component is cancelled out, enabling to extract the second harmonic component.

[0048] The THI image data generation unit 51 extracts second harmonic components from sound ray data output from the transmitting and receiving unit 3, according to the above-described principle of THI, and further performs predetermined processes, such as a bandpass filtering process, a detection process, and a logarithmic compression process, on the second harmonic components, and thereby generates THI image data. A display mode of a THI image represented by the THI image data is grayscale display including two- or three-dimensional image space where the values of R (red), G (green), and B (blue), which are variables when an RGB color system is adopted as a color space, match each other.

[0049] The tissue characteristic image data generation unit 52 generates tissue characteristic image data representing the tissue characteristics of a specimen, based on the feature extracted by the feature extraction unit 43. More specifically, the tissue characteristic image data generation unit 52 assigns a pixel whose feature is smaller than a predetermined threshold, with a pixel value of a color according to the feature; on the other hand, the tissue characteristic image data generation unit 52 assigns a pixel whose feature is greater than or equal to the predetermined threshold, with zero (i.e., black) as a pixel value. Namely, a display mode (second display mode) of the tissue characteristic image is color display. Alternatively, reversely, a pixel whose feature is larger than the predetermined threshold may be assigned a pixel value of a color according to the feature, and on the other hand, a pixel whose feature is less than or equal to the predetermined threshold may be assigned zero as a pixel value.

[0050] The image composition unit 53 generates, using the THI image data generated by the THI image data generation unit 51, image data of a composite image where areas that are specified based on the feature extracted by the feature extraction unit 43 are displayed in a different mode than the display mode of the THI image. In the first embodiment, the image composition unit 53 generates composite image data of a composite image where a THI image and a tissue characteristic image are combined, using the THI image data generated by the THI image data generation unit 51 and the tissue characteristic image data generated by the tissue characteristic image data generation unit 52. The composite image displayed using the composite image data is an image where the color tissue characteristic image is superimposed on the grayscale THI image.

[0051] The input unit 6 is implemented using interfaces such as a keyboard, a mouse, a touch panel, and a card reader, and inputs to the control unit 9 a signal generated according to an operation performed externally by an operator, etc. Specifically, the input unit 6 accepts patient identification information for identifying a patient to be examined, an instruction to set an area of interest, instructions to start various types of operation, etc., and inputs signals representing these information and instructions to the control unit 9. The area of interest as used herein refers to an area in the THI image displayed on the display unit 7 that is specified by the operator of the ultrasound observation apparatus 1 using the input unit 6.

[0052] The storage unit 8 is implemented using, for example, a ROM that pre-stores an operating program for the ultrasound observation apparatus 1 according to the first embodiment, a program for starting up a predetermined OS, etc., and a RAM that stores parameters and data used in processes, etc. More specifically, the storage unit 8 stores sound ray data output from the transmitting and receiving unit 3, window functions (Hamming, Hanning, Blackman, etc.) used by the frequency analysis unit 42, feature extracted by the feature extraction unit 43, THI image data, tissue characteristic image data, and composite image data which are generated by the image processing unit 5, etc.

[0053] The control unit 9 includes a transmission signal controller 91 that controls pulse signal generation operation of the transmission signal generation unit 31; and a display controller 92 that controls display operation of the display unit 7. The transmission signal controller 91 controls the transmission signal generation unit 31 such that transmission operation where a plurality of ultrasound signals having the same waveform and different phases or amplitudes are sequentially transmitted along the same line traveling toward a specimen is considered one set, and a specimen is

scanned by ultrasound which is transmitted and received by this one set of transmission operation.

[0054] Elements of the ultrasound observation apparatus 1 having the above-described functional configuration other than the ultrasound probe 2 are implemented using a computer that includes a CPU having calculation and control functions. The CPU included in the ultrasound observation apparatus 1 performs a calculation process related to a method of operating the ultrasound observation apparatus 1 according to the first embodiment, by reading from the storage unit 8 information and various types of programs including the above-described operating program for the ultrasound observation apparatus 1, which are stored and retained in the storage unit 8.

[0055] Note that the operating program for the ultrasound observation apparatus 1 according to the first embodiment can also be widely distributed by recording the operating program in a computer-readable recording medium such as a hard disk, a flash memory, a CD-ROM, a DVD-ROM, or a flexible disk.

[0056] Next, the operation of the ultrasound observation apparatus 1 will be described. FIG. 6 is a flowchart illustrating the operation of the ultrasound observation apparatus 1.

[0057] First, at step S10, the ultrasound observation apparatus 1 measures a new specimen by the ultrasound probe 2. Specifically, in the first embodiment, the ultrasound observation apparatus 1 sequentially transmits first and second ultrasound signals having the same waveform and the same phase and different amplitudes along the same line traveling toward a specimen, and receives first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals by the specimen. At this time, a frequency band of the first and second ultrasound signals may be a wide band that covers substantially the entire frequency band of the ultrasound transducers. By that, an accurate linear approximation is possible in a process of extracting feature from frequency spectra which will be described later.

[0058] The ultrasound echo signals received by the ultrasound probe 2 are converted into electrical echo signals, and furthermore, the electrical echo signals are subjected to predetermined signal processing such as amplification (STC correction), filtering, A/D conversion, and delay addition, in the received signal processing unit 32. By that, first and second sound ray data (received signals) are generated for each line along which the first and second ultrasound signals are transmitted. Namely, two types of sound ray data are acquired for one line. The ultrasound observation apparatus 1 repeatedly performs such ultrasound transmission/reception and signal processing while moving the line along which ultrasound signals are transmitted.

[0059] At subsequent step S11, the THI image data generation unit 51 generates THI image data, based on the first and second sound ray data which are acquired at step S10. For example, as illustrated in (a) of FIG. 4 and (a) of FIG. 5, when the amplitude of the second ultrasound signal is $1/n$ of the amplitude of the first ultrasound signal (e.g., $n = 2$), the THI image data generation unit 51 multiplies a data value (corresponding to amplitude) of the second sound ray data (second received signal) by a factor of n , and then subtracts the resulting second sound ray data from the first sound ray data (first received signal), and thereby generates third sound ray data (third received signal) corresponding to a second harmonic component. Then, predetermined processes such as a bandpass filtering process, a detection process, and a logarithmic compression process are performed on the third sound ray data, by which THI image data is generated.

[0060] At step S12, the display controller 92 displays a THI image based on the THI image data generated by the THI image data generation unit 51. FIG. 7 is a diagram illustrating exemplary display of the THI image. As described above, the THI image is a grayscale image where the values of R (red), G (green), and B (blue) of each pixel match each other.

[0061] Thereafter, if an area of interest is set through the input unit 6 (step S13: Yes), the computation unit 4 captures the first and second sound ray data acquired at step S10, performs amplification correction on these sound ray data, and then performs frequency analysis by FFT computation, and thereby computes frequency spectra (step S14). At this step S13, it is also possible to set the entire area of the image as an area of interest. On the other hand, if an area of interest is not set (step S13: No) and if an instruction to terminate the operation is input through the input unit 6 (step S15: Yes), the ultrasound observation apparatus 1 terminates its operation. In contrast to this, if an area of interest is not set (step S13: No) and if an instruction to terminate the operation is not input through the input unit 6 (step S15: No), the operation of the ultrasound observation apparatus 1 returns to step S13.

[0062] FIG. 8 is a flowchart illustrating the frequency analysis process performed by the frequency analysis unit 42 at step S14.

[0063] First, at step S140, the frequency analysis unit 42 sets a counter k for identifying a line for sound ray data to be analyzed to k_0 . At subsequent step S141, the frequency analysis unit 42 sets a counter j for identifying sound ray data to be analyzed to 1.

[0064] At subsequent step S142, the frequency analysis unit 42 sets an initial value $Z^{(k(j))}_0$ of a data position (corresponding to reception depth) $Z^{(k(j))}$ that represents a data group (FFT data group) acquired for FFT computation. FIG. 9 is a diagram schematically illustrating a data array of one sound ray. In sound ray data $SR_{k(j)}$ illustrated in the drawing, a white or black rectangle indicates one piece of data. The sound ray data $SR_{k(j)}$ is discretized in time intervals corresponding to the sampling frequency (e.g., 50 MHz) of A/D conversion performed by the received signal processing unit 32. Although FIG. 9 illustrates the case in which the first data position of the sound ray data $SR_{k(j)}$ is set as the initial value $Z^{(k(j))}_0$, the position of the initial value can be set arbitrarily.

[0065] At subsequent step S143, the frequency analysis unit 42 acquires an FFT data group for the data position $Z^{(k(j))}$.

Then, at step S144, a window function stored in the storage unit 8 is applied to the acquired FFT data group. By thus allowing the window function to act on the FFT data group, the FFT data group is avoided from becoming discontinuous at boundaries, enabling to prevent the occurrence of artifacts.

[0066] At subsequent step S145, the frequency analysis unit 42 determines whether the FFT data group for the data position $Z^{(k(j))}$ is a normal data group. Here, the FFT data group needs to have a power-of-two number of data. In the following, the number of data included in an FFT data group is 2^n (n is a positive integer). The FFT data group being normal refers to that the data position $Z^{(k(j))}$ is a 2^{n-1} th position from the top of the FFT data group. In other words, the FFT data group being normal refers to that there are $2^{n-1} - 1$ ($= N$) pieces of data before the data position $Z^{(k(j))}$ and there are 2^{n-1} ($= M$) pieces of data after the data position $Z^{(k(j))}$. In the case illustrated in FIG. 9, FFT data groups F_2 and F_3 are both normal. Note that FIG. 9 exemplifies the case of $n = 4$ ($N = 7$ and $M = 8$).

[0067] If, as a result of the determination at step S145, the FFT data group for the data position $Z^{(k(j))}$ is normal (step S145: Yes), processing of the frequency analysis unit 42 transitions to step S147 which will be described later.

[0068] On the other hand, if, as a result of the determination at step S145, the FFT data group for the data position $Z^{(k(j))}$ is not normal (step S145: No), the frequency analysis unit 42 inserts zero data corresponding to a shortfall and thereby generates a normal FFT data group (step S146). The FFT data group that is determined at step S145 to be not normal is acted on by the window function before adding zero data. Hence, even if zero data is inserted into the FFT data group, discontinuity of data does not occur. After step S146, processing of the frequency analysis unit 42 transitions to step S147 which will be described later.

[0069] At step S147, the frequency analysis unit 42 performs FFT computation using the FFT data group and thereby obtains a frequency spectrum including complex numbers. As a result, for example, a spectrum C_1 such as that illustrated in FIG. 10 is obtained.

[0070] At subsequent step S148, the frequency analysis unit 42 changes the data position $Z^{(k(j))}$ by a step width D . It is assumed that the step width D is pre-stored in the storage unit 8. FIG. 9 exemplifies the case of $D = 15$. It is desirable that the step width D match a data step width which is used when the THI image data generation unit 51 generates THI image data, but when the amount of calculation of the frequency analysis unit 42 needs to be reduced, a larger value than the data step width may be set.

[0071] At subsequent step S149, the frequency analysis unit 42 determines whether the data position $Z^{(k(j))}$ is greater than a maximum value $Z^{(k(j))}_{\max}$ of the sound ray data $SR_{k(j)}$. If the data position $Z^{(k(j))}$ is less than or equal to the maximum value $Z^{(k(j))}_{\max}$ (step S149: No), processing returns to step S143. In this manner, the frequency analysis unit 42 performs FFT computation on $\lceil (Z^{(k(j))}_{\max} - Z^{(k(j))}_0)/D \rceil + 1$ FFT data groups for the sound ray data $SR_{k(j)}$. Here, $[X]$ represents a maximum integer not exceeding X .

[0072] On the other hand, if the data position $Z^{(k(j))}$ is greater than the maximum value $Z^{(k(j))}_{\max}$ (step S149: Yes), then the frequency analysis unit 42 increments the counter j by 1 (step S150).

[0073] At subsequent step S151, the frequency analysis unit 42 determines whether the counter j is greater than a maximum value $j_{\max}$ (step S151). Here, for the maximum value $j_{\max}$, the number of sound ray data acquired for each line is set. For example, when first and second sound ray data are acquired for each line, $j_{\max} = 2$. If the counter j is less than or equal to the maximum value $j_{\max}$ (step S151: No), processing of the frequency analysis unit 42 returns to step S142. On the other hand, if the counter j is greater than the maximum value $j_{\max}$ (step S151: Yes), the frequency analysis unit 42 increments the counter k by 1 (step S152).

[0074] At subsequent step S153, the frequency analysis unit 42 determines whether the counter k is greater than a maximum value $k_{\max}$. If the counter k is greater than the maximum value $k_{\max}$ (step S153: Yes), the frequency analysis unit 42 ends a series of FFT processes and the operation of the ultrasound observation apparatus 1 returns to the main routine. On the other hand, if the counter k is less than or equal to the maximum value $k_{\max}$ (step S153: No), processing of the frequency analysis unit 42 returns to step S141.

[0075] In this manner, the frequency analysis unit 42 performs a plurality of FFT computations for each of the first and second sound ray data which are acquired for each of $(k_{\max} - k_0 + 1)$ lines. Thereafter, the operation of the ultrasound observation apparatus 1 returns to the main routine.

[0076] Note that here the input unit 6 accepts in advance input for setting a specific area of interest and a frequency analysis process is performed only on the area of interest; however, the frequency analysis unit 42 may perform a frequency analysis process on the whole area for which ultrasound echo signals are received.

[0077] At step S16 subsequent to step S14, the feature extraction unit 43 extracts a feature from each sound ray data so as to be associated with pixel positions in a two-or three-dimensional image, based on the results of the frequency analysis at step S14. Information to be associated with each pixel position in image space is determined according to the amount of data of an FFT data group which is used when the frequency analysis unit 42 computes a frequency spectrum. Specifically, for example, a pixel position corresponding to the amount of data of one FFT data group is assigned information (feature) corresponding to feature of a frequency spectrum that is computed for the FFT data group.

[0078] In addition, at this time, when the process is performed for the second sound ray data whose amplitude is $1/n$ of the amplitude of the first sound ray data, the intensities of frequency spectra acquired at step S14 are multiplied by

a factor of n in advance. Note that timing at which the intensities are multiplied by a factor of n may be after an attenuation correction process which will be described later.

[0079] First, the approximation unit 431 performs regression analysis on P frequency spectra which are computed by the frequency analysis unit 42, and thereby extracts pre-correction feature. Specifically, the approximation unit 431 computes, by regression analysis, a linear expression that approximates a frequency spectrum with a frequency band $f_{\text{LOW}} < f < f_{\text{HIGH}}$, and thereby extracts a slope a_0 , an intercept b_0 , and an intensity c_0 that characterize the linear expression, as pre-correction feature. A straight line L_1 illustrated in FIG. 10 is a pre-correction regression line obtained by this process.

[0080] Subsequently, the attenuation correction unit 432 substitutes the value of the data position $Z^{(k(j))}$ into the reception depth z in the above-described equations (2) to (4), and thereby computes a slope a , an intercept b , and an intensity c which are corrected feature. A straight line L_1' illustrated in FIG. 10 is a regression line obtained at step S16.

[0081] In this manner, the approximation unit 431 and the attenuation correction unit 432 extract first and second features from the first and second sound ray data for each of pixel positions corresponding to a plurality of points on each line along which the first and second ultrasound signals are transmitted.

[0082] At subsequent step S17, the weighting and adding unit 433 calculates the weighted sum of the first and second features which are extracted for the same pixel position at step S16 to compute a third feature. Weights assigned to the first and second features both are not particularly limited as long as the weights are greater than zero. For example, when the weights assigned to the first and second features are both set to 1, the third feature which is obtained by addition of the first and second features is obtained. In addition, when the weights assigned to the first and second features are both set to $1/2$, the third feature which is the average of the first and second features is obtained. Of course, the weights assigned to the first and second features may differ from each other. In addition, at this time, when a plurality of types of feature (e.g., features a , b , and c) are extracted by the approximation unit 431 and the attenuation correction unit 432, the weighting and adding unit 433 calculates the weighted sum of the features of the same type (e.g., weighted sum of features a , weighted sum of features b , and weighted sum of features c).

[0083] At subsequent step S18, the tissue characteristic image data generation unit 52 generates tissue characteristic image data based on the third feature computed at step S17. FIG. 11 is a flowchart illustrating a tissue characteristic image data generating process.

[0084] The tissue characteristic image data generation unit 52 performs a process of a loop A on each pixel in the area of interest. First, at step S181, the tissue characteristic image data generation unit 52 determines whether the third feature of a pixel to be processed is smaller than a predetermined threshold.

[0085] If the third feature is smaller than the predetermined threshold (step S181: Yes), the tissue characteristic image data generation unit 52 assigns the pixel with a pixel value of a color according to the value of the third feature (step S182). Specifically, the tissue characteristic image data generation unit 52 assigns each of color components R (red), G (green), and B (blue) of the pixel with a slope a , an intercept b , and an intensity c which are the third feature.

[0086] On the other hand, if the third feature assigned to the pixel to be processed is greater than or equal to the predetermined threshold (step S181: No), the tissue characteristic image data generation unit 52 assigns the pixel with zero as a pixel value (step S183). Namely, the pixel is represented in black. Thereafter, processing of the tissue characteristic image data generation unit 52 transitions to the next pixel to be processed.

[0087] Such processes are performed on all pixels in the area of interest, and then the operation of the ultrasound observation apparatus 1 returns to the main routine.

[0088] FIG. 12 is a diagram illustrating an example of a tissue characteristic image generated based on the tissue characteristic image data generated at step S18. FIG. 13 is a diagram schematically illustrating the image illustrated in FIG. 12 in black and white. An area 100 illustrated in the drawings is an area whose third feature is determined at step S181 to be smaller than the predetermined threshold. The area 100 roughly includes a green-based area 101 and a red-based area 102, and a boundary portion between two areas is displayed in yellow-based color (not displayed in FIG. 13). As illustrated in FIGS. 12 and 13, each area does not include a single color. For example, the green-based area 101 is an area where pixels whose colors are close to green are gathered together. Likewise, the red-based area 102 is an area where pixels whose colors are close to red are gathered together.

[0089] At step S19 subsequent to step S18, the image composition unit 53 generates composite image data using the THI image data generated at step S11 and the tissue characteristic image data generated at step S18. More specifically, for an area in the tissue characteristic image that includes pixels with pixel values other than zero, a pixel value in the composite image is obtained by calculating the weighted sum of pixel values of pixels of the THI image and the tissue characteristic image at corresponding positions. Specifically, a pixel value (values of color components R (red), G (green), and B (blue)) of each pixel in a composite image is computed by the following equation (6):

$$g_{\text{syn}}(x, y) = w \cdot g_{\text{th}}(x, y) + (1 - w) \cdot g_{\text{ch}}(x, y) \quad \dots \quad (6)$$

In equation (6), the symbol $g_{syn}(x, y)$ represents the pixel value of a pixel at coordinates (x, y) of the composite image, the symbol $g_{th}(x, y)$ represents the pixel value of a pixel at coordinates (x, y) of the THI image, and the symbol $g_{ch}(x, y)$ represents the pixel value of a pixel at coordinates (x, y) of the tissue characteristic image. In addition, the symbols w and $(1 - w)$ are weights assigned to the pixel value of the THI image and the pixel value of the tissue characteristic image, respectively, and $0 < w < 1$.

[0090] On the other hand, for an area in the tissue characteristic image that includes pixels with pixel values of 0, pixel values in the THI image are assigned as they are as pixel values in the composite image.

[0091] At step S20, the display controller 92 displays a composite image generated based on the composite image data generated at step S19, on the display unit 7. FIG. 14 is a diagram illustrating an example of the composite image displayed at step S20. FIG. 15 is a diagram schematically illustrating the image illustrated in FIG. 14 in black and white. As illustrated in the drawings, the composite image is an image where the color tissue characteristic image (see FIG. 12) is superimposed on the grayscale THI image (see FIG. 7). An area 100 where the tissue characteristic image is superimposed is in a state in which the THI image and the tissue characteristic image are blended together by the weighted sum given by equation (6). Therefore, in the area 100, too, excellent resolution of the THI image is maintained and furthermore identification of tissue characteristics is also possible.

[0092] FIG. 16 is a diagram illustrating another exemplary display of the composite image. The display controller 92 may display a screen 200 such as that illustrated in FIG. 16, on the display unit 7 using the THI image data generated at step S11 and the composite image data generated at step S19. On the screen 200, the THI image illustrated in FIG. 7 is disposed in one image display area 201, and the composite image illustrated in FIG. 14 is disposed in another image display area 202. In addition to that, a display area 203 that displays information about a patient to be examined (ID, name, gender, etc.) may be provided on the screen 200. By thus displaying the THI image and the composite image side by side, an observation method is possible where microstructures in the body are observed using the THI image and a tissue characteristic for a suspicious area is checked using the composite image. Note that exemplary display of the composite image is not limited to the above-described examples, and for example, three images including the THI image, the tissue characteristic image, and the composite image may be displayed side by side.

[0093] As described above, according to the first embodiment of the present invention, a composite image can be obtained that allows identification of tissue characteristics while using advantages of a THI image having excellent resolution and reduced artifacts. Accordingly, by referring to the composite image, an operator can grasp microstructures of body tissue and grasp differences in tissue characteristics.

[0094] In addition, according to the first embodiment, there is also an advantage in that the image quality of a tissue characteristic image itself can be improved. Here, in a conventional tissue characteristic image, noise may become large in a specific area (e.g., a blood vessel area), depending on the type of frequency feature to be used for image generation. In the first embodiment, on the other hand, first and second features are extracted from first and second sound ray data corresponding to ultrasound echoes of first and second ultrasound signals which are transmitted at different points in time, and the third feature is computed by the weighted sum of the first and second features. Thus, noise components in the feature can be reduced. Therefore, by using such third feature, a tissue characteristic image with suppressed noise can be generated.

[0095] Note that when a tissue characteristic image is generated, frequency analysis may be performed on either one of the first and second sound ray data, feature may be acquired from results of the frequency analysis, and tissue characteristic image data may be generated using the feature. In this case, calculation load of the frequency analysis unit 42 and the feature extraction unit 43 can be reduced.

[0096] In addition, the first embodiment describes an example in which two-dimensional ultrasound images (a THI image, a tissue characteristic image, and a composite image) are generated by scanning a specimen by one-dimensionally moving a line along which ultrasound signals are transmitted and received; however, three-dimensional ultrasound images may be generated by two-dimensionally scanning a specimen using, for example, a 2D array where a plurality of ultrasound transducers are arranged two-dimensionally.

(First Modification)

[0097] Next, a first modification of the first embodiment of the present invention will be described.

[0098] In the first embodiment, at step S18, a pixel whose third feature is smaller than the threshold is assigned a pixel of a color according to the value of the third feature; on the other hand, depending on types of features, types of tissue to be observed, or the like, a pixel whose third feature is larger than the threshold may be assigned a pixel of a color according to the value of the third feature. In this case, a pixel whose third feature is less than or equal to the threshold is assigned zero as a pixel value. Even when tissue characteristic image data is generated in this way, it is possible to generate and display a composite image that allows identification of tissue characteristics.

(Second Modification)

[0099] Next, a second modification of the first embodiment of the present invention will be described.

[0100] When a feature is extracted at step S16 illustrated in FIG. 6, attenuation correction may be performed before performing regression analysis of a frequency spectrum.

[0101] FIG. 17 is a schematic diagram for describing an attenuation correction method of the second modification. When, for example, a frequency spectrum curve C_2 illustrated in FIG. 17 is acquired as a result of frequency analysis at step S14, the feature extraction unit 43 performs correction where the amount of attenuation A in the above-described equation (1) is added to intensity I , on all frequencies f , and thereby obtains a new frequency spectrum curve C_2' . By this, a frequency spectrum where contribution of attenuation association with propagation of ultrasound is reduced can be obtained.

[0102] Thereafter, the feature extraction unit 43 performs regression analysis on the entire frequency spectrum having been subjected to the attenuation correction, and thereby extracts feature of the frequency spectrum. Specifically, the feature extraction unit 43 computes a slope a and an intercept b of a linear expression and an intensity c of a center frequency f_{MID} by regression analysis on the frequency spectrum curve C_2' . A straight line L_2 illustrated in FIG. 17 is a regression line (intercept b_2) which is obtained by performing a feature extraction process on the frequency spectrum curve C_2' . By such a correction method, too, first and second features can be extracted with the same accuracy as in the first embodiment.

(Second Embodiment)

[0103] Next, a second embodiment of the present invention will be described. Note that a configuration of an ultrasound observation apparatus according to the second embodiment is the same as that of the first embodiment (see FIG. 1).

[0104] Although the first embodiment describes the case in which a second ultrasound signal has the same phase as a first ultrasound signal and has an amplitude that is $1/n$ of that of the first ultrasound signal, a combination of the first and second ultrasound signals is not limited thereto. For example, as illustrated in (a) of FIG. 18, a second ultrasound signal having the same amplitude as a first ultrasound signal (see (a) of FIG. 4) and having a phase inverted relative to that of the first ultrasound signal may be transmitted.

[0105] In this case, due to a nonlinear effect of a body signal, as illustrated in (b) of FIG. 18, distortion occurs in the waveform of a second ultrasound echo signal generated by reflection of the second ultrasound signal by a specimen. The second ultrasound echo signal can be represented as a combined wave of a fundamental wave component illustrated in (c) of FIG. 18, a second harmonic component illustrated in (d) of FIG. 18, and furthermore, a third or higher harmonic component. Hence, by adding up a first ultrasound echo signal and the second ultrasound echo signal, the fundamental wave component is cancelled out, enabling to enhance the second harmonic component.

[0106] Next, the operation of the ultrasound observation apparatus according to the second embodiment will be described with reference to FIG. 6. First, at step S10, the ultrasound observation apparatus 1 illustrated in FIG. 1 measures a new specimen by an ultrasound probe 2. At this time, the ultrasound observation apparatus 1 sequentially transmits first and second ultrasound signals having the same waveform and the same amplitude and having phases inverted relative to each other, along the same line traveling toward the specimen, and receives first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals by the specimen. Operation after step S10 is the same as that of the first embodiment.

[0107] At subsequent step S11, a THI image data generation unit 51 generates THI image data based on the first and second sound ray data (received signals) which are acquired at step S10. At this time, the THI image data generation unit 51 adds up the first sound ray data (first received signal) and the second sound ray data (second received signal), and thereby generates third sound ray data (third received signal). Then, predetermined processes such as a bandpass filtering process, a detection process, and a logarithmic compression process are performed on the third sound ray data, by which THI image data is generated. Subsequent steps S12 to S14 are the same as those of the first embodiment.

[0108] At step S16 subsequent to step S14, a feature extraction unit 43 extracts a feature from each sound ray data, based on results of frequency analysis at step S14. A feature extraction process as a whole is the same as that of the first embodiment, but in the second embodiment, there is no need to perform a process where the intensities of frequency spectra which are acquired at step S14 or feature obtained after an attenuation correction process are multiplied by a factor of n , like the first embodiment. Subsequent steps S17 to S20 are the same as those of the first embodiment.

[0109] As described above, according to the second embodiment, as with the above-described first embodiment, a composite image can be obtained that allows identification of tissue characteristics while using advantages of a THI image having excellent resolution and reduced artifacts. In addition, by using a third feature image obtained by the weighted sum of the first and second features, noise in a tissue characteristic image can also be suppressed.

(Third Embodiment)

[0110] Next, a third embodiment of the present invention will be described. Note that a configuration of an ultrasound observation apparatus according to the third embodiment as a whole is the same as that of the first embodiment (see FIG. 1), but a tissue characteristic image data generation unit 52 does not necessarily need to be provided.

[0111] FIG. 19 is a flowchart illustrating the operation of the ultrasound observation apparatus according to the third embodiment of the present invention. Note that steps S10 to S17 illustrated in FIG. 19 are the same as those of the first embodiment. Alternatively, for steps S10, S11, and S16, those of the second embodiment may be applied.

[0112] At step S31 subsequent to step S17, an image composition unit 53 generates composite image data, using THI image data generated at step S11 and the third feature computed at step S17. More specifically, the image composition unit 53 first performs a threshold process on the third feature and thereby extracts an area including pixels whose third feature is smaller than a predetermined threshold. Then, for pixels in a THI image corresponding to respective pixel positions in the extracted area, the percentage of R, G, and B components is changed according to the value of the third feature. Namely, the hue of a pixel that is originally a grayscale pixel is changed such that the pixel becomes a color pixel. Note that for each pixel in the THI image corresponding to a pixel whose third feature is greater than or equal to the predetermined threshold, the pixel remains as its original grayscale pixel. Subsequent step S20 is the same as that of the first embodiment.

[0113] By such a technique, too, an image can be generated that allows identification of tissue characteristics while maintaining excellent resolution of a THI image.

[0114] Although the first to third embodiments and modifications of the present invention are described above, the present invention is not limited to the first to third embodiments and the modifications, and various inventions can be formed by appropriately combining together a plurality of elements which are disclosed in each embodiment or modification. For example, an invention may be formed by excluding some elements from all elements shown in each embodiment or modification, or an invention may be formed by appropriately combining together elements shown in different embodiments or modifications.

Reference Signs List

[0115]

1	ULTRASOUND OBSERVATION APPARATUS
2	ULTRASOUND PROBE
3	TRANSMITTING AND RECEIVING UNIT
4	COMPUTATION UNIT
5	IMAGE PROCESSING UNIT
6	INPUT UNIT
7	DISPLAY UNIT
8	STORAGE UNIT
9	CONTROL UNIT
21	SIGNAL CONVERTER
31	TRANSMISSION SIGNAL GENERATION UNIT
32	RECEIVED SIGNAL PROCESSING UNIT
41	AMPLIFICATION CORRECTION UNIT
42	FREQUENCY ANALYSIS UNIT
43	FEATURE EXTRACTION UNIT
51	THI IMAGE DATA GENERATION UNIT
52	TISSUE CHARACTERISTIC IMAGE DATA GENERATION UNIT
53	IMAGE COMPOSITION UNIT
91	TRANSMISSION SIGNAL CONTROLLER
92	DISPLAY CONTROLLER
100, 101, and 102	AREA
200	SCREEN
201 and 202	IMAGE DISPLAY AREA
203	DISPLAY AREA
431	APPROXIMATION UNIT
432	ATTENUATION CORRECTION UNIT
433	WEIGHTING AND ADDING UNIT

Claims

1. An ultrasound observation apparatus comprising:

an ultrasound probe configured to transmit a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and to receive first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal;

an image data generation unit configured to perform a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute a third received signal, and to generate image data of an ultrasound image having a predetermined display mode, using the third received signal;

a feature extraction unit configured to extract at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and

an image composition unit configured to generate image data of a composite image using the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

2. The ultrasound observation apparatus according to claim 1, wherein

the feature extraction unit is configured to compute the first and second features, and to calculate a weighted sum of the first and second features using positive weights to extract a third feature, and

the image composition unit is configured to generate the image data of the composite image having the display mode different from the predetermined display mode at the pixel position where the third feature is not less than or not more than the predetermined threshold.

3. The ultrasound observation apparatus according to claim 1 or 2, further comprising a second image data generation unit configured to generate image data of a second ultrasound image having a second display mode different from the predetermined display mode at the pixel position where the value based on at least one of the first and second features is not less than or not more than the predetermined threshold, wherein

the image composition unit is configured to combine the ultrasound image with the second ultrasound image to generate the image data of the composite image.

4. The ultrasound observation apparatus according to claim 3, wherein

the predetermined display mode is grayscale display,

the second display mode is color display, and

the second image data generation unit is configured to:

assign, to a first pixel whose value in the second ultrasound image based on at least one of the first and second features is not less than or not more than the predetermined threshold, a pixel value of a color according to the value; and

assign a pixel value of zero to a second pixel other than the first pixel in the second ultrasound image.

5. The ultrasound observation apparatus according to claim 4, wherein

the image composition unit is configured to:

assign a pixel value obtained by a weighted sum of pixel values of pixels of the ultrasound image and the second ultrasound image at corresponding positions, to a pixel in the composite image at a corresponding position to the first pixel; and

assign a pixel value of the second pixel to a pixel in the composite image at a corresponding position to the second pixel.

6. The ultrasound observation apparatus according to claim 1 or 2, wherein

the predetermined display mode is grayscale display, and

for a pixel in the ultrasound image corresponding to the pixel position where the value based on at least one of the

first and second features is not less than or not more than the predetermined threshold, the image composition unit is configured to change a percentage of each color component in the grayscale display according to the value to generate the image data of the composite image.

- 5 7. The ultrasound observation apparatus according to any one of claims 1 to 6, further comprising a display unit configured to display the composite image.
8. The ultrasound observation apparatus according to claim 7, wherein the display unit is configured to display the composite image and the ultrasound image side by side.
- 10 9. The ultrasound observation apparatus according to any one of claims 1 to 8, wherein the second ultrasound signal has a same phase as that of the first ultrasound signal and has an amplitude that is $1/n$ of an amplitude of the first ultrasound signal ($n > 0$ and $n \neq 1$), and the image data generation unit is configured to compute the third received signal by multiplying the second received signal by a factor of n and subtracting the resulting second received signal from the first received signal.
- 15 10. The ultrasound observation apparatus according to any one of claims 1 to 8, wherein the second ultrasound signal has a same amplitude as that of the first ultrasound signal and has a phase inverted relative to a phase of the first ultrasound signal, and the image data generation unit is configured to generate the third received signal by adding up the first and second received signals.
- 20

11. A method for operating an ultrasound observation apparatus, the method comprising:

25 an ultrasound transmitting and receiving step of, by an ultrasound probe, transmitting a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and receiving first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal;

30 an image data generation step of, by an image data generation unit, performing a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute a third received signal, and generating image data of an ultrasound image having a predetermined display mode, using the third received signal;

35 a feature extraction step of extracting, by a feature extraction unit, at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and

40 an image composition step of generating, by an image composition unit, image data of a composite image using the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

12. An operating program for an ultrasound observation apparatus, the program instructing the ultrasound observation apparatus to execute:

45 an ultrasound transmitting and receiving step of, by an ultrasound probe, transmitting a first ultrasound signal and a second ultrasound signal to a specimen along a same line, and receiving first and second ultrasound echo signals generated by reflection of the first and second ultrasound signals from the specimen, the second ultrasound signal having a different phase or amplitude from the first ultrasound signal;

50 an image data generation step of, by an image data generation unit, performing a calculation process on first and second received signals, which respectively correspond to the first and second ultrasound echo signals received by the ultrasound probe, to compute a third received signal, and generating image data of an ultrasound image having a predetermined display mode, using the third received signal;

55 a feature extraction step of extracting, by a feature extraction unit, at least one of a first feature and a second feature so as to be associated with a pixel position in an image, the first feature being based on a frequency spectrum of the first ultrasound echo signal and the second feature being based on a frequency spectrum of the second ultrasound echo signal; and

an image composition step of generating, by an image composition unit, image data of a composite image using

the image data of the ultrasound image, the composite image having a display mode different from the predetermined display mode at a pixel position where a value based on at least one of the first and second features is not less than or not more than a predetermined threshold, and the composite image having the predetermined display mode at another pixel position.

5

10

15

20

25

30

35

40

45

50

55

FIG.1




FIG.2




FIG.3




FIG.4




FIG.5




FIG.6




FIG.7




FIG.8




FIG.9




FIG.10




FIG.11




FIG.12




FIG.13




FIG.14




FIG.15




FIG.16




FIG.17




FIG.18





FIG.19



INTERNATIONAL SEARCH REPORT

International application No.

PCT/JP2014/078998

A. CLASSIFICATION OF SUBJECT MATTER

A61B8/08(2006.01)i, A61B8/12(2006.01)i, A61B8/14(2006.01)i

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

A61B8/08, A61B8/12, A61B8/14

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Jitsuyo Shinan Koho 1922-1996 Jitsuyo Shinan Toroku Koho 1996-2015

Kokai Jitsuyo Shinan Koho 1971-2015 Toroku Jitsuyo Shinan Koho 1994-2015

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y A	WO 2012/063976 A1 (Olympus Medical Systems Corp.), 18 May 2012 (18.05.2012), paragraphs [0029], [0037] to [0042], [0072] to [0091]; fig. 12 to 20 & US 2012/0310087 A1 & CN 102802536 A	1, 3-12 2
Y	WO 2014/007100 A1 (Hitachi Aloka Medical, Ltd.), 09 January 2014 (09.01.2014), paragraphs [0003], [0005] to [0007] (Family: none)	1, 3-12

☐ Further documents are listed in the continuation of Box C.☐ See patent family annex.

* Special categories of cited documents:

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier application or patent but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

"&" document member of the same patent family

Date of the actual completion of the international search
15 January 2015 (15.01.15)Date of mailing of the international search report
10 February 2015 (10.02.15)Name and mailing address of the ISA/
Japan Patent Office
3-4-3, Kasumigaseki, Chiyoda-ku,
Tokyo 100-8915, Japan

Authorized officer

Telephone No.

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- JP 2013056033 A [0004]
- JP 5114609 B [0004]

专利名称(译)	超声波观察装置，超声波观察装置的操作方法，超声波观察装置的操作程序		
公开(公告)号	EP3097859A1	公开(公告)日	2016-11-30
申请号	EP2014879286	申请日	2014-10-24
[标]申请(专利权)人(译)	奥林巴斯株式会社		
申请(专利权)人(译)	OLYMPUS CORPORATION		
当前申请(专利权)人(译)	OLYMPUS CORPORATION		
[标]发明人	MIYAKI HIRONAKA		
发明人	MIYAKI, HIRONAKA		
IPC分类号	A61B8/08 A61B8/12 A61B8/14		
CPC分类号	A61B8/5223 A61B8/08 A61B8/4461 A61B8/4477 A61B8/461 A61B8/463 A61B8/5207 A61B8/5246 G01N29/0609 G01N29/0654 G01N29/4418 G01N29/46 G01N2291/02475 G01N2291/02491 G01S7/52033 G01S7/52038 G01S15/8963		
优先权	2014011840 2014-01-24 JP		
其他公开文献	EP3097859A4		
外部链接	Espacenet		

摘要(译)

一种超声波观察装置，包括：THI图像数据生成单元，基于通过将具有不同相位或幅度的第一和第二超声信号沿着所述第一和第二超声信号发送到样本而获得的第一和第二接收信号，生成灰度THI图像的图像数据。线和接收从样本反射的第一和第二超声回波信号；特征提取单元，基于第一和第二超声回波信号的频谱计算第一和第二特征；组织特征图像数据生成单元，其在基于第一特征和第二特征的第三特征的值大于或等于或小于等于的区域中生成具有彩色显示模式的组织特征图像的图像数据。预定阈值；以及图像合成单元，其生成合成图像和组织特征图像的合成图像的图像数据。

FIG.15

