



(11) **EP 1 829 483 A1**

(12) **EUROPEAN PATENT APPLICATION**

(43) Date of publication: **05.09.2007 Bulletin 2007/36** (51) Int Cl.: **A61B 8/12 (2006.01)**

(21) Application number: **07004259.3**

(22) Date of filing: **01.03.2007**

(84) Designated Contracting States:
AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HU IE IS IT LI LT LU LV MC MT NL PL PT RO SE SI SK TR
Designated Extension States:
AL BA HR MK YU

- **Shikata, Hiroyuki**
Otawara-shi
Tochigi 324-8550 (JP)
- **Kosaku, Hideki**
Otawara-shi
Tochigi 324-8550 (JP)
- **Ogawa, Takashi**
Otawara-shi
Tochigi 324-8550 (JP)
- **Aoki, Minoru**
Otawara-shi
Tochigi 324-8550 (JP)

(30) Priority: **02.03.2006 JP 2006056464**

(71) Applicants:
• **KABUSHIKI KAISHA TOSHIBA**
Tokyo 105-8001 (JP)
• **Toshiba Medical Systems Corporation**
Otawara-shi,
Tochigi-ken 324-8550 (JP)

(74) Representative: **Kramer - Barske - Schmidtchen**
Radeckestrasse 43
81245 München (DE)

(72) Inventors:
• **Oonuki, Yutaka**
Otawara-shi
Tochigi 324-8550 (JP)

(54) **Ultrasonic diagnostic apparatus and ultrasonic diagnostic processing method**

(57) A generating unit (11) generates a plurality of ultrasonic cross-sectional images at a plurality of angles when an ultrasonic wave emitting surface in an ultrasonic probe (2) is rotated. A regenerating unit (12) saves information on the angle of the ultrasonic wave emitting sur-

face when a predetermined condition is satisfied, and rotates the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.




FIG. 3

EP 1 829 483 A1

Description

[0001] The present invention relates to an ultrasonic diagnostic apparatus and an ultrasonic diagnostic processing method in which a surface emitting ultrasonic waves emitted from an ultrasonic probe is rotated on an axis of this ultrasonic wave emitting surface and this rotation is used to generate a plurality of ultrasonic cross-sectional images from a plurality of angles to a subject.

[0002] An ultrasonic diagnostic apparatus scans the inside of a specimen with ultrasonic waves generated from an ultrasonic probe, and receives reflected waves from the inside of the specimen. The ultrasonic diagnostic apparatus acquires an ultrasonic diagnostic image of the inside of the specimen on the basis of a reception signal generated from the reflected waves from the inside of the specimen. In a diagnosis using the ultrasonic diagnostic apparatus, when a specimen such as one human body is inspected, a plurality of ultrasonic diagnostic images of the inside of the specimen may be generated. In this case, the ultrasonic diagnostic images may be saved.

[0003] Some of the ultrasonic diagnostic apparatuses conduct a transesophageal ultrasonic diagnosis. A transesophageal ultrasonic diagnostic apparatus comprises an ultrasonic probe. The ultrasonic probe allows the rotation of an ultrasonic wave emitting surface on the direction of this ultrasonic wave emitting surface. That is, the ultrasonic probe comprises a mechanism for rotating the ultrasonic wave emitting surface. For example, as shown in FIG. 10, the transesophageal ultrasonic diagnostic apparatus sets the angle of the ultrasonic wave emitting surface in the ultrasonic probe, generates an ultrasonic diagnostic image at the set angle, and then saves the generated ultrasonic diagnostic image.

[0004] In a diagnosis using the transesophageal ultrasonic diagnostic apparatus, a plurality of ultrasonic diagnostic images may be sequentially generated from a plurality of angular directions to a certain region of the inside of a specimen during the inspection of the specimen. In a series of operations for sequentially generating the ultrasonic diagnostic images from the plurality of angular directions, the ultrasonic diagnostic images may be generated in the following manner. First, an ultrasonic diagnostic image is saved at an arbitrary angle. Then, an ultrasonic diagnostic image is generated at an angle different from the arbitrary angle. Subsequently, an ultrasonic diagnostic image is again generated returning to the arbitrary angle.

[0005] The ultrasonic diagnostic image again generated corresponds to a region of the specimen which is important, for example, in diagnosing the specimen and which requires another observation. When an ultrasonic diagnostic image at the arbitrary angle is again generated, a rotating knob or switch provided in the ultrasonic probe is operated. Thus, the angle in the direction of the ultrasonic wave emitting surface is adjusted, and the ultrasonic diagnostic image at the arbitrary angle is generated.

[0006] Some of the ultrasonic diagnostic apparatuses comprise a mechanism which mechanically swings the ultrasonic probe in a slicing direction. In a diagnosis using such an ultrasonic diagnostic apparatus, an observation and a diagnosis are conducted concerning ultrasonic diagnostic images from a plurality of angular directions to, for example, a target within the specimen. In the above observation and diagnosis, an ultrasonic diagnostic image is observed and diagnosed at an arbitrary angle. Then, an ultrasonic diagnostic image is observed and diagnosed at an angle different from the arbitrary angle. Again, an ultrasonic diagnostic image at the arbitrary angle is observed and diagnosed. In this case, the operation of again generating the ultrasonic diagnostic image at the arbitrary angle is carried out looking at the ultrasonic diagnostic images.

[0007] An angle at which the ultrasonic diagnostic image is generated in a rotated direction of the ultrasonic wave emitting surface is saved as angle information. The angle information for the ultrasonic diagnostic image is added to the ultrasonic diagnostic image and saved together when this ultrasonic diagnostic image is saved. Then, the angle information is generally displayed together with the display of the ultrasonic diagnostic image.

[0008] According to the disclosure of Jpn. Pat. Appln. KOKAI Publication No. 2000-79121, when the updating of the display of the images is temporarily stopped in accordance with an instruction from an operator during a cine-display in which a plurality of images stored in a memory are sequentially displayed in a time-series order, the image displayed when the updating is stopped is automatically saved, and the automatically saved image is displayed in accordance with a call operation from the operator.

[0009] Therefore, when the ultrasonic diagnostic images at a plurality of angles are sequentially generated and the ultrasonic diagnostic image at the angle at which the diagnosis has been completed is again generated, it is necessary to operate the rotating knob or switch provided in the ultrasonic probe and carry out the regenerating looking at the ultrasonic diagnostic images. Thus, the operation once performed has to be performed again for the regenerating of the ultrasonic diagnostic image. This leads to an increase in diagnostic time.

[0010] It is an object of the present invention to provide an ultrasonic diagnostic apparatus and an ultrasonic diagnostic processing method which enable the regenerating of an ultrasonic diagnostic image at a desired angle with a simple operation to achieve a reduction in inspection time.

[0011] An ultrasonic diagnostic apparatus according to a first aspect of the present invention comprises: an ultrasonic probe in which a surface emitting ultrasonic waves is rotatable on an axis in the direction of this ultrasonic wave emitting surface; a generating unit which generates a plurality of ultrasonic cross-sectional images at a plurality of angles when the ultrasonic wave emitting surface in the ultrasonic probe is rotated; and a regen-

erating unit which saves information on the angle of the ultrasonic wave emitting surface when a predetermined condition is satisfied and which rotates the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

[0012] An ultrasonic diagnostic processing method according to a second aspect of the present invention comprises: rotating a surface emitting ultrasonic waves emitted from an ultrasonic probe on the direction of this ultrasonic wave emitting surface; generating a plurality of ultrasonic cross-sectional images at a plurality of angles when the ultrasonic wave emitting surface in the ultrasonic probe is rotated; saving information on the angle of the ultrasonic wave emitting surface when a predetermined condition is satisfied; and rotating the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

[0013] The invention can be more fully understood from the following detailed description when taken in conjunction with the accompanying drawings, in which:

FIG. 1 is an external configuration diagram showing a first embodiment of an ultrasonic diagnostic apparatus according to the present invention;

FIG. 2 is a schematic diagram showing the rotation of a surface emitting ultrasonic waves transmitted from the distal end of an ultrasonic probe in the same apparatus;

FIG. 3 is a block configuration diagram showing a diagnostic processing device in the same apparatus; FIG. 4 is a diagram showing one example of angle information displayed on a display screen of an operation panel in the same apparatus;

FIG. 5 is a diagram showing the flow of a method of saving the angle information for an ultrasonic diagnostic image in the same apparatus;

FIG. 6 is a diagram showing the flow of a method of saving the angle information for the ultrasonic diagnostic image in a second embodiment of the ultrasonic diagnostic apparatus according to the present invention;

FIG. 7 is a diagram showing the flow of a method of saving the angle information for the ultrasonic diagnostic image in a third embodiment of the ultrasonic diagnostic apparatus according to the present invention;

FIG. 8A is a diagram showing touch buttons and reduced ultrasonic cross-sectional images displayed on the display screen of the operation panel in the ultrasonic diagnostic apparatus according to the present invention;

FIG. 8B is a diagram showing the reduced ultrasonic cross-sectional images doubling as the touch buttons displayed on the display screen of the operation

panel in the ultrasonic diagnostic apparatus according to the present invention;

FIG. 9 is a diagram showing a return touch button displayed on the display screen of the operation panel in the ultrasonic diagnostic apparatus according to the present invention; and

FIG. 10 is a diagram showing the flow of a conventional operation method of generating an ultrasonic diagnostic image.

[0014] Hereinafter, a first embodiment of the present invention will be described with reference to the drawing.

[0015] FIG. 1 shows a diagram of the external appearance of an ultrasonic diagnostic apparatus applied to a transesophageal ultrasonic diagnosis. A ultrasonic probe 2 is connected to an ultrasonic diagnostic apparatus 1. As shown in FIG. 2, a plurality of ultrasonic wave oscillating elements 3 are provided at the distal end of the ultrasonic probe 2. Each of the ultrasonic wave oscillating elements 3 transmits ultrasonic waves 4.

[0016] The ultrasonic wave oscillating elements 3 are, for example, one-dimensionally arranged. An acoustic lens 3a is provided in front of the ultrasonic wave oscillating elements 3. The acoustic lens 3a converges the ultrasonic waves 4 emitted from each of the ultrasonic wave oscillating elements 3. Each of the one-dimensionally arranged ultrasonic wave oscillating elements 3 can mechanically rotate, for example, in a direction of an arrow R_2 . When each of the ultrasonic wave oscillating elements 3 mechanically rotates in the direction of the arrow R_2 , an ultrasonic wave emitting surface 4a rotates in a direction of an arrow R_1 .

[0017] Alternatively, the ultrasonic wave oscillating elements 3 are arranged, for example, on a two-dimensional plane. The ultrasonic wave oscillating elements 3 arranged on the two-dimensional plane are electrically driven/controlled so that some of the ultrasonic wave oscillating elements 3 emit the ultrasonic waves 4 and some of the ultrasonic wave oscillating elements 3 do not emit the ultrasonic waves. Thus, each of the ultrasonic wave oscillating elements 3 rotates, in the direction of the arrow R_1 , the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 to be emitted.

[0018] Each of the one-dimensionally or two-dimensionally arranged ultrasonic wave oscillating elements 3 rotates, in the direction of the arrow R_1 , the ultrasonic wave emitting surface 4a on an axis x included in the direction of the surface emitting the ultrasonic waves 4. An axis x is set within the ultrasonic wave emitting surface 4a of the ultrasonic waves 4. The axis x is set to be, for example, in the center of the ultrasonic wave oscillating elements 3, and to be vertical to the one-dimensional direction or two-dimensional plane in which the ultrasonic wave oscillating elements 3 are arranged.

[0019] An operation portion 5 is provided in a grip of the ultrasonic probe 2. The operation portion 5 is provided with, for example, various switches for operating the ultrasonic probe 2. The operation portion 5 is provided with

a switch 5a for regenerating.

[0020] FIG. 3 is a block configuration diagram of the ultrasonic diagnostic apparatus 1. The ultrasonic diagnostic apparatus 1 computer-processes an output signal of the ultrasonic probe 2 to obtain an ultrasonic cross-sectional image of a specimen. The ultrasonic diagnostic apparatus 1 has a main control unit 6 configured by, for example, a CPU. To the main control unit 6, there are connected an input/output unit 7, an angle information memory 8, a program memory 9, an image memory 10, a generating unit 11, a regenerating unit 12, an angle information display unit 13, an image display 14 and an operation panel 15. In addition, regarding the regenerating unit 12 and the angle information display unit 13 in FIG. 3, functions when a program for ultrasonic diagnostic processing is processed by a computer are shown in block diagrams.

[0021] The main control unit 6 executes the program for the ultrasonic diagnostic processing stored in the program memory 9 in order to save data in or read data from the angle information memory 8 and the image memory 10, to send operation commands to the input/output unit 7, the regenerating unit 12 and the angle information display unit 13, to display images on the image display 14, and to receive the operation commands from the operation panel 15.

[0022] The ultrasonic probe 2 is connected to the input/output unit 7. The ultrasonic probe 2 is provided with an angle sensor 16. The angle sensor 16 detects the mechanical angle of each of the ultrasonic wave oscillating elements 3 which are, for example, one-dimensionally arranged. For example, the angle sensor 16 detects the angle of rotation in, for example, a slicing direction of each of the one-dimensionally arranged ultrasonic wave oscillating elements 3, and outputs an angle signal corresponding to this angle. Further, the angle sensor 16 detects the angle of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 emitted from each of the ultrasonic wave oscillating elements 3 which are, for example, two-dimensionally arranged. In the case of the two-dimensionally arranged ultrasonic wave oscillating elements 3, the angle sensor 16 detects the angle of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 rotated in the direction of the arrow R_1 , for example, in accordance with the driving/control of each of the ultrasonic wave oscillating elements 3, and outputs an angle signal corresponding to the detected angle.

[0023] Whether the ultrasonic wave oscillating elements 3 are arranged one-dimensionally or two-dimensionally, the angle sensor 16 detects the angle in the direction of the arrow R_1 of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 emitted from each of the ultrasonic wave oscillating elements 3. The angle of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 is, for example, an angle to a preset referential angle position.

[0024] The angle information memory 8 stores as angle information the angle signal output from the angle

sensor 16.

[0025] The program memory 9 stores the program for the ultrasonic diagnostic processing executed by the CPU of the computer. The program for the ultrasonic diagnostic processing causes the ultrasonic cross-sectional images from a plurality of angles to be generated on the image display 14 when the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 is rotated on the axis x. The program for the ultrasonic diagnostic processing causes the information on the angle of the ultrasonic wave emitting surface 4a to be saved in the angle information memory 8 when a predetermined condition is satisfied, and causes the ultrasonic wave emitting surface 4a to be rotated from the angle of the ultrasonic wave emitting surface 4a different from the saved angle information to an angle based on the angle information saved in the angle information memory 8. The program for the ultrasonic diagnostic processing includes a program for saving the angle information in the angle information memory 8 when the angle at which the ultrasonic cross-sectional image is generated is maintained for a given time.

[0026] When the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 is rotated on the axis x, the generating unit 11 generates ultrasonic cross-sectional images from a plurality of angles, and generates the ultrasonic cross-sectional images on the image display 14.

[0027] The regenerating unit 12 judges, for example, whether the angle at which the predetermined condition is satisfied, for example, at which the ultrasonic cross-sectional image is generated on the image display 14 is maintained for a given time. Specifically, the regenerating unit 12 judges whether the angle of the ultrasonic wave emitting surface 4a at which an ultrasonic cross-sectional image of a certain region within the specimen is generated remains unchanged for a preset given time. When the result of the judgment is that the angle is maintained for a given time, the regenerating unit 12 saves the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated as angle information in the angle information memory 8. Regarding the angle information, the angle signal output from the angle sensor 16 is stored as the angle information in the angle information memory 8.

[0028] The regenerating unit 12 judges whether the switch 5a for regenerating provided in the operation portion 5 is operated. If the switch 5a for regenerating is operated, the regenerating unit 12 reads the angle information saved in the angle information memory 8. In accordance with the read angle information, the regenerating unit 12 rotates the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 emitted from the ultrasonic probe 2 in the direction of the arrow R_1 shown in FIG. 2. Then, the regenerating unit 12 sets the angle of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 to the angle information read from the angle information memory 8. The regenerating unit 12 redraws a particular ultrasonic cross-sectional image corresponding to the

angle information on the image display 14.

[0029] The regenerating unit 12 judges whether there is a touch operation for angle information 18 displayed on a display screen 17 of the operation panel 15, for example, as shown in FIG. 4. When a touch operation for the angle information 18 is performed, the regenerating unit 12 receives coordinates corresponding to the angle information 18 sent from the operation panel 15, and rotates the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 in the direction of the arrow R_1 shown in FIG. 2 in accordance with the angle information 18 corresponding to the coordinates. Then, the regenerating unit 12 sets the angle of the ultrasonic wave emitting surface 4a of the ultrasonic waves 4 to the angle information read from the angle information memory 8, in the same manner as described above. The regenerating unit 12 redraws a particular ultrasonic cross-sectional image corresponding to the angle information on the image display 14.

[0030] The angle information display unit 13 displays the angle information saved in the angle information memory 8 on the display screen 17 of the operation panel 15. FIG. 4 shows one example of the angle information 18 displayed on the display screen 17 of the operation panel 15. The angle information 18 indicates, for example, four angles A to D. The display form of the angle information 18 for the angles A to D is not limited to the rectangular form shown in FIG. 4, and the angle information 18 may be displayed in a desired shape, display color, layout, etc. The angle information 18 indicates, for example, touch buttons 18a to 18d for the angles A to D. Coordinates to be displayed on the display screen 17 are set in the touch buttons 18a to 18d.

[0031] The operation panel 15 is formed by, for example, a touch command screen (TCS). Therefore, when one indication of the angle information 18 displayed on the display screen 17 of the operation panel 15 is touched, the operation panel 15 reads, for example, the coordinates on the display screen 17 corresponding to the touched indication of the angle information 18, and sends the coordinates corresponding to this indication of the angle information 18 to the regenerating unit 12.

[0032] Next, the operation of the redrawing in the apparatus configured as described above will be described.

[0033] The ultrasonic probe 2 scans the inside of the specimen with the ultrasonic waves 4, and receives reflected waves from the inside of the specimen. The ultrasonic probe 2 outputs a reception signal generated from the reflected waves from the inside of the specimen. The reception signal from the ultrasonic probe 2 is input to the generating unit 11 through the input/output unit 7, and the generating unit 11 acquires an ultrasonic diagnostic image of the inside of the specimen, and then displays the ultrasonic diagnostic image on the image display 14.

[0034] In a diagnosis using the transesophageal ultrasonic diagnostic apparatus, the ultrasonic wave emitting surface 4a is rotated in the direction of the arrow R_1 on

the axis x set in the direction of the ultrasonic wave emitting surface 4a as shown in FIG. 2, during the diagnosis of the specimen. Thus, a plurality of ultrasonic diagnostic images from a plurality of angles to a certain region within the specimen are sequentially generated on the image display 14.

[0035] During the diagnosis of a certain region within the specimen, the regenerating unit 12 judges whether the angle of the ultrasonic diagnostic image of a certain region within the specimen generated on the image display 14 remains unchanged for a preset given time, for example, as shown in FIG. 5.

[0036] The fact that the angle of the ultrasonic diagnostic image remains unchanged for a preset given time means that the region in which the ultrasonic diagnostic image is generated is important in diagnosing the specimen and that this angle is the angle at which the observation of the region is needed.

[0037] Therefore, when the angle at which the ultrasonic diagnostic image is generated remains unchanged for a preset given time, the regenerating unit 12 saves, in the angle information memory 8, the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic diagnostic image is generated on the image display 14, for example, the angle information for an angle A of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2. The angle information display unit 13 displays the angle information saved in the angle information memory 8 as the angle information 18 on the display screen 17 of the operation panel 15, for example, as shown in FIG. 4. The angle information 18 is indicated, for example, by the touch button 18a for the angle A.

[0038] When a plurality of ultrasonic diagnostic images from a plurality of angular directions to a certain region within the specimen are sequentially generated on the image display 14 in the same manner as described above, there may be a plurality of ultrasonic diagnostic images in which the angle of the ultrasonic wave emitting surface 4a at which these ultrasonic diagnostic images are generated remains unchanged for a preset given time. In this case, the regenerating unit 12 sequentially saves the information on the angle of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 in the angle information memory 8 every time the angle of the ultrasonic wave emitting surface 4a remains unchanged for a given time.

[0039] The angle information display unit 13 displays a plurality of pieces of angle information saved in the angle information memory 8 as the angle information 18 on the display screen 17 of the operation panel 15, for example, as shown in FIG. 4. The angle information 18 is composed of, for example, the touch buttons 18a to 18d for the four angles A to D.

[0040] For example, the generating unit 11 draws an ultrasonic diagnostic image at the angle A of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2. If the angle A of the ultrasonic wave emitting surface 4a remains unchanged for a given time, the regenerating

unit 12 saves the angle A of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 as the angle information in the angle information memory 8. The angle A of the ultrasonic wave emitting surface 4a at this point corresponds to the touch button 18a.

[0041] Next, the generating unit 11 draws the ultrasonic cross-sectional image at the angle B of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2. If the angle B of the ultrasonic wave emitting surface 4a remains unchanged for a given time, the regenerating unit 12 saves the angle B of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 as the angle information in the angle information memory 8. The angle B of the ultrasonic wave emitting surface 4a at this point corresponds to the touch button 18b.

[0042] Subsequently, the regenerating unit 12 saves the angle C of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 as the angle information in the angle information memory 8 in a similar manner. The angle C of the ultrasonic wave emitting surface 4a at this point corresponds to the touch button 18c. The regenerating unit 12 saves the angle D of the ultrasonic wave emitting surface 4a of the ultrasonic probe 2 as the angle information in the angle information memory 8. The angle D of the ultrasonic wave emitting surface 4a at this point corresponds to the touch button 18c.

[0043] Then, the ultrasonic cross-sectional image of the region may be again diagnosed at the angle of the direction of the emitting surface corresponding to the angle information 18 for the angle A displayed on the display screen 17 of the operation panel 15. In this case, the touch button 18a for the angle A is touched out of the angle information 18 displayed on the display screen 17 of the operation panel 15. The operation panel 15 reads, for example, the coordinates on the display screen 17 corresponding to the angle information 18 for the touched touch button 18a, and sends the coordinates corresponding to the angle information 18 to the regenerating unit 12.

[0044] The regenerating unit 12 receives the coordinates sent from the operation panel 15, and recognizes the angle A in the direction of the emitting surface corresponding to the touch button 18a on the basis of the coordinates. The regenerating unit 12 sends a command of the angle A to the ultrasonic probe 2. The ultrasonic probe 2 rotates the direction of the emitting surface for the ultrasonic waves 4 on the axis x in the direction of the arrow R_1 , and sets the emitting surface at the angle A. Thus, the ultrasonic probe 2 scans the inside of the specimen with the ultrasonic waves 4 at the angle A, receives reflected waves from the inside of the specimen, and outputs a reception signal generated from the reflected waves. The reception signal from the ultrasonic probe 2 is input to the generating unit 11 through the input/output unit 7, and the generating unit 11 acquires an ultrasonic diagnostic image at the angle A, and then redraws the ultrasonic diagnostic image on the image display 14. As a result, the ultrasonic diagnostic images at the angle A which is important in diagnosing the spec-

imen can be again diagnosed.

[0045] The ultrasonic cross-sectional images of the region may be again diagnosed at the angles in the direction of the emitting surface corresponding to the angle information 18 for the angles A and B displayed on the display screen 17 of the operation panel 15. In this case, the touch button 18a for the angle A is first touched out of the angle information 18 displayed on the display screen 17 of the operation panel 15. The operation panel 15 reads, for example, the coordinates on the display screen 17 corresponding to the angle information 18 for the touched touch button 18a, and sends the coordinates corresponding to the angle information 18 to the regenerating unit 12. The regenerating unit 12 receives the coordinates sent from the operation panel 15, and, in the same manner as described above, acquires an ultrasonic diagnostic image at the angle A, and then redraws the ultrasonic diagnostic image on the image display 14.

[0046] Next, the touch button 18b for the angle B is touched out of the angle information 18 displayed on the display screen 17 of the operation panel 15. The operation panel 15 reads, for example, the coordinates on the display screen 17 corresponding to the angle information 18 for the touched touch button 18b, and sends the coordinates corresponding to the angle information 18 to the regenerating unit 12. The regenerating unit 12 receives the coordinates sent from the operation panel 15, and, in the same manner as described above, acquires an ultrasonic diagnostic image at the angle B, and then redraws the ultrasonic diagnostic image on the image display 14.

[0047] As a result, the ultrasonic cross-sectional images, for example, when the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 is at the angles A and B can be regenerated on the image display 14. The ultrasonic diagnostic images at the angles A and B which are important in diagnosing the specimen can be again diagnosed.

[0048] On the other hand, in the same manner as described above, when the specimen is diagnosed, there is an angle of the ultrasonic wave emitting surface 4a in a certain region which is important in diagnosing the specimen and which requires another observation. For example, there is a case where the ultrasonic diagnostic image which has been generated is regenerated, for example, at one angle before the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic diagnostic image is currently being generated on the image display 14. In this case, the switch 5a for redrawing provided in the operation portion 5 is operated.

[0049] In response to the operation of the switch 5a for redrawing, the regenerating unit 12 reads, from the angle information memory 8, the previous angle information, for example, reads the angle such as the angle A which is one angle before the angle at which the ultrasonic diagnostic image is currently generated. The regenerating unit 12 redraws, on the image display 14, the ultrasonic diagnostic image at the angle in the direction

of the emitting surface corresponding to the angle information for the angle A.

[0050] In this case, the regenerating unit 12 sends a command of the angle A to the ultrasonic probe 2. The ultrasonic probe 2 rotates the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 on the axis x in the direction of the arrow R_1 , and sets the ultrasonic wave emitting surface 4a at the angle A. Thus, the ultrasonic probe 2 scans the inside of the specimen with the ultrasonic waves 4 at the angle A. The ultrasonic probe 2 receives reflected waves from the inside of the specimen, and outputs a reception signal generated from the reflected waves. The reception signal from the ultrasonic probe 2 is input to the generating unit 11 through the input/output unit 7. The generating unit 11 acquires an ultrasonic diagnostic image at the angle A of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4, and redraws the ultrasonic diagnostic image on the image display 14.

[0051] As a result, the region which is important in diagnosing the specimen and which requires another observation is regenerated on the image display 14. Thus, it is possible to diagnose the ultrasonic diagnostic image at a desired angle, for example, at the angle A of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4.

[0052] In this manner, according to the first embodiment described above, when the angle of the direction of the emitting surface for the ultrasonic waves 4 at which the ultrasonic cross-sectional image is generated is maintained for a given time, the angle information is saved in the angle information memory 8. Thus, it is possible to redraw the ultrasonic diagnostic image at the desired angle of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 with a simple operation and to achieve a reduction in diagnostic time.

[0053] For example, in the diagnosis using the transesophageal ultrasonic diagnostic apparatus, the ultrasonic wave emitting surface 4a is rotated on the axis in the direction of the arrow R_1 as shown in FIG. 2 during the diagnosis of the specimen, and a plurality of ultrasonic diagnostic images are sequentially generated on the image display 14 from a plurality of angular directions to a certain region within the specimen. During such an inspection of a certain region of the specimen, the angle of the ultrasonic wave emitting surface 4a remains unchanged for a preset given time in the diagnosis of the region which is important in diagnosing the specimen. This ensures that the angle information for the ultrasonic diagnostic image of the region which is important in diagnosing the specimen can be saved in the angle information memory 8.

[0054] Therefore, when the ultrasonic diagnostic image of the region which is important in diagnosing the specimen is regenerated from a desired angle, it is possible to redraw, on the image display 14, the ultrasonic diagnostic image of the region which is important in diagnosing the specimen with a simple operation; for ex-

ample, operating the switch 5a for redrawing or touching the touch buttons 18a to 18d on the display screen 17 of the operation panel 15. As a result, it is possible to again diagnose the ultrasonic diagnostic image at the angle which is important in diagnosing the specimen.

[0055] The ultrasonic diagnostic image at a desired angle can be regenerated with a simple operation to achieve a reduction in inspection time, such that a certain region of the specimen can be diagnosed in real time owing to the regenerated ultrasonic diagnostic image at the desired angle. For example, when a region of the specimen is treated, for example, burned off, it is possible to administer a proper treatment by observing the regenerated ultrasonic diagnostic image in real time. Moreover, it is possible to redraw and observe in real time the ultrasonic diagnostic image of the region of the specimen after the treatment such as the burn-off has been administered.

[0056] In addition, the first embodiment described above may be modified in the following manner.

[0057] The regenerating unit 12 redraws the ultrasonic diagnostic image at one angle before the angle of the currently generated ultrasonic diagnostic image in accordance with the operation of the switch 5a for redrawing. This is not a limitation. For example, when the switch 5a for redrawing is operated N (= integer number) times in a row, the regenerating unit 12 may redraw the ultrasonic diagnostic image at the angle N times before the angle of the currently generated ultrasonic diagnostic image. In this case, the angle information N times before back in time is read out of the angle information saved in the angle information memory 8 in a time-series manner, and the ultrasonic cross-sectional image for the read angle information is regenerated.

[0058] Next, a second embodiment of the present invention will be described with reference to the drawing. It is to be noted that the same signs are assigned to the same parts as those in FIGS. 1 to 3, and these parts will not be described in detail.

[0059] When a predetermined condition is satisfied, for example, when the ultrasonic cross-sectional image generated on the image display 14 is saved in the image memory 10, the regenerating unit 12 saves the angle information for the saved ultrasonic cross-sectional image in the angle information memory 8.

[0060] The program for the ultrasonic diagnostic processing stored in the program memory 9 includes a program which causes the information on the angle of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 at which the saved ultrasonic cross-sectional image is generated to be saved in the angle information memory 8 when the ultrasonic cross-sectional image generated on the image display 14 is saved in the image memory 10.

[0061] In such a configuration, the ultrasonic wave emitting surface 4a is rotated on the axis x in the direction of the arrow R_1 as shown in FIG. 2 during the diagnosis of the specimen in the same manner as described above.

During the rotation of the ultrasonic wave emitting surface 4a, a plurality of ultrasonic diagnostic images are sequentially generated on the image display 14 from a plurality of angular directions to a certain region within the specimen.

[0062] During this diagnosis, if the ultrasonic diagnostic image at the angle which is important in diagnosing the specimen is generated on the image display 14 among the ultrasonic diagnostic images at the respective angles to the specimen, an operation is performed on the operation panel 15 for a save instruction. In response to the save instruction from the operation panel 15, the main control unit 6 saves the ultrasonic diagnostic image at the angle which is important in diagnosing the specimen in the image memory 10.

[0063] At this point, the regenerating unit 12 saves, in the angle information memory 8, the information on the angle of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 at which the ultrasonic cross-sectional image saved in the image memory 10 is generated, as shown in FIG. 6. In addition, when there are a plurality of ultrasonic cross-sectional images saved in the image memory 10, the regenerating unit 12 sequentially saves, in the angle information memory 8, the information on the angle of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 at which each ultrasonic cross-sectional image is generated every time each ultrasonic cross-sectional image is saved in the image memory 10.

[0064] Along with this, the angle information display unit 13 displays a plurality of pieces of angle information saved in the angle information memory 8, for example, by indicating the touch buttons 18a to 18d for the four angles A to D as the angle information 18 on the display screen 17 of the operation panel 15 as shown in FIG. 4.

[0065] Next, in the case of redrawing the ultrasonic diagnostic image generated at the angle to the region which is important in diagnosing the specimen and which requires another observation, for example, at one angle before that of the ultrasonic diagnostic image currently generated on the image display 14, the switch 5a for redrawing provided in the operation portion 5 is operated, as in the first embodiment described above.

[0066] On the other hand, in the case of again diagnosing the ultrasonic diagnostic image of the region corresponding to the angle information 18 for the angle to the region which is important in diagnosing the specimen and which requires another observation, for example, the angle A displayed on the display screen 17 of the operation panel 15, the touch button 18a for the angle A is touched out of the angle information 18 displayed on the display screen 17 of the operation panel 15, as in the first embodiment described above. As a result of this operation, the ultrasonic diagnostic image at a desired angle to the region which is important in diagnosing the specimen, for example, at the angle A of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 is regenerated. Consequently, it is possible to again diagnose the ultrasonic diagnostic images at the angle A which is

important in diagnosing the specimen.

[0067] In this manner, according to the second embodiment, when the ultrasonic cross-sectional image generated on the image display 14 is saved in the image memory 10, the information on the angle of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 at which this ultrasonic cross-sectional image is generated is saved in the angle information memory 8. Thus, the second embodiment can provide effects similar to those in the first embodiment described above.

[0068] Next, a third embodiment of the present invention will be described with reference to the drawing. It is to be noted that the same signs are assigned to the same parts as those in FIGS. 1 to 3, and these parts will not be described in detail.

[0069] When a predetermined condition is satisfied, for example, when the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated on the image display 14 is changed, the regenerating unit 12 saves, in the angle information memory 8, the angle information indicating the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated before the change of the angle of the ultrasonic wave emitting surface 4a.

[0070] The program for the ultrasonic diagnostic processing stored in the program memory 9 includes a program which causes the angle information indicating the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated before the change of the angle of the ultrasonic wave emitting surface 4a to be saved in the angle information memory 8 when the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated on the image display 14 is changed.

[0071] In such a configuration, the ultrasonic wave emitting surface 4a is rotated on the axis x in the direction of the arrow R_1 as shown in FIG. 2 during the inspection of the specimen in the same manner as described above, and a plurality of ultrasonic diagnostic images are sequentially generated on the image display 14 from a plurality of angular directions to a certain region within the specimen.

[0072] During this diagnosis, when the angle for the ultrasonic diagnostic image is changed, the regenerating unit 12 saves, in the angle information memory 8, the angle information for the ultrasonic cross-sectional image generated before the change of the angle, as shown in FIG. 7. In addition, if the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic diagnostic image is generated is changed a plurality of times, the regenerating unit 12 sequentially saves, in the angle information memory 8, the information on the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic diagnostic image is generated, every time the angle is changed.

[0073] Along with this, the angle information display unit 13 displays a plurality of pieces of angle information

saved in the angle information memory 8, for example, by indicating the touch buttons 18a to 18d for the four angles A to D as the angle information 18 on the display screen 17 of the operation panel 15 as shown in FIG. 4.

[0074] Next, in the case of redrawing the ultrasonic diagnostic image generated at the angle to the region which is important in diagnosing the specimen and which requires another observation, for example, at one angle before that of the ultrasonic diagnostic image currently generated on the image display 14, the switch 5a for redrawing provided in the operation portion 5 is operated, as in the first embodiment described above.

[0075] On the other hand, in the case of again diagnosing the ultrasonic diagnostic image of the region corresponding to the angle information 18 for the angle to the region which is important in diagnosing the specimen and which requires another observation, for example, the angle A displayed on the display screen 17 of the operation panel 15, the touch button 18a for the angle A is touched out of the angle information 18 displayed on the display screen 17 of the operation panel 15, as in the first embodiment described above.

[0076] As a result of this operation, the ultrasonic diagnostic image at a desired angle to the region which is important in diagnosing the specimen is regenerated. It is possible to again diagnose the ultrasonic diagnostic images at the angle which is important in diagnosing the specimen.

[0077] In this manner, according to the third embodiment, when the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic cross-sectional image is generated on the image display 14 is changed, the angle information indicating the angle of the ultrasonic wave emitting surface 4a at which the ultrasonic diagnostic image is generated before the angle of the ultrasonic wave emitting surface 4a is changed is saved in the angle information memory 8. Thus, the third embodiment can provide effects similar to those in the first embodiment described above.

[0078] It is to be noted that the present invention is not limited to the embodiments described above, and may be modified in the following manner.

[0079] For example, the angle information display unit 13 displays the touch buttons 18a to 18d for the four angles A to D as the angle information 18 on the display screen 17 of the operation panel 15 as shown in FIG. 4. This is not a limitation. For example, the angle information display unit 13 may display reduced ultrasonic cross-sectional images (thumbnail images) Da to Dd corresponding to the four angles A to D as well as the touch buttons 18a to 18d on the display screen 17 of the operation panel 15 as shown in FIG. 8A. The reduced images Da to Dd are displayed side by side with the touch buttons 18a to 18d, respectively.

[0080] The angle information display unit 13 may also display the reduced ultrasonic cross-sectional images (thumbnail images) Da to Dd corresponding to the four angles A to D alone on the display screen 17 of the op-

eration panel 15 as shown in FIG. 8B. Each of the reduced images (thumbnail images) Da to Dd functions as a touch button. For example, when the reduced image Da is touched, the regenerating unit 12 redraws the ultrasonic diagnostic image at the angle A of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4 on the image display 14.

[0081] The angle information display unit 13 displays a plurality of pieces of angle information saved in the angle information memory 8, for example, by indicating the touch buttons 18a to 18d for the four angles A to D as the angle information 18 on the display screen 17 of the operation panel 15 as shown in FIG. 4 and FIG. 8A. This is not a limitation, and the angle information display unit 13 may display the touch buttons 18a to 18d for the four angles A to D as the angle information 18 on the image display 14. The image display 14 displays, for example, an angle for the diagnosis of the specimen and a rotation angle of the transesophageal probe, and may display the angle information 18 together.

[0082] For example, a return touch button 19 may be displayed on the display screen 17 of the operation panel 15 as shown in FIG. 9. The return touch button 19 sends, to the regenerating unit 12, a command to sequentially return to the previous drawing of the ultrasonic cross-sectional image at each operation. For example, the angle information for the four angles A to D is sequentially saved in the angle information memory 8. In this case, when the return touch button 19 is operated, the regenerating unit 12 draws the previous angle information from the angle information memory 8 in response to each operation of the return touch button 19. For example, if the ultrasonic diagnostic image at the angle D is generated at present, the regenerating unit 12 then draws the ultrasonic diagnostic image at the angle C. If the return touch button 19 is continuously operated, the regenerating unit 12 draws, for example, the ultrasonic diagnostic image at the angle D, then draws the ultrasonic diagnostic image at the angle C, then draws the ultrasonic diagnostic image at the angle B, and then draws the ultrasonic diagnostic image at the angle A.

[0083] A return switch is not limited to the return touch button 19 displayed on the display screen 17. The return switch may be provided on a surface other than the display screen 17 in the operation panel 15. For example, as shown in FIG. 1, a return switch 20 may be provided on a main unit front surface 21 of the ultrasonic diagnostic apparatus 1. The return switch 20 is, for example, a mechanical switch such as a push button or switch button, or a touch button exclusive to the return switch.

[0084] For example, when one return switch 20 is provided, return angle information allocated to the return switch 20 is displayed on, for example, the display screen 17 in the operation panel 15. If the return switch 20 is operated, the regenerating unit 12 redraws a particular ultrasonic cross-sectional image generated one image before.

[0085] When a plurality of mechanical return switches

20 such as the push buttons or switch buttons are provided on the main unit front surface 21 of the ultrasonic diagnostic apparatus 1, return angle information allocated to each of the return switches 20 is displayed on, for example, the display screen 17 in the operation panel 15. The return angle information includes, for example, the four angles A to D of the ultrasonic wave emitting surface 4a for the ultrasonic waves 4.

[0086] It is explicitly stated that all features disclosed in the description and/or the claims are intended to be disclosed separately and independently from each other for the purpose of original disclosure as well as for the purpose of restricting the claimed invention independent of the compositions of the features in the embodiments and/or the claims. It is explicitly stated that all value ranges or indications of groups of entities disclose every possible intermediate value or intermediate entity for the purpose of original disclosure as well as for the purpose of restricting the claimed invention, in particular as limits of value ranges.

Claims

1. An ultrasonic diagnostic apparatus **characterized by** comprising:

an ultrasonic probe (2) in which a surface emitting ultrasonic waves is rotatable on an axis in the direction of this ultrasonic wave emitting surface;

a generating unit (11) which generates a plurality of ultrasonic cross-sectional images at a plurality of angles when the ultrasonic wave emitting surface in the ultrasonic probe is rotated; and
a regenerating unit (12) which saves information on the angle of the ultrasonic wave emitting surface when a predetermined condition is satisfied and which rotates the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

2. An ultrasonic diagnostic apparatus **characterized by** comprising:

an ultrasonic probe (2) in which a plurality of ultrasonic wave oscillating elements are one-dimensionally arranged;

an acoustic lens (3a) which converges ultrasonic waves emitted from the ultrasonic probe;
a generating unit (11) which mechanically rotates the ultrasonic probe with respect to the acoustic lens and which rotates a surface emitting the ultrasonic waves to generate a three-dimensional ultrasonic cross-sectional image; and

a regenerating unit (12) which saves information on the angle of the ultrasonic wave emitting surface when a predetermined condition is satisfied and which rotates the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

3. An ultrasonic diagnostic apparatus **characterized by** comprising:

an ultrasonic probe (2) in which a plurality of ultrasonic wave oscillating elements are two-dimensionally arranged;

a generating unit (11) which controls the operation of the ultrasonic probe to rotate the direction of a surface emitting ultrasonic waves on an axis of this ultrasonic wave emitting surface and which generates a three-dimensional ultrasonic cross-sectional image; and
a regenerating unit (12) which saves information on the angle of the ultrasonic wave emitting surface when a predetermined condition is satisfied and which rotates the ultrasonic wave emitting surface from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

4. The ultrasonic diagnostic apparatus according to claim 1, 2 or 3, **characterized in that:**

the regenerating unit (12) saves the angle information in the case where the angle of the ultrasonic wave emitting surface is maintained for a given time when the ultrasonic cross-sectional image is generated.

5. The ultrasonic diagnostic apparatus according to claim 1, 2 or 3, **characterized in that:**

the regenerating unit (12) saves the angle information when the ultrasonic cross-sectional image is saved.

6. The ultrasonic diagnostic apparatus according to claim 1, 2 or 3, **characterized in that:**

the regenerating unit (12) saves the angle information before the angle is changed in the case where the angle of the ultrasonic wave emitting surface is changed.

7. The ultrasonic diagnostic apparatus according to claim 1, 2 or 3, **characterized by** further comprising:

an operation unit (5) which operates the ultra-

- sonic probe; and
 a switch (15a) provided in the operation unit,
 wherein the regenerating unit (12) rotates the
 ultrasonic wave emitting surface to the angle
 based on the saved angle information when the
 switch is operated. 5
8. The ultrasonic diagnostic apparatus according to
 claim 1, 2 or 3, **characterized by** further comprising:
 10
 an operation panel (15) having a display screen;
 and
 an angle information display unit (13) which dis-
 plays the saved angle information on the display
 screen of the operation panel,
 wherein the regenerating unit (12) rotates the
 ultrasonic wave emitting surface to the angle
 based on the selected angle information when
 the angle information displayed on the display
 screen is selected. 20
9. The ultrasonic diagnostic apparatus according to
 claim 8, **characterized in that:**
 25
 the angle information display unit (13) displays,
 on the display screen, the angle information and
 the reduced ultrasonic cross-sectional image
 corresponding to the angle information.
10. The ultrasonic diagnostic apparatus according to
 claim 1, 2 or 3, **characterized by** further comprising:
 30
 an operation panel (15) having a display screen,
 wherein the regenerating unit (12) displays, on
 the display screen of the operation panel, a re-
 turn switch which instructs to return to the pre-
 vious angle information, and regenerates the
 previously generated particular ultrasonic
 cross-sectional image when the return switch is
 operated. 40
11. The ultrasonic diagnostic apparatus according to
 claim 1, 2 or 3, **characterized by** further comprising:
 45
 an operation panel (15) having a display screen;
 and
 a return switch (19) which is provided on a sur-
 face other than the display screen in the opera-
 tion panel and which instructs to return to the
 previous angle information,
 wherein the regenerating unit (12) regenerates
 the ultrasonic cross-sectional image corre-
 sponding to the previous angle information
 when the return switch is operated. 50
12. The ultrasonic diagnostic apparatus according to
 claim 10, **characterized in that:**
 55
 the regenerating unit (12) sequentially regener-
 ates the ultrasonic cross-sectional image corre-
 sponding to the previous angle information eve-
 ry time the return switch (19) is operated.
13. The ultrasonic diagnostic apparatus according to
 claim 11, **characterized in that:**
 60
 the regenerating unit (12) sequentially regener-
 ates the ultrasonic cross-sectional image corre-
 sponding to the previous angle information eve-
 ry time the return switch (19) is operated.
14. An ultrasonic diagnostic processing method **char-**
acterized by comprising:
 65
 rotating a surface emitting ultrasonic waves
 emitted from an ultrasonic probe (2) on an axis
 of this ultrasonic wave emitting surface (4a);
 generating a plurality of ultrasonic cross-sec-
 tional images at a plurality of angles when the
 ultrasonic wave emitting surface in the ultraso-
 nic probe is rotated;
 saving information on the angle (18) of the ul-
 trasonic wave emitting surface when a prede-
 termined condition is satisfied; and
 rotating the ultrasonic wave emitting surface
 (4a) from the angle of the ultrasonic wave em-
 itting surface different from the saved angle in-
 formation to the angle based on the saved angle
 information.
15. An ultrasonic diagnostic processing method **char-**
acterized by comprising:
 70
 mechanically rotating, with respect to an acous-
 tic lens (3a), an ultrasonic probe (2) in which a
 plurality of ultrasonic wave oscillating elements
 are one-dimensionally arranged, and rotating a
 surface emitting ultrasonic waves emitted from
 an ultrasonic probe on an axis of this ultrasonic
 wave emitting surface to generate a three-di-
 mensional ultrasonic cross-sectional image;
 saving information on the angle (18) of the ul-
 trasonic wave emitting surface when a prede-
 termined condition is satisfied; and
 rotating the ultrasonic wave emitting surface
 (4a) from the angle of the ultrasonic wave em-
 itting surface different from the saved angle in-
 formation to the angle based on the saved angle
 information.
16. An ultrasonic diagnostic processing method **char-**
acterized by comprising:
 75
 controlling the operation of an ultrasonic probe
 (2) in which a plurality of ultrasonic wave oscil-
 lating elements are two-dimensionally ar-

ranged, and rotating a surface emitting ultrasonic waves (4a) emitted from an ultrasonic probe on an axis of this ultrasonic wave emitting surface to generate a three-dimensional ultrasonic cross-sectional image;
saving information on the angle (18) of the ultrasonic wave emitting surface when a predetermined condition is satisfied; and
rotating the ultrasonic wave emitting surface (4a) from the angle of the ultrasonic wave emitting surface different from the saved angle information to the angle based on the saved angle information.

5

10

15

20

25

30

35

40

45

50

55




FIG. 1




FIG. 2




FIG. 3




FIG. 4




FIG. 5









FIG. 9





DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (IPC)
X	EP 0 514 584 A (HEWLETT PACKARD CO [US]) 25 November 1992 (1992-11-25) * abstract *	1,4-7,14	INV. A61B8/12
Y	* column 4, line 1 - column 9, line 7 * * figures 1,3,5A-B *	2,3,15, 16	
Y	----- US 4 543 960 A (HARUI NORIO [US] ET AL) 1 October 1985 (1985-10-01) * column 2, line 1 - line 35; figure 1 *	2,15	
Y	----- US 2005/277836 A1 (PROULX TIMOTHY L [US] ET AL) 15 December 2005 (2005-12-15) * abstract * * paragraph [0013] - paragraph [0019] * * figures 1,2 *	3,16	
X	----- US 5 402 793 A (GRUNER GEORGE P [US] ET AL) 4 April 1995 (1995-04-04) * abstract *	1,6,7, 11,13,14	
A	* column 1, line 43 - column 5, line 18 * * column 7, line 63 - column 8, line 57 * * figures 1,2a-b,6a,7a,11 *	2,3, 8-10,12, 15,16	
D,A	----- EP 0 984 298 A (YOKOGAWA MEDICAL SYST [JP]) 8 March 2000 (2000-03-08) * the whole document *	1-3,5-8, 11,13-16	
A	----- US 5 207 225 A (OAKS FRANK B [US] ET AL) 4 May 1993 (1993-05-04) * column 1, line 56 - column 3, line 28 * * column 6, line 1 - column 8, line 3 * * figures 1,2a-b,9-11 *	1,2,7-9, 14,15	
The present search report has been drawn up for all claims			
Place of search Munich		Date of completion of the search 8 May 2007	Examiner ARTIKIS, T
CATEGORY OF CITED DOCUMENTS X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document		T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons ----- & : member of the same patent family, corresponding document	

5
EPO FORM 1503 03.82 (P04C01)

ANNEX TO THE EUROPEAN SEARCH REPORT
ON EUROPEAN PATENT APPLICATION NO.

EP 07 00 4259

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report. The members are as contained in the European Patent Office EDP file on The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

08-05-2007

Patent document cited in search report		Publication date	Patent family member(s)	Publication date
EP 0514584	A	25-11-1992	JP 5161653 A	29-06-1993
			US 5181514 A	26-01-1993

US 4543960	A	01-10-1985	NONE	

US 2005277836	A1	15-12-2005	NONE	

US 5402793	A	04-04-1995	AT 250890 T	15-10-2003
			CA 2135692 A1	20-05-1995
			DE 69433198 D1	06-11-2003
			DE 69433198 T2	09-09-2004
			EP 0654245 A1	24-05-1995
			JP 7250836 A	03-10-1995
			JP 2005040619 A	17-02-2005
			US 5555887 A	17-09-1996
			US 5479930 A	02-01-1996

EP 0984298	A	08-03-2000	CN 1250928 A	19-04-2000
			JP 3321103 B2	03-09-2002
			JP 2000079121 A	21-03-2000
			KR 20000022899 A	25-04-2000
			US 6217515 B1	17-04-2001

US 5207225	A	04-05-1993	NONE	

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- JP 2000079121 A [0008]

专利名称(译)	超声波诊断装置和超声波诊断处理方法		
公开(公告)号	EP1829483A1	公开(公告)日	2007-09-05
申请号	EP2007004259	申请日	2007-03-01
[标]申请(专利权)人(译)	株式会社东芝 东芝医疗系统株式会社		
申请(专利权)人(译)	株式会社东芝 东芝医疗系统公司		
当前申请(专利权)人(译)	东芝医疗系统公司 株式会社东芝		
[标]发明人	OONUKEI YUTAKA SHIKATA HIROYUKI KOSAKU HIDEKI OGAWA TAKASHI AOKI MINORU		
发明人	OONUKEI, YUTAKA SHIKATA, HIROYUKI KOSAKU, HIDEKI OGAWA, TAKASHI AOKI, MINORU		
IPC分类号	A61B8/12		
CPC分类号	A61B8/4461 A61B8/12 A61B8/4483 A61B8/4488 A61B8/463 A61B8/465 A61B8/483 G01S7/52074 G01S15/894		
代理机构(译)	KRAMER - HARSH - 施密特陈		
优先权	2006056464 2006-03-02 JP		
其他公开文献	EP1829483B1 EP1829483B8		
外部链接	Espacenet		

摘要(译)

当超声波探头 (2) 中的超声波发射表面旋转时, 产生单元 (11) 以多个角度产生多个超声波横截面图像。当满足预定条件时, 再生单元 (12) 保存关于超声波发射表面的角度的信息, 并且将超声波发射表面从与保存的角度信息不同的超声波发射表面的角度旋转到基于角度的角度。在保存的角度信息上。




FIG. 3